Controlling Dental Enamel–Cavity Ablation Depth with Optimized Stepping Parameters Along the Focal Plane Normal Using a Three Axis, Numerically Controlled Picosecond Laser

Fusong Yuan, PhD,¹ Peijun Lv, MD, PhD,¹ Dangxiao Wang, PhD,²
Lei Wang, MSc,² Yuchun Sun, MD, PhD,^{1,*} and Yong Wang, PhD^{1,*}

Abstract

Objective: The purpose of this study was to establish a depth-control method in enamel-cavity ablation by optimizing the timing of the focal-plane-normal stepping and the single-step size of a three axis, numerically controlled picosecond laser. **Background data:** Although it has been proposed that picosecond lasers may be used to ablate dental hard tissue, the viability of such a depth-control method in enamel-cavity ablation remains uncertain. **Methods:** Forty-two enamel slices with approximately level surfaces were prepared and subjected to two-dimensional ablation by a picosecond laser. The additive-pulse layer, n , was set to 5, 10, 15, 20, 25, 30, 35, 40, 45, 50, 55, 60, 65, 70. A three-dimensional microscope was then used to measure the ablation depth, d , to obtain a quantitative function relating n and d . Six enamel slices were then subjected to three dimensional ablation to produce 10 cavities, respectively, with additive-pulse layer and single-step size set to corresponding values. The difference between the theoretical and measured values was calculated for both the cavity depth and the ablation depth of a single step. These were used to determine minimum-difference values for both the additive-pulse layer (n) and single-step size (d). **Results:** When the additive-pulse layer and the single-step size were set 5 and 45, respectively, the depth error had a minimum of $2.25\ \mu\text{m}$, and $450\ \mu\text{m}$ deep enamel cavities were produced. **Conclusions:** When performing three-dimensional ablating of enamel with a picosecond laser, adjusting the timing of the focal-plane-normal stepping and the single-step size allows for the control of ablation-depth error to the order of micrometers.

Introduction

LASER DENTISTRY IS ATTRACTING INCREASING ATTENTION as an alternative to conventional dental-drilling procedures, primarily in terms of precision and reduced procedural time. However, millisecond, microsecond, and nanosecond lasers have limitations when used in the ablation of hard dental tissues. For example, thermal ablation can produce lateral collateral damage, whereas the ablation range of WaterLase ablation is difficult to control.^{1,2} In recent years, researchers studying shape preparation for the ablation of human hard dental tissue have begun using ultrashort pulse lasers, such as femtosecond and picosecond lasers.³ Because of the refined degree of control of ablation accuracy in these types of lasers, it may be possible to apply multi-axis,

numerically controlled galvanometer technology to the automated ablation process. Such technology may help in overcoming issues such as difficulty in observing ablation position by visual inspection (in real time) caused by the inherent instability of handheld laser processes.

In terms of laser dentistry, we note that human enamel is an uneven material. Therefore, when using ultrashort-pulse laser ablation, there can be problems in controlling ablation accuracy. The output form of the laser energy is a light spot, which means that point ablation occurs. Therefore, the majority of ablation error is reflected in the ablation depth.⁴ Consequently, controlling the depth of ablation is of primary importance.

In this light, we examine here the quantitative relationship between the timing of the focal-plane-normal stepping of a

¹Center of Digital Dentistry, Faculty of Prosthodontics, Peking University School and Hospital of Stomatology and National Engineering Laboratory for Digital and Material Technology of Stomatology and Research Center of Engineering and Technology for Digital Dentistry of Ministry of Health, Beijing, China.

²Robotics Institute, Beihang University, Beijing, China.

*These two authors contributed equally.

three axis, numerically controlled picosecond laser and the single-step size of the depth error for human enamel cavity preparation. Further, we attempt to establish a depth-control method in enamel cavity preparation via focal-plane-normal step-parameter optimization of the laser. Enamel is mainly removed in tooth preparation; therefore, the evaluation of precise depth-control methods for ultra-short-pulse laser enamel ablation can effectively avoid damaging the surrounding hard and soft tissues while achieving highly accurate three-dimensional laser grinding. Such an approach may provide a strong foundation for future developments in oral clinical precise and automatic tooth preparing.

Materials and Methods

Equipment and methods

Equipment and software. Laser system: picosecond laser (wavelength, 1064 nm; pulse width, 15 psec; repetition frequency, 100 kHz; power, 30 W; laser focal-light spot diameter, 38 μm).

We used a three axis laser scanning device with a mini-numerical-control galvanometer that was developed and constructed in-house (lens focal length, 175 mm; light-spot scanning speed for the two-dimensional scanning galvanometer, 1900 mm/sec; light-spot overlapping rate within and between the scanning lines, 50%; minimum step size along the z axis, 0.1 mm; maximum step size, 10 mm). We also programmed the numerical control software used to operate the laser system.

The sample-stage position-control device was a spiral micrometer (Mitutoyo Corporation, Japan; accuracy, 0.01 mm).

The three-dimensional measuring device was a three-dimensional laser scanning microscope (Keyence Corporation, Japan, VK- X200).

Materials. A total of 24 freshly removed mandibular first molars were collected from the Oral and Maxillofacial Surgery Clinic at the Peking University Hospital of Stomatology.

Methods

Sample preparation. Calculus and soft tissues on the dental surfaces were removed using an ultrasonic scaler and the surfaces were rinsed clean with saline. The crowns and the roots were cut transversely along the cemento-enamel separation using a diamond wire cutter (Shenyang Kejing Instrument Co., Ltd, STX-202). The diamond wire was

moved parallel to the cross-section towards the crown, producing dental hard tissue sample slices of 2 mm thickness. These slices consisted of a round inner portion of dentin surrounded by an external enamel layer of at least 2 mm width (Fig. 1). The surfaces for ablation were polished using 800 and 1000 grit waterproof sandpaper to form essentially flat surfaces; they were then submerged in formalin solution until use. The samples were divided into two groups: 42 sample slices from the first group were used in ablation with a fixed focal plane position and 6 sample slices from the second group were used in stepwise ablation along the focal plane normal.

Ablation with fixed focal plane position. The sample was mounted on an electrical spiral micrometer sample stage; the ablation surface was oriented perpendicular to the picosecond laser beam and positioned at the focal point in front of the center of the lens using two vertically placed spiral micrometers (Fig. 2). The picosecond-laser path was modeled computationally using software (Fig. 3). The frame for the laser was a square with an outer-side length of 5 mm and inner-side length of 3 mm, and the ablation section covered both the dentin and enamel regions. The space between the frames was set to 19 μm . The light-spot diameter was 38 μm , the surface scanning speed was 1900 mm/sec within controllable scanning lines, and the light-spot overlapping rate was 50%. The control software automatically calculated the number of frames to be scanned, and it also performed two-dimensional scanning and ablation of the laser spot from the outside to the inside, starting from the focal plane. The number of additive pulses was n (one additive-pulse layer denotes the completion of one ablation motion from the outer to the inner frame by the laser, and, therefore, n represents the layer number). The 14 enamel-dentin slice surfaces were scanned for $n = 10, 15, 20, 25, 30, 35, 40, 45, 50, 55, 60, 65$, and 70 layers in the given order. The ablation of each additive pulse layer value was repeated three times, yielding a total of 42 samples.

Measurement of two-dimensional scanning-ablation depth. A three-dimensional shape-measurement laser microscope was used to measure the actual ablation depth, d , for each slice (actual two-dimensional scan-ablation depth with fixed focal plane position). In total, 40 points were selected at the base of the ablation of each enamel section (the points were distributed as evenly as possible on the cleaved surface). With the upper surface as a reference, the

FIG. 1. Dental hard tissue sample slice.

FIG. 2. Experiment platform and spot diameter measurement.

depths of the 40 points were measured to calculate their average. Three repeated ablations were made for each n (additive-pulse layer) value to calculate an average. Hence, actual ablation depths, corresponding to each n value, were recorded. A maximum ablation depth value for the two-dimensional ablation was thus obtained, along with a plot of the quantitative relationship between ablation depth and additive-pulse layer.

Three-dimensional scanning and ablation in stepwise increments along focal plane normal. The parameters for this process, such as the focal-plane two-dimensional scanning path, were set as described previously. The timing for the stepwise increment along the normal was set to 5, 10, 15, 20, 25, 30, 35, 40, 45, and 50 layers. The single-step size was obtained from the linear regression function of the quantitative relationship between ablation depth and additive pulse layer. The number of steps was set to 9, 4, 3, 2, 2, 1, 1, 1, 1, and 1. Ten square cavities (2×2 mm) were cut into the enamel regions in each of the enamel–dentin mixed slices. This was repeated for six sample slices.

Measurement of three-dimensional ablation depth. A three-dimensional shape-measurement laser microscope was used to measure the actual total ablation depth, d_2 , of each square cavity using the method described previously. The theoretical total ablation depth $d_1 = d_0 \times (t + 1)$, where d_0 denotes the theoretical single-step ablation depth and t denotes the number of steps. The error in the total depth, e , was also calculated. The error in the focal-plane single-step ablation depth was then calculated as $e_1 = (d_2 / t + 1) - d_0$. The relationship among e , e_1 , n , and d was subsequently analyzed.

Results

Two-dimensional ablation

Table 1 lists the n and d values corresponding to each laser ablation in the enamel region. The maximum d value was $674.24 \mu\text{m}$, with $n = 60$.

The quantitative relationship between n and d showed an increasing trend within the range of the maximum ablation depth. Through linear fitting, the function equation and fitting coefficient were computed as $d = 10.16n - 6.04$ and

FIG. 3. Two-dimensional motion paths of picosecond laser spot.

TABLE 1. CORRELATION BETWEEN THE NUMBER OF TWO-DIMENSIONAL SCANNING LAYERS, N , AND THE ABLATION DEPTH, d , IN ENAMEL

n (layer)	d (μm)
5	48.21 ± 0.041
10	106.72 ± 0.052
15	158.26 ± 0.048
20	212.84 ± 0.046
25	234.56 ± 0.053
30	293.43 ± 0.054
35	340.28 ± 0.049
40	387.54 ± 0.047
45	410.91 ± 0.053
50	500.12 ± 0.043
55	524.42 ± 0.046
60	674.24 ± 0.052
65	600.25 ± 0.051
70	584.26 ± 0.055

$R^2=0.9778$. Figure 4 shows the qualitative relationship between n and d , as well as a linear-fitting plot.

Three-dimensional ablation

A three-dimensional shape measurement laser microscope was used to measure d_2 . The values of d_1 , e , d_0 , and e_1 were calculated, and these values are listed in Table 2.

From the table, we note that for $n=5$, the error in the total depth reached a minimum value of $2.25 \mu\text{m}$. Therefore, the depth error was controllable on the micrometer scale.

Discussion

In this section, we attempt to relate our findings to other studies in this area. Sun et al.⁴⁻⁸ used a Ti:sapphire femtosecond laser system (CPA-2001, Clark-MXR, USA) to conduct a preliminary study on the surface appearance and changes in dental tissue morphology as well as the control of the medullary cavity temperature and ablation efficiency with dental hard tissue. Their results showed that numerically controlled femtosecond laser ablation yielded a smooth

FIG. 4. Linear fitting plot and quantitative relationship between the number of focal plane additive-pulse layers, n , and ablation depth, d .

TABLE 2. PARAMETER SETTINGS AND RESULTS OF THREE-DIMENSIONAL ENAMEL ABLATION

n (layer)	d_0 (μm)	t (steps)	d_1 (μm)	d_2 (μm)	e (μm)	e_1 (μm)
5	45	10	450	452.25 ± 0.041	2.25	0.04
10	96	5	480	501.43 ± 0.039	21.43	0.43
15	146	4	584	619.12 ± 0.037	35.12	0.62
20	197	3	591	642.34 ± 0.032	51.34	0.86
25	248	3	744	819.24 ± 0.035	75.24	1.00
30	299	2	598	698.51 ± 0.029	100.51	1.67
35	350	2	700	841.62 ± 0.038	141.62	2.02
40	401	2	802	982.91 ± 0.026	180.91	2.26
45	451	2	902	1115.56 ± 0.040	213.56	2.37
50	502	2	1004	1254.64 ± 0.036	250.64	2.51

surface and boundaries with clear lines and corners in hard dental tissue, and that the increase in medullary cavity temperature could be controlled to within 5.5°C , using compressed-air cooling at room temperature. In addition, it was found that the ablation efficiency for dentin was significantly higher than that for enamel. However, the purchase and operation of femtosecond laser equipment is prohibitive. In this light, the application of another type of ultrashort pulse laser, the picosecond laser, has improved the prospects for the practical application of dental ablation. Current research in China and elsewhere mainly focuses on the micromorphology, thermal effects, and ablation efficiency of picosecond-laser hard-dental-tissue ablation.^{9–12} On the other hand, there have been fewer studies on ablation-depth control methods with picosecond systems.

Studies by Lizarelli et al.^{13,14} showed that, at equal energy densities, picosecond laser ablation efficiency on dentin was approximately eight times that on enamel. This is primarily because dentin contains a large amount of collagen compared with enamel, making it easier to ablate. Sun et al.⁸ conducted a preliminary study on the depth-error control method of a numerically controlled picosecond laser in human dentin ablation. They found that when scan-ablating dentin using a 30 W three-axis picosecond laser at a wavelength of 1064 nm and a repetition frequency of 100 kHz, the focal plane additive-pulse layer, the errors of the step increment along the normal and single-layer ablation depth, and the total ablation depth error showed positive correlation. By adjusting the single-step size and the step timing along the focal plane normal when ablating dentin using a numerically controlled picosecond laser, the single-step ablation depth error was controllable at the micrometer level. Based on this study, our research group further investigated the dentin-cavity-ablation depth-control method and evaluated its effect based on three-axis, numerically controlled picosecond laser focal-plane-normal step-parameter optimization. The experimental results showed that when the timing of the stepwise increment and single-step size along the focal plane normal were set at five additive pulse layers and 45 μm , respectively, the single-step ablation depth error was minimized (0.05 μm); then, when a square cavity was ablated with a depth of 420 μm , the depth error was 2.75 μm .

Next, we discuss the need for superimposing the focal and ablation planes. Laser energy exhibits a non-uniform Gaussian distribution. When a dentin surface moves out of the focal plane, the energy density at the periphery of the light spot decreases; whereas at the center, this energy may still be

above the dentin-ablation threshold. With existing methods, the single-spot ablation surface may take on the appearance of volcanic craters,¹⁵ and the successive superposition of many light spots will lead to an increased depth error in the ablation surface, which is unfavorable in regard to ablation accuracy control. Therefore, in theory, the ablation surface of interest should be kept superimposed onto the focal plane (along its normal) during ablation in order to ensure that each dentin surface element targeted by light spots can reach the same ablation depth (or the same depth with minimal difference), thus reducing the single-step ablation-depth error. The results of this study showed that by reducing the additive-pulse layers of the two-dimensional scan with a fixed focal plane and focal-plane-normal single-step size, the single-step dentin-ablation depth error was controllable on a micrometer scale.

Finally, we discuss the rationale underlying our choice of light-spot diameter and overlapping rate. The picosecond laser is a pulse laser of which each pulse can be considered a light spot. Previous experiments have demonstrated that a 50% light spot overlapping rate can produce an ablation surface in dental hard tissue that is smoother and flatter than those yielded by other laser treatments.⁶ During linear scanning, the light-spot diameter, scanning diameter, and scanning speed all influence the light-spot overlapping rate; whereas during surface scanning, the distance between the scanning lines determines the inter-line light-spot overlapping rate. In this study, we set the light-spot diameter to 38 μm , the linear scanning speed to 1900 mm/sec, the repetition frequency to 100 kHz, and the light-spot overlapping rate with lines at $(1900 \times 1000 \mu\text{m} / 100,000 \text{ pulses}) / 38 \mu\text{m} = 50\%$. The distance between the scanning lines was set to 19 μm , and the light-spot overlapping inter-line rate was set to $19 \mu\text{m} / 38 \mu\text{m} = 50\%$. Using the aforementioned settings, we obtained smooth and flat ablation surfaces.

Although we obtained such smooth, flat ablation surfaces for dentin in our study, we note that during actual clinical dental preparation, the hard dental tissues to be removed contained both enamel and dentin. Meanwhile, the composition of enamel and dentin, which includes the density of dentinal tubules, as well as the amount of water and organic material, are different from one tooth to another, and also from the different parts of the same tooth. It may have great influence on the precise ablation. The choice of whether to adjust the laser output parameters according to the ablation object under study or to use a single set of strongly compatible parameters may determine the means for highly accurate ablation on the two types of dental hard tissues and

may produce a preparatory shape that satisfies clinical requirements; this constitutes a topic for further research.

Conclusions

Our experimental results demonstrate that when ablating enamel two dimensionally, using a numerically controlled picosecond laser, the number of focal-plane additive pulse layers and the ablation depth exhibit an overall positive linear correlation. In three-dimensional ablation, when the single-step size along the focal plane normal is set to be equal with the two-dimensional scanning-ablation depth, the number of focal plane additive pulse layers and the single-step size along the normal are positively correlated with the single-step ablation-depth error and the total ablation-depth error.

By adjusting the timing of stepwise increments along the focal plane normal and the single-step size, the depth error of ablating human enamel via a numerically controlled picosecond laser can be limited to within $0.04\text{ }\mu\text{m}$ per step. When the timing of the stepwise increment and single-step size along the focal plane normal were set to five additive pulse layers and $45\text{ }\mu\text{m}$, respectively, the ablation of a square cavity with a depth of $450\text{ }\mu\text{m}$ had an enamel-ablation depth error limited to within $2.25\text{ }\mu\text{m}$. In light of these results, we believe that this work may significantly contribute to the development of laser dentistry.

Acknowledgment

This study was supported by the National Science & Technology Pillar Program during the 12th Five-Year Plan (Grant no. 2012BAI07B04), the National High Technology Research and Development Program ("863" Program) of China (Grant no. 2013AA040802), and the National Natural Science Foundation of China (Grant no. 81271181).

Author Disclosure Statement

No competing financial interests exist.

References

- Sharma N, Williams C, Angadi P, Jethlia H. Application of lasers in dentistry: a review. *Res Rev J Dent Sci* 2013;1:22–25.
- Lizarelli RFZ, Kurachi C, Misoguti L, Bagnato VS. A comparative study of nanosecond and picosecond laser ablation in enamel: morphological aspects. *J Clin Laser Med Surg* 2000;18:151–157.
- Braun A, Krillke RF, Frentzen M, Bourauel C, Stark H, Schelle F. Heat generation caused by ablation of dental hard tissues with an ultrashort pulse laser (USPL) system. *Lasers Med Sci* 2013 [Epub ahead of print].
- Sun YC, Yuan FS, Lv PJ, Wang DX, Wang L, Wang Y. Method to control depth error when ablating human dentin with numerically-controlled picosecond laser: a preliminary study. *Lasers Med Sci* 2014;29:1–7.
- Sun YC, Vorobyev A, Wang Y, Li H. Influence of intrapulpal temperature when using femtosecond laser in specific parameters to prepare cavities in tooth enamel: an in vitro study. *Beijing Da Xue Xue Bao* 2013;45:286–290.

- Sun YC, Vorobyev A, Lv PJ, Liu J. Femtosecond pulsed laser ablation of dental hard tissues with numerical control: a roughness and morphology study. *Zhonghua Kou Qiang Yi Xue Za Zhi* 2012;47:486–489.
- Sun YC, Vorobyev A, Lv PJ, Chen H. Ablation of enamel and dentin with computerized numerical controlled femtosecond pulsed laser: a primary study on the ablation rate. *Zhonghua Kou Qiang Yi Xue Za Zhi* 2013;48:58–59.
- Sun YC, Chen H, Vorobyev AY. High-intensity femtosecond laser ablation of human enamel and dentin. *J Med Imaging Health Information* 2014;4:422–426.
- Lizarelli RFZ, Moriyama LT, Bagnato VS. Temperature response in the pulpal chamber of primary human teeth exposed to Nd:YAG laser using a picosecond pulsed regime. *Photomed Laser Surg* 2006;5:610–615.
- Lorenzo MC, Portillo M, Moreno P, et al. In vitro analysis of femtosecond laser as an alternative to acid etching for achieving suitable bond strength of brackets to human enamel. *Lasers Med Sci* 2014;29:897–905.
- Luengo MCL, Portillo M, Sánchez JM. Evaluation of micromorphological changes in tooth enamel after mechanical and ultrafast laser preparation of surface cavities. *Lasers Med Sci* 2013;28:267–273.
- Strassl M, Kopecek H, Weinrotter M. Novel applications of short and ultra-short pulses. *Appl Surf Sci* 2005;247:561–570.
- Lizarelli RFZ, Kurachi C, Misoguti L. Characterization of enamel and dentin response to Nd:YAG picosecond laser ablation. *J Clin Laser Med Surg* 1999;17:127–131.
- Lizarelli RFZ, Moriyama LT, Bagnato VS. Ablation rate and micromorphological aspects with Nd:YAG picosecond pulsed laser on primary teeth. *Lasers Surg Med* 2002;31:177–185.
- Neev J, Da Silva LB, Feit MD. Ultrashort pulse lasers for hard tissue ablation. *IEEE J Sel Top Quant Electron* 1996;2:790–800.

Address correspondence to:

Yong Wang

or

Yuchun Sun

Center of Digital Dentistry

Faculty of Prosthodontics

Peking University School and Hospital of Stomatology

and

National Engineering Laboratory for Digital

and Material Technology of Stomatology

and

Research Center of Engineering and Technology

for Digital Dentistry of Ministry of Health

22 Zhongguancun Nandajie

Haidian District, Beijing 100081

China

E-mail: kqcadc@bjmu.edu.cn

or

E-mail: polarshining@163.com