Published in final edited form as:

J Prosthet Orthot. 1999 ; 11(1): 15–19.

The Effect of Ankle-Foot Orthoses on Balance Impairment: Single-Case Study

Noel Rao, MD and

Chair of physical medicine & rehabilitation OPD at the Marianjoy Rehabilitation Hospital and Clinics in Wheaton, Illinois

Alexander Aruin, PhD

Senior research scientist at the Rehabilitation Foundation, Inc., 26W171 Roosevelt Rd., P.O. Box 675, Wheaton, Illinois 60189. Phone: (630) 462-4277; FAX: (630) 4624547

Alexander Aruin: aruin@rfi.org

Abstract

Instability and balance impairment is common in patients with diabetic sensory neuropathy placing them at a higher risk of falling when performing more challenging daily tasks. The report describes the results of a dynamic balance tests of a subject with neuropathy due to the long-standing diabetes. The Computerized Dynamic Posturography was performed with and without ankle-foot orthoses (AFOs). The apparatus provided six test conditions designed to systematically manipulate vestibular, somatosensory, or visual information. With no orthoses the patient had falls performing most of the tests. Bilateral orthoses improved his balance: a derived composite balance score increased four times. In view of these findings, AFOs, in addition to correcting the patient's foot placement during locomotion, could also be expected to improve the maintenance of balance during quiet stance or dynamic perturbation.

Keywords

Diabetic neuropathy; balance; ankle-foot orthosis

The remarkable ability of the body to maintain balance is based on the efficient detection and integration of information from the vestibular, visual, and somatosensory systems. Dysfunction of any of the three sensory systems (which may happen when information from one or more of the perceptual systems is in conflict with the information from the other perceptual systems) results in spatial disorientation, varying degrees of unsteadiness, or imbalance. When either the visual field or support surface is moving, the somatosensory and visual systems dominate the control of balance. This happens because the vestibular system is less sensitive to slow and quick changes of environment than the visual or somatosensory systems. The vestibular system acts as an internal reference for inaccuracies that may result from operating both the somatosensory and visual systems. The challenge of equilibrium maintenance increases considerably and depends more on vestibular input when both visual and somatosensory information are compromised. The most serious consequences of errors

in the vestibular, visual, and proprioceptive perception of the environment are faulty or delayed corrective responses leading to falls. In particular, sensory deprivation studies have shown that deficit of the proprioceptive perception (i.e., seen in patients with peripheral neuropathy and associated with a loss of distal sensation) is associated with increased risk of falls.¹⁻³

One of the major complications associated with diabetic neuropathy is bilateral loss of somatosensory information in the hands and feet. Such a somatosensory deficit in the feet compromises functional postural stability of patients with diabetic neuropathy and might place them at a higher risk of falling when performing more challenging daily tasks.¹⁻⁶ It was also shown that diabetic subjects with peripheral neuropathy have demonstrated a significant loss of ankle movement perception,⁷ have larger ranges of postural sway,^{4,8} and more likely to use hip control balance strategy. Because of the percentage of individuals diagnosed with diabetes who develop polyneuropathy after 20 years reaches 50%,¹⁰ the number of balance studies involving diabetic patients is growing.^{3,5,6}

AFOs are prescribed to control and limit movement at the ankle and knee, thus improving gait abnormalities in patients with hemiplegia, and peroneal and tibial nerve paralysis. AFOs are also prescribed to have an impact on speed and energy cost of hemiparetic ambulation.¹¹ The importance of orthotic use is well documented.¹²⁻¹⁶ In particular, it was demonstrated in a single-subject design study that standing balance of a 4.5-year-old boy with cerebral palsy improved while using an AFO.¹⁷ In particular, improvements were noted in duration of independent standing during with-orthoses condition, in the symmetry of the stance pattern, and in the ease with which the subject maintained independent standing. The effect of wearing ankle orthoses has also been shown in healthy volunteers tested while standing upright on a statokinesimetric platform with and without orthoses.¹⁸ The study provided information about the prophylactic effect of bilateral orthoses for subjects with major variations in postural equilibrium. However, objective functional measurement of dynamic balance of diabetic patients with and without AFOs has not been reported.

The purpose of the study is to assess the effect of AFOs on balance impairment using computerized dynamic posturography. Dynamic posturography provides information on the use of sensory cues to maintain postural stability and on the automatic postural responses to maintain stability after translation or rotation of the support surface.¹⁹

Methodology

The subject was a 48-year-old male with a 23 year history of insulin-dependent diabetes mellitus. The subject had peroneal nerve palsy and impaired sensations of both hands and feet. His light touch was impaired below the ankles and below the wrists bilaterally. He also was without proprioception and had diminished vibratory sense in these areas. Motor tone was normal in all extremities, and reflexes were normal, with the exception of absent ankle jerks and diminished knee jerks. Romberg test was positive and the patient was unable to perform tandem gait. He was unable to stand on either the right or left foot for more than one second without losing his balance. Gait without AFOs revealed that the patient during initial stance and mid-stance exhibited bilateral equino varus ankle-foot deformity followed

by excessive pronation during terminal stance. He also exhibited hyperextension of both knees during mid- and terminal-stance phases. Steppage gait was observed during swing phase.

Four days prior to the study, the patient was provided with a bilateral rigid polypropylene AFOs (thickness of plastic was 3/16 in) with the trim line medially and laterally at the apex of the malleoli, proximally 1 in below fibular head and distally at the toe sulcus and the ankle set at 2° of dorsiflexion (Figure 1). The orthoses were beneficial in correcting the patient's gait deviations, providing limitation of plantar flexion and dorsiflexion by its biomechanical configuration. Gait evaluation with orthoses revealed that the ankle and knee control improved, the patient no longer exhibited bilateral equino varus at the ankle. He had heel-to-toe gait without hyperextension at the knees, and step-page gait was not observed. The patient had no complaints of dizziness or balance difficulty and was independent in ambulation with the orthoses without any assistive devices.

In the dynamic posturography, the subject stands on a dual forceplate enclosed by visual surround. Both the forceplate and the surround can be made to move with the subject's anteroposterior sway or independent of the sway, thus enabling programmed disturbances of the equilibrium. The dual force-plate records the vertical forces between feet and ground as well as shear forces, therefore allowing estimation of the position of the swaying body and the pattern of sway in terms of hip or ankle strategy.

Dynamic posturography was performed with and without AFOs. The subject was placed on the dual force-plate and secured with a harness to prevent a fall. The Sensory Organization Test (SOT) was performed in a clinically routine manner. The SOT included six tests conditions (Figure 2). The first three involved the patient standing on a fixed platform with eyes open (SOT 1), eyes closed (SOT 2), and using sway-referenced vision (SOT 3). Changes in platform sway referencing (SOT 4,5,6) introduced changes in somatosensory input while alterations in vision were similar to previous three series.

Posturography scores (that range from 0 to 100, with 0 representing a fall (protected by harness), and 100 representing perfect stability) were based on the database established by the initial users' group. In all trials, the position of the subject on the dual forceplate was the same: the medial malleolus of each foot were centered directly over the stripe on the dual forceplate in such a way that the distance between two feet (measured as the distance between the midlines of the two heels) was 0.15 m (or 8.5% of standing height of the subject) which corresponds to the recommended foot placement for balance testing.²⁰ The experimenter checked the positioning of the subject during all parts of the experiment to make sure that feet position and the sway-referenced stimuli were equivalent for both orthotic and nonorthotic trials.

Results

Figure 3 represents equilibrium scores of the patient measured with the computerized dynamic posturography technique. The scores reflect how much the patient swayed during each trial of the six sensory conditions. The equilibrium scores were calculated by

comparing the patient's anteroposterior sway during each trial to a theoretical sway stability limit of 12.5°. In the first test with no orthoses, the subject demonstrated scores over 75 in three trials. However, only 19% of the trials were successful (he fell in 13 of 16 trials), and as a result, the composite score (calculated by averaging the scores for all conditions) reached a value of only 14 (Figure 3a). Because of the harness, there was no danger of falling for the patient even though equilibrium was lost. In contrast, during the second test with bilateral orthoses the patient was more stable: in 73% of trials (11 of 15), the patient showed normal posturography patterns (Figure 3b) resulting in improvement of integral composite score nearly 4 times higher while using orthoses. The composite score for the second test reached the magnitude of 59, which is still below the magnitude established for healthy subjects, but above the magnitude which is considered to be abnormal^a.

Discussion

The results of the study show that the overall balance test performance of the patient with severely reduced foot sensation could be improved by using AFOs. Both experimental series were performed under the same conditions, except that the subject was wearing AFOs in the second series. The patient had falls performing most of the sensory organization test conditions without orthoses and improved his performance while wearing orthoses.

It appears reasonable to believe that the patient with loss of afferent function could rely on more proximal cues provided by the orthoses. Indeed, it is widely known that stability of posture increased while an ambulatory aid such as a cane, walker, etc. is used.²¹ The ambulatory devices, in addition to providing an extended base of support,²² may furnish somatosensory information to the proximal parts of the body that are still functioning normally.²³

It has been shown that somatosensory stimulation from contact of the feet with the support surface plays an important role in maintaining upright stance.²⁴ Also, touch and pressure cues from any part of the body in contact with a stable external surface positively influence apparent body orientation.²⁵ Even a very slight finger contact with a stable surface attenuates body sway in blind individuals.²⁶

In view of these findings, AFOs, in addition to correcting the patient's foot placement during locomotion, could also be expected to improve the maintenance of balance during quiet stance or dynamic perturbation. It is quite possible that mechanical connection of the intact parts of the body to the surface area could help in getting somatosensory feedback necessary for balance control. This is consistent with the finding that peripheral neuropathy patients improved their unipedal proprioception while using a firm lateral knee pad during a stance test.²⁷

Conclusion

This study provides information about the prophylactic effect of wearing AFOs by a patient with severely reduced foot sensation due to diabetic neuropathy. The orthoses were

^aNeuroCom International, Inc., Lawnfield Rd, Clackamas, OR 97015.

beneficial not only in correcting the patient's gait and providing independence in ambulation, they also changed for better his balance performance: improvement of the overall balance score performance measured with the Computerized Dynamic Posturography test was seen with bilateral orthoses as compared with results of the test with no orthoses.

This single-subject research design supports the efficiency of AFOs in improving standing balance for a patient with diabetic neuropathy but does not provide information for making any statistical statement. Our findings, nevertheless, encourage us to pursue future research to statistically document the effectiveness of using orthoses on balance improvement.

Acknowledgments

The study was in part supported by grant HD-37141 from the National Center for Medical Rehabilitation Research, NIH.

References

1. Richardson JK, Ching C, Hurvitz EA. The relationship between electromyography documented peripheral neuropathy and falls. *J Am Geriatr Soc.* 1992; 40:1008–1012.
2. Richardson JK, Ashton-Miller JA. Peripheral neuropathy: an often-overlooked cause of falls in the elderly. *Postgrad Med.* 1996; 99(6):161–172. [PubMed: 8668629]
3. Simoneau GG, Utbrecht JS, Derr JT, Becker MB, Cavanagh PR. Postural instability in patients with diabetic sensory neuropathy. *Diabetes Care.* 1994; 17:1411–1421. [PubMed: 7882810]
4. Boucher P, Teasdale N, Courtemance R, Bard C, Fleury M. Postural stability in diabetic polyneuropathy. *Diabetes Care.* 1995; 18:638–45. [PubMed: 8586001]
5. Uccioli L, Giacomini PG, Monticone G, Margini A, Durola L, Bruno E, et al. Body sway in diabetic neuropathy. *Diabetes Care.* 1995; 18(3):339–344. [PubMed: 7555477]
6. Simmons RW, Richardson C, Rozos R. Postural stability of diabetic patients with and without cutaneous sensory deficit in the foot. *Diabetes Research and Clinical Practice.* 1997; 36:153–160. [PubMed: 9237781]
7. Simoneau GG, Derr JA, Ulbrecht JS, Becker MB, Cavanagh PR. Diabetic sensory neuropathy effect on ankle joint movement perception. *Archives of Physical Medicine & Rehabilitation.* 1996; 77:453–460. [PubMed: 8629921]
8. Lord SR, Caplan GA, Colagiuri R, Ward JA. Sensory-motor function in older persons with diabetes. *Diabet Med.* 1993; 10:614–618. [PubMed: 8403821]
9. Giacomini PG, Bruno E, Monticone G, Di Girolamo S, Margini A, Parisi L, et al. Postural rearrangement in IDDM patients with peripheral neuropathy. *Diabetes Care.* 1996; 19(4):372–374. [PubMed: 8729163]
10. Birke JA, Sims DS. Plantar sensory threshold in the ulcerative foot. *Lepr Rev.* 1986:261–267. [PubMed: 3784758]
11. Corcoran PH, Jepsen RH, Brengelmann GL, Simons BC. Effects of plastic and metal leg braces on speed and energy cost of hemiparetic ambulation. *Arch Phys Med Rehabil.* 1970; 51:69–77. [PubMed: 5437126]
12. Lehmann JF, Condon SM, de Lateur BJ, Price R. Gait abnormalities in peroneal nerve paralysis and their corrections by orthoses: A biomechanical study. *Arch Phys Med Rehabil.* 1986; 67:380–386. [PubMed: 3718197]
13. Lehmann JF, Condon SM, Price R. Gait abnormalities in hemiplegia: Their correction by ankle-foot orthoses. *Arch Phys Med Rehabil.* 1987; 68:763–771. [PubMed: 3675173]
14. Hesse S, Luecke D, Jahnke MT, Mauritz KH. Gait function in spastic hemiparetic patients walking barefoot, with firm shoes, and with ankle-foot orthosis. *Int J Rehabil Res.* 1996; 19(2):133–141. [PubMed: 8842827]

15. Carlson WE, Vaughan CL, Damiano DL, Abel MF. Orthotic management of gait in spastic diplegia. *Am J Phys Med Rehabil.* 1997; 76:219–225. [PubMed: 9207708]
16. Dieli J, Ayyappa E, Hornbeak S. Effect of dynamic AFOs on three hemiplegic adults. *Journal of Prosthetics and Orthotics.* 1997; 9:82–89.
17. Harris SR, Rifflee K. Effect of inhibitive Ankle-Foot Orthoses on standing balance in a child with cerebral palsy. *Physical Therapy.* 1986; 66:663–667. [PubMed: 3703929]
18. Calmels P, Escafit M, Domenach M, Minaire P. Posturographic evaluation of the proprioceptive effect of ankle orthoses in healthy volunteers. *Int Disability Study.* 1991; 13:42–45.
19. Nashner LM, Peters JF. Dynamic Posturography in the diagnosis and management of dizziness and balance disorders. *Neurot Clinic.* 1990; 8:331–347.
20. McIlroy WE, Maki BE. Preferred placement of the feet during quiet stance: development of a standardized foot placement for balance testing. *Clin Biomech.* 1997; 12:66–70.
21. Deathe AB, Hayes KC, Winter DA. The biomechanics of canes, crutches, and walkers. *Crit Rev Phys Rehabil Med.* 1993; 5:15–29.
22. Milszarek JJ, Kirby RL, Harrison ER, MacLeod DA. Standard and four-footed canes: their effect on the standing balance of patients with hemiparesis. *Arch Phys Med Rehabil.* 1993; 74:281–284. [PubMed: 8439256]
23. Feuerbach JW, Grabiner MD, Koh TJ, Weiker GG. Effect of an ankle orthosis and ankle ligament anesthesia on ankle joint proprioception. *Am J Sports Med.* 1994; 22(2):223–229. [PubMed: 8198191]
24. Dinier HC, Dichgans J, Guschlbauer B, Mau H. The significance of proprioception on postural stabilization as assessed by ischemia. *Brain Res.* 1984; 296:103–109. [PubMed: 6713202]
25. Lackner JR. Multimodal and motor influences in orientation: implications for adapting to weightless and virtual environments. *J Vestib Res.* 1992; 2:307–322. [PubMed: 1342405]
26. Jeka JJ, Easton RD, Bentzen BL, Lackner JR. Haptic cues for orientation and postural control in sighted and blind individuals. *Perceptions & Psychophysics.* 1996; 58:409–423.
27. Van den Bosch GG, Gilsing M, Lee S-G, Richardson JK, Ashton-Miller JA. Peripheral neuropathy effect on ankle inversion and eversion detected thresholds. *Arch Phys Med Rehabil.* 1995; 76:850–856. [PubMed: 7668957]

Figure 1.
Polypropylene ankle-foot orthosis.

Figure 2.

The six different sensory conditions of dynamic posturography. Under SOT 1 condition eyes are open and all sensory systems are available. Under SOT 2 condition (eyes closed) and SOT 3 condition (vision distorted through sway-referencing of the visual surround) anteroposterior sway is measured while the subjects stands on a fixed support. Conditions SOT 4, SOT 5, and SOT 6 duplicate conditions SOT 1, SOT 2, and SOT 3, respectively, regarding vision except the subject's support surface is sway-referenced. SOT 5 and SOT 6 allow to evaluate contribution of the vestibular system because both vision and somatosensation are compromised.

Figure 3. Sensory organization patterns of the subject on computerized dynamic posturography measured with no orthosis (A) and with bilateral rigid ankle-foot orthoses (B). The six sensory conditions refer to those shown in Figure 1. The ordinate refers to equilibrium score. Dotted background indicates lower limit of normal sway. Each bar represents results from a single 20-second trial of a given condition. The letters “N/S” on the chart indicate “no score,” which means the investigators did not run that trial. The word “fall” on the chart indicates that the trial was stopped because of the equilibrium loss and was marked as a fall. The “composite” bar on the right of the charts is averaged of scores of all single trials.