Published in final edited form as:

Phys Med Biol. 2015 January 7; 60(1): 301–307. doi:10.1088/0031-9155/60/1/301.

Consideration of the effects of intense tissue heating on the RF electromagnetic fields during MRI: Simulations for MRgFUS in the hip

Sherman Xuegang Xin^{1,2,*}, Shiyong Gu¹, Giuseppe Carluccio², and Christopher M. Collins²

¹Biomedical Engineering School of the Southern Medical University, Guangzhou city, China. Guangdong Province, China. 510515

²Bernard and Irene Schwartz Center for Biomedical Imaging, New York University School of Medicine, New York, New York, USA. 10016

Abstract

Due to the strong dependence of tissue electrical properties on temperature, it is important to consider the potential effects of intense tissue heating on the RF electromagnetic fields during MRI, as can occur in MR-guided focused ultrasound surgery. In principle, changes of the RF electromagnetic fields could affect both efficacy of RF pulses, and the MRI-induced RF heating (SAR) pattern. In this study, the equilibrium temperature distribution in a whole-body model with 2-mm resolution before and during intense tissue heating up to 60 °C at the target region was calculated. Temperature-dependent electric properties of tissues were assigned to the model to establish a temperature-dependent electromagnetic whole-body model in a 3T MRI system. The results showed maximum changes in conductivity, permittivity, $|B_1^+|$, and SAR of about 25%, 6%, 2%, and 20% respectively. Though the B_1 field and SAR distributions are both temperature-dependent, the potential harm to patients due to higher SARs is expected to be minimal and the effects on the B_1 field distribution should have minimal effect on images from basic MRI sequences.

Keywords

temperature-dependent electric properties; magnetic resonance imaging (MRI); radiofrequency (RF) electromagnetic fields; specific absorption rate (SAR)

Introduction

With magnetic resonance imaging (MRI) it is possible to map temperature change in vivo non-invasively (Rieke and Butts Pauly, 2008). MRI-guided focused ultrasound surgery (MRgFUS) is a relatively new technology (Bradley, 2009), used for the thermal ablation of uterine fibroid, pancreas tumors, etc. (Jenne *et al.*, 2012), or used for the palliative treatment

*To whom correspondence should be addressed. Sherman Xuegang Xin, the Biomedical Medical Engineering school, Southern Medical University, Guangzhou 510515 China, and the Department of Radiology, New York University, New York, NY, 10016 USA. shermansheen@gmail.com.

The authors report no conflicts of interests.

of painful bone metastases (Gianfelice *et al.*, 2008). In this technique, MRI and MR thermography can guide and monitor focused ultrasound thermal surgery to ensure that temperature is safely controlled. Based on a recent study of the measurement of dielectric properties of biological tissues, electrical conductivity of some tissues could change more than 20% during the process of MRgFUS (Fu *et al.*, 2014). Due to the strong dependence of tissue electrical properties on temperature (Lazebnik *et al.*, 2006; Zurbuchen *et al.*, 2010), it is important to consider the potential effects of intense tissue heating on the RF electromagnetic fields during MRI, as can occur in MRgFUS. Changes of the RF electromagnetic fields resulting from the changes in tissue electrical properties during an MR scan may affect, for example, both the efficacy of RF pulses and the MRI-induced specific absorption rate (SAR) pattern (Collins and Wang, 2011). To the best of our knowledge, however, a study of the potential effects of intense tissue heating on the RF electromagnetic fields during MRI has not previously been performed.

In this study, an equilibrium temperature distribution inside a body model with a resolution of 2 mm was calculated as may occur in focused ultrasonic heating for palliative care of patients with bone metastases (Gianfelice *et al.*, 2008). Simulations of RF electromagnetic fields in the body model before and during intense focal tissue heating related to thermal ablation were then performed. Temperature-induced changes of the distribution of RF electromagnetic fields and SAR in MRI are presented and discussed.

Methods

In this work, the temperature distribution in a whole-body model with 2-mm resolution during MRgFUS therapy was first calculated using a previously published algorithm (Collins *et al.*, 2004). Then the temperature-dependent dielectric properties of tissues were assigned to the whole-body model according to the calculated temperature distribution to establish a temperature-dependent electromagnetic whole body model. Finally, numerical simulations were performed to determine the changes of RF electromagnetic fields due to the intense tissue heating. These steps are described in further detail below.

A temperature distribution map from MRgFUS designed for pain palliation in a patient with iliac bone metastases was simulated using a 2mm numerical model of the human body, NAOMI (Dimbylow, 1997, 1998), by defining a focal region in the iliac bone where the temperature was fixed at 60°C (Gianfelice *et al.*, 2008). The focal region was spherical, with a diameter of 75 mm, and included only cortical bone. The equilibrium temperature distribution throughout the rest of body was calculated using a self-convergent iterative algorithm which considers thermal conduction, blood perfusion, and other relevant factors (Collins *et al.*, 2004). The algorithm uses a finite difference implementation of the Pennes bioheat equation (Pennes, 1948):

$$\rho C \left(\frac{dT}{dt} \right) = \nabla \cdot (k \nabla T) + [-\rho_{blood} W C_{blood} (T - T_{core})] + Q_m \quad [1]$$

where ρ is the material density, C is the heat capacity, k is the thermal conductivity, W is the perfusion by blood, Q_m is the heat generated by metabolism. All the values of the above

parameters and other details of the algorithm were obtained from the literature (Collins *et al.*, 2004).

After finishing the calculation of the equilibrium temperature distribution with and without the region of focal heating, recently-measured tissue-specific temperature-dependent electric properties in agreement with published values were assigned to the voxels of the body model in the region of high temperature to create a second model with electrical property distribution as could be expected in the high-temperature state (Fu *et al.*, 2014). The temperature-dependent dielectric properties of three tissues, muscle, fat, and bone, were needed in this study. While the temperature dependent dielectric properties of muscle and fat at 128 MHz have been reported (Fu *et al.*, 2014), data for bone at this frequency is not currently available from the literature. Because cortical bone has low water content, it (like fat) is expected to have relatively low electric permittivity and conductivity at any frequency or temperature, and thus its variation should have relatively little effect on the RF field distribution. In this work the temperature-dependent dielectric properties of bone were assigned those of fat. A 16-rung, low-pass birdcage coil with a diameter of 60 cm and a height of 40 cm loaded with the modified whole body model at 128 MHz (3 T) was used in the simulation (Fig 1). The finite difference time domain (FDTD) method for electromagnetics (Kane, 1966) was used to determine the steady-state RF electromagnetic field distributions. The spatial resolution of the calculation region was 2 mm in all three directions. All FDTD calculations were performed using commercial software (XFDTD; Remcom, Inc.; State College, PA) (Trakic *et al.*, 2004).

The circularly polarized RF transmit magnetic field $|\mathbf{B}_1^+|$, was calculated as (Hoult, 2000)

$$\mathbf{B}_1^+ = \frac{1}{2}(\tilde{\mathbf{B}}_x + i\tilde{\mathbf{B}}_y) \quad [2]$$

where $\tilde{\mathbf{B}}_x$ and $\tilde{\mathbf{B}}_y$ are the complex magnetic flux density of RF electromagnetic field along orthogonal directions in the transverse plane as obtained from the FDTD calculation. SAR_n was calculated at every location in the model as (Collins and Smith, 2001)

$$\text{SAR}_n = \frac{\sigma}{2\rho_n}(|\mathbf{E}_x|^2 + |\mathbf{E}_y|^2 + |\mathbf{E}_z|^2) \quad [3]$$

where $\mathbf{E}$, σ , and ρ refer to the RF electric field, tissue conductivity, and density, respectively. The subscript n indicates the voxel in the simulation region.

A gradient echo MR image was simulated by multiplying the amplitude of $\mathbf{B}_1^+$ by a value roughly proportional to tissue proton density content. This method assumes that a low excitation flip angle, a reconstruction method which removes the weighting of the RF receive coil distribution (Pruessmann *et al.*, 1999), and an effective method of fat suppression are used. Compared to an excitation flip angle near 90° , will accentuate effects of $\mathbf{B}_1^+$ inhomogeneity on the simulated image.

Results and Discussion

The comparison of temperature, conductivity, relative permittivity, $|B_1^+|$, and SAR before and after heating is shown in Fig 2. A temperature gradient distribution after equilibrium was established inside the body centered at the region of interest, extending into the surrounding muscle and fat tissues. The difference in temperature between the state with and without the ultrasonic heating was as high as 23 °C, and the percent of temperature difference on the Celsius scale was more than 60% (Fig 2). The maximum change in conductivity was more than 25%, while that in permittivity was only about 6%. The maximum change in $|B_1^+|$ was less than 2%, and that in the phase of the $|B_1^+|$ field was less than 6 degrees, indicating that only minor influence of the temperature-dependence of the $|B_1^+|$ field. The maximum relative change of SAR was about 20%. This occurred in the defined focal region, as indicated in Fig 2. In this case the region of ultrasonic heating (also the region of maximum change in SAR) was not near the region of maximum local SAR, and had no effect on this measure of RF safety. Even in the event that the region of ultrasonic heating would coincide with that of maximum local SAR, it is expected that maximal MR-induced heating (on the order of 1 °C) would be a small fraction of that due to the ultrasonic heating and would likely not be a major consideration for MR safety.

Simulated images from before and after ultrasonic heating are shown in Figure 3. As expected from the minimal changes in B_1^+ alone, the effects on image intensity due to temperature-dependent tissue electrical properties are very small. It is important to note, however, that this analysis, assuming a strongly proton density weighted image, does not consider effects of temperature on T_1 and T_2 that may affect signal intensity in images from other sequences. The effects of heating on the phase of the B_1^+ field should also have only a small effect on the accuracy of conventional phase-based MR thermography. For example, assuming a proton resonance shift based method of MR thermography (Reike and Butts Pauly, 2008) with a temperature-dependent chemical shift coefficient of -0.01 ppm and a gradient-echo sequence with a TR of about 20ms, a temperature increase of 20 °C should result in a change in phase of a little more than 180°. Thus the temperature-induced changes in the RF fields of about 6° could produce about an error of about 3% (or 0.6 degrees for the assumed 20 ° change), which is close to the uncertainty of *in vivo* MR thermography itself at 3T. The size of this relative error would be inversely proportional to the length of TE chosen for the MR thermography sequence.

This work represents an initial study into the potential effects of intense tissue heating on the RF fields and SAR in MRI, and was performed on a single model of a single human subject. Though the $|B_1^+|$ field and SAR distribution may vary with different human bodies having different morphologies (Liu *et al.*, 2005), the scale of temperature dependence should be similar. The temperature changes studied here, ranging to about 30 °C, may not cover the range for all treatment paradigms. Further study including a larger range would also require characterization of the temperature-dependence of tissue dielectric properties at MR frequencies over a larger range than is currently available. Also, while different MR sequences may produce $|B_1^+|$ field and SAR with different amplitudes, the distribution

patterns will not change such that the alterations of $|B_1^+|$ field and SAR in percentage before and after heating will be the same for any MR sequence.

This study does not consider the potential effect of cavitation bubbles, which may be produced by high intensity ultrasound pulses as another treatment strategy. Exploration of the possible impact of cavitation on the $|B_1^+|$ field and SAR in MRI would require further study and development of different model.

In thermal ablation, some thermally-induced physiological changes, such as those in blood perfusion or due to coagulation, will result in alterations of local volume magnetic susceptibility of tissues (Sprinkhuizen *et al.*, 2012). Effects of susceptibility on the $|B_1^+|$ field are only on the order of a few parts per million (ppm), and are negligible compared to the changes shown here (de Lacheisserie *et al.*, 2005). However, the alterations of the static magnetic (B_0) field of the order of ppm due to the change of local volume susceptibility of tissues cannot be neglected. The impact of temperature-dependent volume susceptibility on MR thermometry has been noticed and reported (Sprinkhuizen *et al.*, 2010).

Usually there are no metal parts of the ultrasound probes that could contact with or be inside the human tissues during the MRgFUS process. In rare cases, the MRgFUS surgery could be performed in patients who were already implanted metallic devices. MR imaging of patients with implanted metallic devices can result in very different SAR distributions than seen here (Mattei *et al.*, 2008), and is beyond the scope of this work. In the absence of implants and for lower temperature increases, field and temperature simulations like those seen here are seen to be in reasonably good agreement with measurement in phantoms and in vivo (Oh *et al.*, 2014).

Conclusion

Results show that the B_1 field and SAR distributions are both temperature-dependent. The potential harm to patients due to higher SAR is expected to be minimal in a situation such as MRgFUS, where temperature increase is dominated not by the RF fields of the MR system, but by the energy from the intense focused ultrasound system. The effects on the B_1 field distribution should have minimal localized effect on images and even MR thermometry measurements acquired with commonly used MRI sequences.

Acknowledgements

This study was partially supported by the National Natural Science Foundation of China (grant number 61172034), Science and Technology Program of Guangzhou, China (No.2014J4100160), and the National Institute of Health (grant numbers NIH R01 EB000454 and NIH R01EB011551).

References

- Bradley WG Jr. MR-guided focused ultrasound: a potentially disruptive technology. *Journal of the American College of Radiology : JACR*. 2009; 6:510–513. [PubMed: 19560068]
- Collins CM, Liu WZ, Wang JH, Gruetter R, Vaughan JT, Ugurbil K, Smith MB. Temperature and SAR calculations for a human head within volume and surface coils at 64 and 300 MHz. *J Magn Reson Imaging*. 2004; 19:650–656. [PubMed: 15112317]

- Collins CM, Smith MB. Signal-to-noise ratio and absorbed power as functions of main magnetic field strength, and definition of "90°" RF pulse for the head in the birdcage coil. *Magnetic Resonance in Medicine*. 2001; 45:684–691. [PubMed: 11283997]
- Collins CM, Wang Z. Calculation of radiofrequency electromagnetic fields and their effects in MRI of human subjects. *Magnetic Resonance in Medicine*. 2011; 65:1470–1482. [PubMed: 21381106]
- de Lacheisserie, EdT; Gignoux, D.; Schlenker, M. *Magnetism: Materials and applications*. Vol. 2. Springer; 2005.
- Dimbylow PJ. FDTD calculations of the whole-body averaged SAR in an anatomically realistic voxel model of the human body from 1 MHz to 1 GHz. *Phys Med Biol*. 1997; 42:479–490. [PubMed: 9080530]
- Dimbylow PJ. Induced current densities from low-frequency magnetic fields in a 2 mm resolution, anatomically realistic model of the body. *Phys Med Biol*. 1998; 43:221–230. [PubMed: 9509522]
- Fu FR, Xin SX, Chen WF. Temperature- and frequency-dependent dielectric properties of biological tissues within the temperature and frequency ranges typically used for magnetic resonance imaging-guided focused ultrasound surgery. *International journal of hyperthermia*. 2014; 30:56–65. [PubMed: 24417349]
- Gianfelice D, Gupta C, Kucharczyk W, Bret P, Havill D, Clemons M. Palliative treatment of painful bone metastases with MR imaging--guided focused ultrasound. *Radiology*. 2008; 249:355–363. [PubMed: 18695209]
- Hoult DI. The principle of reciprocity in signal strength calculations - A mathematical guide. *Concept Magnetic Res*. 2000; 12:173–187.
- Jenne JW, Preusser T, Gunther M. High-intensity focused ultrasound: principles, therapy guidance, simulations and applications. *Zeitschrift fur medizinische Physik*. 2012; 22:311–322. [PubMed: 22884198]
- Kane Y. Numerical solution of initial boundary value problems involving maxwell's equations in isotropic media. *Antennas and Propagation, IEEE Transactions on*. 1966; 14:302–307.
- Lazebnik M, Converse MC, Booske JH, Hagness SC. Ultrawideband temperature-dependent dielectric properties of animal liver tissue in the microwave frequency range. *Phys Med Biol*. 2006; 51:1941–1955. [PubMed: 16552116]
- Liu W, Collins CM, Smith MB. Calculations of B1 distribution, specific energy absorption rate, and intrinsic signal-to-noise ratio for a body-size birdcage coil loaded with different human subjects at 64 and 128 MHz. *Applied Magnetic Resonance*. 2005; 29:5–18. [PubMed: 23565039]
- Mattei E, Triventi M, Calcagnini G, Censi F, Kainz W, Mendoza G, Bassen HI, Bartolini P. Complexity of MRI induced heating on metallic leads: experimental measurements of 374 configurations. *BioMedical Engineering OnLine*. 2008; 7:11. [PubMed: 18315869]
- Oh S, Ryu Y-C, Carluccio G, Sica CT, Collins CM. Measurement of SAR-induced temperature increase in a phantom and in vivo with comparison to numerical simulation. *Magn Reson Med*. 2014; 71:1923–1931. [PubMed: 23804188]
- Pennes HH. Analysis of Tissue and Arterial Blood Temperatures in the Resting Human Forearm. *J Appl Physiol*. 1948; 1:93–122. [PubMed: 18887578]
- Pruessmann KP, Weiger M, Scheidegger MB, Boesiger P. SENSE: sensitivity encoding for fast MRI. *Magn Reson Med*. 1999; 42:952–962. [PubMed: 10542355]
- Rieke V, Butts Pauly K. MR thermometry. *J Magn Reson Imaging*. 2008; 27:376–390. [PubMed: 18219673]
- Sprinkhuizen SM, Bakker CJ, Ippel JH, Boelens R, Viergever MA, Bartels LW. Temperature dependence of the magnetic volume susceptibility of human breast fat tissue: an NMR study. *Magnetic Resonance Materials in Physics, Biology and Medicine*. 2012; 25:33–39.
- Sprinkhuizen SM, Konings MK, van der Bom MJ, Viergever MA, Bakker CJ, Bartels LW. Temperature-induced tissue susceptibility changes lead to significant temperature errors in PRFS-based MR thermometry during thermal interventions. *Magnetic Resonance in Medicine*. 2010; 64:1360–1372. [PubMed: 20648685]
- Trakic A, Crozier S, Liu F. Numerical modelling of thermal effects in rats due to high-field magnetic resonance imaging (0.5–1 GHz). *Physics in medicine and biology*. 2004; 49:5547. [PubMed: 15724541]

Zurbuchen U, Holmer C, Lehmann KS, Stein T, Roggan A, Seifarth C, Buhr HJ, Ritz JP.
Determination of the temperature-dependent electric conductivity of liver tissue ex vivo and in vivo: Importance for therapy planning for the radiofrequency ablation of liver tumours.
International journal of hyperthermia. 2010; 26:26–33. [PubMed: 20100050]

Author Manuscript

Author Manuscript

Author Manuscript

Author Manuscript

Figure 1.

Top: 3D rendering of the whole-body 3D model with 2-mm resolution loaded in birdcage coil. Bottom: Coronal slice chosen to show a variety of internal tissues.

Figure 2. Comparison of temperature, conductivity, relative permittivity, amplitude of B_1^+ , phase of B_1^+ , and SAR before and during heating. The distribution, the absolute difference, and the percent of difference of these parameters are shown on a transverse plane passing through the middle of the region of maximum temperature increase.

Figure 3.
Simulated proton density weighted gradient echo images before and during focal heating,
and the percent difference between them.