A simple technique to reduce surgeon radiation exposure during the use of the mini C-arm fluoroscope

J Smith, R Frood, S Southern

Mid Yorkshire Hospitals NHS Trust, UK

CORRESPONDENCE TO

James Smith, E: james.smith@doctors.org.uk

None of the authors have any commercial interest in the product described in this technical note.

BACKGROUND

The mini C-arm fluoroscope can provide intraoperative contemporaneous single and continuous screening radiographic images, and has become an important resource for hand surgery. There have been limited studies investigating the amount of radiation to which the surgical team is exposed when using the mini C-arm fluoroscope.^{1,2} For this reason, the lead author has developed a technique allowing mobilisation of the insensate patient hand when screening the stability of a fracture.

TECHNIQUE

Three ½ inch × 4 inch Suture Strip® wound closure strips (Derma Sciences, Maidenhead, UK) are attached to the fractured finger: two in a longitudinal fashion to the distal phalanx and one circumferentially at the level of the distal interphalangeal joint. The surgeon can hold or place a clip on to the end of the Suture Strips®. A second clip, forceps or marker is placed into the palm at the level of the metacarpal phalangeal joint, allowing for mobilisation of the finger without exposure of the surgeon to the radiation field (Fig 1).

DISCUSSION

This simple technique minimises surgeon exposure to radiation when using a mini C-arm fluoroscope.

References

1. Giordano BD, Ryder S, Baumhauer JF, DiGiovanni BF. Exposure to direct and scatter radiation with use of mini-c-arm fluoroscopy. *J Bone Joint Surg Am* 2007; **89**: 948–952.
2. Giordano BD, Baumhauer JF, Morgan TL, Reichtine GR. Patient and surgeon radiation exposure: comparison of standard and mini-c-arm fluoroscopy. *J Bone Joint Surg Am* 2009; **91**: 297–304.

Figure 1 Demonstrating the position of the Suture Strips® on the finger in full extension (left) and flexion (right)

Stryker® intracompartmental pressure monitor in the triage of circumferential limb burns

J May¹, A Sadri², F Wood³

¹Newcastle upon Tyne Hospitals NHS Foundation Trust, UK

²Frimley Health NHS Foundation Trust, UK

³Royal Perth Hospital, Australia

CORRESPONDENCE TO

Jolyon May, E: jolyonmay@yahoo.co.uk

None of the authors have any commercial interest in the product described in this technical note.

BACKGROUND

Compartment syndrome is a life and limb threatening emergency associated with circumferential limb burns. Immediate surgical decompression is the treatment of choice but the diagnosis can be difficult to make in the unconscious patient.

Following a boat explosion, 23 patients, with a mean burn area of 24%, were transferred to Royal Perth Hospital in 12 hours. Many of the patients were sedated. Seven had circumferential limb burns that would potentially require decompression. To assist with the clinical assessment, the Stryker® (Kalamazoo, MI, US) pressure device (Fig 1) was used.

TECHNIQUE

The device was inserted into the compartments of the affected limbs (16 in total). A pressure of 30mmHg was used as the threshold for

Figure 1 Stryker® device

Figure 2 Lower limb escharotomies