Available online at www.sciencedirect.com

ScienceDirect

journal homepage: www.elsevier.com/locate/mjafi

Case Report

Salivary reservoir denture – A novel approach to battle xerostomia

Maj Kamal Verma^{a,*}, Col E. Mahesh Gowda^b, Brig V.R. Pawar^c,
Maj Ashish Kalra^a

^a Resident (Prosthodontics), Army Dental Centre (R&R), New Delhi 110010, India

^b Dy Comdt & Senior Resident (Prosthodontics), Army Dental Centre (R&R), New Delhi 110010, India

^c Commandant and HOD, (Orthodontics), Army Dental Centre (R&R), New Delhi 110010, India

ARTICLE INFO

Article history:

Received 3 November 2011

Received in revised form

16 February 2012

Accepted 9 March 2012

Available online 10 September 2012

Keywords:

Salivary reservoir

Xerostomia

Artificial saliva

Introduction

Xerostomia is the subjective feeling of dryness in the mouth, often referred to as reduced salivary flow. Individuals usually detect they have a “dry mouth” when the flow of saliva decreases to approximately half the normal unstimulated rate of around 0.3 ml/min.¹

Causes of xerostomia are anxiety, Sjogren's syndrome, salivary gland diseases, drug induced side effects, sequelae to head and neck radiation and general medical conditions such as diabetes mellitus. Patients suffering from xerostomia may

complain of not only a dry mouth, but also of difficulty in normal oral and oropharyngeal functions including eating, speaking and swallowing. Extreme discomfort in wearing dentures is a common complaint.^{2,3}

Depending upon the cause, a variety of treatment options are available. In cases where all treatment modalities have proven unsuccessful, the incorporation of reservoirs containing salivary substitutes into dentures has been proposed.

This paper presents a case report where a reservoir denture was used successfully in rehabilitating a xerostomic complete edentulous patient.

* Corresponding author. Tel.: +91 9650255935.

E-mail address: kamalverma_in@yahoo.com (K. Verma).

0377-1237/\$ – see front matter © 2012, Armed Forces Medical Services (AFMS). All rights reserved.

doi:[10.1016/j.mjafi.2012.03.001](https://doi.org/10.1016/j.mjafi.2012.03.001)

Fig. 1 – a: Lower denture base fabrication with Lego® blocks and precision buttons. b: Reservoir space cut out on both sides of clear acrylic base.

Case report

A 54 years old male reported with chief complaints of missing all upper and lower teeth and dry mouth. Medical history revealed that the patient had been operated for carcinoma base of tongue followed by radiation therapy 2 years back resulting in reduced saliva. Intra oral examination revealed edentulous maxillary and mandibular arches, dry tongue, and minimal saliva in oral cavity.

Primary impressions were made in high fusing impression compound (Pinnacle, DPI, India) and final impressions were made by using special trays in light body polysiloxane impression material (Aquasil, Denstply) since zinc oxide eugenol paste causes dryness of mucosa and burning sensation. Maxillomandibular relationships were recorded and models were articulated. A clear acrylic base was fabricated with press fit buttons and plastic (Lego®) blocks as shown in Fig. 1a kept parallel to each other on either side and at centre and waxed in such a way that only the extensions of the plastic blocks were above the wax rims on the mandibular cast. The height of the base was determined by measuring the total height of the mandibular bite rim (a) and subtracting the desired height of the teeth to be set plus 3 mm (b) added on to allow for sufficient acrylic under the teeth for strength [Fig. 2]. Shorter teeth were

used in the lower rim to allow a deeper area for the future placement of reservoirs.

The clear acrylic base was resealed on the duplicated mandibular cast and the assembly was duplicated using addition silicone material (Elite Double 32, Zhermack) and a stone cast was retrieved. In order to articulate this stone duplicated cast in the corresponding position to the clear acrylic base, a wax squash bite was made on the articulator, between the upper wax rim and the clear acrylic base. The clear acrylic base was then removed and the stone duplicated cast placed in its position and articulated. Teeth arrangement was done and tried in patient's mouth [Fig. 3].

The flasking was done in heat cure (Trevalon, Denstply) acrylic resin. After careful deflasking and polishing of upper and lower segment of the mandibular denture, the two sections should 'click' into place at this point. Polishing was done with the segments together to ensure a flush, smooth finish and no damage to the edges [Fig. 4].

The reservoirs were cut on both sides of the clear acrylic base while maintaining sufficient thickness of the denture walls for strength. A minimum thickness of 2 mm was maintained for the reservoir walls [Fig. 1b]. Once the reservoirs were placed, a 0.5 mm diameter acrylic bur was used to drill a drainage hole from the inferior aspect of the lingual flange of the denture into the reservoirs on both sides.

Fig. 2 – Showing the available space for denture base (a-b) and teeth arrangement.

Fig. 3 – Try in patient's mouth.

Fig. 4 – Showing the split and polished mandibular denture.

Fig. 5 – Showing the comparative evaluation of patients comfort.

Instructions were given to the patient about how to separate and clean the two halves of the denture. The patient was instructed to flush out the reservoirs weekly with 1% sodium hypochlorite solution and clean the reservoirs daily with either a small toothbrush or bottlebrush. Fine orthodontic wire was provided to clean the drainage holes if they became blocked. The patient needed to refill the reservoirs twice per day with artificial saliva (Wet Mouth, ICPA). The patient found a great reduction in his symptoms of Xerostomia and was comfortable in his recalled visits [Fig. 5].

Discussion

Upon review of the literature two main approaches to radiation-induced xerostomia surfaced, prevention and treatment. In mild xerostomic cases, gustatory stimulation of the salivary glands by mastication of sugar-free chewing gums or lozenges is helpful.⁴ In severe cases, saliva substitutes may be utilized. The advantage of this split denture technique lies in the ready access to the reservoirs. The use of clear acrylic for the base section permits the clinician to determine the best size and position for placement of the reservoirs. Additionally, it also enables the patient to clearly visualize the levels of salivary substitute within the chamber.⁵

Case selection is also extremely important. Cutting reservoirs into the denture weakens its structure, so only cases with no bony undercuts and sufficient vertical dimension, and thickness are suitable.⁶ Additionally, patients must have the manual dexterity to separate and rejoin the two segments of the split denture. Poor maintenance of denture by patient leads to fungal growth specially *Candida* biofilms.⁷ Evaluation of study showed that the metal base surface exhibited significantly less growth of *Candida* as compared to resin.⁸ From a clinician's perspective, little additional chair-side time is required. However, laboratory stages are time consuming and precision is essential. Additionally, repairs and relines of a split denture become more complex and collection of foods and fluids might clog the holes of reservoir. Many other materials like magnets, precision attachments are also used in the literature.⁹

Previous studies suggest that reservoirs in the palatal aspect of maxillary dentures may also be useful in treating xerostomia.¹⁰

Conclusion

In summary, this paper presents a novel technique for the fabrication of a mandibular denture incorporating a salivary reservoir. Xerostomic patients form a unique group in whom prosthodontic treatment is challenging. Successful treatment depends on

knowledge and recognition of their particular problems and methods of prevention combined with skilful prosthodontics.

REFERENCES

1. Greenspan D. Xerostomia: diagnosis and management. *Oncology*. 1996;10:7–11.
2. Sinclair GF, Frost PM, Walter JD. New design for an artificial saliva reservoir for the mandibular complete denture. *J Prosthet Dent*. 1996;75:276–280.
3. Vergo TJ, Kadish SP. Dentures as artificial saliva reservoirs in the irradiated edentulous cancer patient with xerostomia: a pilot study. *Oral Surg Oral Med Oral Pathol*. 1981;51:229–233.
4. Toljanic JA, Zucuskie TG. Use of a palatal reservoir in denture patients with xerostomia. *J Prosthet Dent*. 1984;52: 540–544.
5. Mendoza AR, Tomlinson MJ. The split denture: a new technique for artificial saliva reservoirs. *Aust Dent J*. 2003;48(3): 190–194.
6. Hirvikangas M, Posh J, Makila E. Treatment of xerostomia through use of dentures containing reservoirs of saliva substitute. *Proc Finn Dent Soc*. 1989;85:47–50.
7. Ramage G, VandeWalle K, Wickes BL, Lopezribot JL. Biofilm formation by *Candida dubliniensis*. *J Clin Microbiol*. 2001;39:3234–3240.
8. Pryor WJ. Swaged denture bases. *J Am Dent Assoc*. 1928;15:1281–1288.
9. Gary FS, Peter MF, John WD. A new design for an artificial saliva reservoir in mandibular complete denture. *J Prosthet Dent*. 1996;75:276–280.
10. Vissink A, Huisman MC. Construction of artificial saliva reservoir in maxillary denture. *J Prosthet Dent*. 1986;56:70–74.