Received:
8 February 2015

Revised:
15 April 2015

Accepted:
26 May 2015

doi: 10.1259/bjr.20150126

Cite this article as:

Ma WK, Aspin R, Kelly J, Millington S, Hogg P. What is the minimum amount of simulated breast movement required for visual detection of blurring? An exploratory investigation. *Br J Radiol* 2015; **88**: 20150126.

FULL PAPER

What is the minimum amount of simulated breast movement required for visual detection of blurring? An exploratory investigation

¹W K MA, MSc, ²R ASPIN, PhD, ³J KELLY, DCR, MSc, ³S MILLINGTON, BSc (Hons), PGD and ¹P HOGG, FCR,

¹Directorate of Radiography, University of Salford, Salford, UK

²Directorate of Computer Science & Software Engineering, School of Computing, Science and Engineering, University of Salford, Salford, UK

³Department of Radiography, Countess of Chester Hospital, Chester, UK

Address correspondence to: Mr Wang Kei Ma

E-mail: caraby2000@gmail.com

Objective: Image blurring in mammography can cause significant image degradation and interpretational problems. A potential source is due to paddle movement during image formation. Paddle movement has been shown to be as much as 1.5 mm. No study has yet been performed to determine how much motion would be noticeable visually. The aim of this study is to determine the minimum amount of simulated breast movement at which blurring can be detected visually.

Methods: 25 artefact-free mammogram images were selected. Mathematical simulation software was created to mimic the effect of blurring produced by breast movement during exposure. Motion simulation was imposed to 15 levels, from 0.1 to 1.5 mm stepping through 0.1 mm increments. 15 degraded images and 1 without blurring were de-identified, randomized and assessed on

a blinded basis by two clinical experts to determine the presence or absence of blurring. Statistical testing was carried out to determine the consistency between the two observers.

Results: The probability of simulated blurred image detection is the highest for the gaussian method and the lowest for soft-edged mask estimation.

Conclusion: The amount of simulated breast movement at which blurring can be detected visually for gaussian blur, hard-edge mask estimation and soft-edge mask estimation is 0.4, 0.8 and 0.7 mm, respectively. Cohen's kappa for all the levels of simulated blurring is 0.689 ($p < 0.05$).

Advances in knowledge: This research establishes the concept of using probability to represent visual detection of blurring rather than defining a hard cut-off level.

Image blurring in mammography causes image degradation, which can lead to problems with interpretation. Blurring has been detected during routine clinical practice, and according to the results of a local audit, it was the reason behind 87% of repeat images.¹ Within the UK, blurred mammograms have been increasingly noticed, particularly since the introduction of full-field digital mammography (FFDM).¹ Image blurring was probably present on conventional film, but it is easier to detect by FFDM owing to improvements in contrast resolution. A study by Saunders et al² shows that the presence of noise and blurring would significantly reduce lesion detection accuracy; blurring may mask certain types of lesions, which can result in false-negative or false-positive outcomes.

Image blurring may be caused by a number of factors such as geometric distortion from the finite focal spot size, digitizer blur, patient movement and undercompression.^{3,4} Studies have shown that using too little compression can result in blurring;

also, one study has demonstrated that compression force is not evenly distributed across the breast, and for the mediolateral oblique view, a lot of compression is lost in the pectoral muscle leading to undercompression in the rest of the breast.⁴

Human perception of blurring could be affected by the distribution of contrast across different spatial frequencies. For instance, if the high spatial frequencies are filtered by a low-pass spatial filter, the image will appear blurred, while the image will appear sharp if low spatial frequencies are filtered by a high-pass spatial filter. Psychophysical studies suggest that perception of blurring can be controlled by adaptive adjustment of the spatial frequencies.⁵ Literature on motion perception concurred with the psychophysical findings and suggested that image-processing algorithms could affect observer's perception of motion because the noise suppression for these processing algorithms not only changes the spatial frequencies but also reduces sharp transitions between pixel intensities; these

Figure 1. 1 of the 25 blur-free images. RMLO, right mediolateral oblique view.

modifications affect how human observers perceive motion in the presence of image degradation.^{6,7}

Blurring can result in mammography images needing to be repeated; this increases client radiation dose and anxiety. Recently,

paddle movement has been suggested as a potential source of blurring. Ma *et al*⁸ observed movement attributed to the paddle using an experimental approach in which metal ball-bearing diameters altered during compression of a deformable breast phantom in FFDM; change in diameter was considered to be the result of paddle motion. In a multicentre study, the extent of the paddle movement was observed to be as much as 1.5 mm during the acquisition/exposure period, in the vertical plane.⁹

Whilst these observations have been reported, no study has yet investigated how much motion needs to occur before the effect is visually noticeable on a mammogram. Therefore, the aim of this exploratory investigation was to determine the minimum amount of breast movement that needs to occur for blurring to be visually evident. Software-based image processing has been used to replicate the visual effect of breast movement. This study is a compromise because of the practical difficulties encountered when trying to control motion to 0.1-mm precision on an FFDM mammographic machine, so a physical approach for creating motion was not adopted.

METHODS AND MATERIALS

Image acquisition

Mammography images were acquired on a Selenia® Dimensions® (Hologic®, Bedford, MA) FFDM machine within the National Health Service Breast Screening Programme (NHSBSP). Routine equipment quality assurance had been performed, and the results complied with manufacturer specifications.¹⁰

Images selection

Images were only included if blur was not present.^{11,12} Blur-free images were identified by selecting images in which breast anatomical structures had distinct/sharp edges (Figure 1). Hundreds of images were reviewed initially on 5-MP monitors

Figure 2. Simulation of pixel intensity accumulation while a pixel is under random motion using a uniform intensity block to represent a single pixel within the image. The spread of the pixel is represented by the spread dot in the lower part of the figure and analysed by sampling a diagonal cross-section through the spread point plotted in the upper chart with a smooth gaussian function matched to the sample line. This is the basis of the hard-edged mask; a profile dot of this spread point is presented in figure 5.

Figure 3. Simulation of pixel intensity accumulation while a pixel is under random motion using a gaussian distribution to represent a single pixel within the image. The spread of the pixel is represented by the spread dot in the lower part of the figure and analysed by sampling a diagonal cross-section through the spread point plotted in the upper chart with a smooth gaussian function matched to the sample line. This is the basis of the soft-edged mask; a profile plot of this spread point is presented in [Figure 6](#).

within the breast unit, and 100 met the inclusion criteria. These had been scrutinized by imaging practitioners as part of the routine clinical processes for breast screening within the UK and

also through additional quality review as part of this study. A university professor who regularly conducts and publishes research into dose and image quality optimization identified the

Figure 4. Gaussian mask for 1.5-mm simulated motion.

Figure 5. Hard-edged pixel mask for 1.5-mm simulated motion.

25 images that were included in this study. Selection of the final 25 images did not involve the observers who later assessed the 1200 images for the absence/presence of blurring.

Motion simulation

It is extremely difficult to control submillimetre movement of a mammogram accurately in a physical study. Software was therefore developed and validated to simulate motion in mammogram image data sets. Research by Ma *et al*⁸ suggests that the amount of movement during a mammography exposure in a female breast ranges from 0.4 to 1.5 mm, in the vertical plane. Our software motion simulation was designed to impose 15 levels of motion, from 0.1 to 1.5 mm stepping through 0.1 mm. Using this approach, it is possible to create 15 blurred images for each image.

Blurring was created by the accumulated pixel points moving under random motion. After sampling blurred image data from real patient images, we identified three mathematical techniques that are suitable to simulate motion blurring: gaussian, hard-edge mask estimation and soft-edge mask estimation.^{13,14} Currently, there is a paucity of literature about motion simulation techniques;

therefore, all three mathematical simulation techniques were used in this study.

Evaluation of pixel motion was made by accumulating the pixel intensity of randomized microsteps, within 1.5 mm of the motion boundary. Propagation of the pixel intensity to the accumulation plane was initially enacted as hard-edged block pixel and soft-edged block pixel, respectively, with a further refinement to apply a gaussian distribution centred on the pixel step position, with mask dimension proportional to the original pixel size. The resultant accumulation of intensity was sampled by a diagonal section of the spread pixel to determine the profile of intensity spread owing to motion. Matching this to the standard gaussian distribution, commonly used to enact blurring in image processing, showed that for the motion enacted by the actual pixel a significantly steeper drop-off and narrower spread of intensity was presented (Figures 2 and 3).

Averaged data from multiple runs of this experiment were evaluated using curve-fitting processes to determine a curve function to represent the transition. This function was used to populate convolution masks for enacting blurring in the sample

Figure 6. Soft-edged pixel mask for 1.5-mm simulated motion.

images (Figures 4–6). Blurred images were generated for all three approaches (Figures 7–9) by applying the convolution masks to the source images.

For each source image (image without blurring), 15 processed images were created with between 0.1 and 1.5 mm of simulated movement, and these were exported as a randomized digital imaging and communications in medicine (DICOM) sequence for evaluation by two observers.

Visual assessment of images

MediViewer viewing software (Schaeff Systemtechnik, Petersaurach, Germany) was used to display the images. 1200 images were reviewed for the 3 simulation methods; 25 DICOM image sets and 15 levels of motion plus 1 without blurring ($3 \times 25 \times 16$). Ambient light level in the reading room was set to be less than 10 lux.¹⁵ On a blinded and individual basis, the images were displayed in a random order to the two observers. The two observers were qualified radiographers with specialist post-graduate qualifications in mammography imaging; they had been trained to interpret mammograms, and they had been doing this role independently for 7 and 15 years, respectively. As part of this role, they participate annually in the PERFORMS (Loughborough University, UK)¹⁶ image reading analysis; this independent analysis has always resulted in their scores being well above the expected mark for

interpreting mammography images within the NHSBSP. Both reviewers therefore had proven technical and clinical skills in image quality appraisal from technical and diagnostic points of view.

The presence or absence of blurring was a purely binary decision—indicating whether it was present (yes) or not (no). Kappa analysis was carried out to determine variability between the two observers. The presence of motion blurring was assessed by the observers with reference to the guidelines published by the NHSBSP and through determining whether breast anatomical structures had distinct/sharp edges.¹⁷ To minimize eye strain/fatigue, the observers did not score images for longer than 30 min in any one session. Images were viewed on a 21.3-inch liquid crystal display monitor (EIZO, Ishikawa, Japan) with a resolution of 2560×1920 (5 MP). The viewing monitors were calibrated to the DICOM Grayscale Standard Display Function.¹⁸

RESULTS

The probability of detecting blurring for gaussian blur, hard-edge mask estimation and soft-edge mask estimation is shown in Tables 1–3. All the non-motion images were identified correctly by the observers. Cohen's kappa (κ) for all the levels of simulated blurring is 0.689 ($p < 0.05$), representing good agreement for the two observers.¹⁹ The average probability of

Figure 7. Blurred image generated by gaussian for 1.5-mm simulated motion. RMLO, right mediolateral oblique view.

detecting blurring rises from 40% (0.1 mm) to 92% (0.3 mm) for gaussian blurring (Table 1); from 10% (0.1 mm) to 98% (0.7 mm) for hard-edge mask estimation (Table 2); and from 4% (0.1 mm) to 98% (0.6 mm) for soft-edge mask estimation (Table 3). The average probability of detecting blurring is the highest in gaussian blur (40%), which means when the amount of breast movement is 0.1 mm, there is 40% chance that the blurring can be detected visually. The amount of simulated breast movement at which 100% of blurred images can be detected visually for gaussian blur, hard-edge mask estimation and soft-edge mask estimation is 0.4, 0.8 and 0.7 mm, respectively.

DISCUSSION

The rationale for choosing the image algorithms

The gaussian algorithm was chosen as it is the standard approach for image blurring; it replicates the effect of putting a translucent film over the image and scatters light uniformly. Initial review by the clinicians showed that this was similar to the blurring effect seen but that it was creating too much blur for the movement levels.

The pixel walking application was then set up to examine what happens to a point in the breast under motion during the (pixel) acquisition. The hard-edged mask estimation made the assumption that a single point in the image would map to a single pixel at every point of its motion. The softened mask estimation further refined this by applying a convolution function based on a gaussian distributed pixel under simulated motion, instead of a uniform intensity pixel

Figure 8. Blurred image generated by hard-edged estimation for 1.5-mm simulated motion. RMLO, right mediolateral oblique view.

as used in the hard-edged simulation. This produced a slight relaxation to the simulated motion profile, and resultant convolution mask, but more closely reflected the actual

Figure 9. Blurred image generated by soft-edged estimation for 1.5-mm simulated motion. RMLO, right mediolateral oblique view.

Table 1. Probability of detecting blurring for gaussian blurring

Movement (mm)	0.1	0.2	0.3	0.4	0.5	0.6	0.7	0.8	0.9	1.0	1.1	1.2	1.3	1.4	1.5
Observer 1, probability (%)	24	64	84	100	100	100	100	100	100	100	100	100	100	100	100
Observer 2, probability (%)	56	100	100	100	100	100	100	100	100	100	100	100	100	100	100
Average probability (%)	40	82	92	100	100	100	100	100	100	100	100	100	100	100	100

energy propagation in the capture system. Validation was also performed by calculation against grid images to ensure that the amount of spread (size of mask) was appropriate.

Blurring masks

While the convolution mask used in the gaussian blurring approach is the same size as for the other approaches, with the size of the mask defined by the level of motion simulated, the more gradual spread of the function creates a greater distribution of the pixel intensity within the specific motion. Gaussian blurring is commonly used to simulate blurring in image processing, and it assumes a model similar to placing an opaque diffusion filter over the image. However, the blur within the images here is based on an actual motion of the subject during the time the pixel was captured. Therefore, it is not a good fit for replicating the characteristic blur seen in these image types. The simulation process we undertook, which replicates the subject motion as a stochastic sampling process, shows that the profile of distribution within a motion-based blur is characterized by a function in which the intensity drop-off is much more rapid (Figures 2 and 3). This means that for the same level of simulated motion, application of a gaussian-based function delivers a greater level of visual blurring than our evaluated mask suggests, explaining why the detection of blur within the gaussian image sequences is more detectable for smaller levels of motion.

The two simulated convolution masks evolved as part of an iterative process. The initial prototype considered a sample of the same size as the capture pixel that was moved during the simulated walking by fixed step sizes of the same unit. This demonstrated the discrepancy between the general case gaussian mask profile and motion-based simulation within the characteristics of the imaging system. Given this initial validation, a refined variant of the simulation process was evolved in which a soft-edged subject point was translated by variable

step sizes within the motion walking process to more accurately reflect the analogue nature of motion within the imaging process. For completeness in the evaluation, images from both the initial prototype and the refined (soft-edged) approach were used. The more subtle simulation of motion within the soft-edged approach delivered a slightly more diffuse distribution of the pixel blur, which makes the perception of blur slightly more prevalent at lower motion levels (Table 3) than for the hard-edged approach (Table 2), but this latter approach is more reflective of the actual physical effect of motion within the imaging process.

Perception of blurring

Visual perception of blurring can vary between individuals, and factors such as visual acuity, experience/ability and circadian rhythm variations can account for these differences.²⁰ Consequently, in this study, we used probability to represent blurring detection rather than defining a hard cut-off level. The level of agreement between the two observers is reasonable ($\kappa = 0.698$); however, their level of agreement was not perfect ($\kappa = 1$). Nevertheless, the level of agreement is good and provides a basis upon which conclusions can be drawn. Further research needs to be conducted to identify and assess the factors, which may affect the ability of observers to visually detect blurring.

Full-field digital mammography imaging and display system resolution

In this study, we have shown that only small amounts of simulated motion are needed in FFDM images for blurring to be seen visually. If our simulation models reflect actual motion, then our results could have clinical importance, as they could help to explain why there has been an increase in blurred images since the introduction of FFDM. Furthermore, as technical advances continue to be made in acquisition and display resolution, then blurred images could become more

Table 2. Probability of detecting blurring for hard-edged mask estimation

Movement (mm)	0.1	0.2	0.3	0.4	0.5	0.6	0.7	0.8	0.9	1.0	1.1	1.2	1.3	1.4	1.5
Observer 1, probability (%)	12	28	32	56	72	72	96	100	100	100	100	100	100	100	100
Observer 2, probability (%)	8	80	84	88	100	100	100	100	100	100	100	100	100	100	100
Average probability (%)	10	54	58	72	86	86	98	100	100	100	100	100	100	100	100

Table 3. Probability of detecting blurring for soft-edged mask estimation

Movement (mm)	0.1	0.2	0.3	0.4	0.5	0.6	0.7	0.8	0.9	1.0	1.1	1.2	1.3	1.4	1.5
Observer 1, probability (%)	8	32	52	72	96	96	100	100	100	100	100	100	100	100	100
Observer 2, probability (%)	0	60	96	100	100	100	100	100	100	100	100	100	100	100	100
Average probability (%)	4	46	74	86	98	98	100	100	100	100	100	100	100	100	100

problematic. Further research is needed to understand whether this would be the case.

An additional problem related to image blurring concerns the resolution of the monitors used in clinical room, as it is here where the images are checked for technical accuracy prior to sending the client home. Although monitor specification has risen in recent years, for monitors used for checking images prior to the client leaving the actual resolution can be as low as 1.8 MP. We speculate that imaging practitioners might not see visual blurring on monitors at this resolution, and as local audits show, it is only when images are displayed on 5-MP screens that blurring becomes apparent. This disparity might be a cause of technical recalls noted in audits. Further research needs to be conducted to assess higher and lower monitor resolution with respect to blurring detection. In light of this, we have already commenced a project, using simulated blurred images, to investigate the effect of monitor resolution on blurring detection.

Lesion detection performance

In this study, we have demonstrated that simulated blurring can be detected visually even when there is only a small amount of breast movement (e.g. 0.1 mm). We have not addressed the most important question of whether blurring affects lesion detection performance and lesion characterization. Consequently, further research needs conducting to address this.

Ways to reduce blurring

Sources of blurring can be summarized into three categories: patient, practitioner and machine related. Breast compression can only deal with patient-related factors such as patient movement. Practitioner-related factors can be minimized by good patient positioning skills; for example, patient movement could be avoided through sufficient compression and good

patient communication.²¹ Movement at the paddle is an example of a machine-related factor; according to the research by Ma et al,⁸ paddle movement is the highest in the first 10 s. Therefore, if practitioners could wait a few seconds after the compression force ceases to be applied before making the exposure, blurring could be minimized. Further research could be carried out to investigate the optimum waiting time to cease the compression.

Limitations

One of the limitations of our work is the use of motion-simulated images. Motion simulation might not fully represent real/physically blurred images because mammography image blurring may fully or partly affect the image; our software simulation only fully blurred the images. Since conducting this study, we have refined our mathematical model to introduce regional blurring to better reflect clinical reality. Another limitation is the number of observers involved. Since this is an exploratory investigation, only 2 observers and 25 original images were involved. We have already extended our work to include a larger sample of observers.

CONCLUSION

All non-motion images were identified by the observers. The probability of simulated blurred image detection is the highest for the gaussian method and the lowest for soft-edged mask estimation. The amount of simulated breast movement at which 100% of blurred images can be detected visually for gaussian blur, hard-edge mask estimation and soft-edge mask estimation is 0.4, 0.8 and 0.7 mm, respectively. Our results could have clinical importance, and they reinforce the need to minimize patient motion during the acquisition process. It is likely that image blurring will become even more apparent when improvements are made in FFDM acquisition and display system resolution.

REFERENCES

- Kelly J, Hogg P, Szczepura K, McGeever G, Wilcock C, Tinston S, et al. *The blurring issue in mammography*. Mammographic Imaging Seminar; December 2011. Salford, UK: University of Salford; 2011.
- Saunders RS, Baker JA, Delong DM, Johnson JP, Samei E. Does image quality matter? Impact of resolution and noise on mammographic task performance. *Med Phys* 2007; **34**: 3971. doi: [10.1118/1.2776253](https://doi.org/10.1118/1.2776253)
- Highnam R, Brady JM. *Computational imaging and vision mammographic: image analysis*. London, UK: Springer; 1999.
- Dustler M, Andersson I, Brorson H, Fröjd P, Mattsson S, Tingberg A, et al. Breast compression in mammography: pressure distribution patterns. *Acta Radiol* 2012; **53**: 973–80. doi: [10.1258/ar.2012.120238](https://doi.org/10.1258/ar.2012.120238)
- Webster MA, Georgeson MA, Webster SM. Neural adjustments to image blur. *Nat Neurosci* 2002; **5**: 839–40. doi: [10.1038/nn906](https://doi.org/10.1038/nn906)

6. Massanes F, Brankov G. Motion perception in medical imaging. In: Manning DJ, Abbey CK, eds. *Proceedings of SPIE. Medical imaging 2011: image perception, observer performance, and technology assessment*. Vol. 7966; 2011. p. 796610.
7. Gonzalez RC, Woods RE. *Digital image processing*. 3rd edn. London, UK: Pearson; 2008.
8. Ma WK, Hogg P, Kelly J, Millington S. A method to investigate image blurring due to mammography machine compression paddle movement. *Radiography* 2015; **21**: 36–41. doi:10.1016/j.radi.2014.06.004
9. Ma WK, Brett D, Howard D, Kelly J, Millington S, Hogg P. Extra patient movement during mammographic imaging: an experimental study. *Br J Radiol* 2014; **87**: 20140241. doi: 10.1259/bjr.20140241
10. Moore AC, Dance DR, Evans DS, Lawinski CP, Pitcher EM, Rust A, et al. *The commissioning and routine testing of mammographic X-ray systems: a technical quality control protocol*. Report No. 89. York, UK: IPeM; 2005.
11. Taplin SH, Rutter CM, Finder C, Mandelson MT, Houn F, White E. Screening mammography: clinical image quality and the risk of interval breast cancer. *AJR Am J Roentgenol* 2002; **17**: 797–803. doi: 10.2214/ajr.178.4.1780797
12. Samei E. Performance of digital radiographic detectors: quantification and assessment methods. *Advances in digital radiography: RSNA categorical course in diagnostic radiology physics*. Oak Brook, IL: Radiological Society of North America; 2003. pp. 37–47.
13. Young SS, Driggers GR, Jacobs LE. *Signal processing and performance analysis for imaging systems*. New York, NY: Artech House; 2008.
14. Dougherty G. *Digital image processing for medical applications*. Cambridge, UK: Cambridge University Press; 2009.
15. Perry N, Broeders M, Wolf C, Törnberg S, Holland R, Karsa L, et al. *European guidelines for quality assurance in breast cancer screening and diagnosis*. 4th edn. Luxembourg: European Communities; 2006.
16. Gale AG, Scott H. Measuring radiology performance in breast screening. In: Michell M, ed. *Contemporary issues in cancer imaging —breast cancer*. Cambridge, UK: Cambridge University Press; 2010.
17. Pritchard J; The National Health Service Breast Screening Programme. *Quality assurance guidelines for mammography*. Vol. 63. Sheffield, UK: National Health Service Breast Screening Programme; 2006.
18. National Electrical Manufacturers Association (NEMA). *Digital imaging and communications in medicine (DICOM). Part 14: Grayscale standard display function*. Arlington, VA: National Electrical Manufacturers Associations; 2011.
19. Hicks CM. *Research methods for clinical therapists: applied project design and analysis*. 5th edn. Edinburgh, UK: Churchill Livingstone; 2009.
20. Lanca C, Lanca L, Thompson J, Hogg P. A method to assess visual function prior to conducting medical imaging research using perceptual methodologies. *Radiol Tech* 2015; in press.
21. Hogg P, Kelly J, Claire E, eds. *Digital mammography: a holistic approach*. 1st edn. London, UK: Springer; 2015.