Published in final edited form as:

Rep U S. 2013 November ; 2013: 3071–3076. doi:10.1109/IROS.2013.6696791.

A Cooperatively Controlled Robot for Ultrasound Monitoring of Radiation Therapy

H. Tutkun en¹, Muyinatu A. Lediju Bell¹, Iulian Iordachita², John Wong³, and Peter Kazanzides¹

¹H. Tutkun en, Muyinatu A. Lediju Bell and Peter Kazanzides are with the Dept. of Computer Science, Johns Hopkins University, 3400 N. Charles Street, Baltimore, MD, USA

²Iulian Iordachita is with the Dept. of Mechanical Engineering, Johns Hopkins University, 3400 North Charles Street, Baltimore, MD USA iordachita@jhu.edu

³John Wong is with the Dept. of Radiation Oncology, The Johns Hopkins Hospital, 600 N. Wolfe Street, Baltimore, MD USA jwong35@jhmi.edu

Abstract

Image-guided radiation therapy (IGRT) involves two main procedures, performed in different rooms on different days: (1) treatment planning in the simulator room on the first day, and (2) radiotherapy in the linear accelerator room over multiple subsequent days. Both the simulator and the linear accelerator include CT imaging capabilities, which enables both treatment planning and reproducible patient setup, but does not provide good soft tissue contrast or allow monitoring of the target during treatment. We propose a cooperatively-controlled robot to reproducibly position an ultrasound (US) probe on the patient during simulation and treatment, thereby improving soft tissue visualization and allowing real-time monitoring of the target. A key goal of the robotic system is to produce consistent tissue deformations for both CT and US imaging, which simplifies registration of these two modalities. This paper presents the robotic system design and describes a novel control algorithm that employs virtual springs to implement guidance virtual fixtures during “hands on” cooperative control.

I. INTRODUCTION

Radiation therapy is commonly used as a treatment for cancer. The goal is to direct sufficient radiation to kill the tumor cells, without harming the healthy surrounding tissue. This treatment is usually fractionated; that is, the patient receives multiple radiation treatments, typically over several days. We consider image-guided radiation therapy (IGRT), which involves two main steps: (1) planning/simulation and (2) treatment delivery. Typically, planning includes the use of a *simulator*, which is a large-bore CT scanner, to provide the image for dose planning (possibly in conjunction with other images, such as MRI) and to guide the setup of the patient for the subsequent radiation treatments. The radiation treatments are performed with a linear accelerator (LINAC); modern LINACs

hsen1@jhu.edu, mledijubell@jhu.edu, pkaz@jhu.edu

*This work was supported by NIH R01 CA161613

include on-board cone beam CT (CBCT) imaging to directly show the setup of the patient in the treatment room frame of reference. Nevertheless, two major deficiencies have become apparent when CBCT is applied to verify radiotherapy: (1) CBCT only provides a “snapshot” of patient information at the time of imaging, but not during actual radiation delivery, and (2) CBCT often does not provide sufficient contrast to discriminate soft tissue targets. Ultrasound imaging can overcome these deficiencies, but the ionizing radiation precludes an ultrasonographer from holding the ultrasound probe on the patient during treatment. This has motivated the development of at least one telerobotic system for ultrasound monitoring during radiotherapy [8], [9]. Robotic systems for ultrasonography have been developed for other applications, as well [4], [1], [10]. Because ultrasound imaging requires contact between the probe and patient, all of these robotic systems include a force sensor for monitoring and/or controlling the contact force.

Our goal is to construct a robotically-controlled, integrated 3D x-ray and ultrasound imaging system to guide radiation treatment of soft-tissue targets. We are especially interested in registration between the ultrasound images and the CT images (from both the simulator and accelerator), because this enables the treatment plan and overall anatomy to be fused with the ultrasound image. We note, however, that ultrasound image acquisition requires relatively large contact forces between the probe and patient, which leads to tissue deformation. One approach is to apply a deformable (non-rigid) registration between the ultrasound and CT, but this is technically challenging. Our approach is to apply the same tissue deformation during CT image acquisition, thereby removing the need for a non-rigid registration method. We use a model (fake) ultrasound probe to avoid the CT image artifacts that would result from using a real probe. Thus, the requirement for our robotic system is to enable an expert ultrasonographer to place the probe during simulation, record the relevant information (e.g., position and force), and then allow a less experienced person to use the robot system to reproduce this placement (and tissue deformation) during the subsequent fractionated radiotherapy sessions. We do not attempt to do this autonomously, but rather employ a cooperative control strategy, where the robot shares control of the ultrasound probe with a human operator. One important task for the robot is to help the less experienced user reproduce the original setup. This paper presents the design of the robotic system, describes the novel cooperative control strategy that was developed, and demonstrates its performance with phantom experiments.

II. SYSTEM DESCRIPTION

The proposed setup is illustrated in Fig. 1, where the robotic unit is attached to the LINAC table via a bridge that can slide along the bed with the help of the rails attached on the sides. At its end-effector, the robot holds an ultrasound (US) probe that obtains real-time 3D US images of the patient’s target anatomy while the radiation beams are being emitted from the collimator of the gantry in the radiation therapy process.

As previously mentioned, the image-guided radiation therapy process consists of two major phases: (1) treatment planning in the simulator room, and (2) treatment delivery in the treatment (LINAC) room. We assume that the treatment room and simulator room each contain a camera system to simultaneously track the US probe, the robot base, and the

couch. In our setup, the tracker (camera) and US probe are provided by the Clarity System (Elekta AB, Stockholm, Sweden). In both rooms, the cameras are calibrated to the room isocenter and therefore provide a consistent reference coordinate system, within the accuracy of the calibration.

In the treatment planning phase, an ultrasonographer finds the anatomical target using the US probe attached at the tip of the robot and saves the 3D US image. From now on, this image will be referred as the “day-1 image” or the “reference image”. Then, the US probe is replaced by an x-ray compatible model probe and, with the assistance of the robot, is brought to exactly the same relative position with the target area as the real probe, to create the same local soft tissue deformation. After that, the couch is moved inside the CT scanner and the target area is scanned to get a CT image which will then be registered to the reference US image. Now, all information is available to plan the radiation therapy process. In the planning phase, the configuration of the robot should also be taken into account when determining the beam directions to avoid irradiating the robot or US probe.

The treatment phase can be repeated on multiple days depending on the therapy planning details. In the remaining parts of this paper, the treatment phase will be referred to as “day-2”. In this phase, again with robot assistance, the US probe or model probe is brought to the same relative position with the target organ as in day-1. As in the treatment planning phase, the model probe is used when a CBCT image is acquired, in order to reduce the metal artifact in the CT image. After the clinician confirms proper setup of the patient, based on the CBCT and 3D US images, the radiation beams are delivered to annihilate the malignant anatomical structure. Because the robot continues to hold the US probe on the patient, the clinician can monitor the target to ensure that proper setup is maintained (e.g., that the target has not significantly shifted from its initial position).

III. SYSTEM DESIGN

To carry out the robotic tasks mentioned in Section II, a 5 degree-of-freedom (dof) prototype robotic system is designed, Fig. 2. As can be seen, the robot consists of two motorized units connected through a passive arm whose configuration can be changed and locked manually. For the current experiments, the 3 dof linear stage shown in the CAD drawing (Fig. 1) was replaced by an existing linear stage, but this does not affect the performance of the control algorithms described in this paper. The 2 dof rotary stages are used to set the orientation of the US probe. Mechanically, this is not a remote center of motion (RCM) device, but RCM motion can be obtained in software by coordinating motion of the linear and rotary stages. In a closer view, the first rotary stage and the parallelogram mechanism which are responsible for accurate orientation motion can be seen in Fig. 3. In the design of the mechanism, extra care is taken to minimize the use of metals to avoid introducing significant artifacts in the CT images. A force sensor is installed between the US probe and the robot to enable the system to measure the forces applied on the probe. The use of a single force sensor minimizes the cost of the system, but makes it difficult to distinguish between forces applied by the user’s hand and by the patient’s tissue. This affects the manner in which the system is used, as described in Section IV.

The system can be considered as two fully-encoded mechanical subsystems connected by an unencoded passive arm. Ideally, the existing camera system would provide the homogeneous transformation matrix between the two ends of the passive arm. In these experiments, however, the camera system was not available and we therefore used the force sensor to obtain a coarse estimate of the 3×3 rotation component of the transformation matrix. To do this, we pushed along two perpendicular axes in the robot base frame and recorded the measured force vectors. By normalizing and orthogonalizing these force vectors, two columns of the rotation matrix corresponding to the pushed axes are estimated. The third column is obtained by taking the cross product of the first two column vectors.

The passive arm enables us to use more compact linear stages, which considerably decreases the weight of the robot attached to the bridge Fig. 1. The passive arm is mainly used to bring the US probe to the vicinity of the target area, and to provide the 6th degree of freedom (rotation about the probe axis). For the experiments in this paper, the passive arm is configured only at the beginning of each experiment and kept locked for the duration of that experiment.

IV. CONTROL SCHEMES

There are three basic procedures that must be performed with the robotic US system:

- On the treatment planning day, which we call day-1, the ultrasonographer finds the target with the US probe attached to the robotic system. This free space motion is provided through an unconstrained cooperative control process. Once the user places the probe on the tissue, he/she disables cooperative control and lets go of the ultrasound probe. At this point, the force sensor measures the reaction force due to the contact between the US probe and patient's tissue.
- Still on day-1, the user replaces the US probe with a model (fake) probe and reproduces the original probe position and force with the assistance of the robot. This requires a constrained cooperative control mode, where the constraints help the user to reproduce the original probe location. Details of this cooperative control method are provided below. Note that the user cannot view real-time US images during this procedure because a model probe is installed.
- On day-2, a different person uses the force and position information from day-1 to reproduce the original US imaging conditions. This requires a constrained cooperative control mode, as above, but in this case it is reasonable to place less confidence in the recorded probe position information due to: (1) differences between the treatment room and simulator room coordinate systems, (2) errors in patient setup, and (3) physiologic changes in the patient. We assume, however, that the recorded force information is still valid. This procedure may be performed with the US or model probe, depending on whether or not CBCT will be acquired. If the US probe is installed, the user can also compare the current US image to the day-1 US image.

The following sections describe the two basic control modes.

A. Unconstrained Cooperative Control

This mode is active when the user wants to move either the real or the model US probe freely. In this cooperative control paradigm [11], the clinician and robot share control of the US probe mounted on the robot's end-effector. The robot system senses the forces applied by the clinician and moves its end-effector with a velocity that is proportional to the applied force (i.e., an admittance control law). Mathematically, this can be expressed as:

$$\dot{q} = J^{-1}(q) \times G \times \begin{bmatrix} F_w \\ T_w \end{bmatrix} \quad (1)$$

where $\dot{q}$ is the goal velocity in the robot joint space, G is the diagonal matrix of admittance gains, J^{-1} is the inverse Jacobian matrix that transforms Cartesian velocities into joint velocities, and F_w and T_w are the measured forces and torques in the robot base frame. The Jacobian, and inverse Jacobian, matrices depend on the kinematic structure of the robot and its measured joint angles, q . To make the guiding operation easier for the user, the translation and rotation stages to be maneuvered can be enabled and disabled separately.

A nonlinear admittance gain (see Fig. 4) is used to enable fine control for small applied forces, while also allowing faster motions for higher applied forces, as proposed in [3]. This method also helps to reduce the effect of noise in the force sensor readings. The nonlinear gain is an experimentally-determined function of the measured force.

B. Cooperative Control with Virtual Spring Constraints

This controller is used to help the user reproduce the initial placement of the US probe. It is used in day-1 to place the model probe for CT imaging, with the goal of obtaining the same tissue deformation as previously obtained with the real US probe. It is also used in day-2 to reproducibly place the real or model probe prior to treatment.

On day-1, we define a goal frame whose origin is at the tip of the US probe, with its axes determined by the orientation of the US probe. The constrained controller should then help the user to move the US probe tip to the $x = y = 0$ position of the goal frame. Next, the controller should help the user to reproduce the correct orientation by aligning the probe with the goal frame. Finally, the controller should enter a mode that helps the user to apply the specified (previously recorded) force on the patient. Note that this sequence of actions mirrors the actions typically performed during image-guided needle interventions: (1) position the needle tip at the entry point, (2) align the needle axis with the insertion axis, and (3) insert the needle along its axis until it reaches the target. When the user switches from unconstrained cooperative control mode to guidance mode, the system remains in the unconstrained cooperative control mode until the US probe crosses the $x = 0$ or $y = 0$ plane. Once a plane is crossed, a virtual spring is created and attached between the plane and the probe tip, thereby pulling the probe toward the plane (i.e., a soft virtual fixture). When both planes have been crossed, the result is two orthogonal springs that pull the probe toward the line formed by the intersection of the two planes (i.e., the line that passes through the goal position at $x = y = 0$). These virtual springs are visualized in Fig. 5. The virtual spring force is proportional to the displacement on that particular axis.

When the linear springs are enabled, the user can switch to the rotation mode to get guidance around the x and y axes via the use of torsional springs. This procedure is similar to the linear springs, where the user must cross the goal orientation before a torsional spring is created and applied. It should be noted that the requirement to pass through the goal position or orientation before activating the spring prevents a discontinuity in the motion of the robot. For example, if a spring is enabled when the probe is far from the desired position, the controller would immediately apply a large force to move the probe toward the goal.

Between the treatment planning day and the treatment day, the existence of anatomical changes on the target organ (e.g., tumor shrinkage) or physical changes on the patient (e.g., gas in the abdominal area) raise the importance of the virtual spring approach. In such a case, the recorded goal position at day-1 will not create the same soft tissue deformation. However, it should be noted that these displacements are expected to be relatively small (i.e., on the order of 1-2 mm). With this approach, the clinician will be guided to the target area via springs with the flexibility to modify the final position and orientation.

When all springs are activated and the desired position and orientation on the x and y axes are set, which is ideally the same position and orientation that was found on day-1, the user can turn on the hybrid position/force control mode [7]. In this mode, the controller keeps the force component in the probe's z axis equal to the desired force from day-1, while maintaining the same position on all other degrees of freedom. With this motion the robot marches into the body along the US probe z -axis. It should be emphasized that force control along the direction normal to the patient's body is essential to compensate for patient motion during radiation therapy (e.g., breathing motion).

V. EXPERIMENTS

The experiment setup in Fig. 6 is used to evaluate the effectiveness of the system and the control approach. In the experiments, 2D B-Mode images are generated to simulate the treatment procedure. In order to more easily identify and analyze US images (and to visualize deformations), we created a 4×4 grid of markers inside the phantom. These markers were actually segments of pencil lead that were implanted using a transrectal ultrasound (TRUS) robot for prostate brachytherapy [2]. An example of a reference US image, representing the day-1 treatment planning image, can be observed in Fig. 7. The performance of the virtual spring control algorithm and force control along the probe direction are evaluated via two different experiments. First, a reference US image (Fig. 7) is obtained using the cooperative control mode; this represents the day-1 reference image. At this robot configuration, the position of the robot and the force applied on the US probe are recorded as the goal information.

A. Experiment 1

This experiment is performed to compare the virtual spring control approach with an unconstrained cooperative control approach (free hand mode) when the recorded goal position is not exact, which is the case for the day-2 positioning tasks. Essentially, this experiment checks whether the virtual fixture created by the virtual springs is “soft” enough to allow the user to compensate for relatively small errors using the US image feedback.

Here, an initial goal position (and force) is selected and a reference US image is acquired. The goal position is then randomly disturbed in the x-y plane by an amount between 3 mm and 5 mm on each axis, which emulates the relative motion of the target area with respect to robot base between the treatment planning day and the treatment day. The user is asked to place the US probe to obtain an US image that is as close to the reference image as possible with two methods: unconstrained cooperative control (free hand) and virtual spring control. In unconstrained cooperative control, the user does not get any feedback from the robot. In virtual spring control, however, the user will be guided to the incorrect (disturbed) goal and therefore the user must pull against the virtual springs to displace the probe until the correct image is obtained. This procedure is repeated a total of 9 times for each method, and the following information is recorded: (1) the time it takes the user to find the goal, and (2) the position error, which is the difference between the user's final position and the original goal position. As seen in Fig. 8 and Fig. 9, the virtual spring control shows an improvement in both time and accuracy compared to free hand control.

B. Experiment 2

To analyze the force control algorithm and to assess the repeatability of creating the same soft tissue deformation, we let the US probe go up and down to the target position on its own when all virtual springs are activated. All virtual springs are enabled to ensure that there will not be major displacements in the x-y plane. This procedure is repeated 6 times and the force applied to the probe by the phantom is recorded (Fig. 10). In this plot, "Experiment Number-0" corresponds to the force applied on the treatment planning day, so this is the force that was taken as a reference. It is seen that the controller is successful in applying pressure on the phantom by an amount close to the reference value.

After having satisfactory results on the force control, the next thing is to determine if the US probe created similar soft tissue deformations. For this, the centers of each marker in the 4×4 grid are located in each of the 6 target images, as shown in Fig. 11. The error in reproducing the deformation is defined as the distance between the center of each grid marker and its corresponding center in the reference image. The results show that the mean absolute error in the x and z directions is $0.9 \pm 0.5 [mm]$ and $0.3 \pm 0.3 [mm]$, respectively.

As an alternative method to measure the error in reproducing the distortion, the raw RF ultrasound data for each image is obtained and compared to the reference image. With this method, we obtain a mean normalized cross-correlation coefficient of 0.91 ± 0.01 , which indicates very good agreement (a coefficient of 1.0 would indicate perfect agreement).

VI. DISCUSSION AND CONCLUSIONS

We developed a robot system for integrating 3D ultrasound (US) imaging with x-ray (CT) guidance for external beam radiotherapy. Because ultrasound provides better soft tissue contrast, it allows the clinician to better visualize the treatment target and surrounding organs. Furthermore, US does not introduce additional radiation and can be used to continuously monitor the treatment. In particular, this enables the clinician to ensure that the target remains in the x-ray beam path.

Because US probe placement causes tissue deformation, it is desirable to create the same deformation when acquiring CT images. This simplifies the US-CT registration problem and enables the live ultrasound image to be viewed with respect to the treatment plan. But, it is not practical to acquire a CT image with the ultrasound probe in place due to the image artifact caused by the probe. We therefore must replace the real US probe by a model (fake) probe and use the robot to ensure that this model probe is placed in same position, and exerts the same force, as the real probe.

This repeated switching between real and model US probes must be performed reproducibly on the same day (intra-fraction) and on different days (inter-fraction). While one might reasonably assume that the patient has not moved on the same day, it is likely that the position of the patient will vary from day to day, even considering the clinical techniques currently used for patient setup. Thus, it is not advisable for the robot to autonomously return to the same position, or to impose a hard virtual fixture that forces the clinician to return the probe to that position in a cooperative control mode. We therefore created virtual springs to implement a soft virtual fixture that guides the clinician to the previously recorded position, but enables him/her to deviate from that position by exerting forces against the springs. In particular, the clinician can compensate for the placement error by comparing the live ultrasound image to the recorded reference image. We performed experiments with a phantom to demonstrate that our virtual fixtures are “soft” enough to enable the user to effectively compensate for these setup errors. We also verify that our method produces repeatable forces between the probe and phantom, and that our approach leads to reproducible deformations. Future work includes in-vivo experiments in an animal model.

References

- [1]. Abolmaesumi P, Salcudean SE, Zhu W-H, Sirouspour MR, DiMaio SP. Image-guided control of a robot for medical ultrasound. *IEEE Trans on Robotics and Automation*. 2002; 18(1):11–23.
- [2]. Fichtinger G, Fiene J, Kennedy C, Kronreif G, Iordachita I, Song D, Burdette E, Kazanzides P. Robotic assistance for ultrasound-guided prostate brachytherapy. *Medical Image Analysis*. Oct; 2008 12(5):535–545. [PubMed: 18650122]
- [3]. Kazanzides, P.; Zuhars, J.; Mittelstadt, B.; Taylor, R. Force sensing and control for a surgical robot. *IEEE Intl. Conf. on Robotics and Automation*; Nice, France: May. 1992 p. 612–617.
- [4]. Pierrot F, Dombre E, Degoulange E, Urbain L, Caron P, Garipey J, Louis Megnien J. Hippocrate: A safe robot arm for medical applications with force feedback. *Medical Image Analysis*. 1999; 3:285–300. [PubMed: 10710297]
- [5]. Pratt J, Chew C-M, Torres A, Dilworth P, Pratt G. Virtual model control: An intuitive approach for bipedal locomotion. *The International Journal of Robotics Research*. 2001; 20(2):129–143.
- [6]. Pratt, J.; Dilworth, P.; Pratt, G. Virtual model control of a bipedal walking robot. *Robotics and Automation*, 1997; Proceedings., 1997 IEEE International Conference on, volume 1; IEEE; 1997. p. 193–198.
- [7]. Raibert MH, Craig JJ. Hybrid position/force control of manipulators. *Trans. of the ASME*. Jun. 1981 102
- [8]. Schlosser J, Salisbury K, Hristov D. Telerobotic system concept for real-time soft-tissue imaging during radiotherapy beam delivery. *Medical Physics*. 2010; 37:6357–6367. [PubMed: 21302793]
- [9]. Schlosser J, Salisbury K, Hristov D. Tissue displacement monitoring for prostate and liver IGRT using a robotically controlled ultrasound system. *Medical Physics*. 2011; 38:3812.
- [10]. Vilchis A, Trocaz J, Cinquin P, Masuda K, Pellissier F. A new robot architecture for tele-echography. *IEEE Trans on Robotics and Automation*. 2003; 19(5):922–926.

- [11]. Xia T, Baird C, Jallo G, Hayes K, Nakajima N, Hata N, Kazanzides P. An integrated system for planning, navigation and robotic assistance for skull base surgery. *Intl J Medical Robotics and Computer Assisted Surgery*. 2008; 4(4):321–330.

Author Manuscript

Author Manuscript

Author Manuscript

Author Manuscript

Fig. 1.
CAD model presenting the overall concept of the robotically assisted image guided radiation therapy system

Fig. 2.
5 dof robotic system designed for IGRT

Fig. 3.
The mechanism for the rotary stages of the designed robot

Fig. 4.

The model for diagonal elements of the admittance matrix for force; a similar function is used for the torque admittance gain

Fig. 5. Guiding the clinician to the goal position and force using virtual springs, which are also used in [6] and [5].

Fig. 6.
Experiment setup

Fig. 7.
The reference B-Mode US image used in the experiments, showing the 4×4 grid of markers

Fig. 8.
Evaluation of virtual spring control algorithm

Fig. 9.
Comparison of unconstrained cooperative control (free hand) and virtual spring control

Fig. 10.
Steady-state force applied to the phantom

Fig. 11.
The centers of each 4×4 grid point superimposed on a single plot for 6 US images