Received:
22 December 2014

Revised:
25 March 2015

Accepted:
9 September 2015

doi: 10.1259/bjr.20140862

Cite this article as:

Lee J, Kim J-I, Ye S-J, Kim HJ, Carlson J, Park JM. Dosimetric effects of roll rotational setup errors on lung stereotactic ablative radiotherapy using volumetric modulated arc therapy. *Br J Radiol* 2015; **88**: 20140862.

FULL PAPER

Dosimetric effects of roll rotational setup errors on lung stereotactic ablative radiotherapy using volumetric modulated arc therapy

^{1,2}JAEGI LEE, BS, ^{2,3,4,5}JUNG-IN KIM, PhD, ^{1,3}SUNG-JOON YE, PhD, ^{3,4,6}HAK JAE KIM, MD, ^{1,2}JOEL CARLSON, BS and ^{2,3,4,5}JONG MIN PARK, PhD

¹Program in Biomedical Radiation Sciences, Department of Transdisciplinary Studies, Seoul National University Graduate School of Convergence Science and Technology, Seoul, Korea

²Biomedical Research Institute, Seoul National University College of Medicine, Seoul, Korea

³Department of Radiation Oncology, Seoul National University Hospital, Seoul, Korea

⁴Institute of Radiation Medicine, Seoul National University Medical Research Center, Seoul, Korea

⁵Center for Convergence Research on Robotics, Advanced Institutes of Convergence Technology, Suwon, Korea

⁶Department of Radiation Oncology, Seoul National University College of Medicine, Seoul, Korea

Address correspondence to: Dr Jong Min Park

E-mail: leodavinci@naver.com

Objective: To evaluate the dosimetric effects of roll-rotational setup errors of stereotactic ablative radiotherapy (SABR) for lung cancer using volumetric modulated arc therapy (VMAT).

Methods: A total of 23 lung SABR cases were evaluated retrospectively. Each of the planning CT images was intentionally rotated by $\pm 1^\circ$, $\pm 2^\circ$ and $\pm 3^\circ$. After that, to simulate the translational couch correction, rotated CT images were moved along the x, y and z axis to match the centroid of the target volume in the rotated CT images with that in the original CT images. The differences in $D_{95\%}$ and $V_{100\%}$ of the target volume, $D_{0.35cc}$ of spinal cord, $D_{0.35cc}$ and D_{5cc} of oesophagus and V_{20Gy} of lung between the original and the rotated CT images were calculated.

Results: The average differences in $D_{95\%}$ and $V_{100\%}$ of target volume, $D_{0.35cc}$ of spinal cord, $D_{0.35cc}$ and D_{5cc} of oesophagus and V_{20Gy} of lung were $-0.3\% \pm 0.4\%$ and $-0.7\% \pm 2.4\%$, 1.6 ± 27.9 cGy, -1.6 ± 37.6 cGy, 15.9 ± 25.3 cGy and $0.0\% \pm 0.1\%$, respectively. The dosimetric changes in organs at risk (OARs) near the target volume were sometimes considerable due to roll-rotational setup errors, despite the translational correction, and those were patient specific.

Conclusion: In the case of coplanar VMAT for lung SABR, dosimetric changes to the target volume due to roll-rotational setup errors could be compensated by translational correction, whereas those to the OARs could not in some cases.

Advances in knowledge: Roll-rotational setup errors would increase the dose to OARs despite the translational correction.

INTRODUCTION

Stereotactic ablative radiotherapy (SABR) has been shown to be effective in treating oligometastatic lesions to the lung, as well as early-stage non-small cell lung cancer.^{1–6} Since lung SABR delivers much a higher dose per fraction than conventional radiation therapy, accurate positioning of a patient is highly important. Therefore, the positioning of the patient is performed with image guidance and generally corrected by couch movements to match the treatment plan.^{7–10} However, most commercial couch systems support only translational movement and yaw-rotational movement. Although six degree-of-freedom couch systems are able to correct setup errors including both roll- and pitch-rotational errors automatically, it has not yet come into wide use in the clinic.¹¹ Few studies have investigated the dosimetric effects due to rotational setup errors for lung SABR.^{12–14} Shang

et al¹⁵ showed that rotational setup errors of $-0.11 \pm 0.82^\circ$ in the roll direction and $0.07 \pm 1.03^\circ$ in the pitch direction existed by analysing cone beam CT (CBCT) during lung SABR. Yang et al¹⁶ evaluated dosimetric effects due to rotational setup errors with three-dimensional conformal radiation therapy (3D-CRT) or intensity modulated radiation therapy (IMRT) techniques for medially located lung tumour. Galerani et al¹⁷ analysed dosimetric effects of rotational setup errors with CBCT for lung SABR. However, no study has yet demonstrated the dosimetric effects of rotational setup errors for lung SABR using volumetric modulated arc therapy (VMAT) technique.

VMAT can reduce normal tissue toxicity by delivering beams from numerous angles, as well as reduce the treatment time significantly. Reducing treatment time is

beneficial in terms of the reduction of intrafractional errors as well as supporting frail patients who cannot maintain their fixed positions during SABR.¹⁸ Therefore, VMAT has gained popularity and has recently been widely adopted in the clinic for lung SABR.^{19–24} The changes in dosimetric effects of 3D-CRT or IMRT due to rotational setup errors might be different from those of VMAT, since VMAT delivers beams from much more varied directions around the patient than do 3D-CRT or IMRT. Since the rotational setup errors during coplanar VMAT cause differences not only in penetration depths but also in the material being penetrated by each beam in an arc, the changes in dose distributions due to rotational setup errors are unclear. Therefore, we analysed the dosimetric effects of roll-rotational setup errors on lung SABR using coplanar VMAT technique by rotating planning CT images.

METHODS AND MATERIALS

Patient selection

After an institutional review board approval, a total of 23 patients who received SABR for lung cancer from May 2013

to September 2013 were selected retrospectively for this study. The mean and median ages were 69.5 and 71 years, respectively. Patient information is summarized in Table 1.

Volumetric modulated arc therapy plans for lung stereotactic ablative radiotherapy

Ten phase 4D-CT scanning (Real-time Position Management™ System; Varian Medical Systems, Palo Alto, CA, and Brilliance CT Big Bore; Philips Healthcare, Cleveland, OH) with 2-mm slice thickness was performed for each patient to define the internal target volume (ITV). Among ten phase 4D-CT images, 80% phase CT images were used for the planning of lung SABR according to the institutional protocol.²⁵ The clinical target volume (CTV) was defined to be the same as the ITV, and the planning target volume (PTV) was defined with isotropic margins ranging from 3 to 7 mm from the CTV. Doses of 48–60 Gy in four fractions were prescribed to the PTV except for one case (Patient 8), for whom the prescription dose was 45 Gy in five fractions because the tumour was located near the pericardium. Two coplanar partial arcs with a 6 MV flattening filter free beam

Table 1. Summary of patient information

Patient number	Sex	Age (years)	Stage	Fractions	Prescription (Gy)	Tumour position (Lobe)	Tumour volume (cm ³)	Tumour margin (cm)
1	Female	54	IV	4	54	RLL	26.6	0.5
2	Female	79	II	4	54	RML	1.3	0.5
3	Male	74	I	4	54	RUL	11.4	0.5
4	Male	70	I	4	54	RUL	5.6	0.7
5	Female	79	I	4	54	LLL	6.2	0.5
6	Male	54	IV	4	54	RML	2.0	0.7
7	Male	78	I	4	54	LUL	7.8	0.7
8	Male	75	IV	5	45	LUL	8.6	0.5
9	Female	49	IV	4	54	LUL	11.1	0.5
10	Male	90	II	4	54	RML	9.0	0.7
11	Male	67	II	4	60	LUL	25.1	0.3
12	Male	62	IV	4	54	LUL	7.4	0.5
13	Male	76	II	4	48	RML	12.3	0.5
14	Male	76	IV	4	60	RLL	12.9	0.3
15	Male	66	IV	4	48	LUL	2.1	0.7
16	Female	52	I	4	48	LUL	12.0	0.5
17	Male	63	I	4	48	LUL	2.8	0.7
18	Male	66	I	4	48	RLL	3.2	0.5
19	Male	71	IV	4	48	RLL	0.4	0.5
20	Male	72	I	4	54	RUL	7.5	0.5
21	Male	68	I	4	48	LUL	4.2	0.5
22	Male	77	I	4	54	RUL	19.4	0.5
23	Female	80	I	4	54	LUL	7.4	0.5

LLL, left lower lobe; LUL, left upper lobe; RLL, right lower lobe; RML, right middle lobe; RUL, right upper lobe.

Table 2. The dosimetric changes in target volume and organs at risk by roll-rotational setup errors

Structure	Dose-volumetric parameter	Couch translational correction	Dosimetric differences compared with the original plan					
			Clockwise rotation			Counter-clockwise rotation		
			-3°	-2°	-1°	1°	2°	3°
Target volume	PTV $D_{95\%}$ (%)	Before TC	-11.3 ± 11.3	-5.5 ± 7.2	-1.1 ± 1.4	-3.3 ± 4.0	-9.6 ± 10.3	-16.3 ± 14.7
		After TC	-0.2 ± 0.5	-0.2 ± 0.4	-0.2 ± 0.3	-0.3 ± 0.3	-0.4 ± 0.3	-0.4 ± 0.4
	CTV $D_{100\%}$ (%)	Before TC	-3.8 ± 6.8	-0.9 ± 2.3	-0.2 ± 0.8	-0.1 ± 0.5	-1.5 ± 3.9	-6.5 ± 10.7
		After TC	0.0 ± 0.4	0.0 ± 0.3	-0.1 ± 0.2	-0.1 ± 0.2	-0.2 ± 0.3	-0.3 ± 0.4
	PTV $V_{100\%}$ (%)	Before TC	-17.7 ± 17.8	-10.3 ± 12.9	-3.2 ± 4.1	-5.3 ± 5.1	-11.7 ± 10.3	-18.4 ± 15.1
		After TC	0.4 ± 2.7	-0.1 ± 1.7	-1.0 ± 1.4	-0.9 ± 1.4	-1.3 ± 1.9	-1.8 ± 3.6
	CTV $V_{100\%}$ (%)	Before TC	-5.5 ± 13.8	-2.5 ± 8.0	-0.7 ± 2.6	0.0 ± 0.4	-0.6 ± 2.7	-4.5 ± 13.1
		After TC	0.8 ± 3.5	0.5 ± 2.4	0.3 ± 1.6	-0.3 ± 1.5	-0.5 ± 2.4	-1.0 ± 4.5
	Conformity index	Before TC	-4.4 ± 15.2	-3.2 ± 10.6	-1.5 ± 4.3	-0.3 ± 2.7	-0.2 ± 5.0	-0.4 ± 7.9
		After TC	0.2 ± 3.6	0.2 ± 2.3	-0.2 ± 1.8	-1.0 ± 1.7	-1.5 ± 2.2	-2.2 ± 4.0
	Gradient measure (cm)	Before TC	3.4 ± 14.0	2.2 ± 8.6	0.8 ± 3.2	-0.3 ± 1.6	-0.5 ± 3.2	-0.4 ± 4.8
		After TC	-0.2 ± 1.2	-0.3 ± 0.9	0.0 ± 0.6	0.2 ± 0.6	0.4 ± 0.9	0.7 ± 1.5
OAR	Spinal cord max. (cGy)	Before TC	55.7 ± 114.5	37.4 ± 71.6	16.0 ± 40.7	-20.7 ± 45.4	-32.4 ± 76.0	-48.1 ± 105.3
		After TC	23.6 ± 59.3	16.7 ± 37.3	8.8 ± 16.0	-10.0 ± 17.2	-8.0 ± 30.4	-11.3 ± 41.3
	Spinal cord $D_{0.35cc}$ (cGy)	Before TC	40.8 ± 80.0	29.4 ± 55.8	12.9 ± 27.8	-13.7 ± 27.8	-26.0 ± 50.0	-34.5 ± 74.1
		After TC	20.6 ± 36.1	14.4 ± 23.1	5.7 ± 12.7	-6.4 ± 11.2	-10.7 ± 22.0	-14.1 ± 32.2
	Oesophagus max. (cGy)	Before TC	0.1 ± 48.2	-0.1 ± 35.8	-5.8 ± 27.8	-12.0 ± 37.7	-14.1 ± 44.7	-15.0 ± 52.4
		After TC	-14.3 ± 74.1	-13.0 ± 50.9	-11.4 ± 35.7	-2.7 ± 36.8	4.8 ± 43.6	6.1 ± 51.1
	Oesophagus $D_{0.35cc}$ (cGy)	Before TC	6.2 ± 29.7	5.7 ± 17.2	3.2 ± 11.8	-6.1 ± 21.2	-7.9 ± 23.8	-7.4 ± 30.2
		After TC	-19.1 ± 53.6	-13.2 ± 34.1	-5.1 ± 21.1	2.4 ± 22.0	9.1 ± 27.4	16.3 ± 42.6
	Oesophagus D_{5cc} (cGy)	Before TC	16.4 ± 20.1	16.2 ± 18.8	16.2 ± 20.1	15.2 ± 23.4	14.9 ± 25.7	12.4 ± 27.1
		After TC	10.3 ± 24.6	11.9 ± 22.8	14.9 ± 23.7	17.4 ± 22.3	20.1 ± 24.4	20.6 ± 30.8
	Heart max. (cGy)	Before TC	6.1 ± 206.9	10.3 ± 195.8	-11.2 ± 68.9	15.2 ± 99.2	17.0 ± 171.0	19.1 ± 240.6
		After TC	-2.6 ± 34.6	1.1 ± 69.8	-3.1 ± 33.1	16.7 ± 95.3	8.9 ± 86.9	19.8 ± 108.1
	Skin max. (cGy)	Before TC	4.7 ± 213.3	5.7 ± 160.9	11.2 ± 107.8	92.1 ± 281.6	127.5 ± 359.1	154.4 ± 408.9
		After TC	62.2 ± 104.8	45.4 ± 73.2	40.6 ± 73.2	45.1 ± 115.8	37.6 ± 57.3	42.7 ± 83.7
	Lung mean (cGy)	Before TC	-2.7 ± 16.1	-1.3 ± 10.8	-0.6 ± 5.7	0.3 ± 6.1	0.2 ± 11.4	0.1 ± 16.6
		After TC	-0.3 ± 3.4	-0.2 ± 2.7	0.1 ± 2.3	0.1 ± 2.1	0.2 ± 2.5	0.6 ± 3.0

(Continued)

Table 2. (Continued)

Structure	Dose–volumetric parameter	Couch translational correction	Dosimetric differences compared with the original plan					
			Clockwise rotation			Counter-clockwise rotation		
			−3°	−2°	−1°	1°	2°	3°
	Lung $V_{20\text{Gy}}$ (%)	Before TC	0.0 ± 0.5	0.0 ± 0.3	0.0 ± 0.2	0.0 ± 0.2	0.0 ± 0.3	0.0 ± 0.5
		After TC	0.0 ± 0.1	0.0 ± 0.1	0.0 ± 0.1	0.0 ± 0.1	0.0 ± 0.1	0.0 ± 0.1

CTV $D_{100\%}$, the dose received at least 100% volume of the clinical target volume; CTV $V_{100\%}$, the volume received 100% of prescription dose on the clinical target volume; Lung $V_{20\text{Gy}}$, the volume received 20 Gy on the ipsilateral lung; Max., the maximum dose; OAR, organs at risk; PTV $D_{95\%}$, the dose received at least 95% volume of the planning target volume; PTV $V_{100\%}$, the volume received 100% of prescription dose on the planning target volume; TC, translational correction.

of Truebeam STxTM (Varian Medical Systems) were used to generate VMAT plans using the EclipseTM system (Varian Medical Systems). The isocenters of every plan were located inside the PTV. The angles of partial arcs were determined to reduce dose to the contralateral lung while considering the tumour location for each patient. Each of the VMAT plans were optimized with the progressive resolution optimizer (PRO3, v. 10; Varian Medical Systems) and dose distributions were calculated with the anisotropic analytic algorithm (AAA, v. 10; Varian Medical Systems) with a calculation grid of 1 mm.

Simulation of roll-rotational errors

To simulate the roll-rotational setup errors, we rotated the planning CT images in the roll direction using an in-house program written in MATLAB[®] v. R2012a (MathWorks[®] Inc., Natick, MA). Roll-rotating planning CT images were chosen instead of using CBCT images acquired before treatment in order to eliminate disturbance factors on dose distributions, such as differences in CT numbers between the planning CT and the CBCT, patient anatomy changes and artefacts in CBCT. The rotation angles were set as $\pm 1^\circ$, $\pm 2^\circ$ and $\pm 3^\circ$ with the rotation axis at the centre of the seventh thoracic vertebra.²⁶ Positive angles indicate counter-clockwise rotation. The structures in the original CT images were registered to the rotated CT images using MIM MaestroTM (MIM SoftwareTM Inc., Cleveland, OH). Since the changes in volumes and Hounsfield unit (HU) of each structure should be minimal after registration, the volumes of each structure in the original CT images were compared with those in the rotated CT images. Dose distributions of the rotated CT images were calculated with a SABR plan identical to the plan in the original CT images. After that, to simulate the real situation in the clinic, we moved the rotated CT images along x, y and z axis to match the centroid of the PTV in the original CT images to that in the rotated CT images. This was a simulation of the translational couch correction, and the dose distributions were calculated.

Evaluation of dosimetric changes due to roll-rotations

For the PTVs, the minimum dose received by at least 95% volume of the target volume ($D_{95\%}$) and the target volume receiving 100% of the prescription dose ($V_{100\%}$) were calculated for the original CT images as well as for every rotated CT image before and after translational correction. For the CTVs, $D_{100\%}$ and $V_{100\%}$ were acquired in the same way. The values of conformity index (CI)²⁷ and gradient measure (GM)²⁸ were defined as follows and calculated.

$$\text{Conformity Index (CI)} = \frac{V_{100\%} \text{ of PTV}}{\text{Volume of PTV}} \quad (1)$$

$$\text{Gradient Measure (GM)} = r_{V_{50\%}} - r_{V_{100\%}} \quad (2)$$

where $r_{V_{n\%}}$ is the equivalent sphere radius of the $n\%$ of the prescription isodose. The equivalent sphere is a sphere with the same volume as one inside a given isodose surface. For organs at risk (OARs), the maximum dose to the spinal cord, oesophagus, heart and skin, as well as mean dose to the lung were calculated

Figure 1. Changes in the dose received at least 95% volume ($D_{95\%}$) of planning target volume (PTV) (a), the volume received by 100% of prescription dose ($V_{100\%}$) of PTV (c), $D_{100\%}$ of clinical target volume (CTV) (e), CTV $V_{100\%}$ (g), conformity index (CI) (i) and gradient measure (GM) (k) by roll-rotational setup errors before translational corrections are shown. After translational corrections, changes in PTV $D_{95\%}$ (b), PTV $V_{100\%}$ (d), CTV $D_{100\%}$ (f), CTV $V_{100\%}$ (h), CI (j) and GM (l) are also shown.

and evaluated based on the National Comprehensive Cancer Network Clinical Practice Guideline in Oncology (NCCN Guidelines[®]; National Comprehensive Cancer Network, Fort Washington, PA) for non-small cell lung cancer (http://www.nccn.org/professionals/physician_gls/pdf/nscl.pdf). The values of $D_{0.35cc}$ of spinal cord, $D_{0.35cc}$ and D_{5cc} of oesophagus and V_{20Gy} of lung were also evaluated. The differences in dose–volumetric parameters between the original and the rotated CT images before and after translational correction were calculated.

For the target volume and each OAR, the correlations between the magnitudes of dose–volumetric parameter differences and the distances from the rotation axis to the centroid of the PTV were investigated. In addition, the correlations between the magnitudes of dose–volumetric parameter differences and the distance from the centroid of the PTV to the centroid of each OAR were also investigated. Pearson correlation coefficients were calculated for the correlation analysis.

RESULTS

Differences in structure volume between the original and the rotated CT images

The average differences in the volume and HU per volume of PTV, CTV, spinal cord, oesophagus, heart, skin and lung between the original and rotated CT images were 0.4% and 0.5%, respectively. After a manual review, we concluded that these differences came from the

quantization effect due to the voxelized grids of the structures, and the dosimetric changes due to these differences were minimal.

Dosimetric changes in target volume

The dosimetric changes in the target volumes with roll-rotational setup errors before and after translational corrections are shown in Table 2 and Figure 1. Generally, the more the degree of rotation increased, the more the dosimetric changes increased. The magnitudes of dosimetric changes due to roll-rotations were reduced by the translational corrections. Before translational correction, the average values for changes in $D_{95\%}$ and $V_{100\%}$ of PTV, $D_{100\%}$ and $V_{100\%}$ of CTV, CI and GM due to roll-rotations were $-7.9\% \pm 10.6\%$, $-11.1\% \pm 13.3\%$, $-2.2\% \pm 6.0\%$, $-2.3\% \pm 8.8\%$, $-1.7\% \pm 8.9\%$ and $0.9\% \pm 7.4\%$, respectively. One case with a small target volume and large distance from the target volume to rotation axis showed more than 60% change in $D_{95\%}$ of the PTV at 3° rotation.

After translational correction, those values were reduced to $-0.3\% \pm 0.4\%$, $-0.7\% \pm 2.4\%$, $-0.1\% \pm 0.3\%$, $0.0\% \pm 2.9\%$, $-0.7\% \pm 2.9\%$ and $0.1\% \pm 1.1\%$, respectively. No changes larger than 1.5% in PTV $D_{95\%}$ and 3.3% in PTV $V_{100\%}$, except in one case (Patient 8, -18.4%), were observed in the rotated CT images. In the case of CTV, no changes larger than 1.3% in $D_{100\%}$ and 0.8% in $V_{100\%}$, except in one case (Patient 8, -21.9%), were observed. The differences in CI and GM were

below 5.0% and 2.2%, respectively, except in one case (Patient 8, -18.7% for CI and 7.2% for GM at 3°).

No correlation with statistical significance was observed between the magnitudes of the differences in any of the dose–volumetric parameters and the distances from the rotation axis to the centroid of the PTV.

Dosimetric changes in organs at risk

The changes in the maximum dose to the spinal cord, oesophagus, heart, skin and the mean dose to the lung as well as $D_{0.35cc}$

of spinal cord, $D_{0.35cc}$ and D_{5cc} of oesophagus and V_{20Gy} of lung by the roll-rotations are shown in Table 2 and Figures 2–4. Similar to the results for the target volumes, the more the degree of rotation increased, the more the dosimetric changes increased in general.

The average differences in the maximum dose to the spinal cord, oesophagus, heart and skin and the mean dose to the lung due to roll-rotations before translational corrections were 1.3 ± 88.8 cGy, -7.8 ± 42.4 cGy, 9.4 ± 174.9 cGy, 66.0 ± 283.2 cGy and -0.7 ± 12.0 cGy, respectively. Those values after translational

Figure 2. Changes by roll-rotational setup errors in the maximum dose to the spinal cord (a), oesophagus (c), heart (e) and skin (g) before translational corrections are shown. After translational corrections, changes in the maximum dose to the spinal cord (b), oesophagus (d), heart (f) and skin (h) are also shown.

Figure 3. Changes by roll-rotational setup errors in the mean dose to the ipsilateral lung (a) and lung V_{20Gy} (c) before translational corrections are shown. After translational correction, changes in the mean dose to the ipsilateral lung (b) and lung V_{20Gy} (d) are also shown.

corrections were 3.3 ± 39.2 cGy, -5.1 ± 51.1 cGy, 6.8 ± 77.4 cGy, 45.6 ± 87.3 cGy and 0.1 ± 2.7 cGy, respectively. The average differences in $D_{0.35cc}$ of spinal cord before and after translational correction were 1.5 ± 63.2 cGy and 1.6 ± 27.9 cGy, respectively (Figure 4). The average differences in $D_{0.35cc}$ and D_{5cc} of oesophagus before translational correction were -1.1 ± 24.2 cGy and 15.2 ± 22.9 cGy, respectively (Figure 4). After translational correction, those were -1.6 ± 37.6 cGy and 15.9 ± 25.3 cGy, respectively (Figure 4). The average differences in V_{20Gy} of lung before and after translational correction were $-0.0\% \pm 0.4\%$ and $0.0\% \pm 0.1\%$, respectively (Figure 3). The case of Patient 18 showed a spinal cord dose difference of 540.2 cGy at -3° rotation before translation correction, and this was reduced into -91.2 cGy by translational correction (Figure 5). In the case of Patient 10, the maximum dose to the spinal cord before translational correction was <32.7 cGy; however, it was increased by 173.4 cGy due to rotation by -3° after translational correction (Figure 5). In the case of the oesophagus, no distinctive tendency of reduction of differences by translational correction was observed. Patient 1 showed a difference in maximum dose to the oesophagus of <21.4 cGy at 3° rotation before translational corrections, however, that difference was increased by 120.5 cGy after translational correction (Figure 5). For the heart, two cases (Patients 6 and 9) showed more than 800 cGy differences before translational correction (Figure 5). Those were reduced to 384.4 and 387.4 cGy after translational correction. In these cases, the locations of the target volumes were close to the heart such that the maximum doses to the heart were already 3643.3 cGy for Patient 6 and 3915.7 cGy for Patient 9 in the original CT images. Even after translational corrections, the maximum doses to the heart were 4027.7 cGy for Patient 6 at 2° rotation and 4303.1 cGy for Patient 9 at 3° rotation. For the skin, Patient 3 showed a maximum skin dose

increase of 1800 cGy due to rotation since the target volume was located nearby the skin (Figure 5). After translational correction, the dose difference was reduced by less than 550 cGy. The differences in mean lung dose were minimal due to roll-rotations.

Correlation analysis

Before translational correction, Pearson correlation coefficients between the difference in PTV $D_{95\%}$, PTV $V_{100\%}$, CTV $D_{100\%}$ and CTV $V_{100\%}$ and the distances from the rotation axis to the centroid of the PTV were 0.095, 0.077, -0.136 and -0.149 , respectively. After translational correction, those values were -0.129 , -0.047 , 0.014 and 0.207, respectively. In the cases of OARs, correlation coefficients of the differences in the maximum dose to spinal cord, oesophagus, heart and skin, to the distances from the centroid of the PTV to each OARs were 0.076, 0.252, -0.361 and 0.105, respectively, before translational correction. After translational correction, those values were -0.076 , 0.059, -0.340 and 0.095, respectively. No correlations were observed, always showing correlation coefficients <0.4 .

DISCUSSION

Reviewing the results of the target volumes, the dosimetric differences were large before translational corrections due to the differences induced in the location between the target volume and the volume of the prescription isodose by roll-rotations especially when the target volume was located far from the rotation axis. However, both average differences and their standard deviations were significantly reduced by translational corrections, since the target volumes of lung SABR and their corresponding prescription isodose volumes were generally round in shape. The exceptional case (Patient 8) had an irregularly shaped prescription isodose owing to the close proximity of the aorta to the target

Figure 4. Changes by roll-rotational setup errors in the spinal cord $D_{0.35cc}$ (a) and oesophagus $D_{0.35cc}$ (c) and D_{5cc} (e) before translational corrections are shown. After translational correction, changes in the spinal cord (b) and oesophagus $D_{0.35cc}$ (d) and D_{5cc} (f) are also shown.

volume (Figure 5). Therefore, if the target volume and the corresponding volume of the prescription isodose are round in shape, as in the case of lung SABR, the dosimetric effects on the target volume due to roll-rotational setup errors could be minimized to clinically acceptable levels by NCCN Guidelines for non-small cell lung cancer with translational corrections. The changes in radiological penetrating depths due to roll-rotations of up to 3° during VMAT SABR did not significantly contribute to the changes in dose–volumetric parameters of the target volume. In accordance with the previous study,¹⁶ no correlations were observed between the changes in dose–volumetric parameters and the distances from the rotation axis to the centroid of the target volume before and after translational corrections. Such no correlations might be caused by the large variation in target volumes (from 0.4 to 26.6 cm³) in this study. Even though the locations of the target volumes are the same, if the volumes are different, the coverage by roll-rotation would be different, resulting in no correlations between the dosimetric change and the distances from the rotation axis to the centroid of the PTV.

In the case of OARs, both average differences and their standard deviations in dose–volumetric parameters induced by

roll-rotations were generally reduced by translational corrections. The reductions of the standard deviations were much higher than those of the average values of differences by translational corrections. However, in the few cases where the OARs were located nearby the target volume, the magnitudes of dosimetric changes after translational corrections were still large, thus we cannot guarantee the dose to the OARs when there were roll-rotational setup errors and these were compensated with translational corrections. This might be due to the irregular shape of isodoses in each VMAT plan due to intensity modulation during the optimization procedure. Those irregular shapes were patient specific and various such that some exceptional cases were observed. In addition, since the dose gradients were very steep in the region near the target volume, the dosimetric changes due to roll-rotational setup errors were more sensitive when the OARs were nearby the target volume. In this case, OARs should be more carefully considered than those located far from the target volume since they are expected to receive relatively high doses. Therefore, for OARs near the target volume, roll-rotations should be carefully corrected instead of using simple translational corrections.

Figure 5. The cases when the target volumes are located nearby the organs at risk are shown. Colour wash level is 20–110% of the prescription dose.

As in the previous study,¹⁶ no correlations were observed between the dosimetric changes and the distances between the rotation axis to the centroid of the PTV, nor were there correlations between the distance between the centroid of the PTV and the centroid of the OAR before and after translational corrections. This might be due to the irregular shape of the isodose line in the same context as previously mentioned. We

could not predict the dosimetric changes in OARs due to roll-rotations with certain distances since the isodose irregularities are patient specific.

The results of this study assumed consistent roll-rotations during whole fractions of treatment, which is not necessarily realistic. In reality, the roll-rotational setup errors could have

variations at each fraction, thus those could be averaged out through entire course resulting in much smaller dosimetric changes than those presented in this study. A previous study¹⁵ has demonstrated that the net averages of roll-rotations in lung SABR were close to zero. Although the situations in this study are unlikely, it still has a value as we showed that the changes in penetrating depth and materials due to roll-rotations for VMAT have minimal effects on dosimetric parameters of the target volume up to 3° rotations, as well as that the OARs nearby the target volume need careful considerations for roll-rotations. By rotating CT images to eliminate disturbance factors, we could analyse the changes in dose–volumetric parameters purely due to roll-rotations. The limitation of this study was only the roll-rotational setup errors were simulated. Further study including pitch-rotational setup errors will be performed as a future work.

The dosimetric changes in OARs for lung SABR VMAT could be considerable due to small roll-rotational setup errors up to 3°, despite translational corrections. According to the NCCN Guidelines suggesting the tolerance doses for each OAR during lung SABR, the maximum doses to spinal cord, heart and skin should be less than 26, 34 and 36 Gy, respectively (four fraction

SABR). This study showed that larger doses than the tolerance level by NCCN Guidelines could be delivered to OARs due to roll-rotational setup errors in some cases, which might result in complications. One spinal cord case, two heart cases and two skin cases from a total of 23 cases violated the tolerance dose. A full correction with a six degree-of-freedom couch system is preferred. If it is not available, careful consideration during patient setup or margins for OARs accounting for roll-rotations is needed for lung SABR with VMAT technique.

CONCLUSION

Roll-rotational setup errors had a minimal influence on the coverage of the target volume during lung SABR with coplanar VMAT technique when translational corrections were applied, with the exception of a VMAT plan with an irregularly shaped prescription isodose volume. However, the dosimetric changes in OARs nearby the target volume were sometimes considerable due to roll-rotational setup errors, despite the translational correction, and those were patient specific. Therefore, the roll-rotational setup errors should be carefully corrected for lung SABR with coplanar VMAT technique to avoid normal tissue complications.

REFERENCES

1. Fakiris AJ, McGarry RC, Yiannoutsos CT, Papiez L, Williams M, Henderson MA, et al. Stereotactic body radiation therapy for early-stage non-small-cell lung carcinoma: four-year results of a prospective phase II study. *Int J Radiat Oncol Biol Phys* 2009; **75**: 677–82. doi: [10.1016/j.ijrobp.2008.11.042](https://doi.org/10.1016/j.ijrobp.2008.11.042)
2. Onishi H, Shirato H, Nagata Y, Hiraoka M, Fujino M, Gomi K, et al. Stereotactic body radiotherapy (SBRT) for operable stage I non-small-cell lung cancer: can SBRT be comparable to surgery?. *Int J Radiat Oncol Biol Phys* 2011; **81**: 1352–8. doi: [10.1016/j.ijrobp.2009.07.1751](https://doi.org/10.1016/j.ijrobp.2009.07.1751)
3. Timmerman R, Paulus R, Galvin J, Michalski J, Straube W, Bradley J, et al. Stereotactic body radiation therapy for inoperable early stage lung cancer. *JAMA* 2010; **303**: 1070–6. doi: [10.1001/jama.2010.261](https://doi.org/10.1001/jama.2010.261)
4. Rusthoven KE, Kavanagh BD, Burri SH, Chen C, Cardenes H, Chidel MA, et al. Multi-institutional phase I/II trial of stereotactic body radiation therapy for lung metastases. *J Clin Oncol* 2009; **27**: 1579–84. doi: [10.1200/JCO.2008.19.6386](https://doi.org/10.1200/JCO.2008.19.6386)
5. Wulf J, Haedinger U, Oppitz U, Thiele W, Mueller G, Flentje M. Stereotactic radiotherapy for primary lung cancer and pulmonary metastases: a noninvasive treatment approach in medically inoperable patients. *Int J Radiat Oncol Biol Phys* 2004; **60**: 186–96. doi: [10.1016/j.ijrobp.2004.02.060](https://doi.org/10.1016/j.ijrobp.2004.02.060)
6. Xia T, Li H, Sun Q, Wang Y, Fan N, Yu Y, et al. Promising clinical outcome of stereotactic body radiation therapy for patients with inoperable stage I/II non-small-cell lung cancer. *Int J Radiat Oncol Biol Phys* 2006; **66**: 117–25. doi: [10.1016/j.ijrobp.2006.04.013](https://doi.org/10.1016/j.ijrobp.2006.04.013)
7. Grills IS, Hugo G, Kestin LL, Galerani AP, Chao KK, Wloch J, et al. Image-guided radiotherapy via daily online cone-beam CT substantially reduces margin requirements for stereotactic lung radiotherapy. *Int J Radiat Oncol Biol Phys* 2008; **70**: 1045–56. doi: [10.1016/j.ijrobp.2007.07.2352](https://doi.org/10.1016/j.ijrobp.2007.07.2352)
8. Bissonnette JP, Purdie TG, Higgins JA, Li W, Bezjak A. Cone-beam computed tomographic image guidance for lung cancer radiation therapy. *Int J Radiat Oncol Biol Phys* 2009; **73**: 927–34. doi: [10.1016/j.ijrobp.2008.08.059](https://doi.org/10.1016/j.ijrobp.2008.08.059)
9. Benedict SH, Yenice KM, Followill D, Galvin JM, Hinson W, Kavanagh B, et al. Stereotactic body radiation therapy: the report of AAPM Task Group 101. *Med Phys* 2010; **37**: 4078–101. doi: [10.1118/1.3438081](https://doi.org/10.1118/1.3438081)
10. Li W, Purdie TG, Taremi M, Fung S, Brade A, Cho BCJ, et al. Effect of immobilization and performance status on intrafraction motion for stereotactic lung radiotherapy: analysis of 133 patients. *Int J Radiat Oncol Biol Phys* 2011; **81**: 1568–75. doi: [10.1016/j.ijrobp.2010.09.035](https://doi.org/10.1016/j.ijrobp.2010.09.035)
11. Guckenberger M, Meyer J, Wilbert J, Baier K, Sauer O, Flentje M. Precision of image-guided radiotherapy (IGRT) in six degrees of freedom and limitations in clinical practice. *Strahlenther Onkol* 2007; **183**: 307–13. doi: [10.1007/s00066-007-1695-0](https://doi.org/10.1007/s00066-007-1695-0)
12. Josipovic M, Persson GF, Logadottir Á, Smulders B, Westmann G, Bangsgaard JP. Translational and rotational intra- and inter-fractional errors in patient and target position during a short course of frameless stereotactic body radiotherapy. *Acta Oncol* 2012; **51**: 610–17. doi: [10.3109/0284186X.2011.626448](https://doi.org/10.3109/0284186X.2011.626448)
13. de Boer HCJ, van Sörnsen de Koste JR, Senan S, Visser AG, Heijmen BJM. Analysis and reduction of 3D systematic and random setup errors during the simulation and treatment of lung cancer patients with CT-based external beam radiotherapy dose planning. *Int J Radiat Oncol Biol Phys* 2001; **49**: 857–68. doi: [10.1016/S0360-3016\(00\)01413-9](https://doi.org/10.1016/S0360-3016(00)01413-9)
14. Guckenberger M, Meyer J, Vordermark D, Baier K, Wilbert J, Flentje M. Magnitude and clinical relevance of translational and rotational patient setup errors: a cone-beam CT study. *Int J Radiat Oncol Biol Phys* 2006; **65**: 934–42. doi: [10.1016/j.ijrobp.2006.02.019](https://doi.org/10.1016/j.ijrobp.2006.02.019)
15. Shang C, Williams T, Kasper M. SU-E-J-135: clinical efficacy of using kV-CBCT imaging guided 6D robotic couch in lung SBRT. *Med Phys* 2011; **38**: 3473. doi: [10.1118/1.3611902](https://doi.org/10.1118/1.3611902)
16. Yang Y, Catalano S, Kelsey CR, Yoo DS, Yin FF, Cai J. Dosimetric effects of rotational

- offsets in stereotactic body radiation therapy (SBRT) for lung cancer. *Med Dosim* 2014; **39**: 117–21. doi: [10.1016/j.meddos.2013.11.002](https://doi.org/10.1016/j.meddos.2013.11.002)
17. Galerani AP, Grills I, Hugo G, Kestin L, Mohammed N, Chao KK, et al. Dosimetric impact of online correction via cone-beam CT-based image guidance for stereotactic lung radiotherapy. *Int J Radiat Oncol Biol Phys* 2010; **78**: 1571–8. doi: [10.1016/j.ijrobp.2010.02.012](https://doi.org/10.1016/j.ijrobp.2010.02.012)
 18. Zhang GG, Ku L, Dilling TJ, Stevens CW, Zhang RR, Li W, et al. Volumetric modulated arc planning for lung stereotactic body radiotherapy using conventional and unflattened photon beams: a dosimetric comparison with 3D technique. *Radiat Oncol* 2011; **6**: 152–7. doi: [10.1186/1748-717X-6-152](https://doi.org/10.1186/1748-717X-6-152)
 19. Bedford JL, Nordmark Hansen V, McNair HA, Aitken AH, Brock JEC, Warrington AP, et al. Treatment of lung cancer using volumetric modulated arc therapy and image guidance: a case study. *Acta Oncol* 2008; **47**: 1438–43. doi: [10.1080/02841860802282778](https://doi.org/10.1080/02841860802282778)
 20. Verbakel WF, Senan S, Cuijpers JP, Slotman BJ, Lagerwaard FJ. Rapid delivery of stereotactic radiotherapy for peripheral lung tumors using volumetric intensity-modulated arcs. *Radiother Oncol* 2009; **93**: 122–4. doi: [10.1016/j.radonc.2009.05.020](https://doi.org/10.1016/j.radonc.2009.05.020)
 21. Ong CL, Verbakel WF, Cuijpers JP, Slotman BJ, Lagerwaard FJ, Senan S. Stereotactic radiotherapy for peripheral lung tumors: a comparison of volumetric modulated arc therapy with 3 other delivery techniques. *Radiother Oncol* 2010; **97**: 437–42. doi: [10.1016/j.radonc.2010.09.027](https://doi.org/10.1016/j.radonc.2010.09.027)
 22. McGrath SD, Matuszak MM, Yan D, Kestin LL, Martinez AA, Grills IS. Volumetric modulated arc therapy for delivery of hypofractionated stereotactic lung radiotherapy: A dosimetric and treatment efficiency analysis. *Radiother Oncol* 2010; **95**: 153–7. doi: [10.1016/j.radonc.2009.12.039](https://doi.org/10.1016/j.radonc.2009.12.039)
 23. Holt A, van Vliet-Vroegindeweij C, Mans A, Belderbos JS, Damen EMF. Volumetric-modulated arc therapy for stereotactic body radiotherapy of lung tumors: A comparison with intensity-modulated radiotherapy techniques. *Int J Radiat Oncol Biol Phys* 2011; **81**: 1560–7. doi: [10.1016/j.ijrobp.2010.09.014](https://doi.org/10.1016/j.ijrobp.2010.09.014)
 24. Yea JW. Comparison of the dose distributions with beam arrangements in the stereotactic body radiotherapy (SBRT) for primary lung cancer. *Prog Med Phys* 2014; **25**: 110–15. doi: [10.14316/pmp.2014.25.2.110](https://doi.org/10.14316/pmp.2014.25.2.110)
 25. Yeo SG, Kim ES. Efficient approach for determining four-dimensional computed tomography-based internal target volume in stereotactic radiotherapy of lung cancer. *Radiat Oncol J* 2013; **31**: 247–51. doi: [10.3857/roj.2013.31.4.247](https://doi.org/10.3857/roj.2013.31.4.247)
 26. Purdie TG, Bissonnette JP, Franks K, Bezjak A, Payne D, Sie F, et al. Cone-beam computed tomography for on-line image guidance of lung stereotactic radiotherapy: localization, verification, and intrafraction tumor position. *Int J Radiat Oncol Biol Phys* 2007; **68**: 243–52. doi: [10.1016/j.ijrobp.2006.12.022](https://doi.org/10.1016/j.ijrobp.2006.12.022)
 27. Shaw E, Kline R, Gillin M, Souhami L, Hirschfeld A, Dinapoli R, et al. Radiation Therapy Oncology Group: radiosurgery quality assurance guidelines. *Int J Radiat Oncol Biol Phys* 1993; **27**: 1231–9. doi: [10.1016/0360-3016\(93\)90548-A](https://doi.org/10.1016/0360-3016(93)90548-A)
 28. Varian Medical Systems, Inc. *Treatment planning for external beam, Eclipse reference guide*. Palo Alto, CA: Varian Medical Systems, Inc.; 2010.