Sentinel lymph node biopsy under fluorescent indocyanine green guidance: Initial experience

Fatih Aydoğan¹, Akif Enes Arıkan², Erman Aytaç², Mehmet Velidedeoğlu², Mehmet Halit Yılmaz³, Muhammet Sait Sager⁴, Varol Çelik¹, Cihan Uras¹

ABSTRACT

Objective: Sentinel lymph node biopsy can be applied by using either blue dye or radionuclide method or both in breast cancer. Fluorescent imaging with indocyanine green is a new defined method. This study evaluates the applicability of sentinel lymph node biopsy via fluorescent indocyanine green.

Material and Methods: IC-VIEW (Pulsion Medical Systems AG, Munich, Germany) infrared visualization system was used for imaging. Two mL of indocyanine green was injected to visualize sentinel lymph nodes. After injection, subcutaneous lymphatics were traced and sentinel lymph nodes were found with simultaneous imaging. Sentinel lymph nodes were excised under fluorescent light guidance, and excised lymph nodes were examined histopathologically. Patients with sentinel lymph node metastases underwent axillary dissection.

Results: Four patients with sentinel lymph node biopsy due to breast cancer were included in the study. Sentinel lymph nodes were visualized with indocyanine green in all patients. The median number of excised sentinel lymph node was 2 (2-3). Two patients with lymph node metastasis underwent axillary dissection. No metastasis was detected in lymph nodes other than the sentinel nodes in patients with axillary dissection. There was no complication during and after the operation related to the method.

Conclusion: According to our limited experience, sentinel lymph node biopsy under fluorescent indocyanine green guidance, which has an advantage of simultaneous visualization, is technically feasible.

Keywords: Fluorescent light, indocyanine green, sentinel lymph node biopsy

INTRODUCTION

Sentinel lymph node biopsy (SLNB) is an effective method of determining the need for axillary dissection in the surgical treatment of early stage breast cancer (1-4). The quality of life can be improved by preventing axillary lymph node dissection (ALND) related morbidity with SLNB (5). Radiocolloid and blue dye are the currently used methods for sentinel lymph node (SLN) detection. The two techniques have their own advantages and disadvantages. The radiocolloid method, which has a better SLN detection rate as compared to blue dye, requires collaboration with nuclear medicine and availability of a gamma probe as well as some measures for radiation exposure. Although blue dye does not have such issues, it may lead to skin discoloration, decrease in saturation, allergies, anaphylactic reactions, injection site nodules and fat necrosis (6-10). The SLN detection rates are lower in SLNB with blue dye alone as compared to radiocolloid use (11-13). Indocyanine green (ICG) is a non-toxic and non-ionized substance that not extravasates by binding to lipoproteins, has a short life-time has a fluorescence in near infrared (NIR) wavelengths, and enables tracing the lymphatic flow through the tissue with charged-couple device (CCD) (14).

This study aimed to evaluate the feasibility of fluorescent light displayed ICG guided SLNB, which was introduced as an alternative to the two current SLNB techniques.

MATERIAL AND METHODS

Four patients who had been operated for early stage breast cancer with clinically (physical examination and radiology) negative axilla in Istanbul University, Cerrahpaşa Medical Faculty, Department of General Surgery, Breast Diseases Clinic between August 25th 2010 and September 14th 2010 were included in the study. There were no pathologic lymph nodes detected in the preoperative examination and imaging of patients.

Simultaneous radionuclide material was applied in two, and blue dye was applied in the remaining two patients.

An infrared imaging system (IC-VIEW, Pulsion Medical Systems AG, Munich, Germany) was used to visualize indocyanine green fluorescence. Following induction of general anesthesia and proper surgical field preparation, 2 cc of indocyanine green (ICG-Pulsion, Pulsion Medical Systems, Munich, Germany)

¹Department of General Surgery, Breast Diseases Service, Istanbul University, Cerrahpaşa Medical Faculty Istanbul, Turkey

²Department of General Surgery, Istanbul University, Cerrahpaşa Medical Faculty Istanbul, Turkey

³Department of Radiology, Istanbul University, Cerrahpaşa Medical Faculty Istanbul, Turkey

⁴Department of Nuclear Medicine, Istanbul University, Cerrahpaşa Medical Faculty Istanbul, Turkey

Address for Correspondence
Akif Enes Arıkan
e-mail: enesarikan@yahoo.com

Received: 24.07.2014

Accepted: 26.12.2014

Available Online Date: 18.08.2015

©Copyright 2016
by Turkish Surgical Association
Available online at
www.uluscerrahidergisi.org

Figure 1. Subareolar injection of indocyanine green and follow-up of lymphatic drainage by infrared camera

Figure 2. Evaluation of the excised lymph node by infrared camera in terms of indocyanine green uptake

Figure 3. Infra-red camera view of the excised lymph node with indocyanine green uptake

was injected with subareolar or peritumoral injections, and subcutaneous lymphatics were traced by simultaneous imaging to detect the SLNs (Figure 1, 2). The axilla was accessed through an incision performed just below the axillary hairline in two patients with ICG fluorescence in the axilla, and through the lateral side of a Stewart incision in the other two (due to their not being suitable for breast-conserving surgery). In

addition to the lymph nodes visualized by ICG, the blue dye stained lymph nodes and those showing radioactivity were also excised. The excised lymph nodes were confirmed by infrared imaging before being sent to histopathology (Figure 3). The excised SLNs were sent for intraoperative histopathologic examination. Axillary lymph node dissection was performed in two cases that had positive intraoperative results.

Istanbul University, Cerrahpaşa Medical Faculty, Clinical Researchs Ethical Review Committee approved this study. Informed written consent was obtained from the patients.

RESULTS

The median age of four patients included in the study was 67.50 (51-72) years. The median number of excised lymph nodes was 2.5 (2-3) for ICG, 0.5 (0-1) for blue dye, and 1.5 (1-2) for radiocolloid. All lymph nodes detected by radiocolloid has been identified by ICG while one of the lymph nodes stained with blue dye has not been detected by ICG. Patient information is presented in Table 1, and SLNB results are depicted in Table 2. Preoperative lymphoscintigraphy of patients who underwent SLNB with radioisotope revealed one focal uptake in the left axilla. Patients in whom visualization with ICG was sustained and who had a positive intraoperative frozen-section evaluation of the SLNB underwent ALND. In two patients who underwent to axillary dissection, metastasis was not detected in the nodes except the SLN (28 and 16 lymph nodes). The paraffin examination of the SLNs without evidence of metastases on intraoperative study, did not detect any additional metastases.

During the four-year median follow-up of patients, there were no loco/regional recurrences. None of the patients had allergic reactions, and no complications related to the method developed during or after surgery.

DISCUSSION

Giuliano et al. (15) have first defined SLNB in breast cancer. Sentinel lymph node biopsy is an independent prognostic factor for axillary lymph node staging in early stage breast cancer, and has become the standard of care (4, 16-19). Currently, blue dye and radiocolloid techniques are being used for SLNB. Using radiocolloid alone gives more accurate results than blue dye and its application by the surgeon is easy. Nevertheless, it is invisible it exposes the patients and health workers to radiation, although at a low amount and it requires collaboration with nuclear medicine unit (20, 21). Its use is mainly limited by the availability of a nuclear medicine unit in the hospital.

In the blue dye method, these disadvantages related to radiocolloid are not encountered, it is a method that can be visualized with bare eyes but requires the use of high-density blue dye. Adverse events such as skin discoloration, allergic reactions, fat necrosis and nodule formation in the injection site can be observed (6, 8). Nevertheless, any side effect related to either radiocolloid substance or blue dye (with the exception of skin discoloration) use was not observed in our study group.

In order to reduce the rate of false negative results in SLNB, methods that are more precise are required. To this end, recently, ICG that can be displayed simultaneously and has a good lymphatic drainage has been used for SLNB in breast cancer. Similar to this study, Murawa et al. (22) reported no

Table 1. Patient characteristics

Patient number	Age	Lesion location	Lesion diameter (mm)	Type of surgery	AD	AD positive LN number (SLN excluded)	AD LN number (SLN excluded)	Adjuvant therapy
1	67	Left, UOQ	15	Partial mastectomy	Yes	0	16	CT, RT
2	72	Left, IMQ	20	Partial mastectomy	No	-	-	CT, RT
3	51	Left, UOQ [‡]	15	MRM	Yes	0	28 (0)	CT, HT, RT
4	68	Right, UMQ	40	Mastectomy	No	-	-	CT, HT

[‡]multifocal. UOQ: upper outer quadrant; IMQ: inner middle quadrant; UMQ: upper middle quadrant; HT: hormonotherapy; CT: chemotherapy; RT: radiotherapy; MRM: modified radical mastectomy; AD: axillary dissection; SLN: sentinel lymph node

Table 2. Details of sentinel lymph node biopsy

Patient number	Number of SLN	Second technique	Number of LNs detected by ST	Number of LNs detected by ICG	Number of LNs detected by both techniques	SLN intraoperative evaluation (LN number; technique)	SLN pathologic evaluation method
1	3	M	0	3	0	Positive (1, ICG)	Imprint
2	2	R	2	2	1	Negative	Imprint
3	3	R	1	3	1	Positive (1, R+ICG*)	Imprint, Frozen
4	3	M	1	2	0	Negative	Imprint

*The positive lymph node is the node detected by both radionuclide and indocyanine green. AD: axillary dissection; ST: second technique; LN: lymph node; ICG: indocyanine green; B: blue dye; R: radionuclide; SLN: sentinel lymph node

toxic or allergic effects of ICG. Indocyanine green, that can be visualized simultaneously and has a good lymphatic drainage, can provide similar or even better results to the currently routine used radiocolloid and blue dye in combination (23). Sugie et al. (24) reported the SLN detection rate as 78% with use of blue dye as compared to the 99% with ICG use. In the same study, difference of the median number of SLNs detected by the two methods was as one more in the ICG method than blue dye method. Similarly, the number of SLNs detected by ICG in this study was higher than the blue dye method. Jung et al. (25) compared radiocolloid with combined methods (radiocolloid, blue dye and ICG) in terms of identifying SLN and yielded the number of SLNs was higher in the combined method (combined method: 3.4 ± 1.37 , radiocolloid: 2.3 ± 1.04 , $p < 0.001$) similar to this study.

However, the requirement of a near infrared camera for the ICG method and its being detected up to a depth of 21 mm are its most important disadvantages (26). This disadvantage is particularly problematic in overweight patients. Kitai et al. (27) defined an axillary compression technique to address this issue where a cone-shaped transparent plastic device was compressed against the chest wall. The patients in this study have not encountered such a problem. Another disadvantage of ICG is that the surgical lights prevents visualization of ICG's fluorescence and to prevent this use of low-spectrum light sources like laser or interference filters as excitation light source has been suggested (14). In this study, room lights in the operating room were turned off to prevent this problem and the SLNs were visualized transcutaneously by ICG fluorescence.

The limited number of cases constitutes a limitation to this study but still is acceptable for an initial experience.

CONCLUSION

According to the literature and our limited experience, ICG-guided SLN biopsy that has an advantage of simultaneous visualization is considered technically feasible. New studies on this method could yield promising results for the future.

Ethics Committee Approval: Ethics committee approval was received for this study from the ethics committee of İstanbul University Cerrahpaşa Medical Faculty

Informed Consent: Written informed consent was obtained from patients who participated in this study.

Peer-review: Externally peer-reviewed.

Author Contributions: Concept - F.A.; Design - V.C.; Supervision - C.U.; Funding - C.U.; Materials - F.A., E.A.; Data Collection and/or Processing - A.E.A., M.H.Y., M.S.S.; Analysis and/or Interpretation - E.A.; Literature Review - F.A.; Writer - A.E.A., F.A.; Critical Review - F.A., V.C., A.E.A.; Other - M.V.

Conflict of Interest: No conflict of interest was declared by the authors.

Financial Disclosure: The authors declared that this study has received no financial support.

REFERENCES

- Veronesi U, Paganelli G, Viale G, Luini A, Zurrada S, Galimberti V, et al. A randomized comparison of sentinel-node biopsy with routine axillary dissection in breast cancer. *N Engl J Med* 2003; 349: 546-553. [\[CrossRef\]](#)
- Krag DN, Anderson SJ, Julian TB, Brown AM, Harlow SP, Costantino JP, et al. Sentinel-lymph-node resection compared with conventional axillary-lymph-node dissection in clinically node-negative patients with breast cancer: overall survival findings from the NSABP B-32 randomised phase 3 trial. *Lancet Oncol* 2010; 11: 927-933. [\[CrossRef\]](#)
- Mansel RE, Fallowfield L, Kissin M, Goyal A, Newcombe RG, Dixon

- JM, et al. Randomized multicenter trial of sentinel node biopsy versus standard axillary treatment in operable breast cancer: the ALMANAC Trial. *J Natl Cancer Inst* 2006; 98: 599-609. [\[CrossRef\]](#)
4. Zavagno G, De Salvo GL, Scalco G, Bozza F, Barutta L, Del Bianco P, et al. A randomized clinical trial on sentinel lymph node biopsy versus axillary lymph node dissection in breast cancer: results of the Sentinella/GIVOM trial. *Ann Surg* 2008; 247: 207-213. [\[CrossRef\]](#)
 5. Fleissig A, Fallowfield LJ, Langridge CI, Johnson L, Newcombe RG, Dixon JM, et al. Post-operative arm morbidity and quality of life. Results of the ALMANAC randomised trial comparing sentinel node biopsy with standard axillary treatment in the management of patients with early breast cancer. *Breast Cancer Res Treat* 2006; 95: 279-293. [\[CrossRef\]](#)
 6. Aydoğan F, Celik V, Uras C, Salihoglu Z, Topuz U. A comparison of the adverse reactions associated with isosulfan blue versus methylene blue dye in sentinel lymph node biopsy for breast cancer. *Am J Surg* 2008; 195: 277-278. [\[CrossRef\]](#)
 7. Aydoğan F, Salihoglu Z, Uras C, Karabıcak I, Celik V, Cercel A, et al. Intraparenchymal methylene blue injection for sentinel lymph node biopsy in breast cancer patients does not interfere with the pulse oximetry readings. *J Breast Cancer* 2009; 12: 32-35. [\[CrossRef\]](#)
 8. Parvaiz MA, Isgar B. Anaphylaxis and blue urticaria associated with Patent Blue V injection. *Anaesthesia* 2012; 67: 1275-1276. [\[CrossRef\]](#)
 9. Rutgers EJ. Guidelines to assure quality in breast cancer surgery. *Eur J Surg Oncol* 2005; 31: 568-576. [\[CrossRef\]](#)
 10. Somashekhar SP, Zaveri Shabbir S, Udupa Venkatesh K, Venkatachala K, Parameshwaran R, Vasan Thirumalai MM. Sentinel lymph node biopsy in early breast cancer using methylene blue dye and radioactive sulphur colloid - a single institution Indian experience. *Indian J Surg* 2008; 70: 111-119. [\[CrossRef\]](#)
 11. Hayashida T, Jinno H, Sakata M, Takahashi M, Onishi T, Seki H, et al. Superiority of radioisotope over blue dye for sentinel lymph node detection in breast cancer. *Eur Surg Res* 2010; 44: 111-116. [\[CrossRef\]](#)
 12. Meyer-Rochow GY, Martin RC, Harman CR. Sentinel node biopsy in breast cancer: validation study and comparison of blue dye alone with triple modality localization. *ANZ J Surg* 2003; 73: 815-818. [\[CrossRef\]](#)
 13. Radovanovic Z, Golubovic A, Plzak A, Stojiljkovic B, Radovanovic D. Blue dye versus combined blue dye-radioactive tracer technique in detection of sentinel lymph node in breast cancer. *Eur J Surg Oncol* 2004; 30: 913-917. [\[CrossRef\]](#)
 14. Alander JT, Kaartinen I, Laakso A, Patila T, Spillmann T, Tuchin VV, et al. A review of indocyanine green fluorescent imaging in surgery. *Int J Biomed Imaging* 2012; 2012: 940585. [\[CrossRef\]](#)
 15. Giuliano AE, Kirgan DM, Guenther JM, Morton DL. Lymphatic mapping and sentinel lymphadenectomy for breast cancer. *Ann Surg* 1994; 220: 391-401. [\[CrossRef\]](#)
 16. Bonnema J, van de Velde CJ. Sentinel lymph node biopsy in breast cancer. *Ann Oncol* 2002; 13: 1531-1537. [\[CrossRef\]](#)
 17. Cox CE, Pendas S, Cox JM, Joseph E, Shons AR, Yeatman T, et al. Guidelines for sentinel node biopsy and lymphatic mapping of patients with breast cancer. *Ann Surg* 1998; 227: 645-53. [\[CrossRef\]](#)
 18. Carter CL, Allen C, Henson DE. Relation of tumor size, lymph node status, and survival in 24,740 breast cancer cases. *Cancer* 1989; 63: 181-187. [\[CrossRef\]](#)
 19. Layeequr Rahman R, Siegel E, Boneti C, Ingram M, Kepple J, Henry-Tillman RS, et al. Stage migration with sentinel node biopsy in breast cancer. *Am J Surg* 2009; 197: 491-496. [\[CrossRef\]](#)
 20. Upponi SS, McIntosh SA, Wishart GC, Balan KK, Purushotham AD. Sentinel lymph node biopsy in breast cancer—is lymphoscintigraphy really necessary? *Eur J Surg Oncol* 2002; 28: 479-480. [\[CrossRef\]](#)
 21. Dauphine CE, Khalkhali I, Vargas MP, Isaac NM, Haukoos J, Vargas HI. Intraoperative injection of technetium-99m sulfur colloid is effective in the detection of sentinel lymph nodes in breast cancer. *Am J Surg* 2006; 192: 423-426. [\[CrossRef\]](#)
 22. Murawa D, Hirche C, Dresel S, Hunerbein M. Sentinel lymph node biopsy in breast cancer guided by indocyanine green fluorescence. *Br J Surg* 2009; 96: 1289-1294. [\[CrossRef\]](#)
 23. Kitai T, Inomoto T, Miwa M, Shikayama T. Fluorescence navigation with indocyanine green for detecting sentinel lymph nodes in breast cancer. *Breast Cancer* 2005; 12: 211-215. [\[CrossRef\]](#)
 24. Sugie T, Sawada T, Tagaya N, Kinoshita T, Yamagami K, Suwa H, et al. Comparison of the indocyanine green fluorescence and blue dye methods in detection of sentinel lymph nodes in early-stage breast cancer. *Ann Surg Oncol* 2013; 20: 2213-2218. [\[CrossRef\]](#)
 25. Jung SY, Kim SK, Kim SW, Kwon Y, Lee ES, Kang HS, et al. Comparison of sentinel lymph node biopsy guided by the multimodal method of indocyanine green fluorescence, radioisotope, and blue dye versus the radioisotope method in breast cancer: a randomized controlled trial. *Ann Surg Oncol* 2014; 21: 1254-1259. [\[CrossRef\]](#)
 26. Pleijhuis RG, Langhout GC, Helfrich W, Themelis G, Sarantopoulos A, Crane LM, et al. Near-infrared fluorescence (NIRF) imaging in breast-conserving surgery: assessing intraoperative techniques in tissue-simulating breast phantoms. *Eur J Surg Oncol* 2011; 37: 32-39. [\[CrossRef\]](#)
 27. Kitai T, Kawashima M. Transcutaneous detection and direct approach to the sentinel node using axillary compression technique in ICG fluorescence-navigated sentinel node biopsy for breast cancer. *Breast Cancer* 2012; 19: 343-348. [\[CrossRef\]](#)