Published in final edited form as:

Bio Protoc. 2016 April 5; 6(7): .

Extraction and Quantification of Tryptophan and Kynurenine from Cultured Cells and Media Using a High Performance Liquid Chromatography (HPLC) System Equipped with an Ultra-Sensitive Diode Array Detector

Jeffrey Kim¹, Benjamin Stewart², and Robert H Weiss^{1,3,4,*}

¹Division of Nephrology, Department of Internal Medicine, University of California, Davis, USA

²Lawrence Livermore National Laboratory, Livermore, USA

³Cancer Center, University of California, Davis, USA

⁴Medical Service, Sacramento VA Medical Center, Sacramento, USA

Abstract

Evidence of the involvement of tryptophan and its metabolite, kynurenine, in various biological processes including cancer is constantly expanding. Analysis of cell extracts and culture media can allow for quick snapshots of the metabolic fluctuations occurring *in vitro*. Here, we describe a method for metabolite extraction from mammalian cells and analysis of extracted metabolites and cell culture media by HPLC with detection using an ultra-sensitive diode array detector.

Materials and Reagents

1. Disposable 2 ml plastic microcentrifuge tubes
2. 1.5 ml glass vials (Agilent Technologies, catalog number: 5182-0715), screw top caps (Agilent Technologies, catalog number: 5182-0717) and 200 µl glass inserts (Agilent Technologies, catalog number: 5182-0549)
3. AcroVac Filter Unit (Pall Life Sciences, catalog number: AVFP02L)
4. Chloroform (HPLC grade) (Thermo Fisher Scientific, catalog number: C607)
5. L-Tryptophan (Sigma-Aldrich, catalog number: T0254)
6. L-Kynurenine (Sigma-Aldrich, catalog number: K8625)
7. 3-Nitro-L-tyrosine (nitrotyrosine) (Sigma-Aldrich, catalog number: N7389)
8. Methanol (HPLC grade) (Thermo Fisher Scientific, catalog number: A452)
9. Acetonitrile (HPLC grade) (Thermo Fisher Scientific, catalog number: A998)

*For correspondence: ; Email: rhweiss@ucdavis.edu.

10. Potassium phosphate monobasic anhydrous (KH_2PO_4) (Thermo Fisher Scientific, catalog number: P285-500)
11. Potassium phosphate dibasic anhydrous (K_2HPO_4) (Thermo Fisher Scientific, catalog number: P288-500)
12. PIPES (Thermo Fisher Scientific, catalog number: BP292450)
13. EDTA (Thermo Fisher Scientific, catalog number: S311-500)
14. Kynurenic acid (Sigma-Aldrich, catalog number: K3375)
15. Quinolate (Sigma-Aldrich, catalog number: P63204)
16. Phosphate buffered saline (PBS)
17. 50% methanol /50% 3 mM PIPES-3 mM EDTA (pH 7.4) (see Recipes)
18. Potassium phosphate (pH 6.4) with 2.7% (v/v) acetonitrile (see Recipes)
19. Standard mix (see Recipes)

Equipment

1. Refrigerated tabletop centrifuge
2. HPLC system with binary pump, autosampler and UV diode array detector (DAD) (Agilent Technologies, model: 1100)
3. Hypersil GOLD C18 column (50 mm with 2.1 mm diameter) (Thermo Fisher Scientific, catalog number: 25003-052130)

Procedure

A. Tryptophan and kynurenine extraction

1. The protocol requires a minimum of 1×10^6 cells to accurately detect tryptophan and kynurenine levels. If tryptophan and kynurenine quantification is desired from fresh media, collect 2 ml of fresh media prior to cell addition for subsequent analysis and freeze at -80°C . This sample will be used to measure the baseline levels of tryptophan and kynurenine in the media.
2. If cells of interest are adherent cells, then proceed to step A2a. If cells of interest are suspension cells, then proceed to step A2b. If desired, the cell culture can be spiked with an internal standard such as 3-nitro-L-tyrosine (a final concentration of 20–50 μM is generally appropriate) at this time. The internal standard can be used to quantify metabolite recovery.
 - a. At the completion of the treatment schedule, collect 2 ml media in 2 ml microcentrifuge tubes. Centrifuge at $2,000 \times g$ at 4°C in centrifuge to pellet cells/debris from the media. Without disturbing the pellet, transfer media into two labeled tubes of equal volume. One will be used for immediate analysis while the other will serve as a backup. At this point, cell-free, centrifuged culture media can be stored at -80°C until analysis. Media can

maintained at 30 °C. Use an isocratic mobile phase of 15 mM potassium phosphate (pH 6.4), with 2.7% (v/v) acetonitrile with a flow rate of 0.8 ml/min. The run time should be set for 10 min.

2. Quantify the area under the peaks that display absorbance at 286 nm and 360 nm to calculate the concentration of tryptophan and kynurenine, respectively, in the samples. Generate standard curves ranging from 0 to 200 μ M for both tryptophan and kynurenine and other metabolites of interest as well as any internal standard (if applicable) using authentic standards. Standards of each concentration should be prepared in triplicate and measured in triplicate. A recipe for a standard mix is included in the Recipes section.

Note: Details of standard curve generation will vary depending on the chromatography system and data analysis software used. Consult the manufacturer's documentation to generate a standard curve with your instrument's software.

Representative data

Figure 1. HPLC chromatogram showing separation of tryptophan, kynurenine, kynurenic acid and nitrotyrosine standards

Quinolate is also present in the standard mix but below the limit of quantification. The upper panel shows absorbance at 360 nm, and the lower panel shows absorbance at 286 nm. The tryptophan metabolite kynurenic acid and 3-nitro-L-tyrosine (nitrotyrosine), which can be used as an internal standard, are also separated. All analytes are present at a concentration of 12.5 μ M. The standard solution was prepared as described in the Recipes section.

Notes

This method can be modified for compatibility with liquid chromatography-mass spectrometry by using a mobile phase consisting of 10 mM ammonium formate (pH 6.4),

containing 2.7% acetonitrile. Other chromatographic conditions remain unchanged. If desired, a known quantity of 3-nitro-L-tyrosine can be added to samples prior to extraction for use as an internal standard.

Recipes

1. 50% methanol / 50% 3 mM PIPES-3 mM EDTA (pH 7.4)
 - a. Prepare 3 mM PIPES-3 mM EDTA (pH 7.4) (for 500 ml).
Mix 0.453 g PIPES and 0.438 g EDTA in 400 ml of sterile water (HPLC grade). After the solution becomes completely dissolved, complete to 500 ml with sterile water. Filter the solution.
 - b. Mix the above with 500 ml of Methanol (HPLC grade).
 - c. Degas and store at 4 °C.
2. 15 mM potassium phosphate, pH 6.4, with 2.7% (v/v) acetonitrile (for 1 L)
Make 1 M stock solutions of both monobasic anhydrous (KH_2PO_4) and dibasic anhydrous (K_2HPO_4).
 - a. Mix 72.2 ml of KH_2PO_4 and 27.8 ml of K_2HPO_4 dilute with 900 ml of sterile water (HPLC grade) to make 100 mM stock.
 - b. Take 150 ml of the 100 mM stock and dilute with 850 ml of sterile water (HPLC grade) to make 15 mM potassium phosphate stock. Filter the solution.
 - c. Add appropriate amount of acetonitrile (HPLC grade) fresh when running HPLC separation.
 - d. Degas and store at room temperature.
3. Standard mix (200 μM each of tryptophan, nitrotyrosine, kynurenine, kynurenic acid, and quinolinate)
 - a. Prepare each standard separately at a concentration of 1 mM in 50 ml mobile phase [15 mM potassium phosphate (pH 6.4) with 2.7% acetonitrile]. For each solution, weigh out an appropriate quantity of each standard and place in a 50 ml volumetric flask. Record the actual mass of each standard.
 - b. To each volumetric flask, add 50 ml mobile phase and mix. The standards, especially kynurenic acid, may require more than an hour to go into solution. A stir bar can be added to each flask and stirred on a stir plate to facilitate dissolution of the standard.
 - c. When each solution is ready, add 2 ml of each standard 1 mM standard solution to a 15 ml tube and vortex mix. The final concentration of each standard in this standard mix will be 200 μM . Perform replicate serial dilutions as required to make a series of standards for use in generation of a standard curve.

Note: Serial dilutions of standards are generally satisfactory but can propagate concentration errors if the stock solution concentration is not correct.

Acknowledgments

This work was supported by NIH grants 1R01CA135401-01A1, and 1R01DK082690-01A1, the Medical Service of the US Department of Veterans' Affairs, and Dialysis Clinics, Inc. (DCI) (all to R. H. W.). This work was performed in part at the Research Resource for Biomedical Accelerator Mass Spectrometry, which is operated at LLNL under the auspices of the U. S. Department of Energy under contract DEAC52-07NA27344. The Research Resource is supported by the National Institutes of Health, National Institute of General Medical Sciences under Grant P40 RR13461. The extraction protocol was adapted from Villas-Boas *et al.* (2005). The HPLC method was adapted from Laich *et al.* (2002).

References

1. Laich A, Neurauter G, Widner B, Fuchs D. More rapid method for simultaneous measurement of tryptophan and kynurenine by HPLC. *Clin Chem.* 2002; 48(3):579–581. [PubMed: 11861457]
2. Villas-Boas SG, Hojer-Pedersen J, Akesson M, Smedsgaard J, Nielsen J. Global metabolite analysis of yeast: evaluation of sample preparation methods. *Yeast.* 2005; 22(14):1155–1169. [PubMed: 16240456]
3. Wettersten HI, Hakimi AA, Morin D, Bianchi C, Johnstone ME, Donohoe DR, Trott JF, Aboud OA, Stirdivant S, Neri B, Wolfert R, Stewart B, Perego R, Hsieh JJ, Weiss RH. Grade-dependent metabolic reprogramming in kidney cancer revealed by combined proteomics and metabolomics analysis. *Cancer Res.* 2015; 75(12):2541–2552. [PubMed: 25952651]