Denoising of Ultrasound Cervix Image Using Improved Anisotropic Diffusion Filter

R Jemila Rose¹, S Allwin²

ABSTRACT

Objective: The purpose of this study was to evaluate an improved oriented speckle reducing anisotropic diffusion (IADF) filter that suppress the speckle noise from ultrasound B-mode images and shows better result than previous filters such as anisotropic diffusion, wavelet denoising and local statistics.

Methods: The clinical ultrasound images of the cervix were obtained by ATL HDI 5000 ultrasound machine from the Regional Cancer Centre, Medical College campus, Thiruvananthapuram. The standardized ways of organizing and storing the image were in the format of bmp and the dimensions of 256×256 with the help of an improved oriented speckle reducing anisotropic diffusion filter. For analysis, 24 ultrasound cervix images were tested and the performance measured.

Results: This provides quality metrics in the case of maximum peak signal-to-noise ratio (PSNR) of 31 dB, structural similarity index map (SSIM) of 0.88 and edge preservation accuracy of 88%.

Conclusion: The IADF filter is the optimal method and it is capable of strong speckle suppression with less computational complexity.

Keywords: Anisotropic diffusion filter, denoising, speckle noise, ultrasound image

Eliminación del Ruido de la Imagen Ecográfica del Cuello Uterino Usando un Filtro de Difusión Anisotrópica Mejorado

R Jemila Rose¹, S Allwin²

RESUMEN

Objetivo: El propósito de este estudio fue evaluar un filtro de difusión anisotrópica mejorado para la reducción del speckle (DARS) orientado. El filtro elimina el speckle (moteado o ruido granular) de las imágenes ecográficas en modo B, y muestra mejores resultados que los filtros previos, tales como la difusión anisotrópica, la transformada de ondeletas (wavelets) y las estadísticas locales.

Métodos: Las imágenes de ecografía clínica del cuello uterino fueron obtenidas mediante el equipo de ultrasonido ATL HDI 5000 del Centro Regional de Cáncer, Universidad Médica, Thiruvananthapuram. Las formas estándares de organizar y almacenar la imagen fueron el formato bmp y las dimensiones de 256×256 con la ayuda de un filtro de difusión anisotrópica mejorado para la reducción del speckle orientado. Para el análisis, se sometieron a prueba 24 imágenes de ultrasonido del cuello uterino, y se midieron los resultados.

Resultados: Esto proporciona métricas de calidad en el caso de una relación señal a ruido de pico (RSRP) máxima de 31 dB, un mapa de índice de similitud estructural (SSIM) de 0.88, y una precisión de la preservación del borde de 88%.

Conclusión: El filtro DARS constituye el método óptimo, y es capaz de suprimir un fuerte ruido speckle con menos complejidad computacional.

Palabras claves: Filtro de difusión anisotrópica, eliminación del ruido, ruido speckle, imagen de ultrasonido

West Indian Med J 2015; 64 (4): 376

From: ¹Department of Information Technology, St Xavier's Catholic College of Engineering, Chunkankadai, Nagercoil – 629 003, Kanyakumari District, India and ²Department of Computer Science and Engineering, Infant Jesus College of Engineering, Vallanadu, Tamil Nadu 628 851, Thoothukudi District, India.

Correspondence: Ms R Jemila Rose, Department of Information Technology, St Xavier's Catholic College of Engineering, Chunkankadai, Nagercoil – 629003, Kanyakumari District, India. E-mail: jemila.rose@gmail.com

INTRODUCTION

Cervical cancer is the second most common type of cancer among women, with more than 250 000 deaths every year (1). The mortality rate of cervical cancer in India is more than the mortality rate of child birth [0.8 per cent vs 0.6 per cent] (2). Moreover, it is estimated that 70 to 80% of the total female population in India is found to be affected by cervical cancer (3, 4). In South India, it is expected that the death rate will increase further and it is believed that carcinoma of the uterine cervix is the major cause (5). Cervical cancer can be cured, if detected in the early stages. In India, magnetic resonance imaging (MRI) scan is preferred for identification of cervical cancer, however, the cost of the system is high; hence, identification through ultrasound can reduce the cost. But ultrasound images are highly affected by artefacts. So by introducing improved anisotropic diffusion filter denoising approach, the quality of the image can be improved.

Denoising of an image is a vital image processing task, both as a method itself, and as a part in other processes (5). There are many ways to denoise an image. The major property of an image denoising model is that it will remove noise, preserve edges and enhance maximum peak signal to noise ratio (PSNR).

In the presently offered medical imaging modalities, ultrasound imaging is considered non-invasive, practically harmless to the human body, portable, accurate and cost effective (6). These characteristics have made the ultrasound imaging the foremost prevalent diagnostic tool in nearly all hospitals around the world. Therefore, in the past few decades, considerable efforts in the field of ultrasound imaging are directed at development of signal processing techniques (7).

Ultrasound images are mainly corrupted through intrinsic artefact called "speckle" (7), which is the outcome of the constructive and destructive coherent abstraction of ultrasound resonances. These resonances or echoes can be due to device malfunctions or because of an inexperienced radiologist.

The evaluation of post-processing speckle reducing technique (8) employs the most common techniques of anisotropic diffusion (9), wavelet denoising and local statistics (10). Other approaches are proposed, together with the non-local means method. Generally, the anisotropic diffusion (11) [particularly oriented speckle reducing anisotropic diffusion – OSRAD], nonlinear multi-wavelet diffusion (NMWD) filter and other wavelet-based methods are not sufficient for a speckle removal. In the existing device, the OSRAD filter is considered the appropriate method for clinical application, based on consideration of its performance on both simulated and clinical data, and additional evaluation of its computational necessities (12). However, the OSRAD filter shows poor visual quality of images with low structural similarities.

In this paper, an improved anisotropic diffusion filter (IADF) was designed to exhibit strong speckle suppression by changing the eigenvalues, and the performance of the improved filter is evaluated. The IADF is an iterative process, everywhere kernel information is processed and regular update

is carried out at the kernel set. This technique can also be accurately deployed for a large number of cases.

METHODS

The multiplicative Lee filter approximates with a linear model to get the signal estimate, and the weighting function is described previously (10). The filter projected by Kuan *et al* comes by transforming into a signal-dependent additive noise formulation rather than the linear approximation utilized in the Lee filter (7). An equivalent general form of the Lee filter is used, but with a weighting function described by Kuan *et al* (7) and it effectively controls the amount of smoothing applied to the image by the filter (13). Frost *et al* estimated the noise-free image by convolving the observed image with a spatially-varying kernel, as cited by Kuan *et al* (7). Yu and Acton developed a diffusion approach better suited to speckle noise removal. The diffusion partial differential equations (PDE) is employed, the variation coefficient utilized in synthetic aperture radar (SAR) filtering strategies as a signal/edge discriminator for filtering processes where the ratio of local standard deviation to mean is given (14). Detail preserving anisotropic diffusion (DPAD) filter is proposed that improves upon the operation of the SRAD filter (12).

Weickert introduced a coherence enhancing diffusion (CED) tensor-valued diffusion function, permitting the level of smoothing to vary directionally (15). The nonlinear coherent diffusion (NCD) technique of Abd-Elmoniem *et al* tries to discriminate between different levels of speckle, based on the similarity to completely developed speckle. Image regions closely resembling totally developed speckle are mean filtered, while those dissimilar remain unchanged (11). As with the CED technique, this approach utilizes a tensor-valued diffusion function (16), calculated as a component-wise convolution of a Gaussian kernel with the structure tensor (11). Here, the initial stage of smoothing performed is not employed; therefore, the structure matrix represents gradient information from image details of even the smallest size.

Oriented speckle reducing anisotropic diffusion Krissian *et al* extended the SRAD technique to a matrix diffusion scheme (16). This permits the speckle adaptive diffusion to vary in strength within the contour and curvature directions. The enhancements of the DPAD method are utilized in this filter, like the employment of a larger window to estimate $q(x,y;t)$ and the median estimation of $q_0(t)$. The OSRAD diffusion function $c(q)$ relies on the Kuan *et al* filter (7).

The local directional variance is interrelated to the local geometry of the image. The extension of the SRAD technique to a matrix scheme is performed by finding the local directions of the gradient and curvature. This could be performed utilizing the Hessian matrix, as here, the structure tensor T_p is employed, like the CED and NCD methods. As in the CED technique, the eigenvectors of T_p are accustomed to construct the diffusion matrix, D . The eigenvalues of D , defining the strength of diffusion within the gradient and curvature direc-

tions, are given as

$$\lambda_1 = C_{SRAD} \quad [1]$$

$$\lambda_2 = C_{tang} \quad [2]$$

Where λ_1 gradient direction, λ_2 curvature direction, C_{SRAD} is the SRAD diffusion [c (q)] and C_{tang} is a constant. Diffusion is then performed.

Improved oriented speckle reducing anisotropic diffusion

Krissian *et al* extended the SRAD method to a matrix diffusion scheme (16). This permits the speckle adaptive diffusion to differ in power in the contour and curvature directions. The improvements of the DPAD technique are used in this filter, such as the usage of a larger window to evaluate $q(x,y;t)$ and the median approximation of $q_0(t)$. The OSRAD diffusion function $c(q)$ is based on the Kuan *et al* filter (7).

The local directional variance is related to the local geometry of the image. The extension of the SRAD technique to a matrix scheme is achieved by finding the local directions of gradient and curvature. This can be performed using the Hessian matrix, but here, the structure tensor is used, similar to the CED and NCD methods. As in the CED method, the eigenvectors of T_p are used to construct the diffusion matrix, D . The eigenvalues of D , describing the strength of diffusion in the gradient and curvature directions, are specified as (18):

$$\lambda_1 = c_{SRAD} \quad \text{and} \quad \lambda_2 = c_{DPAD} \quad [3]$$

where

$$c_{SRAD} = \frac{q(x,y;t)}{I(x,y;t)} \quad [4]$$

$$c_{DPAD} = c[q(x,y;t), q_0(t)] = \frac{1 + 1/q^2(x,y,t)}{1 + 1/q_0^2(t)} \quad [5]$$

Diffusion steps

Improved anisotropic diffusion filter scheme the separate set image gradient ($R_x I$, $R_y I$, $L_x I$, $L_y I$) and apply eigenvalues to improve the performance of the similarity index and the edge preservation accuracy.

1. Calculate the image gradient ($R_x I$, $R_y I$, $L_x I$, $L_y I$)

$$R_x I = \begin{bmatrix} 0 & 0 & 0 \\ 0 & -1 & 1 \\ 0 & 0 & 0 \end{bmatrix}$$

$$R_y I = \begin{bmatrix} 0 & 1 & 0 \\ 0 & -1 & 0 \\ 0 & 0 & 0 \end{bmatrix}$$

$$L_x I = \begin{bmatrix} 0 & 0 & 0 \\ -1 & 1 & 0 \\ 0 & 0 & 0 \end{bmatrix}$$

$$L_y I = \begin{bmatrix} 0 & 0 & 0 \\ 0 & 1 & 0 \\ 0 & -1 & 0 \end{bmatrix}$$

2. Make the kernel window co-ordinates (X,Y)
3. Build the Gaussian second derivatives filters (D_{xx} , D_{yy} , D_{xy}) $D_{Gaussyy} = D_{Gaussxx}$
4. Compute the eigenvectors (V_x, V_y)
5. Normalize the values and check whether eigenvectors are orthogonal
 $v1x = -v2y$
 $v1y = v2x$
6. Compute the eigenvalues (μ_1, μ_2)
 $\lambda_1 = \mu_1$
 $\lambda_2 = \mu_1$
7. Sort eigenvalues by absolute value $abs(\lambda_1) < abs(\lambda_2)$
8. Check $abs(\mu_1) < abs(\mu_2)$
9. Replace the value of OSRAD filter λ_1 and λ_2 with $\lambda_1 = c_{SRAD}$ and $\lambda_2 = c_{DPAD}$ mentioned as part of the proposed system
10. Update the kernel information based on the Gaussian second order derivatives and λ_1 and λ_2
11. Reconstruct the image using the updated kernel information from step 10
12. Display the reconstructed image as denoised results

In the existing approach, $\lambda_2 = c_{tang}$ is a constant, but in the proposed approach, eigenvalues are taken as the extension of DPAD. To measure the variation caused by filtering, the difference between the filtered and reference value is used. This is normalized relative to the value in the reference image.

Filter performance evaluation

The effects of filtering are calculated by application of image quality metrics. Image performance metrics are applied to the ultrasound images. The performance of each filter is evaluated quantitatively for ultrasound image with speckle noise using quality metrics such as maximum peak signal-to-noise ratio (PSNR), structural similarity index (SSIM) and figure of merit (FoM).

i) Peak signal-to-noise ratio

The PSNR indicates the level of chosen signal to the level of background noise and is computed using

$$PSNR = 20 \log_{10}(g_{max}^2 / MSE) \quad [6]$$

where MSE is defined as mean square error and g_{max}^2 is the maximum intensity in the unfiltered images (18). The larger PSNR values show good quality image.

ii) Structural similarity

The SSIM measure is a way for assessing the similarity between two images. It is used to assess the preservation of spatial information of the pixel in the filtering process.

$$SSIM = \frac{1}{M} \sum \frac{(2\mu_1\mu_2 + C_1)(2\sigma_{12} + C_2)}{(\mu_1^2 + \mu_2^2 + C_1)(\sigma_1^2 + \sigma_2^2 + C_2)} \quad [7]$$

where μ_1, μ_2 and σ_1, σ_2 are the means and standard deviations of the images being compared, and σ_{12} is the covariance between them; M is the total number of pixels. The SSIM has values in the -1 to 1 range, with unity representing structurally similar images. The SSIM lies between -1 for a bad and 1 for a good similarity between the original and despeckled images (11).

iii) Figure of merit

The FOM indicates edge pixel displacement between each filtered image I_{filt} and reference I_{ref} .

$$\text{FOM}(I_{\text{filt}}, I_{\text{ref}}) = \frac{1}{\max(N_{\text{filt}}, N_{\text{ref}})} \sum_{i=1}^{N'} \frac{1}{1+d_i^2 \alpha} \quad [8]$$

where N_{filt} and N_{ref} are the number of edge pixels in edge maps of I_{filt} and I_{ref} . Parameter α is set to a constant $1/9$, and d_i is the Euclidean distance of the i^{th} detected edge pixel and the nearest ideal edge pixel. The FOM has a range between 0 and 1, with 1 representing perfect edge preservation (11).

RESULTS

In this experiment, the clinical ultrasound images of cervix dataset obtained from the Regional Cancer Centre, Thiruvananthapuram, is shown in Fig. 1. The standardized ways of

Fig. 1: Ultrasound cervix dataset test images (TI1–TI24).

organizing and storing the image are in the format of bmp and the dimensions of 256×256 . For analysis, 24 ultrasound images were tested at different times. Figure 2 shows the different levels of speckle noise such as 0.01, 0.02 and 0.03 of test image TI1 from the dataset. Figure 3 demonstrates the seven types of filter, processed from speckle noise 0.03 (level 3) ultrasound test image from the dataset. Tables 1–3 show the performances (PSNR, SSIM and edge preservation accuracy [FOM]) for all test images with IADF.

Fig. 2: Noisy input image TI1. a) Noisy input image at level 0.01, b) noisy image at level 0.02, c) noisy image at level 0.03.

Fig. 3: Denoising of cervix input image (level 3) of different filters. a) Lee filter, b) Frost filter, c) Kuan filter, d) speckle reducing anisotropic diffusion (SRAD) filter, e) detail preserving anisotropic diffusion (DPAD) filter, f) oriented speckle reducing anisotropic diffusion (OSRAD) filter, g) improved anisotropic diffusion filter (IADF) (proposed).

From Tables 1–3, there are a few observations. The outcome shows that best performance for PSNR (31 dB) and SSIM (0.88) were achieved, respectively from the test images of TI2, TI3, TI5, TI7, TI10, TI11, TI13, TI16, TI17, TI19, TI21, TI22, TI23 and TI5, TI6, TI10, TI11, TI12, TI16, TI17, TI20, TI21. Similarly, edge preservation accuracy (88%) was obtained from the test images of TI6, TI10, TI13, TI14, TI19, TI20, TI21, TI24, respectively.

Table 1: Peak signal-to-noise ratio (PSNR) of improved anisotropic diffusion filter (IADF)

Performance matrix	PSNR
Test images	IADF (proposed)
TI2, TI3, TI5, TI7, TI10, TI11, TI13, TI16, TI17, TI19, TI21, TI22, TI23	31 dB
TI1, TI4, TI6, TI8, TI9, TI12, TI14, TI15, TI18, TI20, TI24	30 dB

Table 2: Structural similarity of improved anisotropic diffusion filter (IADF)

Performance matrix SSIM

Test images	IADF (proposed)
TI5, TI6, TI10, TI11, TI12, TI16, TI17, TI20, TI21	0.88
TI3, TI8, TI14, TI18, TI23, TI24	0.85
TI2, TI7, TI13, TI22	0.84
TI4, TI9, TI15, TI19	0.82
TI1	0.80

SSIM: structural similarity index map

Table 3: Edge preservation accuracy of improved anisotropic diffusion filter (IADF)

Performance matrix	Edge preservation accuracy
Test images	IADF (proposed)
TI6, TI10, TI13, TI14, TI19, TI20, TI21, TI24	88%
TI4, TI11, TI17, TI22	86%
TI5, TI12, TI18, TI23	84%
TI3, TI9, TI16	82%
TI1, TI2, TI7, TI8, TI15	81%

These observations show the IADF achieved best results for all cases. Some images show the predominant results for proposed filter and as an average, the best overall filter performances were obtained by the combination of all the datasets. Overall, experimental results proved that the IADF suppresses the speckle noise higher than previous filters. Figure 4 compares the PSNR rate of different filters with IADF. Figure 5 compares the structural similarity ratio of different filters with IADF. It is noted that IADF shows better performance. Figure 6 compares the accuracy rate of different filters with IADF, and IADF shows better performance in all cases.

Fig. 4: Extracted maximum peak signal-to-noise ratio (PSNR) rate of different filters.

Fig. 5: Extracted maximum structural similarity index map (SSIM) value of different filters.

Fig. 6: Extracted maximum edge preservation accuracy of different filters.

In the case of improved anisotropic diffusion, the filter suppresses the speckle more but the similarity between the two images shows poor performance. Table 4 shows PSNR values of all 24 images and the filters. Table 5 shows structural similarity between all the filters, while Table 6 shows the edge preservation accuracy of all filters. The low PSNR rate of 30 dB and SSIM (0.80) is noted in some test cases. The results also show low edge preservation accuracy (81%) noted in some test cases. Separate set of image gradient and eigenvalues improve the accuracy rate of PSNR, SSIM and edge preservation accuracy.

Table 4: Quality matrix of peak signal-to-noise ratio (PSNR)

Performance matrix		PSNR (dB)					
Test images (TI)	Lee	Frost	Kuan	DPAD	SRAD	OSRAD	IADF (proposed)
TI1	23.0164	22.3975	22.4512	22.341	22.2361	24.1209	30.023
TI2	22.8471	21.7312	21.4012	23.1756	23.1765	23.1771	30.5114
TI3	21.4045	20.1289	21.145	24.1472	24.1231	24.1785	30.7482
TI4	22.0164	21.3975	21.4512	21.341	21.2361	21.1209	30.024
TI5	22.0271	21.6312	21.1432	23.1435	23.1765	23.1771	30.7414
TI6	20.4124	20.1354	21.132	24.1874	24.1681	24.4532	30.2482
TI7	22.0164	21.3975	21.4512	21.341	22.2361	23.1209	30.7812
TI8	21.8471	20.1354	20.4012	22.1756	22.1765	22.1771	30.3114
TI9	20.4045	21.3975	20.145	23.1472	23.1231	23.1785	30.4578
TI10	22.0164	20.7312	21.4512	21.341	21.2361	23.1209	30.523
TI11	20.4045	19.1289	20.4012	22.1756	22.1765	22.1771	30.5123
TI12	21.8471	21.3975	21.145	23.1472	23.1231	24.1785	30.2873
TI13	22.0164	20.7312	20.4012	22.1756	22.1765	22.1771	30.5123
TI14	21.8471	19.1289	22.4512	21.341	21.2361	23.1209	30.0451
TI15	20.4045	20.7312	20.4012	22.1756	22.1765	22.1771	30.3451
TI16	21.8471	21.3975	21.4012	22.1756	23.1231	24.1209	30.8482
TI17	23.0164	20.7312	20.4012	22.1756	22.1765	22.1771	30.5123
TI18	21.8471	20.7312	21.145	20.1472	22.2361	22.1771	30.023
TI19	21.4045	19.1289	20.145	22.341	22.1765	24.1785	30.5123
TI20	22.8471	22.3975	20.4012	22.1756	23.1765	24.1209	30.3114
TI21	21.8471	20.7312	21.4512	24.1472	22.1765	23.1785	30.5123
TI22	21.8471	21.7312	20.4012	22.1756	24.1231	22.1771	30.7482
TI23	23.0164	20.7312	22.4512	23.1756	22.1765	23.1785	30.8482
TI24	22.8471	20.1289	21.145	22.1756	22.2361	23.1771	30.3114

DPAD: detail preserving anisotropic diffusion; SRAD: speckle reducing anisotropic diffusion; OSRAD: oriented speckle reducing anisotropic diffusion; IADF: improved anisotropic diffusion filter

Table 5: Quality matrix of structural similarity index map (SSIM)

Performance matrix	SSIM						
Test images (TI)	Lee	Frost	Kuan	DPAD	SRAD	OSRAD	IADF (proposed)
TI1	0.8738	0.8517	0.8281	0.8424	0.8704	0.1265	0.8001
TI2	0.3641	0.4134	0.6885	0.2563	0.2009	0.0781	0.8445
TI3	0.821	0.9293	0.5656	0.5986	0.4897	0.0563	0.8513
TI4	0.8602	0.8193	0.8136	0.7107	0.8246	0.8153	0.8217
TI5	0.0683	0.0703	0.9266	0.1404	0.0939	0.5298	0.878
TI6	0.442	0.4859	0.2113	0.4409	0.4265	0.3943	0.8753
TI7	0.3641	0.4134	0.6885	0.2563	0.2009	0.0781	0.8415
TI8	0.821	0.9293	0.5656	0.5986	0.4897	0.0563	0.8513
TI9	0.8602	0.0703	0.8136	0.7107	0.8246	0.8153	0.8247
TI10	0.0683	0.4859	0.9266	0.1404	0.0939	0.5298	0.878
TI11	0.442	0.4134	0.2113	0.4409	0.4265	0.3943	0.8753
TI12	0.3641	0.9293	0.6885	0.2563	0.2009	0.0781	0.8753
TI13	0.3641	0.8193	0.5656	0.5986	0.4897	0.0563	0.8465
TI14	0.821	0.0703	0.9266	0.7107	0.4265	0.8153	0.8513
TI15	0.8602	0.4134	0.2113	0.1404	0.8246	0.5298	0.8247
TI16	0.0683	0.4859	0.8281	0.4409	0.0939	0.3943	0.878
TI17	0.442	0.9293	0.6885	0.8424	0.4265	0.5298	0.8753
TI18	0.3641	0.8193	0.5656	0.2563	0.4897	0.3943	0.8513
TI19	0.821	0.0703	0.8136	0.5986	0.8246	0.0781	0.8247
TI20	0.8602	0.4859	0.2113	0.7107	0.0939	0.8153	0.878
TI21	0.0683	0.4134	0.6885	0.5986	0.4897	0.5298	0.8753
TI22	0.442	0.4859	0.5656	0.5986	0.0939	0.3943	0.8415
TI23	0.821	0.0703	0.6885	0.2563	0.4265	0.0781	0.8513
TI24	0.8602	0.8193	0.8281	0.1404	0.8246	0.8153	0.8513

DPAD: detail preserving anisotropic diffusion; SRAD: speckle reducing anisotropic diffusion; OSRAD: oriented speckle reducing anisotropic diffusion; IADF: improved anisotropic diffusion filter

Table 6: Quality matrix of figure of merit (FOM)

Performance matrix	Edge preservation accuracy (%)						
Test images (TI)	Lee	Frost	Kuan	DPAD	SRAD	OSRAD	IADF (proposed)
TI1	0.6531	0.7197	0.6377	0.4227	0.1506	0.7925	0.8111
TI2	0.6215	0.6803	0.6089	0.3807	0.1359	0.7294	0.8148
TI3	0.6286	0.6527	0.5688	0.3576	0.1231	0.746	0.8243
TI4	0.6993	0.7375	0.6384	0.5144	0.3118	0.7735	0.8641
TI5	0.7003	0.7375	0.677	0.5525	0.3272	0.7854	0.8439
TI6	0.6929	0.7575	0.6751	0.5385	0.3438	0.7889	0.8845
TI7	0.6531	0.7197	0.6751	0.3807	0.1506	0.7925	0.8111
TI8	0.6215	0.6803	0.6377	0.4227	0.1359	0.7294	0.8148
TI9	0.6286	0.6527	0.6089	0.3807	0.1231	0.746	0.8243
TI10	0.6531	0.7197	0.5688	0.3576	0.3118	0.7925	0.8811
TI11	0.6993	0.7375	0.6384	0.5144	0.3118	0.7735	0.8641
TI12	0.7003	0.7553	0.677	0.5525	0.3272	0.7854	0.8439
TI13	0.6929	0.7575	0.6751	0.5385	0.3438	0.7889	0.8845
TI14	0.6531	0.7197	0.6089	0.3576	0.1506	0.7925	0.8811
TI15	0.6929	0.6803	0.6377	0.4227	0.1359	0.7294	0.8148
TI16	0.6215	0.6527	0.6089	0.3807	0.1231	0.746	0.8243
TI17	0.6286	0.6803	0.5688	0.3576	0.3118	0.7735	0.8631
TI18	0.6993	0.7375	0.6384	0.5144	0.3272	0.7854	0.8439
TI19	0.7003	0.7553	0.677	0.5525	0.3272	0.7889	0.8845
TI20	0.6929	0.7575	0.6751	0.5385	0.3438	0.7294	0.8758
TI21	0.6993	0.7375	0.6384	0.5144	0.3438	0.7889	0.8845
TI22	0.7003	0.7553	0.6384	0.5385	0.3118	0.7735	0.8641
TI23	0.6929	0.7575	0.677	0.5525	0.3272	0.7854	0.8439
TI24	0.6993	0.7553	0.6751	0.5385	0.3438	0.7889	0.8845

DPAD: detail preserving anisotropic diffusion; SRAD: speckle reducing anisotropic diffusion; OSRAD: oriented speckle reducing anisotropic diffusion; IADF: improved anisotropic diffusion filter

DISCUSSION

In this work, we presented an efficient IADF that gives better results, and this approach represents some real time performance for denoising an ultrasound image. The proposed approach provides the quality metrics of 31 dB, SSIM of 0.88 and also the edge preservation accuracy of 88%. Additionally, our approach shows better improvement in terms of SSIM with the original image reconstructed.

Further investigations into the nature and uses of IADF are analysed. The computational efficiency and robustness of the proposed approach is compared with several state-of-the-art filters in ultrasound images. The experimental results verify that the proposed technique achieves the highest accuracy among all the strategies under comparison. Lastly, the IADF is tested with mostly real time ultrasound images and shows better improvements with less time for execution.

REFERENCES

- World Health Organization. Comprehensive cervical cancer control: a guide to essential practice. Geneva: WHO; 2006.
- India Medical Times. Cervical cancer leading cancer-killer among Indian women. Aalattimes Media Venture; 2012 Mar 28 [cited 2013 Jul 10]. Available from: <http://www.aalattimes.com/2012/03/28/cervical-cancer-leading-cancer-killer-among-indian-women>
- Parkin DM, Bray F, Ferlay J, Pisani P. Global cancer statistics, 2002. CA Cancer J Clin 2005; **55**: 74–108.
- Castellsague X, de Sanjose S, Aguado T, Louie KS, Bruni L, Munoz J et al. HPV and cervical cancer in the world 2007 report. Vaccine 2007; **25**: C1–C26.
- Shanta V. Perspectives in cervical cancer prevention in India. Brussels: International Network for Cancer Treatment and Research; 2003.
- Oviredy S, Muthusamy E. Speckle suppressing anisotropic diffusion filter for medical ultrasound images. Ultrason Imaging 2014; **36**: 112–32.
- Kaur J, Kaur J, Kaur M. Survey of despeckling techniques for medical ultrasound images. Int J Comp Tech Appl 2011; **2**: 1003–7.
- Aja-Fernandez S, Alberola-Lopez C. On the estimation of the coefficient of variation for anisotropic diffusion speckle filtering. IEEE Trans Image Process 2006; **15**: 2694–701.
- Loizou CP, Pattichis CS, Member S, Christodoulou CI, Istepanian RSH, Pantziaris M et al. Comparative evaluation of despeckle filtering in ultrasound imaging of the carotid artery. IEEE Trans Ultrason Ferroelectr Freq Control 2005; **52**: 1653–69. Available from: <http://ieeexplore.ieee.org/stamp.jsp?arnumber=1561621>
- Lee JS. Digital image enhancement and noise filtering by use of local statistics. IEEE Trans Pattern Anal Mach Intell 1980; **2**: 165–8.
- Abd-Elmoniem K, Youssef A-B, Kadah Y. Real-time speckle reduction and coherence enhancement in ultrasound imaging via nonlinear anisotropic diffusion. IEEE Trans Biomed Eng 2002; **49**: 997–1014.
- Finn S, Glavin M, Jones E. Echocardiographic speckle reduction comparison. IEEE Trans Ultrason Ferroelectr Freq Control 2011; **58**: 82–101.
- Shanthi I, Valarmathi ML. Speckle noise suppression of SAR color image using hybrid mean median filter. Int J Comput Appl 2011; **31**: 13–22.
- Yu Y, Acton S. Speckle reducing anisotropic diffusion. IEEE Trans Image Process 2002; **11**: 1260–70.
- Liu F, Liu J. Anisotropic diffusion for image denoising based on diffusion tensors. J Vis Commun Image R 2012; **23**: 516–21.

16. Krissian K, Westin C-F, Kikinis R, Vosburgh K. Oriented speckle reducing anisotropic diffusion. *IEEE Trans Image Process* 2007; **16**: 1412–24.
17. Hafizah WM, Supriyanto E. Comparative evaluation of ultrasound kidney image enhancement techniques. *Int J Comput Appl* 2011; **21**: 15–9.
18. Jemila Rose R, Allwin S. Speckle suppressing improved oriented speckle reducing anisotropic diffusion (IOSRAD) filter for medical ultrasound images. *Appl Mech Mater* 2014; **626**: 106–10.