Published in final edited form as:

J Fluoresc. 2003 September ; 13(5): 371–374. doi:10.1023/A:1026103804104.

A Glucose Sensing Contact Lens: A Non-Invasive Technique for Continuous Physiological Glucose Monitoring

Ramachandram Badugu¹, Joseph R. Lakowicz^{1,3}, and Chris D. Geddes^{1,2,3}

¹Center for Fluorescence Spectroscopy, Department of Biochemistry and Molecular Biology, Medical Biotechnology Center, University of Maryland School of Medicine, Baltimore, Maryland

²Institute of Fluorescence and Center for Fluorescence Spectroscopy, Medical Biotechnology Center, University of Maryland Biotechnology Institute, Baltimore, Maryland

Abstract

We have developed a range of glucose sensing contact lenses, using a daily, disposable contact lens embedded with newly developed boronic acid containing fluorophores. Our findings show that our approach may be suitable for the continuous monitoring of tear glucose levels in the range 50–1000 μ M, which typically track blood glucose levels, which are $\approx$ 5–10 fold higher. Our non-invasive approach may well offer an alternative solution to current invasive glucose monitoring techniques for diabetes, such as “finger pricking.”

Keywords

Glucose monitoring; non-invasive; continuous monitoring; contact lens

INTRODUCTION

Diabetes results in long term health disorders including cardiovascular disease and blindness. One of the major challenges in the management of diabetes is the monitoring of glucose concentrations.

The most commonly used technology for blood glucose determination is an enzyme based method, which requires frequent blood sampling and therefore blood drawing. While “finger pricking” is a relatively painless process, this method does suffer from a few practical problems. The first one is inconvenience and the required compliance by patients, while the second is that this is not a continuous monitoring method. Despite intensive efforts [1–11], no method is currently available for the continuous non-invasive monitoring of blood glucose.

Elevated tear glucose during hyperglycaemia was first demonstrated by Michail *et al.* as early as 1937 [12,13], as tear glucose levels track blood levels, in an analogous manner to the equilibrium that normally exists between blood and tissue fluid [14]. Since that time tear

³To whom correspondence should be addressed. cfs@cfs.umbi.umd.edu.

glucose has received very little attention for assessing patients for hyperglycaemia [15–19], and those studies undertaken have employed invasive and non-continuous methodologies.

We have embraced the notion of elevated tear glucose levels during hyperglycaemia and investigated the possibility, for the first time, of monitoring tear glucose, and therefore blood glucose, using a disposable, off-the-shelf contact lens. We chose to adopt commonly used plastic contact lenses,⁴ for the primary reason that their physiological compatibility has already been assessed and optimised with regard to vision correction, size and oxygen/analyte permeability etc. However, our ongoing studies and an intense feasibility study have revealed that published boronic acid containing fluorophores (BAFs), which are well-known to sense monosaccharides [20–26], do not respond well from within the lenses. This is currently thought due to the low internal pH and polarity of the contact lens polymers [27], combined with the fact that the pKa's of the BAFs have been mostly designed to operate at physiological blood pH, i.e., pH 7.2 [20–26].

Subsequently, we have designed a range of new BAFs, based on the quaternized form of the quinolinium nucleus, which respond well to glucose within the contact lens and which also display a high monosaccharide affinity, afforded by the stabilization of the anionic glucose bound form (boronate ester form) by the cationic quaternary nitrogen heterocyclic center, Fig. 1. In addition we have synthesized the respective control compounds, which do not contain the boronic acid moiety, and are therefore not sensitive to monosaccharides, but which have allowed us to additionally rationale and correct for potential chloride, oxygen, fluoride, albumin and pH interferences, Fig. 1—bottom.

Figure 2 shows the pH dependence of the emission intensities of probes 1 and 4 in the presence of glucose, fructose and in buffer. Interestingly, the pKa of the glucose bound forms are substantially reduced, pKa 6 → 7, as compared to typical BAFs reported in the literature, pKa's usually $\gg 7$. This reduction in probe-bound sugar pKa affords for a substantial sugar response in the contact lens polymers studied here,⁵ which we have shown have an internal pH in the range 5.5–6.5 and a lens polarity similar to that of methanol [27].

Figure 3 shows the response of the contact lenses doped with both probes 1 and 4, in the tear glucose range of physiological importance, i.e. <1 mM glucose. A similar response can be observed for a probe 1 doped lens towards both glucose and fructose, while a higher fructose affinity can be observed for a probe 4 doped lens, Fig. 3 bottom. Interestingly, the concentration of fructose in blood is typically 10 fold lower than glucose, which suggests that fructose is unlikely to interfere in the glucose response of a probe 1 doped contact lens.

⁴The contact lenses were supplied by CIBA Vision, Atlanta, USA, and were stirred in 500 mL water, 20°C for 24 hr before post-doping. The contact lens is a polyvinyl alcohol type polymer which swells slightly in water. Its hydrophilic character readily allows for the diffusion of the aqueous analytes in tears. Doping the contact lenses was undertaken by incubating the lenses in a high concentration of the respective BAF solution for 24 hr, before being rinsed in Millipore water. Doped lenses were then allowed to leach probe for 1 hr into a large volume. Leaching was typically complete after 15–30 min, evident by no further change (loss) in lens fluorescence intensity.

⁵See FN4.

CONCLUSIONS

In this rapid communication we have briefly described a range of boronic acid containing fluorophores which have been both designed and synthesised to respond to physiological tear glucose in a disposable, plastic, off-the-shelf contact lens. The probes have been designed to be compatible with both the internal pH and polarity within the lens, as well as to respond well too physiological tear glucose, which is typically present at a 10-fold lower concentration in tears, than in blood. The reversible nature of our glucose sensing contact lens, their sensitivity, and their requirement for non-UV excitation, strongly suggests that our approach could be used for the continuous and noninvasive monitoring of physiological glucose, Fig. 4. We envisage that multiple modes of sensing can be undertaken in the contact lens, such as ratiometric, lifetime and (visual) polarisation based glucose fluorescence sensing [28], Fig. 4, for both doped lenses and lenses containing unique sensing regions (spots). Further details will be reported in a full paper in due course.

Acknowledgments

The authors wish to thank the University of Maryland Biotechnology Institute and the National Center for Research Resources, RR-08119, for financial support.

Abbreviations

BAFs Boronic acid containing fluorophores

References

1. Robinson MR, Eaton RP, Haaland DM, Koepp GW, Thomas EV, Stallard BR, Robinson PL. Non-invasive glucose monitoring in diabetic patients: A preliminary evaluation. *Clin Chem.* 1992; 38:1618–1622. [PubMed: 1525990]
2. Heise HM, Marbach R, Koschinsky TH, Gries FA. Non-invasive blood glucose sensors based on near-infrared spectroscopy. *Ann Occup Hyg.* 1994; 18:439–447.
3. March WF, Rabinovitch B, Adams R, Wise JR, Melton M. Ocular glucose sensor. *Trans Am Soc Artif Intern Organs.* 1982; 28:232–235. [PubMed: 7164243]
4. Rabinovitch B, March WF, Adams RL. Non-invasive glucose monitoring of the aqueous humor of the eye, Part 1, Measurement of very small optical rotations. *Diabetes Care.* 1982; 5:254–258. [PubMed: 7172992]
5. Schier GM, Moses RG, Gan IET, Blair SC. An evaluation and comparison of reflolux Iiand Glucometer II, two new portable reflectance meters for capillary blood glucose determination. *Diabetes Res Clin Pract.* 1988; 4:177–181. [PubMed: 3359917]
6. Clarke W, Becker DJ, Cox D, Santiago JV, White NH, Betschart J, Eckenrode K, Levandoski LA, Prusinski EA, Simineiro LM, Snyder AL, Tideman AM, Yaegar T. Evaluation of a new system for self blood glucose monitoring. *Diabetes Res Clin Pract.* 1988; 4:209–214. [PubMed: 3359921]
7. Trettnak W, Wolfbeis OS. Fully reversible fiber-optic glucose biosensor based on the intrinsic fluorescence of glucose-oxidase. *Anal Chim Acta.* 1989; 221:195–203.
8. Meadows D, Schultz JS. Fibre optic biosensor based on fluorescence energy transfer. *Talanta.* 1988; 35:145–150. [PubMed: 18964483]
9. Tolosa L, Malak H, Rao G, Lakowicz JR. Optical assay for glucose based on the luminescence decay time of the long wavelength dye Cy5. *Sens Actuators B.* 1997; 45:93–99.
10. Tolosa L, Gryczynski I, Eichorn LR, Dattelbaum JD, Castellano FN, Rao G, Lakowicz JR. Glucose sensors for low cost lifetime-based sensing using a genetically engineered protein. *Anal Biochem.* 1999; 267:114–120. [PubMed: 9918662]

11. D'Auria S, Dicesare N, Gryczynski Z, Gryczynski I, Rossi M, Lakowicz JR. A thermophilic apoglucose dehydrogenase as a nonconsuming glucose sensor. *Biochem Biophys Res Commun*. 2000; 274:727–731. [PubMed: 10924344]
12. Michail D, Zolog N. Sur l'élimination lacrymale du glucose au cours de l'hyperglycémie alimentaire. *C R Soc Biol Paris*. 1937; 126:1042.
13. Michail D, Vancea P, Zolog N. Sur l'élimination lacrymale du glucose chez les diabétiques. *C R Soc Biol, Paris*. 1937; 125:1095.
14. Van Haeringen NJ. Clinical biochemistry of tears. *Surv Ophthalmol*. 1981; 29(2):84–96. [PubMed: 7034254]
15. Daum KM, Hill RM. Human tear glucose. *Invest Ophthalmol Vis Sci*. 1982; 22(4):509–514. [PubMed: 7199517]
16. Gasser AR, Braverman LE, Fleming MC, Arky RA, Alter BR. Tear glucose detection in hyperglycemia. *Am J Ophthalmol*. 1968; 65(3):414–420. [PubMed: 5643199]
17. Das BN, Sengupta S, Das BK, Goswami NR. Tear Glucose estimation—An alternative to blood glucose estimation. *J Indian Med Assoc*. 1995; 93(4):127–128. [PubMed: 8699035]
18. Giardai A, Roberts JRE. Concentration of glucose and total chloride in tears. *Br J Ophthalmol*. 1950; 34:737–743. [PubMed: 14821266]
19. Ridley F. The intra-ocular pressure and drainage of aqueous humour. *Br J Exp Pathol*. 1930; 11:217–222.
20. James TD, Sandanayake KRAS, Shinkai S. *Angew Chem Int Ed Engl*. 1994; 33:2207.
21. James TD, Sandanayake KRAS, Iguchi R, Shinkai S. *J Am Chem Soc*. 1995; 117:8982.
22. Bielecki M, Eggert H, Norrild JC. *J Chem Soc Perkin Trans*. 1999; 2:449.
23. Dicesare N, Lakowicz JR. Evaluation of two synthetic glucose probes for fluorescence-lifetime based sensing. *Anal Biochem*. 2001; 294:154–160. [PubMed: 11444811]
24. Dicesare N, Lakowicz JR. Wavelength-ratiometric probes for saccharides based on donor–acceptor diphenylpolyenes. *J Photochem Photobiol A: Chem*. 2001; 143:39–47.
25. Dicesare N, Lakowicz JR. New color chemosensors for monosaccharides based on Azo dyes. *Org Lett*. 2001; 3(24):3891–3893. [PubMed: 11720562]
26. Dicesare N, Lakowicz JR. Chalcone-analogue fluorescent probes for saccharides signaling using the boronic acid group. *Tetrahedron Lett*. 2002; 43:2615–2618.
27. Badugu R, Lakowicz JR, Geddes CD. Towards the non-invasive continuous monitoring of physiological glucose using a novel monosaccharide-sensing contact lens. *J Phys Chem B* Submitted. in press.
28. Gryczynski Z, Gryczynski I, Lakowicz JR. Fluorescence sensing methods. *Methods Enzymol*. 2002; 360:44–74. [PubMed: 12622146]

Boronic acid probes				
Entry	R ₁	R ₂	R ₃	R ₄
1	-OCH ₃	-B(OH) ₂	H	H
2	-OCH ₃	H	-B(OH) ₂	H
3	-OCH ₃	H	H	-B(OH) ₂
4	-CH ₃	-B(OH) ₂	H	H
5	-CH ₃	H	-B(OH) ₂	H
6	-CH ₃	H	H	-B(OH) ₂
Control compounds				
7	-OCH ₃	H	H	H
8	-CH ₃	H	H	H

Fig. 1.
Structural formula of new boronic acid containing fluorophores (BAFs) and their respective control compounds which do not contain the boronic acid moiety.

Fig. 2.

Emission intensity at $\lambda_{\text{max}} = 450 \text{ nm}$, I , divided by the initial emission intensity, I_0 , as a function of pH for Probe 1 (Top) and Probe 4, $\lambda_{\text{max}} = 427 \text{ nm}$ (Bottom) for the fructose-bound, glucose-bound and free probe (buffer) forms.

Fig. 3. Response of a Probe 1 (Top) and a Probe 4 (Bottom) doped contact lens towards both glucose and fructose. I —emission intensity in the presence of sugar and I' in the absence of sugar.

Fig. 4. Potential methods for non-invasive continuous tear glucose monitoring. (Top)—BAF doped contact lens as described here, and (Bottom)—Sensor spots on the surface of the lens to additionally monitor other analytes in addition to glucose, such as chloride or oxygen. Sensor spot regions may also allow for ratiometric, lifetime or polarization based fluorescence glucose sensing.