Method in Medicine

Equipment Review

BISPECTRAL INDEX - A latest technique to monitor depth of anaesthesia state

A new technology called "Bispectral index" (BIS) monitoring helps the anaesthetist monitor the effects of volatile and non-volatile intravenous anaesthetic agents on central nervous system and improve the titration of anaesthetic agents. BIS is a processed EEG parameter that measures the hypnotic effects of anaesthetic and sedative agents on the brain of anaesthetised patients. It is based on measurements of frequency, amplitude and coherence of the EEG.

The conventional processed EEG parameters like Spectral Edge Frequency (SEF) are based solely on power spectral analysis. SEF quantifies the gross EEG change to lower frequencies as anaesthesia deepens, but is not likely to correlate with the level of sedation. The BIS has undergone more clinical research than any other EEG parameter and is strongly correlated with clinical measurement of sedation and hypnosis.

General anaesthesia is associated with a decrease in EEG frequency and an increase in average power. This is called the power spectrum and is represented by a single number ranging from 100 for wide awake and 0 in the absence of brain activity. BIS measures the state of the brain and not the concentration of the anaesthetic agent in the brain. It measures the current hypnotic state and not the depth of anaesthesia. A low value of BIS indicates hypnosis irrespective of how it was achieved. BIS monitoring may be a guide to the titration of anaesthetic drugs to achieve an effective dose requirement without the risk of awareness and help better balance of hypnotic and analgesic administration. Potential benefits of BIS monitoring are low risk of awareness, better recovery and a more rational selection of anaesthetic drugs thereby achieving an effective anaesthetic regimen. BIS monitoring can be used in ambulatory surgical procedures, in cardiac surgery and in the technique of total intravenous anaesthesia (TIVA). In addition it can be used in patients with a previous history of awareness under anaesthesia, on patients with drug tolerance and in surgeries with an increased risk of awareness.

An anaesthetised patient during recovery shows return of consciousness above the BIS of 60. In balanced anaesthesia the BIS is maintained within a range

of 45 to 60, rising to above 60 in the last 15 minutes of anaesthesia. A BIS level of less than 60 gives adequate hypnosis where opioids are used for analgesic effect and an infusion of propofol is given for its hypnotic effect. A BIS level of less than 40 is required when an inhalational agent is used to produce hypnosis and analgesia. In day care surgery a BIS value of 55-60 is adequate. In case of intra-ocular surgery where sudden awakening is likely to initiate the oculo-cardiac reflex, a lower value of BIS is mandatory.

BIS is displayed as a numerical value obtained over a period of 30 seconds and is updated every 2-3 seconds thereby preventing fluctuations in value and effects of electrical cautery. Cerebral ischaemia due to any cause and hypothermia (35°C) can also lead to a low BIS value reflecting a low cerebral metabolic rate.

Artifacts and poor signal quality as a result of poor skin contact, muscle rigidity, head movement, sustained eye movements, improper sensor placement and unusual electrical interference may all lead to inappropriate BIS values. Reliance on the BIS value alone for intra-operative anaesthetic management is not advised. Although BIS, a processed EEG measure, has demonstrated its clinical utility in the management of anaesthesia and got the clearance from FDA, additional widespread clinical experience is required to determine its clinical utilization. In addition to the operation theatre, it can be used in the ICU set up, conscious sedation technique in dental surgery and endoscopy.

References

1. Gan TJ, Glass PS, Windsor A, et al. Bispectral Index Monitoring allows faster recall and improved recovery from Propofol, Alfentanil and Nitrous Oxide anaesthesia. *Anaesthesiology* 1997; 87(4): 808-15.
2. Kearse L, Roscow C, Zaslavsky A, et al. Bispectral analysis of the electroencephalogram predicts conscious processing of information during propofol sedation and hypnosis. *Anesthesiology* 1998; 88: 25-34.

Contributed by

Col PK ROY

Associate Professor, Department of Anaesthesiology and Critical Care Medicine, Armed Forces Medical College, Pune - 40