Published in final edited form as:

Trends Biotechnol. 2016 August ; 34(8): 665–678. doi:10.1016/j.tibtech.2016.05.003.

Micro- and Nano-scale Technologies for Delivery into Adherent Cells

Wonmo Kang^{a,b,†}, Rebecca L. McNaughton^b, and Horacio D. Espinosa^{a,b,c,*}

^aDepartment of Mechanical Engineering, Northwestern University, Evanston, IL 60208, USA

^biNfinesimal LLC, Skokie, IL 60077, USA

^cInstitute for Cellular Engineering Technologies, Northwestern University, Evanston, IL 60208, USA

Abstract

Several recent micro- and nano-technologies have provided novel methods for biological studies of adherent cells because the small features of these new biotools provide unique capabilities for accessing cells without the need for suspension or lysis. These novel approaches have enabled gentle, yet effective delivery of molecules into specific adhered target cells, with unprecedented spatial resolution. Here we review recent progress in the development of these technologies with an emphasis on *in vitro* delivery into adherent cells utilizing mechanical penetration or electroporation. We discuss major advantages and limitations of these approaches and propose possible strategies for improvements. Finally, we discuss the impact of these technologies on biological research concerning cell-specific temporal studies, e.g., non-destructive sampling and analysis of intracellular molecules.

Need For Techniques To Study Adherent Cells—A mechanistic understanding of cell biology is often limited by both the complexity of the processes and limitations of commonly available research tools that lack temporal or spatial resolution. The lack of tools capable of providing cell-specific, non-destructive biomolecular delivery and analysis is a particular barrier for advancing fundamental discoveries of cell heterogeneity, single-cell behavior within a complex environment, and the mechanisms that govern disease states, responses to drugs or other stimuli, and differentiation of stem cells. To gain new mechanistic understanding, advances in methods for precise intracellular delivery and non-destructive biochemical analyses of non-secretory molecules (e.g., mRNA and proteins) are greatly needed so that individual cells can be experimentally controlled and repeatedly analyzed over time and/or within a particular location of the cell. For example, developing neurons must undergo a series of sequential changes in gene expression to achieve a mature phenotype; hence, understanding the process will require the ability to accurately monitor the sequence of intracellular events, within individual cells, in a non-destructive manner. In addition, neuronal maturation is influenced by interactions with surrounding cells and with

*Correspondence: espinosa@northwestern.edu (H.D. Espinosa).

†Currently at the US Naval Research Laboratory

Publisher's Disclaimer: This is a PDF file of an unedited manuscript that has been accepted for publication. As a service to our customers we are providing this early version of the manuscript. The manuscript will undergo copyediting, typesetting, and review of the resulting proof before it is published in its final citable form. Please note that during the production process errors may be discovered which could affect the content, and all legal disclaimers that apply to the journal pertain.

extracellular matrix, so it is necessary to be able to simultaneously monitor events occurring in multiple cells that are interacting with each other and with the matrix. While the requirements are challenging, these experimental capabilities would provide unprecedented insight into the determinants of both the timing of cellular processes and their phenotype, the principles of cell heterogeneity, and the role of cell-cell communication in homogeneous cell populations and co-cultures.

Because most cells adhere to a substrate or to other cells during their growth or differentiation [1], it is advantageous for new technologies to be capable of accessing adhered cells to avoid the need to disrupt cell processes by suspension and replating. Several technologies for studying adhered cells are currently being developed, and due to the need for individual cell access and non-destructive probing, micro- and nano-technologies are a natural choice because they interact with cells at the appropriate length scale, reduce the working volume of expensive reagents, require less time and space for replicates, allow for automation and integration of sequential analyses, enable portability, and reduce waste [2, 3]. Here we present an overview of recently developed micro- and nano-tools, with a focus on trends in intracellular delivery for *in vitro* studies of adhered cells, and highlight major advantages/disadvantages of these technologies with respect to features such as individual cell selectivity, spatial resolution, non-destructive cell analysis, and potential for high throughput or automation. Finally, we discuss the exciting promise for these technologies to cause a paradigm shift in biological research by providing methods to study cells over time at the individual cell level.

Keywords

single cell access; electroporation; cell specificity; adherent cells; non-destructive analysis; nanobiotechnology

Technologies For *In Vitro* Studies Of Adherent Cells

Traditionally, molecules have been delivered into adhered cells by viral or chemical methods, micropipette injection, and electroporation, which often is significantly toxic and produces heterogeneous delivery results. These deleterious outcomes limit their usefulness for cell biology and biotechnology applications where high cell viability, dosage precision, and selectivity within a population are desired. By contrast, micro- and nano-technologies offer unprecedented levels of spatiotemporal control and cell stress minimization, which enables high efficiency high viability delivery of biomolecules and in some cases non-destructive live-cell analyses that could be transformative for exploring time-dependent phenotypes, heterogeneity, and differentiation mechanisms. Several recent micro- and nano-technologies have demonstrated promising potential as alternative methods for molecular delivery into adhered cells utilizing working principles that include: mechanical penetration and localized electroporation. Because studying a specific adhered cell during its natural state of growth requires accessing the cell individually, these technologies currently present a trade-off between experimental throughput and cell specificity or spatial resolution as summarized in Table 1. Nevertheless, further development of these technologies promises to increase their capabilities to study, analyze, and control adhered cells.

Mechanical Penetration

Arguably the simplest mechanism to deliver molecules into cells is by microinjection, which is performed by mechanically piercing the cell membrane using a needle-like structure with a sharp tip, e.g., a glass micropipette, which is positioned manually using a micromanipulator (Figure 1). Despite its instrumental simplicity, microinjection has several disadvantages associated with the size and shape of traditional micropipettes. The typical tip diameter is approximately 1 μm , comparable to the diameter of small cells (5–15 μm), and it has a tapered geometry, which limits this technique to use on larger cells. As the tip is inserted deep into a cell, the size of the pierced area on a cell membrane increases, which can be a significant perturbation to the cell. In addition, inserting the micropipette is difficult when a target cell is small (spherical diameter of <15 μm) or flat (thickness of <5 μm), where the micropipette tip may contact the substrate before the piercing is complete due to the compliant nature of the cell. The solution to overcome these disadvantages is to decrease the size of the tip such that it is much smaller than the size of the cell, which also minimizes the applied force needed to pierce into a cell membrane.

One-dimensional nanoprobess, cantilevers with nanoscale tips, and nanopipettes have been developed to increase spatial resolution and minimize cell perturbation during membrane piercing, Figure 1. This allows for intracellular delivery and live-cell probing without lysing or permanently damaging the cell [4, 11, 12, 15, 29–31]. The small size of the nanoprobe allows for extremely precise spatial positioning and long incubation times within the cell. The most common 1D nanostructures for nanoijection are nanotubes and nanowires with dimensions of 1–750 nm in diameter and 0.5–20 μm in length assembled on a linear wire or cantilever support [32, 33]. Common nanoprobess are fabricated using carbon nanotubes [9], boron nitride nanotubes [4], and silicon nanowires [34]. The dimensions and choice of material must be balanced to provide robustness such that the nanostructure possesses sufficient bending stiffness to effectively pierce a cell membrane (and if applicable, withstand capillary forces during immersion into aqueous solutions), but at the same time have enough strength to prevent mechanical failure during injection [35]. Although fabrication of the nanostructures has become easier with advancements in micro/nanotechnology, assembly with macroscopic support structures remains time-consuming.

Like conventional microinjection, the position of the nanostructure tip is controlled using a micromanipulator while manually monitoring it relative to the target cell via an optical microscope (Figure 1, left column). However, the limited resolution of optical microscopes makes monitoring the tip a challenge when the tip diameter is <30 nm [32], which is one reason that integration of the nanostructures with a cantilevered support has become increasingly popular for use within an atomic force microscope (AFM) (Figure 1, right column) [8, 9, 30, 34, 35]. In addition to cell injection, this AFM-based approach offers capability for mechanical quantification of cell-tip interaction as discussed in Box 1.

From non-fluidic to fluidic

Nanoijection with a non-fluidic nanostructure (Figure 1, bottom row) requires chemical functionalization of the tip such that molecules can be released into the cell or biomarkers can be captured by binding to the functionalized surface [4–6]. Thus, this approach is

limited to small dosage delivery of molecules that can be coordinated to the nanostructure surfaces, and it lacks precise control for releasing the molecules into the cell. In addition, because non-fluidic nanoinjection often relies on passive desorption of molecules, this technology requires relatively long dwell time, i.e., typically the tip remains inside each target cell for several minutes [8, 36] and, as a result, intrinsically limits its throughput. To provide active control of fluid delivery and increase the number of cells that can be treated per experiment, fluidic nanopipettes and nanoprobe are increasingly being used (Figure 1, top row).

While technology to reduce the size of glass nanopipette tips has improved such that diameters of <100 nm are routinely used [12, 37, 38], the fragility of the tip often results in mechanical failure during insertion into cells, and this has prompted innovations such as assembly of a carbon nanotube within a glass nanopipette, by either magnetic assembly [39], flow-through [10], or chemical vapor deposition [14], to take advantage of the superior mechanical flexibility and strength of the carbon nanotube [10, 13, 14, 39]. Additional utility is provided in this approach by correlating the change in electrical impedance and interfacial capacitance between a cell and a carbon nanopipette during nanoinjection to independently detect penetration of the cell and nucleus membranes, a critical capability to achieve automated injection [13, 14]. Another approach to increase the mechanical stability of glass nanopipettes is to reinforce them with a thin-film coating, which can also provide useful electrical properties [40]. In practice, however, difficulties in fabricating continuous nanotube structures and clogging of the nanochannels have been reported as major challenges [41].

An important advance that combines microscale fluidics with nanoscale tips, and eliminates the time-consuming fabrication of individual probes, is the development of cantilever-based technologies that are batch-fabricated from silicon wafers to include built-in microchannels, microreservoirs, and protruding tips with nanoscale opening [15, 19]. These versatile probes can be positioned by a manipulator or an AFM such that single cell molecular delivery into the cytoplasm or the nucleus can be achieved [15, 18, 20]. Because these microfluidic devices generally rely on pressure-driven injection utilizing an external pump or electrophoretic transport, precise control of injected volume (0.5 fL – 3 pL) has been achieved with much reduced dwell time inside a cell (<1 sec [20]) compared to non-fluidic approaches. While most fluidic nanoprobe technologies consist of one probe and one reservoir, it is worth noting that the wafer-based fabrication approach enables the design and use of multiple cantilevers or reservoirs on the same chip as demonstrated by the nanofountain probe (NFP) technology [42], which was fabricated with twelve cantilevers connected via embedded microchannels to two microreservoirs on a single chip [43]. This provides unique possibilities for on-chip multiplexing for delivery of different molecules, or delivery followed by analysis on the same chip.

It is important to note that similar to microinjection cases, tip size and shape play a key role in nanoinjection. For example, penetration of cell and nuclear membranes using a fluidic AFM cantilever with a conventional pyramid-shaped tip, requires about 30 and 60 nN, respectively [20], while the corresponding forces using an AFM probe with 1-D nanowire tip are about 0.5 and 1 nN [44]. Because fluidic AFM probes use considerably large forces,

compared to some of the traditional AFM probes (1–3 nN [36]), perturbation to the cell membrane could be significant and its biological implication should be appropriately investigated in the future.

Arrays of 1D nanostructures

Microfabricated substrates containing arrays of 1D nanostructures such as nanowires [22, 45–48] and nanostraws [24, 25] have been employed for delivery to and/or analysis of a population of adhered cells [33, 49]. When cells are cultured on top of the microfabricated substrates, the arrays of 1D nanostructures interact with the cells, although the exact mechanism of penetration of the cell membrane is currently being elucidated [50, 51]. By selecting the proper aspect ratio of the nanostructures (length to diameter), probe insertion deep into the cytosol can be obtained. The arrays are generally fabricated by either lithography-based techniques (top-down) or nanowire synthesis from a substrate using deposition techniques (bottom-up). The detailed review on these well-established micro- and nano-fabrication techniques can be found in [52]. In general, use of arrays of 1D nanostructures offers higher throughput, when compared to nanoinjection approaches, and a simpler experimental protocol. The remaining challenges are fabricating uniform arrays of 1D nanostructures and controlling the nature of the cell-nanostructure interface, e.g., the number of nanowires contacting or penetrating each cell. Also, unlike nanoinjection, it has been reported that arrays of 1D nanostructures may not penetrate through nuclear membranes, which prevents delivery of molecules directly into the nucleus [53].

Despite successful biological studies using various 1D nanostructures [22, 24, 25, 45–47], the governing mechanism(s) of interaction between the nanostructure and cell membrane are currently being investigated [22, 24, 50, 51]. For example, spontaneous endocytosis was proposed as the internalization mechanism when using arrays of nanowires (200 nm) or nanotubes (100 nm) [22, 24], yet high-resolution transmission electron microscope images of the interface between cortical neurons and SiO₂ nanowires (50–300 nm) indicate that cells surround the nanowires rather than penetrate the cell membranes [54]. Theoretical modeling using a mechanical continuum model of elastic cell membrane penetration predicts that gravitational force alone is not sufficient to trigger penetration of nanowires with diameters of approximately 50 nm [55], but the authors speculate that membrane piercing may be favorable with an additional external force, e.g., cell adhesion. This assertion is supported by another *in situ* experimental characterization that found only approximately 7% of 100 nm-diameter nanostraws penetrate cells and the penetration is adhesion dependent [51]. The influence of 1D nanostructures on cell phenotype is somewhat controversial because deleterious effects to the cells such as slow growth and abnormal division, development of irregular contours, lipid scrambling, and DNA damage have been observed in some cases [53, 56, 57]. Thus, further studies are needed toward fundamental understanding of cell-nanostructure interaction and its effect on cells.

Current and future trends in mechanical penetration

Low cell viability and low throughput have limited the expansion of conventional injection methods to biological applications beyond *in vitro* fertilization, but these limitations are being overcome by novel fabrication approaches and use of nanomaterials in the

manufacturing of nanoprobes and nanopipettes. Likewise, advances in instrumentation for increased force control provide a significantly less invasive means to penetrate cells, and as a result, cell viability after mechanical penetration has improved from < 50% to > 90%. Indeed, these advancements have allowed insertion of nanostructures into cells for over an hour while maintaining cell viability [34], which presents unique opportunities beyond delivery of molecules toward temporal live-cell analysis *in vitro*, potentially without repeated penetration of the cell membrane. As an example, selective capture and analysis of biomarkers is possible by coating 1-D nanowires with molecules such as antibodies [5] or with gold nanoparticles to monitor DNA and proteins using surface-enhanced Raman spectroscopy [6]. However, experimental throughput is still limited due to time-consuming positioning of the tips, and as a result, the practical applications remain in the realm of single-cell studies. This is an acceptable limitation for some applications because the unique capability of using mechanical penetration to obtain *in vitro* force measurements can provide insight into the role of mechanical signals that influence cell migration, cell growth, stem cell differentiation, and regulation of disease states [58].

Electroporation

Since its introduction in 1982 [59], bulk electroporation has become increasingly popular because it is reproducible, versatile (almost no limitation on cell type and size), and easy to use compared to alternative approaches [60]. Electroporation causes transient nanopores to form in the cell membrane when the cell is subjected to a sufficiently large electric field through which molecules can be delivered inside cells [3, 17, 18, 61–63]. As the electric field is typically created by applying an input voltage between two electrodes, the appropriate placement of the electrodes is crucial for reproducibility and consistent yield. The mechanism of cell electroporation involves three steps: membrane charging, pore nucleation, and pore evolution. During the membrane charging, an electrically non-conductive cell membrane behaves as a capacitor between conductive cell culture media and cytoplasm [64, 65]. Formation of nanopores (with estimated diameters of 2–50 nm [65]) is triggered when the electric potential difference across the cell membrane reaches 0.2–1 V [66]. Because a whole membrane of individual cells is subjected to an applied external electric field, bulk electroporation tends to create large numbers of small pores over a large fraction of a cell membrane. The key parameters that govern density, location, and size of nanopores in the cell membrane are amplitude, duration, frequency, and shape of the electrical input signals [67]. Transport of molecules into cells through the nanopores involves several mechanisms, including diffusion, convection, electrophoresis, electroosmosis, endocytosis, and micropinocytosis [64, 68, 69]. After the electric pulse, the nanopores shrink and reseal on the order of seconds. When an excessively high input voltage is applied, electroporation becomes irreversible. For example, application of several hundreds to thousands V/cm to cells results in cell lysing [70].

Bulk electroporation systems suffer from the need for high input voltage, non-uniformity of the electric field, formation of bubbles, and variations in local pH and temperature [60, 71, 72]. These challenges can be overcome by miniaturization to reduce the required input voltage, create a more uniform electric field, and rapidly dissipate heat due to the large surface-to-volume ratio [60]. Applying this miniaturization toward adhered cells has resulted

in electroporation methods that have recently been developed in configurations suitable for single cells using fluidic nanopipettes (Figure 2, right column) and for multiple cells using a lab-on-a-chip (LOC) approach (Figure 2, left column) [3, 18, 73].

Key attributes of localized electroporation

In localized electroporation, the applied electric field is focused to a small area (typically ranging from tens to hundreds of nanometers in diameter [3, 18, 25]) on the cell membrane that is electrically sealed by a micro- or nanochannel in contact with the cell membrane (Figure 2). Due to the focused electric field, this technique results in formation of relatively larger pores in a small area while allowing use of applied voltages that are orders of magnitude less than typical bulk electroporation. As a result, delivery efficiency and cell viability exceed 90% for most cell types and biomolecules. The system typically consists of metallic electrodes, a pump for fluidic control, and either a fluidic cantilever/pipette with nanoscale tip or a microporous substrate integrated within a microfluidic device.

Localized electroporation shares the same governing equation for the electric field as bulk electroporation, but accurate prediction of the focused electric field requires accounting for the effects of cell and nanochannel shape, size, and interface. Because the effective electric field applied to target cells depends on the microdevice architecture and dimensions, numerical analyses of localized electroporation are often utilized to optimize and quantify the local electric field needed for poration [3, 18, 74, 75]. Recently, experimental studies indicate that a highly focused electric field can alter the mechanism of molecular transport, in particular, for large molecules such as DNA plasmids and quantum dots. For example, use of a nanochannel creates a large electric field (70 MV/m for 200 V) only in the area of the cell membrane adjacent to the channel. This extremely large electric field, which is, interestingly, orders of magnitude larger than typical inputs for electrical cell lysis techniques, likely results in formation of large pores and a strong electrophoretic force that can transport plasmids directly into the cytoplasm [76]. Further studies to fully quantify the pore size, understand the dynamic evolution of membrane pores and corresponding molecular transport mechanism(s), when cells are subjected to highly focused and intense electric fields, are highly needed.

Single-cell selective localized electroporation

Techniques for single-cell localized electroporation have been developed using either glass nanopipettes or microfluidic cantilevers (Figure 2, right column) that are integrated with an electrode to apply the electrical signal through the fluidic channel and are positioned in contact with the membrane of an adhered cell using a micromanipulator [16, 18, 77, 78]. This experimental setup and capability share many similarities to microfluidic nanoinjection including high spatial resolution, dosage control, and versatility for delivering different types and sizes of molecules into adherent cells. However, there are two important differences that provide distinct advantages: (1) localized electroporation requires only gentle contact of the nanostructure tip with the cell membrane rather than mechanical penetration, which reduces the stress exerted on the cell and diminishes the influence of size and shape of the cell, and (2) electrical feedback-controlled positioning of the nanostructure tip can be used to detect contact with the cell membrane, which eliminates the need to rely on a skilled operator and

provides a method for automation that significantly increases throughput and delivery efficiency [77]. The nanofountain probe electroporation (NFP-E) system, a key example that demonstrates these advantages, is highlighted in Box 2.

Lab-on-a-Chip platforms for populations of cells

Lab-on-a-chip platforms offer electroporation of a population of adhered cells by incorporating microwell arrays or perforated substrates into the chip and coupling it with electrodes and fluidic controls (Figure 2, left column). Pneumatic valves are often integrated with these small-scale devices for automated, multiplexed, and high-throughput microfluidic control [79]. Because cells are cultured directly on these devices, biocompatibility of each component must be appropriately considered, in particular, for studies of sensitive cells and long-term cell behavior. Detailed reviews of materials and fabrication for microfluidic systems can be found elsewhere [80, 81].

Microwell-based microfluidic electroporation devices (see Microwell-E in Figure 2) typically consist of an array of microwells, with diameters in the range of 100–500 μm , assembled with microelectrodes (patterned metallic films on Si or glass substrates) [26, 27]. Input voltages of <30 V are typically used due to the small distance between electrodes, which is often referred to as microelectroporation. Despite the small working distance, the entire cell membrane is exposed to the electric field, which may compromise cell viability, e.g., up to 93% using optimal input signals [27], compared to localized electroporation methods. Owing to the array structure, these devices are often designed to be compatible with conventional microarray readers for high throughput analysis. Recently, a cell arraying-assisted electroporation (CAE) chip [26] was developed where cells can be effectively positioned into each microwell using dielectrophoretic and hydraulic forces. For selective electroporation, an array of microwells were registered with an array of independent microelectrodes such that a small population of cells in a particular microwell can be selectively electroporated.

In an effort to achieve gentler electroporation by focusing the electric field in a small region of the cell membrane, microfluidic devices using a perforated substrate (with pore diameters 0.2–2 μm) were developed [3, 73]. These perforated-substrate electroporation systems consist of a porous substrate containing nanochannels (see LEPD in Figure 2), which form the substrate for cell culture on the device. Integrated microchannels are used to load and circulate cell media for long-term cell culture on the device and/or transport of a solution with molecules to be delivered. After cells are plated into the cell culture chamber, the cells adhere to the perforated substrate and electrically seal the nanochannels such that an applied electric field is focused within the area of the cell membrane-nanochannel junction [3]. This unique feature offers effective electroporation of adhered cells in their natural *in vitro* state while maintaining high cell viability (~97%), even when transfecting sensitive cells such as embryonic stem cells and neurons [3, 28]. However, similar to an array of 1D nanostructures, cells are randomly plated on a perforated membrane or patterned microelectrodes, and therefore, achieving cell selectivity and consistent cell-nanochannel interfaces are remaining challenges.

Current and future trends in localized electroporation

Due to miniaturization and device architecture, localized electroporation requires much lower input voltage compared to its bulk counterpart and, as a result, achieves significantly improved cell viability. As discussed above, localized electroporation offers a versatile technique as it can be implemented in either nanoprobe or LOC configurations for different biological applications, i.e., in the scope of selected single cells or a population of cells, respectively. Localized electroporation technologies for single cells structurally share many similarities with microfluidic-based mechanical penetration approaches; however, localized electroporation presents unique features due to the built-in electric circuits. For example, when comparing fluidic probes used for electroporation and mechanical penetration, the former presents much greater potential in overcoming the throughput limitation by incorporating electrical feedback signals in addition to optical images. In addition to throughput, microfluidic-based mechanical penetration approaches are frequently complemented by localized electroporation to overcome their limitation on delivery efficiency. For example, 250 nm-diameter nanostraws were combined with a platform for electroporation, resulting in increased plasmid delivery efficiency from 10% to 81% into CHO cells [24, 25]. Exploiting high delivery efficiency, cell viability, and moderate experimental throughput, localized electroporation technologies show promising potential for unprecedented temporal studies of gene expression and cell phenotype on adhered cells that can provide data on a scale that is useful for systems biology analyses [82, 83].

Concluding Remarks And Future Perspectives

The vibrant, ongoing innovation, and development of probes and microfluidic systems continue to offer new capabilities for biological research, and as they are adopted into mainstream use, they will cause a paradigm shift in biological studies of adherent cells by achieving precise transfection/sampling while maintaining the cells' natural state of growth or differentiation. They also offer the potential for automation leading to hands-free cellular manipulation and analysis systems (see Outstanding Questions). In this regard, we envision that integration of micro/nanodevices with different functions, such as cell sorting and long-term culture, transfection, sampling, and biomolecule detection with single-cell specificity and high throughput (Figure 3), would provide powerful biotools for advancing applications in therapeutics, diagnostics, and drug discovery, particularly for cellular engineering involving somatic cell reprogramming, stem cell differentiation, and gene editing. As an example, these integrated microfabricated biotools could greatly advance fundamental understanding of stochastic cell signaling pathways—the link between inputs and outputs through interconnected molecular interactions—e.g., during stem cell differentiation. It is known that stem cells display stochastic behavior within pathways during differentiation due to cross talk between multiple pathways, localization of reactions, and the low concentration of molecules involved in signaling [84–86]. One way to explore such complex stochastic behavior is to integrate unique capabilities of microdevices (i.e., the *cell access and analysis module* in Figure 3) to include long-term cell culture [3], non-destructive cell access [16, 18, 25, 73], and highly sensitive sensing capabilities [87] with temporal and spatial resolution. Several other microfluidic-based methods, such as isotachopheresis [88], dielectrophoresis

[89], and bio-barcodes [90]), could also be integrated for detection and analysis of biosamples by modular assembly.

To understand complex intracellular input-output relationships and to develop mathematical descriptions of cellular behavior, it is essential to have tools for multiplexed cell manipulation and analyses, e.g., by adding cell sorting [91] and multiplexing [92] schemes to the cell access and analysis module (Figure 3), that can be performed with throughput that is statistically significant and practical with respect to research time per data point (1000 cells have been suggested as a reasonable goal for statistical relevance using single-cell technologies [2, 93]). In addition, systems biology analyses of the temporal data are needed to systematically process large sets of data from multiplexed cell analysis to elucidate the cellular behavior or mechanisms being studied [82, 83]. These integrated systems will have a significant impact on fundamental biological studies and lead to advancements in our ability to understand cell phenotypes and develop predictive analyses for engineering higher-level systems such as tissues and therapeutics [94].

Acknowledgments

H.D.E. acknowledges extensive discussions with John Kessler (Northwestern University) and Nick Melosh (Stanford University) as well as the support of the National Science Foundation under award IIP-1330151 and the National Institutes of Health under awards 1R41GM101833 and NS20778. A McCormick School of Engineering Catalyst award is also acknowledged. W.K. is grateful for the postdoctoral fellowship from the American Society for Engineering Education-Naval Research Laboratory and generous funding from the Naval Research Laboratory's Institute for Nanoscience.

REFERENCES

1. Discher DE, Janmey P, Wang YL. Tissue cells feel and respond to the stiffness of their substrate. *Science*. 2005; 310:1139–1143. [PubMed: 16293750]
2. Lindström S, Andersson-Svahn H. Miniaturization of biological assays -- overview on microwell devices for single-cell analyses. *Biochimica et biophysica acta*. 2011; 1810:308–316. [PubMed: 20451582]
3. Kang W, Giraldo-Vela JP, Nathamgari SSP, McGuire T, McNaughton RL, Kessler JA, Espinosa HD. Microfluidic device for stem cell differentiation and localized electroporation of postmitotic neurons. *Lab on a Chip*. 2014; 14:4486–4495. [PubMed: 25205561]
4. Yum K, Na S, Xiang Y, Wang N, Yu MF. Mechanochemical Delivery and Dynamic Tracking of Fluorescent Quantum Dots in the Cytoplasm and Nucleus of Living Cells. *Nano Letters*. 2009; 9:2193–2198. [PubMed: 19366190]
5. Vo-Dinh T, Alarie JP, Cullum BM, Griffin GD. Antibody-based nanoprobe for measurement of a fluorescent analyte in a single cell. *Nature Biotechnology*. 2000; 18:764–767.
6. Niu JJ, Schrlau MG, Friedman G, Gogotsi Y. Carbon Nanotube-Tipped Endoscope for In Situ Intracellular Surface-Enhanced Raman Spectroscopy. *Small*. 2011; 7:540–545. [PubMed: 21360811]
7. Vitol EA, Orynbayeva Z, Bouchard MJ, Azizkhan-Clifford J, Friedman G, Gogotsi Y. In Situ Intracellular Spectroscopy with Surface Enhanced Raman Spectroscopy (SERS)-Enabled Nanopipettes. *ACS Nano*. 2009; 3:3529–3536. [PubMed: 19891490]
8. Han S-W, Nakamura C, Kotobuki N, Obataya I, Ohgushi H, Nagamune T, Miyake J. High-efficiency DNA injection into a single human mesenchymal stem cell using a nanoneedle and atomic force microscopy. *Nanomedicine : nanotechnology, biology, and medicine*. 2008; 4:215–225.
9. Chen X, Kis A, Zettl A, Bertozzi CR. A cell nanoinjector based on carbon nanotubes. *Proceedings of the National Academy of Sciences of the United States of America*. 2007; 104:8218–8222. [PubMed: 17485677]

10. Singhal R, Orynbayeva Z, Sundaram RVK, Niu JJ, Bhattacharyya S, Vitol EA, Schrlau MG, Papazoglou ES, Friedman G, Gogotsi Y. Multifunctional carbon-nanotube cellular endoscopes. *Nature Nanotechnology*. 2011; 6:57–64.
11. Babakinejad B, Jönsson P, López Córdoba A, Actis P, Novak P, Takahashi Y, Shevchuk A, Anand U, Anand P, Drews A, Ferrer-Montiel A, Klenerman D, Korchev YE. Local delivery of molecules from a nanopipette for quantitative receptor mapping on live cells. *Analytical chemistry*. 2013; 85:9333–9342. [PubMed: 24004146]
12. Actis P, Maalouf MM, Kim HJ, Lohith A, Vilozny B, Seger RA, Pourmand N. Compartmental genomics in living cells revealed by single-cell nanobiopsy. *ACS nano*. 2014; 8:546–553. [PubMed: 24279711]
13. Anderson SE, Bau HH. Carbon nanoelectrodes for single-cell probing. *Nanotechnology*. 2015; 26:185101. [PubMed: 25876625]
14. Anderson SE, Bau HH. Electrical detection of cellular penetration during microinjection with carbon nanopipettes. *Nanotechnology*. 2014; 25:245102. [PubMed: 24859799]
15. Loh O, Lam R, Chen M, Moldovan N, Huang H, Ho D, Espinosa HD. Nanofountain-probe-based high-resolution patterning and single-cell injection of functionalized nanodiamonds. *Small*. 2009; 5:1667–1674. [PubMed: 19437464]
16. Giraldo-Vela JP, Kang W, McNaughton RL, Zhang X, Wile BM, Tsourkas A, Bao G, Espinosa HD. Single-Cell Detection of mRNA Expression Using Nanofountain-Probe Electroporated Molecular Beacons. *Small*. 2015
17. Kang W, McNaughton RL, Yavari F, Minary-Jolandan M, Safi A, Espinosa HD. Microfluidic parallel patterning and cellular delivery of molecules with a nanofountain probe. *Journal of laboratory automation*. 2014; 19:100–109. [PubMed: 23897012]
18. Kang W, Yavari F, Minary-Jolandan M, Giraldo-Vela JP, Safi A, McNaughton RL, Parpoil V, Espinosa HD. Nanofountain Probe Electroporation (NFP-E) of Single Cells. *Nano letters*. 2013
19. Meister A, Gabi M, Behr P, Studer P, Vörös J, Niedermann P, Bitterli J, Polesel-Maris J, Liley M, Heinzelmann H, Zambelli T. FluidFM: combining atomic force microscopy and nanofluidics in a universal liquid delivery system for single cell applications and beyond. *Nano letters*. 2009; 9:2501–2507. [PubMed: 19453133]
20. Guillaume-Gentil O, Potthoff E, Ossola D, Dorig P, Zambelli T, Vorholt JA. Force-Controlled Fluidic Injection into Single Cell Nuclei. *Small*. 2013; 9:1904–1907. [PubMed: 23166090]
21. Guillaume-Gentil O, Potthoff E, Ossola D, Franz CM, Zambelli T, Vorholt JA. Force-controlled manipulation of single cells: from AFM to FluidFM. *Trends in Biotechnology*. 2014; 32:381–388. [PubMed: 24856959]
22. Shalek AK, Robinson JT, Karp ES, Lee JS, Ahn D-R, Yoon M-h, Sutton A, Jorgolli M, Gertner RS, Gujral TS, MacBeath G, Yang EG, Park H, Seok J, Gyeong E. Vertical silicon nanowires as a universal platform for delivering biomolecules into living cells. *Proceedings of the National Academy of Sciences of the United States of America*. 2010; 107:1870–1875. [PubMed: 20080678]
23. Peer E, Artzy-Schnirman A, Gepstein L, Sivan U. Hollow Nanoneedle Array and Its Utilization for Repeated Administration of Biomolecules to the Same Cells. *ACS Nano*. 2012; 6:4940–4946. [PubMed: 22632128]
24. VanDersarl JJ, Xu AM, Melosh NA. Nanostraws for direct fluidic intracellular access. *Nano letters*. 2012; 12:3881–3886. [PubMed: 22166016]
25. Xie X, Xu AM, Leal-Ortiz S, Cao Y, Garner CC, Melosh Na. Nanostraw-electroporation system for highly efficient intracellular delivery and transfection. *ACS Nano*. 2013; 7:4351–4358. [PubMed: 23597131]
26. Xu YC, Yao HF, Wang L, Xing WL, Cheng J. The construction of an individually addressable cell array for selective patterning and electroporation. *Lab on a Chip*. 2011; 11:2417–2423. [PubMed: 21625729]
27. Jain T, McBride R, Head S, Saez E. Highly parallel introduction of nucleic acids into mammalian cells grown in microwell arrays. *Lab on a Chip*. 2009; 9:3557–3566. [PubMed: 20024036]

28. Fei ZZ, Wu Y, Sharma S, Gallego-Perez D, Higuera-Castro N, Hansford D, Lannutti JJ, Lee LJ. Gene Delivery to Cultured Embryonic Stem Cells Using Nanofiber-Based Sandwich Electroporation. *Analytical Chemistry*. 2013; 85:1401–1407. [PubMed: 23237665]
29. Nawarathna D, Turan T, Wickramasinghe HK. Selective probing of mRNA expression levels within a living cell. *Applied physics letters*. 2009; 95:83117–83117. [PubMed: 19777090]
30. Silberberg YR, Mieda S, Amemiya Y, Sato T, Kihara T, Nakamura N, Fukazawa K, Ishihara K, Miyake J, Nakamura C. Evaluation of the actin cytoskeleton state using an antibody-functionalized nanoneedle and an AFM. *Biosensors & Bioelectronics*. 2013; 40:3–9. [PubMed: 22784496]
31. Deladi S, Tas NR, Berenschot JW, Krijnen GJM, de Boer MJ, de Boer JH, Peter M, Elwenspoek MC. Micromachined fountain pen for atomic force microscope-based nanopatterning. *Applied Physics Letters*. 2004; 85:5361–5363.
32. Yum K, Wang N, Yu M-F. Nanoneedle: a multifunctional tool for biological studies in living cells. *Nanoscale*. 2010; 2:363–372. [PubMed: 20644817]
33. Bonde S, Buch-Manson N, Rostgaard KR, Andersen TK, Berthing T, Martinez KL. Exploring arrays of vertical one-dimensional nanostructures for cellular investigations. *Nanotechnology*. 2014; 25
34. Han S, Nakamura C, Obataya I, Nakamura N, Miyake J. Gene expression using an ultrathin needle enabling accurate displacement and low invasiveness. *Biochemical and biophysical research communications*. 2005; 332:633–639. [PubMed: 15925564]
35. Vakarelski IU, Brown SC, Higashitani K, Moudgil BM. Penetration of living cell membranes with fortified carbon nanotube tips. *Langmuir : the ACS journal of surfaces and colloids*. 2007; 23:10893–10896. [PubMed: 17894512]
36. Cuierrier CM, Lebel R, Grandbois M. Single cell transfection using plasmid decorated AFM probes. *Biochemical and biophysical research communications*. 2007; 355:632–636. [PubMed: 17316557]
37. Guerrette JP, Zhang B. Scan-Rate-Dependent Current Rectification of Cone-Shaped Silica Nanopores in Quartz Nanopipettes. *Journal of the American Chemical Society*. 2010; 132:17088–17091. [PubMed: 21073197]
38. Knoblauch M, Hibberd JM, Gray JC, van Bel AJE. A galinstan expansion femtosyringe for microinjection of eukaryotic organelles and prokaryotes. *Nature Biotechnology*. 1999; 17:906–909.
39. Freedman JR, Mattia D, Korneva G, Gogotsi Y, Friedman G, Fontecchio AK. Magnetically assembled carbon nanotube tipped pipettes. *Applied Physics Letters*. 2007; 90
40. Schrlau MG, Dun NJ, Bau HH. Cell Electrophysiology with Carbon Nanopipettes. *Acs Nano*. 2009; 3:563–568. [PubMed: 19309170]
41. Wallace EJ, Sansom MSP. Blocking of carbon nanotube based nanoinjectors by lipids: A simulation study. *Nano Letters*. 2008; 8:2751–2756. [PubMed: 18665655]
42. Kim, K-H.; Ke, C.; Moldovan, N.; Espinosa, HD. the 2003 SEM Annual Conference and Exposition on Experimental and Applied Mechanics. Charlotte, North Carolina, USA: 2003. Massively parallel multi-tip nanoscale writer with fluidic capabilities—fountain pen nanolithography (FPN); p. 235-238.
43. Safi A, Kang W, Czaplinski D, Divan R, Moldovan N, Espinosa HD. Optimization of nanofountain probe microfabrication enables large-scale nanopatterning. *Journal of Micromechanics and Microengineering*. 2013; 23
44. Liu H, Wen J, Xiao Y, Liu J, Hopyan S, Radisic M, Simmons CA, Sun Y. In situ mechanical characterization of the cell nucleus by atomic force microscopy. *ACS Nano*. 2014; 8:3821–3828. [PubMed: 24673613]
45. Na Y-R, Kim SY, Gaubblomme JT, Shalek AK, Jorgolli M, Park H, Yang EG. Probing enzymatic activity inside living cells using a nanowire-cell "sandwich" assay. *Nano letters*. 2013; 13:153–158. [PubMed: 23244056]
46. Shalek AK, Gaubblomme JT, Wang L, Yosef N, Chevrier N, Andersen MS, Robinson JT, Pochet N, Neuberger D, Gertner RS, Amit I, Brown JR, Hacohen N, Regev A, Wu CJ, Park H. Nanowire-mediated delivery enables functional interrogation of primary immune cells: application to the

analysis of chronic lymphocytic leukemia. *Nano letters*. 2012; 12:6498–6504. [PubMed: 23190424]

47. Robinson JT, Jorgolli M, Shalek AK, Yoon MH, Gertner RS, Park H. Vertical nanowire electrode arrays as a scalable platform for intracellular interfacing to neuronal circuits. *Nature Nanotechnology*. 2012; 7:180–184.
48. Xie C, Lin ZL, Hanson L, Cui Y, Cui BX. Intracellular recording of action potentials by nanopillar electroporation. *Nature Nanotechnology*. 2012; 7:185–190.
49. Yan L, Yang Y, Zhang WJ, Chen XF. Advanced Materials and Nanotechnology for Drug Delivery. *Advanced Materials*. 2014; 26:5533–5540. [PubMed: 24449177]
50. Gao HJ. Probing mechanical principles of cell-nanomaterial interactions. *Journal of the Mechanics and Physics of Solids*. 2014; 62:312–339.
51. Xu AM, Aalipour A, Leal-Ortiz S, Mekhdjian AH, Xie X, Dunn AR, Garner CC, Melosh NA. Quantification of nanowire penetration into living cells. *Nature Communications*. 2014; 5
52. Biswas A, Bayer IS, Biris AS, Wang T, Dervishi E, Faupel F. Advances in top-down and bottom-up surface nanofabrication: Techniques, applications & future prospects. *Advances in Colloid and Interface Science*. 2012; 170:2–27. [PubMed: 22154364]
53. Persson H, Kobler C, Molhave K, Samuelson L, Tegenfeldt JO, Oredsson S, Prinz CN. Fibroblasts Cultured on Nanowires Exhibit Low Motility, Impaired Cell Division, and DNA Damage. *Small*. 2013; 9:4006–4016. [PubMed: 23813871]
54. Hanson L, Lin ZC, Xie C, Cui Y, Cui B. Characterization of the cell-nanopillar interface by transmission electron microscopy. *Nano Letters*. 2012; 12:5815–5820. [PubMed: 23030066]
55. Xie X, Xu AM, Angle MR, Tayebi N, Verma P, Melosh NA. Mechanical model of vertical nanowire cell penetration. *Nano Letters*. 2013; 13:6002–6008. [PubMed: 24237230]
56. Qi SJ, Yi CQ, Ji SL, Fong CC, Yang MS. Cell Adhesion and Spreading Behavior on Vertically Aligned Silicon Nanowire Arrays. *Acs Applied Materials & Interfaces*. 2009; 1:30–34. [PubMed: 20355748]
57. van Engeland M, Nieland LJW, Ramaekers FCS, Schutte B, Reutelingsperger CPM. Annexin V-affinity assay: A review on an apoptosis detection system based on phosphatidylserine exposure. *Cytometry*. 1998; 31:1–9. [PubMed: 9450519]
58. Hoffman BD, Crocker JC. Cell Mechanics: Dissecting the Physical Responses of Cells to Force. *Annual Review of Biomedical Engineering*. 2009; 11:259–288.
59. Neumann E, Schaefferidder M, Wang Y, Hofschneider PH. Gene-Transfer into Mouse Lyoma Cells by Electroporation in High Electric-Fields. *Embo Journal*. 1982; 1:841–845. [PubMed: 6329708]
60. Geng T, Lu C. Microfluidic electroporation for cellular analysis and delivery. *Lab on a chip*. 2013; 13:3803–3821. [PubMed: 23917998]
61. Kim SH, Yamamoto T, Fourmy D, Fujii T. Electroactive Microwell Arrays for Highly Efficient Single-Cell Trapping and Analysis. *Small*. 2011; 7:3239–3247. [PubMed: 21932278]
62. Zhan Y, Martin Va, Geahlen RL, Lu C. One-step extraction of subcellular proteins from eukaryotic cells. *Lab on a chip*. 2010; 10:2046–2048. [PubMed: 20548993]
63. Zhan Y, Sun C, Cao Z, Bao N, Xing J, Lu C. Release of intracellular proteins by electroporation with preserved cell viability. *Analytical chemistry*. 2012; 84:8102–8105. [PubMed: 22974138]
64. Weaver JC, Chizmadzhev YA. Theory of electroporation: A review. *Bioelectrochemistry and Bioenergetics*. 1996; 41:135–160.
65. Krassowska W, Filev PD. Modeling electroporation in a single cell. *Biophysical Journal*. 2007; 92:404–417. [PubMed: 17056739]
66. Teissie J, Rols MP. An Experimental Evaluation of the Critical Potential Difference Inducing Cell-Membrane Electroporation. *Biophysical Journal*. 1993; 65:409–413. [PubMed: 8369446]
67. Gabriel B, Teissie J. Time courses of mammalian cell electroporation observed millisecond imaging of membrane property changes during the pulse. *Biophysical Journal*. 1999; 76:2158–2165. [PubMed: 10096909]
68. Chen C, Smye SW, Robinson MP, Evans JA. Membrane electroporation theories: a review. *Medical & Biological Engineering & Computing*. 2006; 44:5–14. [PubMed: 16929916]

69. Escoffre JM, Portet T, Wasungu L, Teissie J, Dean D, Rols MP. What is (Still not) Known the Mechanism by Which Electroporation Mediates Gene Transfer and Expression in Cells and Tissues. *Molecular Biotechnology*. 2009; 41:286–295. [PubMed: 19016008]
70. Brown RB, Audet J. Current techniques for single-cell lysis. *Journal of the Royal Society Interface*. 2008; 5:S131–S138.
71. Olofsson J, Nolkranz K, Ryttsén F, Lambie Ba, Weber SG, Orwar O. Single-cell electroporation. *Current Opinion in Biotechnology*. 2003; 14:29–34. [PubMed: 12565999]
72. Wang M, Orwar O, Olofsson J, Weber SG. Single-cell electroporation. *Analytical and bioanalytical chemistry*. 2010; 397:3235–3248. [PubMed: 20496058]
73. Chang LQ, Bertani P, Gallego-Perez D, Yang ZG, Chen F, Chiang CL, Malkoc V, Kuang TR, Gao K, Lee LJ, Lu W. 3D nanochannel electroporation for high-throughput cell transfection with high uniformity and dosage control. *Nanoscale*. 2016; 8:243–252. [PubMed: 26309218]
74. Khine M, Lau A, Ionescu-Zanetti C, Seo J, Lee LP. A single cell electroporation chip. *Lab on chip*. 2005; 5:38–43.
75. Khine M, Ionescu-Zanetti C, Blatz A, Wang L-P, Lee LP. Single-cell electroporation arrays with real-time monitoring and feedback control. *Lab on a chip*. 2007; 7:457–462. [PubMed: 17389961]
76. Boukany PE, Morss A, Liao W-C, Henslee B, Jung H, Zhang X, Yu B, Wang X, Wu Y, Li L, Gao K, Hu X, Zhao X, Hemminger O, Lu W, Lafyatis GP, Lee LJ. Nanochannel electroporation delivers precise amounts of biomolecules into living cells. *Nature Nanotechnology*. 2011; 6:747–754.
77. Sakaki K, Dechev N, Burke RD, Park EJ. Development of an Autonomous Biological Cell Manipulator With Single-Cell Electroporation and Visual Servoing Capabilities. *Ieee Transactions Biomedical Engineering*. 2009; 56:2064–2074.
78. Iwata F, Yamazaki K, Ishizaki K, Ushiki T. Local electroporation of a single cell using a scanning ion conductance microscope. *Japanese Journal of Applied Physics*. 2014; 53
79. Melin J, Quake SR. Microfluidic large-scale integration: The evolution of design rules for biological automation. *Annual Review of Biophysics and Biomolecular Structure*. 2007; 36:213–231.
80. van Midwoud PM, Janse A, Merema MT, Groothuis GMM, Verpoorte E. Comparison of Biocompatibility and Adsorption Properties of Different Plastics for Advanced Microfluidic Cell and Tissue Culture Models. *Analytical Chemistry*. 2012; 84:3938–3944. [PubMed: 22444457]
81. Ren KN, Zhou JH, Wu HK. Materials for Microfluidic Chip Fabrication. *Accounts of Chemical Research*. 2013; 46:2396–2406. [PubMed: 24245999]
82. Bagheri N, Shiina M, Lauffenburger DA, Korn WM. A Dynamical Systems Model for Combinatorial Cancer Therapy Enhances Oncolytic Adenovirus Efficacy by MEK-Inhibition. *Plos Computational Biology*. 2011; 7
83. Han Q, Bagheri N, Bradshaw EM, Hafler DA, Lauffenburger DA, Love JC. Polyfunctional responses by human T cells result from sequential release of cytokines. *Proceedings of the National Academy of Sciences of the United States of America*. 2012; 109:1607–1612. [PubMed: 22160692]
84. Boyer LA, Lee TI, Cole MF, Johnstone SE, Levine SS, Zucker JR, Guenther MG, Kumar RM, Murray HL, Jenner RG, Gifford DK, Melton DA, Jaenisch R, Young RA. Core transcriptional regulatory circuitry in human embryonic stem cells. *Cell*. 2005; 122:947–956. [PubMed: 16153702]
85. Bernstein BE, Mikkelsen TS, Xie XH, Kamal M, Huebert DJ, Cuff J, Fry B, Meissner A, Wernig M, Plath K, Jaenisch R, Wagschal A, Feil R, Schreiber SL, Lander ES. A bivalent chromatin structure marks key developmental genes in embryonic stem cells. *Cell*. 2006; 125:315–326. [PubMed: 16630819]
86. Xu Y, Shi Y, Ding S. A chemical approach to stem-cell biology and regenerative medicine. *Nature*. 2008; 453:338–344. [PubMed: 18480815]
87. Eid C, Palko JW, Katilius E, Santiago JG. Rapid Slow Off-Rate Modified Aptamer (SOMAmer)-Based Detection of C-Reactive Protein Using Isotachopheresis and an Ionic Spacer. *Analytical Chemistry*. 2015; 87:6736–6743. [PubMed: 26024067]

88. Kaigala GV, Bercovici M, Behnam M, Elliott D, Santiago JG, Backhouse CJ. Miniaturized system for isotachopheresis assays. *Lab on a chip*. 2010; 10:2242–2250. [PubMed: 20571691]
89. Gong JR. Label-Free Attomolar Detection of Proteins Using Integrated Nanoelectronic and Electrokinetic Devices. *Small*. 2010; 6:967–973. [PubMed: 20209654]
90. Fan R, Vermesh O, Srivastava A, Yen BKH, Qin L, Ahmad H, Kwong Ga, Liu C-C, Gould J, Hood L, Heath JR. Integrated barcode chips for rapid, multiplexed analysis of proteins in microliter quantities of blood. *Nature Biotechnology*. 2008; 26:1373–1378.
91. Nathamgari SSP, Dong BQ, Zhou F, Kang WM, Giraldo-Vela JP, McGuire T, McNaughton RL, Sun C, Kessler JA, Espinosa HD. Isolating single cells in a neurosphere assay using inertial microfluidics. *Lab on a Chip*. 2015; 15:4591–4597. [PubMed: 26511875]
92. Unger MA, Chou HP, Thorsen T, Scherer A, Quake SR. Monolithic microfabricated valves and pumps by multilayer soft lithography. *Science*. 2000; 288:113–116. [PubMed: 10753110]
93. Svahn HA, van den Berg A. Single cells or large populations? *Lab on a chip*. 2007; 7:544–546. [PubMed: 17476370]
94. Di Carlo D, Lee LP. Dynamic single-cell analysis for quantitative biology. *Analytical Chemistry*. 2006; 78:7918–7925. [PubMed: 17186633]
95. Varadarajan N, Kwon DS, Law KM, Ogunniyi AO, Anahtar MN, Richter JM, Walker BD, Love JC. Rapid, efficient functional characterization and recovery of HIV-specific human CD8+ T cells using microengraving. *Proceedings of the National Academy of Sciences of the United States of America*. 2012; 109:3885–3890. [PubMed: 22355106]
96. Lee CC, Sui GD, Elizarov A, Shu CYJ, Shin YS, Dooley AN, Huang J, Daridon A, Wyatt P, Stout D, Kolb HC, Witte ON, Satyamurthy N, Heath JR, Phelps ME, Quake SR, Tseng HR. Multistep synthesis of a radiolabeled imaging probe using integrated microfluidics. *Science*. 2005; 310:1793–1796. [PubMed: 16357255]
97. Xue M, Wei W, Su YP, Kirn J, Shin YS, Mai WX, Nathanson DA, Heath JR. Chemical Methods for the Simultaneous Quantitation of Metabolites and Proteins from Single Cells. *Journal of the American Chemical Society*. 2015; 137:4066–4069. [PubMed: 25789560]
98. Uehara H, Kunitomi Y, Ikai A, Osada T. mRNA detection of individual cells with the single cell nanoprobe method compared with in situ hybridization. *J Nanobiotechnology*. 2007; 5:7. [PubMed: 17925043]

Box 1: Force control measurement

AFMs allow quantification of the force signature associated with the interaction between a probe tip and cell during mechanical penetration of the cell membrane [8, 9, 34, 35]. This quantitative force measurement provides key mechanical information including the onset of cell-tip contact, penetration of the cell membrane, and dissociation—all signatures that can, in principle, provide a method for computer-aided operation for repeatable and less user-dependent protocols. Pioneering single-cell injection using an AFM was first performed with conventional AFM tips [36, 98], whereas recent work has focused on using cantilevers with 1D nanostructure tips [8, 9, 34, 35] fabricated using either a bottom-up or top-down approach, i.e., assembly of an AFM cantilever with a nanostructure within a scanning electron microscope [9, 35] or fabrication of a nanoprobe tip via etching techniques [8, 34], respectively. Force-controlled nanoinjection allows reduced mechanical force on the nanostructure tip (e.g., <200 pN force or <200 nm cell deformation [35] compared to 1–3 nN or >1 μm using conventional pyramid-shaped AFM tips [36]), significantly increasing the success of nanoinjection and likelihood of cell recovery after injection. In addition, this technique allows for direct insertion of molecules into the nucleus, due to the nanoscale tip size and spatial resolution, resulting in higher efficiency (about 74% GFP expression) than other conventional approaches including lipofection (~50%) and microinjection (~10%) [8]. With increasing application of AFM to biological studies, conventional AFM systems are now frequently integrated with an inverted fluorescence microscope for simultaneous optical monitoring of cells, providing real-time imaging of tip insertion into a cell.

Box 2: Localized electroporation with probe-cell membrane proximity detection

The nanofountain probe electroporation (NFP-E) system: By packaging batch-fabricated nanofountain probes with a fluidic circuit containing a wire electrode, electrical resistance measurement, local membrane nanoporation, and electrophoretic transport of charged biomolecules is achieved. The electrical resistance measurement enables accurate probe-cell membrane proximity detection, which is ideal for automation and throughput. In this configurations, the molecules to be delivered are loaded into the low-volume NFP chip, and cells are cultured in standard Petri dishes or in arrays of microwells or stamped matrix proteins. The arrayed cell protocol provides a multiplexing capability to conduct many experiments of interest in parallel. The NFP-E has proven to be a versatile technique for molecular delivery to many cell types including immortalized cell lines, stem cells, and immune cells [16, 18]. Successful delivery of proteins, DNA and RNA hairpin molecules, and plasmids have been achieved with high rates of efficiency and viability (>90% for small molecules) into the cytoplasm or the nucleus of the various cells, indicating the broad applicability of the NFP-E system as a robust and versatile biotool [18].

Outstanding Box

- **How can high throughput transfection and analysis be achieved in the context of single-cell studies?** Current micro/nanotechnology-based biotools generally suffer from limited throughput, which must be addressed by adopting automation and/or multiplexing strategies for comprehensive and statistically relevant studies of biological complexity such as cell heterogeneity, cell differentiation, and disease mechanisms.
- **How can micro/nanotechnology-based biotools be used to provide unprecedented capabilities for novel biological studies and medical applications?** We foresee that micro/nanotechnology-based biotools, which offer precise cell transfection and non-destructive biomolecular analysis, will enable new approaches to address challenges in probing biological processes of cell reprogramming and stem cell differentiation, and eventually, the generation of personalized diagnostics and cell therapeutics.

Trends Box

Miniaturization of biotools

Micro/nano-scale biotools for cell transfection and analysis are being developed to achieve cell-specific experimental capabilities and localized cell-tool interfaces. This allows for minimal perturbation to cells and unprecedented spatial resolution, which is essential for fundamental cell studies.

Biotools for adherent cells

Despite the risk for altering phenotype or stressing cells, conventional biotechnologies often require suspending and replating cells during *in vitro* studies. Novel micro/nano-technologies are being developed to transfect and analyze adhered cells, which is particularly advantageous for longitudinal studies of individual cells or investigating cell mechanisms.

Combination of micro/nano-technologies with conventional biotechnologies

Various strategies use micro/nano-technologies with conventional analytical tools such as fluorescence array readers and atomic force microscopes. This assembly approach promises revolutionary advances in biology and medicine.

Figure 1.

Mechanical penetration. For delivery of molecules into cells, the membrane of a target cell can be mechanically pierced using a needle-like structure with a sharp tip including (counterclockwise) nanopipettes [5], one-dimensional nanowires [4], or cantilevers with either 1D nanoscale tips [8] or embedded microchannels [15, 19]. These figures are reproduced with permission of Macmillan Publishers Ltd, American Chemical Society, Elsevier, and American Chemical Society, respectively.

Figure 2.

Electroporation. When a cell is subjected to a sufficiently large electric field, transient nanopores are formed in the cell membrane through which molecules can be delivered into the cell. Electroporation has been used in both lab-on-a-chip (left column) and nanoprobe (right column) configurations including (counterclockwise) localized electroporation device (LEPD), and microwell-, nanopipette-, and nanofountain-electroporation (Microwell-E, Nanopipette-E, and NFP-E), respectively. These figures are reproduced from [3] and [26], [78], and [18] with permission of The Royal Society of Chemistry, The Japan Society of Applied Physics, and American Chemical Society, respectively.

Figure 3.

Example of envisioned integrated micro/nanofluidic platform for transfection, sampling, biomolecular detection, sorting, and on-chip cell culture [3, 15, 35, 88, 91, 95–97]. Through temporal analysis, intracellular processes leading to mechanistic understanding through systems biology is possible [82, 83]. (Reproduced from [88] and [91], [35] and [97], [96], and [83] with permission of The Royal Society of Chemistry, American Chemical Society, The American Association for the Advancement of Science, and the National Academy of Sciences, respectively)

Table 1

Micro- and nano- technologies for cell transfection and analysis of adherent cells

Mechanism	Fabricated Device	Cells	Use Demonstrated	Comments	Reference
Non-fluidic Nanoprobes	1D probes	HeLa	Delivery of fluorescent QD [†]	Single cell; position controlled by a manipulator	[4]
		Rat liver epithelial cell	Detection of benzopyrene tetrol	Single cell; manipulator	[5]
		HeLa	Detection of trace of intracellular molecules using SERS	Single cell; manipulator	[6, 7]
Fluidic Nanoprobes	Cantilever	MSC, HEK293	Delivery of DNA	Single cell; AFM; force measurement	[8]
		HeLa	Delivery of QD	Single cell; AFM	[9]
	1D probes	HeLa, HOS	Delivery of fluorescent particles, aspiration of Ca ²⁺ -labeled cytosol	Single cell; manipulator; SERS and electrochemical measurement	[10, 11]
		DRG neurons	Delivery of capsaicin to the surface of neurons	Single cell; position feedback control using SICM [‡]	[11]
		Human fibroblast, HeLa	Aspiration of RNA samples	Single cell; position feedback control using SICM	[12]
		U2OS	Delivery of iRNA	Single cell; detection for penetration of cell and nuclear membrane; manipulator	[13, 14]
		HeLa, HT1080	Delivery of dextran, bovine serum albumin, RNA- /DNA-based MB ^Δ , DNA plasmids, and nanoparticles	Single cell; localized electroporation or injection; AFM or manipulator; force measurement	[15–18]
	Cantilever	Myoblast, neuroblastoma, HeLa	Delivery of Lucifer yellow CH, DNA, FITC	Single cell; AFM; force measurements; cell manipulation	[19–21]

Mechanism	Fabricated Device	Cells	Use Demonstrated	Comments	Reference
Arrays 1D-vertical structures	Non-fluidics	HeLa, fibroblast, NPC, neurons	Delivery of siRNA, peptide, DNA, protein	Population of cells; NW-induced delivery	[22]
	Fluidics	Hollow nano needle array (500 nm)	Delivery of dextran, DNA	Population of cell; saponin, cell membrane permeation promoter, was used	[23]
		Nanostraw (d=100~750 nm)	Delivery of Ion, PI **, plasmid DNA	Population of cells; nanostraw-induced delivery or localized electroporation	[24, 25]
Lab-on-a-chip platforms	Microwell	HeLa	Delivery of PI, DNA	Population of cells; EP for cell seeding; electroporation of each well	[26]
		HEK 293T, primary mouse macrophages	Delivery of PI, siRNA, DNA	Population of cells	[27]
	Perforated-substrate	HeLa, HT1080, neurons	Delivery of PI, DNA plasmids	Population of cells; long-term cell culture; on-chip differentiation; localized electroporation	[3]
		Mouse embryonic stem cells	Delivery of DNA	Population of cells	[28]

SERS*: surface enhanced Raman Spectroscopy, PI**: propidium iodide, QD[†]: quantum dot, SICM[‡]: scanning ion channel microscopy, EP : electrophoretic force, MB[§]: Molecular beacon