Effect of sodium azide on EIA (Corzyme)

Sera tested with sodium azide	Corab (RIA)	Corzyme (EIA)
1	+	+
2	+	+
3	+	+
4	—	+
5	+	+
6	—	+
7	—	+
8	—	+

RK CHAUDHARY
C PARKER

*Viral Hepatitis Section
Laboratory of the National Hepatitis
Reference Centre
Viral Diagnostic Services Division
Bureau of Microbiology
Laboratory Centre for Disease Control
Health and Welfare Canada
Ottawa, Ontario Canada*

Reference

- Richardson TC, Chapman DV, Heyderman E. Immunoperoxidase techniques: the deleterious effect of sodium azide on the activity of peroxidase conjugates. *J Clin Pathol* 1983;36:411-4.

Quality assurance and medical audit in histopathology

In the Bulletin of the Royal College of Pathologists,¹ there is a report under this heading that "Council has given consideration to the means for encouraging quality assurance and medical audit in histopathology. It recommends that histopathologists should form 'slide-clubs' at which material of interest and importance can be studied. Experts should be invited to discuss specific topics and exchanges of material between different slide-clubs encouraged. The geographical distribution of these clubs should be such that all histopathologists should be able to participate on a regular basis."

Is this really the sum of what Council has achieved in its discussion of this important issue? The suggestion that histopathologists should form slide-clubs certainly lacks originality, but in any case the way in which these are advised to function can in no way be regarded as a form of quality assurance or medical audit. What constitutes medical audit in histopathology is perhaps debatable² but there is a wide range of options available for both macro-

scopic and microscopic work (see Bibliography).

BC MORSON

*Department of Pathology,
St Mark's Hospital,
City Road, London ECV1 2PS*

References

- Bulletin of the Royal College of Pathologists. April 1983; No. 42, 9.
- Morson BC, Frei JV. Letter *J Clin Pathol* 1982;35:476.

Bibliography

- Cocker J, Fox H, Langley FA. Consistency in the histological diagnosis of epithelial abnormalities of the cervix uteri. *J Clin Pathol* 1968;21:67-70.
- Klionsky B. Editorial statement of the ad hoc committee on reproducibility of histopathological diagnosis of uterine cervical lesions. *Am J Clin Pathol* 1969;52:358-9.
- Iversen OH, Sandnes K. The reliability of pathologists. *Acta Pathol Scand* 1971;79A:330-4.
- Owen DA, Tighe JR. Quality evaluation in histopathology. *Br Med J* 1975;i:149-50.
- Sissons HA. Agreement and disagreement between pathologists in histological diagnosis. *Postgrad Med J* 1975;51:685-9.
- Henson DE, Codling BW, Macartney JC. *Interlaboratory histological evaluation: a new approach to quality control in anatomic pathology*. College of American Pathologists, 1976 (65 pp).
- Langley FA. Quality control in histopathology and diagnostic cytology. *Histopathology* 1978;2:3-18.
- Ringsted J, Amtrup F, Asklund C, et al. Reliability of histopathological diagnosis of squamous epithelial changes of the uterine cervix. *Acta Pathol Microbiol Scand* 1978;86:273-8.
- Hutter RVP, et al. Patterns of care study: improving cancer care by assuring the quality of pathology reports. American College of Pathologists: *Pathologist* 1979;33:536-8.
- Blenkinsopp WK, Stewart-Brown S, Blesovsky L, Kearney G, Fielding LP. Histopathology reporting in large bowel cancer. *J Clin Pathol* 1981;34:509-13.
- Morson BC. Histopathology reporting in large bowel cancer. *Br Med J* 1981;283:1493.
- Schwartz WB, Wolfe HJ, Panker SG. Pathology and probabilities. *N Engl J Med* 1981;305:917-23.
- Letters responding to the above. *N Engl J Med* 1981;306:303-5.
- Barr WT, Williams ED. Value of external quality assessment of the technical aspects of histopathology. *J Clin Pathol* 1982;35:1050-6.
- Frei JV, Morson BC. Medical audit of rectal biopsy diagnosis of inflammatory bowel disease. *J Clin Pathol* 1982;35:341-4.
- Derman H. Histopathology reporting in large bowel cancer. *Br Med J* 1982;284:346.

- Thomas GDH, Dixon MF, Smeeton NC, Williams NS. Observer variation in the histological grading of rectal carcinoma. *J Clin Pathol* 1983;36:385-91.

- Holman CDJ, Matz LR, Finlay Jones LR, et al. Interobserver variation in the histopathological reporting of Hodgkin's disease: an analysis of diagnostic subcomponents using kappa statistics. *Histopathology* 1983;7:399-407.

Confusion of terms "birefringence" and "optical activity"

There is a tendency for authors of histopathological studies to use the term "birefringence" when "optical activity" is intended. The first term means double refraction or having two refractive indices, which is shown by passing a beam of light through a crystalline solid and observing that the beam is split into two rays. Optical activity is a different property which applies to some solids and liquids and indicates that a beam of polarised light is rotated when passing through them. This is the phenomenon studied in tissue sections when examining for foreign bodies, crystals, amyloid etc.

The apparent confusion between these two properties of matter possibly arises because each of the two rays issuing from a birefringent crystal—for example, a Nicol prism—consist of polarised light and may be used as a source of such light.

PN COWEN

*Department of Pathology,
University of Leeds,
Leeds LS2 9JT*

Platelet storage in PL 146, CLX and Biotest 733822 Plastics

The Oxford Transfusion Centre recently confirmed the superiority of the PL 732 (Travenol Laboratories Ltd) polyolefin packs over the conventional PL 146 packs for platelet storage over five days.¹ Since the PL 732 packs are now no longer available in the UK we have now tested two other packs designed for extended platelet storage; CLX from Cutter Ltd and 733822 a new pack from Biotest-Folex Ltd.

Whole blood (450 ml) was collected from 31 healthy donors who had taken no drugs during the previous seven days. All donations were taken into CPD anticoagulant, 7 into Travenol PL 146 packs serving as controls, 12 into Biotest-Folex 733822 packs and 12 into Cutter CLX packs. Each platelet concentrate