Video Article

Designing Microfluidic Devices for Studying Cellular Responses Under Single or Coexisting Chemical/Electrical/Shear Stress Stimuli

Tzu-Yuan Chou¹, Yung-Shin Sun², Hsien-San Hou³, Shang-Ying Wu¹, Yun Zhu¹, Ji-Yen Cheng³, Kai-Yin Lo¹

¹Department of Agricultural Chemistry, National Taiwan University

²Department of Physics, Fu-Jen Catholic University

³Research Center for Applied Sciences, Academia Sinica

Correspondence to: Kai-Yin Lo at kaiyin@ntu.edu.tw

URL: <http://www.jove.com/video/54397>

DOI: [doi:10.3791/54397](https://doi.org/10.3791/54397)

Keywords: Bioengineering, Issue 114, microfluidic chip, cell migration, reactive oxygen species, electric field, electrotaxis, shear stress

Date Published: 8/13/2016

Citation: Chou, T.Y., Sun, Y.S., Hou, H.S., Wu, S.Y., Zhu, Y., Cheng, J.Y., Lo, K.Y. Designing Microfluidic Devices for Studying Cellular Responses Under Single or Coexisting Chemical/Electrical/Shear Stress Stimuli. *J. Vis. Exp.* (114), e54397, doi:10.3791/54397 (2016).

Abstract

Microfluidic devices are capable of creating a precise and controllable cellular micro-environment of pH, temperature, salt concentration, and other physical or chemical stimuli. They have been commonly used for *in vitro* cell studies by providing *in vivo* like surroundings. Especially, how cells response to chemical gradients, electrical fields, and shear stresses has drawn many interests since these phenomena are important in understanding cellular properties and functions. These microfluidic chips can be made of glass substrates, silicon wafers, polydimethylsiloxane (PDMS) polymers, polymethylmethacrylate (PMMA) substrates, or polyethyleneterephthalate (PET) substrates. Out of these materials, PMMA substrates are cheap and can be easily processed using laser ablation and writing. Although a few microfluidic devices have been designed and fabricated for generating multiple, coexisting chemical and electrical stimuli, none of them was considered efficient enough in reducing experimental repeats, particular for screening purposes. In this report, we describe our design and fabrication of two PMMA-based microfluidic chips for investigating cellular responses, in the production of reactive oxygen species and the migration, under single or coexisting chemical/ electrical/shear stress stimuli. The first chip generates five relative concentrations of 0, 1/8, 1/2, 7/8, and 1 in the culture regions, together with a shear stress gradient produced inside each of these areas. The second chip generates the same relative concentrations, but with five different electric field strengths created within each culture area. These devices not only provide cells with a precise, controllable micro-environment but also greatly increase the experimental throughput.

Video Link

The video component of this article can be found at <http://www.jove.com/video/54397/>

Introduction

In vivo cells are surrounded by a variety of biomolecules including extracellular matrix (ECM), carbohydrates, lipids, and other cells. They functionalize by responding to micro-environmental stimuli such as interactions with ECM and responses to chemical gradients of various growth factors. Traditionally, *in vitro* cell studies are conducted in cell culture dishes where the consumption of cells and reagents is large and cells grow in a static (non-circulating) environment. Recently, micro-fabricated devices integrated with fluidic components have provided an alternative platform for cell studies in a more controllable way. Such devices are capable of creating a precise micro-environment of chemical and physical stimuli while minimizing the consumption of cells and reagents. These microfluidic chips can be made of glass substrates, silicon wafers, polydimethylsiloxane (PDMS) polymers, polymethylmethacrylate (PMMA) substrates, or polyethyleneterephthalate (PET) substrates¹⁻³. PDMS-based devices are transparent, biocompatible, and permeable to gases, making them suitable for long-term cell culture and studies. PMMA and PET substrates are cheap and easy to be processed using laser ablation and writing.

Microfluidic devices should provide cells with a stable and controllable micro-environment where cells are subject to different chemical and physical stimuli. For example, microfluidic chips are used to study chemotaxis of cells. Instead of traditional methods that employ Boyden chamber and capillary^{4,5} these miniaturized fluidic devices can generate precise chemical gradients for studying cells' behaviors^{1,6,7}. Another example is to study cells' directional migration under electric fields (EFs), a phenomenon named electrotaxis. Electrotactic behaviors of cells were reported to be related to nerve regeneration⁸, embryonic development⁹, and wound healing^{10,11}. And many studies have been performed to investigate the electrotaxis of various cell types including cancer cells^{12,13}, lymphocytes^{14,15}, leukemia cells¹¹, and stem cells¹⁶. Conventionally, Petri dishes and cover glasses are used to construct electrotactic chambers for generating EFs¹⁷. Such simple setups pose problems of medium evaporation and imprecise EFs, but they can be overcome by microfluidic devices of enclosed, well-defined fluidic channels^{12,18,19}.

To systematically study cellular responses under precise, controllable chemical and electrical stimuli, it would be of great use to develop microfluidic devices capable of providing cells with multiple stimuli at the same time. For example, Li *et al.* reported a PDMS-based microfluidic device for creating single or coexisting chemical gradients and EFs²⁰. Kao *et al.* developed a similar microfluidic chip to modulate the chemotaxis of lung cancer cells by EFs⁶. Moreover, to increase the throughput, Hou *et al.* designed and fabricated a PMMA-based multichannel-dual-

electric-field chip to provide cells with 8 different combined stimuli, being (2 EF strengths x 4 chemical concentrations)²¹. To further increase the throughput and add the shear stress stimulus, we developed two PMMA-based microfluidic devices for studying cellular responses under single or coexisting chemical/electrical/shear stress stimuli.

Reported by Lo *et al.*^{22,23}, these devices contain five independent cell culture channels subject to continuous fluidic flowing, mimicking the *in vivo* circulatory system. In the first chip (the chemical-shear stress chip or the CSS chip), five relative concentrations of 0, 1/8, 1/2, 7/8, and 1 are generated in the culture regions, and a shear stress gradient is produced inside each of the five culture areas. In the second chip (the chemical-electric field chip or the CEF chip), by using one single set of electrodes and 2 syringe pumps, five EF strengths are generated in addition to five different chemical concentrations within these culture areas. Numerical calculations and simulations are performed to better design and operate these chips, and lung cancer cells cultured inside these devices are subject to single or coexisting stimuli for observing their responses with respect to the production of reactive oxygen species (ROS), the migration rate, and the migration direction. These chips are demonstrated to be time-saving, high-throughput and reliable devices for investigating how cells respond to various micro-environmental stimuli.

Protocol

1. Chip Design and Fabrication

1. Draw patterns to be ablated on PMMA substrates and double-side tapes using commercial software²⁴.
 1. To study the effects of chemical concentrations and shear stresses, draw a "Christmas tree" pattern with a varying width at its end in each of the five culture areas (**Figure 1A** and **1B**).
 2. To study the effects of chemical concentrations and electric fields, draw a "Christmas tree" pattern with two more fluidic channels for salt bridges (**Figure 2A** and **2B**).
2. Scribe an individual pattern on a PMMA sheet or a double-side tape by loading the corresponding file into the CO₂ laser scriber.
 1. Turn on the laser scriber, and check its connection to the computer. Open the pattern to be ablated using the commercial software.
 2. Place a PMMA sheet or a double-sided tape on top of the stage of the scriber. Adjust the focus of the CO₂ laser if necessary using the calibration bar and the visible He-Ne laser.
 3. Load the pattern into the scriber for ablation on the PMMA sheet or the tape.
 4. Pick up the patterned sheet or tape, remove unwanted pieces, and clean the surface with nitrogen blowing.

NOTE: The thickness of the PMMA sheet is 1 mm, and that of the double-sided tape is 0.07 mm or 0.22 mm.

2. Chip Assembly

1. With super glue, attach acrylic adaptors (length x width x height = 10 mm x 10 mm x 5 mm, with screw threads in the middle) to the top-most layer of the chip by aligning the screw threads and the holes on the top-most layer. These adaptors serve as medium inlets/outlets and salt bridge connectors.
2. Assemble the microfluidic chip inside a laminar flow hood.
3. Assemble the chemical-shear stress microfluidic chip (CSS chip).
 1. Attach 3 sheets of scribed PMMA sheets using 2 pieces of scribed double-sided tape. (**Figure 1A** and **1B**).
 2. Add one more piece of the 0.07 mm thick double sided tape to the bottom of the chip and attach the chip to a 10 cm diameter Petri dish.
 3. Put the assembly chip in the vacuum chamber for overnight.
4. To assemble the chemical-electric field chip (CEF chip), attach the PMMA sheet to a double-sided tape of thickness = 0.22 mm (**Figure 2A** and **2B**).
 1. Attach the scribed PMMA sheet and double-sided tape to a 10 cm diameter Petri dish using this same piece of tape.
5. Leave the assembled chip inside the hood and expose it to UV for 30 min for sterilization.
6. Connect the inlets to two 3-ml syringes via plastic tubes and finger-tight nuts. Connect the outlet to a waste tube via a plastic tube and a finger-tight nut.

NOTE: Autoclave all tubes and nuts at 121 °C for 15 min prior to usage.
7. Slowly depress the plunger on the syringe to prime the fluidic channels with PBS. Push syringes back and forth to remove bubbles.
8. Put the chips inside an incubator overnight at 37 °C under 5% CO₂.

NOTE: These two steps are aimed to wash the chip and remove any remaining bubbles within the chip.
9. Take the chip out of the incubator.
10. Prepare agar salt bridges for generating electric fields in the CEF chip.
 1. Dissolve 3% low melting point agarose in 1x phosphate buffered saline (PBS) buffer using a microwave.
 2. Inject solution-phased agarose into the salt bridge channel and insert the electrodes before solution solidifies.
 3. Place silver/silver-chloride electrodes into the tubular nuts.
11. Connect the syringes to the syringe pumps, and continuously flow the culture medium (Dulbecco's Modified Eagle's medium, DMEM) into the fluidic channel for 10 min at a flow rate of 20 µl/min to replace PBS.

3. Cell Preparation and Experimental Setup

NOTE: Pre-warm 1x PBS, culture medium (DMEM plus FBS), and trypsin in a 37 °C water bath before usage.

1. Plate 2×10^5 lung cancer CL1-5 cells^{25,26} in a 10 cm Petri dish supplied with DMEM plus 10% fetal bovine serum (FBS). Incubate the cells inside an incubator under 5% CO₂ at 37 °C until 90% confluence.
2. Aspirate the medium and wash the cells once with pre-warmed 1x PBS. Add 2 ml of 0.05% trypsin buffer to the cells and wait for 2 to 3 min at 37 °C to detach the cells.
3. Transfer the cells to a 15 ml sterile centrifugation tube and add 6 ml of culture medium into the tube. Gently invert the tube for mixing and take out 5 µl of the cell-containing medium for counting the cell number in a hemocytometer.
4. Centrifuge the tube at 300 x g for 5 min. Suspend 10^6 cells in 1 ml of culture medium and place the cell-containing medium in a 1 ml syringe.
5. Infuse the cell-containing medium into the microfluidic chip from the outlet and make sure that the solution distributes all through the five culture areas. Incubate the chip inside an incubator under 5% CO₂ at 37 °C for 2 hr.

4. Experimental Setup

1. Take the chip out of the incubator and place it on top of a transparent indium tin oxide (ITO) glass heater.
NOTE: The ITO glass is connected to a proportional-integral-derivative (PID) controller for maintaining the temperature at 37 ± 0.5 °C via feedback from a thermal coupler clamped tightly between the ITO heater and the chip.
2. Put the chip-heater assembly on top of a motorized XY stage of an inverted microscope for tracking cell migration or a fixed XY stage of an inverted fluorescent microscope for measuring the production of ROS.
3. To generate different chemical concentrations, fill two syringes with 5 ml of chemical solutions of relative concentrations 0 and 1 (dissolved in culture medium) and pump them into the chip at desired flow rates: in the CSS chip, at a flow rate of 0.3 ml/min for 1 hr; in the CEF chip, at a flow rate of 20 µl/min for the first 20 min and a flow rate of 20 µl/hr for another 120 min (total time = 2 hr).
4. In the CEF chip, for tracking cell migration, program the XY stage of the microscope to repeat taking photos, via a digital single-lens reflex (DSLR) camera, of certain field of views (FOVs) within culture areas every 15 min for 2 hr.
5. For measuring the production of ROS, use the fluorescence-based indicator 2'-7'-dichlorodihydrofluoresce diacetate (2'-7'-DCFDA).
 1. Prepare the stock of 2'-7'-DCFDA at 10 mM in molecular biology grade dimethyl sulfoxide (DMSO). Dilute DCFDA in DMEM only without serum (5 µM in DMEM). The potential deacetylase could increase the background signals and decrease the signals in cells.
 2. After 1 hr of shear stress stimulus in the CSS chip or after 2 hr of EF stimulus in the CEF chip, pump 2'-7'-DCFDA (5 µM in DMEM) into the chip at a flow rate of 20 µl/min for the first 20 min and a flow rate of 20 µl/hr for another 20 min. For washing, pump DMEM into the chip at a flow rate of 20 µl/hr for 20 min.
 3. Take photos, via a charge-coupled device (CCD) camera, of certain FOVs within culture areas for analyzing the fluorescent intensities.

5. Calculations of Chemical Concentrations, Shear Stresses, and Electric Fields

1. In both the CSS chip and the CEF chip, calculate the chemical concentrations in the five culture areas. For example, by injecting H₂O₂ of concentrations 0 and 200 µM from the two inlets, concentrations of 0, 25, 100, 175, and 200 µM are generated.
NOTE: By assuming that all liquids split-flow smoothly and equally around the fork, the relative concentrations in the five culture areas are 0, 1/8, 1/2, 7/8, and 1, respectively.
2. In the CSS chip, calculate the shear stress (τ) within each of the culture areas using $\tau = 12 \frac{Q\eta}{h^2w}$ ²⁷, where Q is the volume flow rate, η is the fluidic viscosity, h is the height of the channel, and w is the width of the channel.
NOTE: By setting $Q = 0.3$ ml/min in each inlet ($Q = 0.12$ ml/min in each culture area), $\eta = 0.0008$ Pa·s for culture medium, $h = 1$ mm, and $w = 1 \sim 4$ mm, the shear stress is calculated to range from 0.0048 Pa (4 mm wide region) to 0.0192 Pa (1 mm wide region).
3. In the CEF chip, calculate the EF strength within each of the culture areas using $E = I/(\sigma A_{eff})$ (Ohm's law), where I is the electric current flowing across the fluidic channel, σ is the electrical conductivity of the culture medium, and A_{eff} is the effective cross-sectional area of the channel.
NOTE: Using $\sigma = 1.38 \Omega^{-1}m^{-1}$ for culture medium and $A_{eff} = 0.22$ mm² (width = 1 mm and height = 0.22 mm), the EF strength is calculated to be E (mV/mm) = I (A) $\times 3.3 \times 10^6$.
 1. As shown in **Figure 2D**, treat the equivalent circuit as five C-section circuits with four (8 + 35 + 8) segments and one (5 + 35 + 5) segment.
NOTE: By analyzing this parallel circuit according to Kirchhoff's voltage law and Ohm's law, currents flowing across five culture areas, I_1 through I_5 from bottom to top, are calculated to be around 0.49I (area 1), 0.25I (area 2), 0.13I (area 3), 0.08I (area 4), and 0.05I (area 5), respectively, where I is the total direct current (dc). With an applied dc of 0.157 mA, EFs of 254, 130, 67, 41, and 26 mV/mm are generated within the five culture areas.
NOTE: For a simplified electrical analysis of the microfluidic network, all fluidic segments are considered as resistors with resistance proportional to their lengths.

6. Data Analysis

Note: Data analysis is performed using the ImageJ software.

1. Analyze the Production of ROS.
 1. Run the ImageJ software. Go to "File" → Open to load a fluorescent image to be analyzed.
 2. Go to "Image" → "Type" → "16-bit" to change the image to a gray scale.
 3. Draw a polygon to enclose a cell. Go to "Analyze" → "Measure" to measure the mean fluorescent intensity of the cell.
 4. Repeat 6.1.3 to collect intensities from at least 50 cells of three independent experiments, and calculate the mean intensity with standard error of mean (SEM).
 5. Repeat 6.1.1-6.1.4 for each experimental condition.

2. Analyze Cell Migration.

1. Run the ImageJ software. Go to "File" → "Open" to load an image taken at time = 0 to be analyzed.
2. Draw a polygon to enclose a cell. Go to "Analyze" → "Measure" to measure the center of mass of the cell as (x_1, y_1) .
3. Repeat 6.2.1- 6.2.2 to measure the center of mass of the same cell as (x_2, y_2) from another image take at time = t .
4. Calculate the migration rate (in $\mu\text{m/hr}$) of this cell as $(x_2 - x_1) / \sqrt{(x_2 - x_1)^2 + (y_2 - y_1)^2}$.
5. Repeat 6.2.1-6.2.4 to collect migration rates from at least 50 cells of three independent experiments, and calculate the mean migration rate with standard error of mean (SEM).
6. From 6.2.2 and 6.2.3, calculate the migration directedness of this cell as cosine θ or $\sqrt{(x_2 - x_1)^2 + (y_2 - y_1)^2} / t$, where θ is the angle between the vector of the applied EF (from positive to negative) and the vector from the start to the end position of the cell (**Figure 2G**).
7. Repeat 6.2.6 to collect migration directedness from at least 50 cells of three independent experiments, and calculate the mean migration directedness with standard error of mean (SEM).
8. Calculate the mean migration rate with SEM and the mean migration directedness with SEM for each experimental condition.
NOTE: A directedness of +1 indicates that all cells migrate toward the cathode, and a -1 value indicates that all cells migrate toward the anode. The directedness of a group of randomly moving cells is close to 0.

3. Analyze Cell Alignment.

1. Run the ImageJ software. Go to "File" → "Open" to load an image to be analyzed.
2. Treat the cell as an ellipse and draw a line to indicate the long axis of a cell. Go to "Analyze" → "Measure" to measure the angle β between the line and the horizontal EF direction.
3. Repeat 6.3.2 to collect β from at least 50 cells of three independent experiments, and calculate mean cosine β with standard error of mean (SEM).
4. Repeat 6.3.1-6.3.3 for each experimental condition.
NOTE: A $\cos\beta$ of +1 indicates that all cells align in parallel to the applied EF, and a 0 value indicates that all cells align perpendicularly to the applied EF.

Representative Results

The Chemical-shear Stress (CSS) Chip

The CSS chip is made of three PMMA sheets, each of thickness 1 mm, attached together via two double-sided tapes, each of thickness 0.07 mm (**Figure 1A** and **1B**). The "Christmas tree" structure generates five relative concentrations of 0, 1/8, 1/2, 7/8, and 1 in the five culture areas. By designing the culture area as a triangle, a shear stress gradient, with a magnitude related the volume flow rate, the fluidic viscosity, and the dimension of the fluidic channel, is created within each of the areas. This chip is then attached, via another 0.07 mm thick double-sided tape, to a Petri dish for culturing lung cancer CL1-5 cells and the production of ROS was observed in response to different chemical concentrations and shear stresses. The production of ROS in lung cancer cells is highly related to lung cancer metastasis and development, and certain chemicals and shear stress were demonstrated to be involved in ROS generation²³.

First, to study the effect of H_2O_2 , a chemical stimulus, on the production of ROS, cells were incubated with continuous flowing of H_2O_2 solutions at 0, 25, 100, 175, and 200 μM . As shown in **Figure 1C**, the fluorescent intensity increased as the concentration of H_2O_2 increased, indicating that H_2O_2 stimulated the production of ROS. Next, to investigate the effect of shear stress on the production of ROS, cells were exposed to a shear stress gradient of 0.0048 Pa to 0.0192 Pa. **Figure 1D** shows that the fluorescent intensity increased as the shear stress increased (the shear stresses were the highest and the lowest in the Front and Back areas, respectively), suggesting that higher shear stress induced more ROS production. Also, this CSS chip was used to study the production of ROS in response to different concentrations of α -tocopherol, an antioxidant of a form of vitamin E. Cells were stimulated by shear stress plus α -tocopherol of 0, 3.2, 12.5, 21.9, and 25 $\mu\text{g/ml}$. As shown in **Figure 1E**, for α -tocopherol concentrations lower than 21.9 $\mu\text{g/ml}$, the fluorescent intensity decreased as the concentration increased, indicating the effect of α -tocopherol in reducing the production of ROS. However, as the concentration increased to 25 $\mu\text{g/ml}$, the mean intensity also increased, suggesting that this high concentration of α -tocopherol did not eliminate much ROS compared to lower concentrations.

The Chemical-electric Field (CEF) Chip

The CEF chip is made of a 1 mm-thick PMMA and a 0.22-mm double-sided tape with fluidic channels patterned on it (**Figure 2A** and **2B**). Similarly, the "Christmas tree" structure creates five relative concentrations of 0, 1/8, 1/2, 7/8, and 1 in the five culture areas. In addition, these five parallel areas are connected perpendicularly to form a fluidic path analogous to an electric circuit. The electric field generated within a culture area is related to the conductivity of the fluid, the cross-sectional area of the channel, and the direct current (dc) passing through the area. According to Kirchhoff's voltage law and Ohm's law, the currents in the five culture areas are 0.49I, 0.25I, 0.13I, 0.08I, and 0.05I, respectively, where I is the applied dc (**Figure 2C**). This chip is attached, via the 0.22 mm thick double-sided tape, to a Petri dish for culturing lung cancer CL1-5 cells to observe cell migration and the production of ROS in response to different chemical concentrations and electric fields.

First, this chip was used to study the production of ROS in response to different concentrations of honokiol, different strengths of EFs, and combined treatments of both. Honokiol, a small-molecule polyphenol isolated from the genus *Magnolia*, was found to have antiangiogenic, anti-inflammatory, and antitumor properties in preclinical studies²⁸. As shown in **Figure 2D**, the ROS produced was almost the same for EFs lower than 67 mV/mm, but was increased with increasing EFs above this value. **Figure 2E** shows that under combined treatments of electric fields and honokiol, the ROS level stayed almost the same, indicating that honokiol inhibited the production of exogenous ROS (*i.e.*, ROS related to EF stimulus), especially under higher EFs. The cell migration under single or coexisting chemical/electrical stimuli was also investigated using this CEF chip. As seen in **Figure 2F**, without the addition of honokiol, the migration rate increased as the EF strength increased. After adding honokiol of different concentrations, the migration rate decreased in general, suggesting that honokiol reduced cell migration possibly via inhibiting the production of ROS. The migration directedness is shown in **Figure 2G**. In the presence of EF only, lung cancer cells showed prominent directional migration toward the anode (with negative values). After adding honokiol of different concentrations, there was a slight decrease in migration directedness for all EF-stimulated areas.

Figure 1: The Chemical-shear Stress (CSS) Chip Used to Study the Effects of Chemicals and Shear Stresses on Lung Cancer Cells.²³ (A) Three layers of PMMA substrates are bound together via two double-sided tapes to form the CSS chip. (B) The integrated CSS chip. (C) Top: Fluorescent and bright-field images of CL 1-5 cells after being treated with different concentrations of H_2O_2 . Scale bar = 50 μm . Bottom: Mean fluorescent intensity with SEM plotted at different concentrations of H_2O_2 . (D) Top: Fluorescent images of CL 1-5 cells under different shear stresses. Scale bar = 50 μm . Bottom: Mean fluorescent intensity with SEM plotted at different shear stresses. (E) Top: Fluorescent and bright-field images of CL 1-5 cells after being stimulated with shear stress with different concentrations of α -tocopherol. Scale bar = 50 μm . Bottom: Mean fluorescent intensity with SEM plotted at different concentrations of α -tocopherol. Data presented here was originally published in reference²³. [Please click here to view a larger version of this figure.](#)

Figure 2: The Chemical-electric field (CEF) Chip Used to Study the Effects of Chemicals and Electric Fields on Lung Cancer Cells²². (A) One layer of PMMA is bound onto a culture dish via a double-sided tape to form the CEF chip. (B) The fluidic pattern of the CSS chip. (C) The equivalent electrical circuit of the microfluidic chip. (D) Left: Fluorescent and bright-field images of CL 1-5 cells after being treated with different strengths of EFs. Scale bar = 50 μ m. Right: Mean fluorescent intensity with SEM plotted at different strengths of EFs. (E) Left: Fluorescent and bright-field images of CL 1-5 cells after being treated with combined honokiol and EFs. Scale bar = 50 μ m. Right: Mean fluorescent intensity with SEM plotted at combined honokiol and EFs. (F) The migration rates of CL1-5 cells after being treated with EFs only (marked as control) and combined honokiol and EFs (marked as honokiol). (G) The migration directedness of CL1-5 cells after being treated with EFs only (marked as control) and combined honokiol and EFs (marked as honokiol). Data presented here was originally published in reference ²². Please click here to view a larger version of this figure.

Discussion

PMMA-based chips are fabricated using laser ablation and writing which are cheaper and easier methods when compared to PDMS-based chips which require more complicated soft lithography. After designing a microfluidic chip, the fabrication and assembly can be done within just 5 min. There are some critical steps that attention should be paid to in performing the experiment. The first is the "assembling" issue. The adaptors should be glued properly to the top-most layer of the chip. Glue could leak into the fluidic channels if too much is applied, and liquid could leak out of the chip if too little glue is used. Also, in assembling the PMMA substrates and the double-sided tapes, it is important to apply pressure to press the whole chip tightly to prevent any liquid leakage. The second is the "bubble" issue. To remove bubbles, 1x PBS is continuously flowed into the fluidic channel for 1 or 2 min at a flow rate of 20 μ l/min and then the chip is placed inside an incubator overnight under 5% CO₂ at 37 °C. If there are bubbles remaining within the fluidic channels, the flow will be non-uniform, causing nonhomogeneous chemical concentration and EF distribution. The third is the "cell loading" issue. The number of cells loaded into the culture areas should be well controlled. If the cell number is too low, many experimental runs should be conducted to collect enough data for statistical purpose. If the cell number is too high, it will be difficult to distinguish individual cells and quantify their migration.

In the CSS chip, a "Christmas tree" structure combined with a triangular culture area generates different chemical concentrations in five areas, each having a shear stress gradient. Fluidic shear stresses were known to affect the attachment, the morphology, the migration, and the activity of cells. Shear stresses of as low as 0.25-0.6 Pa could interface cell attachment, and even higher values of stresses (0.5 -10 Pa) were reported

to remove adherent cells²⁹. Laminar shear stresses ranging from 0.8 to 1.5 Pa induced cell alignment in the direction of flow³⁰, and even lower values (0.1-1 Pa) were known to affect cellular morphology and permeability²⁹. Moreover, smooth muscle cells subjected to shear stresses of 2-120 Pa experienced a decline in cell number³¹. Lu *et al.* reported the design and construction of microfluidic shear devices for quantitative analysis of cell adhesion²⁷. The shear stress within the fluidic channel was modified by changing the dimensions of the channel. Chin *et al.* described a hemodynamic Lab-on-a-chip system for controlling the flow rate of the culture medium in the micro-channel to mimic the flow profile of the blood in the vessel³². The shear stress within the fluidic channel was modified by changing the flow rate of the medium. In these two devices, even though a shear stress of any values could be generated, only one single value was available within one culture area at one time. By comparison, the present CSS chip has the advantage of providing a shear stress gradient in one triangular area, increasing the experimental throughput especially for a screening purpose.

In the CEF chip, the culture areas of the "Christmas tree" structure are perpendicularly connected to form a fluidic path analogous to an electric circuit. At one applied direct current, five EF strengths are generated inside the culture areas with each having a flow of a specific chemical concentration. Conventionally, electrotactic experiments were performed on Petri dishes where problems of imprecise EFs, medium evaporation, large cell/reagent consumption, and increased Joule heating could occur. Enclosed, miniature microfluidic devices efficiently overcome these drawbacks. For example, to study the electrotaxis of human blood memory T cells, Lin *et al.* reported a plastic microfluidic device containing two identical, side-by-side micro-channels, two modified pipette tips served as medium reservoirs, and two platinum electrodes for EF application¹⁵. To increase the throughput, PMMA-based electrotactic chips were fabricated to provide multiple EF strengths in one single device^{12,33}. Also, microfluidic devices capable of generating controllable chemical and electrical stimuli simultaneously are of high interest. Li *et al.* fabricated a PDMS-based microfluidic device to generate single or coexisting chemical gradients/EFs for investigating the chemotactic or electrotactic migration of T cells¹⁴. Kao *et al.* employed a similar microfluidic device to modulate the chemotaxis of lung cancer cells by using EFs¹². In these two devices, a Y-shape structure was applied to create a stable concentration gradient in one single applied EF. Hou *et al.* reported a multichannel-dual-electric-field chip to study the concurrent effect of chemicals and EFs on lung cancer cells²¹. This device was able to provide 8 combinations of electrical/chemical stimuli (2 electrical x 4 chemical) in one experiment. Although the throughput increased, this chip's capacity was limited to (1) only one EF strength (plus one zero as a control), and (2) four syringe pumps are required for pumping chemicals into four independent channels. By comparison, the present CEF chip has the advantage of generating five EF strengths together with five chemical concentrations via one single set of electrodes and 2 syringe pumps.

Even though these chips are easy to be fabricated, they have some limitations. First, only five "specific" concentrations and five "specific" EF strengths can be generated. Second, the width of the fluidic channel cannot be less than 1 mm due to the focusing of the CO₂ laser beam. However, by precisely controlling the chemical concentration in the inlet and the electric current passing through the fluidic channel, any concentrations and EF strengths can be attained in the culture areas. In conclusion, the present CSS and CEF chips are capable of providing cells with controllable single or coexisting chemical/electrical/shear stress stimuli. In one single CSS chip, five different chemical concentrations in combination with a shear stress gradient can be generated. In addition, in one single CEF chip, five different chemical concentrations in combination with five EF strengths can be created. By increasing the branches of the "Christmas tree" structure, the throughput of these chips can be further increased. These chips are demonstrated to be an easy, time-saving, reliable, and high-throughput platform for studies of cellular behavior under chemical/electrical/shear stress stimuli.

Disclosures

The authors have nothing to disclose.

Acknowledgements

This work was financially supported by the Ministry of Science and Technology of Taiwan under Contract No. MOST 104-2311-B-002-026 (K. Y. Lo), No. MOST 104-2112-M-030-002 (Y. S. Sun), and National Taiwan University Career Development Project (103R7888) (K. Y. Lo). The authors also thank the Center for Emerging Material and Advanced Devices, National Taiwan University, for the use of the cell culture room.

References

- Cheng, J. Y., Yen, M. H., Kuo, C. T., & Young, T. H. A transparent cell-culture microchamber with a variably controlled concentration gradient generator and flow field rectifier. *Biomicrofluidics*. **2** (2008).
- Terry, S. C., Jerman, J. H., & Angell, J. B. Gas-Chromatographic Air Analyzer Fabricated on a Silicon-Wafer. *Ieee T Electron Dev*. **26**, 1880-1886 (1979).
- Whitesides, G. M., Ostuni, E., Takayama, S., Jiang, X., & Ingber, D. E. Soft lithography in biology and biochemistry. *Annu Rev Biomed Eng*. **3**, 335-373 (2001).
- Adler, J. Chemoreceptors in bacteria. *Science*. **166**, 1588-1597 (1969).
- Boyden, S. The chemotactic effect of mixtures of antibody and antigen on polymorphonuclear leucocytes. *J Exp Med*. **115**, 453-466 (1962).
- Kao, Y. C. *et al.* Modulating chemotaxis of lung cancer cells by using electric fields in a microfluidic device. *Biomicrofluidics*. **8**, 024107 (2014).
- Walker, G. M. *et al.* Effects of flow and diffusion on chemotaxis studies in a microfabricated gradient generator. *Lab Chip*. **5**, 611-618 10.1039/B417245k (2005).
- Al-Majed, A. A., Neumann, C. M., Brushart, T. M., & Gordon, T. Brief electrical stimulation promotes the speed and accuracy of motor axonal regeneration. *J Neurosci*. **20**, 2602-2608 (2000).
- Nuccitelli, R. Endogenous electric fields in embryos during development, regeneration and wound healing. *Radiat Prot Dosim*. **106**, 375-383 (2003).
- McCaig, C. D., Rajnicek, A. M., Song, B., & Zhao, M. Controlling cell behavior electrically: Current views and future potential. *Physiol Rev*. **85**, 943-978 (2005).

11. Tai, G., Reid, B., Cao, L., & Zhao, M. Electrotaxis and wound healing: experimental methods to study electric fields as a directional signal for cell migration. *Methods Mol Biol.* **571**, 77-97 (2009).
12. Huang, C. W., Cheng, J. Y., Yen, M. H., & Young, T. H. Electrotaxis of lung cancer cells in a multiple-electric-field chip. *Biosens Bioelectron.* **24**, 3510-3516 (2009).
13. Yan, X. L. *et al.* Lung Cancer A549 Cells Migrate Directionally in DC Electric Fields With Polarized and Activated EGFRs. *Bioelectromagnetics.* **30**, 29-35 (2009).
14. Li, J. *et al.* Activated T lymphocytes migrate toward the cathode of DC electric fields in microfluidic devices. *Lab Chip.* **11**, 1298-1304 (2011).
15. Lin, F. *et al.* Lymphocyte electrotaxis in vitro and in vivo. *J Immunol.* **181**, 2465-2471 (2008).
16. Zhang, J. *et al.* Electrically Guiding Migration of Human Induced Pluripotent Stem Cells. *Stem Cell Rev.* (2011).
17. Song, B. *et al.* Application of direct current electric fields to cells and tissues in vitro and modulation of wound electric field in vivo. *Nat Protoc.* **2**, 1479-1489 (2007).
18. Sun, Y. S., Peng, S. W., & Cheng, J. Y. In vitro electrical-stimulated wound-healing chip for studying electric field-assisted wound-healing process. *Biomicrofluidics.* **6**, 34117 (2012).
19. Sun, Y. S., Peng, S. W., Lin, K. H., & Cheng, J. Y. Electrotaxis of lung cancer cells in ordered three-dimensional scaffolds. *Biomicrofluidics.* **6** (2012).
20. Li, J., Zhu, L., Zhang, M., & Lin, F. Microfluidic device for studying cell migration in single or co-existing chemical gradients and electric fields. *Biomicrofluidics.* **6**, 24121-2412113 (2012).
21. Hou, H. S., Tsai, H. F., Chiu, H. T., & Cheng, J. Y. Simultaneous chemical and electrical stimulation on lung cancer cells using a multichannel-dual-electric-field chip. *Biomicrofluidics.* **8**, 052007 (2014).
22. Lo, K. Y., Wu, S. Y., & Sun, Y. S. A microfluidic device for studying the production of reactive oxygen species and the migration in lung cancer cells under single or coexisting chemical/electrical stimulation. *Microfluid Nanofluidics.* **20**, 15 (2016).
23. Lo, K. Y., Zhu, Y., Tsai, H. F., & Sun, Y. S. Effects of shear stresses and antioxidant concentrations on the production of reactive oxygen species in lung cancer cells. *Biomicrofluidics.* **7**, 64108 (2013).
24. Cheng, J. Y., Wei, C. W., Hsu, K. H., & Young, T. H. Direct-write laser micromachining and universal surface modification of PMMA for device development. *Sensor Actuat B-Chem.* **99**, 186-196 (2004).
25. Chen, J. J. W. *et al.* Global analysis of gene expression in invasion by a lung cancer model. *Cancer Res.* **61**, 5223-5230 (2001).
26. Shih, J. Y. *et al.* Collapsin response mediator protein-1 and the invasion and metastasis of cancer cells. *J Natl Cancer I.* **93**, 1392-1400 (2001).
27. Lu, H. *et al.* Microfluidic shear devices for quantitative analysis of cell adhesion. *Anal Chem.* **76**, 5257-5264 (2004).
28. Fried, L. E., & Arbiser, J. L. Honokiol, a Multifunctional Antiangiogenic and Antitumor Agent. *Antioxid Redox Sign.* **11**, 1139-1148 (2009).
29. Chisti, Y. Hydrodynamic damage to animal cells. *Crit Rev Biotechnol.* **21**, 67-110 (2001).
30. Davies, P. F., Remuzzi, A., Gordon, E. J., Dewey, C. F., Jr., & Gimbrone, M. A., Jr. Turbulent fluid shear stress induces vascular endothelial cell turnover in vitro. *Proc Natl Acad Sci U S A.* **83**, 2114-2117 (1986).
31. Zoro, B. J., Owen, S., Drake, R. A., & Hoare, M. The impact of process stress on suspended anchorage-dependent mammalian cells as an indicator of likely challenges for regenerative medicines. *Biotechnol Bioeng.* **99**, 468-474 (2008).
32. Chin, L. K. *et al.* Production of reactive oxygen species in endothelial cells under different pulsatile shear stresses and glucose concentrations. *Lab Chip.* **11**, 1856-1863 (2011).
33. Tsai, H. F., Peng, S. W., Wu, C. Y., Chang, H. F., & Cheng, J. Y. Electrotaxis of oral squamous cell carcinoma cells in a multiple-electric-field chip with uniform flow field. *Biomicrofluidics.* **6**, 34116 (2012).