Published in final edited form as:

Am J Surg. 2016 December ; 212(6): 1140–1146. doi:10.1016/j.amjsurg.2016.09.013.

Utilization of Iron (III) Doped Nanoshells for in vivo Marking of Non-palpable Tumors Using a VX2 Rabbit Model

Erin Ward¹, James Wang², Natalie Mendez³, Jian Yang³, Chris Barback¹, Jessica Wang-Rodriguez⁴, William Trogler³, Andrew C. Kummel³, and Sarah Blair¹

¹Moore's Cancer Center, University of California, San Diego, CA 92093, USA

²Department of Nanoengineering, University of California, San Diego, CA 92093, USA

³Department of Chemistry and Biochemistry, University of California, San Diego, CA 92093, USA

⁴Department of Pathology, University of California, San Diego, CA 92093, USA

Abstract

INTRODUCTION—We aimed to evaluate the potential for ultrasound (US) visible biodegradable nanoshells (NS) as an alternative to wire guided localization (WL) for non-palpable tumors in vivo.

METHODS—VX2 tumor was injected in bilateral thighs of 22 New Zealand rabbits and after 5–10 days one tumor was marked with a wire as a control and the contralateral tumor was injected with 1mL of 500 nm gas-filled silica NS under Doppler US. Tumors were excised after 24 hours. Chi square was utilized for significance, $p=.05$.

RESULTS—One rabbit was excluded on postoperative day 1 due to equipment failure, no ill effects were observed from the NS. The NS were utilized to localize and resect 100% of marked tissue, 4/21 wires were displaced ($p<.05$).

CONCLUSIONS—We have shown that preoperatively injected US visible silica NS can be successfully utilized to mark non-palpable tumors in vivo more consistently than WL.

Keywords

breast cancer; nanoshells; silica nanoparticles; tumor localization; non-palpable

Introduction

Breast conserving surgery (BCS) is the standard treatment for early stage breast cancer and to 25–30% of diagnosed and treated breast cancers are non palpable requiring pre-operative localization for successful resection.^{1–4} Current strategies for marking these tumors include

Corresponding Author: Sarah L Blair MD, Moore's Cancer Center, 3855 Heath Science Dr MC 0987, La Jolla CA 92093-0987, slblair@ucsd.edu, (858) 822-1230.

Publisher's Disclaimer: This is a PDF file of an unedited manuscript that has been accepted for publication. As a service to our customers we are providing this early version of the manuscript. The manuscript will undergo copyediting, typesetting, and review of the resulting proof before it is published in its final citable form. Please note that during the production process errors may be discovered which could affect the content, and all legal disclaimers that apply to the journal pertain.

utilizing a wire or implanted radioactive seeds to guide the surgeon to the biopsy proven tumor.

The current gold standard and most frequently utilized method for localization is image guided wire localization (WL). This technique relies on a thin hooked wire that is preoperatively inserted into the tumor under ultrasound or stereotactic guidance by a radiologist to guide the surgeon to the lesion. WL has inherent limits to precise localization because the external portion is vulnerable to displacement and it only provides a single dimension of reference for a 3D tumor volume; consequently, reported rates of positive margins range from 14 to 47%.⁵ To minimize the risk of displacement, the wire must be placed shortly prior to the procedure. Patients also report high levels of discomfort even associated with the wire.⁶ Other techniques that address some of these concerns include radio guided seed localization (RSL) with a internally placed seed. Unfortunately, RSL has been shown to have similar high rates of positive margins requiring re-excision and or mastectomy.^{5,78} Just as with “wire only”, with “RSL only” a single plane of the tumor can be marked with the seed consistent with the continued high rates of positive margins.⁹ The irregular growth of non-palpable tumors is postulated to contribute to the high rates of positive margins for both techniques.^{9,10}

Intraoperative ultrasound guided lumpectomy has demonstrated excision rates comparable to WL.¹¹ Studies reveal that up to 95-100% of non-palpable lesions can be successfully located with ultrasound.¹² This present study demonstrates the utility of localizing non-palpable tumors with ultrasound visible gas filled nanoshells to allow for rapid intraoperative localization and excision of non-palpable tumors. This technique allows for the tumors to be marked multiple days prior and does not rely on radioactive seeds or vulnerable external wires.^{13,14}

Methods

Nanoshells

The 500 nm hollow ultrathin iron (III) doped silica nanoshells (NS) were synthesized as previously described.^{15,16,17} The hollow silica NS are subjected to at least three cycles of vacuum desorption followed by perfluoropentane filling. The perfluoropentane gas filled NS are suspended in degassed water at a concentration of 4 mg/ml

Tumor Model

The VX2 rabbit model was employed as rabbits were deemed the smallest species available (a) with an area of tissue of sufficient size to be able to implant and excise a tumor without severely impacting the mobility of the animal, (b) have tumors which were non-palpable but visible with ultrasound, and (c) have sufficient muscle mass to allow for negative margins to potentially be achieved. The VX2 rabbit model has the additional advantage that it can be introduced and grown in a consistent time line rather than forming spontaneously. This allows for multiple rabbits to have the same timeline for injection, marking, excision and evaluation for regrowth.

This study was approved by the Institutional Animal Care and Use Committee of the University of California San Diego. 22 female New Zealand White rabbits age 10–12 weeks were housed in an approved animal housing facility and kept at 20 °C with a 12 h light/dark cycle. The animals were fed a commercial pelleted diet (Harlan Teklad) ad libitum. Rabbits were anesthetized with isoflurane gas with oxygen during the procedures and excisions of primary tumors.

After the prescribed 72-hour acclimation period, the rabbits were injected with 1cc of VX2 slurry in bilateral rear thighs. The tumors were allowed to grow between 5 and 10 days to a non-palpable tumor approximately 2-5 mm in size. Twenty-four hours prior to excision, each rabbit had one tumor marked utilizing B mode ultrasound guided wire placement to serve as the control. The contralateral thigh was marked with approximately 0.5 to 1.0cc of 500 nm gas filled silica NS with Doppler guided ultrasound (Figure 1). During imaging of the NS, the US parameters were optimized in order to maximize signal from particles while reducing background in all cases, as is done clinically. A frequency of 7 MHz and a mechanical index (MI) of 1.9, were chosen for Doppler imaging parameters, as in previous studies.¹³ The US machine employed was the Siemens Acuson Sequoia 512 with the 15L8 transducer which is traditionally used for breast US imaging.

The marking of the tumors in this model requires general anesthesia for the animals and recovery time was required prior to the second general anesthesia for tumor excision, thus the wires were implanted 24 hours prior to tumor excision.. This approach allowed for both marking strategies to be evaluated in an equal fashion and while more time was allotted to pass between marking and excision than is clinically standard, the transfers and potential manipulation of the wire was limited to minimize the potential risk of wire displacement. In addition, this necessary strategy allowed evaluation of the NS as a potential tumor marker that could be decoupled with the day of surgery.

The tumors were excised under closely monitored isoflurane anesthesia. The wire side was excised first to minimize any manipulation of the wire. After the wire side was closed, the nanoparticle side was imaged with the optimized Doppler parameters. The tumor was excised based on the Doppler image. The excised tissue was imaged and evaluated for evidence of NS in the sample, as was the tumor cavity. All tissue with evidence of remaining NS was excised at that time (Figure 2).

Post operatively, the rabbits were allowed to recover for approximately 3 weeks until the animals were sacrificed to allow for evaluation of regrowth in the separate tumor beds. All pathology specimens were evaluated by a board certified pathologist and tissue from the surgeries were evaluated for evidence of excised tumor and positive margins, which was defined as tumor touching ink (Figure 3).

Statistics

Chi square statistics were employed to compare the success of excision between the wire and the NS as well as the percent regrowth at a p value of .05.

Results

All 22 rabbits survived the initial surgery; one rabbit was euthanized on post-operative day 1 secondary due to a burn from equipment failure and was excluded from the study. No direct ill effects were observed from the injected NS. After marking the tumors, the rabbits were allowed to recover for approximately 24 hours prior to the excision of the tumor. Despite extensive efforts to minimize the displacement of the wire, the wire was displaced in 4 out of the 21 surgeries. The NS that were injected in the tumor were successfully visualized through the skin and within the tumor cavity in 100% of the surgeries, significantly more than for the WL ($p<.05$) (Figure 4).

NS were confirmed within the initial tissue excision in 100% of NS marked tumors. After excision of the specimen on the particle side, the cavity was explored with the Doppler ultrasound to evaluate for any additional visible NS and an additional margin was taken to remove the any ultrasound positive marked. (see Figure 2). At the conclusion of the procedure, the cavity was verified to be clean of all visible NS in 21/21 NS based excisions, as there is currently no way to preform such confirmation with WL, this was significantly different from the WL ($p<.05$).

After the excision, the specimen was evaluated for both presence of tumor within the specimen and evidence for positive margins. Of the 21 WL based excisions, tumor was identified in 17 samples and margins were negative for 5 of the 17 successfully excised tumors. In contrast, 19 of the 21 NS marked tumors were excised and of these 9 had negative margins. While the NS overall had a higher percentage of negative margins on excision than the WL, addition rabbit experiments will be needed for definitive statistics since the p value was modest. ($p=.27$).

After 17-21 days post surgery the animals were euthanized to allow for exploration of bilateral tumor cavity for regrowth. A pathologist reviewed the excised tissue for evidence of regrowth. Overall 17 of the 21 WL tumors regrew and 16 of the 21 NS marked tumors regrew. This had no statistical difference. ($p=.71$).

Discussion

Breast cancer accounts for 1 in 3 of all cancers in the US and current statistics support that 1 in 8 American women will develop breast cancer during their lifetime.¹⁸ Screening mammography and public awareness has led to an increased number of breast cancer diagnoses and a shift towards diagnosis at an earlier stage.³ The primary approach to early stage, small breast cancers is breast-conserving treatment with adjuvant radiation and surgeons rely on preoperative localization to obtain negative margins when resecting non-palpable tumors.⁴ Currently the gold standard is WL but positive margin rates are reported as high as 47% in reports.¹⁹

A VX2 tumor in a rabbit model was employed to compare the efficacy of 500 nm silica NS as an ultrasound visible marker for non-palpable breast cancer localization to the gold standard WL. 19% of the wires were displaced and unable to be utilized for intraoperative localization. Conversely, using the same localization procedure, a 24 hour delay after the

injection of the NS did not cause displacement of the NS allowing for 100% of all marked tissue to be identified intraoperatively for the 21 NS marked tumors.

The inherent flaws of WL include the propensity of the wire to migrate from the initial placement and that wire is limited in its ability to represent a 3D structure that may be irregular in shape as it only identifies one segment of the lesion.⁹¹⁰ While lesion bracketing with multiple wires has been utilized to try to overcome this challenge, Liberman et al. did not find improvement in rates of negative margins.²⁰

In clinical settings, wires must be placed the day of surgery and all efforts must be made to minimize the delay in excision after wire placement leading to significant scheduling difficulties and often operating room delays. Studies comparing the RSL to WL have shown that because the radioactive seed can be placed prior to the day of the surgery, waiting times, patient satisfaction and scheduling difficulties are significantly reduced with RSL.⁸

The NS as a marker overcomes many limitations as the NS are injected intratumorally and previous studies have demonstrated that not only are the NS stationary, they are also visible up to 10 days after initial injection.¹³¹⁴ The NS have been shown to be non-toxic and biodegradable in previous animal studies.¹³¹⁴ The NS are less restrictive in the way a tumor is marked as it allows for a multidimensional marking. The injection of the particles is flexible in that a radiologist could inject NS to mark multiple tumors and fill irregular shaped tumors, minimizing the amount of interpretation that must be done intra-operatively. This ability to fill the entirety of the tumor may allow the NS to outperform RSL in future studies. The NS ability to remain stationary, even in the face of extreme movement of this animal model supports that this technique is immune to the traditional vulnerability of displacement of the wire. In addition as the NS can be injected days prior to the procedure, it can be assumed that scheduling benefits seen with the RSL could also be generalized to the this strategy

While other US contrast agents have been developed, the longevity and the ease of visualization are not comparable to the NS.¹³ With NS, the surgeon is able to visualize the particles within the tumor initially through the skin and then again when imaging directly in the tumor bed with ultrasound, as was demonstrated in this model. This model also demonstrated the potential to be able to utilize the US within the tumor cavity after excision to verify that all tumor that was marked has been excised and verify that there are visible NS within the tissue excised.

The rates of positive margins in this model for both the WL and NS based excision is likely secondary to the limitations of the model because the model has inherent limitations for achieving successful complete excision with negative margins. Unlike in human breast lumpectomies where cosmetic outcome is often the main factor for consideration in limiting total tissue resected, in this model the amount of tissue excised is limited due to concerns for mobility. To allow for survival surgery, the animal must be able to ambulate post-operatively, which limits the extent of the excision. Therefore, the rate of positive margins for both the NS and the WL was likely influenced by the model in addition to the technique utilized to mark the tumors. The wire was secured as much as possible utilizing multiple tools but again

the model requires the placement of the wire in the thigh of a mobile animal likely contributed to the relatively high rate of loss the wire in 19% of the WL excisions. In the same circumstances the NS were found to be unaffected by the animal's mobility prior to the excision of the marked tumor. The elevated rates of regrowth in both WL and NS may also be secondary to the fact that the rabbits did not undergo post operative radiation that is known to be critical to treat the micro-disease¹⁰.

The technique of injecting the tumor is a relatively facile procedure that does only requires equipment that is readily available to both surgeons and radiologists. The NS are easily visualized utilizing Doppler US on clinical US machines and allow for the radiologist marking the tumor to communicate through a stable mechanism a 3D structure up to 10 days prior to a planned OR procedure eliminating many of the inherent flaws of WL.

Conclusion

Localization of non-palpable tumors remains critical to effective surgery and thus treatment of early stage breast cancer. It has been demonstrated that preoperatively injected silica nanoparticles can be successfully utilized to mark non-palpable tumors in vivo more consistently than the current gold standard wire based platform. This study establishes a new category of tumor marking strategies using materials based on nanoparticles with the ability to mark 3-D physiological tumors structures that are not always captured well using traditional marking techniques.

Acknowledgements

This research has been funded by Covidien/RF surgical/Medtronic and supported by the following grants:

NIH IMAT: 1R21CA151140-01A2

NIH IMAT: 1R33CA177449-01A1

NIH NCI: 1F31CA174276-01A1

References

1. Yang SH, et al. Breast conservation therapy for stage I or stage II breast cancer: a meta-analysis of randomized controlled trials. *Ann Oncol*. 2008; 19:1039–1044. doi:10.1093/annonc/mdm573. [PubMed: 18187486]
2. Gangnon RE, et al. The contribution of mammography screening to breast cancer incidence trends in the United States: an updated age-period-cohort model. *Cancer Epidemiol Biomarkers Prev*. 2015; 24:905–912. doi:10.1158/1055-9965.EPI-14-1286. [PubMed: 25787716]
3. Myers ER, et al. Benefits and Harms of Breast Cancer Screening: A Systematic Review. *JAMA*. 2015; 314:1615–1634. doi:10.1001/jama.2015.13183. [PubMed: 26501537]
4. Fisher B, et al. Twenty-year follow-up of a randomized trial comparing total mastectomy, lumpectomy, and lumpectomy plus irradiation for the treatment of invasive breast cancer. *N Engl J Med*. 2002; 347:1233–1241. doi:10.1056/NEJMoa022152. [PubMed: 12393820]
5. Lovrics PJ, et al. A Multicentered, Randomized, Controlled Trial Comparing Radioguided Seed Localization to Standard Wire Localization for Nonpalpable, Invasive and in situ Breast Carcinomas. *Annals of Surgical Oncology*. 2011; 18:3407–3414. doi:10.1245/s10434-011-1699-y. [PubMed: 21533657]

6. Pan H, et al. Intraoperative ultrasound guidance is associated with clear lumpectomy margins for breast cancer: a systematic review and meta-analysis. *PloS one*. 2013; 8:e74028. doi:10.1371/journal.pone.0074028. [PubMed: 24073200]
7. Chan BK, Wiseberg-Firtell JA, Jois RH, Jensen K, Audisio RA. Localization techniques for guided surgical excision of non-palpable breast lesions. *Cochrane Database Syst Rev*. 2015; 12:CD009206. doi:10.1002/14651858.CD009206.pub2.
8. Sharek D, et al. Radioactive seed localization versus wire localization for lumpectomies: a comparison of outcomes. *AJR Am J Roentgenol*. 2015; 204:872–877. doi:10.2214/AJR.14.12743. [PubMed: 25794081]
9. Lovrics PJ, Cornacchi SD, Vora R, Goldsmith CH, Kahn moui K. Systematic review of radioguided surgery for non-palpable breast cancer. *Eur J Surg Oncol*. 2011; 37:388–397. doi:10.1016/j.ejso.2011.01.018. [PubMed: 21333482]
10. Ananthakrishnan P, Balci FL, Crowe JP. Optimizing surgical margins in breast conservation. *Int J Surg Oncol*. 2012; 2012:585670. doi:10.1155/2012/585670. [PubMed: 23304479]
11. Eggemann H, Costa SD, Ignatov A. Ultrasound-Guided Versus Wire-Guided Breast-Conserving Surgery for Nonpalpable Breast Cancer. *Clin Breast Cancer*. 2016; 16:e1–6. doi:10.1016/j.clbc.2015.09.001. [PubMed: 26439275]
12. Volders JH, Haloua MH, Krekel NM, Meijer S, van den Tol PM. Current status of ultrasound-guided surgery in the treatment of breast cancer. *World J Clin Oncol*. 2016; 7:44–53. doi:10.5306/wjco.v7.i1.44. [PubMed: 26862490]
13. Liberman A, et al. Hollow silica and silica-boron nano/microparticles for contrast-enhanced ultrasound to detect small tumors. *Biomaterials*. 2012; 33:5124–5129. doi:10.1016/j.biomaterials.2012.03.066. [PubMed: 22498299]
14. Liberman A, et al. Color Doppler ultrasound and gamma imaging of intratumorally injected 500 nm iron-silica nanoshells. *ACS Nano*. 2013; 7:6367–6377. doi:10.1021/nn402507d. [PubMed: 23802554]
15. Yang J, et al. Red-luminescent europium (III) doped silica nanoshells: synthesis, characterization, and their interaction with HeLa cells. *J Biomed Opt*. 2011; 16:066012. doi:10.1117/1.3593003. [PubMed: 21721813]
16. Martinez HP, et al. Hard shell gas-filled contrast enhancement particles for colour Doppler ultrasound imaging of tumors. *Medchemcomm*. 2010; 1:266–270. doi:10.1039/c0md00139b. [PubMed: 21841967]
17. Liberman A, et al. Mechanically Tunable Hollow Silica Ultrathin Nanoshells for Ultrasound Contrast Agents. *Adv Funct Mater*. 2015; 25:4049–4057. doi:10.1002/adfm.201500610. [PubMed: 26955300]
18. DeSantis C, Ma J, Bryan L, Jemal A. Breast cancer statistics, 2013. *CA Cancer J Clin*. 2014; 64:52–62. doi:10.3322/caac.21203. [PubMed: 24114568]
19. Mullenix PS, et al. Secondary operations are frequently required to complete the surgical phase of therapy in the era of breast conservation and sentinel lymph node biopsy. *Am J Surg*. 2004; 187:643–646. doi:10.1016/j.amjsurg.2004.01.003. [PubMed: 15135683]
20. Liberman L, et al. Bracketing wires for preoperative breast needle localization. *AJR Am J Roentgenol*. 2001; 177:565–572. doi:10.2214/ajr.177.3.1770565. [PubMed: 11517048]

Figure 1. Experimental Design, Nanoshells, and Timeline

(A) Schematic of experimental design. Each New Zealand White rabbit was injected with 1cc of VX2 tumor in bilateral thighs. Tumors were marked with NS or wire for subsequent surgical removal under ultrasound guidance. (B) TEM image of the 500 nm NS prior to injection. (C) Timeline for tumor injection, growth, marking and recovery of the VX2 model in 21 white New Zealand rabbits.

Figure 2. Color Doppler In-Vivo Nanoshell Imaging

Intraoperative ultrasound image of non palpable tumor (arrows) after marking with 500nm silica nanoshells 24 hours prior to excision in (A) Doppler mode and (B) simultaneous B mode. (C) Intraoperative image of nanoshells (NS) in tumor through skin. (D) Same tumor with probe directly on the muscle surrounding marked tumor. (E) Representative image of excised tumor with NS. (F) Representative image of NS found in cavity of excised tumor for margin removal.

Figure 3. Tissue histology of wire guided (WG) and nanoshell (NS) localized surgeries (A) WG excised tumor with positive margins at the arrow (100×). (B) Infiltrating tumor seen in NS marked sample invading the surrounding muscle at the arrow (100×). (C) Post mortem example of regrowth after excision in WG surgery (200×) and (D) Post mortem example of a tumor bed without evidence of regrowth (200×).

Figure 4. Histogram Comparison between Wire and nanoshell (NS) strategies for ultrasound guided tumor marking and resection

The tumors marked with nanoshells (NS) were able to be successfully seen in 100% of all excisions while the wire was displaced and unable to be utilized in 4 surgeries ($p < .05$). The tumor was successfully excised in comparable numbers when comparing the wire to the NS. Margin analysis and regrowth analysis found that NS were similar for both types of marking strategies.