Published in final edited form as:

Inf Process Med Imaging. 2015 ; 24: 662–674.

Segmenting the Brain Surface from CT Images with Artifacts Using Dictionary Learning for Non-rigid MR-CT Registration

John A. Onofrey¹, Lawrence H. Staib^{1,2,3}, and Xenophon Papademetris^{1,3}

¹Department of Diagnostic Radiology, Yale University, New Haven CT 06520, USA

²Department of Electrical Engineering, Yale University, New Haven CT 06520, USA

³Department of Biomedical Engineering, Yale University, New Haven CT 06520, USA

Abstract

This paper presents a dictionary learning-based method to segment the brain surface in post-surgical CT images of epilepsy patients following surgical implantation of electrodes. Using the electrodes identified in the post-implantation CT, surgeons require accurate registration with pre-implantation functional and structural MR imaging to guide surgical resection of epileptic tissue. In this work, we use a surface-based registration method to align the MR and CT brain surfaces. The key challenge here is not the registration, but rather the extraction of the cortical surface from the CT image, which includes missing parts of the skull and artifacts introduced by the electrodes. To segment the brain from these images, we propose learning a model of appearance that captures both the normal tissue and the artifacts found along this brain surface boundary. Using clinical data, we demonstrate that our method both accurately extracts the brain surface and better localizes electrodes than intensity-based rigid and non-rigid registration methods.

Keywords

segmentation; dictionary learning; skull stripping; computed-tomography; non-rigid registration; multi-modal; image-guided surgery

1 Introduction

Epilepsy is a neurological disorder in which brain function is temporarily disrupted by seizures. For certain patients who do not respond favorably to medication, surgical resection of ictal tissue, *e.g.* tissue involved in the generation of the seizures, can be an effective method to reduce or eliminate seizures. As part of a diagnostic procedure, neurosurgeons sometimes perform a craniotomy and surgically implant electrodes on the brain surface (intra-cranial) and in the brain (depth) throughout regions of the brain suspected to be causing the seizures. Clinicians monitor brain activity from these electrodes to localize ictal events [15]. However, before performing a resection, surgeons must determine if the electrode locations correspond to functionally eloquent brain tissue, *e.g.* motor, sensory, and language regions, which may be identified by pre-operative imaging, including anatomical magnetic resonance imaging (MRI) and functional imaging modalities such as functional MRI (fMRI), PET, and SPECT. Therefore, accurate localization of the electrodes with respect to this pre-operative imaging is crucial for surgical planning.

The current best method to co-register the implanted electrodes and the pre-implantation imaging involves acquisition of both an anatomical MRI and an X-ray computed tomography (CT) image after implantation. The electrodes are easily identified in the post-implantation CT images and can be interactively labeled by a trained technologist. While the post-op CT image can easily be fused with the post-op MRI using intensity-based registration methods to visualize electrodes within the post-op MRI [2] (negligible non-rigid deformations occur between the two image acquisitions), we ultimately want to visualize electrodes with respect to intra-subject, multi-modal, pre-implantation imaging in order to integrate this information into the surgical plan. However, the post-op CT image cannot easily be registered with the pre-op MRI because the surgical procedure can result in non-rigid deformations of the brain surface often larger than 1cm [8]. Intensity-based registration of pre-op MRI and post-op CT is difficult because of (i) CT's poor soft tissue contrast lacking salient anatomical structure, (ii) missing anatomical correspondences, such as removal of the skull during surgery, and (iii) imaging artifacts and intensity inhomogeneities caused by the presence of the implanted electrodes. Therefore, the CT image is first rigidly registered to the post-op MRI, and then the post-op MRI is non-rigidly registered to the pre-op MRI image to compensate for the post-surgical deformations. While the post-op MRI also contains artifacts caused by the implanted electrodes, the excellent soft tissue contrast provided by MRI makes enough anatomical structure clearly visible that accurate non-rigid registration to the relatively high quality pre-op MRI is possible. The identified electrodes are then transformed to the pre-implantation imaging space and may be co-visualized with the multi-modal pre-implantation image studies.

Unfortunately, acquiring a post-implantation MRI presents both an inconvenience for the patient, who has electrode wires sticking out of their head, and an additional potential source of infection for the patient as they are moved about the hospital and scanner, as well as an additional expense. Thus, the post-implantation MRI is not available at all institutions. A method for accurate non-rigid registration of post-op CT to the pre-op MR is necessary and remains an open area of research. Intensity-based registration of pre-op MRI and post-op CT is difficult because the post-op CT is of poor quality due to artifacts and relatively poor soft tissue contrast. With these confounding factors, non-rigid registration was actually shown to perform worse than rigid registration in terms of electrode localization error [10]. Onofrey *et al.* [10] attempt to bypass this post-op MRI by building a statistical modeled of non-rigid procedure-induced deformation to directly register the post-op CT image to the pre-op MRI. While this method successfully reduced electrode localization errors, it relies upon registration of all images to a common reference space to model the statistical deformations, which makes it complex for clinical use. In this work, we propose using a surface-based registration scheme to directly non-rigidly register the post-op CT and pre-op MRI.

The key challenge in registering the post-implantation CT and the pre-implantation MRI is not the registration itself, as numerous surface-based registration methods exist [4,11]. Instead, the extraction of the cortical surface from the CT image presents a challenge due to (i) the absence of parts of the boundary at some points, *e.g.* sections where the skull was removed for craniotomy; and (ii) the presence of imaging artifacts, *e.g.* due to implanted electrodes. Well-tested methods exist for extracting the brain surface from MRIs [14], but no tool exists yet to accurately extract the brain surface from post-implant CT images, where

the appearance of the border is varied and likely non-Gaussian. To extract the brain surface from post-implantation CT images, we propose learning a sparse representation of the brain cortical surface appearance in a data-driven manner. Huang *et al.* [9] use a dictionary-learning framework to model textural appearance both inside and outside cardiac boundaries for intra-subject tracking of those contours in time sequence ultrasound images. In Sec. 2.2 we adopt this methodology to train two dictionaries that model the appearance of regions inside and outside the brain surface. This work differs from Huang *et al.* [9] in that we propose using locally oriented image patches along the surface, as opposed to image patches canonically aligned with the image axes, to provide a more invariant model of appearance relative to the surface of interest. Brown *et al.* [3] made use of locally oriented image patches for image matching, but our method differs in that we use the geometry of our estimated surface segmentation to orient the patches instead of using image intensity information to determine the orientation. We also train our dictionary appearance model for inter-subject segmentation by building the appearance dictionaries using a set of clinical training data, and estimate the surface differently. Sec. 2.3 details how we use these appearance dictionaries to classify points along the brain surface as belonging to either inside or outside the brain surface and use this information to extract the cortical surface. We then describe how we use the segmented cortical surface to non-rigidly register post-implantation CT images with pre-implantation MR images in Sec. 2.4. Fig. 1 illustrates and summarizes our proposed training, segmentation, and registration framework. Finally, Sec. 3.1 presents our results demonstrating accurate segmentation of the cortical surface from post-implantation CT images. Sec. 3.2 then shows improved registration performance using this segmented surface to directly register post-op CT and pre-op MRI compared to intensity-based registration.

2 Methods

From a clinical database of $N = 18$ epilepsy patients at our institution, we have a set of images $\mathcal{I} = \{I_{MR,i}^1, I_{MR,i}^2, I_{CT,i}^2 \mid i=1, \dots, N\}$, where $I_{m,i}^t$ denotes pre-op images acquired at time $t = 1$ and post-op images acquired at time $t = 2$ using imaging modality m for patient i . Following the current practice, for each patient we (i) create pre-op and post-op brain surfaces S_i^1 and S_i^2 from the MR images by extracting isosurfaces of the brain masks generated by using the Brain Extraction Tool (BET) [14], (ii) rigidly register $I_{CT,i}^2$ to $I_{MR,i}^2$ to produce the transformation $T_{CT \rightarrow MR,i}$ by maximizing the normalized mutual information (NMI) similarity metric [16], and (iii) non-rigidly register $I_{MR,i}^2$ to $I_{MR,i}^1$ using a free-form deformation (FFD) [13] with 30mm B-spline control point spacing and maximizing their NMI to produce the transformation $T_{MR \rightsquigarrow MR,i}$. We make use of this data to train our model of brain surface appearance in post-surgical CT images.

2.1 Oriented local image appearance

Let $I: \Omega_I \subset \mathbb{R}^3 \mapsto \mathbb{R}$ be the 3D image that maps points from the spatial domain Ω_I to image intensity values. We define the set of intensities in an orientable local image patch Φ centered about $\mathbf{u} \in \Omega_I$.

$$\Phi(\mathbf{u}, \sigma) = \{I(\mathbf{u} + \sigma \mathbf{R} \mathbf{t}) | \forall \mathbf{t} \in \Theta\}. \quad (1)$$

Here, $\mathbf{R} \in \mathbb{R}^{3 \times 3}$ is a rotation matrix consisting of a set of orthonormal basis vectors and Φ is a set of d image intensity values sampled at patch template points $\Theta = \{\mathbf{t}_i | \mathbf{t}_i \in \mathbb{R}^3, i = 1, \dots, d\}$, whose physical size is controlled by a scale term $\sigma \in \mathbb{R}$. Typically, *standard* image patches are aligned with the image axes such that $\mathbf{R}$ uses the standard basis $\mathbf{R} = \mathbf{I}$, the identity matrix, and Φ consists of an isotropic grid of sample points centered about the patch origin. For example, a $5 \times 5 \times 5$ isotropic image patch Φ consists of $d = 125$ sample points Θ arranged in a grid about the patch origin.

In this work, we use *oriented* image patches, where the orientation of the patch $\mathbf{R}$ is determined by the data. Since we are interested in building a model of image appearance both inside and outside the brain surface, we orient image patches according to local surface geometry. For each point on the surface of interest $\mathbf{u} \in \mathcal{S}$, we compute the local surface normal $\mathbf{n} \in \mathbb{R}^3$ and the directions of principal curvature $\mathbf{k}_1, \mathbf{k}_2 \in \mathbb{R}^3$ [6]. These normalized vectors form an orthonormal basis $\mathbf{R} = [\mathbf{n} | \mathbf{k}_1 | \mathbf{k}_2]$ with which we orient the patch $\Phi(\mathbf{u})$ in (1). Fig. 2 provides an illustrative comparison between standard and oriented patches at corresponding points along a surface $\mathbf{u} \in \mathcal{S}$. By orienting the patches in this manner, the texture patterns in our patches exhibit greater invariance to changes in location along the surface $\mathcal{S}$.

2.2 Training the cortical surface appearance model

To learn our model of brain boundary appearance, we first map the segmented post-op brain surface to the post-op CT image space, $S_i^{2'} = T_{CT \rightarrow MR, i}^{-1} \circ S_i^2$, where $\circ$ is the transformation operator. We then create a sparse representation model $\mathbf{D}_{in}$ (dictionary) of the intensities inside of the brain and a model $\mathbf{D}_{out}$ of the region just outside the brain [9]. These dictionaries will capture the varieties of textural appearance found near the brain surface boundary. We create these models by extracting a training set of overlapping local image patches inside, $\Phi_{in}(\mathbf{u} - \alpha \mathbf{n})$, and outside, $\Phi_{out}(\mathbf{u} + \alpha \mathbf{n})$, points on the surface $\mathbf{u} \in S_i^{2'}$, where $\mathbf{n}$ is the outward facing local surface normal and $0 < \alpha \leq 3.0\text{mm}$ defines a narrow band region. As in Sec. 2.1, we use the local surface geometry at $\mathbf{u}$ to orient the image patches.

For each appearance class $c = \{\text{in}, \text{out}\}$, we create appearance vectors $\mathbf{p}_c \in \mathbb{R}^d$ for each image patch by concatenating the patch values from the set Φ_c , where d is the sample dimensionality determined by the chosen patch sampling template Θ in Sec. 2.1, and normalize $\mathbf{p}_c$ to have unit length. Then, we model the distribution of $\mathbf{p}_c$'s from all N training images using an overcomplete dictionary $\mathbf{D}_c \in \mathbb{R}^{d \times n}$ such that $\mathbf{p}_c \approx \mathbf{D}_c \boldsymbol{\gamma}$ [1]. Here, n is the number of dictionary atoms and $\boldsymbol{\gamma}$ is the sparse dictionary weighting coefficients. To reconstruct a given appearance sample $\mathbf{p}$ (normalized to have unit length) from the dictionary with a given target sparsity constraint Γ_0 , we solve the sparse coding problem

$$\min_{\gamma} \|\mathbf{p} - \mathbf{D}\gamma\|_2^2 \text{ s.t. } \|\gamma\|_0 \leq \Gamma_0, \quad (2)$$

using an orthogonal matching pursuit (OMP) algorithm[12]. Next, we define the residual error

$$R_c(\mathbf{p}) = \|\mathbf{p} - \mathbf{D}_c\gamma\|_2 \quad (3)$$

for both the inside and outside region classes. Using normalized appearance vectors, $0 \leq R_c(\mathbf{p}) \leq 1$, where values of 0 correspond to perfect signal reconstruction (strong membership to class c) and values of 1 indicate $\mathbf{p}$ could not be reconstructed by $\mathbf{D}_c$ (poor membership to class c). Intuitively, $R_{\text{in}}(\mathbf{p}) < R_{\text{out}}(\mathbf{p})$ for points inside the cortical surface boundary. In contrast to Huang *et al.* [9]'s multi-scale modeling where multi-scale data is concatenated into appearance vectors, we perform this sampling and training procedure individually for each level $k = 1, \dots, K$ in a multi-resolution Gaussian image pyramid, which means that we train K dictionary pairs $\{\mathbf{D}_{\text{in}}^k, \mathbf{D}_{\text{out}}^k\}$.

2.3 Segmenting the cortical surface in post-implantation CT images

Given a new pair of pre-implantation MRI and post-implantation CT test images not in the training set $I_{\text{MR}}^1, I_{\text{CT}}^2 \notin \mathcal{I}$, we perform an initial brain surface segmentation estimate $\hat{S}$ using an intensity-based rigid registration of I_{MR}^1 and I_{CT}^2 and transforming the segmented MR surface S^1 to post-op imaging space, *i.e.* $\hat{S} = T_{\text{MR} \rightarrow \text{CT}} \circ S^1$. For a point on the estimated surface $\mathbf{u} \in \hat{S}$, we extract the oriented local image patch $\Phi(\mathbf{u})$ to create the appearance vector $\mathbf{p}$, and then compute the difference of the appearance model residuals from (3) such that

$$D(\mathbf{p}) = R_{\text{out}}(\mathbf{p}) - R_{\text{in}}(\mathbf{p}). \quad (4)$$

Intuitively, if $\mathbf{u}$ lies within the true boundary of the cortical surface in the CT image then $D(\mathbf{p}) > 0$, and if $\mathbf{u}$ is outside the true boundary then $D(\mathbf{p}) < 0$. Thus, the cortical surface boundary is located at the point $\mathbf{u}$ where $D(\mathbf{p}) = 0$. We therefore seek the surface that minimizes the objective function

$$\hat{S} = \min_S \int_S \|D\|_2 dS. \quad (5)$$

To solve this optimization problem, we proceed to segment the brain from I_{CT}^2 by iteratively updating the surface in a greedy manner. At each iteration t , our surface update follows

$$\mathbf{u}_{t+1} = \mathbf{u}_t \sigma(D(\mathbf{p}_t)) \mathbf{n}_t + \mathbf{r}_t, \forall \mathbf{u}_t \in \hat{S}_t \quad (6)$$

where $\sigma \in \mathbb{R}_{>0}$ is a scale term determined by the image resolution, and $\mathbf{r}_l$ is a regularization update vector that maintains surface smoothness with respect to the local surface curvature as done by BET [14]. As the surface estimate approaches the true boundary of the cortical surface, $\|\sigma D(\mathbf{p})\mathbf{n}_l\|$ tends to zero, and the algorithm converges. Because $-1 \leq D(\mathbf{p}) \leq 1$ from (3), this update in the direction of the surface normal $\mathbf{n}_l$ can have a maximum magnitude σ . We update the surface estimate starting at the lowest level of resolution in our multi-resolution Gaussian image pyramid representation of I_{CT}^2 and proceed for a fixed number of iterations before switching to the next higher resolution level. Proceeding in this multi-resolution manner, the algorithm first updates the surface with large changes to the surface and then successively refines the results at high levels of resolution until converging.

2.4 Non-rigid MR-CT registration

Once we have a final estimate of the cortical brain surface $\hat{S}$ in the post-implantation CT image I_{CT}^2 , we then seek to non-rigidly register this image to the pre-implantation MR image I_{MR}^1 with previously segmented surface S^1 . To do so, we use a non-rigid robust point matching (RPM) algorithm [4,11] that uses a free-form deformation (FFD) model for surface-based registration of $\hat{S}$ to S^1 . We choose RPM for this task because the RPM metric allows for fuzzy correspondences between points on the two surfaces and can accommodate outlier points, which makes the algorithm robust to small segmentation errors that could exist in the reference and transform surfaces. The RPM registration algorithm results in a transformation $T_{CT \rightsquigarrow MR}$ with which we can then directly warp the post-implantation CT image data back to the pre-op MR image space $I_{CT \rightsquigarrow MR} = T_{CT \rightsquigarrow MR} \circ I_{CT}^2$.

3 Results

To validate our proposed approach, we selected 18 patients with pre-op and post-op MR images, and post-op CT images from a clinical database of epilepsy patients who had electrodes surgically implanted. A trained technician segmented the pre-op and post-op MR brain masks using BET [14], and performed manual corrections, if necessary, according to our clinical protocol. For computational efficiency, we resampled all images to have 1mm^3 isotropic resolution. All images had different volume dimensions. In this work, we parameterized each of our surfaces S as a triangulated mesh. We performed 6-fold cross-validation, where we partitioned the 18 datasets into 6 subsets of $N=15$ subjects for training and 3 subjects for testing. We quantitatively evaluated segmentation accuracy between two surface points A and B using the following metrics: (i) Dice overlap

$$\text{DICE}(A, B) = \frac{2V_A \cap V_B}{V_A + V_B},$$

where V_S is the volume enclosed by surface S ; (ii) Hausdorff Distance

$$\text{HD}(A, B) = \max \left\{ \max_{\mathbf{a} \in A} d(\mathbf{a}, B), \max_{\mathbf{b} \in B} d(\mathbf{b}, A) \right\};$$

and (iii) Mean Absolute Distance

$$\text{MAD}(A, B) = \frac{1}{2} \left\{ \frac{1}{N_A} \sum_{\mathbf{a} \in A} d(\mathbf{a}, B) + \frac{1}{N_B} \sum_{\mathbf{b} \in B} d(\mathbf{b}, A) \right\},$$

where $d(\mathbf{a}, B) = \min_{\mathbf{b} \in B} \|\mathbf{b} - \mathbf{a}\|_2$. Additionally, we specifically quantified segmentation errors at the identified electrode locations $\mathbf{x}_e \in E \subset \Omega_f$. We define the Mean Electrode to Surface Distance

$$\text{MESD}(E, S) = \frac{1}{N_E} \sum_{\mathbf{x}_e \in E} d(\mathbf{x}_e, S)$$

to be a measure of the distribution of electrode distances from the given surface S . We expect implanted intra-cranial cortical electrodes to lie on the surface of the brain, thus the MESD metric quantifies how close the electrodes are to the surface S . High MESD values indicate that the distribution of electrodes is far from the surface, which means they could appear as floating above the surface or sunk within. To calculate MESD, we extracted only the largest 8×8 electrode grids (64 total electrodes per patient) from each patient's set of electrodes as these grids lie on the surface closest to the craniotomy location where we expect the largest amount of brain deformation [7].

3.1 CT cortical surface segmentation

For each fold in our cross-validation study, we tested how our proposed cortical boundary learning (Sec. 2.2) and segmentation (Sec. 2.3) performed using oriented local image patches (Sec. 2.1) by comparing our results to those found using local image patches using standard orientation, *e.g.* $\mathbf{R} = \mathbf{I}$ in (1). For each of these two methods, we used a multi-resolution Gaussian image pyramid with $K = 3$ levels. For each level, we created appearance dictionaries of size $n = 256$ with sparsity constraint $\Gamma_0 = 2$. Local image patches $\Phi(\mathbf{u}, \sigma)$ used point sample templates Θ with $5 \times 5 \times 5$ isotropic grids centered around each surface vertex $\mathbf{u} \in S$, thus the appearance samples had dimensionality $d = 125$. We adjusted the spacing between the Θ sample points for each resolution level k , *e.g.* $k = 1, 2, 3$ had scales $\sigma = 1, 2, 4\text{mm}$, respectively. By doing so, image patches at high image scales have larger physical sizes. The training surface meshes $S_i^{2'}$ had on average 2685 ± 188 vertices, and we extracted 3 times this many patch samples from the N training CT images to train the inside and outside dictionaries in Sec. 2.2.

To segment the 3 test images in each cross-validation fold using the method described in Sec. 2.3, we set the number of update iterations to $t = 150$ for each resolution level k . The testing surface meshes $\hat{S}$ used in Sec. 2.3 had on average 2740 ± 197 vertices. Fig. 3 shows example qualitative segmentation results from 4 subjects using both the standard patch orientation and our proposed oriented patches. Segmentation using oriented patches appeared to extract the cortical surface better than using standard image patches as the standard patch method had trouble correctly localizing the cortical surface in areas where the skull was removed.

To quantify how well our proposed method extracted the cortical surface, we evaluated our segmented surface estimate $\hat{S}_i$ with respect to the rigidly registered post-op MR surface $S_i^{2'}$ as a gold-standard. For each subject i , we computed our 3 surface evaluation measures comparing our surface estimate $\hat{S}_i$ with the reference surface $S_i^{2'}$. We compared three different estimated test surfaces $\hat{S}_i$: (i) the initial, rigidly registered pre-op surface, *i.e.* no segmentation updates; (ii) the surface estimate using standard image patch orientation; and (iii) the surface estimate using oriented image patches. Table 1 summarizes the distribution of these results. Using oriented image patches provided significantly better segmentation results (two-tailed paired t-test $p < 0.05$) in terms of higher $\text{DICE}(\hat{S}_i, S_i^{2'})$, and lower $\text{HD}(\hat{S}_i, S_i^{2'})$, $\text{MAD}(\hat{S}_i, S_i^{2'})$, and $\text{MESD}(E_i, \hat{S}_i)$ values on average. Fig. 4 shows an example comparing these 3 surface estimates along with electrodes identified in the CT images to illustrate how our proposed method estimates a more physically plausible surface on which the electrodes lie.

We implemented the training and segmentation algorithms in MATLAB. The running time for the segmentation using oriented patches was 11.46 ± 0.82 minutes on average, and 2.94 ± 0.31 minutes on average using standard patch orientation. The main reason for the increase in computation time using oriented patches is computation of the principal curvatures.

3.2 Non-rigid MR-CT registration

Using the CT brain surface segmentation results $\hat{S}_i$ found by our proposed algorithm, we registered the post-implantation CT image I_{CT}^2 directly to the pre-implantation MRI I_{MR}^1 with a robust point matching (RPM) surface registration algorithm (Sec. 2.4). We compared our results to intensity-based registration where we registered the post-op CT image to the pre-op MRI using (i) rigid, (ii) affine, and (iii) non-rigid free-form deformation (FFD) [13] transformations T_i . Both the non-rigid RPM and non-rigid intensity FFDs had isotropic B-spline control point spacing set to 30mm. The intensity registration methods used a multi-resolution image pyramid and conjugate gradient optimization to maximize the normalized mutual information (NMI) similarity metric [16].

We evaluated registration performance with respect to the pre-op MR brain surface S_i^1 and the transformed post-op CT surface estimate $T_i \circ \hat{S}_i$ by calculating $\text{DICE}(S_i^1, T_i \circ \hat{S}_i)$ and $\text{MAD}(S_i^1, T_i \circ \hat{S}_i)$. Table 2 shows that on average non-rigid RPM registration of the cortical surface segmented using our proposed method outperforms intensity-based non-rigid registration. More importantly, our proposed method had a much smaller standard deviation of Dice values and greater minimum Dice value, 92.08 for our method compared to 82.72 for intensity FFD. Interestingly, intensity FFD had worse minimum performance than intensity-based rigid and affine registration. Non-rigid RPM registration also had lower mean MAD performance than intensity FFD (two-tailed paired t-test $p = 0.13$). Additionally, we quantified how close the electrodes E_i identified in the CT image were to the MR brain surface after registration, *i.e.* $\text{MESD}(T_i \circ E_i, S_i^1)$. Table 2 shows that electrodes were significantly closer to the segmented pre-op MR surface using non-rigid RPM than the intensity-based registration methods (two-tailed paired t-test $p < 0.05$). These results were to

be expected given that the RPM registration explicitly seeks to align the two surfaces, while intensity-based registration does not enforce such a constraint.

4 Conclusion

The method proposed in this paper can be used to solve an actual clinical problem using imperfect clinical data. Our proposed method accurately extracts the brain surface from post-electrode implantation CT images that contain many artifacts, distortions, and missing anatomical features. By orienting local images, we create a dictionary-based model of the brain boundary's appearance that is more invariant to image orientation changes in comparison to appearance models that use local image patches aligned with the image axes. The resulting appearance model, when used with our proposed surface estimation method, significantly improves segmentation performance (Table 1). Then, using this brain surface estimate, we demonstrate improved non-rigid registration of post-implantation CT images to pre-implantation MRI in comparison to intensity-based registration methods. Further testing is necessary with respect to the algorithm parameters, in particular, the size of the dictionaries, the sparsity constraint term, and the local image patch size, in order to determine optima settings. Furthermore, our segmentation method does not incorporate any prior information regarding the shape of the brain surface. In the future, we aim to combine our locally oriented appearance model with a statistical model of brain shape [5] to better constrain the surface estimation process. The approach presented in this paper could be applied to segment other anatomical structures of interest whose appearance has a highly non-Gaussian distribution, *e.g.* due to post-surgical appearance changes caused by scar tissue or implants.

References

1. Aharon M, Elad M, Bruckstein A. K -svd: An algorithm for designing overcomplete dictionaries for sparse representation. *Signal Processing, IEEE Transactions on*. 2006; 54(11):4311–4322.
2. Azarion AA, Wu J, Davis KA, Pearce A, Krish VT, Wagenaar J, Chen W, Zheng Y, Wang H, Lucas TH, Litt B, Gee JC. An open-source automated platform for three-dimensional visualization of subdural electrodes using CT-MRI coregistration. *Epilepsia*. 2014
3. Brown M, Szeliski R, Winder S. Multi-image matching using multi-scale oriented patches. *CVPR*. 2005; 1:510–517.
4. Chui H, Rangarajan A. A new point matching algorithm for non-rigid registration. *Computer Vision and Image Understanding*. 2003; 89(2–3):114–141.
5. Cootes T, Taylor C, Cooper D, Graham J. Active shape models-their training and application. *Computer Vision and Image Understanding*. 1995; 61(1):38–59.
6. Do Carmo, MP. *Differential geometry of curves and surfaces*. Englewood Cliffs: Prentice-Hall; 1976.
7. Hartkens T, Hill D, Castellano-Smith A, Hawkes D, Maurer CRJ, Martin A, Hall W, Liu H, Truwit C. Measurement and analysis of brain deformation during neurosurgery. *IEEE TMI*. 2003; 22(1): 82–92.
8. Hill DL, Maurer CRJ, Maciunas RJ, Barwise JA, A Fitzpatrick MJ, Wang MY. Measurement of intraoperative brain surface deformation under a craniotomy. *Neurosurgery*. 1998; 3:514–526.
9. Huang X, Dione DP, Compas CB, Papademetris X, Lin BA, Bregasi A, Sinusas AJ, Staib LH, Duncan JS. Contour tracking in echocardiographic sequences via sparse representation and dictionary learning. *Medical Image Analysis*. 2014; 18(2):253–271. [PubMed: 24292554]

10. Onofrey, JA., Staib, LH., Papademetris, X. Learning nonrigid deformations for constrained multi-modal image registration. Vol. 8151. MICCAI; 2013. p. 171-178.
11. Papademetris X, Jackowski AP, Schultz RT, Staib LH, Duncan JS. Computing 3D non-rigid brain registration using extended robust point matching for composite multisubject fMRI analysis. MICCAI. 2003; 2879:788–795.
12. Rubinstein R, Zibulevsky M, Elad M. Efficient implementation of the K-SVD algorithm using batch orthogonal matching pursuit. Tech. rep., Technion. 2008
13. Rueckert D, Sonoda L, Hayes C, Hill D, Leach M, Hawkes D. Nonrigid registration using free-form deformations: application to breast MR images. IEEE TMI. 1999; 18(8):712–721.
14. Smith SM. Fast robust automated brain extraction. Human Brain Mapping. 2002; 17(3):143–155. [PubMed: 12391568]
15. Spencer, SS., Sperling, M., Shewmon, A. Intracranial electrodes. Philadelphia: Lippincott-Raven; 1998. Epilepsy, a Comprehensive Textbook, chap; p. 1719-1748.
16. Studholme C, Hill D, Hawkes D. An overlap invariant entropy measure of 3D medical image alignment. Pattern Recognition. 1999; 32(1):71–86.

Fig. 1.

Dictionary-based appearance learning and segmentation of post-surgical cortical brain surface in CT images for direct non-rigid registration of with pre-surgical MRI. We use a training set of post-implantation CT image to learn two models of image appearance, one inside the cortical surface and one outside. We then use these models to segment the cortical surface from test CT images. Using the extracted post-op surface, we perform surface-based registration with the pre-op MRI in order to co-visualize post-implantation electrodes with pre-op imaging data.

Fig. 2.

Examples of local oriented image patches compared to their corresponding standard local image patches oriented along the image axes. For the surface S , the local surface normal $\mathbf{n}$ and directions of principal curvature $\mathbf{k}_1$, $\mathbf{k}_2$ define a local orthonormal basis by which we orient the image patches. We show patch examples from both inside and outside the cortical surface.

Fig. 3.

Cortical surface segmentation results using (i) our proposed appearance method using local image patches oriented by the surface estimate's geometry (orange contour), and (ii) the standard appearance method using local image patches oriented along the image axes (blue contour). Sagittal (top row), coronal (middle row), and axial (bottom row) slice of 4 example subjects. Arrows highlight areas where our approach better identified the cortical surface.

Fig. 4.

Visualizing electrodes on the segmented cortical brain surface mesh. In this example (subject 2), (left) the electrodes initially appear embedded within the surface, and (middle) using the standard segmentation approach only partially corrects the surface. Our proposed method (right) correctly segments the cortical surface and the electrodes appear to sit correctly on the surface.

Table 1

Segmentation evaluation quality measures for cortical surface segmentation using our proposed method compared to the initial surface and dictionary-based segmentation using standard images patches.

Method	Dice (%)	HD (mm)	MAD (mm)	MESD (mm)
Initial	$93.38 \pm 2.34^{\dagger, \ddagger}$	$17.29 \pm 8.20^{\ddagger}$	3.22 ± 0.91	$3.16 \pm 1.15^{\ddagger}$
Standard	$94.24 \pm 1.38^{\ddagger, *}$	$16.84 \pm 5.06^{*}$	$3.32 \pm 0.53^{*}$	$2.83 \pm 1.57^{*}$
Oriented	$95.08 \pm 0.82^{\ddagger, *}$	$13.21 \pm 5.16^{\ddagger, *}$	$2.99 \pm 0.36^{*}$	$0.95 \pm 0.25^{\ddagger, *}$

Values are expressed as Mean $\pm$ SD.

$\dagger, \ddagger$ and $*$ indicate statistically significant results (two-tailed paired t-test $p < 0.05$) between the two methods indicated by the symbol.

Table 2

Registration evaluation quality measures for direct non-rigid registration of post-op CT images to pre-op MRI. We computed rigid, affine, and non-rigid registration results using (i) RPM registration of the CT brain surface (found using our proposed segmentation method) to the segmented MR brain surface, and (ii) intensity-based registration of the two images.

Method	Dice (%)	Min. Dice (%)	MAD (mm)	MESD (mm)
Intensity Rigid	93.40 ± 2.36	87.27	3.02 ± 0.99	3.12 ± 1.16
Intensity Affine	92.69 ± 2.89	84.48	3.21 ± 1.15	3.56 ± 1.17
Intensity FFD	93.25 ± 3.45	82.72	3.07 ± 1.37	2.31 ± 0.91
RPM Rigid	92.84 ± 2.42	86.22	3.21 ± 1.05	2.78 ± 1.06
RPM Affine	93.23 ± 1.36	90.21	3.10 ± 0.84	1.85 ± 0.59
RPM Non-rigid	94.44 ± 0.93	92.08	2.72 ± 0.65	1.55 ± 0.37

Values are expressed as Mean±SD.