Published in final edited form as:

Methods Cell Biol. 2014 ; 123: 135–151. doi:10.1016/B978-0-12-420138-5.00008-2.

Assessing and benchmarking multiphoton microscopes for biologists

Kaitlin Corbin*, Henry Pinkard*, Sebastian Peck*,¹ Peter Beemiller†, and Matthew F. Krummel*

*Biological Imaging Development Center and Department of Pathology, University of California, San Francisco, California, USA

†Department of Pathology, University of California, San Francisco, California, USA

Abstract

Multiphoton microscopy has become staple tool for tracking cells within tissues and organs due to superior depth of penetration, low excitation volumes, and reduced phototoxicity. Many factors, ranging from laser pulse width to relay optics to detectors and electronics, contribute to the overall ability of these microscopes to excite and detect fluorescence deep within tissues. However, we have found that there are few standard ways already described in the literature to distinguish between microscopes or to benchmark existing microscopes to measure the overall quality and efficiency of these instruments. Here, we discuss some simple parameters and methods that can either be used within a multiphoton facility or by a prospective purchaser to benchmark performance. This can both assist in identifying decay in microscope performance and in choosing features of a scope that are suited to experimental needs.

INTRODUCTION: PRACTICAL QUANTITATIVE 2P BENCHMARKING

Benchmarking and comparison of multiphoton microscopes have traditionally had little rhyme or reason. It is not uncommon for a biologist to make claims of depth of penetration such as his or her microscope “is sensitive to 500 μm ” as an attempted method of comparison. However, such a metric clearly depends on many factors, not the least of which is the nature of the sample. Specifically, intensity of the fluorophore, intrinsic autofluorescence and particularly dispersion and scatter within a tissue of interest all contribute extensively to such a metric. It is also possible to illuminate a biological sample with sufficient power to make a single observation at significant depth, but which effectively destroys the sample in the process. Multiphoton illumination does produce photodamage, of course, only less than equivalent single-photon illumination that would be required to illuminate a fluorophore when dispersion and scatter are present.

We have found that the lack of routine quantitative measurements of key components of microscope systems makes rational purchasing decisions difficult and troubleshooting/

¹Present address: Nikon Instruments, Inc., 1300 Walt Whitman Road, Melville, New York, USA.

The authors declare no competing financial interests.

maintenance uncertain. The former is important when one wishes to independently assess the claims of commercial scopes. The latter is important for keeping microscopes in optimal order and in evaluating the source of poor imaging quality from users of a given microscope. In this review, we discuss some of the methods that we have come to use that allow us to keep track of the quality of microscopes within a lab or shared facility. We have also used these methods for purchasing decisions and we discuss both applications.

8.1 PART I: BENCHMARKING INPUTS

Benchmarking a microscope is similar to conducting a controlled experiment; the most important aspect is to keep key parameters constant. One crucial example is laser power at the sample; every microscope can produce brighter images with lower noise using higher laser power, but such power increase comes with a predictable and fairly routine increase in photodamage and photobleaching. Below, we introduce three parameters that we keep constant when benchmarking or testing scopes. The values of the parameters that we keep constant are similar but not identical to conditions used in everyday biological experiments. For example, since the increase in power per unit area within an illumination pixel in a scanning system is likely to produce fairly uniform increases across microscopes, we start by choosing a value for this parameter and holding it constant across all measurements. Below, we define three key parameters that are either maintained identically between benchmarking sessions or that can be used routinely between microscopes to allow comparison.

8.1.1 LASER POWER AT THE SAMPLE

Laser power at the sample is a measure of photon flux to the sample and produces the largest impact on sample viability of all the parameters we discuss here. It is therefore the most important parameter to keep constant when comparing instrument performance. It also represents a quick check of laser excitation alignment. Decay in the amount of excitation light reaching the sample may occur slowly on a day-to-day basis, but over long periods of time will have a negative effect on the system's overall efficiency unless the system is routinely measured for misalignment or reduction of the light reaching the sample. This measure can be a quick diagnosis for some of the most common problems on a multiphoton microscope. The long path length in multiphoton microscopes as the beam is routed on the bench top creates a number of opportunities for misalignment such as: temperature variations, accidental knocks of beam steering mirrors, or malfunctions in the laser excitation pathway.

To obtain a baseline for the performance of the laser, a power meter is placed just after the power modulator, in our case an electro-optical modulator (EOM), and the maximum output wattage is recorded. Because the maximum wattage will vary across the laser's tunable range, this is done using several commonly used wavelengths about 100 nm apart as benchmarks. Using the same wavelengths and a known amount of power after the EOM, it is expected that there will be a decrease in the laser power reaching the sample due to the objective transmission capability, overfilling the back aperture, and reflection or transmission efficiency of the primary dichroic. In our system, we have observed this decay can be as high as 40% depending upon the wavelength. Such a decrease in the amount of

light reaching the sample without a corresponding drop in laser output at the EOM can be indicative of clipping in the excitation path or dirty or misaligned optics. To measure the output then at the objective for comparison, place the power meter below the objective (on an upright microscope; above on an inverted stand) at approximately its specified working distance and record the wattage delivered at maximum output. For this purpose, we use a Thor Labs brand power meter that integrates at 20 Hz. Whichever brand of power meter selected, the device should be calibrated over time and used similarly for each benchmark so as not to introduce this as an unknown variable in comparisons. It is also important that the power meter chosen is able to handle the very high peak powers of titanium sapphire lasers. Note that some power meters will produce readings that vary slightly when used under a microscope objective so it is important to establish a routine. We have found that securing the detector to the microscope stage and moving the stage incrementally allows us to find the optimal position from which we record the highest stable value.

We note that it is important to consider the response of the modulator being used to adjust laser power. Commonly used power modulators such as EOMs and acousto-optic tunable filters often do not attenuate laser power linearly as voltage is applied, resulting in a response curve similar to Fig. 8.1. While not essential for comparing between microscopes, the generation of graphs such as Fig. 8.1 is helpful in selecting appropriate settings for biological samples when the wattage applied can adversely affect the life of the sample and fluorescence of interest. Set the modulator so that the same laser power is delivered to the sample on each microscope during a continuous scanning mode or when a single point is being continuously scanned. In our hands, 2 mW as measured at the objective is sufficient for the bead gel assay described below and is a value that we routinely use. The lymph node, our standard biological sample described, is routinely imaged using 12 mW for benchmarking.

8.1.2 PHOTOMULTIPLIER SETTINGS

Photomultiplier tubes (PMTs) are quite variable when implemented in a microscope setting, even comparing those of identical manufacturer “minimum specification,” and so benchmarking and standardizing for them can be one of the most difficult aspects of this process. This is easier when comparing the performance of a microscope over time, and we recommend choosing a single applied voltage for benchmarking and keeping it constant. When comparing entire microscope rigs with different makes or models of PMTs with different gain voltages or characteristics, this can be a bit more difficult.

To consider the problem is to consider signal and noise features in detectors generally. PMTs, like other detectors, contribute two distinct types of noise to the images they form. Dark noise can be observed when images are captured while no light falls on the detectors. Dark noise generates low or zero intensity values for the vast majority of pixels but yields stochastic, high intensity pixels that do not coincide in position with the location of fluorescent objects. Frame averaging can be an excellent method of removing this high intensity speckling since these bright pixels rarely happen in the same place across multiple frames. Shot noise (Poisson noise; Chapters 1 and 3), on the other hand, is signal dependent. As voltage is increased in an effort to increase the signal intensity, noise also increases. The

higher voltages applied in this scenario also produce higher gains for signal so some excellent images can be produced under these conditions. There are many excellent reviews (Yotter & Wilson, 2003), which are helpful in understanding the source and quantification of noise.

Although we will discuss frame averaging later, we have observed in practice that some users can mine very weak signals to detect and measure objects of interest (e.g., GFP-positive cells “identified” deep in tissue) when gains are used that produce speckles of high intensity. Thus, although the average or standard deviation of the noise may still be low, some pixels of very high intensity are found but can sometimes be accommodated. This is due to the higher gain in such a PMT giving rise to higher overall signals for the true luminescent objects and the ability to “average” out this noise at acquisition or remove single-pixel noise in postprocessing using a Gaussian kernel. But if very small objects are ultimately to be best spatially resolved it remains best to minimize this noise. Since frame averaging is essentially applied identically across all microscopes and should also never vary in how it functions (it is simply a mathematical average), for the purpose of this discussion, we will assume that a user will always do benchmarking with the same frame averaging used. We recommend single frame collection at the same frame rate in a resonant scanning system is used if testing the same microscope over time, or when comparing between scopes to determine efficiency of light detection/capture, identical criteria (notably pixel dwell time) are applied.

There are, however, the practical issues of simply obtaining the best, unambiguous detection of the objects of interest and of obtaining images that will yield quantitative and revealing data about the response of the PMTs and their role in the system as a whole. For this reason, we describe two approaches to benchmarking: the “fixed PMT voltage” method is faster and easier to perform, and the “variable PMT bias voltage” method which is more time consuming but yields a more complete characterization of the detectors, which can be useful for choosing parameters for imaging actual tissues. In general, benchmarks via the first method will mirror those in the second.

8.1.2.1 Method 1—Fixed PMT voltage—Often, end users raise PMT voltage to compensate for a poorly performing microscope in an effort to obtain adequate signal, which results in noisy images. Although having to apply high PMT voltage to detect signal may be indicative of PMT insensitivity brought on by age or damage, the source of the problem may be found elsewhere. To differentiate a loss of PMT sensitivity from a change in actual signal generated or poor collection efficiency due to alignment, we have devised a simple test.

For the purposes of the bead gel assay that we describe below, we choose a target signal intensity for the most superficially detected beads that precludes pixel saturation and adjust the PMT gain to achieve the desired value, keeping the laser power constant. In an 8-bit system (intensity values 0–255), we typically set the PMT bias voltage so that the target intensity is 240. Once this parameter has been set, we collect Z-stacks of both the dispersive and nondispersive bead samples, at approximately the Nyquist rate (Chapter 1) for the given system, that extend from most superficial beads detected until signal can no longer be distinguished from noise. Our standard data sets extend 500 μm in Z at approximately 500

and 250 nm pixel lateral resolution using a $20 \times$ or $25 \times$ water-dipping lens with a long working distance.

Using this method, the signal-to-noise ratio (SNR; Chapter 1) at a given sample depth can be calculated, and changes in the instrument's ability to detect objects deep within a sample can be identified as discussed below. The large Z-stack collected here will also be used to evaluate the point spread function (PSF; Chapters 1 and 10) of the instrument as another measure of performance.

8.1.2.2 Method 2—PMT voltage range—The second method of detector benchmarking, which more fully characterizes the response of PMTs, allows more direct comparison and perhaps optimization of individual detectors. As with the previous method, the same laser power should be used across all tests, but in this method, we will use a range of PMT voltages. Note that we will be testing changes in SNR as a function of voltage change, it is therefore important to avoid saturation (e.g., values >255 for 8-bit images). If the bead intensities become saturated, there will be no gains in signal, while noise will continue to increase, changing the response profile. Begin by collecting a small ($\sim 20 \mu\text{m}$) representative Z-stack at maximum PMT voltage, sampling at approximately the Nyquist rate. Repeat the same Z-stack acquisition adjusting the PMT voltage by 10% through the PMT's entire range. Some users may choose to acquire the same series of images using a standard biological sample as well due to greater heterogeneity in fluorescence, as it will alter the amount of shot noise produced and give a more accurate picture of how the instrument behaves in experimental settings. In this case, it may be acceptable to allow some saturation in bright areas as it allows dimmer, biologically relevant features to be detected.

8.1.3 STANDARD SAMPLES

Perhaps, the most valuable part of benchmarking is the establishment of a standard sample that can be used over time or at different physical sites. The key features are reproducibility, and representative samples tailored to the types of challenges to which multiphoton is best applied. To this end, while a standard $10\text{-}\mu\text{m}$ thick histology slide, a pollen grain slide, or immobilized bead standards can provide some insight, the best samples are thick specimens set into media that is dispersive to the same extent as biological specimens. We recommend that one establish two such samples. The first, which we have mentioned in passing above, is reproducible fluorophore-impregnated beads distributed in dispersive nonfluorescent beads, all embedded in a thick slab of hydrogel material. The fluorescent beads mimic fluorophores that might be present in a biological tissue and the second set of beads mimic the effects of biological tissues upon the mean-free path of light within tissues (Theer, Hasan, & Denk, 2003). The sample described below may roughly mimic tissue such as a lymph node. If the target tissue or organ for a study is vastly different, a modified sample (e.g., fewer dispersive beads) may be made to better mimic its properties. The second sample we will describe is a "standard biological sample," a tissue whose optical properties very faithfully represent samples to be used in final experiments but critically whose fluorescent intensity can be maintained to very close tolerances over time.

8.1.3.1 A standard three-dimensional sample set with variable dispersive properties

—We recommend, in fact, that two formats of this sample should be made. A first “nondispersive” sample will allow the comparison of the PSF of a multiphoton system. Note that due to the lower numerical aperture (Chapter 2) typically used in multiphoton imaging, and due to the longer wavelength of light, this will typically be inferior to collection in any of the other typical modalities. By collecting a Z-stack of this sample, you will observe that, were tissues not complex (dispersive), imaging depths would be limited only by the working distance of the objective lens. The second variation of this sample, “dispersive,” will demonstrate the degree to which the two-photon effect improves imaging at depths in complex tissues and, in comparison to the first version, will demonstrate the degree to which dispersive objects affect imaging deep within tissues.

8.1.3.1.1 Support protocol: Preparation of PSF beads in a dispersive or nondispersive support: Nondispersive gel:

1.5 ml	Polyacrylamide (40%, 19:1 acrylamide:bis-acrylamide)
2.25 ml	ddH ₂ O (makes 10% polyacrylamide total)
10 µl	Red beads (Invitrogen F13083 1.0 µM 1×10^{10} /ml)
10 µl	Green beads (Invitrogen F13081 1.0 µM 1×10^{10} /ml)
2 ml	Sulfate latex beads (Interfacial Dynamics/Invitrogen: 5 µm diameter. stock: 6.1×10^8)
250 µl	APS (10% solution of powder (kept frozen))
5 µl	TEMED to polymerize the gel

Dispersive gel:

1.5 ml	Polyacrylamide (40%, 19:1 acrylamide:bis-acrylamide)
4.25 ml	ddH ₂ O (makes 10% polyacrylamide total)
10 µl	Red beads (Invitrogen F13083 1.0 µM 1×10^{10} /ml)
10 µl	Green beads (Invitrogen F13081 1.0 µM 1×10^{10} /ml)
250 µl	APS (10% solution of powder (kept frozen))
5 µl	TEMED to polymerize the gel

Mix all ingredients except TEMED into a six-well dish or mold of choice. Then add the TEMED and gently swirl. Allow to polymerize. Figure 8.2 shows samples ready for use.

Note that other beads may be used but should have standardized emission spectra, brightness, and shape. We have selected 1.0 µm beads, as they are small enough to serve as an approximation of the PSF given the magnification and pixel calibration of our systems, yet are large enough to evaluate SNR. Others may choose to prepare separate samples or include smaller (0.1 µm) and larger beads (10 µm) to more critically evaluate these two parameters. Absolute knowledge of the refractive and dispersive properties is important for the sulfate latex beads or equivalent, but the most important consideration for the purposes of your own benchmarking is that the QC of these beads is consistent. Consider purchasing a lot and maintaining it in the fridge over time.

Note also that the mean-free path for the dispersive sample described above is ~300 µm, calculated with a Mie scattering theorem. In our hands, this shows intensity drop-offs with

depths that are similar to mouse lymph nodes. The full calculation of this requires more extensive knowledge of the properties of beads, their precise geometry, etc. and is typically given by:

Mean-free path for scattering

$$s=1/N\sigma_s=4/NK_s\pi d^2$$

where N is the number of spheres per unit volume, σ_s is the scattering cross section, K_s is the scattering coefficient (the ratio of the scattering to the geometrical cross section), and d is the diameter of the spheres (Churchill, Clark, & Sliepcevich, 1960).

For the purposes of direct comparisons between microscopes, the important feature is that the sample remains the same. Ultimately, the most important test of a microscope will use the tissue of interest. Individual tissues obviously have their own characteristics (e.g., mean-free path) as well as additional features such as fluorescent absorption, which may further reduce detection efficiency but should contribute similarly between microscopes.

8.1.3.2 Standard biological samples—A standard biological sample is also extremely useful in multiphoton benchmarking, and the best choices are those identical or very similar to tissues of interest. It is important for such samples to have fluorescence intensities and tissue composition that remain roughly constant over many trials and potentially many years. Each investigator may choose their own but ours consists of a lymph node from a 6- to 12-week-old C57Bl/6 mouse, containing cells from an ActinCFP (Hadjantonakis, Macmaster, & Nagy, 2002) (or UbiquitinGFP; Schaefer, Schaefer, Kappler, Marrack, & Kedl, 2001) transgenic animal. These latter donor mice produce T cells and B cells with consistent fluorescence intensity, and the use of a standard strain and age of donor mice produces organs with consistent size and density and therefore similar optical properties. It is likely that a user can find similar strains of zebrafish or *C. elegans* systems with similarly consistent levels of fluorescent protein expression. As discussed in the next section, the choice of the system that best matches your “typical” experimental system may help to choose the last imaging parameters that you will want to hold constant in your benchmarking.

8.1.4 SAMPLE-DRIVEN PARAMETERS: HOW FAST/HOW LONG

Since the goals of tissue imaging can vary, there is one last factor to consider in determining how to benchmark your microscopes. This is the issue of how much light your sample can tolerate. Laser power is one aspect of this, but damage is a function of both power and time. Some fluorophores and tissues are resistant to long-term exposure, whereas others are very photosensitive. We will not treat this issue in great detail; microscope detection capability can be benchmarked identically. However, scanning details such as long pixel dwell times can be more destructive than rapid sweeps (e.g., resonant or fast galvanometer scans). Especially, when optimizing imaging lengths, it may be useful to have prior knowledge of this feature when choosing laser power and scanning parameters for benchmarking.

As an example, for the lymph node samples, our biological standard, we collect timelapse sequences at 30-s intervals encompassing $\sim 100\ \mu\text{m}$ of z-space at 5- μm intervals. T cells (approximately 5 μm in diameter) will be assessed within the lymph node. The goal of this collection is to assess both the detection depth but also, by collecting a timelapse, to assess whether imaging conditions being tested are compatible with the biology; T cells are typically motile within lymph nodes, and it is a good control that your standard benchmarking imaging conditions for this sample be chosen to be in the range of biological compatibility.

8.2 PART II: BENCHMARKING OUTPUTS

Once a set of standard samples is identified and working parameters are established, it remains to test the microscope for the quality of images that are produced. We chose four parameters to measure routinely. First is the PSF, which allows one to observe the transfer function of a microscope. In other words, for a small point source of fluorescence in a sample, how well the microscope will transfer the intensity to an ideally small region at the detector. Second are the intertwined parameters of total intensity of an object (bead, cell) and the SNR of detection. Lastly, since detection of fluorescence deep within a sample is a desirable feature in multiphoton microscopy, we describe the use of sensitivity as a function of depth as a measured parameter.

8.2.1 THE POINT SPREAD FUNCTION

Degradation of the PSF is often the clearest indication of misalignment and optical aberration in an imaging system that is performing at a suboptimal level (Chapter 7). Figure 8.3A shows the severely aberrated intensity profile of a bead in nondispersive media acquired while the primary dichroic was warped by an aberrant holder. Although the central XY plane is comparable to that shown in Fig. 8.3B, a near ideal PSF, it is clear in the XZ and YZ views that the light propagating axially from the bead is asymmetric, distorting the shape of any detected object (Fig. 8.3C). The cross shape of the PSF is characteristic of an astigmatism (Chapter 2) introduced by a skewed optic or bowing of a mirror. The direction of laser propagation, if it is at a slight angle rather than parallel to the objective as it enters the back aperture, and position of the beam also have a direct impact on the shape of the PSF (Fig. 8.3D). Aberrations may also result from an off-axis optic, for example, in a damaged objective.

Spherical aberration is another prevalent source of degradation of the PSF in tissues. Spherical aberration occurs when light rays that pass through the outside of the optical axis are focused to a different Z position than light rays that pass through the center of the optical axis (Chapter 2). For example, consider a simple curved lens that is ground spherically, the rays that pass through the lens near the outside will be focused differently than the rays that pass through the center. This causes a large spreading of the PSF in the Z axis, and decreases resolution and intensity considerably. Modern optics are corrected for spherical aberration with a specific optical path length, but in practice, this path length invariably changes when imaging through large samples. This is why it is very important to use an objective that has an immersion media matched to the sample, or use an objective that has a correction collar

that allows you to correct for any spherical aberration. Especially, when considering the nonlinear dependence of multiphoton excitation, this can have a huge impact on the brightness and quality of your image.

Using the Z-stack acquired of beads in nondispersive media, we select a representative bead and calculate lateral and axial full-width half maximum (FWHM; Chapter 7) based on the centroid intensity (Fig. 8.3E and F). Although this method may not be sensitive enough to be applied to other systems, particularly those designed for high precision localization, it allows direct comparison between the performance of instruments. Importantly, it provides a quick method of quantification for gains made through incremental tweaks in alignment and optimization and always will affect the other two measurements we will make.

8.2.2 SNR AND TOTAL INTENSITY

Because SNR is a ratio, it is a dimensionless parameter that does not have units; it provides an easy method of comparison of the quality of images. As a first measure of system performance and sensitivity we use the large Z-stack acquired using the dispersive bead gel (method 1) to evaluate the SNR of the most superficial beads. We use Imaris (Bitplane AG) spot detection to identify beads and calculate the average signal intensity in a selected plane. This can also be done in ImageJ using the Track-Mate plug-in or creating masks, which cover the bead area (for our systems $\sim 2 \times 2$ pixel spots). Whichever method is preferred, it should be used identically for all measurements. There are several methods for calculating image noise (Paul, Kalamatianos, Duessman, & Huber, 2008; Yotter & Wilson, 2003) all of which have practical applications, but for our purposes in PMT detection where there can be high shot noise, the standard deviation of pixel intensity in areas without beads is most suitable. Two 140×140 pixel regions of interest (ROIs) in a single plane free of beads were selected and used to calculate the standard deviation of the noise (Fig. 8.4A). We consider that at a minimum standard, a scope should be able to deliver an SNR of no less than 10:1 for the fluorescent beads in dispersive media at superficial depths or in a nondispersive sample. That is, with a target signal intensity of 240, standard deviation of noise should be no greater than 24. In subsequent deeper planes, the SNR can be expected to decrease with scattering and absorption until beads can no longer be detected. SNR as a metric of maximal depth penetration will be discussed below.

While the performance of the PMTs will affect the total intensity and maximal depth of acquisition, any change in these metrics that can be traced back to other elements such as alignment should be corrected prior to evaluating the response of the detectors themselves.

The data set collected in method 2 can be used to generate the response profile of a detector under standardized conditions. We select a representative plane from the acquired Z-stack and repeat the same spot detection and noise calculation steps as described for method 1 for images acquired using each PMT voltage. Both signal (Fig. 8.4B) and noise (Fig. 8.4C and D) intensities increase as more voltage is applied.

As evident visually in Fig. 8.4E, as higher voltage is applied, PMTs generate high intensity noise at higher frequency. Because these high intensity pixels occur stochastically, they are not biased toward particular pixels across frames and their effect on noise standard deviation

and maximum can be mitigated by frame averaging (Fig. 8.4C and D). The cost of this noise reduction is a reduction in frame rate, for example, in Fig. 8.4 speed is reduced from 30 frames per second (video rate) to 10 frames per second. When acquisition speed and photodamage are not primary concerns, averaging of an extremely noisy detector can generate images with acceptable SNRs.

Ideally, as PMT voltage is increased, signal intensity will increase exponentially while noise increases only linearly, improving the SNR. Typically, each detector has a sample-dependent optimal setting for the signal gains beyond which increased voltage produces only marginal returns while continuing to produce noise. Figure 8.4F shows the SNR curve for beads detected at varying voltages. Although the highest SNR is achieved at 0.7 V, by visually inspecting the image, it is clear that several beads that are present are not detected at this setting. Although a higher voltage results in a decrease in SNR (to a level still well above our minimum detection threshold, $\text{SNR} > 4$), the gains in ability to identify objects of interest are more valuable. This trade-off is made in biological samples in an effort to image dimmer, smaller features of interest. We routinely perform the same benchmarking procedure using excised lymph nodes containing actin-CFP as well as CD11c-YFP cells, in addition to bead samples, as we have found the shot noise and signal gains generated are highly context dependent (Fig. 8.4G and I). That is, the beads alone do not give a full picture of how well the detector is truly doing when generating useful experimental images.

Ultimately, the goal of this exercise is to create a historical standard for the signal and noise generated by a PMT at a given voltage in response to a known sample, against which the microscope can be compared on a regular basis. We have found that even on the same instrument, the performance of PMTs of the same make and model varies. Figure 8.5A shows the difference between the SNR of the PMT used as our green channel, which is approximately 3-years old, and a newly replaced PMT used as our red channel. The change in a single PMT's response between years can also be characterized in this way using beads or a standard biological sample (Fig. 8.5B), which we have found more informative about the instrument's practical performance.

8.2.3 MAXIMAL DEPTH OF ACQUISITION

The choice to use two-photon imaging over other optical sectioning techniques is most often due to its ability to penetrate further into tissue, allowing the visualization of biological processes in context without physical disruption. Although the depth at which features will be able to be resolved ($\text{SNR} > 4$) will vary depending on the scattering properties of the tissue and brightness of the features themselves, the microscope should reliably be able to image to a consistent depth in the standard bead sample made of dispersive media. Figure 8.6A (right) shows the falloff of sample intensity with depth. To quantify this parameter, beads were identified in the data set described in PMT Settings, method 1 (500 μm Z-stack, 0.4 μm steps, dispersive sample) using Imaris (Bitplane AG) spot detection. SNR was calculated by dividing mean bead intensity at the intensity center by the standard deviation of the noise. Beads could not reliably be identified at depths beyond 350 μm with $\text{SNR} > 4$ at the chosen laser power and PMT voltages (Fig. 8.6B).

Changes in depth of penetration between benchmarking experiments may be indicative of several independent factors. As PMTs age, they may see increased noise and decreased gains in signal, reducing overall sensitivity and making it more difficult to resolve features deep in the sample. Quantifying and diagnosing PMT changes are discussed in the following section.

Temporal dispersion of the femtosecond laser pulses used in multiphoton systems can decrease the effective power delivered to the focal plane, an affect that is worsened with scattering deep in tissue. Such dispersion results from differential wavelength-dependent refraction of excitation light caused by routing optics or the optical properties of tissues. Some multiphoton lasers include hardware and software packages that allow users to compensate for this dispersion by “chirping” pulses (predispersing laser pulses in the opposite direction to compensate for dispersion introduced by the system).

8.3 TROUBLESHOOTING/OPTIMIZING

Throughout this discussion, there is the presumption that certain parameters can be better in one scope as compared to another or can decay (or improve) in a given scope over time or with change in alignment, etc. A few key parameters to note are the alignment of the system (PSF and intensity), the quality and age of PMTs (intensity and maximal depth of acquisition) and the pulse width of the laser (maximal depth of acquisition) (Zucker & Price, 2001). Of course, many other factors play into the quality achieved by microscope and can be as mundane as the bandpass of detection filters or dust on various optical elements (Chapter 4). The benchmark we describe, however, can be performed and vigorously promoted with a service engineer (in the case of a commercially produced microscope) or with a local technician (in the course of a microscope built or serviced in-house). It is quite important and very helpful that “poor performance” be unambiguously tracked to the microscope itself and not simply a spate of bad samples.

8.4 A RECIPE FOR PURCHASING DECISIONS

Multiphoton microscopes require significant alignment to function optimally and when considering commercially produced microscopes, this is an important issue to understand. The proper perspective on this issue is that you want to have the very best system for your needs, regardless of whether it can be set up in your lab for a 1-week trial or if, conversely, a technician might hand-carry the samples to a distant site. We suggest the latter will be the best way to see the best instrument that a vendor can offer. Based on our experiences, we recommend against having a vendor “demo” microscope on-site since invariably the scope will not function optimally in a short demonstration period, and even if it is functioning to its design optimum, you may wonder if “it might have done better” given stable and optimal conditions. Such is not a rational selection mechanism. A benchmark as used here can be best used during instrument selection to choose ‘the best’ for your uses and then a second time when a purchased instrument is installed to confirm that the instrument that is delivered operates to your required specification.

CONCLUSION

The process of benchmarking is important in the testing and acquisition of a system as well as throughout its life span. It is important to define key collection parameters and samples to use for benchmarking. A standard dispersive bead sample and a standard biological sample represent key steps in establishing a protocol for obtaining and maintaining a microscope with optimal function for biological applications. Ultimately, it helps to determine whether a biological experiment can or cannot be done, and whether the biology or the microscope might be improved to facilitate an experiment.

Acknowledgments

The authors thank the Krummel lab for helpful discussions.

References

- Churchill SW, Clark GC, Sliepcevich CM. Light-scattering by very dense monodispersions of latex particles. *Discussions of the Faraday Society*. 1960; 30:192–199.
- Hadjantonakis AK, Macmaster S, Nagy A. Embryonic stem cells and mice expressing different GFP variants for multiple non-invasive reporter usage within a single animal. *BMC Biotechnology*. 2002; 2:11. [PubMed: 12079497]
- Paul P, Kalamatianos D, Duessman H, Huber H. Automatic quality assessment for fluorescence microscopy images. 8th IEEE international conference on BioInformatics and BioEngineering, BIBE 2008, 8–10, 2008. 2008:1–6. 1–6.
- Schaefer BC, Schaefer ML, Kappler JW, Marrack P, Kedl RM. Observation of antigen-dependent CD8+ T-cell/dendritic cell interactions in vivo. *Cellular Immunology*. 2001; 214:110–122. [PubMed: 12088410]
- Theer P, Hasan MT, Denk W. Two-photon imaging to a depth of 1000 micron in living brains by use of a Ti:Al₂O₃ regenerative amplifier. *Optics Letters*. 2003; 28:1022–1024. [PubMed: 12836766]
- Yotter RA, Wilson DM. A review of photodetectors for sensing light-emitting reporters in biological systems. *IEEE Sensors Journal*. 2003; 3:288–303.
- Zucker RM, Price O. Evaluation of confocal microscopy system performance. *Cytometry*. 2001; 44:273–294. [PubMed: 11500845]

FIGURE 8.1.

Change in laser power measured with corresponding EOM voltage change at the sample with varied wavelength. Measured from Spectra Physics MaiTai XF-1 (710–920 nm).

FIGURE 8.2.
Nondispersive (left) and dispersive (right) bead gels in 35-mm dishes ready for use.

FIGURE 8.3. Acceptable and poor PSFs obtained during benchmarking

(A) Poor PSF suffering from astigmatism. Central XY plane, XZ and YZ views. (B) Acceptable PSF central XY plane, XZ and YZ views. (C) Maximum intensity Z-projections of a single bead (top A, bottom B). (D) XZ view of PSF when the laser beam enters the objective at an angle. (E) FWHM resolution of A. (F) FWHM resolution of B.

FIGURE 8.4.

(A) Spots detected used to calculate signal average intensity and two 140×140 pixel square ROIs used to calculate noise standard deviation. (B) Average bead intensity signal with increased PMT voltage at a single Z-plane ($15 \mu\text{m}$ depth). (C) Standard deviation of noise calculated from two 140×140 pixel ROIs with increased PMT voltage in a single image and from an average of three images at a single Z-plane ($15 \mu\text{m}$ depth). (D) Maximum intensity of noise in two 140×140 pixel ROIs with increased PMT voltage from single image and from an average of three images at a single Z-plane ($15 \mu\text{m}$ depth). (E) Bead images at the same Z-plane at 0.6 V (left), 0.8 V (center), 1.0 V (right) shown at $\sim 2.5 \times$ for clarity. Arrows indicate position of fluorescent beads; intensity values at other pixels are the result of noise. Noise frequency and intensity increases with bias voltage. (F) Signal-to-noise ratio of beads with varied PMT voltage. (G) Average yellow channel signal intensity from a single Z-plane ($20 \mu\text{m}$ depth) of actin-CFP CD11-c-YFP lymph node. (H) Signal-to-noise ratio of yellow channel from single plane of actin-CFP CD11-c-YFP lymph node image. (I) Image of single plane of actin-CFP, CD11-c-YFP lymph node.

FIGURE 8.5.

Comparison of SNR in beads and lymph nodes across years. (A) SNR from red and green (solid and dashed in print version) beads generated under identical conditions between two detectors on the same microscope. (B) SNR of images from the same detector of a CD11c-YFP lymph node under identical conditions taken 1 year apart. Gains in SNR with increased voltage are reduced between benchmarking experiments indicating decreased sensitivity.

FIGURE 8.6.

Z-projections of nondispersive bead, dispersive bead, and a standard biological sample. (A) YZ oriented 3D view of 500 μm thick Z-stack taken with 0.4 μm Z steps in nondispersive gel (left) and dispersive gel (right). The working distance of the objective typically limits Z collection in nondispersive samples. Dispersive media limits the depth at which images can be collected. (B) SNR of beads detected through spot detection algorithm versus depth in dispersive gel media. SNR decreases with imaging depth, and we see a corresponding drop in the number of detected spots. Box indicates beads with $\text{SNR} < 4$, and vertical line indicates “maximal depth” achieved by imaging system.