Published in final edited form as:

Artif Organs. 2017 July ; 41(7): E52–E65. doi:10.1111/aor.12808.

A Physical Heart Failure Simulation System Utilizing the Total Artificial Heart and Modified Donovan Mock Circulation

Jessica R. Crosby¹, Katrina J. DeCook¹, Phat L. Tran^{1,3}, Edward Betterton⁴, Richard G. Smith^{1,3,4}, Douglas F. Larson⁴, Zain I. Khalpey⁴, Daniel Burkho⁵, and Marvin J. Slepian^{1,2,3,*}

¹Biomedical Engineering GIDP, University of Arizona, Tucson, Arizona 85724

²Department of Biomedical Engineering, University of Arizona, Tucson, Arizona 85724

³Department of Medicine, Sarver Heart Center, University of Arizona, Tucson, Arizona 85724

⁴Artificial Heart Department, Banner University Medical Center, University of Arizona, Tucson, Arizona 85724

⁴Department of Surgery, University of Arizona, Tucson, AZ 85724

⁶Columbia University, New York, NY

Abstract

With the growth and diversity of mechanical circulatory support (MCS) systems entering clinical use, a need exists for a robust mock circulation system capable of reliably emulating and reproducing physiologic as well as pathophysiologic states for use in MCS training and inter-device comparison. We report on the development of such a platform utilizing the SynCardia Total Artificial Heart (TAH) and a modified Donovan Mock Circulation System (DMCS), capable of being driven at normal and reduced output. With this platform, clinically relevant heart failure hemodynamics could be reliably reproduced as evidenced by elevated left atrial pressure (+112%), reduced aortic flow (–12.6%), blunted Starling-like behavior and increased afterload sensitivity when compared to normal function. Similarly, pressure-volume relationships demonstrated enhanced sensitivity to afterload and decreased Starling-like behavior in the heart failure model. Lastly, the platform was configured to allow the easy addition of a left ventricular assist device (HeartMate II® at 9600 RPM), which upon insertion resulted in improvement of hemodynamics. The present configuration has the potential to serve as a viable system for training and research, aimed at fostering safe and effective MCS device use.

*Corresponding author: Marvin J. Slepian MD, Sarver Heart Center, 1501 N. Campbell Avenue, Rm 5146, P.O. Box 24-5037, The University of Arizona, Tucson, AZ 85724 USA, chairman.syns@gmail.com.

Summary of Author Contribution:

Jessica R. Crosby – Data Collection, Data Analysis, Draft Manuscript

Katrina DeCook – Data Collection, Data Review, Manuscript Edits

Phat Tran – Experimental Review

Edward Betterton – Manuscript Review, Clinical Input

Richard G. Smith – Experimental and Data Review, Clinical Input

Douglas Larson – Manuscript Review, Experimental Advice

Zain Khalpey – Manuscript Review, Clinical Input

Daniel Burkho⁵ – Experimental Setup Oversight

Marvin Slepian - Experimental and Data Review, Clinical Input, Critical Manuscript Revision

Keywords

total artificial heart; Donovan mock circulation; mechanical circulatory support; heart failure; physiologic simulation; ventricular assist device

Introduction

Mechanical circulatory support (MCS) has emerged as the standard-of-care for advanced heart failure (AHA Stage D, NYHA Class IV) [1–4]. A range of MCS devices are in clinical use in the United States including implantable ventricular assist device (VAD) systems (Thoratec HeartMate II® and HeartWare HVAD®), the total artificial heart (SynCardia TAH-t), and a number of percutaneous/extracorporeal short-term devices, with several other systems in development and testing around the world [5]. With the growth in MCS use and its increasing technical user base, i.e. clinicians, engineers, MCS support staff, nurses, patients, and caregivers, a need exists for an effective, hands-on system for training as to the physiology and hemodynamics associated with use of specific MCS devices under simulated physiological real world conditions [5,6]. In addition, creating a single physical system which could be readily utilized to simulate and assess the performance of a given MCS device under a defined set of clinical conditions prior to *in vivo* use could ultimately lead to a more personalized, scientifically-guided refinement of device parameter settings (i.e., pump speed). Further, such a system would allow for inter-device comparison and provide a means to objectively compare devices under defined hemodynamic conditions.

To date a variety of mock circulation systems have been configured to mimic human circulatory physiology and several of these systems have been utilized for *in vitro* hemodynamic and hemolysis testing of MCS devices [7–19]. However, these mock circulation systems are largely *ad hoc* setups that are assembled from multiple components for lab use, are not readily available to MCS implanting centers, and are typically not utilized by clinicians or MCS support staff to train, personalize or directly compare MCS devices.

Recently we have defined and characterized a full circulatory mock circulation system consisting of the SynCardia Total Artificial Heart (TAH) in combination with a modified Donovan Mock Circulation System (DMCS) [20]. This system is robust, capable of physically reproducing a wide range of hemodynamic parameters – e.g. pressure and load extremes, is relatively transportable, and is presently installed in many clinical TAH implanting sites as part of the regulatory-defined training program [21]. In a previous report from our group, the physiology of the TAH system was evaluated under normal drive conditions over a wide range of loading parameters. In the present study our group builds upon this work by modifying this system to operate under reduced pumping (systolic and diastolic) conditions, thereby emulating left ventricular failure. Here, we hypothesize that the TAH operating with reduced left ventricular output, in conjunction with the DMCS, will provide a reproducible, robust, physical heart failure model, demonstrating hemodynamics consistent with left ventricular failure. Further, we propose that this configuration could be

utilized as a system to evaluate the functional performance of ventricular assist devices under defined clinical hemodynamic scenarios, for training and research purposes.

In this study, we first define and examine the reproducibility and function of normal and heart failure states with the mock system. Second, we characterize the heart failure model over a range of pre- and afterloads, examining hemodynamics as well as pressure-volume (PV) relationships. Finally, we examine the effect of adding a ventricular assist device to the system as to its ability to modulate hemodynamics

Methods

The Total Artificial Heart and The Hydraulic Analog of the Circulatory System

A modified Donovan Mock Circulatory System (DMCS) with an affixed 70cc SynCardia Total Artificial Heart (SynCardia Systems, Inc., Tucson, AZ) was utilized for all studies, as fully detailed in our previous report [20]. Briefly, each of the TAH ventricles supports an inner flexible diaphragm assembly separating each ventricle into two compartments, an air compartment and a blood compartment. The diaphragm assembly is mobilized via pulses of air supplied by an external pneumatic driver. Air in from the driver causes an upward mobilization of the diaphragm assembly which in turn forces blood out of the ventricle during systole, resulting in forward, pulsatile flow [22]. Diastole occurs with air evacuation from the ventricle, allowing the diaphragm assembly to move outward from the blood chamber, creating space in the blood side of the ventricle thus allowing it to fill with blood. The external pneumatic driver allows for the control of drive pressure and vacuum for each ventricle, percent systole, and beat rate [22]. The SynCardia total artificial heart (TAH) as used in normal, clinical use is driven using the partial fill, full eject principle (i.e., the ventricles are partially filled during diastole and fully ejected during systole – all fluid filled during diastole is ejected each cardiac cycle). Operating the TAH in partial fill conditions ensures that there is always reserve volume in the ventricle to accommodate for additional venous return from the body (creating a Starling-like effect). In clinical use, the TAH driver is always set to provide enough driving pressure on the diaphragm to eject all fluid in the ventricle (i.e., ensuring afterload can be overcome) [22].

A DMCS (SynCardia Systems, Inc., Tucson, AZ) was utilized to simulate the systemic and pulmonary portions of the human vasculature for all experiments [21]. The DMCS contains four chambers representative of the following domains: right atrium, pulmonary artery, left atrium and aorta, with the ability to control the flow resistance between aorta and right atrium via a bellows operated valve (simulating systemic vascular resistance) and pulmonary artery and left atrium via a bellows operated valve (simulating pulmonary vascular resistance). Vascular compliance is manipulated via control of the volume of air within each of the chambers representing the four vascular domains [21]. The system was filled with a 35% (v/v) glycerin/deionized water solution, with a viscosity of 3.5 cps at 24°C, as a functional blood equivalent.

Experimental Configuration and Baseline Normal Operating Conditions

All chambers of the DMCS were fitted with pressure transducers (Abbott, Abbott Park, Illinois) allowing pressure determination for each chamber: aortic pressure (AoP), left atrial pressure (LAP), right atrial pressure (RAP), and pulmonary artery pressure (PAP). Four high-fidelity strain gage pressure catheters (SPR-524, Millar Instruments, Inc., Houston TX) connected to a pressure control unit (PCU-2000, Millar Instruments, Inc.) were utilized to monitor continuous aortic pressure, left ventricular pressure, right ventricular pressure, and left-atrial pressure. Three flow meters (ME 25 PXN, Transonic Systems, Inc., Ithaca, NY) were placed throughout the system to measure fluid flow rates of left ventricular inflow, left ventricular outflow, and VAD outflow (Figure 8). Data from all sensors were acquired at 200 Hz with a compact data acquisition board (NI-9219, NI-9211, and two NI-9205, National Instruments, Austin, TX) interfaced with a custom LabVIEW executable.

The TAH was driven by the Companion 2 pneumatic driver (SynCardia Systems, Inc., Tucson, AZ). Baseline normal driver conditions were: left drive pressure 180 mmHg, left vacuum -10 mmHg, right drive pressure 60 mmHg, right vacuum -10 mmHg, heart rate 100 beats per minute, and 50% systole. Once baseline parameters were set on the driver, the DMCS was set to “normotensive” patient conditions: right atrial mean pressure 6 ± 5 mmHg, pulmonary arterial mean pressure 20 ± 5 mmHg, left atrial mean pressure 10 ± 5 mmHg, and aortic mean pressure 95 ± 5 mmHg.

Emulation of Left Ventricular Failure with the DMCS and TAH

To mimic the hemodynamics of left ventricular failure, two parameters were varied on the Companion 2 TAH driver system: left ventricular driving pressure (affecting ventricular systole) and left ventricular vacuum (affecting ventricular diastole). During characterization of the system in reduced output “heart failure” conditions, the DMCS was set to normal operating conditions (settings provided above) as a baseline. The TAH pneumatic driver (Companion 2) left vacuum was then set to 0 mmHg and the left driving pressure was varied between 200 mmHg and 120 mmHg in increments of 10 mmHg; 120 mmHg is the minimum achievable left driving pressure on the Companion 2 Driver (180–200 mmHg is the typical left drive pressure used clinically). Following each incremental adjustment, the DMCS was allowed to equilibrate and stabilize for two minutes, following which data from all channels were recorded at 200 Hz for ten seconds and completed three times. Pressure, flow and percent ventricular emptying were measured as endpoint variables for all protocols completed. (The term “ejection fraction” was avoided in this report – reserving this for a tissue ventricle, as the TAH with a plastic ventricle does not dilate, lacks time varying elastance and is purposely under-filled during clinical use to allow for a Starling-like effect of partial fill full eject, increasing output based on increasing venous return.)

Effect of Varying Afterload and Preload on Heart Failure Model Hemodynamics

The system’s response to an increase in afterload was assessed with the TAH operating under reduced drive conditions, left ventricular drive pressure of 120 mmHg and left ventricular vacuum of 0 mmHg. These reduced output settings were used for heart failure conditions throughout the study. The mean aortic pressure (afterload) was raised in 5 mmHg increments from 95 to 110 mmHg. Similarly, the system’s response to a change in preload

was assessed during reduced driver conditions through variations in the *right* ventricular vacuum, which in turn varied left ventricular end-diastolic volume (EDV). The right vacuum was varied from 0 to 20 mmHg (10 mmHg being the normal vacuum value used clinically) so that the EDV of the left ventricle ranged from 48 to 73 ml. Following each incremental adjustment, the tank was allowed a two-minute interval to reach steady state, following which data from all channels was recorded at 200 Hz for ten seconds. Each test condition was repeated three times.

Addition of a Ventricular Assist Device to the Mock System

The DMCS with attached TAH was altered to allow for incorporation of various MCS devices via modification of the left ventricular outflow tubing. Two T-junctions, separated by a one-way bi-leaflet Open Pivot artificial heart valve (Medtronic, Minneapolis, MN), were placed between the outflow of the left ventricle and the inflow to the aortic (AoP) chamber of the DMCS. This system alteration (i.e., T-junctions with additional valve) essentially “extends” the left ventricle such that the inflow of any LVAD could be incorporated within the mock vasculature loop without damaging the ventricle of the TAH. This additional valve acts as the primary aortic valve when the LVAD is in use (the TAH’s native aortic valve will remain open when the LVAD is in place). Note that the limitation of incorporation of the MCS device in this manner limits the ability of the user to simulate severe suck-down events that could occur in a native ventricle. A flow sensor (ME 25 PXN, Transonic Systems, Inc., Ithaca, NY) was placed immediately distal to the MCS insertion site to measure device outflow.

Effect of VAD Pump Speed on Heart Failure Model Physiology

With the system set to heart failure mode (66% of baseline left drive pressure) a HeartMate II® LVAD was added to the circuit and set to operate over a range of speeds (7000 – 11000 RPM, in increments of 200 RPM). Following each incremental adjustment, the system was allowed a two-minute interval to reach steady state, following which data from all channels was recorded at 200 Hz for ten seconds. Each test condition was repeated three times.

Results

Defining the Heart Failure Model

Incremental decreases in the left pneumatic drive pressure, with the vacuum at 0 mmHg, led to a reduction of TAH filling and ejection. As left drive pressure was decreased with vacuum removed (Figure 2), an effective decline in pumping action, or “model inotropy”, was observed. Reduction of left drive pressures from 200 mmHg to 120 mmHg resulted in a progressive reduction in mean aortic pressure from 89.7 ± 1.8 mmHg to 84.8 ± 0.1 mmHg (–5.4%), and left ventricular output from 4.15 ± 0.2 L/min to 3.6 ± 0.1 L/min (–12.6%), accompanied by an increase in left atrial pressure from 16.2 ± 1.3 mmHg to 34.4 ± 0.1 mmHg (+112%).

When reducing drive pressure, once full eject conditions are not met, stroke volume will decrease with each decrease of drive pressure. However, note that when full-eject conditions are met, changes in drive pressure will not vary pumping characteristics to a large degree as

long as full ejection is occurring. As shown in the presented data, Figure 2, full eject was lost once the drive pressure was reduced to 130 mmHg – this can be observed by a slight increase in left atrial pressure (A), increase in mean left ventricular pressure (B) increased fluid remaining in the ventricle causing increase in the mean pressure, reduction in aortic pressure (C) and reduction of mean left ventricular outflow (D) at 130 mmHg when comparing to the higher drive pressures. In addition, all these effects are exaggerated when the drive pressure is reduced to 120 mmHg.

Emulation of the most severe, i.e. worst case, heart failure hemodynamics obtainable with the TAH + DMCS configuration, as tested, was governed by the maximum degree of drive pressure reduction achievable with the affixed driver. In the present case, utilizing the Companion 2 driver, which has an internally designed reduction limit based on clinical use, this was 120 mmHg. Under these conditions, the DMCS had a mean right atrial pressure of 7.1 ± 0.1 mmHg, mean left ventricular pressure of 69.2 ± 0.7 mmHg, mean aortic pressure of 84.8 ± 0.1 mmHg, mean left atrial pressure of 34.4 ± 0.1 mmHg, and left ventricular outflow of 3.6 ± 0.1 L/min. It was these parameters that were utilized as a baseline “heart failure” condition for the remainder of studies outlined. Of note, standard deviations above and throughout this report represent standard deviation between mean values of each collected data set. The small standard deviation values highlight the reproducibility of the model.

Effect of Afterload Variation on Heart Failure Model Hemodynamics

Afterload was increased in the DMCS by manipulating the bellows-operated valve to restrict flow between the AoP chamber and RAP chamber. Figure 3 (a–e) displays mean right atrial pressure (a), mean right atrial pressure (b), mean left atrial pressure (c) mean left ventricular end diastolic pressure (d) mean left ventricular systolic pressure, and (e) mean left ventricular outflow over a ten second acquisition period with varying mean aortic pressures. As mean aortic pressure was increased from 95 to 110 mmHg, mean right atrial pressure exhibited a 5.2% reduction, from 7.2 to 6.9 mmHg (Figure 3a), average left atrial pressure increased by 30.7%, from 35.4 mmHg to 46.3 mmHg (Figure 3b), and mean left ventricular end diastolic pressure increased from 29.7 to 54.4 mmHg (Figure 3c). *At 110 mmHg, there is a reduction in end diastolic pressure because the ventricle is not filling as quickly once afterload is increased to 110 mmHg, notice that mean cardiac output is less than 1 L/min at this level of afterload. This decreased filling will attribute to the lower LVP at 110 mmHg in comparison to 105 mmHg.* Mean left ventricular systolic pressure remained fairly consistent, between 110.3 and 112.6 mmHg over the range of afterloads tested (Figure 3d). Left ventricular outflow decreased significantly, from an average of 3.59 to 0.66 L/min, an 81% reduction (Figure 3e).

Effect of Afterload Variation on Heart Failure Model Pressure-Volume Relationships

Pressure-volume loops obtained under heart failure conditions at varying afterloads are shown in Figure 4. For ease of interpretation, pressure spikes (artifact) caused by closure of the mitral valve were removed from the pressure tracing in the heart failure model (Figure 4a). Afterload variation under normal operating conditions is shown in Figure 4b. Under heart failure conditions, as afterload was increased, a prominent reduction in stroke volume

(width of the PV loop) was noted, with the smallest stroke volume detected at an aortic pressure of 110 mmHg. This trend correlated with results shown in Figure 3; as afterload increased, output decreased. Additionally, ventricular end systolic volume was between 40 and 50 ml. The ventricle inherently has a residual volume of near 30 ml during full ejection, i.e. when the diaphragm assembly has reached its maximum extended position. In reduced output mode, the end systolic volume was consistently higher than 30 ml, indicating that the ventricle was not fully ejecting the end diastolic volume. Compared to the afterload variations observed under normal operating conditions, Figure 4b, which reveal a consistent stroke volume over varying afterloads, the pressure-volume loops in heart failure conditions (4a) show that the TAH operating under reduced output conditions is sensitive to afterload. The increase in concavity of the loops during systole (top of the loop) in Figure 4a compared to the loops in Figure 4b is likely due to a “lagging” of diaphragm movement under heart failure conditions, as drive pressure is significantly lower than normal operating conditions.

Effect of Preload Variation on the Heart Failure Model

Venous return and preload were altered through adjustment of the right ventricle vacuum pressure. Vacuum was varied between 0 and -20 mmHg, (normal operation being -10 mmHg), resulting in changes in achieved left ventricular end diastolic volumes (EDV). However, no change was seen in left ventricular end diastolic volume with right vacuum parameters greater than -12 mmHg, thus, only data from 0 to -12 mmHg vacuum are presented. The effects of preload variation on mean left atrial pressure (a), left ventricular pressure (b), left ventricular outflow (c), and aortic pressure (d) are shown in Figures 5a-d. As end diastolic volume was increased from 48 to 73 ml, mean right atrial pressure remained relatively constant (Figure 5a), left atrial pressure increased as end diastolic pressure increased, with a 106% increase in LAP, rising from a mean of 8.8 mmHg at an EDV of 48 ml to a mean of 18.5 at an EDV of 73 ml Figure 5b. Mean left ventricular pressure similarly increased, +46%, from 45 to 66 mmHg (Figure 5b) as preload was increased. Accordingly left ventricular end diastolic pressure increased, +41%, from 10.3 mmHg to 24.86 mmHg with augmented preload (Figure 5c) as well as left ventricular systolic pressure, +15.5% from 96.4 mmHg to 111.5 mmHg, (Figure 5d) and left ventricular outflow as well, revealing a significant increase of 86%, from 1.92 L/min to 3.58 L/min, with increasing EDV (Figure 5e).

Effect of Preload Variation on Heart Failure Model Pressure-Volume Relationships

Pressure-volume (PV) loops under various preload conditions are shown in Figure 6. As end diastolic volume was increased in heart failure mode (6a), increases in stroke volume (loop width) between an EDV of 48 to 62 ml were detected. Stroke volume increases between an EDV of 62 and 73 ml were minimal and the end systolic pressure between these two preloads remained similar, though greater than the end systolic pressure generated with an EDV of 48 ml. Figure 6b shows PV loops generated under normal operating conditions, which demonstrate incremental increases in stroke volume with increasing EDV.

A Starling Curve for the variations in preload is presented in Figure 7, comparing normal left ventricular drive conditions to the reduced drive conditions [20]. Under normal drive conditions a higher EDV was achieved when the right ventricular vacuum was set at -20

mmHg, reaching an EDV of 84 ml for normal conditions compared to 73 mL at heart failure conditions. In addition, heart failure conditions produced little increase in left ventricular output once an EDV of 63 ml was reached. At an EDV of 63 ml, left ventricular output averaged 3.45 ± 0.1 L/min across all test runs, while at an EDV of 73 ml, left ventricular output was 3.5 ± 0.1 L/min, $p = 0.6$. As in the human heart, the TAH demonstrates increased stroke volume and increased (left) ventricular pressure during systole with increases in end-diastolic volume. However, in the heart failure model a threshold left ventricular EDV is reached, above which preload augmentation does not increase left ventricular output, as the pump has inadequate pumping force to eject fluid out of the ventricle; this action being similar to a failing heart. Further, the TAH, with a physically set, non-dilatable ventricle, cannot continue to benefit from dilatation and volume augmentation as in the human ventricle.

Addition of a VAD to the Model System: Effect of LVAD Pump Speed Alteration on Heart Failure Model Hemodynamics

A HeartMate II® LVAD was successfully added to the circulatory loop via the addition of two T-junctions and one heart valve; a flow diagram of the configuration is shown in Figure 8. Following establishment of baseline heart failure conditions as above, HeartMate II® speed was adjusted from 7,000 to 11,000 RPM, in 200 RPM increments Figure 9. As LVAD speed was increased the following were observed: an 8% increase in mean right atrial pressure from 6.9 mmHg to 7.5 mmHg, a 68% decrease in mean left ventricular pressure from 62 mmHg to 20 mmHg, limited variation in aortic pressure, a 56% decrease in left atrial pressure from 18 mmHg to 10 mmHg, a 108% decrease in left ventricular outflow 1.7 L/min to -0.15 L/min, yet a significant 204% increase in LVAD outflow from 2.16 to 4.4 L/min. The negative left ventricular outflow observed at higher pump speeds is attributable to LVAD pump suction, as fluid in the ventricle is being diverted from left ventricle ejection via pump suction at high pump speeds; small leakage occurs across the aortic valve, i.e., the additional valve added into our system (from the aorta to the left ventricle). Leakage of the aortic valve does happen with LVAD patients, overtime if the pump speed is too high, creating too low a pressure in the ventricle during diastole. Our mock system actually mimics this occurrence.

Discussion

In the present study, we build upon our prior work of characterizing the physiology of the TAH integrated with a modified Donovan Mock Circulatory System to further define and assess the performance of this configuration as a robust mock circulation system, capable of operating and reproducing both normal and heart failure hemodynamics. Herein, via modifying pump drive parameters, a reproducible left ventricular heart failure model has been demonstrated and physiologically characterized. The value of the platform lies in its biventricular pulsatile heart, hands-on robustness and its conduit designs, allowing for the physical addition and interchange of MCS devices. As such this offers an interactive training tool, affording the ability to examine hemodynamic effects and consequences of pressure and volume variations encountered in patients with heart failure and MCS devices. Further, this system may be utilized for inter-device comparative studies and research, which is

largely impossible in man due to the invasive nature of MCS device implantation, not to mention prohibitive cost and ethical issues. While herein we report only on the system as a model of left ventricular failure, from ongoing studies by our group it is clear that the system can physically emulate right ventricular failure as well as biventricular failure, allowing similar characterization to be performed, training to be accomplished as well as inter-device add-on comparative studies to be undertaken. Reports on these expanded uses will be forthcoming in the future.

A wide variety of physical circulation models and mimics of heart failure hemodynamics have been developed over the years, but most are limited due to either being overly simple – e.g. inability to accurately vary load, inotropy or pulsatility; or overly complex – i.e. with multiple physical elements patched together; pose challenges as to transportability, or ease of use [10,13,17,23]. The present system offers a distinct set of advantages when compared to these earlier systems. The present system affords reproducible vascular loading, resistances and hemodynamic variation of loading via simple mechanical means - without the need for computer or servo control. Use of the TAH allows for true pulsatility with defined impulse wave and dp/dt characteristics, as well as variable “model inotropy.” Being a physically contained unit constructed of polycarbonate, flexible yet robust tubing and the affixed TAH, the system is durable and transportable. As a set system that has been utilized extensively in the lab, by our group and others, with defined standard operating instructions, it allows for ease of use and reproducibility of hemodynamics. Further, as the system is an element of required training for TAH centers, significant availability already exists at numerous medical centers. Additionally, with defined, simple conduit modifications as in the present study, the system allows for addition and interchange of a wide range of augmentative or alternative MCS devices into the circuit, which provides a valuable model system for training and research.

While various physical models of heart failure have been developed [9,10,17,12,13,15], only one other group to date (Senage *et. al.*) has incorporated a pulsatile artificial heart within a mock system. Though these investigators utilized a TAH, they did not fully characterize the physiology of the heart failure model under normal and heart failure operating conditions nor use a clinically relevant pulsatile drive system to pump the paired TAH ventricles. Further, in their study LVAD inflow ports were placed upstream from the left ventricle thus limiting ventricular filling. Herein, that gap is closed through use of a defined, pulsatile, clinical TAH system; characterization of the heart failure model prior to device testing; incorporation of present day LVADs in a more clinically relevant circulation location followed by examination of this addition as to physiologic impact on hemodynamic performance [16]. As such, parameters defined and characterized in our study can be utilized to replicate and establish easy-to-use heart failure models widely, particularly in those centers already in possession of the TAH and DMCS.

As shown in the present study under heart failure conditions the TAH demonstrates Starling-like behavior, although to a lesser degree as compared to non-heart failure conditions, with a reduction seen in the relationship between output and preload, as exists in the failing human heart [24–26]. The Starling-like characteristics of the heart failure model are shown to respond to increases in preload, but only at lower end-diastolic volumes, with output

plateauing at higher end-diastolic volumes, as the ventricle cannot fully eject large volumes during heart failure operation and can not continue to dilate for further augmentation. The system also exhibits afterload sensitivity, with ventricular output and left ventricular ejection decreasing dramatically with increased afterload, thus leading to increased left atrial pressure. The afterload response of the heart failure model thus closely mimics the human heart in failure, which demonstrates a reduction in stroke volume when subjected to increased afterload.

There are clear variations in the shape of pressure-volume loops in the TAH compared to pressure-volume loops of the human heart [27,28]. Under normal operating conditions, the TAH has been shown to be afterload insensitive [20]. However, with the TAH operating in reduced drive heart failure mode, the left ventricle is not always capable of overcoming the afterload and fully ejecting all volume that enters the ventricle during diastole. Reduction in stroke volume, i.e. width of the pressure-volume loop, is seen even with a 5 mmHg increase from baseline mean aortic pressure, with flow decreasing to under 1 L/min at a mean aortic pressure of 110 mmHg.

During preload variation, a left ventricular end-diastolic volume greater than 74 ml was not achieved by right vacuum manipulations. With the left vacuum turned off in the described heart failure model, the diaphragm is incapable of rapidly mobilizing to its “down or fully withdrawn position,” to allow additional filling even with increased preload applied. In a sense this behavior exhibits an element of diastolic dysfunction, with reduced ability to fill. In addition, during preload variations, left ventricular output failed to rise with increasing preload at end-diastolic volumes greater than 63 mL, due to the reduction in pumping power of the left ventricle as occurs in the failing human heart. Despite these “boundary” limitations, the present system is able to mimic many, if not most of the important characteristics of the hemodynamics of the heart failure patient [27,29,30].

With an LVAD connected to the heart failure model system, a clear improvement in hemodynamics occurred. When LVAD speed was increased, elevated atrial pressure and mean left ventricular pressure were reduced and total cardiac output (LVAD + left ventricular outflow) increased. Additionally, back flow was observed with excessive LVAD speed. The determination of best operating LVAD pump speed, when the device is first implanted into a patient, is essential. As the TAH is a biventricular system and has pulsatile pumping action, LVAD displays will show waveforms and values that are typical of what is observed clinically [31–33]. Thus, this system may be utilized as a training tool to mimic various patient parameters and situations that are clinically relevant to the LVAD operating team, allowing training for optimized device performance as well as management of various suboptimal and failure clinical scenarios. For example, in this platform, aortic line variations can be visualized to determine aortic valve opening hemodynamics, flow to or from the device can be completely or partially occluded, device waveforms and pulsatility indices can be visualized, and vascular resistance can be easily be modified, all vital and relevant for real world use. Studies are presently underway examining the comparative effectiveness of differing VAD systems under defined, identical heart failure conditions utilizing this platform.

Clinical Impact

With the rise in use of a wide variety of MCS devices the clinical management personnel involved continue to become increasingly diverse. Presently in addition to physicians, VAD coordinators, nurses, nurse practitioners/physician assistants, engineers and emergency personnel participate in aspects of patient care. As such the present system is well adapted to train these groups. As with all technical systems it has been clearly shown that increased simulation-based training ultimately increases operative skill and management of failure scenarios [34]. To date the unmodified SynCardia Mock system, has been utilized to train more than 100 medical center groups in the US and around the world on proper driver operation of the SynCardia TAH-t. Multiple clinical training scenarios are completed during this training.

In addition to TAH training, the present system, with the modifications discussed in this paper, has been utilized as a model of heart failure alone as a general pathophysiology training tool and simulator for the impact of therapeutic (pharmacologic) manipulation [35].

Further, with the growing variety of commercially available MCS devices, in or newly entering the market, the present platform is well suited to allow training and research as to the relative performance of a given MCS device under defined hemodynamic conditions, which cannot be completed in patients and is difficult in animal models.

Reproducibility

Under all settings and conditions evaluated, normal and heart failure, the system provided reproducible results, as can be observed through the low reported standard deviations. Additionally, no significant drift was observed over time throughout testing. As such, the described configuration represents a robust mock circulation system, capable of operating and consistently reproducing both normal and heart failure hemodynamics.

Limitations

The lowest achievable driving pressure of the Companion II driver (120 mmHg) is also a limitation of the described model. Heart failure severity could be increased in this model if a lower drive pressure was utilized, as is achievable with the original SynCardia TAH driver, the CSS Console [36]. The CSS Console allows for left drive pressure adjustments between 0 and 300 mmHg, whereas the Companion II is limited between 120 and 280 mmHg. In prior and ongoing work by our group it is clear that the envelope of performance of the present described system can be expanded, if need be, through use of a pulsatile pneumatic driver with wider operating range.

Conclusion

A robust, physiologically defined, user-friendly physical left ventricular heart failure model has been developed and characterized utilizing the Donovan Mock Circulation System and the SynCardia Total Artificial Heart. This construct provides physiologically relevant, reproducible pressures and flows representative of clinical left heart failure hemodynamics. The ability to readily manipulate patient parameters on the DMCS and the

straightforwardness in adjusting pumping parameters on the TAH driver provides a versatile system capable of simulating and modeling a multitude of clinical scenarios. The ability to rapidly interpose a wide variety of MCS devices in the flow circuit makes this platform an effective tool for comparative assessment of MCS device performance. Widespread availability of the DMCS and TAH already within the TAH user community further enhances the value of the described system for use as both a hands-on training tool and a research platform, ultimately aimed at enhancing the care and outcomes of patients with MCS devices.

Acknowledgments

Disclosures: Partial support for this research was provided by the Cardiovascular Biomedical Engineering Training Grant, National Institutes of Health T32HL007955

References

1. Stevenson, LW., Pagani, FD., Young, JB., Jessup, M., Miller, L., Kormos, RL., et al. J. Heart Lung Transplant. [Internet]. Vol. 28. International Society for Heart and Lung Transplantation; 2009. INTERMACS Profiles of Advanced Heart Failure: The Current Picture; p. 535-541. Available from: <http://www.ncbi.nlm.nih.gov/pubmed/19481012> [cited 2014 Oct 27]
2. Rose E, Gelijns A. Long-Term Use of a Left-Ventricular Assist Device for End-Stage Heart Failure. N. Engl. J. Med. [Internet]. 2001; 345:1435–1443. [cited 2014 Oct 31] Available from: <http://www.nejm.org/doi/full/10.1056/NEJMoa012175>.
3. Hunt, Sa, Ea, R. The REMATCH trial: Long-term use of a left ventricular assist device for end-stage heart failure. J. Card. Fail. [Internet]. 2002; 8:59–60. [cited 2014 Nov 1] Available from: <http://linkinghub.elsevier.com/retrieve/pii/S107191640280598X>.
4. Hunt, Sa, Baker, DW., Chin, MH., Cinquegrani, MP., Feldmanmd, aM, Francis, GS., et al. ACC/AHA Guidelines for the Evaluation and Management of Chronic Heart Failure in the Adult: Executive Summary A Report of the American College of Cardiology/American Heart Association Task Force on Practice Guidelines (Committee to Revise the 1995 Guideli. Circulation [Internet]. 2001; 104:2996–3007. [cited 2014 Aug 12] Available from: <http://circ.ahajournals.org/cgi/doi/10.1161/hc4901.102568>.
5. Kirklin, JK., Naftel, DC., Pagani, FD., Kormos, RL., Stevenson, LW., Blume, ED., et al. J. Hear. Lung Transplant. [Internet]. Vol. 33. Elsevier; 2014. Sixth INTERMACS annual report: A 10,000-patient database; p. 555-564. Available from: <http://dx.doi.org/10.1016/j.healun.2014.04.010>
6. Van de Bussche TJ, Edwards LG, Elliott T, Harton S, Rivard D, Wolfe AC. Regionalized approach to emergency medical services training for the care of patients with mechanical assist devices. Prog. Transplant. [Internet]. 2010; 20:129–132. Available from: <http://www.ncbi.nlm.nih.gov/pubmed/20642170>.
7. Ferrari G, Lazzari C De, Kozarski M. A hybrid mock circulatory system: testing a prototype under physiologic and pathological conditions. ASAIO [Internet]. 2002 [cited 2014 Oct 31] Available from: http://journals.lww.com/asaiojournal/Abstract/2002/09000/A_Hybrid_Mock_Circulatory_System__Testing_a.9.aspx.
8. Rosenberg G, Phillips W, Landis D, Pierce W. Design and Evaluation of the Pennsylvania State University Mock Circulatory System. ASAIO J [Internet]. 1981 [cited 2014 Apr 2] Available from: <http://scholar.google.com/scholar?hl=en&btnG=Search&q=intitle:Design+and+Evaluation+of+the+Pennsylvania+State+University+Mock+Circulatory+System#0>.
9. Baloa, La, Boston, JR., Antaki, JF. Elastance-Based Control of a Mock Circulatory System. Ann. Biomed. Eng. [Internet]. 2001; 29:244–251. [cited 2014 Mar 27] Available from: <http://link.springer.com/10.1114/1.1355275>.
10. Timms D, Hayne M, McNeil K, Galbraith A. A complete mock circulation loop for the evaluation of left, right, and biventricular assist devices. Artif. Organs [Internet]. 2005; 29:564–572. Available from: <http://www.ncbi.nlm.nih.gov/pubmed/15982285>.

11. Liu Y, Allaire P, Wood H, Olsen D. Design and Initial Testing of a Mock Human Circulatory Loop for Left Ventricular Assist Device Performance Testing. *Artif. Organs* [Internet]. 2005; 29:338–341. Available from: <http://www.ncbi.nlm.nih.gov/pubmed/15787630>.
12. Pantalos GM, Koenig SC, Gillars KJ, Giridharan Ga, Ewert DL. Characterization of an Adult Mock Circulation for Testing Cardiac Support Devices. *ASAIO J.* [Internet]. 2004; 50:37–46. [cited 2013 Oct 8] Available from: <http://content.wkhealth.com/linkback/openurl?sid=WKPTLP:landingpage&an=00002480-200401000-00008>.
13. Timms DL, Gregory SD, Greatrex Na, Percy MJ, Fraser JF, Steinseifer U. A compact mock circulation loop for the in vitro testing of cardiovascular devices. *Artif. Organs* [Internet]. 2011; 35:384–391. [cited 2013 Oct 8] Available from: <http://www.ncbi.nlm.nih.gov/pubmed/20883450>.
14. Ferrari G, Lazzari C. Mock Circulatory System for in vitro reproduction of the left ventricle, the arterial tree and their interaction with a left ventricular assist device. *J. Med. Eng. Technol.* [Internet]. 1994 [cited 2014 Nov 6] Available from: <http://informahealthcare.com/doi/abs/10.3109/03091909409030237>.
15. Ferrari G, Lazzari C De, Mimmo R. A computer controlled mock circulatory system for mono- and biventricular assist device testing. *Int. J. Artif. Organs* [Internet]. 1998; 21:26–36. [cited 2014 May 28] Available from: <http://europepmc.org/abstract/MED/9554823>.
16. Sénage T, Février D, Michel M, Pichot E, Duveau D, Tsui S, et al. A Mock Circulatory System to Assess the Performance of Continuous-Flow Left Ventricular Assist Devices (LVADs): Does Axial Flow Unload Better Than Centrifugal LVAD? *ASAIO J.* [Internet]. 2014; 60:140–147. [cited 2014 Mar 3] Available from: <http://www.ncbi.nlm.nih.gov/pubmed/24577368>.
17. Koenig SC, Pantalos GM, Gillars KJ, Ewert DL, Litwak KN, Etoch SW. Hemodynamic and Pressure-Volume Responses to Continuous and Pulsatile Ventricular Assist in an Adult Mock Circulation. *ASAIO J.* [Internet]. 2004; 50:15–24. [cited 2014 Jul 20] Available from: <http://content.wkhealth.com/linkback/openurl?sid=WKPTLP:landingpage&an=00002480-200401000-00005>.
18. Giridharan, Ga, Sobieski, Ma, Ising, M., Slaughter, MS., Koenig, SC. Blood trauma testing for mechanical circulatory support devices. *Biomed. Instrum. Technol.* [Internet]. 2011; 45:334–339. Available from: <http://www.ncbi.nlm.nih.gov/pubmed/21812617>.
19. Giridharan, Ga, Koenig, SC., Slaughter, MS. Do axial-flow LVADs unload better than centrifugal-flow LVADs? *ASAIO J.* [Internet]. 2014; 60:137–139. [cited 2014 May 27] Available from: <http://www.ncbi.nlm.nih.gov/pubmed/24577367>.
20. Crosby JR, DeCook KJ, Tran PL, Smith RG, Larson DF, Khalpey ZI, et al. Physiological Characterization of the SynCardia Total Artificial Heart in a Mock Circulation System [Internet]. *ASAIO J.* 2015 Available from: <http://content.wkhealth.com/linkback/openurl?sid=WKPTLP:landingpage&an=00002480-900000000-99439>.
21. Donovan, FM. Design of a hydraulic analog of the circulatory system for evaluating artificial hearts; *Biomater. Med. Devices. Artif. Organs* [Internet]. 1975. p. 439-449. Available from: <http://www.ncbi.nlm.nih.gov/pubmed/1225373>
22. Slepian, MJ., Smith, RG., Copeland, JG. The SynCardia CardioWest Total Artificial Heart. In: Baughman, Baumgartner, editor. *Treat. Adv. Hear. Dis.* Taylor and Francis; 2006. p. 473-490.
23. Pantalos GM, Koenig SC, Gillars KJ, Giridharan Ga, Ewert DL. Characterization of an Adult Mock Circulation for Testing Cardiac Support Devices. *ASAIO J.* [Internet]. 2004; 50:37–46. [cited 2014 Mar 27] Available from: <http://content.wkhealth.com/linkback/openurl?sid=WKPTLP:landingpage&an=00002480-200401000-00008>.
24. Komamura K, Shannon RP, Ihara T, Shen YT, Mirsky I, Bishop SP, et al. Exhaustion of Frank-Starling mechanism in conscious dogs with heart failure. *Am. J. Physiol.* [Internet]. 1993; 265:H1119–H1131. Available from: <http://www.ncbi.nlm.nih.gov/pubmed/8238398>.
25. Schwinger RH, Böhm M, Koch a, Schmidt U, Morano I, Eissner HJ, et al. The failing human heart is unable to use the Frank-Starling mechanism. *Circ. Res.* [Internet]. 1994; 74:959–969. Available from: <http://www.ncbi.nlm.nih.gov/pubmed/8156643>.
26. Weil J, Eschenhagen T. Preserved Frank–Starling mechanism in human end stage heart failure. *Cardiovasc. Res.* [Internet]. 1998:541–548. [cited 2014 Oct 28] Available from: <http://cardiovascres.oxfordjournals.org/content/37/2/541.short>.

27. Burkhoff D. Pressure-volume loops in clinical research: a contemporary view. *J. Am. Coll. Cardiol.* [Internet]. 2013; 62:1173–1176. [cited 2014 Mar 26] Available from: <http://www.ncbi.nlm.nih.gov/pubmed/23770176>.
28. Mancini D, Burkhoff D. Mechanical device-based methods of managing and treating heart failure. *Circulation* [Internet]. 2005; 112:438–448. [cited 2014 Sep 18] Available from: <http://www.ncbi.nlm.nih.gov/pubmed/16027272>.
29. McMurray JJV. Systolic Heart Failure. 2013
30. Adams KF, Zannad F, Hill C, Clinical deAdams KF, Zannad F, Hill C. Clinical definition and epidemiology of advanced heart failure, 204–215. finition and epidemiology of advanced heart failure. 1998:204–215. (1998).
31. Slaughter, MS., Pagani, FD., Rogers, JG., Miller, LW., Sun, B., Russell, SD., et al. *J. Heart Lung Transplant.* [Internet]. Vol. 29. Elsevier Inc.; 2010. Clinical management of continuous-flow left ventricular assist devices in advanced heart failure; p. S1-S39. Available from: <http://dx.doi.org/10.1016/j.healun.2010.01.011>
32. Myers, TJ., Bolmers, M., Gregoric, ID., Kar, B., Frazier, OHH. *J. Heart Lung Transplant.* [Internet]. Vol. 28. International Society for Heart and Lung Transplantation; 2009. Assessment of arterial blood pressure during support with an axial flow left ventricular assist device; p. 423-427. Available from: <http://www.ncbi.nlm.nih.gov/pubmed/19416768> [cited 2011 Oct 18]
33. Stanfield JR. *In Vitro* Pulsatility Analysis of Axial-Flow and Centrifugal-Flow Left Ventricular Assist Devices. *J. Biomech. Eng.* [Internet]. 2013; 135:034505. Available from: <http://biomechanical.asmedigitalcollection.asme.org/article.aspx?doi=10.1115/1.4023525>.
34. Schroedl, CJ., Corbridge, TC., Cohen, ER., Fakhraan, SS., Schimmel, D., McGaghie, WC., et al. *J. Crit. Care* [Internet]. Vol. 27. Elsevier Inc.; 2012. Use of simulation-based education to improve resident learning and patient care in the medical intensive care unit: A randomized trial; p. 219.e7-219.e13. Available from: <http://dx.doi.org/10.1016/j.jcrc.2011.08.006>
35. Physical Simulation Session. American Heart Association Annual Scientific Session. 2015 Nov 7–11.
36. Copeland JG, Smith RG, Arabia FA, Nolan PE, Sethi GK, Tsau PH, et al. Cardiac Replacement with a Total Artificial Heart as a Bridge to Transplantation. *N. Engl. J. Med.* 2004; 351:859–867. [PubMed: 15329423]

Fig. 1. The Donovan Mock Circulation System (unmodified)

A) Fluid flow through the total artificial heart (right and left ventricle) and Donovan Mock Circulation System (four chambers). The triangles within the ventricles represent valves of the TAH. B) Close up of the left and right ventricles of the total artificial heart attached to the DMCS. C) The DMCS with attached total artificial heart.

Fig. 2. Effect of Progressive “Failure” (decreasing left drive pressure) on TAH and Systemic Pressures

A–D) Average left atrial pressure, left ventricular pressure, aortic pressure, and left ventricular outflow are shown over a range of left drive pressures. The arrow over 180 mmHg indicates the normal left drive pressure typically used clinically. No driveline vacuum was applied with the driver in order to further decrease function of the TAH in order to mimic reduced function.

Fig. 3. Effect of Afterload Alterations in Heart Failure Conditions

A–D) Plots of mean right atrial pressure, left atrial pressure, left ventricular end diastolic pressure, left ventricular systolic pressure, and left ventricular outflow over a range of mean aortic pressures (afterloads). The afterload was varied using the bellows operated valve in the DMCS.

Afterload Variation Under Reduced Output: Heart Failure Conditions

Afterload Variation Under Normal Operating Conditions

Fig. 4. Pressure-Volume Loops with Varying Afterloads in Heart Failure and Normal Conditions

A) Pressure-Volume loops with 95, 100, and 110 mmHg mean aortic pressures, corresponding to systolic pressures of 111, 112, and 114 mmHg respectively. B) Pressure-Volume loops with 85, 115, and 135 mmHg mean aortic pressures. Notice the difference in loop trends over the varying afterloads. Stroke volume (loop width) *decreases* with increased afterload whereas in normal operating conditions stroke volume remains the same over varying afterloads. Normal condition pressure volume loops (Crosby *et al* 2014). In heart failure conditions, pressure-volume loops with a mean AoP of up to 135 mmHg (as was

tested under normal conditions) was not achievable because flow through the system would cease at higher aortic pressures.

Author Manuscript

Author Manuscript

Author Manuscript

Author Manuscript

Fig. 5. Varying the Preload

Preloads were acquired through variations in the *right* ventricular vacuum pressure. A–D) Plots of mean right atrial pressure, left atrial pressure, left ventricular end diastolic pressure, left ventricular systolic pressure, and left ventricular outflow over a range of preloads.

Preload Variations in Reduced Output, Heart Failure Conditions

Preload Variation in Normal Operating Conditions

Fig. 6. Pressure-Volume Loops with Varying Preloads

Pressure-Volume loops with variations in left ventricular preload. A) With reduced output “heart failure” conditions. B) Under normal operating conditions. Notice the difference in stroke volume (loop width) between heart failure and normal operating conditions. In normal operating conditions, an increase in stroke volume is obvious with increased EDV whereas, in heart failure conditions, increased EDV does not elicit as high of a change in stroke volume.

Fig. 7. Starling Curve

Left Ventricular Output vs Left End Diastolic Volume (fill volume). Variations in preload were achieved by manipulating the right vacuum. Comparison between normal (solid) and reduced drive pressures (dotted).

Fig. 8. The Incorporation of a Ventricular Assist Device

A ventricular assist device was incorporated into the loop via the addition of two T-junctions and the addition of a valve that acts as the aortic valve when the VAD is powered on, essentially extending the volume of the left ventricle. This allows for the addition of a VAD into the system without damaging or altering the TAH. The triangles represent valves. All sensor and flow meter placements are depicted. When the VAD is on, the aortic valve remains open.

Fig. 9. HearMate II Operation at Varying Speeds

Figures A–F display right atrial pressure, left ventricular pressure, aortic pressure, left atrial pressure, left ventricular outflow and VAD outflow as the HearMate II ventricular assist device was placed in-line with our heart failure model. Note the reductions in HF conditions as RPM is increased. At speeds above 9800 RPM, suction is occurring within the left ventricle, which is why left ventricular flow becomes negative.