Published in final edited form as:

J Trauma Acute Care Surg. 2017 July ; 83(1): 139–143. doi:10.1097/TA.0000000000001502.

INCREMENTAL BALLOON DEFLATION FOLLOWING COMPLETE REBOA RESULTS IN STEEP INFLECTION OF FLOW AND RAPID REPERFUSION IN A LARGE ANIMAL MODEL OF HEMORRHAGIC SHOCK

Anders J. Davidson, MD^{1,2}, Rachel M. Russo, MD^{1,2}, Sarah-Ashley E. Ferencz, MD¹, Jeremy W. Cannon, MD^{3,4}, Todd E. Rasmussen, MD⁴, Lucas P. Neff, MD^{1,2,4}, M. Austin Johnson, MD⁵, and Timothy K. Williams, MD⁶

¹Department of Surgery, UC Davis Medical Center, Sacramento, California ²Department of General Surgery, David Grant Medical Center, Travis Air Force Base, California ³Division of Trauma, Surgical Critical Care & Emergency Surgery, Perelman School of Medicine at the University of Pennsylvania, Philadelphia, Pennsylvania ⁴The Norman M. Rich Department of Surgery, the Uniformed Services University of the Health Sciences, Bethesda, Maryland ⁵Department of Emergency Medicine, UC Davis Medical Center, Sacramento, California ⁶Heart, Lung and Vascular Center, David Grant Medical Center, Travis Air Force Base, California

Abstract

Introduction—To avoid potential cardiovascular collapse after Resuscitative Endovascular Balloon Occlusion of the Aorta (REBOA), current guidelines recommend methodically deflating the balloon over 5 minutes to gradually reperfuse distal tissue beds. However, anecdotal evidence suggests that this approach may still result in unpredictable aortic flow rates and hemodynamic instability. We sought to characterize aortic flow dynamics following REBOA as the balloon is deflated in accordance with current practice guidelines.

Methods—Eight Yorkshire-cross swine were splenectomized, instrumented, and subjected to rapid 25% total blood volume hemorrhage. After 30 minutes of shock, animals received 60 minutes of zone 1 REBOA with a low-profile REBOA catheter. During subsequent resuscitation

Corresponding Author: Timothy K. Williams, MD, Heart, Lung and Vascular Center, David Grant USAF Medical Center, Travis Air Force Base, California, 101 Bodin Circle, Travis Air Force Base, CA 94535, Phone: 707-423-2300 timothykeithwilliams@gmail.com.

Level of Evidence: Level V

Disclaimer

The animals involved in this study were procured, maintained, and used in accordance with the Laboratory Animal Welfare Act of 1966, as amended, and NIH 80-23, Guide for the Care and Use of Laboratory Animals, National Research Council. The views expressed in this material are those of the authors, and do not reflect the official policy or position of the U.S. Government, the Department of Defense, the Department of the Air Force, or the University of California Davis. The work reported herein was performed under United States Air Force Surgeon General approved Clinical Investigation No. FDG20160008A. There was no funding from the National Institutes of Health (NIH), Wellcome Trust, or the Howard Hughes Medical Institute (HHMI) for this work. Dr. Cannon received royalty payments from UpToDate for an article on Endovascular Methods for Aortic Control in Trauma. The Clinical Investigation Facility, David Grant Medical Center, Travis Air Force Base, California provided funding for this study.

Presented: A version of this manuscript was presented as a poster presentation at the Eastern Association for Surgery of Trauma Annual Scientific Assembly, January 10–14, 2017 at the Diplomat Resort in Orlando, FL. This work has not been published elsewhere.

with shed blood, the aortic occlusion balloon was gradually deflated in stepwise fashion at the rate of 0.5 mL every 30 seconds until completely deflated. Aortic flow rate and proximal mean arterial pressure (MAP) were measured continuously over the period of balloon deflation.

Results—Graded balloon deflation resulted in variable initial return of aortic flow (median 78 seconds, IQR 68–105sec). A rapid increase in aortic flow during a single balloon deflation step was observed in all animals (Median 819 mL/min IQR 664–1241mL/min) and corresponded with an immediate decrease in proximal MAP (Median 30 mmHg IQR 14.5–37mmHg). Total balloon volume and time to return of flow demonstrated no correlation ($r^2=0.016$).

Conclusion—This study is the first to characterize aortic flow during balloon deflation following REBOA. A steep inflection point occurs during balloon deflation that results in an abrupt increase in aortic flow and a concomitant decrease in MAP. Furthermore, the onset of distal aortic flow was inconsistent across study animals and did not correlate with initial balloon volume or relative deflation volume. Future studies to define the factors that impact aortic flow during balloon deflation are needed to facilitate controlled reperfusion following REBOA.

Introduction

Resuscitative Endovascular Balloon Occlusion of the Aorta (REBOA) is an emerging therapeutic adjunct for non-compressible torso hemorrhage.(1, 2) Despite increasing clinical use, REBOA remains limited by the profound distal ischemia created by complete aortic occlusion.(3) The degree of ischemia corresponds with the duration of balloon occlusion and has led to the recommendation that the REBOA balloon catheter be deflated as soon as possible.(4) Despite the need for balloon deflation as soon as possible, little has been described about the physiologic consequences of balloon deflation after REBOA. As in open aortic cross-clamping, REBOA balloon deflation results in a dramatic decrease in aortic afterload, a rapid washout of ischemic metabolites, and the rapid influx of arterial blood into maximally vasodilated distal vasculature.(5) The aggregate effect of these physiologic derangements has the potential to cause hemodynamic collapse upon balloon deflation.

Initial clinical experiences describe deflating the REBOA balloon in a simple, yet controlled manner.(6) However, this simple method was subsequently questioned after it reportedly led to several instances of hemodynamic collapse that were refractory to balloon reinflation.(7) In an attempt to mitigate the physiologic impact of balloon deflation following REBOA, current practice recommendations advocate slow and methodical balloon deflation to gradually reintroduce distal aortic flow.(8, 9) In theory, graded balloon deflation reduces physiologic derangement by creating a protracted washout of distal ischemic metabolites, while simultaneously providing partial proximal hemodynamic support. Yet, anecdotal evidence suggests even graded balloon deflation may result in hemodynamic instability.(10) Given the lack of clarity from both a translational and clinical perspective, we sought to characterize the hemodynamic impact of incremental balloon deflation following sustained REBOA, particularly with respect to mean arterial pressure and aortic flow.

Materials and Methods

Animal Preparation

The Institutional Animal Care and Use Committee at David Grant Medical Center, Travis Air Force Base, California approved this study. All animal care and use was in strict compliance with the Guide for the Care and Use of Laboratory Animals in a facility accredited by the Association for the Assessment and Accreditation of Laboratory Animal Care International. Healthy adult, castrate male and nonpregnant female Yorkshire-cross swine (*Sus scrofa*), obtained from the University of California, Davis, were acclimated for a minimum of 7 days. At the time of experimentation, animals weighed between 57 and 75 kg. Animals were premedicated with 6.6-mg/kg tiletamine/zolazepam (TELAZOL, Fort Dodge Animal Health, Fort Dodge, IA) intramuscularly. Following isoflurane induction and endotracheal intubation, maintenance anesthesia consisted of 2% isoflurane in 100% oxygen. To offset the vasodilatory effects of general anesthesia, an intravenous infusion of norepinephrine (0.01 µg/kg/h) was instituted upon venous access and titrated to achieve a target mean arterial pressure between 65 and 75 mmHg, which was kept constant for the remainder of the experiment. Animals were mechanically ventilated with tidal volumes of 7–10 mL/kg and a respiratory rate of 10–15 breaths per minute sufficient to maintain end-tidal CO₂ at 40 ± 5 mm Hg. Normal saline was administered at a rate of 5 mL/kg/hour to overcome insensible losses. An underbody warmer was utilized to maintain core body temperature between 35–37°C.

Venous access was obtained through a central line in the right internal jugular vein. Proximal and distal arterial access was obtained through a 7Fr 13 cm introducer sheath (Boston Scientific, Marlborough, MA) placed in the right femoral artery and right and left brachial arteries via surgical cutdowns. The animals then underwent midline laparotomy and a splenectomy was performed to minimize hemodynamic variation from autotransfusion.(11) The supraceliac aorta was then exposed by longitudinally dividing the diaphragm to facilitate placement of a periaortic flow probe (Transonic Systems Inc., Ithaca, NY). A 32mm aortic occlusion balloon (ER-REBOA, Prytime Medical, Boerne, TX) was positioned in the descending thoracic aorta and confirmed to be in Zone 1 by manual palpation. The abdomen was closed with cable ties.

Intervention

Hemorrhagic shock was simulated by withdrawing 25% of estimated blood volume (Body Weight (kg) × 6ml/kg) through an arterial sheath into citrated blood collection bags. The animals underwent a 30-minute equilibration period to simulate the pre-hospital environment. Complete aortic occlusion was then initiated by inflating the ER-REBOA™ balloon catheter until there was loss of a distal arterial waveform. Complete aortic occlusion was sustained for 60 minutes. After 55 minutes of occlusion, shed blood volume was returned through the central venous line via rapid infusion (Belmont Instruments, Billerica, MD) at the rate of 150cc/min. This was then followed by incremental balloon deflation at a rate of 0.5cc every 30 seconds until the balloon was completely deflated. Physiologic parameters and aortic flow measurements were collected in real time using a Biopac MP150 multichannel data acquisition system (Biopac Systems Incorporated, Goleta, California).

Parameters measured included heart rate, blood pressure proximal to the ER-REBOA™ catheter, and aortic flow. Onset of aortic flow during balloon deflation was defined as the time point when flow reached 50mL/min to account for the fidelity of the aortic flow probe at very low flow rates. At the end of the experiment the animals were humanely euthanized.

Conduct of the protocol is depicted in Figure 1. Statistical analysis was conducted with Stata 14 (StataCorp, College Station, TX)

Results

Hemodynamic parameters were similar between animals at baseline, at the conclusion of hemorrhage (T30), and just prior to balloon deflation (T90) (Table 1). The total balloon volume required to eliminate aortic flow varied (Median 6.5mL; range 3–9mL). After 60 minutes of occlusion, graded balloon deflation resulted in variable onset of initial aortic flow (Median 78 seconds; IQR 70–100sec), corresponding to a range of 2–7 deflation steps. There was no correlation between balloon volume and onset of initial aortic flow ($r^2=0.016$ Figure 2).

A rapid increase in aortic flow during a single balloon deflation step was observed in all animals (Median 820mL/min; IQR 683–1242mL/min), corresponding to approximately one-third of native aortic flow (Median 37%; IQR 25%–54%). Furthermore, this rapid increase in flow corresponded with a decrease in proximal MAP in all animals (Median 28.7mmHg; IQR 15.1–37.3mmHg). See Figure 3 for a representative image of the hemodynamics in a study animal.

Discussion

This goal of this study was to characterize the hemodynamic consequences of incremental balloon deflation following sustained REBOA. The intent of gradual balloon deflation is to provide predictable and graded restoration of aortic flow. Despite replicating the current best practices of REBOA balloon deflation consistently, there was wide variation in the hemodynamic response to deflation during this critical period. There were several key observations in this brief study: 1) the timing of initial return of aortic flow was unpredictable, 2) a large proportion of aortic flow returned at a discrete moment, 3) this discrete period of rapid reintroduction of aortic flow corresponded with a significant drop in MAP despite ongoing aggressive blood resuscitation.

Inherently, there are numerous factors influencing the point at which downstream flow is restored during active deflation of a balloon catheter, including the initial balloon filling volume, proximal aortic pressure, cardiac output, ongoing resuscitation, vascular tone, and aortic diameter. Most of these factors, with the exception of balloon volume, remain unmeasurable, unpredictable, and dynamic in nature. Balloon filling volume is certainly a variable within the control of the provider. However, overinflation likely does occur, further complicating the later process of balloon deflation. To minimize the impact of balloon volume on our results, inflation of the balloon in this study was ceased once the distal arterial waveform was lost. Theoretically, this method should mitigate the influence of balloon volume on our results, as well as to minimize the potential for aortic injury or

balloon rupture.(12, 13) However, despite this technique, the onset of flow past the balloon proved to be highly variable with respect to time, absolute volume remaining and percentage of volume remaining (Figure 2). Ultimately, balloon volume and time to return of flow were not correlated, indicating that the onset of aortic flow is likely controlled by factors not captured in the current study design and not routinely monitored in the clinical setting (Figure 4). It worth noting that all animals in this study were actively undergoing resuscitation and had MAPs similar to baseline at the time of balloon deflation (Table 1); indicating that the cause of unpredictable return of aortic flow was not fully explained by differences in proximal MAP or due to differences in resuscitation. Similar hemodynamic instability has been reported in studies of intermittent REBOA (iREBOA). This technique is intended to reduce ischemia from prolonged aortic occlusion by intermittently deflating the balloon by a set volume (usually 1cc) for a short period of time. Hemodynamic instability from unpredictable return of aortic blood flow has been implicated in the death of at least one patient and is a reason cited for unfavorable outcomes associated with iREBOA in animal studies.(14, 15) Our results suggest that relying on balloon volume alone is an inaccurate surrogate marker for return of initial aortic blood flow. Providers should rely on other metrics such as return of a distal arterial waveform measured from the flush port of the arterial introducer sheath to indicate return of aortic flow rather than a pre-specified number of balloon deflation steps.(16)

The primary intent of graded balloon deflation following REBOA is to produce gradual reintroduction of distal flow, thereby minimizing the potential for hemodynamic collapse or clot destabilization. However, we observed a large volume increase in aortic flow during a single balloon deflation step (Table 2, Figure 3). This dramatic increase was observed in all study animals equating to a median of approximately one-third of baseline flow (Table 2). This massive influx of pressurized arterial blood into maximally vasodilated ischemic distal tissue beds results in a rapid redistribution of circulating blood volume and a brisk washout of ischemic metabolites, which may contribute to the hemodynamic collapse sometimes observed following REBOA.(7, 8) Providers should be aware of this phenomenon and prepared to treat the effects of this rapid ischemia-reperfusion injury.

Lastly, gradual balloon deflation after REBOA is thought to reduce rebound hypotension by augmenting proximal MAP during this period. Nevertheless, all animals experienced significant drops in MAP at the time aortic flow rapidly increased, despite receiving aggressive resuscitation with shed blood (Table 2). This relative hypotension and abrupt loss of afterload may worsen perfusion to vital organs at a time when they are concurrently exposed to the systemic effects of ischemia-reperfusion injury. Clinical reports of volatile proximal hemodynamics during step-wise or partial balloon deflation support our finding that graded balloon deflation alone may be insufficient to support proximal MAP.(10, 17) Thus, partial reinflation of the balloon may be required to maintain or restore adequate proximal perfusion pressure to the heart and brain during this process.

This study was focused on a very specific moment during REBOA. As such, it lacks more extensive physiologic outcomes after intervention. Additionally, the absence of comparison group(s) limits any evaluation of alternative balloon deflation techniques. Instead, this small study was designed to provide a descriptive analysis of proximal hemodynamics during

balloon deflation and their relationship to aortic blood flow. Moreover, the present study is the first of its kind to take an in-depth look at the critical period of balloon deflation utilizing a technique currently taught as best practice in REBOA training courses.(9, 18) Our findings highlight that the period of balloon deflation after REBOA is dynamic and difficult to predict. Weaning from REBOA may be an overlooked, yet critical area for future study given the risk of cardiovascular collapse and hypotension reported in the literature. Many complex factors appear to influence the variability in hemodynamics during this period (Figure 4). This study would suggest that weaning from REBOA should be guided by multimodal physiologic and hemodynamic inputs and not based simply on balloon volume manipulation. Providers should be aware of potential hemodynamic instability during balloon deflation and should remain vigilant with adjunctive resuscitative measures at the ready.

Further studies are needed to better define the hemodynamic intricacies of blood flow around aortic occlusion balloons and optimize future transformations of the technique of REBOA. Future studies should promote perfusion optimization by facilitating controlled restoration of distal flow after REBOA.

Acknowledgments

We thank SSG Kelly Caneen, Mr. Carl Gibbins, Ms. Sally Knode, SSgt Elaine Spotts, and SrA Geoffrey O'Hair for their outstanding technical assistance, and the other staff of the Clinical Investigations Facility, David Grant Medical Center for their support.

References

1. Biffl WL, Fox CJ, Moore EE. The role of REBOA in the control of exsanguinating torso hemorrhage. *J Trauma Acute Care Surg.* 2015; 78(5):1054–8. [PubMed: 25909430]
2. Holcomb JB, Fox EE, Scalea TM, Napolitano LM, Albarado R, Gill B, et al. Current opinion on catheter-based hemorrhage control in trauma patients. *J Trauma Acute Care Surg.* 2014; 76(3):888–93. [PubMed: 24553566]
3. Irahara T, Sato N, Moroe Y, Fukuda R, Iwai Y, Unemoto K. Retrospective study of the effectiveness of Intra-Aortic Balloon Occlusion (IABO) for traumatic haemorrhagic shock. *World J Emerg Surg.* 2015; 10(1):1. [PubMed: 25598838]
4. Avaro J-P, Mardelle V, Roch A, Gil C, de Biasi C, Oliver M, et al. Forty-minute endovascular aortic occlusion increases survival in an experimental model of uncontrolled hemorrhagic shock caused by abdominal trauma. *Journal of Trauma and Acute Care Surgery.* 2011; 71(3):720–6.
5. White JM, Cannon JW, Stannard A, Markov NP, Spencer JR, Rasmussen TE. Endovascular balloon occlusion of the aorta is superior to resuscitative thoracotomy with aortic clamping in a porcine model of hemorrhagic shock. *Surgery.* 2011; 150(3):400–9. [PubMed: 21878225]
6. Gupta BK, KHANEJA SC, FLORES L, EASTLICK L, LONGMORE W, SHAFTAN GW. The role of intra-aortic balloon occlusion in penetrating abdominal trauma. *Journal of Trauma and Acute Care Surgery.* 1989; 29(6):861–5.
7. Martinelli T, Thony F, Declety P, Sengel C, Broux C, Tonetti J, et al. Intra-aortic balloon occlusion to salvage patients with life-threatening hemorrhagic shocks from pelvic fractures. *J Trauma.* 2010; 68(4):942–8. [PubMed: 20173661]
8. Stannard A, Eliason JL, Rasmussen TE. Resuscitative endovascular balloon occlusion of the aorta (REBOA) as an adjunct for hemorrhagic shock. *Journal of Trauma and Acute Care Surgery.* 2011; 71(6):1869–72.

9. Brenner M, Hoehn M, Pasley J, Dubose J, Stein D, Scalea T. Basic endovascular skills for trauma course: bridging the gap between endovascular techniques and the acute care surgeon. *Journal of Trauma and Acute Care Surgery*. 2014; 77(2):286–91. [PubMed: 25058255]
10. Davidson AJ, Russo RM, DuBose JJ, Roberts J, Jurkovich GJ, Galante JM. Potential benefit of early operative utilization of low profile, partial resuscitative endovascular balloon occlusion of the aorta (P-REBOA) in major traumatic hemorrhage. *Trauma Surgery & Acute Care Open*. 2016; 1(1):e000028.
11. Ross JD, Burns CJ, Sagini EM, Zarzabal L-A, Morrison JJ. A laparoscopic swine model of noncompressible torso hemorrhage. *Journal of Trauma and Acute Care Surgery*. 2014; 77(3):S77–S82. [PubMed: 25159366]
12. Søvik E, Stokkeland P, Storm B, Åsheim P, Bolås O. The use of aortic occlusion balloon catheter without fluoroscopy for life-threatening post-partum haemorrhage. *Acta Anaesthesiologica Scandinavica*. 2012; 56(3):388–93. [PubMed: 22260088]
13. Matsuda H, Tanaka Y, Hino Y, Matsukawa R, Ozaki N, Okada K, et al. Transbrachial arterial insertion of aortic occlusion balloon catheter in patients with shock from ruptured abdominal aortic aneurysm. *Journal of vascular surgery*. 2003; 38(6):1293–6. [PubMed: 14681630]
14. Ogura T, Lefor AT, Nakano M, Izawa Y, Morita H. Nonoperative management of hemodynamically unstable abdominal trauma patients with angioembolization and resuscitative endovascular balloon occlusion of the aorta. *J Trauma Acute Care Surg*. 2015; 78(1):132–5. [PubMed: 25539214]
15. Morrison JJ, Ross JD, Houston Rt, Watson JD, Sokol KK, Rasmussen TE. Use of resuscitative endovascular balloon occlusion of the aorta in a highly lethal model of noncompressible torso hemorrhage. *Shock*. 2014; 41(2):130–7. [PubMed: 24430492]
16. Johnson MA, Neff LP, Williams TK, DuBose JJ, Group ES. Partial Resuscitative Balloon Occlusion of the AORTA (P-REBOA): Clinical Technique and Rationale. *Journal of Trauma and Acute Care Surgery*. 2016
17. Hörer TM, Cajander P, Jans A, Nilsson KF. A case of partial aortic balloon occlusion in an unstable multi-trauma patient. *Trauma*. 2016 1460408615624727.
18. Villamaria CY, Eliason JL, Napolitano LM, Stansfield RB, Spencer JR, Rasmussen TE. Endovascular Skills for Trauma and Resuscitative Surgery (ESTARS) course: curriculum development, content validation, and program assessment. *Journal of Trauma and Acute Care Surgery*. 2014; 76(4):929–36. [PubMed: 24662854]

Figure 1.
Conduct of the protocol.

Figure 2. Scatter plot of balloon volume vs time to return of flow. Note poor correlation between balloon volume and time to return of flow.

Figure 3.
Hemodynamics of a Representative Study Animal. Note the steep inflection point with dramatic increase in flow with concurrent decrease in proximal MAP.

Figure 4. Determinants of Aortic Flow Conceptual Diagram: Conceptual diagram outlining many of the hypothesized factors that may influence aortic flow during balloon deflation after REBOA.

Table 1

Hemodynamics of Study Animals Prior to Balloon Deflation

	Proximal MAP (mmHg)	Aortic Flow (mL/min)
Baseline	71.1 ± 1.0	2350 ± 119
End of Hemorrhage (T30)	35.8 ± 5.2	1192 ± 117
Prior to Balloon Wean (T90)	76.3 ± 6.1	0

Data presented as mean ± standard error of the mean

MAP=mean arterial pressure

Table 2

Data from Study Animals

Animal	Time to Flow Restoration (Seconds)	# of Steps until Return of Flow	# of Steps until Max Increase in Flow	Max Increase in Flow in 1 Step (mL/min)	Max Increase in Flow as % of Baseline Aortic Flow	Max Decrease in MAP during 1 Deflation Step (mmHg)*	Total Balloon Volume (mL)
1	70	3	4	683	35%	15.1	5
2	62	3	4	1212	67%	6.5	8
3	76	3	3	448	19%	27.8	3
4	185	7	9	607	25%	36.3	6.5
5	100	4	6	720	25%	12.7	7
6	78	3	7	1627	65%	40.2	5.5
7	39	2	6	919	39%	32.3	9
8	119	5	8	1330	51%	57.5	6.5
Median	77.5	3	6	819	37%	30	6.5
IQR	68–105			664–1241	25%–54%	14.5–37	5.25–7.25

* MAP decrease corresponds to shown increase in aortic flow.

IQR = inter-quartile range