Factors Determining *In Vitro* Lung Deposition of Albuterol Aerosol Delivered by Ventolin Metered-Dose Inhaler

Rajoshi Biswas, MS,¹ Nicola A. Hanania, MD,² and Ashutosh Sabharwal, PhD¹

Abstract

Background: The effectiveness of metered-dose inhalers (MDIs) in delivering medication to the lungs highly depends on its correct usage technique. Current guidelines state optimal technique for high lung deposition should include a slow inhalation (>5 seconds) at an inspiratory flow rate of 30 L/min and inhaler actuation at the start of inhalation. However, these recommendations were based on clinical studies using CFC (chlorofluorocarbon)-MDIs and *in vitro* studies of HFA (hydrofluoroalkane)-MDIs using idealized MDI techniques of uniform inhalation and actuation, disregarding the nonuniform techniques of actual patients.

Methods: To better understand the effects of time-varying MDI usage parameters on lung deposition of aerosol delivered by an HFA-MDI, we conducted an *in vitro* study modeled on real-life variable inspiratory flow and actuation techniques recorded from 15 subjects with asthma/chronic obstructive pulmonary disease (COPD). We developed a model representing the time-varying inspiratory flow waveforms and actuation timings based on 43 MDI techniques recorded from patients. Furthermore, we constructed an *in vitro* experimental setup using a mouth–throat cast, programmable MDI actuator, and breath simulator to evaluate lung deposition for the MDI techniques derived from our model.

Results: High inspiratory flow rates, 60–90 L/min, consistently resulted in high *in vitro* lung deposition (>40%) of aerosol (albuterol delivered from Ventolin HFA-MDI) compared to 30 L/min when MDI actuation occurred in the first half of inhalation. Also, positive coordination resulted in higher *in vitro* lung deposition compared with negative or zero coordination (actuating before or at the start of inspiration). Furthermore, variation in coordination affected lung deposition more significantly (23%) than flow rate or duration of inspiration (≤5%).

Conclusions: In an *in vitro* experimental model based on inhalation data from patients with asthma and COPD, we demonstrated that aerosol lung deposition emitted from Ventolin HFA-MDI is most optimal for MDI actuation in the first half of inspiration at high flow rates (60–90 L/min).

Keywords: aerosol, asthma, COPD, inhaler technique, lung deposition, metered-dose inhaler

Introduction

OPTIMAL DELIVERY OF INHALED MEDICATIONS to the lung is highly dependent on the accuracy of the technique using the inhaler device.^(1–5) The metered-dose inhaler (MDI) is the most commonly used device that delivers a variety of rescue and maintenance medications for asthma and chronic obstructive pulmonary disease (COPD).^(1,6) However, several studies have demonstrated that patients often commit various errors using such devices leading to wasted medication and thus, suboptimal control of their disease.^(2–4)

Lung deposition of an aerosol delivered by an MDI may be dependent on all or a combination of the following five factors: (i) inspiratory flow rate, (ii) duration of inhalation, (iii) time of MDI actuation (time of actuating the MDI relative to the start of inspiration), (iv) duration of breath-hold after inhalation, and (v) fine fraction (the fraction of aerosol particles smaller than 5 μm in diameter).^(5,7) The first four factors depend solely on the patient's technique of using MDIs, while the fifth factor, amount of fine fraction released from the MDI, depends on the design of the formulation and construction of the components of the MDI delivery device involved with the aerosol formation process, and not on the patient's MDI

¹Department of Electrical and Computer Engineering, Rice University, Houston, Texas.

²Section of Pulmonary, Critical Care, and Sleep Medicine, Baylor College of Medicine, Houston, Texas.

technique.⁽⁸⁾ A high fine fraction results in higher deposition of medication in the lungs.⁽⁵⁾ Thus, optimum lung deposition is achieved when the patient uses an MDI that releases a large fine fraction with the correct technique.⁽⁵⁾

The proper technique of using an MDI is based on the first four factors, inspiratory flow rate, duration of inhalation, time of MDI actuation and duration of breath-hold after inhalation, and is outlined by published GINA guidelines.⁽⁷⁾ These guidelines are based on the results of numerous *in vivo* and *in vitro* studies previously conducted to quantify lung deposition of aerosol delivered by an MDI.^(9–14)

For instance, *in vivo* studies by Newman et al. and Lawford and McKenzie during the 1980s using CFC (chlorofluorocarbon)-MDIs demonstrated that slow inspiration of duration greater than 4 seconds, and MDI actuation at the start of inhalation resulted in the most optimal lung deposition.^(11–14) Furthermore, the effect of inspiratory flow rate was studied extensively for these CFC inhalers by Newman et al. in the 1990s, who demonstrated that higher lung deposition was achieved at low flow rates (~ 28 L/min).^(9,10) Based on those results, published guidelines for correct use of MDIs⁽⁷⁾ currently recommend a low inspiratory flow rate of 30 L/min, long duration of inhalation of at least 5 seconds and MDI actuation at the start of inhalation, for optimal lung deposition.

However, since the early 2000s, the CFC MDIs have been discontinued due to their potential environmental hazards and were replaced with HFA (hydrofluoroalkane)-MDIs. The aerosol deposition characteristics for medication particles emitted from CFC MDIs are very different from that of HFA-MDIs due to the difference in fine fraction content of the two inhalers.^(15–20) HFA-MDIs have a much higher fine fraction content than the CFC counterparts.^(15–19)

The disparity in lung deposition is demonstrated in more recent *in vivo* studies by Usmani et al., Leach et al. and *in vitro* studies by Cheng et al. and Rahmatalla et al. using HFA-MDIs, which indicate that lung deposition increased with higher inspiratory flow rates, contrary to the lung deposition characteristics of CFC-MDIs.^(15,17,18,20–22) In particular, Usmani et al. showed that for particles of diameter $1.5\ \mu\text{m}$, the lung deposition for fast inhalation was higher than that for slow inhalation.⁽¹⁸⁾ However, as the particle size was increased from 1.5 to $6\ \mu\text{m}$, the lung deposition gradually decreased with increase in inspiratory flow rate.

Recent literature investigating the effects of duration of inhalation and MDI actuation timing for HFA-MDIs has been limited to *in vitro* studies using idealized techniques, having constant inspiratory flow rates, fixed duration of inhalation, and fixed actuation time^(15,20–24); the idealized techniques do not represent actual patient's time-varying inhalation airflow profile and variable MDI actuation timings observed in a real-world setting. The correct HFA-MDI use technique in a real-world setting for both maximizing lung deposition and minimizing oropharyngeal deposition of medication is, therefore, unknown. The effect of the fourth factor, breath-hold after inhalation, has been well-documented through the above-mentioned *in vivo* studies.^(9–14) Holding the breath for up to 10 seconds after inhalation minimizes any portion of the medication to be exhaled out of the lungs and is the recommended current guideline for MDI use.^(5,7,25–27)

In this study, we developed a novel *in vitro* experimental setup to evaluate the lung deposition for time-varying

realistic HFA-MDI techniques derived from techniques recorded from subjects with airway disease (asthma or COPD) using a placebo HFA-MDI in a clinic setting. We conducted the *in vitro* experiments to assess the effects of variation in the steps using the MDI (i–iii) listed in the beginning of this section: inspiratory flow rate, duration of inhalation, and time of MDI actuation on lung deposition of aerosol emitted from a albuterol HFA-MDI, Ventolin (GlaxoSmithKline, Research Triangle Park, NC).

Materials and Methods

This was a two-phase study. We first recorded the MDI technique of 15 subjects, 18–80 years old, with physician-diagnosed asthma or COPD using placebo HFA-MDI. The patients were seen in the Pulmonary Function Testing Laboratory at Ben Taub Hospital, Houston, Texas. We then conducted an *in vitro* lung deposition study, modeled on the recorded real-world HFA-MDI techniques. Both Baylor College of Medicine and Rice University institutional review boards approved the study and all subjects signed an informed consent form. The clinical study was registered at clinicaltrials.gov, with the registration number NCT02447575.

We used a calibrated turbine flow meter attached to a plastic cap with a force sensor, fitted tightly to a placebo HFA-MDI to measure the three factors, inspiratory flow rate as a function of time, duration of inhalation, and time of MDI actuation (Fig. 1). Out of the 15 subjects, 7 took more than one MDI puff with the attached cap, bringing the total number of techniques recorded to 43. The system added an average resistance of $30\ \text{Pa}/(\text{L}\cdot\text{s})$, which did not interfere with inspiration through the MDI. However, the addition of the flow meter to the MDI may have added unfamiliarity to the subjects. Thus, before recording data, the subjects were instructed to actuate the MDI attached with the flow meter setup into the air (away from their mouth) to familiarize them with the system.

FIG. 1. Schematic illustrates the MDI technique recording setup utilized in recording MDI techniques. The MDI is fitted with a plastic cap that connects to a calibrated flow meter. During inhalation through the MDI, the air flows in from the vent at the top of the flow meter and passes through the MDI as shown. The flow meter is also attached to a force sensor that synchronously detects and records the time of MDI actuation. The inhalation and actuation data are stored on the flow meter and are retrieved later on a PC. MDI, metered-dose inhaler; PC, personal computer.

FIG. 2. Histogram of 43 MDI techniques collected from 15 subjects with asthma and COPD. Data collected consist of the inspiratory flow waveform and the time of MDI actuation. The distribution of three MDI technique steps of interest, (a) peak inspiratory flow rate, (b) duration of inhalation, and (c) time of actuation, is shown in (a–c), respectively. The black bars correspond to the correct technique as recommended by GINA guidelines, (a) peak flow rate between 30 and 60 L/min, (b) minimum inhalation of at least 5 seconds, and (c) time of actuation immediately after the start of inhalation (a positive value, 0–0.5 seconds). The gray bars correspond to the errors in MDI use, (a) peak inspiratory flow rate less than 30 L/min or greater than 60 L/min, (b) inhalation of less than 5 seconds, and (c) MDI actuation before the start of inhalation (a negative value) or late in the inhalation (a large positive value, greater than 0.5 seconds). COPD, chronic obstructive pulmonary disease.

Figure 2 illustrates histograms of peak inspiratory flow rates (PIFRs), duration of inspiration, and time of actuation with respect to the start of inhalation, and Table 1 contains the summary statistics for recorded data. Data recorded illustrate significant variability in MDI use techniques among the test subjects and deviations from the current recommendations.⁽⁷⁾ Figure 3a represents a subset of the recorded inhalation waveforms demonstrating variable PIFRs and durations of inhalation, thereby, confirming that actual MDI use is time-varying and not adequately approximated using idealized MDI techniques of constant inspiratory flow rates, duration, and MDI actuation. Figure 3b–d show the MDI technique model derived from the subject data.

Data-driven MDI use technique model

We developed a data-driven model for MDI technique based on the usage data recorded from subjects with asthma and COPD. The model assumes a fixed inspiratory capacity (IC), that is, is the maximum volume of air that can be inspired from a resting respiratory level, of 3 L [IC ranges 2.5–3.5 L in adults⁽²⁸⁾]. The main variables of MDI use examined in our model were (1) PIFR, (2) duration of inhalation, and (3) time of actuation, corresponding to the three out of five factors (i–iii), governing lung deposition. We defined a trapezoidal approximation to the patient inspiratory flow waveforms (Fig. 3a), accounting for the range

TABLE 1. METERED-DOSE INHALER TECHNIQUE DATA RECORDED FROM 15 SUBJECTS WITH ASTHMA AND CHRONIC OBSTRUCTIVE PULMONARY DISEASE

	Peak inspiratory flow (L/min), f_p	Duration of inhalation (seconds), t_d	Time of actuation (seconds), t_a
Mean (SD)	57 (32)	1.85 (0.93)	−0.05 (0.48)
Recommended guidelines	30–60	>5	0–0.5
Percentage with Incorrect MDI uses	60	100	79

MDI, metered-dose inhaler; SD, standard deviation.

FIG. 3. The inspiratory flow waveforms recorded during MDI use by subjects with asthma and COPD, and the trapezoidal approximation to the waveforms are presented. **(a)** Shows a subset of the 43 waveforms collected from 15 subjects, representing the different inspiratory flow rates, duration of inhalation, and waveform shapes. **(b)** Shows the trapezoidal model for a particular 3 L flow waveform. Here, $f_p = 60$ L/min (or 1 L/s), $t_d = 4.5$ seconds and $t_h = 1.5$ seconds. Thus the total inhalation volume is area of trapezoid $= (1/2) f_p (t_d + t_h) = 1/2 \times 1 \times (4.5 + 1.5) = 3$ L. **(c)** Shows the different combinations of 3 L waveforms generated for $f_p = 30, 60,$ and 90 L/min and the ratio $\tau = 1.33, 3,$ and 8 , corresponding to long, , and short duration of inhalation, respectively. As an example, a 90 L/min medium duration ($\tau = 3$) and a 60 L/min short duration ($\tau = 1.33$) are highlighted in bold. **(d)** Shows the trapezoidal approximation (dashed lines) for two of the waveforms recorded.

of inhalation duration and PIFRs observed from our recorded waveforms shown in Figure 3b. The area under the flow-time curve represents the fixed volume (IC of 3 L); we derived the following model for inhalation during MDI use:

$$\text{Area of trapezoid} = \text{IC} = \frac{1}{2} f_p (t_d + t_h) = 3\text{L} \quad (1)$$

where,

f_p = PIFR

t_d = duration of inhalation = base 1 of trapezoid

t_h = hold time = duration that flow rate is held at f_p = base 2 of trapezoid.

We also defined a third parameter, t_r = rise-time, used in the later sections to define the time of MDI actuation. t_r refers to the duration of time it takes for the flow rate to reach f_p from 0 and vice versa (symmetrical waveform). Thus,

$$t_d = 2 \times t_r + t_h \quad (2)$$

The volume constraint Equation (1) is predicated on the assumption that the relationships between f_p , t_d , and t_h are

linearly dependent. We can further rewrite the equation as follows:

$$\text{IC} = \frac{1}{2} f_p t_h \left(1 + \frac{t_d}{t_h} \right) = \frac{1}{2} f_p t_h (1 + \tau) \quad (3)$$

where τ is the ratio of the two durations corresponding to the bases of the trapezoid, $\frac{t_d}{t_h}$.

We considered f_p and τ as two independent factors of MDI technique from our model; t_h is the dependent parameter. The ratio τ is a shape parameter, as it governs the profile of the trapezoid, ranging from 1 to ∞ , from square ($t_h = t_d$) to triangular ($t_h = 0$). Figure 3c demonstrates the range of flow waveforms generated based on the constraint Equation (3) by varying f_p and τ . A higher value of τ for a particular f_p denoted longer inhalation duration (large t_d) with a small t_h , that is, a triangular trapezoid and vice versa.

For example, a 3 L waveform associated with a given f_p with high τ required a long time to attain the peak flow, that is, a high *rise-time*, and resulted in a smaller average flow rate; and vice versa. Thus, the flow waveforms with a particular f_p generated from the trapezoidal model had varying shapes (ranging from square to triangular) and varying

average flow rates (ranging from low to high), due to the different t_r and t_h governed by the value of τ . In our study, we varied the flow waveforms by choosing the factors f_p in the range from 0 to 90 L/min, and the ratio $\tau = 1.33, 3$, and 8, corresponding to a short, medium, and long duration of inhalation, respectively. Figure 3d illustrates how the flow waveforms generated from the trapezoidal waveforms fit inspiratory flow waveforms from actual patients.

The third independent parameter of MDI technique investigated was the time of MDI actuation, t_a , which was the time of firing the MDI relative to the start of inhalation, also known as coordination. In this study, $t_a = 0$ conveyed that the inhaler was actuated at exactly the time instant when the patient started to inhale through the MDI; a negative t_a conveyed that the inhaler was actuated before the start of inhalation; and a positive t_a denoted that the inhaler was actuated after the start of inhalation. In this study, we investigated the effects of coordination on *in vitro* lung deposition by varying t_a for $-0.5, 0$ seconds, $+0.6 \times t_r$, $0.5t_h = 0.5t_d$, and $(t_d - 0.6 \times t_r)$. The three positive t_a values are derived such that the MDI is actuated during the rising ($+0.6 \times t_r$), constant ($0.5t_d$), and falling edges ($t_d - 0.6 \times t_r$) of the trapezoid inspiratory waveform.

Note that the choice of values for f_p , τ , and t_a were based on the data reported in previous literature on the range of PIFRs, duration of inhalation, and MDI actuation time,⁽²⁴⁾ and also on the 43 MDI usage techniques recorded from the 15 patients in our study, as shown in Table 1.

In vitro experimental setup

We constructed an *in vitro* lung deposition experimental setup to generate the time-varying MDI techniques based on f_p , τ , and t_a , and analyzed corresponding lung deposition (see Fig. 4a for a schematic diagram, Fig. 4b for photograph of the setup). A programmable robotic arm was built using

Robotis' Dynamixel RX-64 servomotor (Trossen Robotics, Downers Grove, IL) to actuate the MDI at the predefined values of t_a . We used albuterol sulfate Ventolin HFA-MDIs (GlaxoSmithKline) in our experiments, labeled to dispense 108 mcg of albuterol sulfate at every actuation.

The inhaler was fitted to the same turbine flow meter used to record the maneuvers by the study subjects (Fig. 1), and connected with the anatomically correct mouth–throat cast. The mouth–throat cast used was an adult Alberta Idealized Throat or AIT (manufactured by Copley Scientific Ltd., United Kingdom), the accuracy of which had been validated by previous studies.^(15,21,22,24,29)

The AIT was connected to the collection filter, Respigard 303 (manufactured by Vital Signs), representing *in vitro* lungs. The filter was connected through another flow meter, Thor SpiroTube (manufactured by THOR Research and Development, Hungary), to the ATS and ISO certified pulmonary waveform generator or PWG (manufactured by Piston Ltd., Hungary). The setup was sealed using silicone connectors to ensure the maximum leakage was below 0.15 L (5%). A PC (personal computer) controlled the robotic arm, the two flow meters, and the PWG to generate MDI use-cases, monitor the procedure, and record observations. We performed MDI techniques based on f_p , τ , and t_a , derived from our model [Eq. (3)], using the combination of PWG and a robotic arm in the experimental setup.

Lung deposition testing was performed using AIT according to the guidelines provided by previous *in vitro* studies.^(15,21,22,29) At the beginning of each experiment, we cleaned, dried, and coated the inner surfaces of the AIT with silicone release spray, Molykote 316 (manufactured by Dow Corning, Auburn, MI), to reduce aerosol particle bounce. For each MDI use-case, we repeated the experiment thrice ($n = 3$), including multiple MDI actuations and rinsing of each component to recover the deposited drug. The first 2 repetitions were conducted with 10 MDI actuations and the third

FIG. 4. (a) Schematic to illustrate the *in vitro* experimental setup used in our work. A PC (personal computer) controls the motorized robotic MDI actuator, two flow meters, and the PWG. All the parts of the system are sealed using silicone connectors to minimize air leaks. The MDI is connected to the AIT, which represents the mouth–throat part of the body. The AIT is then connected to the collection filter, which acts as the lungs of the body. The different combinations of MDI techniques, inspiratory flow rates, duration of inhalations, and time of actuation are produced using the combination of the programmable PWG and motorized actuator. (b) Is a photograph of the actual setup used in the *in vitro* lung deposition experiments. AIT, Alberta Idealized Throat; PWG, pulmonary waveform generator.

with 15 actuations to reduce the dose-to-dose variability, as previous studies have shown that the amount of medication delivered by MDIs for each actuation is imprecise.⁽³⁰⁾ Furthermore, we employed a waiting period of 1 minute between actuations and vigorously shook the MDI 20 times before each actuation, as guided in previous studies.⁽³¹⁾

The total drug dispensed by the MDI comprises the amount of drug deposited in the MDI actuator mouthpiece, AIT, and collection filter. Thus the mass of the total drug delivered can be represented as follows:

$$m_t = m_{MP} + m_{AIT} + m_{Filter} \quad (4)$$

where,

m_t is the mass of total delivered drug

m_{MP} is the mass of drug deposited in the MDI actuator mouthpiece

m_{AIT} is the mass of drug deposited in the AIT or the mouth-throat deposition

m_{Filter} is the mass of the drug deposited in the filter or *in vitro* lung deposition.

In this study, the mass of the drug deposited in the filter, m_{Filter} , is an approximation of the total lung deposition and fraction of medication exhaled into the air.

The percentage *in vitro* lung deposition is computed as follows:

$$p_{ivL} = \frac{m_{Filter}}{m_t} \times 100\% = \frac{m_{Filter}}{m_{MP} + m_{AIT} + m_{Filter}} \times 100\% \quad (5)$$

$$= 100 - p_{MP} - p_{AIT} \quad (6)$$

where,

$$p_{MP} = \frac{m_{MP}}{m_t} \times 100\%, p_{AIT} = \frac{m_{AIT}}{m_t} \times 100\% \quad (7)$$

p_{ivL} is the percentage *in vitro* lung deposition of Albuterol

p_{MP} is the percentage drug deposition in the MDI actuator mouthpiece

p_{AIT} is the percentage drug deposition in the AIT

The mouthpiece, AIT, and filter were repeatedly rinsed with deionized water as outlined in previous studies⁽³⁰⁾ and the recovered solutions were analyzed using a UV spectrophotometer, Thermo Scientific™ GENESYS™ 10S UV-Vis Spectrophotometer, at 224 nm.

For our experiments, first f_p was varied from 0 to 90 L/min in steps of 30, and t_d was varied as a short, medium, and long inhalation for the ratio τ of 1.33, 3, and 8, respectively. t_a was kept constant at $0.6t_r$ for all values of f_p and τ , that is, actuation in the rising edge of the waveform or the first half of the inspiration. Next, the 60 L/min long inhalation waveform ($\tau = 8$) was selected and t_a was varied for small intervals from 0.5 seconds before the start of inspiration (-0.5 seconds) to $(t_d - 0.6t_r)$ (falling edge of the waveform at the end of inspiration), to study the effect of coordination on the *in vitro* lung deposition. The *in vitro* lung deposition data obtained were analyzed using one-way ANOVA and Tukey HSD for multiple comparisons at different f_p , τ , and t_a values, where a p -value < 0.05 was considered to be statistically significant.

Results

The results are divided into two categories based on the inhalation variables f_p , τ , and coordination variable t_a . The mean p_{ivL} for $f_p = 0$ L/min was found to be 5.7% and considered the baseline condition, and was, as expected, associated with the lowest drug deposition. The baseline is of similar magnitude to previously reported results, for example, 7% deposition using CFC inhalers.^(32,33)

The first category demonstrates the effect of inhalation waveform parameters, f_p and τ , on lung deposition. Table 2 lists the p_{ivL} , p_{MP} , and p_{AIT} for the MDI use-cases with different values of f_p and τ (and corresponding t_d), at $t_a = 0.6t_r$. There were three repetitions ($n = 3$) of all MDI use-cases except for $f_p = 90$ L/min and $\tau = 8$ ($n = 4$). The results are further visualized graphically in Figure 5a–c, corresponding to the inspiratory flow rates 30, 60, and 90 L/min, respectively.

The results show that mean p_{ivL} was lowest for $f_p = 30$ L/min, always below 40% for all values of τ . Furthermore, the mean p_{ivL} for both $f_p = 60$ L/min and 90 L/min was higher than 40% for all values of τ (correspondingly, all values of t_d) and therefore, significantly higher than the mean p_{ivL} for $f_p = 30$ L/min ($p < 0.001$). The increase in mean p_{ivL} for $f_p = 30$ L/min to corresponding waveforms of $f_p = 60 - 90$ L/min was found to be 5% ($p < 0.001$). There was no significant difference in p_{ivL} for $f_p = 60$ and 90 L/min ($p > 0.05$). Furthermore, Figure 5a shows a steady increase in p_{ivL} for $f_p = 30$ L/min with increasing τ (increase in the duration of inhalation). There was an increase of 3% in mean p_{ivL} for all flow rates due to an increase in τ from 1.33 to 3 and 8 ($p < 0.05$, $n = 3$). There was no significant effect on p_{ivL} for a change in τ from 3 to 8 ($p > 0.05$).

The next category of results demonstrates the effect of coordination, that is, t_a , on lung deposition. As we obtained higher p_{ivL} for $f_p = 60$ and 90 L/min and $\tau = 3$ and 8, we selected the inspiration waveform of $f_p = 60$ L/min and $\tau = 8$ (long duration of 5.33 seconds) to determine p_{ivL} for different values of t_a . Table 3 illustrates the p_{ivL} , p_{MP} , and p_{AIT} data collected for $t_a = -0.5, 0$ seconds, $0.6t_r$, $0.5t_d$, and $(t_d - 0.6t_r)$, and the results are graphically shown in Figure 5c. There were three repetitions ($n = 3$) of all MDI use-cases. We found that the p_{ivL} steadily increased with an increase in t_a for $t_a \leq 0.5t_d$.

Furthermore, the lung deposition value reached a plateau at the highest point in inspiratory flow rate, at $t_a = 0.5t_d$; the increase in p_{ivL} from $t_a = -0.5$ seconds to $t_a = 0.5t_d$ was of 23% ($p < 0.001$). The p_{ivL} for $t_a = -0.5$ seconds was significantly less than p_{ivL} for all other t_a ($p < 0.001$); similarly, the p_{ivL} for $t_a = 0$ seconds was significantly less than p_{ivL} for all positive t_a ($p < 0.001$). There was no significant difference in p_{ivL} for $t_a = 0.6t_r$ and $0.5t_d$. Furthermore, p_{ivL} for $t_a = (t_d - 0.6t_r)$ was significantly smaller than p_{ivL} for $t_a = 0.6t_r$ and $0.5t_d$ ($p < 0.01$).

Discussion

Our study results demonstrate that for a positive coordination in the first half of inspiration, a higher $f_p = 60 - 90$ L/min yielded better lung deposition compared to lower $f_p = 30$ L/min, independent of the duration of inspiration; the increase in mean p_{ivL} was found to be 5%. Thus, the positive effect of high PIFR on lung deposition of Albuterol from Ventolin HFA MDI observed

TABLE 2. PERCENTAGE LUNG DEPOSITION FOR *IN VITRO* EXPERIMENTS FOR VARIATION IN INSPIRATORY PARAMETERS, PEAK INSPIRATORY FLOW RATE (F_p) AND SHAPE (τ), WHERE THE TIME OF ACTUATION IS IN THE FIRST HALF OF INSPIRATION, SET AT $t_A = 0.6t_R$

No.	Trial No.	PIFR (L/min), f_p	Shape parameter, τ	Time duration (seconds), t_d	Time of actuation (seconds), t_a	Percent MP deposition, p_{MP}	Percent AIT deposition, p_{AIT}	Percent in vitro lung deposition, p_{ivL}	Mean percent in vitro lung deposition, mean p_{ivL}
1	1	0	0	0	0	32.5	63.5	4	5.7
	2					32.6	63.4	4	
	3					33.5	58.7	7.8	
	4 ^a					28.6	64.4	7	
2	1	30	1.33	6.85	0.51	15.2	47.8	37	35.1
	2					15.9	49.5	34.6	
	3					13	53.4	33.6	
3	1	30	3	9	1.8	14.3	48.9	36.8	37.9
	2					14.1	48.3	37.6	
	3					16.4	44.4	39.2	
4	1	30	8	10.67	2.8	15.5	45.2	39.3	39.2
	2					14.9	45.8	39.4	
	3					16.6	44.3	39	
5	1	60	1.33	3.42	0.255	13.9	46.5	39.6	40.3
	2					12.7	46.8	40.5	
	3					12.2	47	40.9	
6	1	60	3	4.5	0.9	11.1	43.3	45.7	44.2
	2					10.9	45.5	43.6	
	3					11	45.8	43.3	
7	1	60	8	5.33	1.4	10.6	49.4	40	42.3
	2					12	44.6	43.5	
	3					11.2	45.4	43.4	
8	1	90	1.33	2.28	0.17	13.7	45.9	40.4	41.4
	2					13.8	44.5	41.8	
	3					13.5	44.5	41.9	
9	1	90	3	3	0.6	11.5	46.5	42	43.3
	2					14.7	41.2	44.1	
	3					13.7	42.4	43.9	
10	1	90	8	3.56	0.93	10.6	44.5	44.9	42.8
	2					14.3	44.3	41.5	
	3					13	44.9	42	
	4 ^a					12.7	44.6	42.8	

The trial numbers 1–2 correspond to experiments conducted with 10 doses and the number 3 corresponds to experiments repeated with 15 doses. The first row corresponds to baseline case of no flow rate.

^aNote that the fourth repetition for some of the experiments (with 10 doses) was carried out as we noticed a larger variation for that data point.

AIT, Alberta Idealized Throat; PIFR, peak inspiratory flow rate.

in our *in vitro* experimentation is not only of opposite trend but also less critical (an increase of 5% only) from what has been reported in the past literature.^(9–14)

As mentioned earlier, a number of recent publications, particularly the work of Usmani et al., have suggested that the particle of smaller sizes (1.5 μm) delivered from various HFA inhalers have better lung deposition (4.6% increase) at high mean flow rates ($67.1 \pm 16.7 \text{ L/min}$) compared to low mean flow rates ($30.8 \pm 4.7 \text{ L/min}$).^(15,17–20,34) However, for particles of 3 μm , the lung deposition at high mean flow rates decreased slightly (1.4% decrease) compared to low flow rates, decreasing more dramatically for larger particles of size 6 μm (24.7% decrease).⁽¹⁸⁾

Furthermore, Cheng et al. had demonstrated that, for the *in vitro* study using albuterol sulfate HFA MDI (Proventil) with a mass median aerodynamic diameter (MMAD) of 2.21 μm , the mouth–throat deposition decreased significantly on increasing flow rate from 30 to 90 L/min, consequently increasing the *in vitro* lung deposition.⁽²⁰⁾ A high

oropharyngeal deposition at low flow rates was attributed to the higher speed of aerosol droplets compared to the inspiratory flow rate. Therefore, at higher inspiratory flow rates, a possibly more efficient momentum transfer occurred between the aerosol droplets and air flow, such that a larger fraction of aerosol droplets followed the air stream into the lungs, resulting in a lower mouth–throat deposition.

Furthermore, the authors also suggested that at higher flow rates, the evaporation of aerosol emitted from the MDI is more efficient in generating smaller droplets resulting in lower oropharyngeal deposition. In addition, the authors also stated that the impact force of the aerosol sprays from HFA MDIs are much smaller than CFC MDIs due to the smaller orifice diameters of HFA MDIs, which changes the plume characteristics (softer due to smaller aerosol velocities) and reduces oropharyngeal deposition.

Furthermore, Rahmatalla et al. also demonstrated that with Beclomethasone HFA MDI (QVAR), which has a high fine fraction close to 1 μm , the increased *in vitro* lung

FIG. 5. The percentage deposition in the *in vitro* lung (black bar), AIT (gray bar), and MDI mouthpiece (white bar) for the *in vitro* experiments are presented. The graph shows the results from experiments performed for inhalation (flow, duration)-dependent MDI techniques with MDI actuation fixed at the first half of inspiration. Each group of bars represents one MDI technique; the error bars show the range due to three repetitions of each experiment. The baseline result is shown in the beginning marked 0 L/min. The error margins for *in vitro* lung deposition ranged from $\pm 0.2\%$ to $\pm 2.3\%$, average $\pm 1.3\%$. Three flow rates, 30, 60, and 90 L/min, were used for each of the experiments. (a–c) Show the variation in lung deposition with shape ratio, τ (corresponding to the duration of inhalation), for the three different flow rates 30, 60, and 90 L/min, respectively. The short, medium, and long labels on the x-axis refer to τ equal to 1.33, 3, and 8, respectively. The inhalers were actuated at the beginning of the inhalation, $t_a = 0.6t_r$. We found that the mean *in vitro* lung deposition (black bar) is higher for 60–90 L/min than 30 L/min ($p < 0.001$), always greater than 40%. The maximum mean *in vitro* lung deposition was found to be 44.2% for the case of inspiration with PIFR 60 L/min and duration of 4.5 seconds. PIFR, peak inspiratory flow rate.

deposition at high flow rates was due to alternate mechanisms of aerosol deposition in addition to impaction; such as sedimentation and diffusion.⁽¹⁵⁾ Hoyer et al. also demonstrated that the MMAD of Ventolin Albuterol HFA MDIs was $2.07 \mu\text{m}$.⁽¹⁹⁾ Therefore, we conclude from our observed experimental results and past literature^(15,17–20,34) that due to a combination of mechanisms governing the deposition of albuterol aerosol delivered by Ventolin HFA MDIs, having a small particle size (MMAD $2.07 \mu\text{m}$) compared to CFC counterparts, the *in vitro* lung deposition is slightly higher for PIFRs in the range 60–90 L/min than low PIFRs.

A further extension of our work would be to measure particle size distribution using a cascade impactor for our dynamic trapezoidal inspiration waveforms to confirm higher fine fraction of aerosol delivered by HFA Albuterol MDI (Ventolin) and the corresponding dependency of PIFR on lung deposition obtained in our experiments.

Our results on the study of the effects of coordination on *in vitro* lung deposition demonstrated that the highest lung deposition occurred with positive coordination, especially ac-

tuation in the first half of the inspiratory waveform. The increase in p_{inV} from $t_a = -0.5$ seconds to $t_a = 0.6t_r$ was 23%, suggesting that coordination was the most critical parameter in determining p_{inV} . Furthermore, compared to flow and duration parameters, a negative coordination caused the maximum loss in p_{inV} . As seen from the Figure 6, at negative and zero coordination, the mean p_{MP} is much higher (almost double) than for positive coordination. Thus, when the MDI is actuated without any inspiratory flow rate, about 20%–32% of the medication is deposited in the plastic mouthpiece of the inhaler (compared to 12% for positive coordination), thereby reducing the amount of respirable aerosol medication. The baseline case of $f_p = 0$ L/min, therefore, has the highest amount of p_{MP} (32%) and p_{AIT} (62.5%) of all cases as there is no inspiratory flow rate even after MDI actuation.

Note that for very late actuation, that is actuation in the second half of the inspiration with $t_a = (t_d - 0.6t_r)$, the presence of inspiration at the time of actuation ensured that the p_{MP} is low (13%). However, due to the small volume of air inhaled coupled with a decreasing flow rate postactuation, the

TABLE 3. PERCENTAGE LUNG DEPOSITION FOR *IN VITRO* EXPERIMENTS WITH ACTUATION TIME t_a , RANGING FROM 0.5 SECOND BEFORE START OF INHALATION (-0.5 SECONDS) TO FALLING EDGE OF THE WAVEFORM AT THE END OF INSPIRATION ($t_d - 0.6t_r$), FOR THE INSPIRATORY WAVEFORM WITH $f_p = 60$ L/MIN AND $\tau = 8$ (OR $t_d = 5.33$)

No.	Trial No.	Time of actuation (seconds), t_a	Percent MP deposition, PMP	Percent AIT deposition, PAIT	Percent in vitro lung deposition, p_{ivL}	Mean percent in vitro lung deposition, mean p_{ivL}
1	1	Baseline $f_p = 0$ L/min, $t_a = 0$	32.5	63.5	4	5.7
	2		32.6	63.4	4	
	3		33.5	58.7	7.8	
	4		28.6	64.4	7	
2	1	-0.5	27.1	50.6	22.3	20.8
	2		24.3	55	20.7	
	3		27	53.5	19.5	
3	1	0	20	51.2	28.8	28.7
	2		20	51.1	28.9	
	3		27	44.8	28.3	
4	1	1.4 ($0.6t_r$)	10.6	49.4	40	42.3
	2		12	44.6	43.5	
	3		11.2	45.4	43.4	
5	1	2.67 ($0.5t_d$)	12.7	44.1	43.1	43.8
	2		11.4	45.6	42.9	
	3		13.1	41.4	45.5	
6	1	3.93 ($t_d - 0.6t_r$)	13.5	49.3	37.2	38.9
	2		11.2	50	38.8	
	3		14.1	45.3	40.7	

The trial numbers 1–2 correspond to experiments conducted with 10 doses and the number 3 corresponds to experiments repeated with 15 doses.

p_{AIT} (48%) increased and the p_{ivL} (39%) decreased compared with actuation in the first half of inspiration ($t_a = 0.6t_r$, $0.5t_d$).

The two categories of results investigating the effect of PIFR (f_p), duration of inspiration (t_d), defined in this study

FIG. 6. The percentage deposition in the *in vitro* lung (black bar), AIT (gray bar), and MDI mouthpiece (white bar) as a function of coordination or time of MDI actuation relative to the start of inspiration are presented. Each group of bars represents one MDI technique, the error bars show the range due to three repetitions of each experiment. The baseline result is shown in the beginning marked 0 L/min. The other bars correspond to the MDI technique of PIFR 60 L/min and duration 5.33 seconds, actuated at -0.5 , 0 seconds, $0.6t_r$, $0.5t_d$, and ($t_d - 0.6t_r$). The error margins for *in vitro* lung deposition ranged from $\pm 0.2\%$ to $\pm 2.3\%$, average $\pm 1.3\%$. We found that the mean *in vitro* lung deposition is highest for actuation in the first half of inspiration, 43.8%, and lowest for negative coordination.

by the shape of inspiration waveform τ , and coordination (t_a) suggest that the optimal technique for maximizing the lung deposition of aerosol emitted from an HFA Ventolin MDI for 3 L IC is an inspiration with PIFR of 60–90 L/min of any duration in the range 2.28–5.33 seconds with a positive coordination in the first half of inspiration.

Our study also has some limitations. As is the case with using simplified models to represent complex human maneuvers, our trapezoidal breathing model suffers from some disadvantages. The trapezoid model we developed has a symmetrical form, which is not the case for most inhalation curves recorded; see Figure 3a. The human inhalation waveforms have previously been approximated by asymmetric sinusoidal curves.^(35,36) The asymmetric sinusoidal approximations yield waveforms much closer to actual breathing curves and hence may be a better choice for *in vitro* experimental procedures. However, because MDI actuation in the first half of the inhalation has the highest impact on lung deposition, we believe that the asymmetric trapezoidal waveforms may have similar lung deposition characteristics as that of symmetric trapezoidal waveforms, and thus is a valid first-order approximation.

Finally, the results obtained in our *in vitro* experimentation are an approximation to the actual *in vivo* lung deposition. Further investigation into this work would be to explore *in vitro* lung deposition using a more realistic lung model with lung airways of different sizes and *in vivo* lung deposition on human subjects, to explore the distribution of aerosol in different parts of the lung due to the variation in MDI technique.

In summary, in our *in vitro* experiment modeling of MDI technique based on the maneuvers recorded from subjects with asthma and COPD, we demonstrated that coordination of MDI actuation is the most important factor governing lung deposition of aerosol emitted by an Albuterol HFA-MDI (Ventolin). While higher inspiratory flow rates and

increased duration of inspiration also yielded higher lung deposition, the increase was lower than the effect of coordination. Based on our results, the recommendation for optimal lung deposition for patients using an HFA containing MDI would be to inhale at high inspiratory flow rates, in the range from 60 to 90 L/min, and actuate the MDI in the first half of inspiration.

Acknowledgments

This study was supported by funding from National Institute of Health (Grant: NIH R41DA039450-01), provided by Cognita Labs LLC. The authors particularly thank Mr. Gaurav Patel for his knowledge, guidance, and expertise throughout the study. The authors also gratefully acknowledge Dr. Maura Puerto for her expert knowledge and guidance with the project and experimentation. The authors acknowledge Ms. June Chen for her contributions in the *in vitro* experimentation. The authors also greatly appreciate the help of Hadeel Abdulrazzaq, MD, and Caryn Pope, RCP, for overseeing the patient data collection at Ben Taub Hospital, Houston, Texas.

Author Disclosure Statement

Dr. Ashutosh Sabharwal is a Professor, Department of Electrical and Computer Engineering at Rice University, Houston, Texas. He is also a cofounder of Cognita Labs LLC. Ms. Rajoshi Biswas is a PhD candidate in the Department of Electrical and Computer Engineering at Rice University. She is also an employee of Cognita Labs LLC. Dr. Nicola Hanania is an Associate Professor, Baylor College of Medicine; Director, Airways Clinical Research Center, Houston, Texas. No competing financial interests exist.

References

1. Capstick TG, and Clifton IJ: Inhaler technique and training in people with chronic obstructive pulmonary disease and asthma. *Expert Rev Respir Med*. 2012;6:91–103.
2. Rubin BK: What does it mean when a patient says, “My asthma medication is not working?” *Chest*. 2004;126:972–981.
3. Roche N, Chrystyn H, Lavorini F, and Agusti A: Effectiveness of inhaler devices in adult asthma and COPD. *EMJ Respir*. 2013;1:64–71.
4. Melani AS, Bonavia M, Cilenti V, Cinti C, Lodi M, Martucci P, Serra M, Scichilone N, Sestini P, Aliani M, and Neri M: Inhaler mishandling remains common in real life and is associated with reduced disease control. *Respir Med*. 2011;105:930–938.
5. Laube BL, Janssens HM, De Jongh FHC, Devadason SG, Dhand R, Diot P, Everard ML, Horvath I, Navalesi P, Voshaar T, and Chrystyn H: What the pulmonary specialist should know about the new inhalation therapies. *Eur Respir J*. 2011;37:1308–1331.
6. Bousquet J, and Khaltaev N: Global surveillance, prevention and control of chronic respiratory diseases: A comprehensive approach. *World Health Organ*. 2006;3:215–216.
7. Global Initiative for Asthma: *Pocket Guide for Asthma Management and Prevention*. Global Initiative for Asthma; p. 28, 2016.
8. Smith KJ, Chan HK, and Brown KF: Influence of flow rate on aerosol particle size distributions from pressurized and breath-actuated inhalers. *J Aerosol Med*. 1998;11:231–245.
9. Pauwels R, Newman S, and Borgström L: Airway deposition and airway effects of antiasthma drugs delivered from metered-dose inhalers. *Eur Respir J*. 1997;10:2127–2138.
10. Newman S, Steed K, Hooper G, Källén A, and Borgström L: Comparison of gamma scintigraphy and a pharmacokinetic technique for assessing pulmonary deposition of terbutaline sulphate delivered by pressurized metered dose inhaler. *Pharma Res*. 1995;12:231–236.
11. Newman SP, Pavia D, and Clarke SW: Simple instructions for using pressurized aerosol bronchodilators. *J R Soc Med*. 1980;73:776–779.
12. Newman SP, Pavia D, and Clarke SW: How should a pressurized beta-adrenergic bronchodilator be inhaled? *Eur J Respir Dis*. 1981;62:3–21.
13. Newman SP, Pavia D, Garland N, and Clarke SW: Effects of various inhalation modes on the deposition of radioactive pressurized aerosols. *Eur J Respir Dis Suppl*. 1982;119: 57–65.
14. Lawford P, and McKenzie D: Pressurized aerosol inhaler technique: How important are inhalation from residual volume, inspiratory flow rate and the time interval between puffs? *Br J Dis Chest*. 1983;77:276–281.
15. Rahmatalla MF, Zuberbuhler PC, Lange CF, and Finlay WH: *In vitro* effect of a holding chamber on the mouth-throat deposition of QVAR (hydrofluoroalkane-beclomethasone dipropionate). *J Aerosol Med*. 2002;15:379–385.
16. Lewis DA, Meakin BJ, and Brambilla G: New actuators versus old: Reasons and results for actuator modifications for HFA solution MDIs. *Respir Drug Deliv*. 2006;1:101–110.
17. Leach CL, Davidson PJ, and Boudreau RJ: Improved airway targeting with the CFC-free HFA-beclomethasone metered-dose inhaler compared with CFC-beclomethasone. *Eur Respir J*. 1998;12:1346–1353.
18. Usmani OS, Biddiscombe MF, and Barnes PJ: Regional lung deposition and bronchodilator response as a function of beta2-agonist particle size. *Am J Respir Crit Care Med*. 2005;172:1497–1504.
19. Hoye WL, Mogalian EM, and Myrdal Paul B: Effects of extreme temperatures on drug delivery of albuterol sulfate hydrofluoroalkane inhalation aerosols. *Am J Health Syst Pharm*. 2005;62:227.
20. Cheng YS, Fu CS, Yazzie D, and Zhou Y: Respiratory deposition patterns of salbutamol pMDI with CFC and HFA-134a formulations in a human airway replica. *J Aerosol Med*. 2001;14:255–266.
21. Ruzzycki CA, Golshahi L, Vehring R, and Finlay WH: Comparison of *in vitro* deposition of pharmaceutical aerosols in an idealized child throat with *in vivo* deposition in the upper respiratory tract of children. *Pharm Res*. 2014;31: 1525–1535.
22. Zhang Y, Gilbertson K, and Finlay WH: *In vivo-in vitro* comparison of deposition in three mouth-throat models with Qvar and Turbuhaler inhalers. *J Aerosol Med*. 2007;20: 227–235.
23. Longest PW, Tian G, Li X, Son YJ, and Hindle M: Performance of combination drug and hygroscopic excipient submicrometer particles from a softmist inhaler in a characteristic model of the airways. *Ann Biomed Eng*. 2012; 40:1–15.
24. Olsson B, Borgström L, Lundbäck H, and Svensson M: Validation of a general *in vitro* approach for prediction of total lung deposition in healthy adults for pharmaceutical inhalation products. *J Aerosol Med Pulm. Drug Deliv*. 2013; 26:355–369.

25. *Inhaler Technique in Adults with Asthma or COPD*. National Asthma Council Australia pp. 1–9, 2008.
26. Price D, Bosnic-Anticevich S, Briggs A, Chrystyn H, Rand C, Scheuch G, and Bousquet J: Inhaler competence in asthma: Common errors, barriers to use and recommended solutions. *Respir Med*. 2013;107:37–46.
27. Giraud V, Allaert F-A, and Roche N: Inhaler technique and asthma: Feasibility and acceptability of training by pharmacists. *Respir Med*. 2011;105:1815–1822.
28. Ganong WF, and Barrett KE: *Review of Medical Physiology*. McGraw-Hill Medical, New York; 2005.
29. Shemirani FM, Hoe S, Lewis D, Church T, Vehring R, and Finlay WH: Investigation of the effect of ambient humidity on regional delivered dose with solution and suspension MDIs. *J Aerosol Med Pulm Drug Deliv*. 2012;26: 215–222.
30. Sprigge E, and Affairs P: Non uniform deposition of pMDI aerosol in a large volume spacer. 2014. (Doctoral dissertation, Carleton University Ottawa)
31. Hautmann J, Godoy SE, Marshik P, Chand R, Mcconville J, Krishna S, Krishna S, and Muttill P: Effect of time between actuation on the dose variability for three metered dose inhalers. *Respir Drug Deliv Eur*. 2013;2:429–434.
32. Newman SP, Weisz AW, Talae N, and Clarke SW: Improvement of drug delivery with a breath actuated pressurised aerosol for patients with poor inhaler technique. *Thorax*. 1991;46:712–716.
33. Chrystyn H, and Price D: Not all asthma inhalers are the same: Factors to consider when prescribing an inhaler. *Prim Care Respir J*. 2009;18:243–249.
34. Leach CL, Davidson PJ, Hasselquist BE, and Boudreau RJ: Influence of particle size and patient dosing technique on lung deposition of HFA-beclomethasone from a metered dose inhaler. *J Aerosol Med*. 2005;18:379–385.
35. Delvadia RR, Longest PW, and Byron PR: Tests for aerosol deposition. I: Scaling a physical model of the upper airways to predict drug deposition variation in normal humans. *J Aerosol Med Pulm Drug Deliv*. 2012;25:32–40.
36. Longest PW, Tian G, Walenga RL, and Hindle M: Comparing MDI and DPI aerosol deposition using *in vitro* experiments and a new stochastic individual path (SIP) model of the conducting airways. *Pharm Res*. 2012;29:1670–1688.

Received on October 30, 2015
in final form, December 29, 2016

Reviewed by:
Jolyon Mitchell
Chet Leach

Address correspondence to:
Ashutosh Sabharwal, PhD
Department of Electrical and Computer Engineering
Rice University
6100 Main Street
MS 380
Houston, TX 77005
E-mail: ashu@rice.edu