Published in final edited form as:

Appl Magn Reson. 2015 January ; 46(1): 59–66. doi:10.1007/s00723-014-0612-9.

Design and Test of Magnetic Wall Decoupling for Dipole Transmit/Receive Array for MR Imaging at the Ultrahigh Field of 7T

Xinqiang Yan,

State Key Laboratory of Brain and Cognitive Science, Beijing MRI Center for Brain Research, Institute of Biophysics, Chinese Academy of Sciences, Bldg. 11, 15 Datun Road, Chaoyang District, Beijing 100101, China

Key Laboratory of Nuclear Radiation and Nuclear Energy Technology, Institute of High Energy Physics, Chinese Academy of Sciences, Beijing 100049, China

Beijing Engineering Research Center of Radiographic, Techniques and Equipment, Beijing 100049, China

Xiaoliang Zhang,

Department of Radiology and Biomedical Imaging, University of California San Francisco, Byers Hall, Room 102, 1700 4th ST, San Francisco, CA 94158-2330, USA

UCSF/UC Berkeley Joint Graduate Group in Bioengineering, San Francisco, CA 94158, USA

Long Wei, and

Key Laboratory of Nuclear Radiation and Nuclear Energy Technology, Institute of High Energy Physics, Chinese Academy of Sciences, Beijing 100049, China

Beijing Engineering Research Center of Radiographic, Techniques and Equipment, Beijing 100049, China

Rong Xue

State Key Laboratory of Brain and Cognitive Science, Beijing MRI Center for Brain Research, Institute of Biophysics, Chinese Academy of Sciences, Bldg. 11, 15 Datun Road, Chaoyang District, Beijing 100101, China

Beijing Institute for Brain Disorders, Beijing 100053, China

Abstract

Radio-frequency coil arrays using dipole antenna technique have been recently applied for ultrahigh field magnetic resonance (MR) imaging to obtain the better signal–noise–ratio (SNR) gain at the deep area of human tissues. However, the unique structure of dipole antennas makes it challenging to achieve sufficient electromagnetic decoupling among the dipole antenna elements. Currently, there is no decoupling methods proposed for dipole antenna arrays in MR imaging. The recently developed magnetic wall (MW) or induced current elimination decoupling technique has demonstrated its feasibility and robustness in designing microstrip transmission line arrays, L/C

loop arrays and monopole arrays. In this study, we aim to investigate the possibility and performance of MW decoupling technique in dipole arrays for MR imaging at the ultrahigh field of 7T. To achieve this goal, a two-channel MW decoupled dipole array was designed, constructed and analyzed experimentally through bench test and MR imaging. Electromagnetic isolation between the two dipole elements was improved from about -3.6 dB (without any decoupling treatments) to -16.5 dB by using the MW decoupling method. MR images acquired from a water phantom using the MW decoupled dipole array and the geometry factor maps were measured, calculated and compared with those acquired using the dipole array without decoupling treatments. The MW decoupled dipole array demonstrated well-defined image profiles from each element and had better geometry factor over the array without decoupling treatments. The experimental results indicate that the MW decoupling technique might be a promising solution to reducing the electromagnetic coupling of dipole arrays in ultrahigh field MRI, consequently improving their performance in SNR and parallel imaging.

1 Introduction

Ultrahigh magnetic resonance imaging (MRI) could provide a high signal–noise-ratio (SNR) and a high contrast in human imaging [1–10]. Dipole arrays, which have a simple structure, have been proposed to increase the B_1 field penetration and SNR gain in the deep area of human head or body and have demonstrated promising capability in human imaging at ultrahigh magnetic fields [11, 12]. It is known that the electromagnetic (EM) coupling among coil elements is critical to SNR and parallel imaging performance [13–19] of coil arrays. Due to the unique structure of the dipole antennas, the traditional decoupling methods, such as element overlapping [20] and L/C decoupling network [21–28], face challenges and are not readily feasible for dipole transmit/receive (transceiver) arrays. Insufficient decoupling among dipole array elements in current dipole array designs has negatively impacted the imaging performance. The recently developed magnetic wall (MW) or induced current elimination (ICE) decoupling technique has demonstrated its superior capability of decoupling resonant elements in various coil arrays, including microstrip transmission line (MTL) arrays [29–32], traditional L/C loop arrays [33–35], and monopole arrays [36]. For the MW decoupling method, an independent resonator/antenna was used to eliminate the current induced in the resonant elements by EM coupling. This method does not need physical connection between coil elements and should be promising for dipole transceiver arrays.

In this study, we aim to test the feasibility and investigate the performance of the MW decoupling method in the dipole array. A two-channel MW decoupled dipole array was designed and constructed for 7T MR imaging. As a comparison, we also constructed a 7T dipole array without any decoupling treatments. Bench test, MR imaging and geometry factor results obtained from the two dipole arrays demonstrate that MW decoupling method is capable of obtaining a high degree of EM decoupling between dipole elements and provides enhanced imaging performance at the ultrahigh field of 7T.

2 Materials and Methods

2.1 Design and Construction of the Two-Channel Dipole Array Without and With Magnetic Wall Decoupling Method

Figure 1 shows the photographs and sketches of the constructed two-channel dipole arrays without decoupling treatments and with the MW decoupling method. A simple dipole antenna is equal to $1/2$ wavelength of the electromagnetic wave in the free space [37]. At 297.2 MHz, the Larmor frequency of our 7T MRI scanner, the corresponding length of a simple dipole antenna is about 50 cm, which is a little longer and not suitable for most MRI applications. In this study, the length of each dipole was shortened by two series inductors (Ls, as shown in Fig. 1c, d) to 26 cm, which is suitable for human head MR imaging. Each series inductor was wound with a copper wire of four turns and 4.0 mm in diameter. The distance between two dipole elements of each array is 7 cm. Each dipole element was matched to 50Ω with a trimmer capacitor C_m [12], and was tuned to 297.2 MHz by slightly changing the gap between the turns of Ls. Traditional L/C trap circuits were applied to remove the possible “cable resonance” of each dipole element at the high frequency of 297.2 MHz.

For the magnetic wall decoupled two-channel dipole array, an independent dipole antenna, referred to as the decoupling element, was symmetrically placed between the two dipole elements to reduce the EM coupling. The decoupling element had the same length as the dipole element and was shorted by a series inductor L_d , referred to as the decoupling inductor (Fig. 1d). The decoupling inductor was wound with a copper wire of three turns and 4.0 mm in diameter. The resonance frequency of the decoupling element is about 400 MHz. All conductor strips described above were made of the 10 mm-wide copper tape (3 M, St, Paul, MN).

Figure 2 shows the position of the water phantom vs. dipole arrays. A cylindrical water phantom with an outer diameter of 16 cm and a length of 37 cm was placed 2.5 cm below the dipole arrays. The EM parameters of the water phantom were measured by a dielectric probe (DAK-12, Speag, Switzerland): conductivity $\sigma = 0.59$ S/m; relative permittivity $\epsilon_r = 78$.

2.2 Bench Tests and MR Imaging Experiments

Scattering (S –) parameters, including reflection coefficient (S_{11}) and transmission coefficient (S_{21}), of the dipole elements with and without the MW decoupling method were measured with an Agilent E5071C network analyzer. The reflection coefficient measurements were also used to calculate the coil's Q value.

To demonstrate the decoupling performance of the MW decoupling method for dipole arrays, gradient recalled echo (GRE) images and geometry factor maps on the water phantom using the two dipole elements with and without the decoupling element were measured, calculated and compared. The slice chosen for MR imaging was in the central transverse plane of the dipole elements. The imaging acquisition parameters were: flip angle (FA) = 25° , TR = 100 ms, TE = 10 ms, field of view (FOV) = 250×171 mm², matrix = 256×176 , slice thickness = 5 mm, bandwidth = 320 Hz/pixel, number of excitation (NEX) = 1.

The MRI experiments were performed on a whole-body MRI scanner (7T MAGNETOM, Siemens Healthcare, Erlangen, Germany). During the experiments, one dipole element was used for both transmission and reception with the other one terminated with 50- Ω load. The geometry factor maps were calculated using an radio-frequency coil array design and analysis software Musaik (Speag, Switzerland).

3 Results

3.1 Bench Test Results

The reflection coefficient (S_{11}) of each dipole element is better than -20 dB, as shown in Fig. 3. For the dipole array without decoupling treatments, the transmission coefficient S_{21} between two dipole elements was only about -3.6 dB (Fig. 3a). It is worth noting that no obvious peak split was observed even when the coupling achieved -3.6 dB, which is unlike near-field coil such as loop or MTL coil. The average unloaded Q value (Q_{UL}) and Q value loaded with water phantom (Q_L) of a single element were about 13.3 and 9.5, respectively.

For the MW decoupled two-channel dipole array, the length of the decoupling element was chosen the same as the dipole elements (26 cm). The transmission coefficient S_{21} between the dipole elements could be reduced from -3.6 to -16.5 dB, as shown in Fig. 3b. The average Q_{UL} and Q_L of a single element were 48.1 and 29.2, respectively. Like the results of our previous work [36], the MW decoupled array exhibits higher Q value which might be due to the better decoupling performance and shielding effect of the decoupling element.

3.2 MR Images and Geometry Factor Maps

GRE images from each dipole element of the two-channel arrays were shown in Fig. 4. Figure 4a, b show the images from two coupled dipoles and the 1D profiles of the images. Figure 4c, d show the images from two MW decoupled dipoles and the 1D profile of the images. This significant decoupling effect, as obtained from the S_{21} plots, could also be clearly observed in MR images. Well-defined image profiles of MW decoupled dipole array were obtained, indicating sufficient electromagnetic decoupling between the two elements.

Geometry factor maps with reduced factor (R) of two were shown in Fig. 5. The average and maximum geometry factors of the coupled dipole array are 1.14 and 3.2, respectively; whereas the average and maximum geometry factor of the MW decoupled dipole array are 1.08 and 2.4, respectively. Both average and maximum geometry factor results reveal that the MW decoupled array exhibits better parallel imaging ability, which can also be verified by the comparison of S-parameter results as described above. In parallel imaging, the SNR of accelerated images is inversely proportional to the geometry factor [14]. Therefore, the results of the geometry factor measurements suggest that better quality-accelerated images can be achieved using MW decoupled transceiver dipole array over the array without decoupling treatments.

4 Discussions and Conclusions

The feasibility of magnetic wall decoupling technique for dipole coil arrays has been validated through bench tests and MR imaging experiments at the ultrahigh field of 7T in

this study. The experimental results demonstrate that the transmission coefficient S_{21} of the strongly coupled dipole elements is reduced from -3.6 to -16.5 dB using the MW decoupling technique. The EM isolation between the dipole elements might be further improved by optimizing the length of the MW decoupling element. Due to improved decoupling performance, the MW decoupled dipole elements demonstrate well-defined image profiles and improved geometry factors, indicating that better parallel imaging capability over the dipole array without decoupling treatments can be expected. This study has paved the way for designing MW decoupled multi-channel volume-typed dipole array for human MR imaging at ultrahigh fields.

Acknowledgments

We thank Mr. Lei Shi and Mr. Zihao Zhang from Institute of Biophysics, Chinese Academy of Sciences for their assistance on MR experiments. This work was supported in part by Chinese National Major Scientific Equipment R&D Project (grant number ZDYZ2010-2), the Ministry of Science and Technology (MOST) of China (grant number 2012CB825500), the Chinese Academy of Sciences Strategic Priority Research Program (grant numbers XDB02010001, XDB02050001), National Natural Science Foundation of China Grant 51228702 and National Institutes of Health (NIH) Grants R01EB008699 and P41EB013598.

References

1. Yacoub E, Shmuel A, Pfeuffer J, Van de Moortele PF, Adriany G, Andersen P, Vaughan JT, Merkle H, Ugurbil K, Hu XP. *Magn. Reson. Med.* 2001; 45:588–594. [PubMed: 11283986]
2. Zhang X, Ugurbil K, Chen W. *Magn. Reson. Med.* 2001; 46:443–450. [PubMed: 11550234]
3. Lei H, Zhu X, Zhang X, Ugurbil K, Cheng W. *Magn. Reson. Med.* 2003; 49:199–205. [PubMed: 12541238]
4. Zhu X, Zhang Y, Tian R, Lei H, Zhang N, Zhang X, Merkle H, Ugurbil K, Cheng W. *Proc. Natl. Acad. Sci. USA.* 2003; 100:4352.
5. Zhang X, Ugurbil K, Sainati R, Chen W. *IEEE Trans. Biomed. Eng.* 2005; 52:495–504. [PubMed: 15759580]
6. Zhang X, Zhu XH, Chen W. *Magn. Reson. Med.* 2005; 53:1234–1239. [PubMed: 15844152]
7. Wiggins GC, Potthast A, Triantafyllou C, Wiggins CJ, Wald LL. *Magn. Reson. Med.* 2005; 54:235–240. [PubMed: 15968650]
8. Ibrahim TS, Mitchell C, Schmalbrock P, Lee R, Chakeres DW. *Magn. Reson. Med.* 2005; 54:683–690. [PubMed: 16088934]
9. Thomas BP, Welch EB, Niederhauser BD, Whetsell WO Jr, Anderson AW, Gore JC, Avison MJ, Creasy JL. *J. Magn. Reson. Imaging.* 2008; 28:1266–1272. [PubMed: 18972336]
10. Chang G, Deniz CM, Honig S, Egol K, Regatte RR, Zhu Y, Sodickson DK, Brown R. *J. Magn. Reson. Imaging.* 2014; 39:1384–1393. [PubMed: 24115554]
11. Raaijmakers AJ, Ipek O, Klomp DW, Possanzini C, Harvey PR, Lagendijk JJ, van den Berg CA. *Magn. Reson. Med.* 2011; 66:1488–1497. [PubMed: 21630342]
12. Wiggins GC, Zhang B, Lattanzi R, Chen G, Sodickson D. *Proceedings of the 20th annual meeting of ISMRM; Melbourne.* 2012. p. 541
13. Sodickson DK, Manning WJ. *Magn. Reson. Med.* 1997; 38:591–603. [PubMed: 9324327]
14. Pruessmann KP, Weiger M, Scheidegger MB, Boesiger P. *Magn. Reson. Med.* 1999; 42:952–962. [PubMed: 10542355]
15. Griswold MA, Jakob PM, Heidemann RM, Nittka M, Jellus V, Wang J, Kiefer B, Haase A. *Magn. Reson. Med.* 2002; 47:1202–1210. [PubMed: 12111967]
16. Sodickson DK, McKenzie CA, Ohliger MA, Yeh EN, Price MD. *MAGMA.* 2002; 13:158–163. [PubMed: 11755091]
17. Pang Y, Vigneron DB, Zhang X. *Magn. Reson. Med.* 2012; 67:965–978. [PubMed: 21858863]

18. Wu B, Wang C, Lu J, Pang Y, Nelson SJ, Vigneron DB, Zhang X. IEEE Trans. Med. Imaging. 2012; 31:183–191. [PubMed: 21878410]
19. Wu B, Zhang X, Wang C, Li Y, Pang Y, Lu J, Xu D, Majumdar S, Nelson SJ, Vigneron DB. Magn. Reson. Med. 2012; 68:1332–1338. [PubMed: 22246803]
20. Roemer PB, Edelstein WA, Hayes CE, Souza SP, Mueller OM. Magn. Reson. Med. 1990; 16:192–335. [PubMed: 2266841]
21. Zhang XZ, Webb A. J. Magn. Reson. 2004; 170:149–155. [PubMed: 15324768]
22. Pinkerton RG, Barberi EA, Menon RS. Magn. Reson. Med. 2005; 54:499–503. [PubMed: 16032690]
23. Wu B, Qu P, Wang C, Yuan J, Shen GX. Magn. Reson. Eng. 2007; 31B:116–126.
24. Wu B, Wang C, Krug R, Kelley DA, Xu D, Pang Y, Banerjee S, Vigneron DB, Nelson SJ, Majumdar S, Zhang X. IEEE Trans. Biomed. Eng. 2010; 57:397–403. [PubMed: 19709956]
25. Wu B, Zhang X, Qu P, Shen GX. J. Magn. Reson. 2006; 182:126–132. [PubMed: 16829145]
26. Wu B, Zhang X, Qu P, Shen GX. Magn. Reson. Imaging. 2007; 25:418–424. [PubMed: 17371734]
27. Zuo, Z., Park, J., Li, Y., Li, Z., Yan, X., Zhang, Z., Zhuo, Y., Cho, ZH., Zhou, XJ., Xue, R. Proceedings of the 20th annual meeting of ISMRM; Melbourne. 2012. p. 2804
28. Yan X, Ma C, Shi L, Zhuo Y, Zhou X, Wei L, Xue R. Appl. Magn. Reson. 2014; 45:437–449.
29. Li Y, Xie Z, Pang Y, Vigneron D, Zhang X. Med. Phys. 2011; 38:4086–4093. [PubMed: 21859008]
30. Xie, Z., Zhang, X. Proceedings of the 16th annual meeting of ISMRM; Toronto. 2008. p. 2972
31. Xie, Z., Zhang, X. Proceedings of the 16th annual meeting of ISMRM; Toronto. 2008. p. 1068
32. Yan, X., Zhang, X., Ma, C., Wei, L., Xue, R. Proceedings of the 22nd annual meeting of ISMRM; Milan. 2014. p. 1309
33. Yan X, Zhang X, Feng B, Ma C, Wei L, Xue R. IEEE Trans. Med. Imaging. 2014 [Epub ahead of print]).
34. Yan, X., Zhang, X., Ma, C., Wei, L., Xue, R. in Proceedings of the 22nd annual meeting of ISMRM; Milan. 2014. p. 1306
35. Yan, X., Zhang, X., Ma, C., Wei, L., Xue, R. in Proceedings of the 22nd annual meeting of ISMRM; Milan. 2014. p. 1322
36. Yan X, Zhang X, Wei L, Xue R. Quant. Imaging Mag. Surg. 2014; 4(2):79–86.
37. Kraus, JD., Marhefka, RJ. Antennas: For All Applications. Third. New York: McGraw-Hill Publishers; 2001.

Fig. 1. Photographs and sketches of the two-channel dipole arrays without decoupling treatments (**a** and **c**) and with magnetic wall decoupling (**b** and **d**)

Fig. 2.
Position of the dipole arrays vs. the water phantom

Fig. 3. Measured S_{11} and S_{21} plots vs. frequency of two dipole elements loaded with the water phantom. **a** Without decoupling treatments. **b** With the MW decoupling method. The isolation between the two dipole elements could be reduced from about -3.6 dB to about -16.5 dB, indicating the strong decoupling capability of the MW decoupling method for the dipole array

Fig. 4.

GRE images from each dipole element of the two-channel array without and with the magnetic wall decoupling. **a** Images from two coupled dipole elements. **b** Profile of the red dotted line of the images in **a**. The S_{21} between the two dipole antennas was about -3.6 dB, which indicates that about half of the signal power was coupled to the other element. The strong coupling between the two dipoles could also be observed from figures **a** and **b**. **c** Images from two MW decoupled dipole elements. **d** Profiles of the red dotted line of the

images in **c**. Figures **c** and **d** show the strong decoupling performance between the two dipoles

Author Manuscript

Author Manuscript

Author Manuscript

Author Manuscript

Fig. 5. Geometry factors of the two-channel dipole arrays without (*left*) and with the MW decoupling method (*right*) at reduction factor $R = 2$. The MW decoupled dipole array has better geometry factor results, indicating the image reconstruction related noise amplifications can be reduced