The Carrel patch clamp for renal transplantation

LAC Radford, TF Kleverlaan, SA Hosgood, ML Nicholson

Department of Surgery, University of Cambridge, Addenbrooke's Hospital, Cambridge, UK

CORRESPONDENCE TO

Michael Nicholson, E: mln31@cam.ac.uk

doi 10.1308/rcsann.2017.0080

BACKGROUND

Normothermic machine perfusion (NMP) is a novel technique for assessing and reconditioning transplant kidneys.¹ NMP involves recirculating a red cell-based perfusate through the kidney at near physiological conditions.^{1–3} The NMP arterial cannula has previously been secured with a ligature at the origin of the renal artery. This damages the aortic patch, which must be excised prior to transplantation. We describe a novel clamp used to preserve the aortic Carrel patch.

TECHNIQUE

The Carrel patch is threaded through one end of the clamp (Fig 1). The clamp is closed (Fig 2) and the screw-threaded pin is tightened, creating a secure seal for attachment to the NMP circuit. The patch clamp also allows preservation of the Carrel patch in kidneys with multiple arteries (Figs 2 and 3). Centralisation of the renal artery within the clamp maximises the seal and size of the patch. Single-use plastic or stainless steel clamps which can be sterilised are available in different sizes.

DISCUSSION

A Carrel patch facilitates anastomosis of the renal artery to the iliac arterial system and reduces vascular complications.⁴ Direct renal artery cannulation for NMP necessitates excision of the aortic patch and carries the risk of creating an intimal flap. The Carrel patch clamp is an innovation that preserves the aortic patch and eliminates the risks of direct cannula damage.

Figure 1 Carrel patch threaded through one half of the patch clamp

Figure 2 Clamp closed around Carrel patch (with multiple arteries) and secured with threaded pin

Figure 3 Clamp in situ during normothermic machine perfusion with multiple arteries secured

ACKNOWLEDGEMENTS

The authors wish to thank the Institut Georges Lopez (Lissieu, France) and Organ Assist (Warsaw, Poland) for producing the Carrel patch clamp. The research was funded by the National Institute for Health Research Blood and Transplant Research Unit (NIHR BTRU) in Organ Donation and Transplantation at the University of Cambridge in collaboration with Newcastle University and in partnership with NHS Blood and Transplant (NHSBT). The views expressed are those of the authors and not necessarily those of the NHS, the NIHR, the Department of Health or NHSBT.

References

1. Hosgood SA, Patel M, Nicholson ML. The conditioning effect of ex vivo normothermic perfusion in an experimental kidney model. *J Surg Res* 2013; **182**(1): 153–160.
2. Hosgood SA, Saeb-Parsy K, Hamed MO *et al.* Successful transplantation of human kidneys deemed untransplantable but resuscitated by ex vivo normothermic machine perfusion. *Am J Transplant* 2016; **16**(11): 3,282–3,285.
3. Hosgood SA, Nicholson ML. First in man renal transplantation after ex vivo normothermic perfusion. *Transplantation* 2011; **92**(7): 735–738.
4. Bewick M, Ogg CS, Parsons V *et al.* The arterial complications of 500 renal transplants. *Proc Eur Dialysis Transplant Assoc* 1976; **13**: 186–190.