HHS Public Access

Author manuscript

J Glaucoma. Author manuscript; available in PMC 2019 February 01.

Published in final edited form as:

J Glaucoma. 2018 February ; 27(2): e50. doi:10.1097/IJG.0000000000000844.

Response to Letter to Editor

Tony Realini, MD, MPH¹, Matthew J. Gurka, PhD², and Robert N Weinreb, MD³

¹West Virginia University Eye Institute, Morgantown, WV

²Institute for Child Health Policy and Department of Health Outcomes and Policy, University of Florida, Gainesville, FL

³Hamilton Glaucoma Center, Shiley Eye Institute and Department of Ophthalmology at University of California-San Diego, La Jolla, CA

Dear Editor

We appreciate Drs. De Bernado and Rosa careful reading of our recent study on the reproducibility of central corneal thickness measurements obtained over time using ultrasound pachymetry. We agree with all of their excellent observations. For completeness, we used a solution of fluorescein sodium and benoxinate for all patients at all time points. Immediately after instillation, we performed Goldmann applanation tonometry and then pachymetry, in that order.