Images in Cardiovascular Medicine

OPEN ACCESS

Received: Oct 1, 2017
Accepted: Jan 2, 2018

Correspondence to

Mizuki Miura, MD

Department of Cardiology, Kokura Memorial Hospital, 3-2-1, Asano, Kokurakita-ku, Kitakyushu 802-8555, Japan.
 E-mail: mizumiura-circ@umin.ac.jp

Copyright © 2018. The Korean Society of Cardiology

This is an Open Access article distributed under the terms of the Creative Commons Attribution Non-Commercial License (<https://creativecommons.org/licenses/by-nc/4.0>) which permits unrestricted noncommercial use, distribution, and reproduction in any medium, provided the original work is properly cited.

ORCID iDs

Mizuki Miura
<https://orcid.org/0000-0003-3252-3861>

Conflict of Interest

The authors have no financial conflicts of interest.

Author Contributions

Conceptualization: Miura M, Shirai S, Hayashi M, Jinnouchi H, Isotani A, Kakumoto S, Arai Y, Ando K; Data curation: Miura M; Investigation: Miura M; Project administration: Miura M; Supervision: Shirai S, Ando K; Writing - original draft: Miura M.

Atypical Annulus Rupture after Transcatheter Aortic Valve Implantation

Mizuki Miura , MD¹, Shinichi Shirai, MD¹, Masaomi Hayashi, MD¹, Hiroyuki Jinnouchi, MD¹, Akihiro Isotani, MD¹, Shinichi Kakumoto, MD², Yoshio Arai, MD³, and Kenji Ando, MD¹

¹Department of Cardiology, Kokura Memorial Hospital, Kitakyushu, Japan

²Department of Anesthesiology, Kokura Memorial Hospital, Kitakyushu, Japan

³Department of Cardiovascular Surgery, Kokura Memorial Hospital, Kitakyushu, Japan

An 86-year-old female underwent transcatheter aortic valve implantation (TAVI) for severe aortic stenosis (AS) via a transfemoral approach. The aortic valve was a severely calcified bicuspid valve with fusion of the right and non-coronary cusps (**Figure 1A**). Therefore, a

Figure 1. (A) Multislice computed tomography showing a severely calcified aortic valve with fusion of the right and non-coronary cusps. (B) A 26-mm SAPIEN XT (Edwards Lifesciences, Irvine, CA, USA) deployed via a transfemoral approach. (C) Transthoracic echocardiography showing atypical annulus rupture and a continuous shunt from the Ao to RV (arrow). (D) Aortography showing blood flow from the aortic annulus to RV (arrow). Ao = aorta; LA = left atrial; LV = left ventricle; RV = right ventricle.

SAPIEN XT 26-mm valve (Edwards Lifesciences, Irvine, CA, USA) was implanted using a 1-mL-under balloon volume (**Figure 1B**). Transthoracic echocardiography (TTE) showed no problem one day after TAVI, but predischARGE TTE at 1 week revealed a continuous shunt from the aorta to right ventricle (**Figure 1C, Supplementary Video 1**) and the pulmonary blood flow (Q_p)/systemic blood flow (Q_s) was 1.1. Aortography showed blood flow from the aortic annulus to the right ventricle (**Figure 1D, Supplementary Video 2**). She had no symptoms and was discharged on foot at 19 days after TAVI. At 18 months after TAVI, she was doing well without heart failure.

The silent onset of annulus rupture is very rare. Most annulus ruptures are observed during the initial procedure, and the patient's hemodynamic status is extremely unstable.¹⁾ In our case, the patient was very stable and the complication was late-onset. The mechanism of this complication may be erosion between the implanted valve edge and the aortic wall.²⁾ Especially when there is a bicuspid valve with fusion of the right and non-coronary cusp, we need to carefully think about the type of implanted valve and the method of implantation. In addition, we need to carefully observe the patient for iatrogenic shunt, and percutaneous or surgical closure may be needed if the condition of the patient gets worse.

SUPPLEMENTARY MATERIALS

Supplementary Video 1

Transthoracic echocardiography showing atypical annulus rupture and a continuous shunt from the aorta to right ventricle.

[Click here to view](#)

Supplementary Video 2

Aortography showing blood flow from the aortic annulus to right ventricle.

[Click here to view](#)

REFERENCES

1. Hayashida K, Bouvier E, Lefèvre T, et al. Potential mechanism of annulus rupture during transcatheter aortic valve implantation. *Catheter Cardiovasc Interv* 2013;82:E742-6.
[PUBMED](#) | [CROSSREF](#)
2. Hagiwara K, Saito N, Yamazaki K, Kimura T. Aorto-right ventricular fistula following transcatheter aortic valve implantation using a 29 mm SAPIEN XT valve. *BMJ Case Rep* 2017;2017:bcr-2017-219247.
[PUBMED](#) | [CROSSREF](#)