Received:
08 July 2017

Revised:
17 October 2017

Accepted:
26 October 2017

<https://doi.org/bjr.20170511>

Cite this article as:

Zeinali-Rafsanjani B, Faghihi R, Mosleh-Shirazi MA, Saeedi-Moghadam M, Jalli R, Sina S. Effect of age-dependent bone electron density on the calculated dose distribution from kilovoltage and megavoltage photon and electron radiotherapy in paediatric MRI-only treatment planning. *Br J Radiol* 2018; **91**: 20170511.

FULL PAPER

Effect of age-dependent bone electron density on the calculated dose distribution from kilovoltage and megavoltage photon and electron radiotherapy in paediatric MRI-only treatment planning

¹B ZEINALI-RAFSANJANI, PhD, ^{1,2}R FAGHIHI, PhD, ^{3,4}MA MOSLEH-SHIRAZI, PhD, ³M SAEEDI-MOGHADAM, ³R JALLI, MD and ²S SINA, PhD

¹Department of Nuclear Engineering, School of Mechanical Engineering, Shiraz University, Shiraz, Iran

²Radiation Research Center, School of Mechanical Engineering, Shiraz University, Shiraz, Iran

³Medical Imaging Research Center, Shiraz University of Medical Sciences, Shiraz, Iran

⁴Department of Radiotherapy and Oncology, Physics Unit, Shiraz University of Medical Sciences, Shiraz, Iran

Address correspondence to: Dr MA Mosleh-Shirazi

E-mail: mosleh_amin@hotmail.com

Objective: MRI-only treatment planning (TP) can be advantageous in paediatric radiotherapy. However, electron density extraction is necessary for dose calculation. Normally, after bone segmentation, a bulk density is assigned. However, the variation of bone bulk density in patients makes the creation of pseudo CTs challenging. This study aims to assess the effects of bone density variations in children on radiation attenuation and dose calculation for MRI-only TP.

Methods: Bone contents of <15-year-old children were calculated, and substituted in the Oak Ridge National Laboratory paediatric phantoms. The percentage depth dose and beam profile of 150 kVp and 6 MV photon and 6 MeV electron beams were then calculated using Xcom, MCNPX (Monte Carlo N-particle version X) and ORLN phantoms.

Results: Using 150 kVp X-rays, the difference in attenuation coefficient was almost 5% between an 11-year-old

child and a newborn, and 8% between an adult and a newborn. With megavoltage radiation, the differences were smaller but still important. For an 18 MV photon beam, the difference of radiation attenuation between an 11-year-old child and a newborn was 4% and 7.4% between an adult and a newborn. For 6 MeV electrons, dose differences were observed up to the 2 cm depth. The percentage depth dose difference between 1 and 10-year-olds was 18.5%, and between 10 and 15-year-olds was 24%.

Conclusion: The results suggest that for MRI-only TP of photon- or electron-beam radiotherapy, the bone densities of each age group should be defined separately for accurate dose calculation.

Advances in knowledge: This study highlights the need for more age-specific determination of bone electron density for accurate dose calculations in paediatric MRI-only radiotherapy TP.

INTRODUCTION

The use of automatic computing technology for radiation dose calculations dates back to 1955.¹ MRI was introduced into treatment planning (TP) in the 1990s.^{2–6} Today, TP has evolved towards a more precise target delineation using the advantage of better soft tissue contrast of MRI images. Moreover, more innovative approaches are sought to have the ability to define functional targets, especially in brain radiation therapy by using diffusion/perfusion weighted MRI and magnetic resonance spectroscopy (MRS).^{7–9}

Despite some initial problems, the use of MRI-only TP has been reported and is expected to become increasingly utilized.^{10,11} Electron density extracted information, which is required for dose calculation was one of the problems that MRI-only TP was encountered with, which was addressed by different atlas-based,¹² voxel-based¹³ or bulk density methods.¹⁴ Each method has its own deficiency, e.g. in the bulk density method, manual segmentation and classification of tissues is a tedious and time-consuming task. The atlas-based method has some errors regarding the differences between anatomies; this problem will be more significant in the patients with abnormal anatomies. The voxel-based method solved this geometric problem

Figure 1. The water and mineral content estimated by Woodard *et al* and Guy *et al* studies.^{20,21} The results of each study have numbered as 1 and 2.

by classifying each voxel of the MR image to a category of tissues. Despite the improvements offered by voxel-based TP, this method has some difficulties in distinguishing the air/bone interfaces. The routine MRI pulse sequences cannot provide a useful signal from bones, since the T_2 relaxation time of bone is very short. This problem was also solved by designing ultra-short echo time pulse sequences, which are able to create a suitable signal from bone.^{15,16}

Although most of the problems have now been resolved, there is still an important issue that needs to be addressed. When the bones are segmented using manual/automatic techniques, a bulk density is assigned to them. However, the variation of bone bulk density in patients may be a challenge that can have some influence on the accuracy of the created pseudo CTs.¹⁷

Having an optimum TP for those who have good prognosis and high life expectancy, such as children with benign or low-grade lesions, is particularly important. The above-mentioned problem regarding variation of bone electron density can be worse in patients under the age of 18, since the structure of bone varies in this age range.^{18,19}

Some studies have estimated the mineral and water content of bones in children. Figure 1 shows the water and mineral content of children's cortical bone excluding proteins and collagen.^{20,21}

According to Figure 1, the water content of children's bone decreases, but the mineral content of the body increases during their growth which in turn increases the bone mass and electron density as an important factor in radiotherapy TP.

This study aims to assess the effects of bone density variations on the radiation attenuation and dose calculation in paediatric MRI-only TP. In order to evaluate this difference, the bone content of children was calculated from previous studies to use in ORNL paediatric phantoms as bone material. The percentage depth dose (PDD) and beam profile curves of 150 kVp superficial/orthovoltage and 6 MV Varian linear accelerator photon beams, plus a 6 MeV electron beam, were calculated using the MCNPX Monte Carlo code which is a popular code in dose calculation.^{22–28}

Figure 2. Bone materials used in paediatric phantoms for different ages.

Elements	Newborn	1 y	5, 10 and 15 y
H	8.00E-02	0.07337	7.34E-02
C	9.71E-02	2.55E-01	2.55E-01
N	2.71E-02	3.06E-02	3.06E-02
O	6.68E-01	4.79E-01	4.79E-01
F	----	----	2.50E-04
Na	3.14E-03	3.26E-03	3.26E-03
Mg	1.43E-03	----	1.12E-03
Si	----	2.00E-05	2.00E-05
P	3.71E-02	5.10E-02	5.10E-02
S	3.14E-03	1.73E-03	1.73E-03
Cl	1.40E-03	1.43E-03	1.43E-03
K	1.48E-03	1.53E-03	1.53E-03
Ca	8.00E-02	1.02E-01	1.02E-01
Fe	8.00E-05	8.00E-05	8.00E-05
Zn	----	----	5.00E-05
Rb	----	----	2.00E-05
Sr	----	----	3.00E-05
Pb	----	----	1.00E-05

METHODS AND MATERIALS

The influence of variations in bone electron density in radiotherapy TP was calculated using the MCNPX 2.5 Monte Carlo code. The set of ORNL MIRD paediatric phantom including 1-, 10- and 15-year-old were used in this study.²⁹ These mathematical anatomical phantoms were consisted of skull, chest, lower extremities and some internal organs. They were developed using anatomical references in order to have a standard organ size, shapes and weights regarding to different ages.³⁰

The bone content of these phantoms is shown in Figure 2. Since the bone material content in 5-, 10- and 15-year-old phantoms were the same, we decided to find more differentiated bone contents from the previous studies.

The most important bone mineral is hydroxyapatite [$\text{Ca}_{10}(\text{PO}_4)_6(\text{OH})_2$].^{31,32} There is also some collagen in the bones³³ with the chemical formulation of $\text{C}_2\text{H}_5\text{NOC}_5\text{H}_9\text{NOC}_5\text{H}_{10}\text{NO}_2$.³⁴ The most important bone elements are, therefore, hydrogen, oxygen, nitrogen, carbon, calcium and phosphorus.

The bone components of a newborn, 2–17 days, 2–4 months, 5–9 months, 12–24 months and 11- to 12-year-old children and adults (+18) were calculated using the information from Dickerson, LeGeros *et al* and Pacureanu *et al*, studies.^{32–34} The X-ray attenuation of estimated bone content of different age

Figure 3. The 6 MV X-ray spectrum used in this study (right). The 150 kVp X-ray spectrum used in this study (left).

groups were assessed in photon energies ranging from 0 to 18 MV using Xcom (NIST, USA), an online software of the National Institute of Standards and Technology.³⁵ The uncertainty in the calculation of mass attenuation coefficient of Xcom is almost 0.7%.³⁶

The bone contents of 1-, 10- and 15-year-old children were also included in the ORNL phantoms. The brain material was simulated using 16 elements and density of 1.04 g cm^{-3} . In order to assess the photon attenuation differences regarding different bone components, the energy spectra of superficial/orthovoltage (150 kVp) and linac (6 MV) X-rays were considered to calculate the PDD and beam dose profiles. The 150 kVp beam spectrum, with 3 mm Al + 8 mm Cu filtration, was calculated using a spectrum processor [Report 78 spectrum processor, IPEM (Institute of Physics and Engineering in Medicine, UK)]. The 6 MV X-ray beam of a Varian linear accelerator was used from a previous study³⁷ (Figure 3). In addition to photon beams, the delivered dose distribution of a monoenergetic 6 MeV electron beam was also evaluated. The aforementioned radiation beams were verified by comparing the calculated photon spectra with the extracted data from the spectrum processor, as well as the previous studies. The half-value layer (HVL) of the 150 kVp photon beam was also calculated and compared with the HVL produced by the spectrum processor, and the accuracy of the simulated beam was verified. According to the spectrum processor information, the HVL of this X-ray beam was 16.45 mm Al. Using the MCNP code, the mentioned beam was simulated and the output flux of the beam without and with an Al attenuator with the thickness of 16.45 mm was calculated. The ratio of the flux of open beam (without attenuator) and attenuated beam (using Al attenuator) was 0.51 which shows that the HVL of simulated beam was almost 16.45 mm Al.

To simulate a radiation source, a cone source with the apex angle of 24° at a distance of 50 cm from the skull phantom was considered to produce a circular field of diameter 10.62 cm on the patient's skin. Designing a collimated source such as a cone beam source results in a shorter code runtime, while simulating a reasonable simplification of the real clinical situations.

Mesh tally Type 3 with the keyword of "Total" was used. This tally type calculates the total dose deposited via all types of radiations in the body volume. The dimension of each mesh volume was considered as $0.2 \times 0.2 \times 0.5 \text{ cm}^3$. The codes were run until the relative error of all data points were reduced to less than 2%. This error is as a result of the stochastic nature of the Monte Carlo calculation method. The position of source and phantom

Figure 4. The position of source and phantom and the level of determined percentage depth dose and beam profile.

and the level of determined PDD and beam profile was shown in Figure 4.

RESULTS

The bone components of newborn, 2–17 days, 2–4 months, 5–9 months, 12–24 months, 11- to 12-year-old children and adults (+18) calculated using the previous studies are shown in Figure 5.

The photon attenuations of bone in different age groups, calculated using Xcom, in the X-ray energy range of 1–150 and 0.1–18 MeV are shown in Figures 6 and 7.

The PDD and beam profiles calculated in the ORNL phantoms are shown in Figures 8 and 9.

The calculated PDDs for the 6 MeV electron beam are shown in Figure 10.

DISCUSSION

The MRI-only TP takes advantage of its superior soft tissue contrast to improve tumour delineation. This modality is also useful for multiple adaptive TP, since it does not deliver additional radiation dose to the patient. This issue is specifically important for children and young patients as well as those with benign or low-grade lesions.³⁸ For CT-based TP, the computerized planning systems normally require a conversion table to be entered at the commissioning stage to obtain the required map of mass densities or electron densities the matrix of Hounsfield Units forming the image. The choice of the conversion table affects the dose distribution to various extents depending on the treated site.^{5,6} For MRI-only TP, to build the electron density map of the tissues, often referred to as pseudo CT or

Figure 5. Children's bone component including collagen.

Figure 6. Calculated photon mass attenuation coefficients of the bones of children in different age groups in the X-ray energy range of 1–150 keV (top). The differences between the bones mass attenuation coefficients of different age groups from an adult (bottom).

synthetic CT images, usually the segmented tissues are categorized into various classes and a bulk density is assigned to each of them. In children, whose bones develop rapidly and bone

Figure 7. Calculated photon mass attenuation coefficients of the bones of children in different age groups in the X-ray energy range of 100 keV–18 MeV (top). The inset shows the differences between the bone mass attenuation coefficients of different age groups from an adult.

electron density gradually changes, errors may arise from using this method.

This study aimed to assess the influence of the bone electron density variations in children's bones to highlight the need for a solution to decrease or avoid their effects.

It should be noted that one of the limitations we had was a shortage of the references with information on skull bone components. To the best of our knowledge, there was no reference which provided skull bone components, their weight fractions and their differences during growth. So, we had to use a combination of references^{31–34} to arrive at a reasonable estimation of bone components. Dickerson³³ presented detailed information on bone material of the compact, cancellous and epiphyseal parts of the femur bone. In children's skull bone, all of these components also exist as compact, cancellous and fontanelles.³⁹ In our study, owing to unavailability of data, we had to make an assumption that the chemical compositions of the different parts of the skull bone (compact and spongy bone and fontanelles) vary with age in the same way that occurs in the femur. This is aided by the fact that the bone was defined as a whole in the MCNP input file, as opposed to specifying the individual components (*i.e.* compact and spongy bone and fontanelles). The observed trend is also as expected.

First, the bone components in each age group were extracted from previous studies. The extracted bone component (Figure 5) was considered as cortical plus spongy bone including collagen, which exists in the bone content. As a result, the amount of H and O, shown as H₂O in Figure 5, is more than the estimated amount reported by Guy²¹ and Woodard in Figure 1. It should be emphasized that they had estimated the component of cortical bone, and had excluded the spongy bone and collagen.

According to Figure 5, the water content of bone gradually decreases, but the mineral contents such as Ca and P increases. This changing trend is in agreement with previous studies.²⁰ According to the studies by LeGros *et al* and Dickerson,³³ calcium and phosphorus contribute highly in the main bone formula and structure. The amount of carbon which exist plentifully in collagen also increases by age. The amount of nitrogen also increases, but not as much as the increase by carbon, since there are only 3 nitrogen atoms in each molecule of collagen, compared with 12 carbon atoms.

These components had to be substituted as the bone material of ORNL MIRD phantoms, since the data for these phantoms consider the same bone component for the age range 5- to 15-years-old to simplify simulations. This simplification may provide a good result in most cases, but there is evidence that some small changes occur, which could be important in some especial circumstances.

There is a tradeoff between the relative error and calculation resolution, the smaller voxel sizes the larger relative error. Therefore, in order to have a smaller relative error, a sufficiently large

Figure 8. Percentage depth doses of 150 kVp and 6 MV X-ray beams in different age ORNL phantoms.

voxel size was defined in the mesh while providing adequate calculation resolution.

The simulated photon and electron energies are clinically used in paediatric radiotherapy. The superficial/orthovoltage therapy can be used in the treatment of infant haemangioma.⁴⁰ The skin or other superficial lesions may be treated using either orthovoltage or electron beams.⁴¹ Moreover, orthovoltage photon beam may also be used to deliver high doses to the bones, and the electron therapy can be useful to deliver the boost doses to the tumour bed in intraoperative radiation therapy.⁴² Furthermore, the 6 MeV electron beam is an option for paediatric total skin irradiation.⁴³ Also, the megavoltage photon beam is a choice to treat paediatric rhabdomyosarcoma tumours.⁴⁴

The 150 kVp X-ray spectrum which was used in this study was a heavily filtered beam to avoid much attenuation in the skin and superficial structures (The filtration details had been provided in the text). According to the spectrum processor information and MCNP calculations, the HVL of this X-ray beam was 16.45 mm Al and the mean energy was 116 keV. This is in contrast

to the standard 150 kVp beam with a mean energy of about 50–60 kVp.

The Xcom mass attenuation coefficient calculation results (Figure 7) show that differences in bone component can result in significant differences in attenuation of photons at energies below 100 keV. The difference can even exceed 50%. These results mean that CT scans for TP or treatment verification with 120 kVp X-rays (mean energy of 40–60 keV) results in 29–17% difference between the radiation absorption of bones in different ages. This result implies that substituting CT with MRI in adaptive radiotherapy and/or daily treatment verification can obviate the need for this additional bone radiation dose resulting from CT scans X-ray. This issue can be important in the older children whose bone contains more minerals such as calcium, since it increases the probability of the photoelectric phenomenon. Although the energies lower than 40 keV and higher than 12 MeV are not used clinically, but they were included to show the effect in the lower energies merely from a physics point of view.

Using a filtered 150 kVp beam with the mean energy of 116 keV, the difference in mass attenuation coefficient between

Figure 10. Calculated percentage depth doses of the 6 MeV electron beam.

an 11-year-old and a newborn child was almost 5%. The difference between an adult and a newborn was approximately 8%.

In the megavoltage photon energy range, although the differences are smaller, they still may be important in high energy treatments. For example, in an 18 MeV photon beam, the difference in attenuation between an 11-year-old child and a newborn was 4%. This difference between an adult and a newborn was approximately 7.4%. The results may imply that it is better to take into account the variations in bone components between adults and children even for megavoltage photon energies.

For further evaluation of the dose distributions, the PDDs and beam profiles of 150 kVp and 6 MV photon beams were calculated in different ORNL phantom age groups. The difference between the bone components of a 1-year-old child and 10–15-year-old children resulted in an almost 5% difference in beam profile with the megavoltage beam, although there was less difference between 10- to 15-year-old children. The differences could be greater in kilovoltage X-ray therapy, simply owing to the influence of the photoelectric effect. As an example, the PDD immediately beneath the skull bone in 1, 10 and 15-year-old children were 53, 45, and 42%, respectively. The difference between the PDD of the 1- and 10-year-old children was 15%, and between the 10 and 15-year-old children was almost 5%.

The beam profiles also showed a non-significant difference with the 6 MeV photon beam, but greater observed differences in 150 kVp X-ray beam.

Using the 6 MeV electron beam, the dose differences can be observable up to 1.5 cm beneath the surface of the skull. The PDD of the 1-, 10- and 15-year-old children was 92, 75 and 57%, respectively. The PDDs difference between the 1- and 10-year-old child was 18.5% and the difference between the 10- and 15-year-old children was 24%.

As mentioned before, routine MRI sequences cannot produce a suitable signal from the bone and this pitfall results in some uncertainty in dose calculation in bone and the surrounding tissues. More recently, MRS has been introduced as an adjunct tool for TP. This modality can produce a spectrum from the body metabolites. Some studies have emphasized the potential role of MRS in TP.³¹ MRS may provide the ability to obtain the bone content, which provides a step closer to finding bone electron density. If 3D MRS becomes practical, the need to perform a CT scan diminishes, which also removes the uncertainties owing to image registration.

CONCLUSION

The results suggest that for MRI-only TP of photon- or electron-beam radiotherapy, the bone densities of each age group should be defined separately for accurate dose calculation in paediatric radiotherapy. MRS is also suggested as a possible solution.

REFERENCES

1. Tsien KC. The application of automatic computing machines to radiation treatment planning. *Br J Radiol* 1955; **28**: 432–9. doi: <https://doi.org/10.1259/0007-1285-28-332-432>
2. Ten Haken RK, Thornton AF, Sandler HM, La Vigne ML, Quint DJ, Fraass BA, et al. A quantitative assessment of the addition of MRI to CT-based, 3-D treatment planning of brain tumors. *Radiother Oncol* 1992; **25**: 121–33. doi: [https://doi.org/10.1016/0167-8140\(92\)90018-P](https://doi.org/10.1016/0167-8140(92)90018-P)
3. Thornton AF, Sandler HM, Ten Haken RK, McShan DL, Fraass BA, La Vigne ML, et al. The clinical utility of magnetic resonance imaging in 3-dimensional treatment planning of brain neoplasms. *Int J Radiat Oncol Biol Phys* 1992; **24**: 767–75. doi: [https://doi.org/10.1016/0360-3016\(92\)90727-Y](https://doi.org/10.1016/0360-3016(92)90727-Y)
4. Phillips MH, Kessler M, Chuang FY, Frankel KA, Lyman JT, Fabrikant JJ, et al. Image correlation of MRI and CT in treatment planning for radiosurgery of intracranial vascular malformations. *Int J Radiat Oncol Biol Phys* 1991; **20**: 881–9. doi: [https://doi.org/10.1016/0360-3016\(91\)90036-4](https://doi.org/10.1016/0360-3016(91)90036-4)
5. Mosleh-Shirazi MA, Hansen VN, Childs PJ, Warrington AP, Saran FH. Commissioning and implementation of a stereotactic conformal radiotherapy technique using a general-purpose planning system. *J Appl Clin Med Phys* 2004; **5**: 1–14. doi: <https://doi.org/10.1120/jacmp.v5i3.1948>
6. Bedford JL, Childs PJ, Nordmark Hansen V, Mosleh-Shirazi MA, Verhaegen F, Warrington AP. Commissioning and quality assurance of the Pinnacle³ radiotherapy treatment planning system for external beam photons. *Br J Radiol* 2003; **76**: 163–76. doi: <https://doi.org/10.1259/bjr/42085182>
7. Aramburu Núñez D, Lopez Medina A, Mera Iglesias M, Salvador Gomez F, Dave A, Hatzoglou V, et al. Multimodality functional imaging using DW-MRI and ¹⁸F-FDG-PET/CT during radiation therapy for human papillomavirus negative head and neck squamous cell carcinoma: Meixoeiro hospital of Vigo experience. *World J Radiol* 2017; **9**: 17–26. doi: <https://doi.org/10.4329/wjr.v9.i1.17>
8. Broggi S, Scalco E, Belli ML, Logghe G, Verellen D, Moriconi S, et al. A comparative evaluation of 3 different free-form deformable image registration and contour propagation methods for head and neck MRI: the case of parotid changes during radiotherapy. *Technol Cancer Res Treat* 2017; **16**: 373–81. doi: <https://doi.org/10.1177/1533034617691408>
9. Snelling B, Shah AH, Buttrick S, Benveniste R. The use of MR perfusion imaging in the evaluation of tumor progression in gliomas. *J Korean Neurosurg Soc* 2017; **60**: 15–20. doi: <https://doi.org/10.3340/jkns.2016.0102.001>
10. Kapanen M, Tenhunen M. T1/T2*-weighted MRI provides clinically relevant pseudo-CT density data for the pelvic bones in MRI-only based radiotherapy treatment planning. *Acta Oncol* 2013; **52**: 612–8. doi: <https://doi.org/10.3109/0284186X.2012.692883>
11. Kapanen M, Collan J, Beule A, Seppälä T, Saarialhti K, Tenhunen M. Commissioning

- of MRI-only based treatment planning procedure for external beam radiotherapy of prostate. *Magn Reson Med* 2013; **70**: 127–35. doi: <https://doi.org/10.1002/mrm.24459>
12. Dowling JA, Lambert J, Parker J, Salvado O, Fripp J, Capp A, et al. An atlas-based electron density mapping method for magnetic resonance imaging (MRI)-alone treatment planning and adaptive MRI-based prostate radiation therapy. *Int J Radiat Oncol Biol Phys* 2012; **83**: e5–e11. doi: <https://doi.org/10.1016/j.ijrobp.2011.11.056>
 13. Catana C, van der Kouwe A, Benner T, Michel CJ, Hamm M, Fenchel M, et al. Toward implementing an MRI-based PET attenuation-correction method for neurologic studies on the MR-PET brain prototype. *J Nucl Med* 2010; **51**: 1431–8. doi: <https://doi.org/10.2967/jnumed.109.069112>
 14. Jonsson JH, Johansson A, Söderström K, Askund T, Nyholm T. Treatment planning of intracranial targets on MRI derived substitute CT data. *Radiother Oncol* 2013; **108**: 118–22. doi: <https://doi.org/10.1016/j.radonc.2013.04.028>
 15. Robson MD, Bydder GM. Clinical ultrashort echo time imaging of bone and other connective tissues. *NMR Biomed* 2006; **19**: 765–80. doi: <https://doi.org/10.1002/nbm.1100>
 16. Robson MD, Gatehouse PD, Bydder M, Bydder GM. Magnetic resonance: an introduction to ultrashort TE (UTE) imaging. *J Comput Assist Tomogr* 2003; **27**: 825–46. doi: <https://doi.org/10.1097/00004728-200311000-00001>
 17. Uh J, Merchant TE, Li Y, Li X, Hua C. MRI-based treatment planning with pseudo CT generated through atlas registration. *Med Phys* 2014; **41**: 051711. doi: <https://doi.org/10.1118/1.4873315>
 18. Ahmad MA. *Bone health and body composition of children and adolescents with growth hormone deficiency*. Glasgow, UK: University of Glasgow; 2017.
 19. Boot AM, de Ridder MA, Pols HA, Krenning EP, de Muinck Keizer-Schrama SM. Bone mineral density in children and adolescents: relation to puberty, calcium intake, and physical activity. *J Clin Endocrinol Metab* 1997; **82**: 57–62. doi: <https://doi.org/10.1210/jc.82.1.57>
 20. Woodard HQ, White DR. The composition of body tissues. *Br J Radiol* 1986; **59**: 1209–18. doi: <https://doi.org/10.1259/0007-1285-59-708-1209>
 21. Guy H, Masset C, Baud C-A. Infant taphonomy. *Int J Osteoarchaeol* 1997; **7**: 221–9. doi: [https://doi.org/10.1002/\(SICI\)1099-1212\(199705\)7:3<221::AID-OA338>3.0.CO;2-Z](https://doi.org/10.1002/(SICI)1099-1212(199705)7:3<221::AID-OA338>3.0.CO;2-Z)
 22. Zeinali-Rafsanjani B, Mosleh-Shirazi MA, Faghihi R, Karbasi S, Mosalaei A. Fast and accurate Monte Carlo modeling of a kilovoltage X-ray therapy unit using a photon-source approximation for treatment planning in complex media. *J Med Phys* 2015; **40**: 74. doi: <https://doi.org/10.4103/0971-6203.158676>
 23. Sina S, Saeedimoghadam M, Zeinali B, Kazempour M, Jalli R. A Monte Carlo study of new concrete shielding materials containing Galena and borated minerals for kilo-electron-volt photon energies. *Iran J Med Phys* 2017; **14**.
 24. Hadad K, Saeedi-Moghadam M, Zeinali-Rafsanjani B. Voxel dosimetry: comparison of MCNPX and DOSXYZnrc Monte Carlo codes in patient specific phantom calculations. *Technol Health Care* 2017; **25**: 29–35. doi: <https://doi.org/10.3233/THC-161240>
 25. Faghihi R, Zeinali-Rafsanjani B, Mosleh-Shirazi M, Mosalaei A, Hadad K. Revision of orthovoltage chest wall treatment by Monte Carlo simulations. *Technology and Health Care* 2016; **25**: 1–12.
 26. Zeinali-Rafsanjani B, Mosleh-Shirazi MA, Haghighatafshar M, Jalli R, Saeedi-Moghadam M. Assessment of the dose distribution of Minibeam radiotherapy for lung tumors in an anthropomorphic phantom: a feasibility study. *Technology and Health Care* 2016; **25**: 683–92. doi: <https://doi.org/10.3233/THC-170818>
 27. Kazempour M, Saeedimoghadam M, Shekoohi Shooli F, Shokrpour N. Assessment of the radiation attenuation properties of several lead free composites by Monte Carlo simulation. *J Biomed Phys Eng* 2015; **5**: 67–76.
 28. Tayebi M, Shooli FS, Saeedi-Moghadam M. Evaluation of the scattered radiations of lead and lead-free aprons in diagnostic radiology by MCNPX. *Technol Health Care* 2017; **25**: 513–20. doi: <https://doi.org/10.3233/THC-161293>
 29. Krstić D, Nikezić D, Krstić D. Input files with ORNL—mathematical phantoms of the human body for MCNP-4B. *Comput Phys Commun* 2007; **176**: 33–7. doi: <https://doi.org/10.1016/j.cpc.2006.06.016>
 30. Shoup R, Hwang J, Poston J, Warner G. *Pediatric phantoms for use in dosimetric calculations*. Tennessee, USA: Oak Ridge National Lab; 1976.
 31. Bogdanoviciene I, Beganskiene A, Tõnsuaadu K, Glaser J, Meyer H-J, Kareiva A. Calcium hydroxyapatite, Ca₁₀(PO₄)₆(OH)₂ ceramics prepared by aqueous sol-gel processing. *Mater Res Bull* 2006; **41**: 1754–62. doi: <https://doi.org/10.1016/j.materresbull.2006.02.016>
 32. LeGeros RZ, LeGeros JP. *Calcium phosphate bioceramics: past, present and future*. Key Engineering Materials: Trans Tech Publ; 2003.
 33. Dickerson JW. Changes in the composition of the human femur during growth. *Biochem J* 1962; **82**: 56–61. doi: <https://doi.org/10.1042/bj0820056>
 34. Langer M, Cloetens P, Hesse B, Suhonen H, Pacureanu A, Raum K, et al. Priors for X-ray in-line phase tomography of heterogeneous objects. *Philos Trans A Math Phys Eng Sci* 2014; **372**: 20130129. doi: <https://doi.org/10.1098/rsta.2013.0129>
 35. Berger MJ, Hubbell JH, Seltzer SM, Chang J, Coursey JS, Sukumar R. *XCOM: Photon Cross Section Database*; 2010.
 36. Ali ES, Spencer B, McEwen MR, Rogers DW. Towards a quantitative, measurement-based estimate of the uncertainty in photon mass attenuation coefficients at radiation therapy energies. *Phys Med Biol* 2015; **60**: 1641–54. doi: <https://doi.org/10.1088/0031-9155/60/4/1641>
 37. Ali ES, Rogers DW. Functional forms for photon spectra of clinical linacs. *Phys Med Biol* 2012; **57**: 31–50. doi: <https://doi.org/10.1088/0031-9155/57/1/31>
 38. Mosleh-Shirazi MA. Radiotherapy techniques for intracranial tumours. *J Radiat Res* 2006; **4**: 53–62.
 39. Rogers K. *Bone and Muscle: Structure, Force, and Motion*: The Rosen Publishing Group; 2010.
 40. Ogino I, Torikai K, Kobayasi S, Aida N, Hata M, Kigasawa H. Radiation therapy for life- or function-threatening infant hemangioma. *Radiology* 2001; **218**: 834–9. doi: <https://doi.org/10.1148/radiology.218.3.r01mr04834>
 41. Zeinali Rafsanjani B, Mosleh-Shirazi MA, Faghihi R, Mosalaei A, Omidvari S, Hadad K, et al. Breast cancer and its radiotherapeutic methods. *Iranian Journal of Medical Physics* 2012; **9**: 75–85.
 42. Halperin EC, Constine LS, Tarbell NJ, Kun LE. *Pediatric Radiation Oncology*: Lippincott Williams & Wilkins; 2012.
 43. Bao Q, Hryckushko BA, Dugas JP, Hager FH, Solberg TD. A technique for pediatric total skin electron irradiation. *Radiat Oncol* 2012; **7**: 40. doi: <https://doi.org/10.1186/1748-717X-7-40>
 44. Osman AFI, Kafi ST, Idris OF. The effect of different single doses from X-rays on pediatrics rhabdomyosarcoma tumor cells: in vitro survival curve. *Int J Cancer Ther Oncol* 2016; **2**: 1–4.