Published in final edited form as:

J Struct Biol. 2018 October ; 204(1): 38–44. doi:10.1016/j.jsb.2018.06.007.

Routine Determination of Ice Thickness for Cryo-EM Grids

William J. Rice¹, Anchi Cheng¹, Alex J. Noble¹, Edward T. Eng¹, Laura Y. Kim¹, Bridget Carragher^{*,1,2}, and Clinton S. Potter^{1,2}

¹National Resource for Automated Molecular Microscopy, Simons Electron Microscopy Center, New York Structural Biology Center, New York, United States

²Department of Biochemistry and Molecular Biophysics, Columbia University, New York, NY 10032, USA

Abstract

Recent advances in instrumentation and automation have made cryo-EM a popular method for producing near-atomic resolution structures of a variety of proteins and complexes. Sample preparation is still a limiting factor in collecting high quality data. Thickness of the vitreous ice in which the particles are embedded is one of the many variables that need to be optimized for collection of the highest quality data. Here we present two methods, using either an energy filter or scattering outside the objective aperture, to measure ice thickness for potentially every image collected. Unlike geometrical or tomographic methods, these can be implemented directly in the single particle collection workflow without interrupting or significantly slowing down data collection. We describe the methods as implemented into the Leginon/Appion data collection workflow, along with some examples from test cases. Routine monitoring of ice thickness should prove helpful for optimizing sample preparation, data collection, and data processing.

Keywords

Cryo-EM; energy filter; mean free path; ice thickness; inelastic scattering

Introduction

Recent advances in instrumentation have made calculation of near-atomic resolution structures by electron cryo-electron microscopy (cryo-EM) almost routine (Earl et al., 2017; Murata and Wolf, 2018). Detector improvements now give us the ability to correct for image drift, dose-compensate individual frames, and count electrons directly, greatly improving the DQE of the final images, to the point that information can be collected to nearly the Nyquist limit (Ripstein and Rubinstein, 2016). Modern electron microscopes have improved stability, such as constant power lenses, so they can remain well aligned over a several day data collection period (Bierhoff, 2005). Samples can be kept cold for days at a time through the

*Correspondence: Bridget Carragher, bcarr@nysbc.org.

Publisher's Disclaimer: This is a PDF file of an unedited manuscript that has been accepted for publication. As a service to our customers we are providing this early version of the manuscript. The manuscript will undergo copyediting, typesetting, and review of the resulting proof before it is published in its final citable form. Please note that during the production process errors may be discovered which could affect the content, and all legal disclaimers that apply to the journal pertain

use of autofillers, removing the requirement to fill a dewar every few hours. Various automation programs, such as Leginon (Suloway et al., 2005), SerialEM (Mastronarde, 2005), UCSFImage (Li et al., 2015), and EPU (Thermo Fisher Scientific), have been in place for over a decade, and after initial setup and screening, allow for data collection to proceed over days at a time, providing particle numbers on the order of 10^5 to 10^6 from a single grid. These technological improvements, first available on the highest end microscopes, are now migrating to the microscopes previously thought of as screening instruments (Herzik et al., 2017).

While instrumentation has greatly improved, sample preparation remains a significant bottleneck to the generation of high resolution structures. For some samples it can be challenging to get particles into the ice suspended across the holes in the carbon. The most common way of preparing samples, by blotting followed by plunge freezing into liquid ethane or propane (Adrian et al., 1984), has not changed significantly in over 20 years apart from the development of automated blotting devices (Dandey et al., 2018; Feng et al., 2017; Jain et al., 2012; Razinkov et al., 2016; Wei et al., 2018). The vitrification technique takes some time to master and can produce a wide range of ice thicknesses (Fig. 1) across the grid, the square, and even within individual holes. In addition, it was recently demonstrated that most particles adhere to an air-water interface and if the ice is sufficiently thick this may result in collecting images that contain two layers of particles that will be at different defocus levels and potentially overlap in projection (Noble et al., 2018a). The ideal ice layer is usually considered to be just thick enough to support the particle in ice, and so on the order of the size of the particle (Wu et al., 2016). For these reasons, the ability to determine ice thickness during screening or during data collection is helpful in assessing overall grid quality, and for determining optimal locations for data collection across the grid, and within each hole.

Ice thickness can be measured in several ways. Possibly the most well-defined method is to collect a tomographic tilt series of the desired area, calculate a tomogram, and determine the local ice thickness across the reconstructed area (Noble et al., 2017; Noble et al., 2018b). This approach is much more time consuming than standard single particle data collection and so is only practical for measuring a few areas, perhaps at the beginning of collection to help determine a targeting strategy. A second method requires tilting the sample to 30 degrees, milling a small hole through the ice, tilting to -30 degrees, and taking an image (Angert et al., 1996). The geometry of this scheme means that the ice thickness can be directly determined by measuring the length of the hole in the second image. While this is easier to do than collecting a full tilt series, it is also disruptive during a data collection run and determining the start and end points of the ice tunnel can be difficult (Feja and Aebi, 1999).

The availability of an energy filter allows direct determination of the ice thickness, either through integration of the energy loss spectrum or through comparison of filtered and unfiltered image intensities (Brydson and Royal Microscopical Society (Great Britain), 2001). This requires knowledge of the inelastic mean free path of the electron through the sample, which depends on voltage and sample composition. In practical terms, what is generally measured is the apparent mean free path between inelastic scattering events, which

also has a dependence on objective aperture size and optics of the microscope, and likely is the reason for differences in various measurements at the same voltage (see for example (Angert et al., 1996) and (Grimm et al., 1996)). The relationship is described in equation (1), where d represents the ice thickness, I is the integrated total intensity, I_{ZIP} is the integrated zero-loss peak intensity, and Λ is the apparent mean free path for inelastic scattering.

$$d = \Lambda \ln \frac{I}{I_{ZIP}} \quad (1)$$

Thickness can also be estimated by comparing intensities with and without sample in the electron beam by measuring change in intensity due to electron loss through scattering. Electrons may be lost due to scattering outside the objective aperture and energy filtration if an energy slit is present. This second method has the advantage of not requiring an energy filter. The difference in intensity over vacuum and over sample is described by the Beer-Lambert law, shown in equation 2, where d represents the sample thickness, I_0 is intensity over vacuum, I is the intensity over ice, and λ is the mean free path for electron scattering (Yan et al., 2015).

$$d = \lambda \ln \frac{I_0}{I} \quad (2)$$

The measured value for λ would ideally depend only on voltage and sample composition. In practical terms, what is measured is an apparent mean free path for electron scattering, since a smaller objective aperture or the use of an energy filter will result in a greater loss of electrons outside of these apertures, and thus a smaller value for λ . We will refer to this method as aperture limited scattering (ALS) and λ as the ALS coefficient to emphasize this fact.

To our knowledge, absolute thickness measurements are not routinely determined during single particle experiments, although the ALS method has been successfully used in the grid screening process (Yu et al., 2016). Determination through tilting is understandably more difficult and unlikely to be performed as a matter of course. However, estimates from intensity, either with or without an energy filter, can be used to measure the thickness using automated software with minimal additional configuration and time overhead. What is lacking are accurate values for the inelastic mean free path and ALS coefficients for the microscope being used. These values have previously been determined empirically as well as calculated based on scattering theory, mostly at 80–120 keV, and the published results exhibit a wide variation (see (Vulovic et al., 2013) for a plot of several values). As mentioned above, what is often measured is apparent mean free path for inelastic scattering, since some electrons will also be lost due to elastic scattering outside of the objective aperture, and so measurements for inelastic scattering will have an elastic component (Feja and Aebi, 1999; Grimm et al., 1996). In addition, images are being used to approximate the zero loss peak intensity, which will differ slightly from collecting spectra and integrating the

zero loss peaks. ALS coefficients will depend on the electron optics of the instrument. Nevertheless, for a specific microscope type and objective aperture diameter they should be constant.

We have measured the apparent mean free path for inelastic scattering on an FEI Titan Krios (300keV) equipped with a BioQuantum GIF, and ALS coefficients on an FEI Titan Krios (300keV), an FEI BioTwin T12 (120 keV), and a Tecnai F20 (200 keV). With these values we have implemented a method into our standard Leginon/Appion data collection workflow (Lander et al., 2009; Suloway et al., 2005) so users can routinely monitor ice thickness. Here we will describe the methodology and the results from a few of our standard test specimens. The methods described in this paper could be readily implemented on other microscopes and automated data collection workflows.

Methods

Specimen preparation:

Samples were plunge frozen using standard techniques on a Gatan CP3 plunge freezer or a Leica plunge freezer. Rabbit muscle aldolase was prepared and frozen on gold Ultrafoil grids according to (Herzik et al., 2017). Proteasome grids were prepared and frozen on Quantifoil grids according to (Campbell et al., 2015). Glutamate dehydrogenase was prepared and frozen according to (Merk et al., 2016).

Microscopy:

Images were collected on several microscopes and cameras:

Titan Krios with energy filter, 100 μm objective aperture, nanoprobe mode, Gatan Bioquantum K2, dose rate 8 $\text{e}^-/\text{pix}/\text{sec}$.

Titan Krios with energy filter and Cs corrector, 100 μm objective aperture, nanoprobe mode, Gatan Bioquantum K2, dose rate 8 $\text{e}^-/\text{pix}/\text{sec}$.

Titan Krios, Gatan K2, 100 μm objective aperture, nanoprobe mode, dose rate 8 $\text{e}^-/\text{pix}/\text{sec}$.

Tecnai F20, 70 or 100 μm objective aperture, microprobe mode, DE20 direct detector, dose rate 2 $\text{e}^-/\text{pix}/\text{frame}$.

Tecnai T12, 70 or 100 μm objective aperture, microprobe mode, TVIPS F416 CMOS detector.

In all cases, the beam size was on the order of the hole size and was hitting the surrounding carbon under the exposure conditions.

Tomography:

Tilt series were collected on the Titan Krios microscopes using the Tomography app as implemented in Leginon (Suloway et al., 2009). Tilt series were collected on the T12 and F20 microscopes using SerialEM software (Mastronarde, 2005). In both cases, tilt series were collected bidirectionally between -45 and $+45$ degrees with a starting angle of 0

degrees and an angular increment of 3 degrees. In all cases, tomograms were calculated using Protomo software as implemented in Appion (Noble and Stagg, 2015).

Ice thickness measurement from tomograms:

Based on the results of hundreds of tomograms (Noble et al., 2018b), we know that the vast majority of particles on a vitrified grid are closely associated with the air water interface. We can thus estimate thickness from the z height between proteins (or ice contamination) observed on the two interfaces. For very thin samples the thickness was estimated from this single layer. For thicker samples which had a varying thickness, we chose the approximate average thickness.

Single particle dataset collection:

Images were collected using the Legimon workflow (Suloway et al., 2005). For ice thickness determination by inelastic scattering, two extra 0.5s images were collected; the first image with no slit inserted and the second with a 15 eV slit inserted. Note that our software allows us to collect standard images with a 20 eV slit width then temporarily change it to 15 eV for these two extra images. For ice thickness measurements determined using ALS, several images were taken at the start of the session over vacuum and the mean of these intensities was used as a reference for I_0 .

Image calculations:

Various image calculations and plots were calculated using the EMAN2/Sparx suite of image analysis tools (Hohn et al., 2007; Ludtke, 2016). CTF measurements and Thon ring extent were done using CTFFIND4 (Rohou and Grigorieff, 2015). Curve fitting and plotting was performed using gnuplot, an open source plotting tool.

Results and Discussion

Determination of the apparent mean free path for inelastic scattering.

The only parameter needed for determining ice thickness when using an energy filter is the mean free path for inelastic scattering in ice. This value will vary depending on the protein concentration, but not significantly. We measured ice thickness from tomograms of two samples (T20s proteasome and rabbit muscle aldolase). We then took pairs of images, with and without the energy filter slit inserted, to get values for I and I_{ZLP} . A slit width of 15 eV (*i.e.* 7.5 eV energy cutoff) was chosen to eliminate most of the ice and organic plasmon and structure peaks (Leapman and Sun, 1995). A plot of thickness versus $\ln(I/I_{ZLP})$ is linear, with slope 395 ± 11 nm (Fig. 2A). This value agrees well with the previously determined value of 400 nm by (Yonekura et al., 2006), on a 300 keV Polara F30 with 70 μ m objective aperture. Images of areas of thicker ice occasionally have varying values across the area, which made a single number hard to determine. Nevertheless, since the log ratio of intensities is multiplied directly by this slope value, it should be accurate enough for routine measurement. As a further verification of the method, we acquired pairs of images at 0 degree and 45 degree nominal tilt, and determined thickness using equation (1) and the mean free path as determined above and as expected the thickness increased by a factor of 1.4 (Fig. 2B). We also determined the apparent mean free path at 20 eV slit width by imaging

several holes with both 15 and 20 eV slit widths and comparing ratios of $\ln(I/I_{ZP})$. We found that the wider slit width showed an apparent mean free path 10% higher (Table 1), further demonstrating that these measurements are apparent and not absolute mean free paths. We also found that removing the objective aperture had no measurable effect on this value.

Determination of ALS coefficients

The determination of ice thickness using an energy filter is convenient, but this method has two disadvantages: it requires a microscope equipped with an energy filter, and two extra images must be acquired. These images are taken automatically using a very short exposure time, and need not be done for every exposure, but they do add a slight overhead to the collection time. An alternative option is to measure attenuation due to aperture limited scattering. For this, one must determine the ALS coefficient λ and measure the intensity over vacuum using the same imaging conditions as used for collection. Application of equation (2) then provides the thickness. The disadvantage of this method is that it relies on the beam intensity being constant throughout the collection, although this can be monitored and a new vacuum image can be acquired when necessary.

In order to determine the value of λ on microscopes with an energy filter, we measured ice thickness using the energy filter, then used this value in equation 2 to solve for λ . Imaging in this case was performed with our more commonly used slit width of 20 eV. A plot of thickness versus $\log(I_0/I)$ is shown in Fig. 2C. The plot is highly linear, with a slope of 332 nm. We repeated the experiment three times, using both T20S proteasome and aldolase samples, with an average value of 322 nm (sd=6 nm; n=4).

For microscopes without an energy filter, we used several methods to determine the ALS coefficients for the conditions under study. The first method was to collect tomograms at various locations to calculate thickness, then measure intensities over the same or similar holes and apply equation 2 to determine λ . This was done using both the thin aldolase sample and a thicker proteasome sample. The second method was to use the aldolase sample frozen on gold grids as a ruler to get a defined ice thickness. Our studies of this sample show that we get a very reproducible average ice thickness of 15–20 nm for most images. For each microscope, we collected a reasonably sized dataset (at least 100 images). For each image, we calculated a value of $\log(I_0/I)$, calculated mean and standard deviation values for the set, discarded values more than 2 standard deviations from the mean, and collected a new mean value from the pruned set. Application of equation (2) using the known ice thickness of 15 nm gave us a measure of λ . The final method, used for determining the ALS coefficient for a second objective aperture once it was determined for the first, was to measure intensities over the same area using both apertures, calculate the thickness from the first then determine the coefficient for the second aperture by rearranging equation 2. These results are summarized in Table 2.

These results show that a value of 300 nm is a good starting estimate for ALS coefficients on both the T12 and F20 microscopes or the Titan Krios when an energy filter is inserted, despite voltage differences. For the Titan Krios without an energy filter, the value is 10-fold higher, presumably due to the fact that it includes information out to about 1.4 Å resolution.

Without an objective aperture inserted on this microscope, there is very little loss by scattering and so the method will not work because of the high value of λ .

Application to test samples:

As a test specimen, we collected a high quality dataset on our Titan Krios microscope equipped with Cs corrector and Gatan Bioquantum energy filter. This was an exquisitely thin sample, approximately 15 nm on average, and our standard processing pipeline resulted in a structure of 2.4 Å resolution (further details in (Kim et al., 2018)). A histogram of ice thickness for this sample is shown in Fig. 3A. The majority of the images had extremely thin ice: less than 20 nm as measured by energy filtration. Considering the long dimension of the protein is ~10 nm, the ice was thick enough to just support a single layer of particles. The protein was also very tightly packed together, and perhaps this close packing helps to support an extremely thin layer of ice. The sample was frozen on a gold foil rather than carbon, and gold foils have been shown to result in more consistent ice in general (Russo and Passmore, 2014). We found that, for this sample, the use of carbon grids resulted in either less tightly packed or more aggregated protein and thicker ice overall (Kim et al., 2018).

The proteasome test specimen, in contrast, showed much thicker ice overall and a greater variation in thickness (Fig. 3B). Despite this increased thickness, this dataset also went to high resolution (2.6Å, data not shown). Note that the proteasome is larger, at nearly 20 nm for its longest dimension. Tomography by Noble et al. (Noble et al., 2018a) showed that most of the sample is at one of the air-water interfaces, which means that most of the particles are at the same height despite the increased ice depth. Additionally, it is approximately 700 kDa in molecular weight, making it easier to see in thicker ice than the 150 kDa aldolase protein.

Sorting the images based on ice thickness revealed some interesting trends. Dividing the micrographs into bins of 25 nm thickness (0–25 nm, 25–50 nm, ..., 175–200 nm) and plotting the 1D radial averages of summed power spectra showed that the ice ring at 3.9 Å resolution only appears once the ice is 50–75 nm thick, with increasing prominence as the ice layer thickens (Fig. 3 (C)–(J)). Thus, the appearance of this ring gives immediate feedback about the overall thickness of the sample. This sort of information is also collected in the Focus package (Biyani et al., 2017), where it is listed as an “iciness” metadata value for the micrograph. Plotting ice thickness versus Thon ring extent showed that images with the thinnest ice had the highest resolution Thon rings, apart from where the ice was too thin or absent. As ice becomes thicker, resolution steadily worsens. Interestingly, we often notice two separate lines as ice thickens, indicating a subset of images still goes to high resolution in spite of thicker ice (Fig. 4). We have seen this trend on several samples. While we don’t have a definitive explanation, one possibility is that some areas have most particles on one of the air-water interfaces, thus putting most of the scattering at the same height, Other areas of thick ice may have particles on both interfaces and/or scattered throughout the layer, resulting in a less coherent CTF. Examples of both distributions have been observed (Noble et al., 2017).

Implementation

We have implemented both methods of ice thickness determination into a new node in Leginon named IceT (Fig. 5A), and results are shown in the Leginon web interface (Fig. 5B,C). For thickness determination by energy filtration, the user needs to input the mean free path, exposure times, slit width, and how frequently to take the measurement. This feature could be readily added to other collection suites such as SerialEM (Mastronarde, 2005) and EPU. Thickness determination by ALS can be done on every image with no loss in time, with the caveat that any change in beam intensity will throw off the calculation. On microscopes with an energy filter, we found that measurement using ALS was virtually identical to measurement using the energy filter. Therefore, we now measure thickness with the filter for every image at the start of a session during the initial screen, then reduce to every 10 or 20 images once fully automated collection starts. On high-end microscopes, beam intensity is generally constant throughout a 1 or 2 day imaging session. For screening microscopes, it would be advisable to occasionally check intensity over vacuum and adjust the reference brightness parameter accordingly. We have found that routine monitoring of ice thickness has been useful in both guiding targets and characterizing samples.

Acknowledgements

This work was performed at the Simons Electron Microscopy Center and National Resource for Automated Molecular Microscopy located at the New York Structural Biology Center, supported by grants from the Simons Foundation (349247), NYSTAR, and the NIH National Institute of General Medical Sciences (GM103310) with additional support from the Agouron Institute [Grant Number: F00316] and NIH S10 OD019994-01. We thank Carl Negro (New York Structural Biology Center) for technical assistance with php code. We thank Yifan Cheng (UCSF) for the gift of the T20S proteasome sample.

References

- Adrian M, Dubochet J, Lepault J, McDowell AW, 1984 Cryo-electron microscopy of viruses. *Nature* 308, 32–36. [PubMed: 6322001]
- Angert I, Burmester C, Dinges C, Rose H, Schroder RR, 1996 Elastic and inelastic scattering cross-sections of amorphous layers of carbon and vitrified ice. *Ultramicroscopy* 63, 181–192.
- Bierhoff MPMD,NL), Kooijman, Sander Kees Cornelis (Veldhoven NL), Sanford, Colin August (Atkinson, NH, US). 2005 Magnetic lens.
- Biyani N, Righetto RD, McLeod R, Caujolle-Bert D, Castano-Diez D, Goldie KN, Stahlberg H, 2017 Focus: The interface between data collection and data processing in cryo-EM. *J Struct Biol* 198, 124–133. [PubMed: 28344036]
- Brydson R, Royal Microscopical Society (Great Britain), 2001 Electron energy loss spectroscopy Bios in association with the Royal Microscopical Society, Oxford.
- Campbell MG, Veesler D, Cheng A, Potter CS, Carragher B, 2015 2.8 Å resolution reconstruction of the *Thermoplasma acidophilum* 20S proteasome using cryo-electron microscopy. *Elife* 4.
- Dandey VP, Wei H, Zhang Z, Tan YZ, Acharya P, Eng ET, Rice WJ, Kahn PA, Potter CS, Carragher B, 2018 Spotiton: New features and applications. *J Struct Biol* 202, 161–169. [PubMed: 29366716]
- Earl LA, Falconieri V, Milne JL, Subramaniam S, 2017 Cryo-EM: beyond the microscope. *Curr Opin Struct Biol* 46, 71–78. [PubMed: 28646653]
- Feja B, Aebi U, 1999 Determination of the inelastic mean free path of electrons in vitrified ice layers for on-line thickness measurements by zero-loss imaging. *J Microsc* 193, 15–19. [PubMed: 12558683]
- Feng X, Fu Z, Kaledhonkar S, Jia Y, Shah B, Jin A, Liu Z, Sun M, Chen B, Grassucci RA, Ren Y, Jiang H, Frank J, Lin Q, 2017 A Fast and Effective Microfluidic Spraying-Plunging Method for High-Resolution Single-Particle Cryo-EM. *Structure* 25, 663–670 e663. [PubMed: 28286002]

- Grimm R, Typke D, Barmann M, Baumeister W, 1996 Determination of the inelastic mean free path in ice by examination of tilted vesicles and automated most probable loss imaging. *Ultramicroscopy* 63, 169–179. [PubMed: 8921626]
- Herzik MA, Jr., Wu M, Lander GC, 2017 Achieving better-than-3-Å resolution by single-particle cryo-EM at 200 keV. *Nat Methods* 14, 1075–1078. [PubMed: 28991891]
- Hohn M, Tang G, Goodyear G, Baldwin PR, Huang Z, Penczek PA, Yang C, Glaeser RM, Adams PD, Ludtke SJ, 2007 SPARX, a new environment for Cryo-EM image processing. *J Struct Biol* 157, 47–55. [PubMed: 16931051]
- Jain T, Sheehan P, Crum J, Carragher B, Potter CS, 2012 Spotiton: a prototype for an integrated inkjet dispense and vitrification system for cryo-TEM. *J Struct Biol* 179, 68–75. [PubMed: 22569522]
- Kim LY, Carragher B, Rice WJ, Eng ET, Kopylov M, Cheng A, Raczkowski AM, Jordan KD, Bobe D, Potter C, 2018 Benchmarking cryo-EM single particle analysis workflow. *bioRxiv*.
- Lander GC, Stagg SM, Voss NR, Cheng A, Fellmann D, Pulokas J, Yoshioka C, Irving C, Mulder A, Lau PW, Lyumkis D, Potter CS, Carragher B, 2009 Appion: an integrated, database-driven pipeline to facilitate EM image processing. *J Struct Biol* 166, 95–102. [PubMed: 19263523]
- Leapman RD, Sun S, 1995 Cryo-electron energy loss spectroscopy: observations on vitrified hydrated specimens and radiation damage. *Ultramicroscopy* 59, 71–79. [PubMed: 7571121]
- Li X, Zheng S, Agard DA, Cheng Y, 2015 Asynchronous data acquisition and on-the-fly analysis of dose fractionated cryoEM images by UCSFImage. *J Struct Biol* 192, 174–178. [PubMed: 26370395]
- Ludtke SJ, 2016 Single-Particle Refinement and Variability Analysis in EMAN2.1. *Methods Enzymol* 579, 159–189. [PubMed: 27572727]
- Mastrorade DN, 2005 Automated electron microscope tomography using robust prediction of specimen movements. *J Struct Biol* 152, 36–51. [PubMed: 16182563]
- Merk A, Bartesaghi A, Banerjee S, Falconieri V, Rao P, Davis MI, Pragani R, Boxer MB, Earl LA, Milne JLS, Subramaniam S, 2016 Breaking Cryo-EM Resolution Barriers to Facilitate Drug Discovery. *Cell* 165, 1698–1707. [PubMed: 27238019]
- Murata K, Wolf M, 2018 Cryo-electron microscopy for structural analysis of dynamic biological macromolecules. *Biochim Biophys Acta* 1862, 324–334. [PubMed: 28756276]
- Noble AJ, Stagg SM, 2015 Automated batch fiducial-less tilt-series alignment in Appion using Protomo. *J Struct Biol* 192, 270–278. [PubMed: 26455557]
- Noble AJ, Wei H, Dandey VP, Zhang Z, Potter CS, Carragher B, 2018a Reducing effects of particle adsorption to the air-water interface in cryoEM. *bioRxiv*.
- Noble AJ, Dandey VP, Wei H, Brasch J, Chase J, Acharya P, Tan YZ, Zhang Z, Kim LY, Scapin G, Rapp M, Eng ET, Rice WJ, Cheng A, Negro CJ, Shapiro L, Kwong PD, Jeruzalmi D, des Georges A, Potter CS, Carragher B, 2017 Routine Single Particle CryoEM Sample and Grid Characterization by Tomography. *bioRxiv*.
- Noble AJ, Dandey VP, Wei H, Brasch J, Chase J, Acharya P, Tan YZ, Zhang Z, Kim LY, Scapin G, Rapp M, Eng ET, Rice WJ, Cheng A, Negro CJ, Shapiro L, Kwong PD, Jeruzalmi D, des Georges A, Potter CS, Carragher B, 2018b Routine single particle CryoEM sample and grid characterization by tomography. *Elife* 7.
- Razinkov I, Dandey V, Wei H, Zhang Z, Melnekoff D, Rice WJ, Wigge C, Potter CS, Carragher B, 2016 A new method for vitrifying samples for cryoEM. *J Struct Biol* 195, 190–198. [PubMed: 27288865]
- Ripstein ZA, Rubinstein JL, 2016 Processing of Cryo-EM Movie Data. *Methods Enzymol* 579, 103–124. [PubMed: 27572725]
- Rohou A, Grigorieff N, 2015 CTFFIND4: Fast and accurate defocus estimation from electron micrographs. *J Struct Biol* 192, 216–221. [PubMed: 26278980]
- Russo CJ, Passmore LA, 2014 Electron microscopy: Ultrastable gold substrates for electron cryomicroscopy. *Science* 346, 1377–1380. [PubMed: 25504723]
- Suloway C, Pulokas J, Fellmann D, Cheng A, Guerra F, Quispe J, Stagg S, Potter CS, Carragher B, 2005 Automated molecular microscopy: the new Legimon system. *J Struct Biol* 151, 41–60. [PubMed: 15890530]

- Suloway C, Shi J, Cheng A, Pulokas J, Carragher B, Potter CS, Zheng SQ, Agard DA, Jensen GJ, 2009 Fully automated, sequential tilt-series acquisition with Leginon. *J Struct Biol* 167, 11–18. [PubMed: 19361558]
- Vulovic M, Ravelli RB, van Vliet LJ, Koster AJ, Lazic I, Lucken U, Rullgard H, Oktem O, Rieger B, 2013 Image formation modeling in cryo-electron microscopy. *J Struct Biol* 183, 19–32. [PubMed: 23711417]
- Wei H, Dandey VP, Zhang Z, Raczkowski A, Rice WJ, Carragher B, Potter CS, 2018 Optimizing “self-wicking” nanowire grids. *J Struct Biol* 202, 170–174. [PubMed: 29317278]
- Wu S, Armache JP, Cheng Y, 2016 Single-particle cryo-EM data acquisition by using direct electron detection camera. *Microscopy (Oxf)* 65, 35–41. [PubMed: 26546989]
- Yan R, Edwards TJ, Pankratz LM, Kuhn RJ, Lanman JK, Liu J, Jiang W, 2015 Simultaneous determination of sample thickness, tilt, and electron mean free path using tomographic tilt images based on Beer-Lambert law. *J Struct Biol* 192, 287–296. [PubMed: 26433027]
- Yonekura K, Braunfeld MB, Maki-Yonekura S, Agard DA, 2006 Electron energy filtering significantly improves amplitude contrast of frozen-hydrated protein at 300kV. *J Struct Biol* 156, 524–536. [PubMed: 16987672]
- Yu G, Li K, Huang P, Jiang X, Jiang W, 2016 Antibody-Based Affinity Cryoelectron Microscopy at 2.6-Å Resolution. *Structure* 24, 1984–1990. [PubMed: 27806259]

Figure 1.

(A) Representative grid atlas collected in Leginon, showing an evident gradient in ice thickness. (B) A "square" level image from this same grid shows evidence of varying thickness, including occasional empty holes.

Figure 2.

(A) Determination of mean free path for inelastic scattering by electron tomography. Images were collected for both aldolase and proteasome samples with and without a 15 eV energy slit. Tomograms were collected at the same areas as thickness was measured. Thickness versus $\log (I_{tot}/I_{ZLP})$ is plotted. The line of best fit (green) had a slope of 395 nm. (B) Thickness determined using the energy filter for untilded and 45 degree tilted images. Inset: an overview of 4 holes are shown where the measurements were made, untilded (left) and tilted (right). For each hole, thickness was measured as described for both tilted and untilded images, and plotted as tilted thickness versus untilded thickness (purple crosses). The green line is a plot of the equation $y=1.41x$, which is where the points should ideally lie. (C) Determination of the ALS coefficient using thickness as measured by the energy filter. Thickness was measured over many images using the energy filter and eq. (1). Thickness plotted versus $\log I_0/I$ as described in the text. For this experiment, the slope was 332 nm.

Figure 3.

(A) Histogram of thickness values as measured on a rabbit muscle aldolase sample frozen on gold Ultrafoil grids. (B) Histogram of thickness values as measured on a T20S proteasome test sample frozen on C-flat carbon grids. (C-J) Plots of mean radial intensity of Fourier transforms of image averages versus resolution for various ice thicknesses. Blue bar: 3.9 Å⁻¹ resolution. (C): 0–25 nm thickness. (D): 25–50 nm thickness. (E): 50–75 nm thickness. (F) 75–100 nm thickness. (G): 100–125 nm thickness. (H): 125–150 nm thickness. (I): 150–175 nm thickness. (J): 175–200 nm thickness.

Figure 4: Plots of Thon ring extent, as measured by CTFFIND4 ($\text{\AA}^{-1}$), versus ice thickness for several samples. (A): glutamate dehydrogenase. (B, C): rabbit muscle aldolase. (D): T20S proteasome. Thon ring extent goes to very low resolution where the ice is too thin or absent, is optimal at thin ice, and generally rises as ice thickens.

Figure 5:

Integration of ice thickness determination into Legion. (A) The control panel for the node is shown. The user provides various parameters for thickness determination and chooses how often the thickness is measured. (B) Thickness plots for an experiment as shown in the Appion summary pages. (C) Thickness information is displayed when viewing the image in the Legion web interface.

Table 1:

Apparent mean free path for inelastic scattering

Microscope	Voltage (keV)	Slit Width (eV)	Apparent MFP for inelastic scattering (nm)	Method
Titan Krios	300	15	395 +/- 11	Tomography
Titan Krios	300	20	435 +/- 30	Compare with 15 eV

Table 2:

ALS coefficients

Microscope	Voltage (keV)	Obj. Aperture diameter (μm)	Lambda (nm)	Method
Titan Krios 20 eV slit	300	100	322	compare with EF determination
Titan Krios (no EF)	300	100	3,329	Aldolase thickness
Titan Krios (no EF)	300	none	78,788 *	Aldolase thickness
Tecnai F20	200	100	392	Tomography
Tecnai F20	200	70	302	Compare with 100 μm
Tecnai T12	120	100	319	Tomography
Tecnai T12	120	70	247	Compare with 100 μm

* unusably high