Published in final edited form as:

Intern Med Rev (Wash D C). 2016 December ; 2(11): . doi:10.18103/imr.v2i11.271.

Low Intensity Ultrasound for Promoting Soft Tissue Healing: A Systematic Review of the Literature and Medical Technology

Thomas M. Best, MD, PhD¹, Kevin E. Wilk, DPT², Claude T. Moorman, MD³, and David O. Draper, EdD⁴

¹Ohio State University, Columbus, OH

²Champion Sports Medicine, Birmingham, AL

³Duke University, Durham, NC

⁴Brigham Young University, UT

Abstract

Therapeutic ultrasound has been studied and used for the past seven decades to treat musculoskeletal injuries. Recently, a significant body of animal and human research has focused on the biomechanical effects of daily-applied, low intensity therapeutic ultrasound (LITUS) on soft tissue recovery. We performed a systematic review of the last two decades of LITUS literature to examine the effects on tendon, skeletal muscle, ligament, and tendon-bone junction injuries. LITUS facilitated tendon healing, with increased tensile strength and improved collagen alignment. For skeletal muscle and ligament injuries, LITUS increased cell proliferation during myoregeneration and improved tissue biomechanics (ultimate load, stiffness, energy absorption). LITUS aided tendon-bone junction healing through improved tissue function. Scientific evidence supports the use of LITUS to treat soft tissue injuries, and improve outcomes for musculoskeletal injuries and post-operative recovery. Lastly, we discuss the use of LITUS devices facilitating daily applied therapy in the home setting.

Keywords

Muscle; Tendon; Ligament; Healing; Bioeffects; Rehabilitation; Regeneration; Therapy; Therapeutic Ultrasound; Acoustic; Medicine

Introduction

Soft tissue injuries, both acute or chronic, are among the most frequent issues addressed by physical therapists, athletic trainers and primary care physicians; however, many currently available treatment options are costly, time consuming, and potentially harmful.^{1,2} Nonsteroidal anti-inflammatory pharmacotherapies can be deleterious to soft tissue healing and harmful to the gastrointestinal and renal systems.³ Interventions such as prescription painkillers, corticosteroid injections, and platelet-rich plasma injections can be addictive,

costly, and of questionable efficacy, respectively.⁴ Traditional rest, time-off, and elevation therapy may not be feasible for all soft tissue injuries and patient lifestyles. Many use the regimen of rest, ice, compression, elevation and stabilization (RICES) to treat acute injury, but are unsure what to do if the injury becomes chronic.¹ Whereas, many of these current treatment options address inflammation and pain management, therapeutic ultrasound can both manage pain and facilitate healing. In the following systematic review, we examine the efficacy of low intensity therapeutic ultrasound (LITUS) in promoting healing of soft tissue injuries, and further discuss LITUS delivery systems currently available for researchers, medical practitioners and patients.

Therapeutic ultrasound from 1-3 MHz has been used for over seven decades by physical therapists, athletic trainers and other health care providers to relieve pain and facilitate the healing process.⁵⁻⁷ Traditional therapeutic ultrasound has been used at intensities of 1-4 W/cm² in the physical therapy setting⁷ and at higher intensities up to 15,000 W/cm² for ablation of cancerous tissue,⁸ in either pulsed or continuous treatment modes. More recently, the use of therapeutic ultrasound at lower intensities (0.5-1 W/cm²), in conjunction with other treatment modalities, has been used in both the home and the clinical setting to treat tendon, ligament, and muscle injuries.^{2,9-15} LITUS has been shown to alter tissue biomechanical properties,¹⁶ improve collagen alignment,^{17,18} and stimulate cell proliferation.¹⁹ The majority of the LITUS research in human subjects and animal models has investigated bone fracture healing.²⁰ This systematic review examines the efficacy of LITUS for improving healing in four tissue structures: (1) tendon, (2) muscle, (3) ligament, and (4) tendon-bone junctions. Additionally, two LITUS delivery systems and the benefits of longer duration LITUS treatment are discussed following a summary of the literature.

Methods: Search Strategy and Criteria

A systematic literature review was executed and studies were identified on PubMed, Google Scholar, and in references from review articles that met the following inclusion criteria: 1) the study included at least one experimental arm that was treated with low intensity therapeutic ultrasound (≤ 1.5 W/cm²); 2) the study design had a control group that did not receive an active or alternative treatment to therapeutic ultrasound; 3) the study investigated biomechanical properties, histology (collagen synthesis/alignment), cell proliferation, and/or pain; 4) the study focused on skeletal muscle, tendon, ligament, or tendon-bone injury and healing; 5) the study was an original research article (review articles were excluded); 6) the study had appropriate local institutional review board (IRB) and informed consent procedures for human subjects, and appropriate local Institutional Care and Animal Use Committee (IACUC) approval for animal subjects; 7) the study was not focused only on bone or fracture healing; and 8) the study was written in English. All search results and studies were independently evaluated by two authors (DD, TB) and further reviewed by research associates.

Results

The search yielded 35 applicable studies that met the inclusion criteria. These were subsequently categorized as: tendon (n=16), muscle (n=7), ligament (n=3), or tendon-bone

junction (n=9). It should be noted that the majority of applicable studies utilized animal models, likely due to the experimental techniques required to elucidate the biomechanical and physiological processes resulting from LITUS.

LITUS was shown to have a beneficial effect on tendon strength and collagen synthesis following injury. Rupture strength in tendons treated with LITUS was significantly greater than controls from 5 to 42 days post-injury ($p<0.05$; ^{16-18,21-23} Tensile strength and tendon extensibility in LITUS-treated tendons was also higher than in control tendons.^{16,22-25} Collagen synthesis, measured by conversion of radiolabeled proline to hydroxyproline, increased substantially from day 3 to day 5 post-injury and continued to show benefits through day 21 compared to controls ($p<0.05$; ^{17,18} Collagen type I and III expression was greater following LITUS treatment,²⁶ and higher birefringence (i.e., coherence of collagen alignment) was also observed in treated tendons compared to controls.^{27,28} Additionally, an examination of treatment time and duration found that introducing LITUS in the earlier stages of healing increased tensile strength and matrix synthesis compared to the later stages of healing.^{21,29} In three human studies examining epicondylitis and patellar tendinopathy, tendon pain was significantly decreased in a clinically meaningful way by up to 70% with daily applied continuous LITUS over the course of 6 weeks,⁴⁷ although in two of the studies pulsed LITUS produced similar decreases in pain with the control groups.^{11,14}

Application of LITUS increased cell proliferation and both myogenin and actin protein expression in skeletal muscle following a contusion injury. LITUS nearly doubled satellite cell proliferation compared to controls in an injured gastrocnemius muscle.¹⁹ A higher proliferation rate and cell number at days 6 and 8 were observed following the application of LITUS.³⁰ Cells treated with 8 doses of LITUS demonstrated a 40% increase in myogenin expression and a 47% increase in actin expression compared to controls.³⁰

Cyclooxygenase-2 (COX-2) expression was also reduced in injured muscle tissue (both *in vivo* and *in vitro*) treated with LITUS compared to controls.^{31,32} Two human studies examining pain and trigger point depths demonstrated that LITUS significantly reduces pain,¹² and relaxes muscles, although muscle relaxation did not differ significantly from controls.⁹

Ligament healing benefitted from the application of LITUS. LITUS treated ligaments exhibited superior mechanical properties including ultimate load, stiffness, and energy absorption.³³⁻³⁵ LITUS-treated ligaments from one study were 34.2% stronger, 27.0% stiffer, and could absorb 54.4% more energy compared to sham-treated ligaments after 2 weeks of treatment.³⁵ Another study demonstrated that after six weeks of LITUS treatment, ligaments were 39.5% stronger, 24.5% stiffer, could absorb 69.1% more energy, and were 10.6% larger than sham-treated ligaments.³³ Collagen fibril diameter was larger in a group treated with LITUS compared to controls,³⁴ and there was a greater relative proportion of type II collagen in LITUS-treated ligaments compared to controls at both 3 and 6 weeks.³³

For tendon-bone junction healing, LITUS treatment significantly improved healing and osteogenesis. Application of LITUS resulted in significantly more newly formed bone and improved tissue integration compared to controls.³⁶⁻⁴⁰ In one study,³⁹ reported the amount of new bone formed was 2.6 and 3.0 times greater in the treatment group compared to

controls at weeks 8 and 16, respectively. Vascular endothelial growth factor (VEGF) in the tendon-bone interface was also significantly increased with LITUS treatment, particularly after 4 weeks.^{37,41} Type I and II collagen were increased with LITUS treatment,^{42,43} and collagen fibers demonstrated higher organization.⁴³ Similarly, there was up-regulation of type I collagen gene expression with LITUS treatment compared to controls.⁴⁴ Additionally, in procedures replacing ligaments with tendons (i.e., anterior cruciate ligament surgery; ACL), LITUS treated tendons showed greater stiffness and peak load compared to controls.^{36,40,45}

Discussion

LITUS improves the rate of tissue repair following tendon, skeletal muscle, ligament, and tendon-bone junction injuries. Injured tendons treated with LITUS recover with greater tensile strength and extensibility than similar tendons that received no treatment post-injury.^{16,22-25} Recovering tendons treated with LITUS had greater amounts of collagen,²⁶ and improved organization and aggregation of collagen bundles.^{28,46} Also of importance, the intervention was most effective during the earlier stages of healing, with the most improvement noted during the first 14 days post-injury.²⁹ Tensile strength and collagen cellularity and stainability were more pronounced when LITUS was introduced during the granulation phase rather than the remodeling phase of tendon repair. The collective tendon results are promising when considering that 7% of all primary care physician visits and more than 30% of sports related injuries are related to tendinopathy, but there is currently no gold standard of treatment.⁴⁷ Although the timing of intervention was found to be an important factor, similar effects were noted for continuous or pulsed mode ultrasound with the exception of one study.⁴⁶

Healing of skeletal muscle also benefited from the application of LITUS. Following contusion injuries to skeletal muscle, cell proliferation increased up to 96% in LITUS treated injuries compared to controls during the early stages of regeneration, although there was no effect on myotube production.¹⁹ LITUS produced a higher cell number and proliferation rate at days 6 and 8; two factors related to muscle regeneration.³⁰ Myogenin and actin protein expression were increased at days 4, 6, and 8 in LITUS treated muscle compared to non-treated controls,³⁰ implying the existence of a cell process modulated by LITUS to affect the molecular biology of the cell, improving the rate of muscle repair and regeneration. Additionally, the reduction of inflammatory COX-2,^{31,32} suggests that LITUS accelerated the healing response. LITUS also reduced pain and improved relaxation of trapezius muscle tissue.^{9,12}

The three studies examining ligament healing revealed that LITUS-treated ligaments had superior mechanical properties including tensile strength and energy absorption.³³⁻³⁵ Ligaments treated with NSAIDs absorbed 33.3% less energy than those treated with a control vehicle that contained no COX-2 inhibitors.³⁵ These findings suggest that while NSAIDs address pain management, they may negatively impact the healing process. Comparatively, LITUS has the potential to provide both pain relief and anti-inflammatory effects without diminishing the healing process.

For tendon-bone junction injuries, LITUS treated animals had increased bone formation and greater mineral density than animals that did not receive any treatment.³⁶⁻³⁹ New bone area was 2.6 and 3.0 times greater in LITUS treated groups compared to controls at weeks 8 and 16, respectively.³⁹ Radiographic measurements observed more newly formed bone and increased tissue integration in LITUS treated groups.³⁸ VEGF and collagen expression were also increased in the tendon-bone interface with the use of LITUS.^{37,41-44} Cumulatively, these results support the use of LITUS to shorten post-operative recovery time after procedures such as ACL surgery. Animal models indicate that the tendons used to replace ligaments in ACL surgery that are treated with LITUS are stiffer and can sustain greater peak loads,^{36,40,45} suggesting better clinical outcomes.

LITUS Delivery Systems & Clinical Applications

Collectively, these studies are encouraging for the use of LITUS to treat soft tissue injuries in human; however, the delivery of therapeutic ultrasound has been traditionally applied in the inpatient setting, which limits both the duration of treatment and frequency of application. Currently two wearable LITUS devices designed for daily home use have been approved by the United States FDA and CE-marked, including the Exogen[®] Bone Healing System (Bioventus LLC, Durham, NC, U.S.), and the SAM[®] Sport Sustained Acoustic Medicine System (ZetrOZ Systems, LLC, Trumbull, CT, U.S.) Figure 1. The Exogen device operates a single ultrasonic collimated source at a frequency of 1.5 MHz, an intensity of 0.03 W/cm², temporal average power of 0.117 Watts with a pulsed 20% duty cycle for 20 min per day providing 140 Joules of ultrasound energy per treatment. The Exogen device is typically applied through a hole in a cast to a non-union fracture. The SAM device delivers low intensity continuous ultrasound at 3 MHz frequency, an intensity of 0.132 W/cm², temporal average power of 1.3 Watts for up to 4 hours per day from two divergent ultrasonic sources providing 18,720 Joules of ultrasound energy per treatment. The SAM device is applied with a star-shaped ultrasonic patch typically directly over and proximal to the soft-tissue injury. Both medical devices deliver low intensity ultrasound between 1-3 MHz, however the SAM device is 4.4× greater in intensity and 134× greater in energy delivery compared with Exogen, making the two devices considerably different and serving different clinical purposes.

The Exogen Bone Healing System is used in clinical practice to accelerate the healing of established non-unions (excluding skull and vertebra), in addition to accelerating fracture healing time in fresh, closed, posteriorly displaced distal radius fractures and fresh, closed, or Grade I open tibial diaphysis fractures. The SAM device is used to reduce inflammation and pain, and accelerate the healing of soft tissues such as tendons, ligaments, and muscles. Both Exogen and SAM are non-invasive prescription-use devices that are administered and monitored by a licensed medical professional. Typically, patients prescribed wearable LITUS therapy self-apply the therapy daily in the home-setting, and have regularly scheduled follow up appointments with their healthcare provider who oversees use of the device. Overall, the use of these LITUS delivery systems has the clinical potential to accelerate healing and alleviate pain from a variety of disorders impacting soft tissues as well as bone.

Conclusions & Future Directions

The use of LITUS in the treatment of tendon, muscle, ligament, and tendon-bone junction injuries is supported by the literature. An advantage of using therapeutic ultrasound at lower intensities is that it can be safely used in the home-setting for long treatment times, in some cases up to 8 hours.¹² The vast majority of research to date has used animal models to understand the mechanisms underlying LITUS, and additional research with human subjects is worthwhile to expand our understanding on how physiological observations correlate to improved clinical outcomes of patients. Given the frequency and debilitating nature of soft tissue injuries, LITUS has the potential to reduce healing time, overall healthcare costs, and negative side effects, as well as improve quality of life.

References

1. Knight KL, Draper DO. Therapeutic Modalities: The Art and Science. 2nd ed. Wolters/Kluwer Health, Lippincott Williams & Wilkins Baltimore, 2013.
2. Draper DO. Facts and misfits in ultrasound therapy: steps to improve your treatment outcomes. *European J of Phys and Rehabil Med* 2014;50:209–16. [PubMed: 24594854]
3. Dimmen S, Nordsletten L, Madsen J. Parecoxib and Indomethacin delay early fracture healing: a study in rats. *Clin Orthop* 2009;467:1992–99. [PubMed: 19319614]
4. Gosens T, Peerbooms JC, van Laar W, den Ouden BL. Ongoing positive effect of platelet-rich plasma versus corticosteroid injection in lateral epicondylitis: a double-blind randomized controlled trial with 2-year follow-up. *Am J Sports Med* 2011;39(6):1200–8. [PubMed: 21422467]
5. Dunn F & Frizzell LA. Bioeffects of Ultrasound In Lehman JF, ed. Therapeutic heat and cold. Maryland: Williams & Wilkins, 1990.
6. Lehmann JF & de Lateur BJ. Therapeutic heat In Lehman JF, ed. Therapeutic heat and cold. Maryland: Williams & Wilkins, 1990.
7. Speed CA. Therapeutic ultrasound in soft tissue lesion. *Rheumatology* 2001;40:1331–36. [PubMed: 11752501]
8. Zhou Y-F. High intensity focused ultrasound in clinical tumor ablation. *World J Clin Oncol* 2011;2(1):8–27. [PubMed: 21603311]
9. Draper DO, MaHaffey C, Eggett D. Therapeutic ultrasound softens trigger points in upper trapezius muscles. *Physiother Theory Prac* 2010; 26(3):167–72.
10. Karnes JL, Burton HW. Continuous therapeutic ultrasound accelerates repair of contraction-induced skeletal muscle damage in rats. *Arch Phys Med Rehabil* 2002;83(1):1–4. [PubMed: 11782824]
11. Dvaz AP, Ostor AJK, Speed CA, Jenner JR, Bradley M, Prevost AT, Hazleman BL. Pulsed low-intensity ultrasound therapy for chronic lateral epicondylitis: a randomized controlled trial. *Rheumatology* 2006;45:566–70. [PubMed: 16303817]
12. Lewis GK, Jr, Langer MD, Henderson CR, Jr, Ortiz R. Design and evaluation of a wearable self-applied therapeutic ultrasound device for chronic myofascial pain. *Ultrasound Med Biol* 2013;39(8), 1429–39. [PubMed: 23743101]
13. Ozgonenel L, Aytakin E, Durmosoqlu G. A double-blind trial of clinical effects of therapeutic ultrasound on knee osteoarthritis. *Ultrasound Med Biol* 2009;35(1):44–49. [PubMed: 18829151]
14. Warden SJ, Metcalf BR, Kiss ZS, Cook JL, Purdam CR, Bennell KL, Crossley KM. Low-intensity pulsed ultrasound for chronic patellar tendinopathy: a randomized, double-blind, placebo-controlled trial. *Rheumatol Int* 2008; 47:467–71.
15. Wong RA, Schumann B, Townsend R, Phelps CA. A survey of therapeutic ultrasound use by physical therapists who are orthopaedic certified specialists. *Phys Ther* 2007;87(8): 986–94. [PubMed: 17553923]
16. Enwemeka CS, Rodriguez O, Mendosa S. The biomechanical effects of low-intensity ultrasound on healing tendons. *Ultrasound Med Biol* 1990;16(8):801–7. [PubMed: 2095010]

17. Demir H, Menku P, Kirnap M, Calis M, Ikizceli I. Comparison of the effects of laser, ultrasound, and combined laser + ultrasound treatments in experimental tendon healing. *Lasers Surg Med* 2004;35(1):84–9. [PubMed: 15278933]
18. Jackson BA, Schwane JA, Starcher BC. Effect of ultrasound therapy on the repair of Achilles tendon injuries in rats. *Med Sci Sports Exerc* 1991;23(2):171–6. [PubMed: 2017013]
19. Rantanen J, Thorsson O, Wollmer P, Hurme T, Kalimo H. Effects of therapeutic ultrasound on the regeneration of skeletal myofibers after experimental muscle injury. *Am J Sports Med* 1999;27(1):54–9. [PubMed: 9934419]
20. Bashardoust Tajali S, Houghton P, MacDermid JC, Grewal R. Effects of low-intensity pulsed ultrasound therapy on fracture healing: a systematic review and meta-analysis. *Am J Phys Med Rehabil* 2012;91(4):349–67. [PubMed: 21904188]
21. Fu SC, Shum WT, Hung LK, Wong MW, Qin L, Chan KM. Low-intensity pulsed ultrasound on tendon healing: a study of the effect of treatment duration and treatment initiation. *Am J Sports Med* 2008;36(9):1742–9. [PubMed: 18645043]
22. Jeremias Júnior SL, Camanho GL, Bassit AC, Forgas A, Ingham SJ, Abdalla RJ. Low-intensity pulsed ultrasound accelerates healing in rat calcaneus tendon injuries. *J Orthop Sports Phys Ther* 2011;41(7):526–31. [PubMed: 21335926]
23. Larsen A, Kristensen G, Thorlacius-Ussing O, Oxlund H. The influence of ultrasound on the mechanical properties of healing tendons in rabbits. *Acta Orthop* 2005;76(2):225–30. [PubMed: 16097548]
24. Ng GY & Fung DT. The effect of therapeutic ultrasound intensity on the ultrastructural morphology of tendon repair. *Ultrasound Med Biol* 2007 11;33(11):1750–4. [PubMed: 17630094]
25. Ng CO, Ng GY, See EK, Leung MC. Therapeutic ultrasound improves strength of Achilles tendon repair in rats. *Ultrasound Med Biol* 2003; 29(10):1501–06. [PubMed: 14597348]
26. Kosaka T, Masaoka T, Yamamoto K. Possible molecular mechanism of promotion of repair of acute Achilles tendon rupture by low intensity-pulsed ultrasound treatment in a rat model. *West Indian Med J* 2011;60(3):263–8. [PubMed: 22224336]
27. da Cunha A, Parizotto NA, Vidal Bde C. The effect of therapeutic ultrasound on repair of the achilles tendon (tendo calcaneus) of the rat. *Ultrasound Med Biol* 2001;27(12):1691–6. [PubMed: 11839414]
28. Koeke PU, Parizotto NA, Carrinho PM, Salate AC. Comparative study of the efficacy of the topical application of hydrocortisone, therapeutic ultrasound and phonophoresis on the tissue repair process in rat tendons. *Ultrasound Med Biol* 2005;31(3):345–50. [PubMed: 15749557]
29. Fu SC, Hung LK, Shum WT, Lee YW, Chan LS, Ho G, Chan KM. In vivo low-intensity pulsed ultrasound (LIPUS) following tendon injury promotes repair during granulation but suppresses decorin and biglycan expression during remodeling. *J Orthop Sports Phys Ther* 2010;40(7):422–9. [PubMed: 20479531]
30. Chan Y, Hsu K, Kuo C, Lee S, Chen S, Chen W, Ueng SW. Using low-intensity pulsed ultrasound to improve muscle healing after laceration injury: An in vitro and in vivo study. *Ultrasound Med Biol* 2010 36(5):743–51. [PubMed: 20381949]
31. Nagata K, Nakamura T, Fujihara S, Tanaka E. Ultrasound modulates the inflammatory response Nagata and promotes muscle regeneration in injured muscles. *Ann Biomed Eng* 2013;41(6):1095–105. [PubMed: 23386031]
32. Rennó AC, Toma RL, Feitosa SM, Fernandes K, Bossini PS, de Oliveira P, Parizotto N, Ribeiro DA. Comparative effects of low-intensity pulsed ultrasound and low-level laser therapy on injured skeletal muscle. *Photomed Laser Surg* 2011;29(1):5–10. [PubMed: 21166589]
33. Sparrow KJ, Finucane SD, Owen JR, Wayne JS. The effects of low-intensity ultrasound on medial collateral ligament healing in the rabbit model. *Am J Sports Med* 2005;33(7):1048–56. [PubMed: 15888724]
34. Takakura Y, Matsui N, Yoshiya S, Fujioka H, Muratsu H, Tsunoda M, Kurosaka M. Low-intensity pulsed ultrasound enhances early healing of medial collateral ligament injuries in rats. *J Ultrasound Med* 2002;21(3):283–8. [PubMed: 11883539]

35. Warden SJ, Avin KG, Beck EM, DeWolf ME, Hagemeier MA, Martin KM. Low-intensity pulsed ultrasound accelerates and a nonsteroidal anti-inflammatory drug delays knee ligament healing. *Am J Sports Med* 2006;34(7):1094–102. [PubMed: 16476921]
36. Hu J, Qu J, Xu D, Zhang T, Qin L, Lu H. Combined application of low-intensity pulsed ultrasound and functional electrical stimulation accelerates bone-tendon junction healing in a rabbit model. *J Orthop Res* 2014;32(2):204–9. [PubMed: 24136665]
37. Lovric V, Ledger M, Goldberg J, Harper W, Bertollo N, Pelletier MH, Oliver RA, Yu Y, Walsh WR. The effects of low-intensity pulsed ultrasound on tendon-bone healing in a transosseous-equivalent sheep rotator cuff model. *Knee Surg Sports Traumatol Arthrosc* 2013;21(2):466–75. [PubMed: 22466014]
38. Lu H, Qin L, Fok P, Cheung W, Lee K, Guo X, Wong W, Leung K. Low-intensity pulsed ultrasound accelerates bone-tendon junction healing: a partial patellectomy model in rabbits. *Am J Sports Med* 2006;34(8):1287–96. [PubMed: 16567453]
39. Qin L, Lu H, Fok P, Cheung W, Zheng Y, Lee K, Leung K. Low-intensity pulsed ultrasound accelerates osteogenesis at bone-tendon healing junction. *Ultrasound Med Biol* 2006;32(12):1905–11. [PubMed: 17169702]
40. Walsh WR, Stephens P, Vizesi F, Bruce W, Huckle J, Yu Y. Effects of low-intensity pulsed ultrasound on tendon-bone healing in an intra-articular sheep knee model. *Arthroscopy* 2007;23(2):197–204. [PubMed: 17276228]
41. Lu H, Qin L, Cheung W, Lee K, Wong W, Leung K. Low-intensity pulsed ultrasound accelerated bone-tendon junction healing through regulation of vascular endothelial growth factor expression and cartilage formation. *Ultrasound Med Biol* 2008;34(8):1248–60. [PubMed: 18378382]
42. Naito K, Watari T, Muta T, Furuhashi A, Iwase H, Igarashi M, Kurosawa H, Nagaoka I, Kaneko K. Low-intensity pulsed ultrasound (LIPUS) increases the articular cartilage type II collagen in a rat osteoarthritis model. *J Ortho Res* 2010;28(3):361–9.
43. Wood VT, Pinfildi CE, Neves MAI, Parizoto NA, Hochman B, Ferreira LM. Collagen changes and realignment induced by low-level laser therapy and low-intensity ultrasound in the calcaneal tendon. *Lasers Surg Med* 2010;42:559–65. [PubMed: 20662033]
44. Papatheodorou LK, Malizos KN, Poultsides LA, Hantes ME, Grafanaki K, Giannouli S, Ioannou MG, Koukoulis GK, Protopappas VC, Fotiadis DI, Stathopoulos C. Effect of transosseous application of low-intensity ultrasound at the tendon graft-bone interface healing: gene expression and histological analysis in rabbits. *Ultrasound Med Biol* 2009;35(4):576–84. [PubMed: 19185973]
45. Lu CC, Liu YC, Cheng YM, Chih TT, Tien YC. Augmentation of tendon-bone interface healing with low-intensity pulsed ultrasound. *Orthopedic* 2009;32(3):173.
46. Murtaugh B, Ihm JM. Eccentric training for the treatment of tendinopathies. *Curr Sports Med Rep* 2013;12(3):175–82. [PubMed: 23669088]
47. Best TM, Moore B, Jarit P, Moorman CT, & Lewis GK (2015). Sustained acoustic medicine: wearable, long duration ultrasonic therapy for the treatment of tendinopathy. *The Physician and sportsmedicine*, 43(4), 366–374. [PubMed: 26468991]

Figure 1:
Wearable LITUS devices designed for daily home use: (A) SAM® Sport Sustained Acoustic Medicine System and (B) Exogen® Bone Healing System