The XL probe: A luxury or a necessity? Risk stratification in an obese community cohort using transient elastography

United European Gastroenterology Journal
2018, Vol. 6(9) 1372–1379
© Author(s) 2018
Article reuse guidelines:
sagepub.com/journals-permissions
DOI: 10.1177/2050640618772944
journals.sagepub.com/home/ueg

Rebecca Harris¹ , Timothy R Card², Toby Delahooke³, Guruprasad P Aithal^{1,4}
and I Neil Guha^{1,4}

Abstract

Background: Transient elastography is a non-invasive tool which can stratify patients at risk of chronic liver disease. However, a raised body mass index has been independently associated with a failed or unreliable examination.

Objective: The purpose of this study was to analyse the performance of two probes (M/XL) on a portable transient elastography device within an obese community population.

Method: The method involved a prospective study with recruitment from a primary care practice. Patients identified with a risk factor for chronic liver disease were invited to a community-based risk stratification pathway for transient elastography readings with both probes. A threshold of ≥ 8.0 kPa defined elevated liver stiffness.

Results: A total of 477 patients attended the pathway. Of the patients, 21% had no valid measurements with the M probe. There was a significant difference between the probes in the proportion achieving ≥ 10 valid readings (M versus XL probe: 66.2% versus 90.2%; $p \leq 0.001$) and in their reliability (M versus XL probe: 77.4% versus 98.5%; $p = 0.028$). Unreliable readings with the M probe increased as the body mass index increased. The XL probe re-stratified 5.2% of patients to have a normal reading.

Conclusion: The XL probe on a portable device significantly improves the applicability of transient elastography within a community-based risk stratification pathway.

Keywords

Transient elastography, risk stratification, obesity, community, chronic liver disease

Received: 7 February 2018; accepted: 31 March 2018

Key summary

Summarise established knowledge

1. Transient elastography (TE) is an extensively validated diagnostic tool for stratifying the severity of liver disease.
2. However, a raised body mass index has been independently associated with a failed or unreliable examination, a potential limitation when using transient elastography in a community-based risk stratification pathway.
3. The XL probe has been developed specifically for obese patients and with it now available on the portable device this could increase the applicability of transient elastography as a risk stratification tool.

¹National Institute for Health Research (NIHR) Nottingham Biomedical Research Centre, UK

²Division of Epidemiology and Public Health, University of Nottingham, UK

³Leicester Royal Infirmary, University Hospitals of Leicester NHS Trust, UK

⁴Nottingham Digestive Diseases Centre, University of Nottingham, UK

Corresponding author:

I Neil Guha, E Floor, West Block Queens Medical Centre, Derby Road, Nottingham NG7 2UH, UK.

Email: neil.guha@nottingham.ac.uk

4. Considering its increasing utilisation in clinical care, there remains a paucity of data on the relationship between the M and XL probe when used on the same patient.

Significant/new findings

1. This is the first study which has demonstrated the feasibility and success of using both probes in the community within a risk stratification pathway.
2. Use of the XL probe significantly increased the number of valid (M versus XL probe: 66.2% versus 90.2%) and reliable readings (M versus XL probe: 77.4% versus 98.5%) that were obtained.
3. Use of the correct probe is essential to ensure the patient is risk stratified correctly. The XL probe re-stratified 5.2% of patients to have a normal TE reading according to our definition for clinically significant liver disease.
4. Within a risk stratification pathway, the XL probe is not an optional extra, but a necessity in a population setting where a raised BMI is becoming routine.

Introduction

Chronic liver disease (CLD) has risen up the public health agenda due to increasing mortality rates and the preventable burden of disease within the general population caused in the majority by lifestyle related risk factors.¹

Case finding strategies to tackle these issues have been suggested^{2,3} but are not routinely implemented. Yet, use of non-invasive tests such as transient elastography (TE) as a risk stratification tool for liver disease have previously been demonstrated to be successful.^{4,5} Advantages of TE include its ease of use, provision of timely results, diagnostic accuracy and that it is non-invasive. The convenience of a portable machine also allows risk stratification to occur within the community rather than a hospital setting.⁶

However, a raised body mass index (BMI) has been independently associated with a failed or unreliable TE examination using the standard 'M' probe⁷⁻⁹ with successful readings reported in only 75% of obese patients (BMI ≥ 30 kg/m²).¹⁰ This is a potential limitation when using TE in a risk stratification pathway. This is concerning given the increasing proportion of the general population who are overweight (65% of men and 58% of women in England).¹¹ The XL probe has been developed specifically for obese patients and now that it is available on the portable device this could increase the applicability of TE within a risk stratification pathway.

The aim of this study was to analyse the performance of the M and XL TE probes among those with a BMI ≥ 28 kg/m² within a risk stratification pathway based in the community.

Methods

Study setting and population

This is a prospective study with recruitment from a primary care practice in Leicester, England. The study ran from January 2015–March 2016. Local regulatory

approval was obtained on 10 April 2013 from the Leicester Research Ethics Committee (13/EM/0123) and written informed consent was gained from each patient included in the study. The study protocol conforms to the ethical guidelines of the 1975 Declaration of Helsinki as reflected in a prior approval by the institution's human research committee. The study has been registered on the ClinicalTrials.gov website (NCT02037867).

Clinical, anthropometric and biochemical data was obtained from the electronic primary care records (SystmOne, TPP, UK) within which data is stored as searchable numerical values and prospectively coded 'Read codes'.

Recruitment occurred via an invitation to attend a community-based risk stratification pathway for CLD which has been previously described.⁶ Patients were initially identified from the electronic primary care records. Adults (≥ 18 years) with one or more lifestyle related risk factors for CLD at the start of the study were invited. These included:

1. Hazardous alcohol use – defined as >14 units/week for women and >21 units/week for men, or an Alcohol use disorders identification test (AUDIT) questionnaire score ≥ 8 ,¹² or presence of a Read code for alcohol misuse.
2. Type 2 diabetes – presence of a Read code related to the diagnosis.
3. BMI ≥ 28.0 kg/m² – presence of a numerical value recorded within the past five years.

A lower BMI cut-off for obesity (≥ 28.0 kg/m²) was agreed a priori for all patients included within the study due to the increased prevalence of people with Asian ethnicity in this population. This is in accordance with the World Health Organisation¹³ who recommend different cut-off points for the Asian population due to a higher risk of type 2 diabetes and cardiovascular disease at a lower BMI compared with European populations. A lower cut-off ensured that patients of Asian

ethnicity whose BMI was lower than the traditional cut-off for obesity (30.0 kg/m^2) but who were still at high risk of CLD were invited to attend.

Patients meeting any of the following criteria were ineligible and not invited to attend the risk stratification pathway: (a) contraindication to undertaking a TE reading (e.g. pregnancy, implantable cardiac device); (b) known diagnosis of CLD; (c) known malignancy or other terminal illness; (d) patients unable to consent to investigation or housebound and unable to attend.

TE

TE is a non-invasive diagnostic test which calculates the degree of liver stiffness by propagation of an elastic shear wave. Correlation between its results and stages of liver fibrosis has been extensively evaluated and validated in all major aetiologies of CLD.¹⁴ The TE device can be either static or portable but until recently the XL probe has only been available with the static device. Three experienced operators performed all the TE examinations as per the manufacturer's recommendations and had completed >100 examinations prior to the study using the portable FibroScan 402 device (Echosens, Paris, France). The technique for obtaining a TE reading and the differences between the probes has been previously described.^{15,16} Ten valid measurements were collected with both probes with the median value reported as the liver stiffness measurement in kilopascals (kPa). Examinations unable to record any valid readings were deemed technical failures. As an indicator of variability the ratio of the interquartile (IQR) range of the liver stiffness to the median (M) value (IQR/M) was also recorded. Examinations were compared with the reliability criteria outlined by Boursier et al.¹⁷ and recommended by the manufacturer. To avoid confusion all examinations which are 'very reliable' or 'reliable' will be referred to in the subsequent text as reliable.

A TE reading was attempted with both probes for all patients with a $\text{BMI} \geq 28.0 \text{ kg/m}^2$. A threshold of $\geq 8.0 \text{ kPa}$ was agreed a priori to define elevated liver stiffness consistent with clinically significant liver disease. This threshold has been used within other community-based screening programmes⁴ and has been demonstrated to have a high negative predictive value for advanced fibrosis.¹⁰

Statistical methods

Statistical analysis was completed using Stata version 14.2 (StataCorp LP). Baseline characteristics (Table 1) of the study cohort are presented as numbers (percentage) if categorical data or M (IQR) for non-normally distributed continuous data. A comparison of the

Table 1. Baseline characteristics of the study cohort ($n=477$).

Gender – male, n (%)	239 (50.1%)
Age ^a	58 (47–68)
Ethnicity, n (%)	
White	346 (72.5%)
Asian	96 (20.1%)
Black	19 (4.0%)
Other	16 (3.4%)
Risk factor, n (%)	
$\text{BMI} \geq 28.0 \text{ kg/m}^2$ only	344 (73.7%)
Hazardous alcohol use + $\text{BMI} \geq 28.0$	51 (10.9%)
Type 2 diabetes + $\text{BMI} \geq 28.0$	72 (15.4%)
All 3 risk factors	10 (2.1%)
Metabolic risk factors, n (%)	
Hypertension	192 (40.3%)
Hyperlipidaemia	112 (23.5%)
$\text{BMI} (\text{kg/m}^2)$	
28–29.9	154 (32.3%)
30–34.9	207 (43.4%)
35–39.9	73 (15.3%)
≥ 40	42 (8.8%)
Ischaemic heart disease, n (%)	37 (7.8%)
ALT U/l ^a	26 (19–36)
Platelets $10^9/\text{l}^a$	247.5 (205–290)

ALT: alanine aminotransferase; BMI: body mass index; IQR: interquartile range.

^aMedian (IQR).

performance of both probes was made using the chi-squared test and the Wilcoxon signed rank test for categorical and non-normally distributed continuous data respectively.

The difference in the number of reliable and unreliable readings between the two probes using the criteria outlined by Boursier et al.¹⁷ was compared using the chi-squared test. Readings between the two probes were also compared in accordance with how this would affect risk stratification. A patient was considered to be re-stratified if the XL probe reading was $<8 \text{ kPa}$ when the M probe reading had been $\geq 8 \text{ kPa}$ (in line with our definition of clinically significant liver disease). Correlation between the liver stiffness measurements obtained by both probes was calculated and a linear regression analysis was completed to further characterise this relationship. Multivariable regression analysis was carried out to estimate the effect of potential confounding variables. Agreement between the probes was further analysed using a Bland-Altman plot. To identify variables independently associated with re-stratification, univariate and multivariate logistic regression models including the covariates age, gender, BMI, hypertension, hypercholesterolaemia, the reliability of the M probe reading,

Table 2. A comparison in the performance of the M and XL transient elastography probes.

Characteristic	M probe	XL probe	<i>p</i> Value
≥10 valid measurements	316	430 (90.2%)	<0.001
Reliable LSM (minimum 10 valid measurements and IQR/ M < 0.3 only if LSE > 7.1 kPa)	312 (65.4 %)	425 (89.1%)	<0.001
Reliable LSM ^a (very reliable – IQR/M ≤ 0.1; reliable – 0.1 < IQR/M ≤ 0.3 or IQR/M > 0.3 with LSM < 7.1 kPa; poorly reliable – IQR/M > 0.3 with LSM ≥ 7.1 kPa)	369 (77.4%)	470 (98.5%)	0.028
Median success rate ^b – includes unreliable readings	100% (91–100%)	100% (100–100%)	<0.001
Median IQR/M ^b	13% (9–19%)	15% (10–23%)	0.008
Median liver stiffness ^b	5.3 (4.2–6.7)	4.9 (3.9–6.2)	<0.001

IQR: interquartile range; kPa: kilopascals; LSM: liver stiffness measurement; LSE: liver stiffness evaluation.

^aBoursier et al.¹⁷

^bMedian (IQR).

type 2 diabetes and hazardous alcohol use as risk factors were conducted.

Results

Patient characteristics

The primary care practice had a total adult population of 4150 with 1167 patients identified to have at least one risk factor and eligible to be invited to attend the risk stratification pathway. Of these, 720 patients attended of which 477 had a BMI ≥ 28.0 kg/m² and had TE readings attempted with both probes (patient characteristics outlined in Table 1). Fifty per cent of the patients were male and the median age was 58 years (IQR 47–68). The majority of the patients were white (72.5%) although this was lower than the general population (87.2% in the UK) due to the high percentage of patients with Asian ethnicity (24.6%) in the community in which the risk stratification pathway was implemented. Seventy-three per cent of the cohort had a BMI ≥ 28.0 kg/m² as their only risk factor whilst 10.9% and 15.4% also had hazardous alcohol use or type 2 diabetes as an additional risk factor; 2.1% of patients had all three risk factors. The median BMI was 31.4 (IQR 29.4–34.7).

Reliability of the probes

The XL probe increased the number of valid TE readings (Table 2); 21% of the patients had no valid measurements with the M probe. There was a significant difference in the proportion who had ≥10 valid

Table 3. Reliability of the M versus XL probe using Boursier criteria.

		XL	
		Unreliable	Reliable
M	Unreliable	4/477 (0.8%)	104/477 (21.8%)
	Reliable	3/477 (0.6%)	366/477 (76.7%)

All patients (*n* = 477), *p* value = 0.028.

readings between the M and XL probes (66.2% versus 90.2%; *p* < 0.001); 76% of the patients with <10 valid readings with the M probe had ≥10 readings with the XL probe.

A significant difference was also seen in the reliability with >96% of patients with an unreliable M probe reading obtaining a reliable XL probe reading (Table 3). According to the reliability criteria¹⁷ only 0.84% of the patients did not obtain a reliable reading with either probe (Table 3). The number of reliable measurements increased with the XL probe and a significant difference was observed between the two probes (M versus XL probe: 77.4% versus 98.5%; *p* = 0.028). The number of unreliable readings increased with the M probe as the BMI increased whilst with the XL probe the number of unreliable readings remained low across all BMI categories.

The TE readings between the probes were highly correlated (*R*² = 0.78, *p* value < 0.001) (Figure 1). A Bland-Altman plot (Figure 2) demonstrated this and revealed a larger difference at higher mean values (Pitman's test of

Figure 1. Correlation between the M and XL probes.

Figure 2. Bland Altman plot of the difference between the M and XL probe liver stiffness measurements ($n = 377$). The solid horizontal line equates to the mean difference between the two probes (0.82 kPa 95% confidence interval (CI) 0.58–1.08) and the dotted lines the 95% limits of agreement (−4.14 to 5.79 kPa).

difference in variance: $r = 0.656$, p value ≥ 0.001). In multivariable analysis no appreciable confounding of this relationship was found with any of the studied variables (BMI, gender, ethnicity, age, type 2 diabetes as a risk factor, hazardous alcohol use as a risk factor). In general, the XL probe readings were lower than those obtained with the M probe with linear regression analysis returning the following estimate:

$$XL = (0.59 \times M) + 1.73$$

Table 4. Risk stratification of liver disease using the M and XL probes for all patients ($n = 477$). Includes unreliable readings.

		XL probe	
		Normal	Clinically significant liver disease
M probe	No valid readings	85 (17.8%)	15 (3.2%)
	Normal	320 (67.1%)	1 (0.2%)
	Clinically significant liver disease	25 (5.2%)	31 (6.5%)

Use of both probes within the community risk stratification pathway

The XL probe ensured 21% of patients obtained a TE reading who otherwise would have been deemed a technical failure if only the M probe was available. The XL probe also re-stratified 5.2% of patients to have a normal TE reading according to our definition for clinically significant liver disease (Table 4). The percentage of patients re-stratified increased as the BMI increased. BMI was the only variable in a multivariate logistic regression which was significantly associated with re-stratification. For every 1 kg/m^2 increase in BMI the odds of being re-stratified increase by 19% (Table 5).

Discussion

Key findings

Within a risk stratification pathway, an ideal tool would reliably identify patients at high risk of disease and exclude those who are normal. However, as we have demonstrated use of TE with only the M probe as a risk stratification tool in an obese cohort could potentially lead to a large number of patients with an invalid or unreliable TE reading.

Despite 93.1% of patients having a TE reading with both probes being risk stratified equivalently, one in five patients within this obese community cohort had no valid readings with the M probe. Use of the XL probe significantly improved the number of valid and reliable TE readings that were obtained.

Linear regression analysis suggests there is a good correlation between the probes with the readings from the XL probe relating to that of the M probe in line with the equation

$$XL = (0.59 \times M) + 1.73$$

The XL probe readings are lower than the M probe which is consistent with other findings in the

Table 5. Univariate and multivariate logistic regression analysis of the variables associated with re-stratification. Includes unreliable readings.

Variable	Univariate analysis		Multivariate analysis ^a	
	Odds ratio (95% CI)	<i>p</i> Value	Odds ratio (95% CI)	<i>p</i> Value
Age	1.01 (0.98–1.03)	0.691		
Gender	0.98 (0.43–2.20)	0.957		
BMI	1.17 (1.08–1.26)	<0.001	1.196 (1.10–1.30)	<0.001
Type 2 diabetes as a risk factor for CLD	3.17 (1.33–7.55)	0.009		
Hazardous alcohol use as a risk factor for CLD	2.15 (0.82–5.68)	0.121		
Hypertension	2.99 (1.29–6.98)	0.011		
Hypercholesterolaemia	2.35 (1.03–5.39)	0.043		
Reliability of M probe reading ^b	0.23 (0.05–1.19)	0.080		

BMI: body mass index; CLD: chronic liver disease.

^aAdjusted for age/gender/ethnicity.

^bAccording to Boursier criteria.

literature.^{16,18} The Bland-Altman plot demonstrates this difference when using both probes in the same individual and that this difference is larger the greater the mean reading.

Strengths and limitations

Whilst other studies^{16,19} have compared both probes in a hospital setting, we have demonstrated the feasibility of using both probes in a community risk stratification pathway. Ten valid measurements were obtained in 66.2% of patients using the M probe and in 90.2% of patients with the XL probe. This performance compares favourably with other studies^{16,19} despite being used in a community rather than a hospital setting.

The percentage of reliable readings is lower than other studies which have risk stratified using TE in the community (Baba et al.²⁰ 98.3%; You et al.²¹ 97%). This is likely due to the increased number of patients with obesity in our cohort compared with the unselected general population used within these studies. This highlights the importance of having access to the XL probe in order to maximise the numbers of patients who could be risk stratified.

A limitation is the inability to comment on the diagnostic performance of TE within this population and healthcare setting. We were unable to compare the two TE probe readings against histological findings although a liver biopsy does have its own documented limitations.²² Myers et al.¹⁶ analysed the performance of both TE probes against histological outcomes in an overweight cohort, and identified optimal liver stiffness cut-offs with the M probe (≥ 7.8 kPa) and the XL probe (≥ 6.4 kPa) for diagnosing $\geq F2$ fibrosis. It is worth noting that according to our equation, conversion of

an M probe cut-off of 7.8 kPa yields an XL probe cut-off of 6.3 kPa which is remarkably close to the value derived by Myers et al.¹⁶

However, it has been reported that the disparity between the two probes is eliminated when the M probe is re-calibrated to measure an equivalent depth to the XL probe.¹⁶ Similarly, when the stiffness is made homogeneous using phantom liver models the difference between the probe readings disappears.¹⁶ Histological specimens demonstrate a greater amount of fibrous tissue around the subcapsular region which can be out of proportion to the remainder of the parenchyma.²³ This suggests that the lower readings observed with the XL probe may be due to a deeper area of the liver being scanned. A recent update of the device automatically selects the ‘appropriate’ probe based on skin capsular distance (SCD). Our data shows the difference in readings which might occur if this guidance is not followed and consequently the potential error in risk stratification of patients.²⁴ Thus, choosing the correct probe is not for the sole reason of obtaining a reliable scan. Though our linear regression equation shows a clear relationship numerically between results from the two probes, it does not imply that one probe can be substituted for another. The relationship between the probes at higher stages of fibrosis may differ, as highlighted by the Bland-Altman plot.

At present, it is assumed that the same liver stiffness cut-off can be used if the ‘appropriate’ probe has been chosen. A validation study of the XL probe is required to ensure this assumption is correct and that the liver stiffness measurement obtained is reliable and correlates with the histological diagnosis. As we were unable to measure SCD and did not have any histological comparisons, we are unable to validate the

XL probe in this study but are aware of emerging evidence.

Relevance to clinical practice

There is an urgent need to tackle the growing epidemic of CLD and actively identify those patients who are at increased risk of liver-related disease and death.²⁵ With new guidelines recommending screening for patients at high risk of CLD,³ tools which are accessible and easily utilised are imperative. Our results suggest that a valid and reliable M probe reading is unlikely to underestimate the degree of liver fibrosis and consequently should not mis-stratify patients who have undiagnosed CLD. This is supported by the close correlation observed between the results from the two probes. However, access to the M probe alone would limit the applicability of TE within a community-based risk stratification pathway. The adjunct of the XL probe with the portable device is an important advance to this technology to achieve a non-invasive test which can be universally applied to all patients at risk.²⁶

Whilst the added cost to a healthcare commissioner may be of concern, this must be weighed up against the cost of a false-positive result, a screening failure or even the cost to the healthcare system as a whole if the patient remained undiagnosed. Health economic analyses of proposed case finding strategies would be useful to determine the added economic value of having the XL probe available on a portable TE device.

Even if the XL probe is accessible, there still remains uncertainty over when it should be utilised to ensure that a reliable reading is obtained. The manufacturer recommends measuring the SCD to determine which probe to use but this is not always possible in a community setting. This was indeed the case in our current study. Other studies have demonstrated a patient's BMI to be a reasonable surrogate for SCD. Our results demonstrate that 25% of patients with a BMI > 30 kg/m² had an unreliable reading with the M probe. This therefore could be a practical threshold in which the XL probe should be considered ahead of the M probe. Alternatively, the M probe may soon become redundant if the XL probe is able to provide more reliable readings in a general population who are increasingly overweight and at risk of CLD.

Lastly, assuming that the same liver stiffness cut-off can be used for both probes, identification of the most appropriate probe would also ensure that the patient is risk stratified correctly. Within this study cohort alone, one in 20 patients may have been risk stratified incorrectly. Importantly in the multivariable logistic regression, BMI was the only variable which was significantly

associated with re-stratification. Thus, the XL probe is now not an optional extra but a necessity in a population setting where obesity is becoming routine.

Conclusion

A reliable and valid M probe reading is unlikely to mis-stratify patients within a community-based risk stratification pathway. The addition of the XL probe optimises the applicability of TE within a case finding strategy by significantly increasing the number of patients with reliable readings and ensuring they are risk stratified correctly, particularly in an obese cohort. Clinicians should be aware of the additional benefits the XL probe can offer in a population where obesity prevalence is increasing.

Acknowledgements

The following author contributions were made: R Harris, TR Card, T Delahooke, GP Aithal and IN Guha were involved in the study design and concept, implementation of the study in primary care, interpretation of results and editing of the manuscript. R Harris analysed the data set and wrote the initial manuscript draft. All authors had full access to the data in the study and can take responsibility for the integrity of the data and the accuracy of the data analysis. IN Guha is the guarantor. This paper presents independent research supported by the National Institute for Health Research (NIHR). The views expressed are those of the authors and not necessarily those of the National Health Service, the NIHR or the Department of Health.

Declaration of conflicting interests

All authors declare that they are free from other sources of external funding related to this study.

Funding

Echosens provided a loan of the XL probe at the start of the study. No financial assistance was provided and Echosens had no role in the study design, the collection or interpretation of the data. Funding for the study was provided by (a) the Nottingham Digestive Diseases Centre and National Institute for Health Research (NIHR) Nottingham Biomedical Research Centre part of the Nottingham University Hospitals NHS Trust and University of Nottingham and (b) the East Midlands Academic Health Sciences Network (EMAHSN). The study sponsor is the University of Nottingham which is the data custodian but had no role in the design, analysis or interpretations of the data.

Ethics approval

Leicester Research Ethics Committee (13/EM/0123).

Informed consent

All patients provided written informed consent.

ORCID iD

Rebecca Harris <http://orcid.org/0000-0002-7863-7701>

References

1. Blachier M, Leleu H, Peck-Radosavljevic M, et al. The burden of liver disease in Europe: A review of available epidemiological data. *J Hepatol* 2013; 58: 593–608.
2. EASL-ALEH clinical practice guidelines: Non-invasive tests for evaluation of liver disease severity and prognosis. *J Hepatol* 2015; 63: 237–264.
3. EASL-EASD-EASO clinical practice guidelines for the management of non-alcoholic fatty liver disease. *J Hepatol* 2016; 64: 1388–1402.
4. Roulot D, Costes JL, Buyck JF, et al. Transient elastography as a screening tool for liver fibrosis and cirrhosis in a community-based population aged over 45 years. *Gut* 2011; 60: 977–984.
5. Moessner BK, Jorgensen TR, Skamling M, et al. Outreach screening of drug users for cirrhosis with transient elastography. *Addiction* 2011; 106: 970–976.
6. Harman DJ, Ryder SD, James MW, et al. Direct targeting of risk factors significantly increases the detection of liver cirrhosis in primary care: A cross-sectional diagnostic study utilising transient elastography. *BMJ Open* 2015; 5: e007516.
7. Castera L, Foucher J, Bernard PH, et al. Pitfalls of liver stiffness measurement: A 5-year prospective study of 13,369 examinations. *Hepatology* 2010; 51: 828–835.
8. Foucher J, Castera L, Bernard PH, et al. Prevalence and factors associated with failure of liver stiffness measurement using FibroScan in a prospective study of 2114 examinations. *Eur J Gastroenterol Hepatol* 2006; 18: 411–412.
9. Tapper EB, Challies T, Nasser I, et al. The performance of vibration controlled transient elastography in a US cohort of patients with nonalcoholic fatty liver disease. *Am J Gastroenterol* 2016; 111: 677–684.
10. Wong VW, Vergniol J, Wong GL, et al. Diagnosis of fibrosis and cirrhosis using liver stiffness measurement in nonalcoholic fatty liver disease. *Hepatology* 2010; 51: 454–462.
11. Statistics team, NHS Digital. Health and Social Care Information Centre. Statistics on obesity, physical activity and diet. Government Statistical Service, NHS Digital.
12. Saunders JB, Aasland OG, Babor TF, et al. Development of the Alcohol Use Disorders Identification Test (AUDIT): WHO Collaborative Project on Early Detection of Persons with Harmful Alcohol Consumption-II. *Addiction* 1993; 88: 791–804.
13. World Health Organisation expert consultation. Appropriate body-mass index for Asian populations and its implications for policy and intervention strategies. *Lancet* 2004; 363: 157–163.
14. Friedrich-Rust M, Ong MF, Martens S, et al. Performance of transient elastography for the staging of liver fibrosis: A meta-analysis. *Gastroenterology* 2008; 134: 960–974.
15. Sandrin L, Fourquet B, Hasquenoph JM, et al. Transient elastography: A new noninvasive method for assessment of hepatic fibrosis. *Ultrasound Med Biol* 2003; 29: 1705–1713.
16. Myers RP, Pomier-Layrargues G, Kirsch R, et al. Feasibility and diagnostic performance of the FibroScan XL probe for liver stiffness measurement in overweight and obese patients. *Hepatology* 2012; 55: 199–208.
17. Boursier J, Zarski JP, de Ledinghen V, et al. Determination of reliability criteria for liver stiffness evaluation by transient elastography. *Hepatology* 2013; 57: 1182–1191.
18. de Ledinghen V, Wong VW, Vergniol J, et al. Diagnosis of liver fibrosis and cirrhosis using liver stiffness measurement: Comparison between M and XL probe of FibroScan(R). *J Hepatol* 2012; 56: 833–839.
19. Friedrich-Rust M, Hadji-Hosseini H, Kriener S, et al. Transient elastography with a new probe for obese patients for non-invasive staging of non-alcoholic steatohepatitis. *Eur Radiol* 2010; 20: 2390–2396.
20. Baba M, Furuya K, Bandou H, et al. Discrimination of individuals in a general population at high-risk for alcoholic and non-alcoholic fatty liver disease based on liver stiffness: A cross section study. *BMC Gastroenterol* 2011; 11: 70.
21. You SC, Kim KJ, Kim SU, et al. Factors associated with significant liver fibrosis assessed using transient elastography in general population. *World J Gastroenterol* 2015; 21: 1158–1166.
22. Ratzu V, Charlotte F, Heurtier A, et al. Sampling variability of liver biopsy in nonalcoholic fatty liver disease. *Gastroenterology* 2005; 128: 1898–1906.
23. Standish RA, Cholongitas E, Dhillon A, et al. An appraisal of the histopathological assessment of liver fibrosis. *Gut* 2006; 55: 569–578.
24. Eddowes P, Sasso M, Fournier C, et al. Steatosis and liver stiffness measurements using transient elastography. *Hepatology* 2016; 64: 700.
25. Ekstedt M, Hagstrom H, Nasr P, et al. Fibrosis stage is the strongest predictor for disease-specific mortality in NAFLD after up to 33 years of follow-up. *Hepatology* 2015; 61: 1547–1554.
26. Vuppalanchi R, Siddiqui MS, Van Natta ML, et al. Performance characteristics of vibration-controlled transient elastography for evaluation of nonalcoholic fatty liver disease. *Hepatology* 2018; 67: 134–144.