Validation of foot pitch angle estimation using inertial measurement unit against marker-based optical 3D motion capture system

Shiva Sharif Bidabadi¹ · Iain Murray² · Gabriel Yin Foo Lee^{3,4}

Received: 4 April 2018 / Revised: 1 May 2018 / Accepted: 9 May 2018 / Published online: 17 May 2018
© Korean Society of Medical and Biological Engineering and Springer-Verlag GmbH Germany, part of Springer Nature 2018

Abstract

Gait analysis is relevant to a broad range of clinical applications in areas of orthopedics, neurosurgery, rehabilitation and the sports medicine. There are various methods available for capturing and analyzing the gait cycle. Most of gait analysis methods are computationally expensive and difficult to implement outside the laboratory environment. Inertial measurement units, IMUs are considered a promising alternative for the future of gait analysis. This study reports the results of a systematic validation procedure to validate the foot pitch angle measurement captured by an IMU against Vicon Optical Motion Capture System, considered the standard method of gait analysis. It represents the first phase of a research project which aims to objectively evaluate the ankle function and gait patterns of patients with dorsiflexion weakness (commonly called a “drop foot”) due to a L5 lumbar radiculopathy pre- and post-lumbar decompression surgery. The foot pitch angle of 381 gait cycles from 19 subjects walking trails on a flat surface have been recorded throughout the course of this study. Comparison of results indicates a mean correlation of 99.542% with a standard deviation of 0.834%. The maximum root mean square error of the foot pitch angle measured by the IMU compared with the Vicon Optical Motion Capture System was 3.738° and the maximum error in the same walking trail between two measurements was 9.927°. These results indicate the level of correlation between the two systems.

Keywords Foot drop · Inertial measurement unit · Pitch angle · Gait analysis

1 Introduction

Gait analysis has received significant interest in orthopedics and rehabilitation applications [1]; it has important implications in many clinical settings. The term ‘Gait analysis’ refers to the systematic study of animal

locomotion [2], specifically the study of human motion, using equipments to measure body movements and the activity of the muscles [2]. It involves measurement and analysis of measurable parameters such as absolute and relative angles, positions, movement patterns and joints’ range of motion. However, monitoring and measuring the human gait characteristics is a sophisticated task and thus, still rarely used for routine clinical evaluation [3].

There are three major commonly accepted methods available for gait analysis. The first and most common method is visual inspection. Second method is goniometric measurements, which is performed by goniometer that provides the range of motion in joints. Third is 3-dimensional marker-based optical motion capture (MBMOCAP) system which is an optoelectronic gait analysis method using optical motion capture for real-time 3-dimensional motion analysis [4] and considered as the standard method of movement analysis [5, 6]. This system utilizes retro-reflective markers that allows visualization of multiple body

✉ Shiva Sharif Bidabadi
Shiva.Sharif@curtin.edu.au

Iain Murray
I.Murray@curtin.edu.au

Gabriel Yin Foo Lee
glee.neurosurgery@bigpond.com

¹ Department of Mechanical Engineering, Curtin University of Technology, Perth, Australia

² Department of Electrical and Computer Engineering, Curtin University of Technology, Perth, Australia

³ St John of God Subiaco Hospital, Perth, Australia

⁴ School of Surgery, University of Western Australia, Perth, Australia

regions by set of cameras and is highly reliable and accurate [5].

Current gait analysis methods are subject to various limitations [7]. For instance, relying on visual ability and diagnosis via observation of movements always governed by the medical practitioners' experience and judgment, consequently they are inaccurate and subjective [3]. Additionally, goniometer most often used as a measure for static and passive movement (i.e. not during walking period). Moreover, MBMOCAP systems are usually complex, expensive and laboratory based method which requires software experts to operate the system and apply an inverse kinematic (with a complex exoskeletal model) to get joint angles. Thus its usage is extraordinarily time consuming and not practical for widespread usage in clinical environments [3].

Accurate gait analysis can be done using MBMOCAP, however, there is often a need for gait analysis to be accompanied outside of the laboratory environment, because not only in-laboratory walking varies from everyday typical walking but also gait laboratories do not lend themselves to portability [3].

To overcome these limitations new methodologies have been proposed recently using Inertial Measurement Units (IMU) [2]. In the past few years, several researches have used different types of sensors to identify gait characteristics and events [8, 9, 10, 11]. These studies applied and evaluated a variety of algorithms and sensor attachment locations [12]. For example, angular signals computed from gyroscope attached to the shank [8] and in a similar study, gyroscopes used to recognize normal walking from stair climbing [13]. Additionally, combinational usage of different sensor types suggested in earlier researches evaluating gait events using accelerometer and gyroscope readings for different walking speeds [14]. For example, it has been stated that the accuracy of gait evaluation systems increased by using gyroscopes and accelerometers [15]. The application of these sensors could expand the understanding of Equinus gait characteristics and also allows for an automatic detection of gait events [16]. Equinus defined as congenital deformity of the foot that leads to the walking on the toes without the heel contact to the ground. It is also proposed that the usage of gyroscope and accelerometer as a compact gait-analysis system, for monitoring daily movements in people with a walking disorder, is advantageous in clinical environment [16]. Also these sensor have been widely used in gait-based navigation systems [17, 18].

An IMU consists of accelerometers, gyroscopes and magnetometer sensors in each axis (x, y and z) integrated into single unit [19, 20].

The combination of these three sensors in a single device provides the measurements of angular velocity and

acceleration in the sensor body frame. These, accompanied by micro-controllers, batteries and wireless communication, results in long-term, portable recordings of ambulatory measurements [3].

The case study of the proposed IMU gait analysis method will be in the field of neurosurgery/spinal surgery. One of the key goals of the study is to characterize and objectively document the walking gait cycle of patients with foot drop both pre- and post-lumbar spinal surgery. The patients recruited have a specific lumbar spine pathology whereby the clinical examination findings and imaging studies indicate a compressive lesion of the L5 lumbar nerve root, where possible, the majority of patients have also undertaken confirmatory electrophysiological testing. Patients with a foot drop often develop a characteristic gait disturbance. The proposed IMU solution offers evaluation of these disturbances. In this specific field the level of accuracy is important in assisting specialist in early intervention, quantitative assessment of impairment and indication of the next level of treatment, e.g. surgery [19]. Foot drop is commonly identified based on visual inspection and regardless of the significance of the gait measurements, currently there is no base-line as the accuracy required for foot drop identification.

The purpose of this study is to validate the gait cycle data accuracy captured by IMU sensors in comparison to a MBMOCAP system as the reference. In spite of MBMOCAP's mentioned disadvantages, it is a proven and accepted method with a very high accuracy for the capture and analysis of human gait [5, 21].

The following section provides details of the experiment, followed by the methodologies that have been used to capture and analyze data with IMU and MBMOCAP system. The results of the comparison, conclusion and future works documented and discussed.

2 Experimental design

For validation purposes, an experiment was designed to capture the foot pitch angle using IMU and a MBMOCAP systems simultaneously during walking cycles of subjects with normal gait. Data have been acquired from a group of 19 subjects including 10 male and nine female physically fit, able-bodied, aged between 18 and 25 years, with no reported injury or disability. Any subject with lumbar spine, leg or walking problem was excluded from this study (the relevant ethics obtained). Figure 1 indicates the data collection configuration for each test.

Subjects were asked to walk in a straight line with their natural walking style and initiate walking one second after the recording started. Both MBMOCAP and IMU data were recorded for two trials per subject.

Fig. 1 Data collection setup. **a** IMU attached to the foot. **b** IMU and T-shape reflective markers configuration

In both systems, data logging were manually triggered independently before the subject start the walking trial. However, because of the manual triggering of systems, there might be a short (i.e. less than 2 s) offset between the systems' readings. Because the frequency of the human gait cycle is roughly 2 Hz, sampling rate of both IMU sensors and the MBMOCAP system specified as 100 Hz to capture the details of each gait cycle.

3 Methodology

3.1 MAL

The MBMOCAP system utilized in this study is a Vicon 3D system (Fig. 2). This system consist of 18 semi-infrared camera that tracks infrared reflective markers. The data from these cameras collected by an analogue data collector and transferred to a data acquisition computer. It is also possible to record the trails with digital video camera for farther review.

The components of Vicon 3D system are illustrated in Table 1 [22]. This system is located in the Motion Analyses Laboratory (MAL), Curtin University Perth, Australia [22]. This system is an approximately 5 million AUD

Fig. 2 Vicon 3D motion capture system [10]

setup. It was calibrated prior to each data collection session, to conform to the level of accuracy stated in the system specification [23].

Additionally, a cluster of markers placed on a 3d printed T-shaped holder on top of the IMU system (Fig. 1) to capture the foot pitch angle. Usually the markers should be placed on anatomical landmarks (malleolus, metatarsal etc.) to match the locations of a predefined model, which calculates the angles. However, in this study, with respect to validation purpose, the cluster of markers must be in direct contact with sensor, which placed on foot.

The data from MAL reprocessed using Vicon Nexus software so that the movement can be presented in the form of 3D coordinates of the markers at any given time stamp with respect to the origin point. In the case that any markers being missed by the cameras, Vicon Nexus software provides tools to interpolate the unknown positions of the markers from good, known positions. The markers' coordinates during each trail time exported into a comma separated value (CSV) file.

The desired output from MAL is the angle of the perpendicular vector of the plane that contains A (x_a, y_a, z_a), B (x_b, y_b, z_b) and C (x_c, y_c, z_c), which would be the sensor plane (Fig. 3), with respect to perpendicular vector to the ground surface ($\vec{Z}$) (Fig. 4).

To calculate this angle ($\hat{\theta}$) using the information provided by MAL (coordinate of A, B and C), firstly $\vec{N}$, which is the sensor plane normal vector, calculated as follows

$$\vec{N} = (\vec{AB}) \otimes (\vec{AC}) \quad (1)$$

where $\vec{AB}$ is the vector that connects $\vec{A}$ and $\vec{B}$ and $\vec{AC}$ is the vector that connects $\vec{A}$ and $\vec{C}$, also $\otimes$ represents the cross operation of two vectors. If $\vec{Z}$ is the perpendicular vector to the ground surface, the angle between $\vec{Z}$ and $\vec{N}$ named $\hat{\theta}$ can be calculated from:

$$\hat{\theta} = \cos^{-1} \frac{(\vec{Z} \cdot \vec{N})}{(\|\vec{Z}\| \cdot \|\vec{N}\|)} \quad (2)$$

where “ $\cdot$ ” is the scalar product of the two vector and $\|\vec{Z}\|$ and $\|\vec{N}\|$ indicates the length of $\vec{Z}$ and $\vec{N}$, respectively.

Assuming no-movement within the first 500 ms, the average of $\hat{\theta}$ during this time period was subtracted from the rest of the trail to cancel any offset due to the angle between the T-shape holder and the foot.

As a check for the accuracy of the MAL system, the angle between $\vec{AB}$ and $\vec{AC}$, named $\hat{\beta}$, is also calculated. This angle should be constant at 90° and its variations demonstrates the accuracy of the MAL system.

Table 1 MAL and IMU components

MAL	IMU and receiver
Dual Xeon 6-core Data Acquisition Computer System	Arduino Pro mini board with Atmega328
Two Vicon Ultramet MX HD	MPU-9250 board
Vicon Giganet	NRF24L01
64 channels of analogue data collection	lithium battery
Vicon motion analysis system	

Fig. 3 Labeling the reflective markers. **a** Normal view. **b** Vicon Nexus software view **Fig. 4** Angle and vectors defined for MAL method. **a** Top view. **b** Right side view

3.2 IMU

An in-house designed IMU system has been used in this study as the second method of capturing the gait cycle which consist of three main components: a microcontroller, an IMU sensor and a wireless radio; all powered by a lithium battery. An Arduino Pro mini board (atmega328) [24] used to collect and timestamp measurements. An MPU-9250 board [25] used as the IMU sensor, containing 3-axis accelerometer, 3 axis gyroscope and 3-axis magnetometer sensor, as well as 16-bit analog-to-digital converters [25]. An NRF24L01 [26] used as the wireless communication unit to transfer the data packages into a PC via serial port and recorded as CSV file (Table 1). The final price tag for 3 IMU devices and the wireless dongle was approximately 32 AUD.

Each data package contains a timestamp with millisecond accuracy attached to accelerometer, gyroscope and magnetometer measurements in x, y and z axis.

The IMU system is installed in a 3D printed housing and attached to the right foot of subjects while the sensor's y axis is pointing toward subject's ankle and z axis is pointing upward (Fig. 1).

The 3D printed housing is attached to the foot using double-sided tapes. Although the housing may vibrate at the heel contact to the ground, because the cluster of markers is glued to the 3D housing, this vibration will sensed on both measurement methods. Therefore, when comparing the outcome of two methods against each other, the vibration does not can be neglected.

The desired output from IMU system is the angle between ground plane and sensor plane. This have been done using the data from both accelerometer and gyroscope sensors. The accelerometer which used in this study is highly sensitive to vibration and reports a significant level of noise even when stationary. Moreover, the gyroscope data suffers from drift (i.e. does not return to zero-rate when rotation stops). Nevertheless, by averaging the accelerometer and gyroscope data, a relatively accurate estimate of device orientation can be obtained [26]. To address these matters, a filtering and sensor fusion method used which explained in detail at [27]. Following is an overview about the sensor fusion method.

The accelerometer sensors of the IMU report the normal force (F_n) applied to the IMU plane in the form of a_x , a_y and a_z . Thus, when the sensor is stationary the output of accelerometer is the gravitational force in opposite direction (Fig. 5). It is assumed the gravitational force is always perpendicular to the ground surface [28].

Fig. 5 Angle and vectors defined for IMU Method

The angle of the sensor plane $\hat{\alpha}$ with respect to perpendicular vector to the ground surface $\vec{Z}$ can be calculated using the formula:

$$\hat{\alpha} = \tan^{-1} \frac{a_x}{a_y} \quad (3)$$

However this estimation can only be used if the total acceleration measured by the accelerometer sensor is less than a certain threshold (typically gravitational acceleration plus noise margin). To use accelerometers measurements, two conditions were checked:

First, the stationary condition:

$$\gamma_1 < \sqrt{a_{xk}^2 + a_{yk}^2 + a_{zk}^2} < \gamma_2 \quad (4)$$

For each k time stamp. The parameters γ_1 and γ_2 are selected to accommodate the noise embedded in the accelerometer reading and $\frac{\gamma_1 + \gamma_2}{2}$ to be equal to gravitational acceleration g . The typical value for γ_1 and γ_2 is 8.4 m/s^2 and 11.2 m/s^2 , respectively.

Second, the angular velocity to be close to zero, which indicates that the foot pitch angle is not changing. This condition indicates if foot is stationary and given as:

$$-\varepsilon < \omega_k < \varepsilon \quad (5)$$

where ω_k is the pitch angular velocity measured at time stamp k and ε is selected to accommodate error in gyro data ($3^\circ/\text{s}$).

When the conditions (4) and (5) are satisfied, foot pitch angle update is taken from the accelerometer reading otherwise integrating the gyroscope reading provides the foot pitch angle in that specific axis. The average of

gyroscope reading sampled at time stamps k and $(k-1)$ was used to calculate the foot pitch angle output for k .

Similar to $\hat{\theta}$, to cancel the offset of IMU readings, the average of $\hat{\alpha}$ during the first 500 ms was subtracted from the rest of the trail. By doing so, accelerometer and gyroscope were calibrated prior to each walking trial.

4 Results and discussion

In this section the outcome from both methods of IMU and MAL illustrated and discussed.

Maximum, minimum and standard deviation of $\hat{\beta}$, which are measures for the accuracy of MAL system, calculated as 90.014° , 88.014° and 0.373° , respectively.

The $\hat{\alpha}$ and $\hat{\theta}$ for one of the subjects during the second trial indicated in Figs. 6 and 7, respectively.

In Fig. 6, areas that are labeled “Gyro” and “Acc” indicates if the foot pitch angle is calculated using gyroscope or accelerometer, respectively.

To solve the misalignment caused by manual triggering, an offset has been applied to both signals based on the first peak of the signals (Fig. 8). Using the aligned angles, the level of similarity between them have been compared by three indices: error, RMSE and correlation.

The error between two signals at any given time, t , illustrated in Fig. 9 and calculated as follows:

$$e_t \equiv \text{error}(t) = \theta_t - \alpha_t \quad (6)$$

Fig. 6 $\hat{\alpha}$ versus time for the second trial of the first subject, captured by IMU

Fig. 7 $\hat{\theta}$ versus time for the second trial of the first subject, captured by MAL

Fig. 8 $\hat{\alpha}$ and $\hat{\theta}$ aligned for the second trial of first subject

The minimum and maximum error of all trials for each subject represented in fourth and fifth columns of Tables 2 and 3, respectively.

The correlation coefficient of $\hat{\alpha}$ and $\hat{\theta}$ indicates their linear dependency. If each angle has N scalar observation this measure defined as:

$$\rho(\alpha, \theta) = \frac{1}{N-1} \sum_{i=1}^N \left(\frac{\alpha_i - \mu_\alpha}{\sigma_\alpha} \right) \left(\frac{\theta_i - \mu_\theta}{\sigma_\theta} \right) \quad (7)$$

where $\mu_\alpha, \sigma_\alpha, \mu_\theta$ and σ_θ are the mean and standard deviation of $\hat{\alpha}$ and $\hat{\theta}$ respectively. The average correlation coefficient of $\hat{\alpha}$ and $\hat{\theta}$ for all trials of each subject have been provided in the first column of Tables 2 and 3. Figure 10 illustrates the correlation between $\hat{\alpha}$ and $\hat{\theta}$ signals of the second trial of subject-1 versus time lag (in millisecond). As an example it indicates that with -160 ms lag the correlation is maximum between the two signals.

Also RMSE indicates the root-mean-square level of the error between two signals over the time of the measurement and defined as:

$$RMSE_{(\theta, \alpha)} = \sqrt{\frac{1}{N} \sum_{n=1}^N |\theta_n - \alpha_n|^2} \quad (8)$$

Fig. 9 Error between $\hat{\alpha}$ and $\hat{\theta}$ for the second trial of first subject

The average of RMSE for each subject represented in third column of Tables 2 and 3.

Additionally the standard deviation of the correlation coefficients measured for males group is 0.773% and for females group is 0.851% and for all of the subjects is 0.828%.

The Tables 2 and 3 are representing the measures mentioned above for each female and male subject respectively. The following section discuss the comparison between the two signals.

For the two angles, $\hat{\alpha}$ and $\hat{\theta}$, obtained using two different methods IMU and MAL, the amount of RMSE is relatively low. That indicates the level of similarity between the results of the two system. The maximum value of RMSE among all of the samples in this study is 3.738° and the minimum RMSE value is 1.589° .

Also the correlation coefficient of max 99.542% and min 96.891% indicates the similar trend of both signals during each gait cycle which is also visible in Fig. 9. Moreover, $\hat{\alpha}$ and $\hat{\theta}$ illustrating minor differences during the stand phase of the gait cycle which is caused by the vibration at the contact of the foot and ground. Although the vibration is the same for both systems, it has different effect regarding the accuracy of $\hat{\alpha}$ and $\hat{\theta}$, which is the effect of low pass filter applied on $\hat{\alpha}$. Moreover, the standard deviation of 0.834% represents that the experiment was repeatable among all trails.

As mentioned in the introduction section, the current clinical method to measure foot pitch angle is not accurate and subjective. Therefore, the illustrated results are relatively accurate comparing to the visual inspection method.

Table 2 Maximum and minimum error, correlation and RMSE of $\hat{\alpha}$ and $\hat{\theta}$ for male subjects

Subjects	Mean correlation coefficient	RMSE (°)	Error (°)	
			Min	Max
1	0.9898	2.6042	− 8.529	6.284
2	0.9865	2.783	− 7.937	9.415
3	0.9789	2.923	− 9.655	8.137
4	0.9763	3.019	− 8.541	6.922
5	0.9926	1.974	− 4.293	6.752
6	0.9743	3.188	− 7.526	8.893
7	0.9954	1.812	− 4.184	5.434
8	0.9877	2.602	− 6.002	7.159
9	0.9890	2.754	− 5.513	5.126
10	0.9751	2.928	− 9.436	8.748

Table 3 Maximum and minimum error, correlation and RMSE of $\hat{\alpha}$ and $\hat{\theta}$ for male subjects

Subjects	Mean correlation coefficient	RMSE (°)	Error(°)	
			Min	Max
1	0.9898	2.6042	− 8.529	6.284
2	0.9865	2.783	− 7.937	9.415
3	0.9789	2.923	− 9.655	8.137
4	0.9763	3.019	− 8.541	6.922
5	0.9926	1.974	− 4.293	6.752
6	0.9743	3.188	− 7.526	8.893
7	0.9954	1.812	− 4.184	5.434
8	0.9877	2.602	− 6.002	7.159
9	0.9890	2.754	− 5.513	5.126
10	0.9751	2.928	− 9.436	8.748

Fig. 10 Correlation between $\hat{\alpha}$ and $\hat{\theta}$ versus time lag for the second trial of first subject

5 Conclusion

During this study the results of foot pitch angle measurements from two different gait analysis methods have been compared and validated through three different measures, which are correlation, RMSE and error. The outcome illustrated the level of similarity between the angle measurements from these two methods. Based on the validation results obtained, IMU device is reliable enough to be used in gait analysis applications that require accuracy and portability.

Resulting from the initial study, future improvements to the system were identified. For example, it is desirable to provide visualization feature to the current IMU system. This can be done by adding two digital cameras capturing video in two directions and synchronizing the videos with IMU measurements. Moreover, a user friendly software interface can potentially improve the usability of the system which will ideally get used by medical practitioners and physiotherapists.

Summation, this experiment has shown that the use of IMUs is a possible solution to the issues encountered with MBMOCAP systems such as the lack of portability and expense by providing a portable alternative with a minimal sacrifice in accuracy and a portion of the price for MBMOCAP system.

Acknowledgements We are grateful to The Kailis foundation group, ST John of God healthcare group, for the funding of this project.

Compliance with ethical standards

Ethical approval This study involves Human subjects and the relevant ethical approvals obtained from Curtin University of Technology (Human Research Ethics Office): HR 12/2016 and St John of God healthcare group (HREC): 823.

Conflict of interest The authors report no financial or other conflict of interest relevant to the subject of this article.

References

1. Baker R. Gait analysis methods in rehabilitation. *J Neuroeng Rehabil*. 2006;3:4.
2. Tao W, Liu T, Zheng R, Feng H. Gait analysis using wearable sensors. *Sensors (Basel)*. 2012;12(2):2255–83.
3. Cloete T, Scheffer C. Repeatability of an off-the-shelf, full body inertial motion capture system during clinical gait analysis. In: 2010 annual international conference on IEEE engineering in medicine and biology society EMBC'10, p. 5125–5128; 2010.
4. Zhou H, Hu H. Human motion tracking for rehabilitation—a survey. *Biomed Signal Process Control*. 2008;3(1):1–18.
5. Moeslund TB, Krüger V. A survey of advances in vision-based human motion capture and analysis. *Comput Vis Image Underst*. 2006;104(2):90–126.

6. J. Cockcroft. An evaluation of inertial motion capture technology for use in the analysis and optimization of road cycling kinematics. no. March, 2011.
7. Simon SR. Quantification of human motion: gait analysis—benefits and limitations to its application to clinical problems. *J Biomech.* 2004;37(12):1869–80.
8. K. Tong, M.H. Granat. A practical gait analysis system using gyroscopes.pdf, vol. 21; 1999. p. 87–94.
9. Pappas IPI, Keller T, Mangold S, Popovic MR, Dietz V, Morari M. A reliable gyroscope-based gait-phase detection sensor embedded in a shoe insole. *IEEE Sens J.* 2004;4(2):268–74.
10. Auvinet B, et al. Reference data for normal subjects obtained with an accelerometric device. *Gait Posture.* 2002;16(2):124–34.
11. Aminian K, Najafi B, Büla C, Leyvraz PF, Robert P. Spatio-temporal parameters of gait measured by an ambulatory system using miniature gyroscopes. *J Biomech.* 2002;35(5):689–99.
12. Huang Y, Jirattigalachote W, Cutkosky MR, Zhu X, Shull PB. Novel foot progression angle algorithm estimation via foot-worn, magneto-inertial sensing. *IEEE Trans Biomed Eng.* 2016;63(11):2278–85.
13. Coley B, Najafi B, Paraschiv-Ionescu A, Aminian K. Stair climbing detection during daily physical activity using a miniature gyroscope. *Gait Posture.* 2005;22(4):287–94.
14. Mayagoitia RE, Nene AV, Veltink PH. Accelerometer and rate gyroscope measurement of kinematics: an inexpensive alternative to optical motion analysis systems. *J Biomech.* 2002;35(4):537–42.
15. Wu G, Ladin Z. The study of kinematic transients in locomotion using the integrated kinematic sensor. *IEEE Trans Rehabil Eng.* 1996;4(3):193–200.
16. Lau H, Tong K. The reliability of using accelerometer and gyroscope for gait event identification on persons with dropped foot. *Gait Posture.* 2008;27(2):248–57.
17. Fourati H. Heterogeneous data fusion algorithm for pedestrian navigation via foot-mounted inertial measurement unit and complementary filter. *IEEE Trans Instrum Meas.* 2015;64(1):221–9.
18. F. Woyano, S. Lee, S. Park. Evaluation and comparison of performance analysis of indoor inertial navigation system based on foot mounted IMU. In: International conference on advanced communication technology ICACT, vol. 2016–March; 2016. p. 792–798.
19. Li Q, Young M, Naing V, Donelan JM. Walking speed and slope estimation using shank-mounted inertial measurement units. 2009 IEEE Int Conf Rehabil Robot ICORR. 2009;2009:839–44.
20. Luinge HJ, Veltink PH, Baten CTM. Ambulatory measurement of arm orientation. *J Biomech.* 2007;40(1):78–85.
21. Van Den Noort JC, Kortier HG, Van Beek N, Veeger DHEJ, Veltink PH. Measuring 3D hand and finger kinematics—a comparison between inertial sensing and an opto-electronic marker system. *PLoS ONE.* 2016;11(11):1–16.
22. Curtin university. Motion Analysis Lab, Facilities and Locations, school and departments; 2015. <http://healthsciences.curtin.edu.au/schools-and-departments/physiotherapy-exercise-science/facilities/>.
23. M. Windolf, N. Götz, M. Morlock. Systematic accuracy and precision analysis of video motion capturing systems—exemplified on the Vicon-460 system.
24. S.W. List. Using the Arduino Pro Mini 3.3 V. p. 1–9.
25. R. Chokshi. Datasheet RS-MPU-6000A-00. p. 52, 2015.
26. A. Noureldin, T.B. Karamat, J. Georgy. Fundamentals of inertial navigation, satellite-based positioning and their integration, vol. 58, no. 12. Springer, Berlin, 2013.
27. Sharif Bidabadi S, Murray I, Lee G. The application of inertial measurements unit for the clinical evaluation and assessment of gait events. *J Med Eng Technol.* 2017;41:309–1902.
28. A. Kim, M. Golnaraghi. A quaternion-based orientation estimation algorithm using an inertial measurement unit. In: Position Location and Navigation Symposium, 2004, no. May 2004, p. 268–272.