Received:
02 April 2018

Revised:
06 June 2018

Accepted:
11 June 2018

<https://doi.org/10.1259/bjr.20180317>

Cite this article as:

Al-Murshedi S, Hogg P, England A. An investigation into the validity of utilising the CDRAD 2.0 phantom for optimisation studies in digital radiography. *Br J Radiol* 2018; **91**: 20180317.

FULL PAPER

An investigation into the validity of utilising the CDRAD 2.0 phantom for optimisation studies in digital radiography

SADEQ AL-MURSHEDI, BSc, MSc, PETER HOGG, DCR (R), DRI, Cert RP, BSc, PG Cert, MPhil, FCR and ANDREW ENGLAND, BSc (Hons), PgCert, MSc, PhD

School of Health Sciences, University of Salford, Salford, United Kingdom

Address correspondence to: Mr Sadeq Al-Murshedi
E-mail: s.h.m.al-murshedi@edu.salford.ac.uk

Objectives: To determine if a relationship exists between low contrast detail (LCD) detectability using the CDRAD 2.0 phantom, visual measures of image quality (IQ) and simulated lesion visibility (LV) when performing digital chest radiography (CXR).

Methods: Using a range of acquisition parameters, a CDRAD 2.0 phantom was used to acquire a set of images with different levels of image quality. LCD detectability using the CDRAD 2.0 phantom, represented by an image quality figure inverse (IQF_{inv}) metric, was determined using the phantom analyser software. A Lungman chest phantom was loaded with two simulated lesions, of different sizes/placed in different locations, and was imaged using the same acquisition factors as the CDRAD 2.0 phantom. A relative visual grading analysis (VGA) was used by seven observers for IQ and LV evaluation

of the Lungman images. Correlations between IQF_{inv}, IQ and LV were investigated.

Results: Pearson's correlation demonstrated a strong positive correlation ($r = 0.91$; $p < 0.001$) between the IQ and the IQF_{inv}. Spearman's correlation showed a good positive correlation ($r = 0.79$; $p < 0.001$) and ($r = 0.68$; $p < 0.001$) between the IQF_{inv} and the LV for the first lesion (left upper lobe) and the second lesion (right middle lobe), respectively.

Conclusions: From results presented in this study, the automated evaluation of LCD detectability using CDRAD 2.0 phantom is likely to be a suitable option for IQ and LV evaluation in digital CXR optimisation studies.

Advances in knowledge: This research establishes the potential of the CDRAD 2.0 phantom in digital CXR optimisation studies.

INTRODUCTION

An essential technical priority of any X-ray examination is the production of an image with sufficient quality to answer the clinical question. Balanced against this is the requirement to avoid unnecessary radiation risk to the patient.¹ Within national legislature, a key aim is to identify acquisition parameters which deliver acceptable image quality with the lowest radiation dose practicable.² Methods for patient dose assessment are readily available but the approaches for image quality (IQ) assessment are more complex. Observer performance methods, such as receiver operating characteristic (ROC) and visual grading analysis (VGA), are often utilised for evaluating clinical IQ in research studies.^{3–6} Nevertheless, they are difficult to implement for the purpose of evaluating IQ, within and between hospitals and to use ROC and VGA methods within routine quality assurance programmes would be time consuming and require a large number of images. Physics contrast-detail (CD) phantoms, such as CDRAD 2.0 phantom, are normally

used in medical physics departments⁷ and have been widely used within the literature for optimisation studies and when comparing the performance of imaging systems.^{8–12}

The CDRAD 2.0 phantom does have limitations, it has a uniform background which does not take into account the influence of anatomical noise arising when imaging human anatomy.¹³ The principle of this phantom is based on the Rose model which assumes that the quantum noise is the limiting factor which can influence image quality (LCD detection).¹⁴ This assumption is true for the CDRAD 2.0 phantom, but it could be limited for several common clinical tasks in diagnostic radiology.¹³ Within clinical practice, the limiting factor for LCD detection performance is the anatomical background which is considered to be the most influential factor in CXR IQ rather than the quantum noise.^{15–17} As such, the validity of using the CDRAD 2.0 phantom for representing the chest region for IQ and LV evaluations, including optimisation studies is questionable

and needs further investigation.¹³ It would be extremely useful if this phantom could predict visual (perceptual) IQ and LV in the clinical setting since these parameters are more useful in optimisation studies, when undertaking routine quality assurance and also when evaluating and comparing IQ between systems.

The aim of this study was to investigate the potential of using LCD detectability, using the CDRAD 2.0 phantom, for digital CXR optimisation studies by investigating its correlation with IQ and LV.

MATERIALS AND METHODS

Image acquisition

Two different phantoms were used in this study; (1) an anthropomorphic Lungman chest phantom (Kyoto Kagaku Co., Japan),¹⁸ loaded with two spherical phantom lesions that simulate pulmonary nodules. A 10 mm +100 HU (Hounsfield unit) lesion was placed in left upper lobe and a 12 mm +100 HU lesion was placed in the right middle lobe. Lesion type and placement was based on the need to simulate a clinically relevant scenario and also take into account the complexity of the anatomical surroundings. This approach has been previously reported in the literature.¹⁹ (2) the CDRAD 2.0 phantom (Artinis Medical System, The Netherlands).²⁰ This phantom consists of a square acyclic plastic plate (265 × 265 mm). It is 10 mm thick and has holes of various depths and diameters providing a range of differently sized contrast details against a uniform background. Lungman was required for evaluating IQ and LV, while the CDRAD 2.0 phantom was required for evaluating LCD detectability. Image acquisitions were undertaken using a commercially available X-ray machine (Wolverson X-ray Ltd, Willenhall, West Midlands, UK) with an erect Bucky encompassing an anti scatter radiation grid (10:1 ratio, 40 line/cm frequency) and 3 mm Al inherent X-ray beam filtration and a Konica Minolta Aero DR detector (Konica Minolta Medical Imaging USA INC, Wayne, NJ). 42 images of the Lungman phantom were acquired using a fixed 180 cm source to image receptor distance (SID), no additional filtration and manual exposure termination (fixed mAs). Different settings of tube potential (70, 80, 90, 100, 110, 120 and 130 kVp) and tube current-time (1, 2 and 3.2 mAs), with and without anti scatter grids were applied during image acquisition. For all acquisitions the post-processing was determined as part of the imaging protocol and was based on a PA chest examination. Following this, 42 CDRAD 2.0 phantom images were acquired using the same acquisition parameters that were applied to the Lungman phantom. Each of the CDRAD 2.0 acquisitions were repeated three times as recommended by the CDRAD 2.0 phantom manufacturer.²⁰ The CDRAD 2.0 phantom was placed between 10 cm PMMA slabs to simulate an adult patient.^{21–23} Acquisition parameters were selected to be similar to those used in clinical practice. In addition, using manual exposure control together with an extremely wide range of acquisition factors allowed the acquisition of a series of images with a wide range of image qualities. The detector dose (without backscatter) was measured at the image detector surface using a solid-state dosimeter (RaySafe X2, Unfors RaySafe AB, Billdal, Sweden), additionally the beam quality (half-value layer, HVL) for each kVp value was recorded using the same dosimeter (Supplementary material 1).

Table 1. Visual IQ criteria^{26,27}

Item	Criteria
1	Demonstration of the vascular pattern in the whole lung, particularly the peripheral vessels.
2	Visually sharp demonstration of the trachea and proximal bronchi.
3	Visually sharp demonstration of the borders of the heart and aorta.
4	Visually sharp demonstration of the diaphragm and lateral costo-phrenic angles.
5	Visualisation of the retrocardiac lung and the mediastinum.
6	Visualisation of the spine through the heart shadow.

IQ, image quality.

Image quality evaluation

LCD detectability from the CDRAD 2.0 phantom was represented by an IQF_{inv} value, this was the average of the three repeated images and was produced automatically using the phantom analyser software.²⁰ The IQF_{inv} was calculated using the following equation²⁰:

$$IQF_{inv} = \sum_{i=1}^{15} \frac{1}{C_i \times D_{(i,th)}}$$

$D_{(i,th)}$ represents the lowest diameter (threshold diameter) in column (i) that has a visible hole correctly detected; C_i represents the depth value (contrast) of the object (visible hole) in the column (i). Using the analyser software, the input parameters were set as follows; (1) the alpha level of significance was set at $1e^{-008}$, as recommended in the literature.^{20,24} (2) the priori difference of means was set relative to the image bit depth,²⁰ finally, SID was set to 180 cm.

IQ and LV for the Lungman phantom images were evaluated visually by seven qualified diagnostic radiographers. They applied a relative VGA method and a 5-point Likert scale (1 = much worse, 2 = worse, 3 = equal, 4 = better, 5 = much better) using bespoke computer software.²⁵ The criteria used for IQ and LV evaluations are demonstrated in Tables 1 and 2, respectively and they are based on the European Guidelines²⁶ and the study by Ma et al.²⁷ Two simulated lesions, in two different locations, as previously described were selected as shown in Figure 1. For the relative VGA method, images were presented to observers on a computer monitor, on the left-hand monitor there was a fixed reference image and on the right hand monitor an image was randomly displayed from the experimental image bank. Each observer reviewed the images using a 5 mega-pixel DOME E5

Table 2. Simulated LV criteria²⁷

Item	Criteria
1	Contrast of nodule, against background.
2	Brightness of nodule, against background.
3	Sharpness of nodule edge.

LV, lesion visibility.

Figure 1. An example of a Lungman chest phantom image illustrating the two simulated lesions for the LV evaluation. (a) 10 mm + 100 HU lesion placed in left upper lobe. (b) The 12 mm + 100 HU lesion was placed in the right middle lobe. HU, Hounsfield unit; LV, lesion visibility.

(NDSsi, Santa Rosa, CA) monitors (2048 by 2560 pixels), calibrated to the DICOM Grey Scale Display Function (GSDF) Standard,²⁸ and applied the set of criteria whilst making a comparison against the reference image. The reference image was selected by consensus opinion from expert observers since it had “average” quality when compared with the others. This was to ensure that all of the five points on the Likert scale points were used in the evaluations.^{27,29} Observers were not permitted to adjust the contrast, intensity or magnification of the image. The overall IQ score for each image and for each observer was calculated by adding the scores from all of the IQ criteria. This approach was similar for the calculations of overall LV, for each image, by each observer. Next, the final IQ score for each image was determined by taking the average IQ score across the seven observers, and the same method was used for LV. Ethical approval for this study was granted from the University of Salford (HSR1617-76).

Statistical analysis

Data were analysed using SPSS version 22.0 (IBM Inc, Armonk, NY). Data were first examined to investigate normality based on visual inspection of the data and the Shapiro-Wilk test. This was to determine the most appropriate correlation statistic between IQ_{inv} and IQ and the LV. Data, which were normally distributed, were analysed using Pearson's correlation coefficient; non-parametric data were analysed using Spearman's correlation coefficient. The interpretation of the strength of the correlation (r) was conducted based on recommendations from the literature,^{30,31} in which $r = 0.1$ – 0.29 (small), $r = 0.30$ – 0.49 (medium), and $r = 0.50$ – 1.0 (large).

IQ and the IQ_{inv} had a normal distribution whereas LV showed a non-parametric distribution. To investigate the inter observer variability for the IQ and LV evaluations, an inter class correlation coefficient (ICC) was utilised. ICC values less than 0.5 indicate

poor reliability, values from 0.5 to 0.75 indicate moderate reliability, and values from 0.75 to 0.9 indicate good reliability finally values greater than 0.90 indicates excellent reliability.^{32,33}

RESULTS

The IQ_{inv} values from the CDRAD 2.0 phantom images and visual IQ and LV evaluations, together with the acquisition parameters are illustrated in Table 3. Correlation analysis between IQ and IQ_{inv} demonstrated a strong positive correlation ($r = 0.91$; $p < 0.001$). Spearman's correlation was used to compare the IQ_{inv} and LV. A good correlation was observed for both lesions; lesion one ($r = 0.79$; $p < 0.001$); lesion two ($r = 0.68$; $p < 0.001$). Figure 2 illustrates the linear regression curve between the IQ and IQ_{inv} . Figures 3 and 4 demonstrate the relationship, using a linear regression curve, between the LV for the first lesion and the second lesion again as a function of IQ_{inv} .

Excellent inter observer agreement between the seven observers was observed (ICC = 0.90; 95% CI [0.84–0.94]) for the visual IQ evaluations. A good inter observer agreement was observed for LV of the first lesion (ICC = 0.84; 95% CI: 0.68–0.90) and the second lesion (ICC = 0.83; 95% CI [0.714–0.91])

DISCUSSION

Image quality and lesion visibility studies have a number of applications within medical imaging but are notoriously labour intensive. Automated methods for evaluating IQ are attractive but historically there has been a lack of data on their relationships with perceptual (visual) methods.

It must be acknowledged that there are potential limitations when using the CDRAD 2.0 phantom for IQ evaluation and CXR optimisation studies. The chest region is classified as a quantum-saturated region where the anatomical structures are the main

Table 3. Acquisition parameters used to acquire study images and the resultant mean and SD of IQF_{inv}, IQ and LV values

	Mean (SD) Relative VGA score									
	IV									
	Mean (SD) IQF _{inv} score					Right middle lobe				
kVp/mAs	Left upper lobe					IQ				
	1	2	3.2	1	2	3.2	1	2	3.2	
70	0.28 (0.02)	0.40 (0.07)	0.54 (0.06)	6.43 (2.15)	6.57 (2.44)	7.00 (2.36)	6.00 (0.00)	6.00 (1.13)	6.00 (1.73)	12.14 (1.57)
80	0.79 (0.07)	0.86 (0.03)	1.00 (0.09)	7.43 (1.50)	6.86 (2.27)	7.86 (2.30)	6.29 (1.81)	6.43 (2.12)	7.57 (1.95)	12.71 (3.50)
90	1.00 (0.04)	0.59 (0.06)	0.64 (0.08)	7.57 (2.16)	6.86 (2.70)	8.43 (2.65)	7.29 (1.25)	6.1 (1.81)	6.43 (1.57)	13.43 (3.55)
100	0.89 (0.05)	1.24 (0.06)	1.21 (0.08)	8.00 (3.27)	8.71 (1.25)	8.57 (1.89)	7.0 (2.07)	8.71 (0.76)	9.00 (0.00)	13.43 (4.58)
110	1.41 (0.08)	1.57 (0.10)	0.82 (0.12)	8.29 (1.46)	8.57 (1.89)	8.57 (2.79)	7.43 (1.51)	9.00 (0.00)	6.43 (1.77)	17.43 (1.51)
120	1.17 (0.03)	1.40 (0.05)	1.45 (0.11)	8.14 (1.46)	8.57 (1.27)	8.57 (1.13)	8.00 (2.15)	9.14 (0.83)	9.00 (0.00)	13.86 (2.21)
130	1.59 (0.08)	1.84 (0.18)	2.10 (0.16)	9.00 (1.00)	8.71 (0.76)	8.86 (0.90)	8.57 (1.13)	8.00 (1.41)	9.00 (1.83)	17.86 (0.38)
70*	0.60 (0.09)	0.74 (0.04)	0.88 (0.06)	10.00 (2.63)	10.00 (2.38)	9.57 (3.76)	6.14 (1.81)	8.00 (1.53)	7.00 (2.15)	12.43 (3.26)
80*	0.94 (0.01)	1.15 (0.11)	1.12 (0.03)	9.86 (1.77)	10.00 (1.73)	9.71 (2.06)	7.43 (1.51)	8.29 (2.91)	8.43 (2.58)	13.71 (3.41)
90*	1.21 (0.10)	0.97 (0.04)	1.17 (0.04)	10.57 (1.62)	10.29 (2.14)	10.29 (1.70)	8.71 (2.14)	8.57 (2.45)	9.86 (2.12)	14.71 (1.98)
100*	1.39 (0.13)	1.68 (0.13)	1.65 (0.01)	10.29 (2.51)	10.29 (1.80)	10.14 (2.04)	9.29 (2.79)	9.43 (1.72)	9.00 (1.53)	15.86 (1.35)
110*	1.89 (0.16)	1.75 (0.05)	1.35 (0.15)	10.00 (2.70)	10.86 (1.46)	10.14 (1.86)	9.71 (1.98)	9.29 (3.00)	9.14 (2.27)	16.14 (2.04)
120*	1.57 (0.03)	1.88 (0.12)	2.18 (0.13)	10.29 (1.72)	10.86 (1.15)	10.71 (2.00)	10.14 (2.50)	8.86 (1.68)	9.00 (3.40)	16.57 (0.98)
130*	2.24 (0.13)	2.39 (0.12)	2.38 (0.30)	10.43 (1.40)	10.14 (1.57)	11.29 (1.40)	9.14 (2.69)	9.00 (3.87)	9.14 (3.48)	19.14 (3.24)

IQF_{inv} image quality figure inverse; LV, lesion visibility; SD, standard deviation; VGA, visual grading analysis
 Protocols indicated by a * were undertaken using an anti scatter radiation grid.

Figure 2. Linear regression curve between mean IQF_{inv} scores and mean IQ values. Error bars across the x axis represent the SD of the scores between the seven observers, while the error bars across the y axis represents the SD from the three automated CDRAD 2.0 image scores. IQF_{inv}, image quality figure inverse; LV, lesion visibility; SD, standard deviation.

factor that influence lesion detection and IQ when compared with quantum noise.³⁴ Quantum noise is a limiting factor for physics phantoms (*i.e.* CDRAD 2.0) whereas images acquired on patients would be limited by anatomical noise. Results from our study have demonstrated a strong positive correlation ($r = 0.91$; $p < 0.001$) between the physical evaluation of LCD detectability (IQF_{inv} from the CDRAD 2.0 phantom) and IQ evaluations from Lungman. Based on IQ alone, the CDRAD 2.0 phantom appears to be a valid and a well justified approach and can be used as a surrogate for IQ tasks using an anthropomorphic phantom when evaluating the performance of radiography systems or during optimisation studies. For LV, it is important to know to what extent there is correlation between the IQF_{inv} values and LV from

Figure 3. Linear regression curve between the mean IQF_{inv} scores against the mean LV scores for the first lesion (left upper lobe). Error bars across the x axis represent the SD of the scores between the seven observers, while the error bars across the y axis represents the SD of the three CDRAD 2.0 image scores. IQF_{inv}, image quality figure inverse; LV, lesion visibility; SD, standard deviation.

Figure 4. Linear regression curve between the mean IQF_{inv} scores against the mean LV scores for the second lesion (right middle lobe). Error bars across the x axis represent the SD of the scores between the seven observers, while the error bars across the y axis represents the SD of the three CDRAD 2.0 image scores. IQF_{inv}, image quality figure inverse; LV, lesion visibility; SD, standard deviation.

the Lungman phantom images since the CDRAD 2.0 phantom was designed to examine the performance of imaging systems in visualising LCD objects.

Our findings demonstrated a good positive correlation between IQF_{inv} and LV for the two simulated lesions; the first lesion showed $r = 0.79$ ($p < 0.001$) and the second lesion showed $r = 0.68$ ($p < 0.001$). It was useful to examine the correlation between the IQF_{inv} and the LV when the lesions were located in different positions and with different levels of visibility (high and low). This was necessary to understand how the variations in the location and visualisation of the lesion affect the correlation values of IQF_{inv}. Although the two simulated lesions have the same HU value (+100 HU), the radiographic visibility of the first simulated lesion (in the left upper lobe) was lower than that of the second one (in the right middle lobe). This is related to the complexity of the anatomical background surrounding the first simulated lesion. A relatively large part of this simulated lesion is covered by the first anterior rib and third posterior rib and the HU values for the surrounding lung, the first anterior rib and the third posterior rib were -890, 353 and 575, respectively. In contrast, a small part of the second simulated lesion is covered by the fourth anterior rib. The HU values for the surrounding lung and the fourth anterior rib were -1027 and 467, respectively (Figure 1). In addition, the first simulated lesion had a smaller diameter (10 mm) compared with that of the second simulated lesion (12 mm). It was found that the lesion with lower visibility (left upper lobe) had higher correlation ($r = 0.79$) when compared with that of the higher visibility lesion (right middle lobe, $r = 0.68$). A possible reason behind this could be because the small levels of improvement and degradation of general IQ which would have a higher impact on increasing or decreasing the visualisation of the lower visibility lesion when compared with that of the higher visibility lesion (acquisition parameters had more of an effect on the lower visibility lesion). The good positive correlation observed under these two conditions provides a useful indication on the utility of

the CDRAD 2.0 phantom for investigating the performance of an imaging system in LCD detection.

To our knowledge, only one study undertaken by De Crop and colleagues, attempted to investigate the correlation between the LCD detectability using a CDRAD 2.0 phantom and IQ from cadaveric images using an absolute VGA method.²¹ However, the correlation between LV and LCD detectability from CDRAD 2.0 has not been investigated. Within the work of De Crop et al²¹ comparable results were reported with a strong positive correlation ($r = 0.91$; $p = 0.001$) between the physical evaluation of $IQ_{F_{inv}}$ (CDRAD 2.0) and IQ (cadavers). Our work has built upon the study by De Crop et al³⁵ in that we used a Lungman phantom loaded with simulated pulmonary lesions instead of three normal cadavers with no inherent pathology. The inclusion of pathology provides the opportunity for a more accurate evaluation of IQ and the detection task, and it is an important factor in optimisation studies. We consider our study to have some methodological benefits when compared with De Crop et al.³⁵ Within our work a wider range of image qualities were generated and tested compared with that by De Crop et al³⁵ this can be seen from the lowest value of $IQ_{F_{inv}}$ in our study which was equal to 0.28 compared with 2.56 in study by De Crop et al. The range of image qualities generated in our study closely reflected those acquired in clinical practice but also included both high ($IQ_{F_{inv}} = 2.39$) and low ($IQ_{F_{inv}} = 0.28$) qualities, this was necessary to ensure that the correlations between the $IQ_{F_{inv}}$ and IQ/LV were tested across a full range of different levels of IQ. Image qualities in our study were deemed similar to those acquired in clinical practice for several reasons. First, a group of experienced radiographers were asked to provide acquisition parameters routinely used within their clinical practice. Second, acquisitions parameters were compared with those cited within the literature. Third, resultant images were reviewed by a group of experts and were deemed to demonstrate a range of image qualities which may be expected within the clinical environment. We also used a relative VGA method, without allowing windowing and magnification. This was different to the absolute VGA system, with windowing and magnification, used in the study by De Crop et al.³⁵ Utilising a relative VGA approach is justified because it is likely to be less biased and more sensitive in detection of subtle variations in image quality, this is due to the presence of a reference image. Although the free adjustment of image viewing parameters, such as brightness, contrast and magnification are considered part of a clinical review of an image, it could, however, influence the correlation between the $IQ_{F_{inv}}$ and the IQ and LV and was not permitted within this study. A consistent image display is necessary to ensure that any measured variation in IQ results only from the imaged object and not the adjustment of image viewing parameters such as zooming or windowing.

Geijer et al³⁶ investigated the influence of different tube potentials (by adjusting the mAs values to keep the effective dose constant) on IQ in lumbar spine radiography using two different phantoms, CDRAD 2.0 and the adult anthropomorphic Alderson phantom. Findings of the study by Geijer et al³⁶ demonstrate that the IQ was higher for high kVp values (≥ 96 kVp) for both phantoms. Bacher et al²³ compared the performance of an amorphous

silicon and amorphous selenium flat-panel detector using two different approaches: clinical images generated from patients and using CDRAD 2.0 phantom images. Both methods showed that the amorphous silicon flat-panel system had a lower effective dose, without any reduction of IQ when compared with that of amorphous selenium flat-panel system. However, in these two studies the correlation between the IQ from the Alderson phantom, images from patients and the $IQ_{F_{inv}}$ from the CDRAD 2.0 phantom were not calculated. Within mammography, the use of the CDMAM phantom is well established^{35,37,38} and there is some similarity in the evaluation of image quality with CDMAM and the methods described in this study. To our knowledge, there is no study which has investigated the correlation between the $IQ_{F_{inv}}$ and the LV, as a result there is no data available in the literature to compare with our findings. The good positive correlations reported in our study ($IQ_{F_{inv}}$ and IQ and LV) might also have a value in indicating the detectability of abnormality and possibly this could extend to lesion detection performance; several studies have found that there is a correlation between pathological detectability and the visibility of normal anatomical structure.^{39–42}

In 2011, Moore et al⁴³ created and validated a computer algorithm which was able to simulate computed radiography (CR) chest images from digitally reconstructed radiographs (DRRs) generated by ray tracing virtual X-rays through clinical patient CT datasets. Several studies have been conducted by Moore et al^{44,45} that focused on optimisation of CR CXR imaging using a DRR computer simulation. The simulated clinical experiment (simulated DRR) used by Moore et al has been successfully utilised for deriving an optimal exposure parameter for CXR examinations. Therefore, further studies could be conducted using DRRs or clinical images from patients, applying the same acquisition parameters used in our study, to investigate the validity of using the CDRAD 2.0 phantom as a measure of clinical image quality in optimisation studies. This can provide a further evidence of the validity of using the CDRAD 2.0 phantom for optimisation studies.

Also, based on our findings, future work could focus on establishing a baseline $IQ_{F_{inv}}$ value for CXR that identifies the threshold of LCD detectability that is required for obtaining an adequate IQ for diagnosis. The availability of such baseline measures is highly beneficial for routine quality assurance programs to ensure that the IQ is within accepted limits, and it is also beneficial for optimisation studies as it can facilitate the evaluation of how different techniques can influence IQ in clinical practice.

A limitation of our study was that the correlations between the LCD detectability and IQ and LV were only carried out using a standard size anthropomorphic phantom and the variation in phantom size were not taken into account. Anatomical noise would increase for larger patients and it was not possible to assess the effects of this within our study. A further limitation was the correlation between the LCD detectability and the LV was carried out using only one type of simulated lesions (that simulate a pulmonary nodules) and the other types (such as

micronodules, lines, and reticular opacities) were not considered. Future work can be undertaken taking into account these limitations by using different anthropomorphic phantom sizes, loaded with different simulated lesions.

CONCLUSION

Based on our findings, there is an excellent correlation between LCD detectability using CDRAD 2.0 phantom and IQ. A good correlation existed between LCD detectability and LV. It can be confirmed that the CDRAD 2.0 phantom is valid for evaluating

IQ and LV and could be of use in digital CXR optimisation studies. Also, our findings suggest it is possible that the CDRAD 2.0 phantom can be utilised along with the automated evaluation method for LCD detectability within routine quality assurance and optimisation studies.

ACKNOWLEDGMENT

The author is supported by the Higher Committee for Education Development in Iraq (HCED-Iraq) and he gratefully acknowledges this supporting.

REFERENCES

1. Busch HP, Faulkner K. Image quality and dose management in digital radiography: a new paradigm for optimisation. *Radiat Prot Dosimetry* 2005; **117**(1–3): 143–7. doi: <https://doi.org/10.1093/rpd/nci728>
2. Instrument S. The Ionising Radiation (Medical Exposure) Regulations 2000. 2000.
3. Lanhede B, Båth M, Kheddache S, Sund P, Björnell L, Widell M, et al. The influence of different technique factors on image quality of chest radiographs as evaluated by modified CEC image quality criteria. *Br J Radiol* 2002; **75**: 38–49. doi: <https://doi.org/10.1259/bjr.75.889.750038>
4. Redlich U, Hoeschen C, Doebling W. Assessment and optimisation of the image quality of chest-radiography systems. *Radiat Prot Dosimetry* 2005; **114**: 264–8. doi: <https://doi.org/10.1093/rpd/ncb559>
5. Goo JM, Im JG, Lee HJ, Chung MJ, Seo JB, Kim HY, et al. Detection of simulated chest lesions by using soft-copy reading: comparison of an amorphous silicon flat-panel-detector system and a storage-phosphor system. *Radiology* 2002; **224**: 242–6. doi: <https://doi.org/10.1148/radiol.2241011441>
6. Yano Y, Yabuuchi H, Tanaka N, Morishita J, Akasaka T, Matsuo Y, et al. Detectability of simulated pulmonary nodules on chest radiographs: comparison between irradiation side sampling indirect flat-panel detector and computed radiography. *Eur J Radiol* 2013; **82**: 2050–4. doi: <https://doi.org/10.1016/j.ejrad.2013.05.036>
7. Moore CS, Wood TJ, Beavis AW, Saunderson JR. Correlation of the clinical and physical image quality in chest radiography for average adults with a computed radiography imaging system. *Br J Radiol* 2013; **86**: 20130077. doi: <https://doi.org/10.1259/bjr.20130077>
8. Bacher K, Smeets P, Bonnarens K, De Hauwere A, Verstraete K, Thierens H. Dose reduction in patients undergoing chest imaging: digital amorphous silicon flat-panel detector radiography versus conventional film-screen radiography and phosphor-based computed radiography. *AJR Am J Roentgenol* 2003; **181**: 923–9. doi: <https://doi.org/10.2214/ajr.181.4.1810923>
9. Geijer H, Beckman KW, Andersson T, Persliden J. Image quality vs. radiation dose for a flat-panel amorphous silicon detector: a phantom study. *Eur Radiol* 2001; **11**: 1704–9. doi: <https://doi.org/10.1007/s003300100851>
10. Rong XJ, Shaw CC, Liu X, Lemacks MR, Thompson SK. Comparison of an amorphous silicon/cesium iodide flat-panel digital chest radiography system with screen/film and computed radiography systems--a contrast-detail phantom study. *Med Phys* 2001; **28**: 2328–35. doi: <https://doi.org/10.1118/1.1408620>
11. Veldkamp WJ, Kroft LJ, Boot MV, Mertens BJ, Geleijns J. Contrast-detail evaluation and dose assessment of eight digital chest radiography systems in clinical practice. *Eur Radiol* 2006; **16**: 333–41. doi: <https://doi.org/10.1007/s00330-005-2887-6>
12. Alzimami K, Sassi S, Alkhorayef M, Britten AJ, Spyrou NM. Optimisation of computed radiography systems for chest imaging. *Nuclear Instruments and Methods in Physics Research Section A: Accelerators, Spectrometers, Detectors and Associated Equipment* 2009; **600**: 513–8. doi: <https://doi.org/10.1016/j.nima.2008.12.023>
13. Månsson LG, Båth M, Mattsson S. Priorities in optimisation of medical X-ray imaging--a contribution to the debate. *Radiat Prot Dosimetry* 2005; **114**: 298–302. doi: <https://doi.org/10.1093/rpd/ncb578>
14. Rose A. The sensitivity performance of the human eye on an absolute scale. *J Opt Soc Am* 1948; **38**: 196–208. doi: <https://doi.org/10.1364/JOSA.38.000196>
15. Båth M, Håkansson M, Börjesson S, Kheddache S, Grahn A, Bochud FO, et al. Nodule detection in digital chest radiography: part of image background acting as pure noise. *Radiat Prot Dosimetry* 2005; **114**: 102–8. doi: <https://doi.org/10.1093/rpd/ncb523>
16. Håkansson M, Båth M, Börjesson S, Kheddache S, Johnsson AA, Månsson LG. Nodule detection in digital chest radiography: effect of system noise. *Radiat Prot Dosimetry* 2005; **114**: 97–101. doi: <https://doi.org/10.1093/rpd/ncb525>
17. Samei E, Flynn MJ, Eyler WR. Detection of subtle lung nodules: relative influence of quantum and anatomic noise on chest radiographs. *Radiology* 1999; **213**: 727–34. doi: <https://doi.org/10.1148/radiology.213.3.r99dc19727>
18. K.M. DLA. *The Phantoms of Medical and Health Physics*. New York, NY: Springer; 2014.
19. Jessop M, Thompson JD, Coward J, Sanderud A, Jorge J, de Groot M, et al. Lesion detection performance: comparative analysis of low-dose CT data of the chest on two hybrid imaging systems. *J Nucl Med Technol* 2015; **43**: 47–52. doi: <https://doi.org/10.2967/jnmt.114.147447>
20. Roeland, B., Marianne, F., Martin, T., and Remko, B. Manual CDRAD 2.0 phantom & Analyser software version 2.1. The Netherlands: Artinis Medical Systems; 2014.
21. De Crop A, Bacher K, Van Hoof T, Smeets PV, Smet BS, Vergauwen M, et al. Correlation of contrast-detail analysis and clinical image quality assessment in chest radiography with a human cadaver study. *Radiology* 2012; **262**: 298–304. doi: <https://doi.org/10.1148/radiol.11110447>
22. De Hauwere A, Bacher K, Smeets P, Verstraete K, Thierens H. Analysis of image quality in digital chest imaging. *Radiat Prot Dosimetry* 2005; **117**(1–3): 174–7. doi: <https://doi.org/10.1093/rpd/nci748>
23. Bacher K, Smeets P, Vereecken L, De Hauwere A, Duyck P, De Man R, et al. Image quality and radiation dose on digital

- chest imaging: comparison of amorphous silicon and amorphous selenium flat-panel systems. *AJR Am J Roentgenol* 2006; **187**: 630–7. doi: <https://doi.org/10.2214/AJR.05.0400>
24. Pascoal A, Lawinski CP, Honey I, Blake P. Evaluation of a software package for automated quality assessment of contrast detail images--comparison with subjective visual assessment. *Phys Med Biol* 2005; **50**: 5743–57. doi: <https://doi.org/10.1088/0031-9155/50/23/023>
 25. Hogg P, Blindell. Software for image quality evaluation using a forced choice method, in *United Kingdom RadiolConf. (p. 139). Br. Inst. Radiol. Manchester/London, UK*; 2012.
 26. Commission of the European Communities (CEC). European Guidelines on Quality Criteria for Diagnostic Radiographic Images:(EUR 16260 EN). Luxembourg: CEC; 1996.
 27. Ma WK, Hogg P, Tootell A, Manning D, Thomas N, Kane T, et al. Anthropomorphic chest phantom imaging – The potential for dose creep in computed radiography. *Radiography* 2013; **19**: 207–11. doi: <https://doi.org/10.1016/j.radi.2013.04.002>
 28. The Royal College of Radiologists, Picture archiving and communication systems (PACS) and guidelines on diagnostic display devices. 2012;.
 29. Allen E, Hogg P, Ma WK, Szczepura K. Fact or fiction: An analysis of the 10 kVp ‘rule’ in computed radiography. *Radiography* 2013; **19**: 223–7. doi: <https://doi.org/10.1016/j.radi.2013.05.003>
 30. Field A. *Discovering statistics using SPSS statistics*. Third edit . London: Sage publications Ltd; 2009.
 31. Cohen J. Statistical power analysis for the behavioral sciences. *Hillsdale.* " NJ: Lawrence Earlbaum Associates 1988; **2**.
 32. Koo TK, Li MY. A Guideline of Selecting and Reporting Intraclass Correlation Coefficients for Reliability Research. *J Chiropr Med* 2016; **15**: 155–63. doi: <https://doi.org/10.1016/j.jcm.2016.02.012>
 33. Portney L, Watkins M. *Foundations of clinical research: applications to practice*. New Jersey: Prentice Hall; 2000.
 34. Samei E, Eyler W, Baron L. Effects of anatomical structure on signal detection, in *Handb. Med. imaging. Vol. 1. Phys. Psychophys* 2000; 655–82.
 35. Warren LM, Mackenzie A, Cooke J, Given-Wilson RM, Wallis MG, Chakraborty DP, et al. Effect of image quality on calcification detection in digital mammography. *Med Phys* 2012; **39**(6Part1): 3202–13. doi: <https://doi.org/10.1118/1.4718571>
 36. Geijer H, Persliden J. Varied tube potential with constant effective dose at lumbar spine radiography using a flat-panel digital detector. *Radiat Prot Dosimetry* 2005; **114**: 240–5. doi: <https://doi.org/10.1093/rpd/nch509>
 37. Mackenzie A, Warren LM, Wallis MG, et al. The relationship between cancer detection in mammography and image quality measurements, *Phys. Medica* 2016; **32**: 568–74.
 38. Fausto AM, Lopes MC, de Sousa MC, Furquim TA, Mol AW, Velasco FG. Optimization of Image Quality and Dose in Digital Mammography. *J Digit Imaging* 2017; **30**: 185–96. doi: <https://doi.org/10.1007/s10278-016-9928-3>
 39. Tingberg A, Herrmann C, Besjakov J, et al. Evaluation of lumbar spine images with added pathology. *Proc. SPIE* 2000; **3981**: 34–42.
 40. Tingberg A. Quantifying the quality of medical X-ray images. An evaluation based on normal anatomy for lumbar spine and chest radiography. 2000.
 41. Sund P, Herrmann C, Tingberg A, et al. Comparison of two methods for evaluating image quality of chest radiographs. *Proc. SPIE* 2000; **3981**(April): 251–7.
 42. Tingberg A, Båth M, Håkansson M, et al. Comparison of two methods for evaluation of the image quality of lumbar spine radiographs. *Radiat. Prot. Dosimetry* 2000; **90**: 165–8.
 43. Moore CS, Liney GP, Beavis AW, Saunderson JR. A method to produce and validate a digitally reconstructed radiograph-based computer simulation for optimisation of chest radiographs acquired with a computed radiography imaging system. *Br J Radiol* 2011; **84**: 890–902. doi: <https://doi.org/10.1259/bjr/30125639>
 44. Moore CS, Avery G, Balcam S, Needler L, Swift A, Beavis AW, et al. Use of a digitally reconstructed radiograph-based computer simulation for the optimisation of chest radiographic techniques for computed radiography imaging systems. *Br J Radiol* 2012; **85**: e630–e639. doi: <https://doi.org/10.1259/bjr/47377285>
 45. Moore CS, Wood TJ, Avery G, Balcam S, Needler L, Smith A, et al. Investigating the use of an antiscatter grid in chest radiography for average adults with a computed radiography imaging system. *Br J Radiol* 2015; **88**: 20140613. doi: <https://doi.org/10.1259/bjr.20140613>