Published in final edited form as:

Magn Reson Med. 2019 January ; 81(1): 153–166. doi:10.1002/mrm.27360.

Black Blood Myocardial T₂ mapping

Chengyan Wang^{1,2,*}, Jihye Jang^{1,3,*}, Ulf Neisius¹, Maryam Nezafat¹, Ahmed Fahmy^{1,4}, Jinkyu Kang¹, Jennifer Rodriguez¹, Beth Goddu¹, Patrick Pierce¹, Sophie Berg¹, Jue Zhang², Xiaoying Wang^{2,5}, and Reza Nezafat¹

¹Department of Medicine (Cardiovascular Division), Beth Israel Deaconess Medical Center and Harvard Medical School, Boston, MA

²Academy for Advanced Interdisciplinary Studies, Peking University, Beijing, China

³Department of Computer Science, Technical University of Munich, Munich, Germany

⁴Biomedical Engineering Department, Cairo University, Giza, Egypt

⁵Department of Radiology, Peking University First Hospital, Beijing, China

Abstract

Purpose—To develop a Black blood heart-rate Adaptive T₂-prepared balanced steady-state free-precession (bSSFP) (BEATS) sequence for myocardial T₂ mapping.

Methods—In BEATS, blood suppression is achieved by using a combination of pre-excitation and double inversion recovery pulses. The timing and flip angles of the pre-excitation pulse are auto-calculated in each patient based on heart rate. Numerical simulations, phantom and *in-vivo* studies were conducted to evaluate the performance of BEATS. BEATS T₂ maps were acquired in 36 patients referred for clinical cardiac MRI and in one swine with recent myocardial infarction. Two readers assessed all images acquired in patients to identify presence of artifacts associated with slow blood flow.

Results—Phantom experiments showed that the BEATS sequence provided accurate T₂ values over a wide-range of simulated heart rates. Black blood myocardial T₂ maps were successfully obtained in all subjects. No significant difference was found between the average T₂ measurements obtained from the BEATS and conventional bright-blood T₂; however, there was a decrease in precision using the BEATS sequence. A suppression of the blood pool resulted in sharper definition of the blood-myocardium border and reduced partial voluming effect. The subjective assessment showed that 16% (18 out of 108) of short-axis slices have residual blood artifacts (12 in the apical slice, 4 in the mid-ventricular slice and 2 in the basal slice).

Conclusion—The BEATS sequence yields dark blood myocardial T₂ maps with better definition of the blood-myocardium border. Further studies are warranted to evaluate diagnostic accuracy of black blood T₂ mapping.

Correspondence to: Reza Nezafat, Ph.D., Beth Israel Deaconess Medical Center, 330 Brookline Ave, Boston, MA, 02215, Phone: 617-667-1747; Fax: 617-975-5480, rnezafat@bidmc.harvard.edu.

*The first two authors have equal contribution to this work.

Keywords

black blood; double inversion recovery; myocardial T₂ mapping; partial volume effect; myocardial edema

Introduction

Inflammatory processes underlie a broad spectrum of conditions that injure the myocardium and cause both structural and functional deficits (1). In the presence of inflammation or edema, myocardial T₂ increases and can be detected using T₂-weighted MRI (2,3).

Myocardial T₂ mapping is a quantitative technique that has emerged as a reproducible method to accurately measure T₂ relaxation times (4,5). Compared to conventional T₂-weighted imaging, myocardial T₂ mapping reduces the intersubject and interstudy variability. T₂ mapping has proven advantageous for visualizing myocardial edema in acute MI cases with both high sensitivity and specificity, yielding improved diagnostic accuracy compared to T₂-weighted dark-blood imaging.

T₂ mapping can be performed using a single-shot (SS) turbo spin-echo (TSE) sequence with multiple echoes (6). However, the TSE sequence is limited by cardiac motion artifacts due to a long acquisition window, which greatly affects the accuracy of T₂ measurements. An alternate method used for myocardial T₂ mapping is T₂-preparation (T₂prep) (7–9) with balanced steady-state free precession (bSSFP) imaging or gradient echo (4,10,11). A combination of spin echo excitations with gradient echo readouts (GraSE) was also recently proposed as part of this method? (12). T₂prep-bSSFP based T₂ mapping sequence provides very reproducible T₂ measurements (13) and has replaced the TSE based approach on clinical recommendation (14). In this sequence, T₂prep pulses with different echo times are used to create T₂-decays of the magnetization for the given duration of the echo time. A set of images are acquired with different echo times and used to generate a T₂ map through voxel-wise curve-fitting using either a 2-parameter or 3-parameter exponential model (15). However, T₂quantification in areas bordering the myocardium and the blood pool is challenging due to partial volume errors that cause T₂ measurement errors. While motion correction can improve measurement reproducibility (16), it cannot completely resolve the partial voluming error. One approach to resolve this limitation is to suppress the blood pool in myocardial T₂ mapping. Doing so would effectively reduce the partial volume effects to improve the image contrast at the blood-myocardium boundaries allowing for improved image registration and motion correction. However, blood suppression in myocardial T₂ mapping is challenging and there is currently no available sequence for black blood T₂ mapping.

There are numerous blood suppression preparation sequences in cardiac and vascular imaging (17–25). Double inversion recovery (DIR) (18–25) sequencing is the most widely used technique, in which a non-selective 180° inversion pulse is followed by a slice-selective 180° inversion pulse to invert the blood signal outside the imaging plane (25). Spatial saturation pulses (26,27) in which a volume upstream from the slice is pre-excited and dephased has also been used in cardiac imaging. A motion-sensitizing magnetization

preparation technique (20,21,28,29) can also suppresses the blood pool by dephasing all moving blood spins. However, these preparation sequences are not directly applicable to myocardial T_2 mapping. DIR sequence suffers from the problem of insufficient recovery time for T_2 mapping within a single R-R interval. Spatial saturation pulses cannot robustly suppress the blood chamber due to complex blood motion. DIR preparation also suffers from slow blood flow. Furthermore, motion sensitized techniques cause additional signal suppression in the myocardium that impact T_2 measurements. Therefore, there is an unmet clinical need for developing a robust black blood T_2 mapping sequence.

In this study, we sought to develop a black blood myocardial T_2 mapping sequence. Numerical simulation and phantom experiments were used to evaluate the proposed sequence. *In vivo* studies in healthy subjects, patients, and in an animal model were used to investigate the efficacy of the sequence to suppress the blood pool in myocardial T_2 mapping.

Methods

All the scans were performed on a 1.5T MR system (Philips Achieva, Philips Healthcare, Best, The Netherlands), with maximum amplitude of 40 mT/m and a slew rate of 150 mT/m/ms. A 32-channel cardiac coil was used for signal reception.

Pulse Sequence Design

We propose a “Black blood hEart-rate Adaptive T_2 -prepared bSSFP (BEATS)” sequence (Figure 1(a)) for myocardial T_2 mapping. The pulse sequence consists of 3 preparation pulses: a) a pre-excitation non-selective radio frequency (RF) pulse with a flip angle of β , b) DIR pulses, and c) T_2 prep pulses. Imaging was performed using bSSFP readouts, similar to a conventional bright blood T_2 mapping sequence (15).

A DIR pulse is combined with T_2 preparation pulses to obtain T_2 mapping with blood suppression. However, blood suppression by DIR preparation pulses requires a long delay time for blood magnetization recovery (due to the long blood T_1 relaxation time), which is not feasible within a single R-R interval. Therefore, we propose to insert a pre-excitation pulse with an adaptive flip angle prior to the DIR pulse (Figure 1(a)). The pre-excitation pulse is implemented as a non-selective RF pulse followed by dephasing gradients. The flip angle of this pre-excitation pulse is adaptively selected based on the patient’s heart rate to ensure that the blood signal is fully suppressed within an R-R interval. After the pre-excitation pulse, DIR pulses are performed subsequently with a pair of inversion pulses: a nonselective 180° inversion pulse followed immediately by a slice-selective 180° inversion pulse. The width of the selective inversion pulse is three times that of the imaging slice thickness. Subsequently, T_2 prep pulses, including an initial 90°_x pulse, followed by two composite 180°_y pulses and a final 90°_{-x} flipback pulse (7,15), are applied to generate T_2 weighting with variable echo times (TE). Finally, black blood T_2 -weighted images are acquired with single-shot bSSFP readouts. A rest period of 3 heart beats is used following each image acquisition to ensure sufficient magnetization recovery.

The longitudinal signal changes during the BEATS sequence are shown in Figure 1(b). After the pre-excitation pulse, the longitudinal magnetizations (M_z) are reduced to $\cos(\beta)$ of the initial magnetization M_0 , and then recovered according to the T_1 relaxation times:

$$M_z(TD_\beta) = M_0 + M_0(\cos(\beta) - 1) \times e^{-TD_\beta/T_1}. \quad [1]$$

where M_0 is the initial longitudinal magnetization, β is the flip angle of the pre-excitation pulse, TD_β is the duration between the pre-excitation pulse and the nonselective 180° pulse in DIR. After the application of DIR pulses, the blood signal is immediately inverted while the myocardial signal remains. During the TI time (i.e. the time between the nonselective 180° pulse and the start of T_2 prep pulses), blood and myocardial signals recover as follows:

$$M_{z_{blood}}(TI) = M_{0,blood} - (M_{0,blood} - M_{z_{blood}}(TD_\beta)) \times e^{-TI/T_1}. \quad [2]$$

$$M_{z_{myo}}(TI) = M_{0,myo} - (M_{0,myo} - M_{z_{myo}}(TD_\beta)) \times e^{-TI/T_1}. \quad [3]$$

According to Eq. [2], if TI is selected as $T_1 \times \ln((M_0 + M_{z_{blood}}(TD_\beta))/M_0)$, the blood signal can be fully nulled before the T_2 prep pulses. However, since the heart rate varies among different patients, it is difficult to fix the timing of TI . Therefore, we propose an adaptive timing design in which TI and β are automatically calculated based on the patient heart rate. The TI is set to $T_{dia} - TD_\beta - TE_{T2prep}$ to acquire the data during the diastolic rest phase, where T_{dia} is the trigger delay time and TE_{T2prep} is the echo time of the T_2 preparation pulse. The β is then calculated according to Eq. [1] and Eq. [2], and given by:

$$\beta = \arccos(1 + e^{-TD_\beta/T_1_{blood}} \times (e^{-(T_{dia} - TD_\beta - TE_{T2prep})/T_1_{blood}} - 2)). \quad [4]$$

TI and β are adaptive to different heart rates in the BEATS sequence (Figure 1(c)), with TI_{blood} equal to 1600 ms (30). A pre-excitation pulse duration (~0.2 ms), dephasing gradients (3.1 ms), a non-selective 180° inversion pulse (6.1 ms), and a TD_β of 20 ms (fixed empirically) were applied in this study.

Numerical Simulations

Bloch simulation was performed to investigate the influence of heart rate (beats per minute (bpm)) and tissue T_1 times both on the efficiency of blood suppression and on the recovery of the myocardial signal in the BEATS sequence. The heart rates were simulated from 40 bpm to 130 bpm, while the blood and myocardial T_1 values were simulated ranging from 1200 ms to 2000 ms and from 600 ms to 1600 ms, respectively. The value of β was calculated according to Eq. [4] assuming a fixed arbitrary value for TD_β (20 ms) and TI_{blood} (=1600 ms). TI was estimated as a function of heart rate (HR):

$$TI \cup 60(s)/HR - T_{rest} \cdot \cdot \quad [5]$$

where T_{rest} was the remaining time (ms) in each RR interval (assumed to be 250 ms), including the acquisition window, trigger delay time and pulse durations, and HR is heart rate. The blood and myocardium suppression were evaluated by calculating $SI_{blood}/SI_{blood,0}$ and $SI_{myo}/SI_{myo,0}$, in which SI_{blood} (or SI_{myo}) was the residual blood (or myocardium) signal after the inversion recovery, and $SI_{blood,0}$ (or $SI_{myo,0}$) represented the initial blood (or myocardium) signal intensity before suppression.

Phantom Study

To evaluate the performance of the BEATS sequence, phantom scans were performed on 6 NiCl₂ doped agarose vials, with similar T_2 values as the myocardial T_2 (~40 ms) but different T_1 values (31). Both BEATS and conventional bright blood T_2 mapping were performed on the phantom. The scanning parameters of the phantom were as follows: field of view (FOV) = 320×320 mm², in-plane resolution = 2×2 mm², slice number = 1, slice thickness = 8 mm, SENSE factor = 2.5, bSSFP readout with repetition time/echo time (TR/TE) = 2.8/1.4 ms, flip angle = 60°, 10 linear ramp-up pulses, linear k-space ordering, and acquisition window = 150 ms. A total of 9 different T_2 prep echo times were used, including $TE_{T2prep} = 0$, $TE_{T2prep} = \infty$ and TE_{T2prep} values ranging from 25 ms to 95 ms with linear steps of 10 ms (15). The TE_{T2prep} of ∞ was representative of the long T_2 prep echo time achieved by applying a saturation pulse prior to image acquisition. Heart rates of 60 bpm were simulated for the conventional bright blood T_2 mapping. For the BEATS scans, heart rates of 60, 70, 80, 90, 110 and 130 bpm were simulated to test whether they affected T_2 measurement accuracy. To evaluate the repeatability of the T_2 measurements, the BEATS sequence scan was repeated 3 times with 10 min between scans.

Reference T_2 relaxation times were obtained using a spin-echo (SE) sequence of 32 echoes. The scanning parameters were as follows: FOV = 320×320 mm², in-plane resolution = 2×2 mm², slice thickness = 8 mm, flip angle = 90°, TR = 10 s, TEs from 10 ms to 320 ms with gaps of 10 ms, and number of signal averages (NSA) = 4. Reference T_1 relaxation times were acquired using the SE sequence with 16 inversion times of 50, 100, 200, 300, 400, 500, 600, 700, 800, 900, 1000, 1250, 1500, 1750, 2000, and 3000 ms. The other imaging parameters were: FOV = 320×320 mm², in-plane resolution = 2×2 mm², slice thickness = 8 mm, flip angle = 90°, TR/TE = 10 s/10 ms, and NSA = 4. Both T_1 and T_2 values from the SE images were calculated using a three-parameter fit model (32,33).

In-vivo Human Study

This study was approved by our local Institutional Review Board, and written informed consents were obtained from all participants. *In vivo* human studies were performed on 8 healthy adult subjects (26.0 ± 4.7 years, range: 21–35 years, 2 males) and 36 patients (55.1 ± 16.6 years, range: 20–82 years, 20 males) who were referred for clinical cardiac MR examinations. The patient population consisted of patients with suspected or known hypertrophic cardiomyopathy ($N=5$), non-ischemic cardiomyopathy ($N=9$), suspected or

confirmed ischemic cardiomyopathy ($N=4$), arrhythmias ($N=8$), valvular heart disease ($N=2$), left ventricular hypertrophy ($N=3$), lipomatous atrial septal hypertrophy ($N=1$), athletic cardiomyopathy ($N=1$), suspected or confirmed myo- or pericarditis ($N=2$), and cardiotoxic cardiomyopathy ($N=1$). For each subject, localizer scout images were first acquired to define the mid-ventricular short-axis slice. For both BEATS and conventional bright blood T_2 prep-bSSFP imaging, free-breathing ECG-triggered single-shot bSSFP readout was used for data acquisition. Sequence parameters were as follows: FOV = 320×320 mm², in-plane resolution = 2×2 mm², slice number = 3, slice thickness/gap = 8/4 mm, SENSE factor = 2.5, TR/TE = 2.8/1.4 ms, flip angle = 60°, and 9 T_2 prep echo times including $TE_{T2prep} = 0, 25, 35, 45, 55, 65, 75, 85, 95, \infty$ ms. A two-dimensional spiral-beam navigator was placed at the lung-liver interface of the right hemi-diaphragm, and was used for gating with a window of 5 mm. The typical scan time for each slice was ~40 sec. To assess the scan repeatability of black blood T_2 imaging using the BEATS sequence, each scan was repeated 3 times, 10 min apart, in only healthy adult subjects. Since recruited patients included those referred for clinical cardiac MR examinations, there was not sufficient time to assess repeatability in patients.

Animal Study

The protocol was approved by The Beth Israel Deaconess Medical Center Institutional Animal Care and Use Committee (IACUC) and conformed to the position of the American Heart Association on Research Animal Use and the Declaration of Helsinki. Ischemia-reperfusion mediated myocardial infarction was induced in one Yorkshire swine (35 kg) as previously described (34,35). Under fluoroscopic guidance, an angioplasty balloon was inflated in the mid left anterior descending (LAD) for 180 minutes. The balloon was then deflated and withdrawn to create an ischemia-reperfusion mediated MI. The animal was imaged one week after infarct creation using an imaging protocol, which included both conventional bright blood T_2 mapping and the BEATS sequence with similar imaging parameters as described above in our human study.

Data Analysis

T_2 maps were reconstructed offline using a custom-written program on MATLAB (MathWorks, Natick, Massachusetts, USA). A three-parameter fitting model was applied for pixel-wise T_2 quantification (15):

$$S_{myo}(TE_{T2prep}) = A \times e^{-TE_{T2prep}/T_2} + B. \quad [6]$$

where TE_{T2prep} is the echo time, A is a function of the signal at full-recovery, T_2 is the myocardial T_2 relaxation time, and B characterizes the effect of the bSSFP imaging pulses that were played until the acquisition of the central k -space line. Given a set of images acquired at different (known) echo times, the model in Eq. [6]. is used to fit the intensities at each voxel to estimate the parameter T_2 .

In the phantom experiments, region-of-interest (ROI) analysis was performed for each vial, where the mean value and standard deviation were recorded. Identical ROIs were used in all

the T_2 maps with different simulated heart rates. SNR was calculated in the source images at $TE_{T2prep} = 0$ ms, as the average signal in the phantom divided by the standard deviation of the noise region (33). ROI for the noise measurement was selected from a background area free of image artifacts.

For the *in vivo* study, the acquired images were co-registered using an affine co-registration algorithm (16,36) before T_2 fitting. Voxel-wise curve fitting was then performed on both black blood and conventional bright blood T_2 prep-bSSFP images according to Eq. [5]. To avoid fitting for the residual blood signal in black blood T_2 prep-bSSFP images, a threshold processing was performed. The threshold was selected empirically as 20% of the myocardial signal intensity in the black blood T_2 prep-bSSFP image with $TE_{T2prep} = 0$. Regions with signal intensities below this threshold were avoided by the T_2 fitting. The myocardium was manually traced by drawing the epicardial and endocardial contours. T_2 profiles of a cross section of the mid-ventricular slice were obtained on black blood and conventional bright blood T_2 maps. To evaluate the effectiveness of blood suppression, the blood to myocardial signal ratio was measured for each subject.

Furthermore, a segment-based analysis was performed to investigate the regional T_2 variations. Six segments were used in the mid-ventricular slice, in accordance with the American Heart Association (AHA) segmentation guidelines (37). In each segment, the mean T_2 value and its standard deviation were calculated. In clinical patients, inter-observer agreement for T_2 quantification was assessed by using intraclass correlation coefficient (ICC) analysis from 10 randomly selected patients.

A subjective analysis was conducted by two experienced readers (UN and RN) in consensus to assess artifacts due to blood flow from DIR pulses in BEATS. Readers evaluated the T_2 maps (reconstructed without performing any thresholding of signaling as described above) and original T_2 weighted images to identify slices with artifacts associated with incomplete suppression of the blood pool. The number of slices impacted by the artifacts are reported for apical, mid-ventricular and basal slices. Furthermore, for patients with flow artifacts, readers indicated whether the presence of flow artifact impacted their confidence in reliably measuring myocardial T_2 values.

Statistical Analysis

Data were reported as median [25th percentile; 75th percentile]. The accuracy of the BEATS T_2 measurement was assessed as the difference between the black blood T_2 maps and reference T_2 maps acquired with the standard SE sequence. Precision was assessed as the standard deviation within the ROIs for the given T_2 map. A higher standard deviation indicates lower precision. Repeatability was defined as the standard deviation over the three repeated scans of the spatial average T_2 values in each subject. The accuracy, precision, and repeatability of the values were measured in each segment of the myocardium. The Wilcoxon signed-rank test was conducted to compare the T_2 measurements with different methods. A P -value of less than 0.05 was considered statistically significant.

Results

Numerical Simulation

In the numerical simulation, both heart rate and blood T_1 relaxation time influence blood suppression (Figure 2(a)). The remaining blood signal dropped as the simulated heart rate increased, as consequence of limited T_1 decay time. However, the residual blood signal after inversion recovery was maintained within 22.7% of the original blood signal for a wide range of simulated blood T_1 values ranging from 1200 to 2000 ms.

Myocardial signal recovery was influenced by the heart rate and the myocardial T_1 relaxation time (Figure 2(b)). As expected, the heart rate showed a strong influence on the myocardial signal intensity. For example, at $T_1=1200$ ms, the relative myocardial signal dropped as the simulated heart rate increased (1.00, 0.91, 0.79, 0.68, 0.58, 0.51, 0.44, 0.38, 0.33 for simulated heart rates of 40, 50, 60, 70, 80, 90, 100, 110 and 120 bpm, respectively), as a result of reducing TI. In general, for $T_1 < 1200$ ms and heart rate < 85 bpm, more than 69.3% of the myocardial signal was preserved.

Phantom

T_2 maps generated from the SE, conventional bright blood and BEATS sequences are shown in Figure 3. The T_2 measurements of the phantoms using conventional bright blood and the BEATS sequence with different simulated heart rates are shown in Supporting Figure S1. Compared to T_2 measured from conventional bright blood T_2 mapping, the average T_2 measurements in BEATS T_2 maps were not affected with increased heart rates within a range of 60 to 110 bpm; however higher heart rates decreased measurement precision (observed as increased standard deviation). The SNR within the different vials decreased as the heart rate increased (Figure 4(a)) due to insufficient time for signal recovery. Good repeatability was achieved for the BEATS T_2 mapping, as measured by the standard deviations over repeated scans (0.29 ms, 0.35 ms, 0.35 ms, 0.29 ms, 0.42 ms for heart rate of 60, 70, 80, 90, 100 bpm, respectively). The phantom experiment confirmed the accuracy of the BEATS T_2 mapping (Figure 4(b)). As expected, $T_{2\text{prep}}$ and heart rate had no impact on the accuracy of T_2 measurements in a range of 60–110 bpm, but did have an impact at heart rates higher than 130 bpm. There was no difference in T_2 values in vials measured using the BEATS or conventional bright blood T_2 mapping sequences over heart rates of 60 to 130 bpm.

In-vivo Human Study

Black blood myocardial T_2 images were successfully obtained in all subjects using the BEATS sequence. The average heart rate was 67 ± 5 bpm (range: 55–75 bpm) for healthy subjects and 66 ± 11 bpm (range: 49–96 bpm) for patients. The blood to myocardial signal ratio was reduced from 3.03 [2.56; 3.72] to 0.15 [0.13; 0.17] ($P = 0.01$) in healthy subjects.

T_2 -weighted images with $TE_{T_2\text{prep}} = 0, 25$ and 55 ms and fitted T_2 maps acquired by BEATS and conventional bright-blood T_2 mapping sequences are shown in two healthy subjects with different heart rates: 70 bpm (subject 1) and 55 bpm (subject 2) (Figure 5). In both cases, the BEATS sequence produced T_2 -weighted images without any visual artifact.

Both images demonstrate a relatively high contrast between the myocardium and the blood pool with defined borders. The partial volume effect was significantly reduced with the suppression of intraventricular blood signals. Compared to conventional bright blood T₂ mapping, BEATS T₂ mapping successfully reduced the blood signal, leading to better depiction of the myocardial structures. Signal intensity profiles were drawn in the T₂ maps across the myocardium (Figure 6, red lines). The T₂ profiles have a much sharper transition between blood and myocardium compared to that in conventional bright blood T₂ maps, suggesting a decrease in the partial volume effect of myocardial T₂. Segment-based measurements show comparable T₂ values calculated from BEATS and bright blood T₂ mapping sequences (Figure 6). Example T₂ maps from our patients are presented in Figure 7, and show good image quality and maps. The original T₂-weighted images are included in Supporting Videos S1–2. Myocardial edema was visually present in only one patient with a recent history of myocardial infarction (Figure 8). Arrows show areas of higher T₂ values (59–63ms) located at areas with recent infarct. A signal difference in this region can also be seen on original black blood T₂ weighted images.

No significant difference was found between the average T₂ measurements obtained from the BEATS and conventional bright-blood T₂ maps (48.8 ms [48.2; 50.0] vs. 48.7 ms [48.0; 49.9], $P = 0.67$ in healthy subjects and 48.9 ms [47.1; 51.7] vs. 50.4 ms [48.7; 51.8], $P = 0.05$ in patients) (Figure 9). The interobserver agreement for both T₂ maps was strong with an ICC of 0.74 (95% confidence interval: 0.29 – 0.93) for BEATS, and ICC of 0.93 (95% confidence interval: 0.75 – 0.98) for the bright-blood T₂ map. Homogeneous T₂ maps, as quantified by precision, were observed throughout the myocardium using both techniques, with a decrease in precision for the BEATS sequence due to loss of signal associated with the DIR pulse (healthy subjects: 6.87 ms [6.2; 7.3] vs. 4.9 ms [4.7; 6.6] $P = 0.01$, patients: 8.4 ms [7.3; 11.1] vs. 6.7 ms [5.7; 7.8], $P < 0.001$). Repeatability analysis in healthy subjects shows that the variability of T₂ measurements was low and thus comparable between the two techniques (BEATS, 2.0 ms [1.6; 2.5]; bright-blood T₂, 1.95 ms [1.6; 2.2], $P = 0.58$).

The subjective assessment showed that 16% (18 out of 108) of short-axis slices have artifacts due to a residual blood signal from the DIR pulse in BEATS (12 in the apical slice, 4 in the mid-ventricular slice and 2 in the basal slice). However, the slow flowing blood artifact only impacted measurements in 10 slices (6 at the apex, 3 at the mid-ventricle, and 1 at the base). The flow artifacts often resulted in large T₂ values or noisy areas with random values that could be identified by reviewing original T₂ weighted images. The location of artifacts was random and varied among different patients including middle of the left ventricular cavity, the subendocardial region, and around the papillary muscle where blood is trapped between the papillary muscle and the myocardium (Supporting Figure S2). There was no correlation between presence of flow-related artifacts and heart rate (Spearman's rho = 0.19, $P = 0.27$), and artifacts were present in patients with both low and high heart rates (Supporting Figure S3, Supporting Video S3 shows images of a patient with a heart rate of 96 bpm). In one patient, there was a large subendocardial enhancement in BEATS T₂ maps, consistent with the patient's clinical history; however, readers (UL and RN) could not confirm with confidence if the increased T₂ was due to presence of edema or slow blood flow (Supporting Figure S4, Supporting Video S4).

Animal Study

T_2 weighted images and corresponding T_2 maps in swine confirms presence of edema (red arrows) in the anterior septum, consistent with the location of the infarct in our animal model (Figure 10). The measured T_2 in an ROI within the edematous region was 82 ms for conventional bright blood T_2 mapping and 81 ms for BEATS T_2 mapping. The extent of edema was better depicted in the BEATS T_2 map due to the suppression of the blood signal. Furthermore, the T_2 weighted images with the suppressed blood pool in the BEATS sequence yielded improved depiction of myocardial edema in the anterior septum (red arrows), further confirming the presence of edema.

Discussion

In this study, we present a black blood myocardial T_2 mapping sequence to reduce the partial voluming effect present in conventional bright blood T_2 mapping by suppressing the blood signal using a combination of pre-excitations and DIR pulses. The performance of the BEATS T_2 mapping sequence was evaluated by numerical simulations, phantom and *in vivo* studies. The results of our study demonstrated that the BEATS sequence yields images with suppressed blood pool signal, resulting in better definition of the blood-myocardium border and improved visualization of edema/inflammation.

In the BEATS sequence, selection of appropriate timings of the pre-excitations and DIR pulses are essential for blood suppression and the minimization of SNR loss. These parameters should be adaptively selected based on an individual heart rate to achieve blood suppression. A lower flip angle of the pre-excitation pulse will minimize the myocardial signal loss and improve SNR, but longer TI will be required. In the BEATS sequence, we proposed an adaptive parameter design according to the patient's heart rate. TI and the flip angle of the pre-excitation pulse are auto-calculated in the sequence to achieve optimal blood suppression with minimal myocardial signal loss. In our study, we applied a three-parameter fitting model to 9 T_2 -weighted images with an additional saturated image acquired during free-breathing to estimate the T_2 maps (15). However, T_2 maps can be estimated with as low as 4 images within a single breath-hold, albeit with lower precision (38).

One of the challenges of the DIR preparation pulse in black blood T_2 mapping is the limitation of the number of slices that can be acquired per scan. There have been recent advances in myocardial T_2 mapping to increase coverage and data acquisition efficiency using either whole heart or slice-interleaved acquisition (39–43). Our current implementation of BEATS limits the acquisition to a single slice per acquisition. Improved DIR for multi-slice imaging has been previously demonstrated (22–24), and could potentially be adopted for BEATS T_2 mapping and warrant further investigation.

Inflammation or edema could also occur in the thin wall of the left atrium (LA) or right ventricle (RV). For example, radio-frequency ablation for treatment of atrial fibrillation causes LA edema, which can be detected with T_2 weighted imaging (44–46); however T_2 weighted imaging in LA or RV is challenging and interpretation is subjective. Conventional bright blood T_2 maps do not allow accurate measurement of LA T_2 due to partial voluming.

The proposed black blood T_2 mapping could potentially be used to measure LA/RV T_2 ; however, it will require higher spatial resolution. Further validation and technical improvements are necessary to develop high-resolution black blood T_2 mapping for LA/RV.

We used a combination of a pre-excitation pulses with DIR to suppress the blood pool in BEATS. We only acquired T_2 maps in the short-axis view, since DIR does not effectively null the blood signal if blood flows within the imaging plane (Supporting Figure S5). One main problem with DIR could be slow moving blood along the myocardial border. This is a well-known problem in myocardial regions affected by disease (e.g. dyskinetic myocardium following MI), overall contractile motion anomalies as observed in myocarditis, or significant trabeculations. In such cases, the stagnant blood can appear bright in each individual T_2 Prep-SSFP image. However, blood will experience different preparation pulses in subsequent data acquisitions and the signal may not follow a simple T_2 decay. Therefore, the region may either appear noisy in the estimated T_2 maps or with very high T_2 values. Our preliminary data shows these types of artifacts could occur in different slices, and in patients with different heart rates. A combination of post-processing to take into account blood flow may be used to mitigate the issue. Further studies are needed to minimize artifact in the proposed black blood T_2 mapping sequence.

Our study has limitations. We used a predetermined fixed T_1 and did not measure T_1 values of the blood to optimize the imaging parameters. A patient-specific blood T_1 can be acquired using the T_1 mapping sequence, but this was not studied. For high heart rates >100 bpm, the myocardial signal will decrease, which could impact measurement accuracy and precision. For patients with high heart rates, DIR pulses and imaging readouts can be applied in two subsequent heartbeats; however, this will make the sequence sensitive to heart rate variability. The sample size of our patient data was small, and the majority of patients did not have edema or inflammation. Therefore, we cannot assess the diagnostic accuracy of BEATS; doing so requires a rigorous animal study with histological validation which is not within the scope of our study. There is no gold-standard non-invasive imaging modality with sufficient spatial resolution that can be used to assess sensitivity or specificity of black-blood T_2 mapping sequencing for identifying subendocardial inflammation in patients.

Conclusion

The proposed black blood myocardial T_2 mapping sequence efficiently suppresses the blood signal, resulting in better definition of the blood-myocardium border by reducing the impact of the partial volume effect that is present in conventional bright-blood T_2 mapping. The residual blood flow associated with DIR is a limitation of the proposed technique.

Supplementary Material

Refer to Web version on PubMed Central for supplementary material.

Acknowledgments

Funding: The project described was supported in part by National Institutes of Health 1R01HL129185-01, 1R21HL127650, 1R01HL129157; American Heart Association (AHA) 15EIA22710040; National Science Foundation of China 81627901, 81571666; China Postdoctoral Science Foundation 17Z102060138.

Authors would like to acknowledge Dr. Warren J. Manning for his editorial assistance.

References

1. Marchant DJ, Boyd JH, Lin DC, Granville DJ, Garmaroudi FS, McManus BM. Inflammation in myocardial diseases. *Circ Res*. 2012; 110(1):126–144. [PubMed: 22223210]
2. Mcnamara MT, Higgins CB, Schechtman N, Botvinick E, Lipton MJ, Chatterjee K, Amparo EG. Detection and characterization of acute myocardial infarction in man with use of gated magnetic resonance. *Circulation*. 1985; 71(4):717–724. [PubMed: 3971541]
3. Payne AR, Casey M, McClure J, McGeoch R, Murphy A, Woodward R, Saul A, Bi X, Zuehlsdorff S, Oldroyd KG. Bright-blood T2-weighted MRI has higher diagnostic accuracy than dark-blood short tau inversion recovery MRI for detection of acute myocardial infarction and for assessment of the ischemic area at risk and myocardial salvage. *Circ Cardiovasc Imaging*. 2011; 4(3):210–219. [PubMed: 21427362]
4. Huang T-Y, Liu Y-J, Stemmer A, Poncelet BP. T₂ measurement of the human myocardium using a T₂-prepared transient-state trueFISP sequence. *Magn Reson Med*. 2007; 57(5):960–966. [PubMed: 17457877]
5. Baeßler B, Schaarschmidt F, Stehning C, Schnackenburg B, Giolda A, Maintz D, Bunck AC. Reproducibility of three different cardiac T₂-mapping sequences at 1.5 T. *J Magn Reson Imaging*. 2016; 44(5):1168–1178. [PubMed: 27043352]
6. Kim D, Jensen JH, Wu EX, Sheth SS, Brittenham GM. Breathhold multiecho fast spin-echo pulse sequence for accurate R2 measurement in the heart and liver. *Magn Reson Med*. 2009; 62(2):300–306. [PubMed: 19526516]
7. Brittain JH, Hu BS, Wright GA, Meyer CH, Macovski A, Nishimura DG. Coronary angiography with magnetization-prepared T₂ contrast. *Magn Reson Med*. 1995; 33(5):689–696. [PubMed: 7596274]
8. Nezafat R, Ouwerkerk R, Derbyshire AJ, Stuber M, McVeigh ER. Spectrally selective B₁-insensitive T₂ magnetization preparation sequence. *Magn Reson Med*. 2009; 61(6):1326–1335. [PubMed: 19319903]
9. Nezafat R, Stuber M, Ouwerkerk R, Gharib AM, Desai MY, Pettigrew RI. B₁-insensitive T₂ preparation for improved coronary magnetic resonance angiography at 3 T. *Magn Reson Med*. 2006; 55(4):858–864. [PubMed: 16538606]
10. Giri S, Chung Y-C, Merchant A, Mihai G, Rajagopalan S, Raman SV, Simonetti OP. T₂ quantification for improved detection of myocardial edema. *J Cardiovasc Magn Reson*. 2009; 11(1):56. [PubMed: 20042111]
11. Foltz WD, Al-Kwif O, Sussman MS, Stainsby JA, Wright GA. Optimized spiral imaging for measurement of myocardial T₂ relaxation. *Magn Reson Med*. 2003; 49(6):1089–1097. [PubMed: 12768587]
12. Sprinkart AM, Luetkens JA, Träber F, Doerner J, Gieseke J, Schnackenburg B, Schmitz G, Thomas D, Homs R, Block W. Gradient Spin Echo (GraSE) imaging for fast myocardial T₂ mapping. *J Cardiovasc Magn Reson*. 2015; 17(1):12. [PubMed: 25885268]
13. Bellm S, Basha TA, Shah RV, Murthy VL, Liew C, Tang M, Ngo LH, Manning WJ, Nezafat R. Reproducibility of myocardial T₁ and T₂ relaxation time measurement using slice-interleaved T₁ and T₂ mapping sequences. *J Magn Reson Imaging*. 2016; 44(5):1159–1167. [PubMed: 27043156]
14. Messroghli DR, Moon JC, Ferreira VM, Grosse-Wortmann L, He T, Kellman P, Mascherbauer J, Nezafat R, Salerno M, Schelbert EB. Clinical recommendations for cardiovascular magnetic resonance mapping of T₁, T₂, T₂* and extracellular volume: A consensus statement by the Society for Cardiovascular Magnetic Resonance (SCMR) endorsed by the European Association for Cardiovascular Imaging (EACVI). *J Cardiovasc Magn Reson*. 2017; 19(1):75. [PubMed: 28992817]
15. Akçakaya M, Basha TA, Weingärtner S, Roujol S, Berg S, Nezafat R. Improved quantitative myocardial T₂ mapping: Impact of the fitting model. *Magn Reson Med*. 2015; 74(1):93–105. [PubMed: 25103908]

16. Roujol S, Basha TA, Weingärtner S, Akçakaya M, Berg S, Manning WJ, Nezafat R. Impact of motion correction on reproducibility and spatial variability of quantitative myocardial T₂ mapping. *J Cardiovasc Magn Reson*. 2015; 17(1):46. [PubMed: 26067275]
17. Koktzoglou I, Li D. Diffusion-prepared segmented steady-state free precession: application to 3D black-blood cardiovascular magnetic resonance of the thoracic aorta and carotid artery walls. *J Cardiovasc Magn Reson*. 2007; 9(1):33–42. [PubMed: 17178678]
18. Edelman RR, Chien D, Kim D. Fast selective black blood MR imaging. *Radiology*. 1991; 181(3): 655–660. [PubMed: 1947077]
19. Debbins JP, Riederer SJ, Rossman PJ, Grimm RC, Felmlee JP, Breen JF, Ehman RL. Cardiac magnetic resonance fluoroscopy. *Magn Reson Med*. 1996; 36(4):588–595. [PubMed: 8892212]
20. Nguyen TD, de Rochefort L, Spincemaille P, Cham MD, Weinsaft JW, Prince MR, Wang Y. Effective motion-sensitizing magnetization preparation for black blood magnetic resonance imaging of the heart. *J Magn Reson Imaging*. 2008; 28(5):1092–1100. [PubMed: 18972350]
21. Newbould RD, Owen DR, Shalhoub J, Brown AP, Gambarota G. Motion-sensitized driven equilibrium for blood-suppressed T₂* mapping. *J Magn Reson Imaging*. 2011; 34(3):702–709. [PubMed: 21769964]
22. Saranathan, M; Savin, GS. Multi-slice MR data acquisition in successive heartbeats with black blood contrast. US Patent. 7,412,277. Aug 12, 2008
23. Fayad, ZA; Itskovich, VV; Mani, V; Szimtenings, MM. Rapid multislice black blood double-inversion recovery technique for blood vessel imaging. US Patent. 7,369,887. May 6, 2008
24. Kuribayashi H, Tessier JJ, Checkley DR, Wang YX, Hultin L, Waterton JC. Effective blood signal suppression using double inversion-recovery and slice reordering for multislice fast spin-echo MRI and its application in simultaneous proton density and T₂ weighted imaging. *J Magn Reson Imaging*. 2004; 20(5):881–888. [PubMed: 15503346]
25. Simonetti OP, Finn JP, White RD, Laub G, Henry DA. "Black blood" T₂-weighted inversion-recovery MR imaging of the heart. *Radiology*. 1996; 199(1):49–57. [PubMed: 8633172]
26. Felmlee JP, Ehman RL. Spatial presaturation: a method for suppressing flow artifacts and improving depiction of vascular anatomy in MR imaging. *Radiology*. 1987; 164(2):559–564. [PubMed: 3602402]
27. Nayak KS, Rivas PA, Pauly JM, Scott GC, Kerr AB, Hu BS, Nishimura DG. Real-time black-blood MRI using spatial presaturation. *J Magn Reson Imaging*. 2001; 13(5):807–812. [PubMed: 11329205]
28. Srinivasan S, Hu P, Kissinger KV, Goddu B, Goepfert L, Schmidt EJ, Kozerke S, Nezafat R. Free-breathing 3D whole-heart black-blood imaging with motion sensitized driven equilibrium. *J Magn Reson Imaging*. 2012; 36(2):379–386. [PubMed: 22517477]
29. Wang J, Yarnykh VL, Hatsukami T, Chu B, Balu N, Yuan C. Improved suppression of plaque-mimicking artifacts in black-blood carotid atherosclerosis imaging using a multislice motion-sensitized driven-equilibrium (MSDE) turbo spin-echo (TSE) sequence. *Magn Reson Med*. 2007; 58(5):973–981. [PubMed: 17969103]
30. Nickander J, Lundin M, Abdula G, Sörensson P, Rosmini S, Moon JC, Kellman P, Sigfridsson A, Ugander M. Blood correction reduces variability and gender differences in native myocardial T₁ values at 1.5 T cardiovascular magnetic resonance—a derivation/validation approach. *J Cardiovasc Magn Reson*. 2017; 19(1):41. [PubMed: 28376820]
31. Captur G, Gatehouse P, Keenan KE, Heslinga FG, Bruehl R, Prothmann M, Graves MJ, Eames RJ, Torlasco C, Benedetti G. A medical device-grade T₁ and ECV phantom for global T₁ mapping quality assurance—the T₁ Mapping and ECV Standardization in cardiovascular magnetic resonance (T1MES) program. *J Cardiovasc Magn Reson*. 2016; 18(1):58. [PubMed: 27660042]
32. Weingärtner S, Roujol S, Akçakaya M, Basha TA, Nezafat R. Free-breathing multislice native myocardial T₁ mapping using the slice-interleaved T₁ (STONE) sequence. *Magn Reson Med*. 2015; 74(1):115–124. [PubMed: 25131652]
33. Dietrich O, Raya JG, Reeder SB, Reiser MF, Schoenberg SO. Measurement of signal-to-noise ratios in MR images: Influence of multichannel coils, parallel imaging, and reconstruction filters. *J Magn Reson Imaging*. 2007; 26(2):375–385. [PubMed: 17622966]

34. Tschabrunn CM, Roujol S, Nezafat R, Faulkner-Jones B, Buxton AE, Josephson ME, Anter E. A swine model of infarct-related reentrant ventricular tachycardia: electroanatomic, magnetic resonance, and histopathological characterization. *Heart Rhythm*. 2016; 13(1):262–273. [PubMed: 26226214]
35. Whitaker J, Tschabrunn CM, Jang J, Leshem E, O'Neill M, Manning WJ, Anter E, Nezafat R. Cardiac MR Characterization of Left Ventricular Remodelling in a Swine Model of Infarct Followed by Reperfusion. *J Magn Reson Imaging*. 2018; doi: 10.1002/jmri.26005
36. Roujol S, Foppa M, Weingartner S, Manning WJ, Nezafat R. Adaptive registration of varying contrast-weighted images for improved tissue characterization (ARCTIC): Application to T₁ mapping. *Magn Reson Med*. 2015; 73(4):1469–1482. [PubMed: 24798588]
37. Cerqueira MD, Weissman NJ, Dilsizian V, Jacobs AK, Kaul S, Laskey WK, Pennell DJ, Rumberger JA, Ryan T, Verani MS. Standardized myocardial segmentation and nomenclature for tomographic imaging of the heart. A statement for healthcare professionals from the Cardiac Imaging Committee of the Council on Clinical Cardiology of the American Heart Association. *Circulation*. 2002; 105(4):539–542. [PubMed: 11815441]
38. Basha, T; Akcakaya, M; Roujol, S; Nezafat, R. Precision and reproducibility of T₂ quantifications in myocardial T₂ mapping: impact of the number of echoes and reconstruction model. *Proceedings of the 18th Annual SCMR Scientific Sessions*; Nice, France. 2015. W9
39. Basha TA, Bellm S, Roujol S, Kato S, Nezafat R. Free-breathing slice-interleaved myocardial T₂ mapping with slice-selective T₂ magnetization preparation. *Magn Reson Med*. 2016; 76(2):555–565. [PubMed: 26479866]
40. Kvernby S, Warntjes MJB, Haraldsson H, Carlhäll C-J, Engvall J, Ebbens T. Simultaneous three-dimensional myocardial T₁ and T₂ mapping in one breath hold with 3D-QALAS. *J Cardiovasc Magn Reson*. 2014; 16(1):102. [PubMed: 25526880]
41. Yang HJ, Sharif B, Pang J, Kali A, Bi X, Cokic I, Li D, Dharmakumar R. Free-breathing, motion-corrected, highly efficient whole heart T₂ mapping at 3T with hybrid radial-cartesian trajectory. *Magn Reson Med*. 2016; 75(1):126–136. [PubMed: 25753385]
42. Heeswijk RB, Piccini D, Feliciano H, Hullin R, Schwitter J, Stuber M. Self-navigated isotropic three-dimensional cardiac T₂ mapping. *Magn Reson Med*. 2015; 73(4):1549–1554. [PubMed: 24809849]
43. Akcakaya M, Weingärtner S, Basha TA, Roujol S, Bellm S, Nezafat R. Joint myocardial T₁ and T₂ mapping using a combination of saturation recovery and T₂-preparation. *Magn Reson Med*. 2016; 76(3):888–896. [PubMed: 26418119]
44. Hsing J, Peters DC, Knowles BR, Manning WJ, Josephson ME. Cardiovascular magnetic resonance imaging of scar development following pulmonary vein isolation: a prospective study. *PLoS One*. 2014; 9(9):e104844. [PubMed: 25251403]
45. Dickfeld T, Kato R, Zviman M, Nazarian S, Dong J, Ashikaga H, Lardo AC, Berger RD, Calkins H, Halperin H. Characterization of acute and subacute radiofrequency ablation lesions with nonenhanced magnetic resonance imaging. *Heart Rhythm*. 2007; 4(2):208–214. [PubMed: 17275759]
46. Lardo AC, McVeigh ER, Jumrussirikul P, Berger RD, Calkins H, Lima J, Halperin HR. Visualization and temporal/spatial characterization of cardiac radiofrequency ablation lesions using magnetic resonance imaging. *Circulation*. 2000; 102(6):698–705. [PubMed: 10931812]

Figure 1.

a) pulse sequence diagram of the Black blood hEart-rate Adaptive T_2 -prepared bSSFP (BEATS) sequence, b) blood and myocardial longitudinal signal relaxation curves during the sequence. The pulse sequence consists of the pre-excitation pulse with an adaptive flip angle of β , double inversion recovery (DIR) pulses, a T_2 preparation (T_2 prep) pulse, and a balanced steady-state free-precession (bSSFP) acquisition. The bSSFP readout is synchronized to the end-diastole. The inversion time (TI) after DIR and the flip angle of the pre-excitation pulse are auto-calculated to adapt to the heart rate of individuals. c) shows the calculation of TI and β as a function of different heart rates in the BEATS sequence.

Figure 2.

Numerical simulation for a) the effect of heart rate and blood T₁ on the suppression of the blood signal ($SI_{\text{blood}}/SI_{\text{blood},0}$), and b) the effect of heart rate and myocardial T₁ on the recovery of the myocardial signal ($SI_{\text{myo}}/SI_{\text{myo},0}$). $SI_{\text{blood},0}$ and $SI_{\text{myo},0}$ are the initial blood and myocardial signals, respectively. Note that the remaining blood signal dropped as the simulated heart rate increased due to a limited T₁ decay time. Similarly, the relative myocardial signal dropped as the simulated heart rate increased due to reduction of T₁.

Figure 3.

T_2 maps acquired in a phantom with 6 different vials, measured using the spin-echo, conventional bright blood, and BEATS sequences.

Figure 4. Phantom results of a) SNR and b) T₂ accuracy from six vials measured with different simulated heart rates ranging from 60 to 130 bpm.

Figure 5.

Example short axis mid-ventricular images and T₂ maps obtained from conventional bright blood T₂prep-bSSFP and BEATS sequences. Only 3 (out of 9) representative T₂-weighted images (TE = 0, 25, 55 ms) are displayed. All the grayscale images are displayed at the same scale. Note that the BEATS sequence provides effective suppression of the intraventricular blood, and clear depiction of the ventricular walls.

Figure 6. Examples of T_2 mappings obtained from conventional bright-blood T_2 prep-bSSFP and BEATS sequences, and the T_2 profiles (drawn with interpolation). Segment-based T_2 measurements of both cases are also shown in the bullseye. The red lines in the T_2 mappings indicate where the T_2 profiles are generated. T_2 profiles through the T_2 maps indicate substantially steeper myocardial-blood transitions in the BEATS T_2 maps.

Figure 7. Examples of T₂-weighted images (TE = 0, 55, 95 ms) and T₂ maps acquired with BEATS from two patients with heart rates of 60 bpm (upper) and 70 bpm (lower). Both T₂-weighted images and T₂ maps show good image quality.

Figure 8.

Posterior left ventricular wall edema after a myocardial infarct. The patient presented with new onset chest pain and electrocardiographic findings consistent with a posterior myocardial infarction (A, ST-segment depression and inverted T waves in V2–3, arrows). On coronary angiogram, a large obtuse marginal branch was occluded (B, arrow) and flow was re-established with complete filling of the artery (C, arrow). On day 4 post-infarct, comparison of the myocardial T₂ mappings obtained from conventional bright blood T₂prep-bSSFP and BEATS sequences (D) showed left ventricular subendocardial edema (red arrows). Segment-based T₂ measurements were also shown in the bullseye.

A. Healthy Volunteers

B. Patients

Figure 9.

The bullseye images of T_2 (mean), precision, and repeatability among the 8 healthy subjects and 36 patients in a segment-based analysis using both conventional bright-blood T_2 prep-bSSFP and the BEATS sequence. The center T_2 represents segmental mean T_2 in all 6 segments. No statistical significant differences of T_2 measurements were found.

Figure 10.

Example of short-axis mid-ventricular T_2 maps obtained from conventional bright-blood T_2 prep-bSSFP and BEATS sequences in an animal model with prior myocardial infarction. Edema was observed in anterior septum (red arrows) in both maps (bright-blood: 82 ms, BEATS: 81 ms). Signal intensity profiles drawn across the myocardium in edema (yellow lines) in BEATS shows sharper transition between blood and myocardium.