RESEARCH ARTICLE

Diagnostic reliability of low dose multidetector CT and cone beam CT in maxillofacial trauma—an experimental blinded and randomized study

¹Romke Rozema, Michiel HJ Doff, ^{2,3}Peter MA van Ooijen, ⁴Douwe Postmus, ³Henriëtte E Westerlaan, ⁵Martijn F Boomsma and ¹Baucke van Minnen

¹Department of Oral and Maxillofacial Surgery, University Medical Center Groningen, University of Groningen, Groningen, The Netherlands; ²Center for Medical Imaging - North East Netherlands (CMI-NEN²), University Medical Center Groningen, University of Groningen, Groningen, Groningen, The Netherlands; ³Department of Radiology, University Medical Center Groningen, University of Groningen, Groningen, The Netherlands; ⁴Department of Epidemiology, University Medical Center Groningen, University of Groningen, Groningen, The Netherlands; ⁵Department of Radiology, Isala Hospital, Zwolle, The Netherlands

Objectives: To assess the diagnostic reliability of low dose multidetector CT (MDCT) and cone beam CT (CBCT) for zygomaticomaxillary fracture diagnosis.

Methods: Unilateral zygomaticomaxillary fractures were inflicted on four out of six fresh frozen human cadaver head specimens. All specimens were scanned using four MDCT and two CBCT imaging protocols of which the radiation exposure was systematically reduced. A blinded diagnostic routine was simulated at which 16 radiologists and 8 oral and maxillofacial (OMF) surgeons performed randomized image assessments. We considered the findings during an open operative approach of the zygomatic region as the gold standard.

Results: Zygomaticomaxillary fractures were correctly diagnosed in 90.3% ($n = 130$) of the image assessments. The zygomatic arch was most often correctly diagnosed (91.0%). The zygomatic alveolar crest showed the lowest degree of correct diagnosis (65.3%). Dose reduction did not significantly affect the objective visualization of fractures of the zygomaticomaxillary complex. The sensitivity and specificity also remained consistent among the low dose scan protocols. Dose reduction did not decrease the ability to assess dislocation, comminution, orbital volume, volume rendering and soft tissues. OMF surgeons considered the low dose protocols sufficient for treatment planning.

Conclusions: Dose reduction did not decrease the diagnostic reliability of MDCT and CBCT for the diagnosis of zygomaticomaxillary fractures.

Dentomaxillofacial Radiology (2018) **47**, 20170423. doi: [10.1259/dmfr.20170423](https://doi.org/10.1259/dmfr.20170423)

Cite this article as: Rozema R, Doff MHJ, van Ooijen PMA, Postmus D, Westerlaan HE, Boomsma MF, et al. Diagnostic reliability of low dose multidetector CT and cone beam CT in maxillofacial trauma—an experimental blinded and randomized study. *Dentomaxillofac Radiol* 2018; **47**: 20170423.

Keywords: multidetector CT; cone beam CT; low dose; dose reduction; zygomaticomaxillary fractures

Introduction

Multidetector CT (MDCT) and cone beam CT (CBCT) are the imaging modalities of choice in midfacial trauma

diagnostics and treatment planning.^{1–3} Both modalities are utilized to detect and characterize midfacial fractures.^{2,4,5} The zygomaticomaxillary buttress represents the most common site of facial fractures.^{6,7} Disruption

Correspondence to: Mr Romke Rozema, E-mail: r.rozema@gmail.com

Received 14 November 2017; revised 05 May 2018; accepted 08 May 2018

of the zygomatic position can cause an aesthetic and functional deficit. It is therefore mandatory that zygomatic bone injuries are properly diagnosed and adequately managed.⁸

Although CBCT is considered to produce a generally lower radiation dose, there are some clinical considerations for the common use of MDCT to evaluate blunt facial trauma.^{1,9} First, there is limited availability of CBCT within the emergency department, where the majority of the patients are presented initially. Second, additional traumatic injuries may impede examinations using CBCT in which the patient is scanned in a natural head position. Third, only MDCT allow soft tissue reconstructions, which are appreciated for the ability to detect associated soft tissue injury in midfacial trauma.¹⁰

However, there is an increasing concern regarding the potential radiation induced risks due to the exponentially increased use of MDCT and CBCT in medicine.^{1,11} Keeping the radiation dose as low as reasonably achievable, remains the most important strategy for limiting the risks.^{12,13} There is limited evidence regarding the effect of dose reduction for the diagnosis of maxillofacial trauma. In two studies, evidence for the potential of substantial dose reduction for the diagnosis of midface fractures was successfully provided.^{14,15} However, these studies were not randomized and the observers were not blinded for the outcome of interest. The aim of the present study was to assess the diagnostic reliability of low dose MDCT and low dose CBCT for zygomaticomaxillary fractures using a randomized approach with blinded observers.

Methods and materials

Study subjects

Six fresh frozen human cadaver heads were obtained from the section anatomy of the Department of Neurosciences of the University Medical Center Groningen (Groningen, NL). These cadavers have been donated for medical research and educational purposes according to the legal and ethical guidelines of the Dutch uniform anatomical gift act. Their age ranged from 60 to

96-years-old. Thawing the cadaver specimens to room temperature occurred 48 h prior to experimentation. The specimens were isolated at the craniovertebral junction keeping the intracranial contents intact.

Experimental design

An experiment was designed to systematically inflict unilateral zygomaticomaxillary fractures on four out of six specimens. A blunt facial trauma was simulated using a guided free fall method as an impact model (Figure 1).¹⁵⁻¹⁷ The specimens were placed horizontally on a solid wedge for stabilization. The direction of impact was chosen at a 40 degree angle from the midsagittal section plane in order to obtain perpendicular impact on the malar eminence. A cylindrical 2.0 kg impactor with a flat impacting surface was used as a free falling mass. A drop height estimation of 70 cm was calculated using biomechanical fracture tolerance data where power of collision was measured in dynamic loads.¹⁷⁻¹⁹ An independent consultant oral and maxillofacial (OMF) surgeon physically examined the zygomaticomaxillary complex for fracture presence after impact. The drop height was raised with 10 cm consequently if no discernible evidence for a fracture was found.

MDCT imaging

First, the specimens were scanned using four MDCT protocols (Figure 2). The scan field of view (FOV) was set from the maxillary dentition up to the superior orbital rim of the frontal sinus. Datasets were scanned using a third generation Siemens SOMATOM Force scanner (Siemens Healthcare AG, Erlangen, Germany). Automatic exposure control was used for radiation dose optimization adapting tube current (CARE Dose4D, Siemens) and tube voltage (CARE kV, Siemens). For each specimen the tube voltage was fixed for consecutive scans made from the baseline protocol. Quality reference mAs was used as the only variable to systematically reduce the radiation dose for each MDCT scan protocol. Each specimen was scanned using a quality reference mAs of 50, 40, 30 and the lower limit of 20. Scan parameter details for MDCT are found in Table 1.

Figure 1 Impact model used to inflict unilateral zygomaticomaxillary fractures.

Figure 2 A right-sided zygomaticomaxillary fracture in the zygomatic arch, the anterior and posterior sinus maxillary wall. The MD CT quality reference mAs 50 and 20 (A,C) scan protocols and the CBCT baseline and ultra-low dose (B,D) scan protocols are compared consecutively. CBCT, cone beam CT; MDCT, multidetector CT; ULD, ultra-low dose.

The raw data was reconstructed into 0.6 mm bone and soft tissue volume datasets using the Siemens Hr59d and Hr32d kernels respectively.

CBCT imaging

Second, the specimens were scanned with two CBCT protocols using the ProMax 3D Mid (Planmeca Oy, Helsinki, Finland). The scan FOV was standardized as the MDCT settings. One of these datasets was scanned

using the ultra-low dose (ULD) mode. Both isotropic scan protocols were reconstructed into 0.4 mm slice datasets. Scan parameter details for CBCT are found in [Table 2](#).

Additional image processing

In addition, both MDCT and CBCT bone volumetric datasets were reconstructed into a three-dimensional (3D) volume rendering batch using a bone template.

Table 1 Siemens SOMATOM Force MDCT scan parameters

Tube voltage modulation ^a	CARE kV
Tube voltage	80–100 kV
Tube current modulation ^b	CARE Dose4D
Quality reference mAs ^c	50–20
ADMIRE strength	1
FOV	220.0 mm
Collimation	192 × 0.6 mm
Average scan length	118 mm
Slice thickness	0.6 mm
Position increment	0.4 mm
Grayscale depth	12 bit
Pitch	0.6
Rotation time	0.5 s
Exposure time	0.5 s
Scan time	3.4 s
Matrix	512 × 512

MDCT, multidetector CT; ADMIRE, advanced modeled iterative reconstruction; FOV, field of view.

^aFor each specimen the tube voltage was fixed for consecutive scans made from the baseline protocol.

^bThe quality reference mAs was used as the only variable to systemically reduce the radiation dose for each MDCT scan protocol.

^cRadiation exposure outcomes for all four MDCT scan protocols were presented in Table 7 as [supplementary material](#).

This data was presented in a dynamic 360 degree rotational view round a fixed vertical axis providing the spatial relationship of the osseous midfacial anatomy. All data was exported in a Digital Imaging and Communications in Medicine (DICOM) standard.

Gold standard

A gold standard was created in a consensus meeting between a European certified head and neck radiologist (HW) and an OMF surgeon with specific cranio-maxillofacial trauma experience (BM). Both were not included in the image assessment procedure. The presence of fractures of the zygomatic region were recorded by the radiologist who had access to all scanned MDCT and CBCT data. These findings were validated by the OMF surgeon using a combined local transcutaneous and maxillary vestibular approach of the specimen's zygomaticomaxillary complex ([Figure 3](#)). Discrepancies were discussed and advocated.

Assessment setup

Image assessment were performed by a total of 24 observers. The group consisted of 16 radiologists and 8 OMF surgeons with an average of 10.7 (±7.7) and 12.6 (±11.4) years of experience respectively. Each observer performed image assessments for all six patient cases and each observer assessed only one scan protocol for each of these patients. One of the six scan protocols was randomly allocated for each of these patient cases using a research randomizer.²⁰ The order in which the observer assessed the patient cases was also randomized. Using

Figure 3 Approach used as gold standard. For this specimen the 3D volume rendering (A, C) is compared to the surgical approach (B, D) from the same projection. Fractures were recorded in the zygomatic arch, frontozygomatic suture, infra-orbital rim, anterior maxillary sinus wall and lateral orbital wall. 3D volume, three-dimensional volume.

this study design, each patient case was assessed with a unique scan protocol and therefore, observational selection bias was prevented. In each image assessment, the observer was blinded for the outcome of interest and all scan parameters. Each image assessment was simulated as a diagnostic routine where each specimen was presented as a fictive patient case with a suspected zygomaticomaxillary fracture. The use of a medical diagnostic display was set as a condition for radiologist. These displays all met the acceptance and constancy tests following the American Association of Physicists in Medicine (AAPM) Task Group 18 (TG18) guidelines.²¹ All available DICOM datasets were provided on a mobile TeraRecon Aquarius iNtuition DICOM Viewer (TeraRecon, Foster City, California). This viewer application allowed the features and functionalities that a medical professional would use in the daily clinical care. The observer was allowed to adjust window width

Table 2 Planmeca Promax 3D Mid CBCT scan parameters for the two used protocols (baseline/ULD)

Tube voltage	90 kV
Tube current	10/7 mA
DAP	15.6/3.2 mGy
Exposure time	13.6 s/4.5 s
Grayscale depth	12 bit
Scan time	18–26 s
FOV	200 × 100
Radiation source	Pulsed
Rotation degree	Single 200 degree rotation
Projections per rotation	300
Detector type	Flat panel detector
Voxel size (x,y,z)	0.4 mm ³
Matrix	512 × 512

CBCT, cone beam CT; DAP, dose area product; FOV, field of view; ULD, ultra-low dose.

Table 3 Objective and subjective image assessment parameters used to assess the diagnostic reliability for each of the scan protocols

Objective (presence and location) <i>Question; do the following anatomical sites of the midface comprise a fracture?</i>	Subjective (scan protocol sufficiency) <i>Question; is the image quality sufficient to assess the following parameters?</i>
(yes/no/inconclusive)	(sufficient/not sufficient)
Frontozygomatic suture	Fracture dislocation
Lateral orbital wall	Fracture comminution
Zygomatic arch	Orbital volume
Orbital floor	Volume rendering
Zygomatic alveolar crest	Soft tissue ^a
Anterior maxillary sinus wall	Treatment management sufficiency ^b
Posterior maxillary sinus wall	Fracture assessment sufficiency ^c
Zygomatic corpus	

^aSoft tissue reconstructions were only available for multidetector CT.

^bThis parameter was only assessed by oral and maxillofacial surgeons.

^cThis parameter was only assessed by radiologists.

and window level from a 2000/250 bone and 400/40 soft tissue preset. Volume datasets were presented into orthogonal axial, sagittal and coronal plane where free angle adjustment using the multi planar reconstruction module was allowed.

Image assessment

Each patient case was provided with fictive personal particulars, mechanism of injury, clinical presentation and physical examination findings. Each image assessment was divided into an objective and subjective analysis. The objective analysis consisted of a fracture diagnosis including a location determination using a predefined selection of zygomaticomaxillary fracture associated anatomical sites. The gold standard was used to verify the correctness of these observations. The subjective analysis was used to study the sufficiency of the image quality to assess parameters relevant to diagnosing fractures in the maxillofacial region. Also, the acceptability of the image quality for 3D volume rendering and soft tissue reconstructions was assessed. Within the subjective image assessment, the sufficiency for fracture assessment is only assessed by radiologists and the sufficiency for treatment management is only assessed by OMF surgeons. All parameters of the image assessment are summarized in Table 3 in detail.

Effective dose estimations

Effective dose estimations were only accessible for MDCT using volume-weighted CT dose indices ($CTDI_{vol}$) and Dose Length Products (DLP) provided by the scanner. Head conversion coefficients ($0.0021 \text{ mSv mGy}^{-1} \text{ cm}^{-1}$) quoted by the AAPM were used to convert DLP to effective dose.²²

Statistics

Data analysis was performed using the R environment for statistical computing and graphics (v. 3.3.3).²³ A generalized linear mixed model was constructed to assess the effect of dose reduced scan protocols for the objective fracture detectability outcome. The analysis was performed separately for MDCT and CBCT and the baseline protocols were used as intercept. The diagnostic accuracy of the eight fractures sites, as verified by the gold standard, was used as response variable (Table 3). The variable was defined in binary and inconclusive reported data was statistically handled as an incorrect diagnostic finding as it would otherwise result in a distorted skew towards a diagnostic reliable outcome. The scan protocol covariate was included as a fixed effect and the observer, patient case and fracture location covariates nested within the patient case were included as random effects to account for the correlation among the repeated measurements. The significance level was set at 5%. As an additional outcome for the overall diagnostic reliability, the sensitivity and specificity were calculated for each scan protocol by pooling the verified assessment outcomes for all the anatomical sites of the zygomaticomaxillary complex. The percentage of objective fracture detectability was compared between radiologists and OMF surgeons.

Results

Zygomaticomaxillary fractures were successfully inflicted on all four specimens in varying degree of severity. The consensus meeting corresponded that the baseline protocols of MDCT and CBCT appeared sufficient to visualize all fractures of the anatomical sites selection of the midface. Over all the image assessments, 90.3% ($n = 130$) of the zygomaticomaxillary fractures were correctly diagnosed and 7.6% ($n = 11$) were misdiagnosed. In another 2.1% ($n = 3$) the diagnosis was reported as inconclusive. The four specimens that sustained a fracture were diagnosed correctly by all 24 observers (100%). From the two specimens without a fracture, one was diagnosed correctly by 23 observers (95.8%) whilst the other was diagnosed correctly by 11 observers (45.8%). The objective detectability percentages varied substantially among the assessed anatomical sites (Table 4). Zygomatic arch fractures were most often diagnosed correctly (91.0%). On average, the highest degree of misdiagnosis was found for zygomatic alveolar crest fractures. For patient case 4 there was an outlying degree of misdiagnosis for the anterior and posterior maxillary sinus wall and the orbital floor. In the consensus meeting, these fractures were found to be minimal and non-dislocated. For patient case 6, there was a high degree of zygomaticomaxillary fracture misdiagnosis when compared to the other patient cases. For the subjective analysis, there was a high degree of conformity that the image quality was sufficient to assess the fracture dislocation (95.5%), comminution

Table 4 Outcome of correctly diagnosed fracture sites subdivided for each patient case

	Patient case ^a										Mean (%)
	1	2	3	4	5	6					
Fracture presence ^b	24/24 (100) Y	24/24 (100) Y	24/24 (100) Y	24/24 (100) Y	23/24 (95.8) N	11/24 (45.8) N					90.3
Frontozygomatic suture	21/24 (87.5) Y	23/24 (95.8) Y	12/24 (50.0) N	11/24 (45.8) Y	23/24 (95.8) N	19/24 (79.2) N					75.7
Lateral orbital wall	23/24 (95.8) Y	21/24 (87.5) Y	22/24 (91.7) Y	20/24 (83.3) Y	23/24 (95.8) N	20/24 (83.3) N					89.6
Zygomatic arch	24/24 (100) Y	24/24 (100) Y	24/24 (100) Y	24/24 (100) Y	24/24 (100) N	11/24 (45.8) N					91.0
Orbital floor	19/24 (79.2) Y	24/24 (100) Y	17/24 (70.8) Y	2/24 (8.3) Y	24/24 (100) N	22/24 (91.7) N					75.0
Zygomatic alveolar crest	8/24 (33.3) N	8/24 (33.3) N	13/24 (54.2) Y	21/24 (87.5) N	24/24 (100) N	20/24 (83.3) N					65.3
Anterior maxillary sinus wall	24/24 (100) Y	24/24 (100) Y	24/24 (100) Y	4/24 (16.7) Y	24/24 (100) N	18/24 (75.0) N					81.9
Posterior maxillary sinus wall	21/24 (87.5) Y	20/24 (83.3) Y	23/24 (95.8) Y	2/24 (8.3) Y	24/24 (100) N	17/24 (70.8) N					74.3
Zygomatic corpus	15/24 (62.5) N	15/24 (62.5) N	15/24 (62.5) N	19/24 (79.2) N	24/24 (100) N	22/24 (91.7) N					76.4

^aReported as correctly diagnosed proportions and percentages. For each patient case, the presence of a fracture was verified by the gold standard and provided for each assessed anatomical site of the zygomaticomaxillary complex (Y/N).

^bZygomaticomaxillary fractures were inflicted on the first four patient cases 1–4.

(89.2%) and the orbital volume (95.5%). This was also the case for the 3D volume rendering batch (85.6%) and the soft tissue reconstructions for MDCT (91.8%). The agreement remained consistent among the scan protocols (Table 5).

Effect of dose reduction

Both the outcomes of the objective detectability of the eight anatomical fracture sites and the subjective assessed parameters remained consistent among the six scan protocols (Table 5). The GLMM analysis showed

Table 5 Objective and subjective image assessment proportion outcomes for the four MDCT and two CBCT protocols

	Quality reference mAs MDCT protocol				CBCT protocol	
	50	40	30	20	Baseline	ULD
Effective dose (μSv) ^a	129.9 ± 9.2	99.9 ± 12.0	78.3 ± 6.3	51.0 ± 4.2		
Objective analysis^b						
Fracture presence	21/24 (87.5)	23/24 (95.8)	21/24 (87.5)	22/24 (91.7)	21/24 (87.5)	22/24 (91.7)
Frontozygomatic suture	18/24 (75.0)	18/24 (75.0)	17/24 (70.8)	19/24 (79.2)	18/24 (75.0)	19/24 (79.2)
Lateral orbital wall	21/24 (87.5)	21/24 (87.5)	24/24 (100)	20/24 (83.3)	22/24 (91.7)	21/24 (87.5)
Zygomatic arch	23/24 (95.8)	22/24 (91.7)	21/24 (87.5)	22/24 (91.7)	21/24 (87.5)	22/24 (91.7)
Orbital floor	17/24 (70.8)	18/24 (75.0)	18/24 (75.0)	17/24 (70.8)	19/24 (79.2)	19/24 (79.2)
Zygomatic alveolar crest	15/24 (62.5)	15/24 (62.5)	16/24 (66.7)	18/24 (75.0)	15/24 (62.5)	15/24 (62.5)
Anterior maxillary sinus wall	19/24 (79.2)	19/24 (79.2)	19/24 (79.2)	19/24 (79.2)	21/24 (87.5)	21/24 (87.5)
Posterior maxillary sinus wall	17/24 (70.8)	19/24 (79.2)	16/24 (66.7)	19/24 (79.2)	17/24 (70.8)	19/24 (79.2)
Zygomatic corpus	18/24 (75.0)	20/24 (83.3)	19/24 (79.2)	17/24 (70.8)	15/24 (62.5)	21/24 (87.5)
Subjective analysis^c						
Dislocation	17/18 (94.4)	18/18 (100)	18/19 (94.7)	17/18 (94.4)	18/19 (94.7)	18/19 (94.7)
Comminution	15/18 (83.3)	18/18 (100)	14/19 (73.7)	17/18 (94.4)	17/19 (89.5)	18/19 (94.7)
Orbital volume	16/18 (88.9)	18/18 (100)	17/19 (89.5)	18/18 (100)	18/19 (94.7)	19/19 (100)
Volume rendering	18/18 (100)	17/18 (94.4)	15/19 (78.9)	17/18 (94.4)	13/19 (68.4)	15/19 (78.9)
Soft tissue	17/18 (94.4)	17/18 (94.4)	16/19 (84.2)	17/17 (94.4)	-	-
Fracture assessment	13/16 (81.3)	15/16 (93.8)	14/16 (87.5)	16/16 (100)	13/16 (81.3)	9/16 (56.3)
Treatment management	7/8 (87.5)	8/8 (100)	8/8 (100)	7/8 (87.5)	8/8 (100)	8/8 (100)

CBCT, cone beam CT; MDCT, multidetector CT; ULD, ultra-low dose.

^aOnly provided for MDCT. Effective dose estimations (Mean ± SD μSv) were calculated by converting Dose Length Products (DLP) using head conversion coefficients (0.0021 mSv mGy⁻¹ cm⁻¹) quoted by the American Association of Physicists in Medicine (AAPM).²²

^bReported as correctly diagnosed proportions and percentages as verified by the gold standard.

^cReported as the acceptable reported proportions and percentages. It was only reported when the observer suspected the presence of a zygomaticomaxillary fracture. Soft tissue reconstructions were not available for CBCT. Treatment management was only assessed by oral and maxillofacial surgeons (*n* = 8) whilst the fracture assessment was only assessed by radiologists (*n* = 16).

Table 6 Generalized linear mixed model outcome and sensitivity and specificity for each scan protocol

	OR	CI	p-value ^a	Sensitivity (%) ^b	Specificity (%) ^b
MDCT protocol					
50 (intercept)				78	77
40	0.91	0.47–1.76	0.78	78	81
30	0.75	0.37–1.53	0.44	78	83
20	1.10	0.57–2.14	0.78	76	81
CBCT protocol					
Baseline (intercept)				80	79
ULD	1.43	0.76–2.70	0.27	82	85

CBCT, cone beam CT; CI, confidence interval; MDCT, multidetector CT; OR, odds ratio; ULD, ultra-low dose.

^aGeneralized linear mixed model were used to assess the effect of scan protocols for the outcome of the assessment.

^bThe sensitivity and specificity were calculated pooling the verified outcome for all assessed anatomical sites of the zygomaticomaxillary complex.

that the objective fracture detectability accuracy did not differ significantly between the different MDCT and CBCT scan protocols (Table 6). The sensitivity and specificity also remained consistent among the scan protocols. The effective dose for MDCT examinations (Mean $\pm$ SD) was reduced by 60.7% from 129.9 ± 9.2 to 51.0 ± 4.2 μ Sv (Table 5 and supplementary material).

Profession related outcome

The objective fracture detectability of radiologists (79.6%) remained well in range of OMF surgeons (76.6%). For radiologists, there was a high agreement for the overall sufficiency of the image quality to assess fractures for the MDCT protocols and the baseline CBCT protocol. The CBCT ULD protocol was found sufficient in 56.3%. For OMF surgeons, the treatment management sufficiency remained consistently high for all scan protocols (Table 5).

Discussion

In this study, we assessed the diagnostic reliability of low dose MDCT and CBCT for the diagnosis of zygomaticomaxillary complex fractures in a blinded and randomized study design. We showed that dose reduction did not significantly affect the objective visualization of fractures of the zygomaticomaxillary complex. Above that, the sensitivity and specificity remained consistent among the low dose scan protocols. Similar results were found in a previous study in which low dose protocols were considered sufficient for the diagnosis of both dislocated and non-dislocated midface fractures.¹⁴

Despite these findings, detectability percentages varied substantially among the assessed anatomical sites (Table 4). Potential explanations are discussed. First, patient case 6 was falsely diagnosed with a zygomatic

arch fracture whilst no trauma was inflicted on this specimen. A consolidated fracture was suspected but no medical records were available for verification. Second, the highest degree of misdiagnosis was found for zygomatic alveolar crest fractures. The zygomatic alveolar crest is located in the infrazygomatic region. It is clinically appreciated because the cortical bone thickness provides anchorage for the osteosynthesis screws during surgery.²⁴ We believe that the determination of a fracture in this region was doubtful for the observer, as the zygomatic alveolar crest is unclearly demarcated in anatomical terms. Third, the low detection rates for the orbital floor, the anterior and posterior maxillary sinus wall for patient case 4 could be explained by their minimal and non-dislocated characteristics. Although not evaluated in the present study, the clinical consequences of missing non-dislocated fractures could be argued. Nevertheless, the detection variability among the anatomical sites does emphasize that careful reading and adequate anatomical proficiency is a requisite to assess datasets of a suspected midface fracture sufficiently.

Both MDCT and CBCT dose reduced protocols were found sufficient to assess the fracture dislocation, the extend of comminution and the orbital volume. Also, no decrease was found for the diagnostic quality sufficiency of the volume rendering and soft tissue reconstructions. The use of 3D volume rendering provides a visualization of the spatial relationships of the midface buttresses and is available for both MDCT and CBCT.²⁵ For MDCT, the soft tissue reconstructions allow evaluation of soft tissue injury which, in case of midfacial fractures, is predominantly utilized to assess the extra-ocular muscles.^{10,26,27}

For OMF surgeons, dose reduction did not affect the capability for treatment management. For radiologists, the overall sufficiency to assess the fracture did not decrease for MDCT protocols. Although the objective fracture detectability did not decrease significantly, the ULD protocol of CBCT was only found sufficient subjectively in 56.3% of the assessments. We believe that this is first, due to more evident artifacts in this protocol and second, due to the fact that the application and interpretation of CBCT is dedicated to OMF surgeons. From a clinical perspective, there is a difference in image interpretation between radiologists and OMF surgeons. We think that OMF surgeons predominantly focus on determination of the treatment pathway and planning of potential surgery. This is in contrast to radiologists who are responsible for a correct diagnosis and therefore assess the image quality from a different approach. This allocation of roles should be taken into account when MDCT and CBCT are used for maxillofacial fracture diagnosis.

In this study, effective dose estimations were only provided for MDCT. For CBCT, a previous study used thermoluminescent dosimeters in anthropomorphic Alderson radiation therapy phantoms to estimate effective radiation dose and reported 122 μ Sv and

28 µSv for baseline and low dose medium FOV dento-alveolar protocols using the same CBCT device.²⁸ When compared, the effective dose of MDCT remained well in range of CBCT. Moreover, the effective dose of the MDCT low dose protocols remained below the baseline CBCT protocol. Another study could also successfully confirm dose reduction of MDCT below the levels of CBCT.²⁹ This data is in contrast to the common belief that MDCT delivers a higher radiation exposure compared to CBCT.^{1,5,9,30} The delivered radiation dose is the result of the interplay between a variety of elements. Above all, one should be aware of the large exposure ranges among MDCT and CBCT devices.^{9,31,32} For CBCT, a distinction is needed between small-, medium- and large FOVs.²⁸ Optimization should be performed using an appropriate scan range determination, tube potential and reconstruction algorithm. For MDCT, some studies propose the use of a full head protocol.^{9,14} However, maxillofacial trauma protocols should predominantly be used to visualize the osseous anatomy, requiring lower levels of radiation dose. Also, the use of automatic exposure control improves consistency of the image quality and curtails the radiation dose for each individual patient.¹² In this study this was accomplished for both the tube current (CARE Dose4D) and tube voltage (CARE kV). Furthermore, filtered back projection is currently the well-established reconstruction type for clinical MDCT protocols. The recent availability of iterative reconstruction (IR) algorithms offers new possibilities for radiation dose reduction.^{14,33} In this study, filtered back projection was combined with a model-based IR (ADMIRE 1, Siemens). As there is limited evidence regarding the use of IR algorithms for maxillofacial trauma, further research should focus on the potential of dose reduction, the clinical applicability and computational time requirements.

This study has some limitations. First, the post-mortem status of human cadaver specimen carries process changes that may affect the interpretability of the scans. Second, trauma patients may have indirect predictive CT findings, such as sinus opacification and

periorbital contusion, that have discriminatory benefits for predicting midface fractures in trauma patients.³⁴ These findings are absent in human cadavers. Nevertheless, we believe that, as a surrogate model, it is the most accurate approach of a maxillofacial trauma patient.

Another limitation is that we were only able to use six human cadavers as a representation of a midface trauma patient population. Therefore, the statistical power is limited and there was a lack of variability of fracture patterns. This study focuses on the diagnosis of zygomaticomaxillary fractures only, whilst clinicians are faced with a broad range of fracture patterns such as Le Fort or naso-orbitoethmoid fractures that were not assessed with the present study design. However, we believe that the experimental approach of this study provided substantial evidence to further study the clinical applicability of low dose maxillofacial trauma protocols in human subjects.

In conclusion, the results of this human cadaver study show that dose reduction did not decrease the diagnostic reliability of MDCT and CBCT for the diagnosis of zygomaticomaxillary fractures. Low dose protocols were also considered sufficient for treatment management according to OMF surgeons. Further research should focus on the potential of IR algorithms and the applicability of low dose MDCT and CBCT protocols in a clinical environment.

Acknowledgement

We would like to thank Mr *R Dob* (BSc, Department of Radiology) and Mr *JA Dieters* (BSc, Department of Orthodontics) from the University Medical Center Groningen for their advice and technical assistance during the use of the MDCT and CBCT devices respectively. Also, we would like to thank the section anatomy from the Department of Neurosciences of the University Medical Center Groningen, Groningen, Netherlands for the provision of human cadaver specimen and assistance during the experiments.

References

1. De Vos W, Casselman J, Swennen GR. Cone-beam computerized tomography (CBCT) imaging of the oral and maxillofacial region: a systematic review of the literature. *Int J Oral Maxillofac Surg* 2009; **38**: 609–25. doi: <https://doi.org/10.1016/j.ijom.2009.02.028>
2. Sohns JM, Staab W, Sohns C, Schwarz A, Streit U, Hosseini AS, et al. Current perspective of multidetector computed tomography (MDCT) in patients after midface and craniofacial trauma. *Clin Imaging* 2013; **37**: 728–33. doi: <https://doi.org/10.1016/j.clinimag.2012.11.002>
3. Miracle AC, Mukherji SK. Conebeam CT of the head and neck, part 2: clinical applications. *AJNR Am J Neuroradiol* 2009; **30**: 1285–92. doi: <https://doi.org/10.3174/ajnr.A1654>
4. Choudhary AB, Motwani MB, Degwekar SS, Bhowate RR, Banode PJ, Yadav AO, et al. Utility of digital volume tomography in maxillofacial trauma. *J Oral Maxillofac Surg* 2011; **69**: e135–e140. doi: <https://doi.org/10.1016/j.joms.2010.07.081>
5. Shintaku WH, Venturin JS, Azevedo B, Noujeim M. Applications of cone-beam computed tomography in fractures of the maxillofacial complex. *Dent Traumatol* 2009; **25**: 358–66. doi: <https://doi.org/10.1111/j.1600-9657.2009.00795.x>
6. Gassner R, Tuli T, Hächl O, Rudisch A, Ulmer H. Cranio-maxillofacial trauma: a 10 year review of 9,543 cases with 21,067 injuries. *J Craniomaxillofac Surg* 2003; **31**: 51–61. doi: [https://doi.org/10.1016/S1010-5182\(02\)00168-3](https://doi.org/10.1016/S1010-5182(02)00168-3)
7. Kraft A, Abermann E, Stigler R, Zsifkovits C, Pedross F, Kloss F, et al. Craniomaxillofacial trauma: synopsis of 14,654 cases with 35,129 injuries in 15 years. *Craniomaxillofac Trauma Reconstr* 2012; **5**: 041–50. doi: <https://doi.org/10.1055/s-0031-1293520>
8. Rana M, Warraich R, Tahir S, Iqbal A, von See C, Eckardt AM, et al. Surgical treatment of zygomatic bone fracture using two

- points fixation versus three point fixation--a randomised prospective clinical trial. *Trials* 2012; **13**: 36. doi: <https://doi.org/10.1186/1745-6215-13-36>
9. Loubele M, Bogaerts R, Van Dijk E, Pauwels R, Vanheusden S, Suetens P, et al. Comparison between effective radiation dose of CBCT and MSCT scanners for dentomaxillofacial applications. *Eur J Radiol* 2009; **71**: 461–8. doi: <https://doi.org/10.1016/j.ejrad.2008.06.002>
10. Winegar BA, Murillo H, Tantiwongkosi B. Spectrum of critical imaging findings in complex facial skeletal trauma. *Radiographics* 2013; **33**: 3–19. doi: <https://doi.org/10.1148/rg.331125080>
11. Brenner DJ, Hall EJ. Computed tomography--an increasing source of radiation exposure. *N Engl J Med* 2007; **357**: 2277–84. doi: <https://doi.org/10.1056/NEJMr072149>
12. Yu L, Liu X, Leng S, Kofler JM, Ramirez-Giraldo JC, Qu M, et al. Radiation dose reduction in computed tomography: techniques and future perspective. *Imaging Med* 2009; **1**: 65–84. doi: <https://doi.org/10.1056/NEJMr072149>
13. McCollough CH, Bruesewitz MR, Kofler JM. CT dose reduction and dose management tools: overview of available options. *Radiographics* 2006; **26**: 503–12. doi: <https://doi.org/10.1148/rg.262055138>
14. Widmann G, Dalla Torre D, Hoermann R, Schullian P, Gassner EM, Bale R, et al. Ultralow-dose computed tomography imaging for surgery of midfacial and orbital fractures using ASiR and MBiR. *Int J Oral Maxillofac Surg* 2015; **44**: 441–6. doi: <https://doi.org/10.1016/j.ijom.2015.01.011>
15. Meijer A, Rozema R, Hartman R, van der Duim S, van Minnen B, Krijnen WP, et al. Substantial CT radiation dose reduction does not affect the preference for CT over direct digital radiography to diagnose isolated zygomatic fractures – a study in human cadavers. *Radiography* 2016; **22**: e228–e232. doi: <https://doi.org/10.1016/j.radi.2016.07.007>
16. Verschueren P, Delye H, Depreitere B, Van Lierde C, Haex B, Berckmans D, et al. A new test set-up for skull fracture characterisation. *J Biomech* 2007; **40**: 3389–96. doi: <https://doi.org/10.1016/j.jbiomech.2007.05.018>
17. Hampson D. Facial injury: a review of biomechanical studies and test procedures for facial injury assessment. *J Biomech* 1995; **28**: 1–7. doi: [https://doi.org/10.1016/0021-9290\(95\)80001-8](https://doi.org/10.1016/0021-9290(95)80001-8)
18. Yoganandan N, Pintar FA, Sances A, Walsh PR, Ewing CL, Thomas DJ, et al. Biomechanics of skull fracture. *J Neurotrauma* 1995; **12**: 659–68. doi: <https://doi.org/10.1089/neu.1995.12.659>
19. Yoganandan N, Sances A, Pintar FA, Maiman DJ, Hemmy DC, Larson SJ, et al. Traumatic facial injuries with steering wheel loading. *J Trauma* 1991; **31**: 699–710. doi: <https://doi.org/10.1097/00005373-199105000-00015>
20. Urbaniak G, Plous S. Research Randomizer (Version 4.0). 2015. Available from: <https://www.randomizer.org> [accessed October 23, 2015].
21. Samei E, Badano A, Chakraborty D, Compton K, Cornelius C. Assessment of display performance for medical imaging systems. *Med Phys* 2005; **32**: 4095–155.
22. McCollough C, Cody D, Edyvean S, Geise R. The measurement, reporting, and management of radiation dose in CT. *Report of AAPM Task Group* 2008; 1–34.
23. R Development Core Team. R: A language and environment for statistical computing. R Foundation for Statistical Computing, Vienna, Austria. 2008. Available from: <http://www.R-project.org>.
24. Liou EJ, Chen PH, Wang YC, Lin JC, . A computed tomographic image study on the thickness of the infrazygomatic crest of the maxilla and its clinical implications for miniscrew insertion. *Am J Orthod Dentofacial Orthop* 2007; **131**: 352–6. doi: <https://doi.org/10.1016/j.ajodo.2005.04.044>
25. Saigal K, Winokur RS, Finden S, Taub D, Pribitkin E. Use of three-dimensional computerized tomography reconstruction in complex facial trauma. *Facial Plast Surg* 2005; **21**: 214–20. doi: <https://doi.org/10.1055/s-2005-922862>
26. Guly CM, Guly HR, Bouamra O, Gray RH, Lecky FE. Ocular injuries in patients with major trauma. *Emerg Med J* 2006; **23**: 915–7. doi: <https://doi.org/10.1136/emj.2006.038562>
27. Lee EI, Mohan K, Koshy JC, Hollier LH. Optimizing the surgical management of zygomaticomaxillary complex fractures. *Semin Plast Surg* 2010; **24**: 389–97. doi: <https://doi.org/10.1055/s-0030-1269768>
28. Pauwels R, Beinsberger J, Collaert B, Theodorakou C, Rogers J, Walker A, et al. Effective dose range for dental cone beam computed tomography scanners. *Eur J Radiol* 2012; **81**: 267–71. doi: <https://doi.org/10.1016/j.ejrad.2010.11.028>
29. Widmann G, Fasser M, Schullian P, Zangerl A, Puelacher W, Kral F, et al. Substantial dose reduction in modern multi-slice spiral computed tomography (MSCT)-guided craniofacial and skull base surgery. *Rofo* 2012; **184**: 136–42. doi: <https://doi.org/10.1055/s-0031-1281971>
30. Brisco J, Fuller K, Lee N, Andrew D. Cone beam computed tomography for imaging orbital trauma-image quality and radiation dose compared with conventional multislice computed tomography. *Br J Oral Maxillofac Surg* 2014; **52**: 76–80. doi: <https://doi.org/10.1016/j.bjoms.2013.09.011>
31. Ludlow JB, Timothy R, Walker C, Hunter R, Benavides E, Samuelson DB, et al. Effective dose of dental CBCT-a meta analysis of published data and additional data for nine CBCT units. *Dentomaxillofac Radiol* 2015; **44**: 20140197. doi: <https://doi.org/10.1259/dmfr.20140197>
32. Ludlow JB, Ivanovic M. Comparative dosimetry of dental CBCT devices and 64-slice CT for oral and maxillofacial radiology. *Oral Surg Oral Med Oral Pathol Oral Radiol Endod* 2008; **106**: 106–14. doi: <https://doi.org/10.1016/j.tripleo.2008.03.018>
33. Solomon J, Mileto A, Ramirez Giraldo JC, Samei E. Diagnostic performance of an advanced modeled iterative reconstruction algorithm for multidetector CT scanner: potential for radiation dose reduction in a multireader study 1. *Radiology* 2015; **275**: 735–45.
34. Smith HL, Chrischilles E, Janus TJ, Sidwell RA, Ramirez M, Peek-Asa C, et al. Clinical indicators of midface fracture in patients with trauma. *Dent Traumatol* 2013; **29**: 313–8. doi: <https://doi.org/10.1111/edt.12006>