Does oxidized zirconium make a difference? Midterm cohort survivorship of symmetric posterior condyle posterior-stabilized total knee arthroplasty

A. Ross Demcoe, MD
Eric R. Bohm, MD, MSc
David R. Hedden, MD
Colin D. Burnell, MD
Thomas R. Turgeon, MD, MPH

Accepted July 10, 2018

Correspondence to:

T. Turgeon
310-1155 Concordia Ave
Winnipeg MB R2K 2M9
tturgeon@cjrj.ca

DOI: 10.1503/cjs.007518

Background: A novel symmetric posterior condylar design and option of a femoral component with an outer zirconium oxide layer were introduced to a successful asymmetric condylar total knee arthroplasty system in 2005. Given the paucity of information on this modified design, we performed a study to determine its early to midterm survivorship and performance.

Methods: Patients who received the Genesis II Symmetric Posterior Condyle or Legion Primary total knee (cobalt–chromium [CoCr] or oxidized zirconium–niobium [OxZr]) (Smith & Nephew) implanted at the study centre between March 2007 and December 2013 were enrolled into a prospective database. We retrospectively reviewed the database and performed survival analysis using Kaplan–Meier techniques.

Results: There were 2178 patients (1359 women [62.4%]; mean age 64.6 yr; mean body mass index 35.0) with 2815 knee replacements available for analysis. Survival rates were 98.2% (95% confidence interval [CI] 97.6%–98.7%) for failure for any reason at 2 years and 96.8% (95% CI 96.0%–97.7%) at 5 years. Age (hazard ratio [HR] 0.97, 95% CI 0.94–0.997) and female sex (HR 0.45, 95% CI 0.27–0.75) were protective, whereas body mass index (HR 1.02, 95% CI 0.99–1.05) and OxZr implant (HR 1.11, 95% CI 0.57–2.18) did not influence survivorship. Oxford Knee Score values improved from a mean of 40.2 to 21.8 by 2 years ($p < 0.001$), with no difference between the CoCr and OxZr groups.

Conclusion: The symmetric posterior condylar posterior-stabilized knee offers excellent midterm survivorship. Implant bearing surface did not have an influence on survivorship to 5 years, and, thus, use of OxZr implants may not be justified.

Contexte : Un nouveau modèle à condyles postérieurs symétriques et l'option d'un élément fémoral pourvu d'une couche d'oxyde de zirconium ont été introduits avec succès dans un système de prothèse totale du genou à condyles asymétriques en 2005. Compte tenu du manque de renseignements concernant ce modèle modifié, nous avons procédé à une étude pour en déterminer la survie et le rendement à court et à moyen terme.

Méthodes : Les patients ayant reçu la prothèse à condyles postérieurs symétriques Genesis II ou une prothèse totale primaire Legion (au cobalt–chrome [CoCr] ou au zirconium–niobium oxydé [OxZr]) (Smith & Nephew) au centre de recherche entre mars 2007 et décembre 2013 ont été inscrits dans une base de données prospective. Nous avons procédé à une revue rétrospective de la base de données et effectué une analyse de survie à l'aide de techniques de Kaplan–Meier.

Résultats : On a dénombré 2178 patients (dont 1359 femmes [62,4 %]; âge moyen 64,6 ans; indice de masse corporelle moyen 35,0) totalisant 2815 arthroplasties du genou pour analyse. Les taux de survie ont été de 98,2 % (intervalle de confiance [IC] à 95 % 97,6 %–98,7 %) pour le paramètre défaillance de toutes causes à 2 ans et de 96,8 % (IC à 95 % 96,0 %–97,7 %) à 5 ans. L'âge (risque relatif [RR] 0,97, IC à 95 % 0,94–0,997) et le fait d'être de sexe féminin (RR 0,45, IC à 95 % 0,27–0,75) ont conféré un effet protecteur, tandis que l'indice de masse corporelle (RR 1,02, IC à 95 % 0,99–1,05) et l'implant OxZr (RR 1,11, IC à 95 % 0,57–2,18) n'ont eu aucune influence sur la survie. Les scores Oxford d'évaluation du genou se sont améliorés, passant d'une moyenne de 40,2 à 21,8 en l'espace de 2 ans ($p < 0,001$), sans différence entre les groupes CoCr et OxZr.

Conclusion : La prothèse du genou à condyles postérieurs symétriques postéro-stabilisée possède une excellente survie à moyen terme. La surface portant l'implant n'a pas eu d'influence sur la survie à 5 ans et, donc, l'utilisation d'implants OxZr ne sera peut-être pas justifiée.

The Genesis II (Smith & Nephew) total knee implant system was released to the general market in 1996 and since its release has experienced good overall performance. Case series from the design surgeons^{1,2} as well as registry data^{3,4} have shown excellent survivorship and functional results in patients undergoing primary total knee replacement with this system.⁵ The original system uses built-in external rotation of the femoral component in an attempt to optimize patellofemoral tracking, minimize the risk of femoral notching and prevent tibiofemoral shear forces in extension and flexion.¹ Design variations were introduced into the system in 2005 to accommodate a condylar-constrained polyethylene insert requiring a symmetric posterior condyle design. This variant was initially branded as the Symmetric Posterior Condyle and later as the Legion Primary. In an effort to reduce wear, the components were made available in both standard cobalt–chromium (CoCr) bearing as well as an oxidized zirconium–niobium (OxZr) gradient ceramic bearing surface. The OxZr implants are predominantly zirconium–niobium metal with a 4- to 6-mm gradient layer of zirconium oxide on the articulating surface.⁶ Proposed benefits of the more expensive oxidized zirconium surface are increased wettability,⁷ decreased peak roughness,⁸ decreased adhesive and abrasive wear,⁷ and its applicability in patients with known metal sensitivities given the absence of nickel in the base alloy.^{9–11} The tibial component of this device is a titanium–aluminum–vanadium alloy that does not have the issue of nickel as seen in CoCr alloy femoral components.

There have been reports of seemingly minor changes in implant design having effects on overall survivorship and revision rates in both hip and knee arthroplasty.^{12,13} These occurrences show the need to critically evaluate the effect of minor design changes on implant function and survivorship.

Given the paucity of information regarding the survivorship of this iteration of these modifications to a successful design, we performed a study to determine the early to midterm survivorship of the Symmetric Posterior Condyle total knee arthroplasty components at a single institution. Additional goals were to assess for differences in revision risk between the OxZr and CoCr implants and to evaluate the impact of age, sex and body mass index on revision risk and assess the functional scores in patients in whom the prosthesis has been used.

METHODS

The institution's ethics review board approved the study. All patients received the Genesis II Symmetric Posterior

Condyle or Legion Primary total knee implanted at the study centre between March 2007 and December 2013. All implants in the study were cemented and used a posterior stabilized design. Selection of OxZr versus CoCr was left to surgeon preference in discussion with the patient; however, selection of OxZr was restricted to patients aged less than 60 years with expected higher activity levels and a history of metal sensitivity, based on regional guidelines. Preoperatively, patients provided informed consent and were enrolled into a prospective research database. The database documents patient demographic characteristics, diagnosis, procedure details and outcome scores.

All procedures performed by the 4 participating fellowship-trained arthroplasty surgeons were eligible for inclusion in the study. Patients were followed at 6 weeks, 6 months, 1 year, 2 years and every 2 years thereafter for routine monitoring and surveillance. Follow-up data were captured from March 2007 until August 2014. Age at time of surgery, height and weight were documented, and body mass index was calculated. The bearing surface of the femoral component was also documented. In addition, we analyzed the preoperative, 1-year postoperative and 2-year postoperative Oxford Knee Score values (best possible score 12, worst possible score 60). This instrument has proven to be reliable, valid and responsive to clinical change.^{14–16}

Statistical analysis

We performed survival analysis using Kaplan–Meier techniques for the entire population and then separately for the CoCr and OxZr implants. We compared survivorship of the CoCr and OxZr version using the Cox proportional hazards method to control for the possible confounding effects of age,^{17,18} sex^{17–19} and body mass index^{19,20} on revision risk. We compared demographic characteristics and functional score using *t* tests for continuous data and the χ^2 test for dichotomous data.

RESULTS

A total of 2178 patients with 2815 primary total knee replacements were identified in the database. The demographic data can be found in Table 1. The mean age of the patients at the time of surgery was 64.7 (standard deviation [SD] 10.7) years, and the patients were predominantly female (1359 [62.4%]). The average body mass index of all eligible patients from the database was 35.0 (SD 8.5).

There were 2286 CoCr components and 529 OxZr components implanted. Patients who received the OxZr

implants were younger than those who received the CoCr implants and were predominantly male (Table 2). There was no difference in body mass index between the 2 implant groups.

Survivorship

Kaplan–Meier analysis of the pooled implants showed a 2-year survival rate of 98.2% (95% CI 97.6%–98.7%) for failure for any reason and a 5-year survival rate of 96.8% (95% CI 96.0%–97.7%) (Fig. 1). There were 63 cases that required revision, with infection being the most common reason (Table 3). Cox proportional hazards analysis showed that both female sex and increased age lowered revision risk and that there was a trend toward increased revision risk with increased body mass index, but this did not reach statistical significance (Table 4). The individual CoCr and OxZr Kaplan–Meier survival curves are presented in Figure 2. After adjustment for age, sex and body mass index, we could not detect any difference in survival between the 2 materials.

Outcome scores

There were 1851 total knee replacements with preoperative Oxford Knee Score values available for analysis. The mean preoperative score was 40.2 (SD 8.3). The mean score improved to 22.6 (SD 8.6) at 1 year postoperatively and 21.8 (SD 8.9) at 2 years. The improvement in the score at 1 year was statistically significant (95% CI 17.06–18.09), as was the improvement at 2 years (95% CI 17.06–

Fig. 1. Pooled Kaplan–Meier product-limit survival estimate of the Genesis II Symmetric Posterior Condyle/Legion Primary total knee implant (Smith & Nephew). Grey shading represents 95% confidence intervals.

18.09). No differences were found between patients who received OxZr implants and those who received CoCr implants.

DISCUSSION

We found that the overall (combined CoCr and OxZr) revision rate of the Legion Primary knee system was 1.8% at 2 years and 3.2% at 5 years. The Australian National Joint Replacement Registry shows a cumulative revision rate at 3 years of 1.8% for the posterior-stabilized CoCr bearing models and 3.2% for the posterior-stabilized OxZr models.³ Although this may reflect a difference in the performance of these two materials, it may also result from selection bias from the OxZr products' being used in patients at increased risk for failure, such as those less than age 55. At 5 years, the registry shows cumulative revision rates of 2.1% for CoCr models and 4.1% for OxZr models with posterior-stabilized designs.³ The findings in the current study are consistent with the results from the Australian registry.

Our results also suggest that both age at time of surgery and sex have an influence on overall survivorship. Increased age at time of surgery decreased the revision rate, with a hazard ratio of 0.968, which is in keeping with registry data. Data from the National Joint Registry of England, Wales and Northern Ireland and the Isle of Man demonstrate decreasing cumulative revision rates at each incremental increase in age of 5 years.⁴ Similarly, the New Zealand Joint Registry illustrates a statistically significant decrease in revision rate with each successive incremental increase in age of 10 years.²¹ In our series, female sex had a positive influence on implant survivorship. Again, this is in keeping with reports from several national joint registries.^{3,4,21} The 2016 report from National Joint Registry of England, Wales and Northern Ireland and the Isle of Man

Table 1. Patient demographic and clinical characteristics

Characteristic	No. (%) of patients*
Female sex (<i>n</i> = 2178)	1359 (62.4)
Age, yr, mean ± SD	64.7 ± 10.7
Body mass index, mean ± SD	35.0 ± 8.5
Cobalt–chromium implant (<i>n</i> = 2815)	2286 (81.2)
Length of follow-up, yr, mean ± SD	3.2 ± 1.9
SD = standard deviation.	
*Except where noted otherwise.	

Table 2. Patient demographic characteristics by femoral component material

Characteristic	Material; no. (%) of implants*		<i>p</i> value
	Cobalt–chromium <i>n</i> = 2286	Oxidized zirconium <i>n</i> = 529	
Female sex	1483 (83.4)	296 (16.6)	< 0.001
Male sex	803 (77.5)	233 (22.5)	
Age, yr, mean ± SD	67.4 ± 9.4	52.3 ± 7.5	< 0.001
Body mass index, mean ± SD	35.0 ± 9.0	34.3 ± 8.3	0.1
SD = standard deviation.			
*Except where noted otherwise.			

Table 3. Reasons for revision

Diagnosis	No. of cases <i>n</i> = 63
Infection	23
Instability	9
Stiffness	6
Extensor mechanism injury	5
Aseptic loosening	3
Patellar maltracking	7
Locking mechanism dissociation	2
Hematoma	1
Recurrent hemarthrosis	1
Pain	2
Fracture	1
Patellar clunk	1
Undocumented	2

Table 4. Risk of revision by age, sex, body mass index and femoral implant material

Predictor	HR (95% CI)
Oxidized zirconium femur	1.110 (0.567–2.179)
Age	0.968 (0.939–0.997)
Female sex	0.450 (0.269–0.754)
Body mass index	1.023 (0.994–1.054)

CI = confidence interval; HR = hazard ratio.

Fig. 2. Kaplan–Meier product-limit survival estimates for cobalt-chromium (CoCr) and oxidized zirconium (OxZr) femoral implants. Grey shading represents 95% confidence intervals.

indicates that males are more likely to undergo revision than females of comparable age if they were less than 75 years old at the time of primary surgery.⁴ Given that the predominant indication for revision in our study was infection, this may be related to the finding of Rasouli and colleagues²² that male sex was associated with an increased risk of surgical site infection, with an odds ratio of 1.79.

Femoral bearing surface did not have a statistically significant influence on overall survivorship of the Legion Primary implant in our series, even when we controlled for

the possible confounding effects of age, sex and body mass index. The OxZr implant is a premium implant with an associated increase in cost. The differential varies between centres based on regional contracting; however, the cost of the OxZr femoral component may be as much as 4 times that of the otherwise identical CoCr component. Tribological studies have suggested superior characteristics of OxZr femoral implants in comparison to the standard CoCr femoral implants.^{7,23–25} Interestingly, these superior characteristics have not resulted in consistent published reports of improvement in in vivo wear or superior clinical results in knee arthroplasty. This is in contrast to total hip arthroplasty, where data from the Australian joint registry suggest superiority of this bearing surface in terms of survival.³ Matched comparison retrieval studies have shown a decrease in femoral component damage in OxZr components^{8,26,27} and a corresponding decrease in polyethylene damage.²⁶ Synovial fluid analysis in patients with bilateral total knee replacement failed to show differences in polyethylene particle number, weight, size and shape between the CoCr knee and the OxZr knee,²⁸ and no clinical differences were found between groups.²⁹ A small randomized controlled trial evaluating OxZr and CoCr implants in patients undergoing simultaneous total knee arthroplasty procedures failed to show significant differences in clinical, subjective and radiographic outcomes at 5 years.³⁰ Vertullo and colleagues³¹ reviewed 17 577 cases from the Australian joint registry involving a cruciate-retaining asymmetric-flexion knee and found no difference in implant survival in patients less than 75 years of age, with poorer survival of OxZr implants at or beyond this age. Our study supports these findings in a posterior-stabilized symmetric flexion knee by the same single manufacturer.

CONCLUSION

Although there are theoretical advantages to using the OxZr implant bearing surface, we did not detect an improvement in early to midterm implant survivorship. With no apparent survival benefit to the OxZr surface in knees, the use of these more expensive implants may not be warranted to extend implant longevity until longer-term evidence of superiority is presented. The symmetric posterior condylar posterior-stabilized design as a whole offers excellent early to midterm survivorship, with a cumulative revision rate of 1.8% at 2 years and 3.2% at 5 years in the current study. The findings support the continued use of this design iteration, which is based on a previous implant with an excellent record. Both increased age and female sex had a positive influence on implant survivorship; however, we could not detect any effect of body mass index on survivorship.

Acknowledgements: This study was supported by funds granted from the Thorlakson Chair at the Department of Surgery, University of Manitoba.

Affiliations: From Kelowna Orthopedics, Kelowna, BC (Demcoe); and the Concordia Hip and Knee Institute, Division of Orthopaedic Surgery, University of Manitoba, Winnipeg, Man. (Bohm, Hedden, Burnell, Turgeon).

Competing interests: A.R. Demcoe has no conflicts to declare. E.R. Bohm, D.R. Hedden, C.D. Burnell and T.R. Turgeon all declare institutional research support (unrelated to this project) from DePuy Synthes, Hip Innovation Technology, Smith & Nephew and Zimmer, and institutional fellowship educational funding from DePuy Synthes and Smith & Nephew. E.R. Bohm declares speakers bureau/paid presentations for Zimmer, DePuy Synthes and Stryker and unpaid consultancy for Smith & Nephew. D.R. Hedden declares a paid consultancy for Smith & Nephew. T.R. Turgeon declares speakers bureau/paid presentations for DePuy Synthes and unpaid consultancy for Smith & Nephew.

Contributors: A.R. Demcoe, E.R. Bohm and T.R. Turgeon designed the study. D.R. Hedden and C.D. Burnell acquired the data, which A.R. Demcoe, E.R. Bohm and T.R. Turgeon analyzed. A.R. Demcoe and T.R. Turgeon wrote the article, which all authors reviewed. All authors approved the final version to be published and can certify that no other individuals not listed as authors have made substantial contributions to the paper.

References

- Bourne RB, Laskin RS, Guerin JS. Ten-year results of the first 100 Genesis II total knee replacement procedures. *Orthopedics* 2007;30 (Suppl):83-5.
- Laskin RS, Davis J. Total knee replacement using the Genesis II prosthesis: a 5-year follow up study of the first 100 consecutive cases. *Knee* 2005;12:163-7.
- Annual report 2015*. Adelaide (Australia): Australian Orthopaedic Association National Joint Replacement Registry; 2015.
- 13th annual report*. Hemel Hempstead (UK): National Joint Registry for England, Wales, Northern Ireland and the Isle of Man; 2016.
- Bhandari M, Pascale W, Sprague S, et al. The Genesis II in primary total knee replacement: a systematic literature review of clinical outcomes. *Knee* 2012;19:8-13.
- Spector BM, Ries MD, Bourne RB, et al. Wear performance of ultra-high molecular weight polyethylene on oxidized zirconium total knee femoral components. *J Bone Joint Surg Am* 2001;83-A (Suppl 2 Pt 2):80-6.
- White SE, Whiteside LA, McCarthy DS, et al. Simulated knee wear with cobalt chromium and oxidized zirconium knee femoral components. *Clin Orthop Relat Res* 1994;309:176-84.
- Brandt JM, Guenther L, O'Brien S, et al. Performance assessment of femoral knee components made from cobalt-chromium alloy and oxidized zirconium. *Knee* 2013;20:388-96.
- Laskin RS. An oxidized Zr ceramic surfaced femoral component for total knee arthroplasty. *Clin Orthop Relat Res* 2003;416:191-6.
- Innocenti M, Carulli C, Matassi F, et al. Total knee arthroplasty in patients with hypersensitivity to metals. *Int Orthop* 2014;38:329-33.
- Hofer JK, Ezzet KA. A minimum 5-year follow-up of an oxidized zirconium femoral prosthesis used for total knee arthroplasty. *Knee* 2014;21:168-71.
- Han HS, Kang SB, Yoon KS. High incidence of loosening of the femoral component in Legacy posterior stabilised-flex total knee replacement. *J Bone Joint Surg Br* 2007;89:1457-61.
- Middleton RG, Howie DW, Costi K, et al. Effects of design changes on cemented tapered femoral stem fixation. *Clin Orthop Relat Res* 1998;355:47-56.
- Dawson J, Fitzpatrick R, Murray D, et al. Questionnaire on the perceptions of patients about total knee replacement. *J Bone Joint Surg Br* 1998;80:63-9.
- Garratt AM, Brealey S, Gillespie WJ, et al. Patient-assessed health instruments for the knee: a structured review. *Rheumatology (Oxford)* 2004;43:1414-23.
- Murray DW, Fitzpatrick R, Rogers K, et al. The use of the Oxford hip and knee scores. *J Bone Joint Surg Br* 2007;89:1010-4.
- Dy CJ, Marx RG, Bozic KJ, et al. Risk factors for revision within 10 years of total knee arthroplasty. *Clin Orthop Relat Res* 2014;472: 1198-207.
- Koh JJ, Cho WS, Choi NY, et al.; Kleos Korea Research Group. Causes, risk factors, and trends in failures after TKA in Korea over the past 5 years: a multicenter study. *Clin Orthop Relat Res* 2014;472: 316-26.
- Pfefferle KJ, Gil KM, Fening SD, et al. Validation study of a pooled electronic healthcare database: the effect of obesity on the revision rate of total knee arthroplasty. *Eur J Orthop Surg Traumatol* 2014;24: 1625-8.
- Si HB, Zeng Y, Shen B, et al. The influence of body mass index on the outcomes of primary total knee arthroplasty. *Knee Surg Sports Traumatol Arthrosc* 2015;23:1824-32.
- The New Zealand Joint Registry: sixteen year report. January 1999 to December 2014*. Wellington (NZ): New Zealand Orthopaedic Association; 2015.
- Rasouli MR, Restrepo C, Maltenfort MG, et al. Risk factors for surgical site infection following total joint arthroplasty. *J Bone Joint Surg Am* 2014;96:e158.
- Tsukamoto R, Chen S, Asano T, et al. Improved wear performance with crosslinked UHMWPE and zirconia implants in knee simulation. *Acta Orthop* 2006;77:505-11.
- Ezzet KA, Hermida JC, Colwell CW Jr, et al. Oxidized zirconium femoral components reduce polyethylene wear in a knee wear simulator. *Clin Orthop Relat Res* 2004;428:120-4.
- Ezzet KA, Hermida JC, Steklou N, et al. Wear of polyethylene against oxidized zirconium femoral components effect of aggressive kinematic conditions and malalignment in total knee arthroplasty. *J Arthroplasty* 2012;27:116-21.
- Heyse TJ, Chen DX, Kelly N, et al. Matched-pair total knee arthroplasty retrieval analysis: oxidized zirconium vs. CoCrMo. *Knee* 2011;18:448-52.
- Heyse TJ, Elpers ME, Nawabi DH, et al. Oxidized zirconium versus cobalt-chromium in TKA: profilometry of retrieved femoral components. *Clin Orthop Relat Res* 2014;472:277-83.
- Kim YH, Kim JS, Huh W, et al. Weight of polyethylene wear particles is similar in TKAs with oxidized zirconium and cobalt-chrome prostheses. *Clin Orthop Relat Res* 2010;468:1296-304.
- Kim YH, Park JW, Kim JS. Comparison of the Genesis II total knee replacement with oxidized zirconium and cobalt-chromium femoral components in the same patients: a prospective, double-blind, randomised controlled study. *J Bone Joint Surg Br* 2012;94:1221-7.
- Hui C, Salmon L, Maeno S, et al. Five-year comparison of oxidized zirconium and cobalt-chromium femoral components in total knee arthroplasty: a randomized controlled trial. *J Bone Joint Surg Am* 2011;93:624-30.
- Vertullo CJ, Lewis PL, Graves S, et al. Twelve-year outcomes of an Oxinium total knee replacement compared with the same cobalt-chromium design: an analysis of 17,577 prostheses from the Australian Orthopaedic Association National Joint Replacement Registry. *J Bone Joint Surg Am* 2017;99:275-83.