Published in final edited form as:

J Microbiol Methods. 2019 June ; 161: 71–73. doi:10.1016/j.mimet.2019.04.013.

Chromagar™ requires secondary confirmation strategies to minimize false positive/negative results for detection of *Staphylococcus aureus*

Sirisha Sirobhushanam¹, Navya Parsa², Tamra J. Reed¹, and J. Michelle Kahlenberg^{1,3}

¹Department of Internal Medicine, Division of Rheumatology, University of Michigan, Ann Arbor, MI 48109.

²University of Toledo, Toledo, OH.

Abstract

Chromagar is a medium that is used in culture-based colonization studies of *Staphylococcus aureus*. We have found that *S. aureus* negative colonies often fit the color recommendations for *S. aureus* identification and can cause overestimation of colonization rates. Confirmation of suspect colonies is important to minimize false negative/positive results.

Keywords

Chromagar; Colonization; *Staphylococcus aureus*

Note

Staphylococcus aureus is a pathobiont that colonizes ~30% of the healthy adult population and is the leading cause of skin and soft tissue infections in the world (Balasubramanian et al., 2017; Wertheim et al., 2005). Skewing of the skin microbiome to promote colonization by *S. aureus* is associated with cutaneous disease severity for diseases such as atopic dermatitis (Meylan et al., 2017; Williams et al., 2017). This has resulted in resurgence in *S. aureus* colonization studies (Conti et al., 2016; Hajialilo et al., 2015; Ng et al., 2017; Williams and Gallo, 2017; Wollina, 2017). These investigations often require isolation of the bacterium from survey populations performed by individuals who lack in-depth knowledge of *S. aureus* and its growth characteristics. Thus, methodology that is quick, inexpensive for large numbers of samples, and reliable in results is required for these studies.

3 Corresponding Author: Dr. J M Kahlenberg, Department of Internal Medicine, Division of Rheumatology, University of Michigan, Ann Arbor, MI, 48109, USA. mkahlenb@med.umich.edu.

Publisher's Disclaimer: This is a PDF file of an unedited manuscript that has been accepted for publication. As a service to our customers we are providing this early version of the manuscript. The manuscript will undergo copyediting, typesetting, and review of the resulting proof before it is published in its final citable form. Please note that during the production process errors may be discovered which could affect the content, and all legal disclaimers that apply to the journal pertain.

CONFLICT OF INTEREST

J.M.K. serves on advisory boards for AstraZeneca and Eli Lilly. The other authors have no financial conflicts of interest.

Growth of bacteria on blood agar is effective but does not allow for selection of staphylococci. Growth of bacteria in Mannitol Salt Agar (MSA) selects for staphylococci (high salt media) and differentiates *S. aureus* (mannitol fermentation). Unfortunately, false positive results derived from coagulase negative *staphylococci* cause overestimation of colonization rates (Shittu et al., 2006). Chromagar™ is recognized as an effective medium for identifying *S. aureus* in comparison to MSA by various studies (Carricajo et al., 2002; Flayhart et al., 2004; Gaillot et al., 2006; Han et al., 2007; Luteijn et al., 2011).

We employed rigorous methodology to survey our patient population for *S. aureus* colonization. Patients from the University of Michigan Rheumatology clinic (n=51) were consented according to IRBMED #HUM00116841 and swabbed in the nares and on various areas of their skin using sterile FLOQSwabs (Copan Diagnostics, Murrieta, CA) wetted with sterile PBS. Swabs were placed in PBS and vortexed, 100 µl was spread on plates (Chromagar) and incubated at 35 ± 2 °C for 24 hours. Mauve colored colonies were restreaked onto MSA plates and incubated for 16–20 hours at 37 °C and observed for mannitol fermentation. All colonies were restreaked on Chromagar to present clear photographic evidence (Figures 1 and 2). This was followed by colony PCR for *S. aureus* specific region in the thermonuclease gene using the primers SaNuc forward 5'-GCGATTGATGGTGATACGGTT -3' and reverse 5'-AGCCAAGCCTTGACGAAGTAAAGC -3' as a gold standard to confirm the presence of *S. aureus* (Zhang et al., 2004). Our approach has been summarized as a flowchart in Fig. 1A. Each step of the pathway was accompanied by positive and negative controls using *S. aureus* strain USA300, a community acquired methicillin-resistant *S. aureus*, generously shared by Dr. Alexander R. Horswill, Department of Immunology and Microbiology, University of Colorado, and *Staphylococcus epidermidis* H140 obtained from BEI resources.

S. aureus positive colonies that displayed the expected color on Chromagar are presented in Fig. 1B and *S. aureus* negative colonies that were easily identified are presented in Fig. 1C. However, several colonies did not fit the expected color characteristics. Fig. 2A shows *S. aureus* positive colonies (by PCR) that were atypical in appearance at 24 hrs. Further incubation resulted in persistence of this appearance in some colonies while others showed intense mauve color at 48 hrs (Fig. 2B). The most problematic results were those *S. aureus* negative colonies (confirmed by PCR) that exhibited varying shades of the color mauve/pink and were found to be indistinguishable from typical *S. aureus* positive colonies at both 24 and 48 hrs incubation (Fig. 2C and D).

We also observed that samples that contain high numbers of *S. aureus* were easy to screen and less likely to display false negative results. Presence of other staphylococci including *S. epidermidis* in high numbers made *S. aureus* identification more difficult resulting in higher false negative and false positive results. False positive results were lower in Chromagar (7%) than MSA (11.5%) at 24 hrs but increase dramatically by 48 hrs (13%) (Table. 1). This indicates that although chromagar was more specific than MSA at 24 hrs further incubation results in misidentification. However, false negative results were fairly high for both media types (31% Chromagar and 39% on MSA). This is likely due to *S. aureus* positive colonies that did not show typical appearance and Coagulase negative Staphylococci that ferment mannitol.

In summary, we have identified Chromagar as an excellent media for *S. aureus* identification in line with other groups (Carricajo et al., 2002; Han et al., 2007; Luteijn et al., 2011). However, we also report a 7% false-positive rate and 31% false-negative rate when colonies are judged by color alone. This is in line with published work (both report ~20% false-negative results) (D'Souza and Baron, 2005; Luteijn et al., 2011). In this work we also pursued colonies that had typical *S. aureus* appearance on the media, although not the required color, with colony PCR and identified several that were in fact *S. aureus*. We recommend that colonies exhibiting other features of *S. aureus* colony morphology such as yellow color and circular, pinhead shape (1–2 mm) with a smooth margin, that appear on the agar ought to be pursued by confirmatory methods (Ribeiro de Souza da Cunha, 2017). Confirmation could involve plating on MSA to ensure growth in high salt media that is typical of staphylococci followed by PCR confirmation.

Secondary confirmation strategies proved particularly useful in our study since our initial screen yielded very high positive rates that were later corrected due to our rigorous methodology. A brief search in Pubmed with the term “Chromagar *Staphylococcus aureus*” yielded ~110 publications demonstrating its widespread use. However, addition of terms associated with confirmation such as tube coagulase, catalase or PCR resulted in significantly lower number of results indicating that secondary confirmation was not widely employed. Thus, Chromagar, while useful in screening large cohorts for *S. aureus*, should not be considered a confirmatory test for *S. aureus* and other methods such as PCR must be employed to yield accurate results.

ACKNOWLEDGEMENTS

We appreciate the generosity of the patients from the Michigan Rheumatology Clinics for sharing their samples. This work was funded by the Physician Scientist Development award from the Doris Duke Foundation (PI: JMK) and by the National Institute of Arthritis and Musculoskeletal Diseases via R01 AR071384 (PI: JMK).

REFERENCES CITED

- Balasubramanian D, Harper L, Shopsis B, Torres VJ, 2017 *Staphylococcus aureus* pathogenesis in diverse host environments. *Pathog. Dis* 75, 1–13. 10.1093/femspd/ftx005
- Carricajo A, Treny A, Fonsale N, Bes M, Reverdy ME, Gille Y, Aubert G, Freydiere AM, 2002 Performance of the Chromogenic Medium CHROMagar Staph Aureus and the Staphychrom Coagulase Test in the Detection and Identification of *Staphylococcus aureus* in Clinical Specimens. *J. Clin. Microbiol* 39, 2581–2583. 10.1128/jcm.39.7.2581-2583.2001
- Conti F, Ceccarelli F, Iaiani G, Perricone C, Giordano A, Amori L, Miranda F, Massaro L, Pacucci VA, Truglia S, Girelli G, Fakeri A, Taliani G, Temperoni C, Spinelli FR, Alessandri C, Valesini G, 2016 Association between *Staphylococcus aureus* nasal carriage and disease phenotype in patients affected by systemic lupus erythematosus. *Arthritis Res. Ther* 18, 177 10.1186/s13075-016-1079-x [PubMed: 27475749]
- D'Souza HA, Baron EJ, 2005 BBL CHROMagar Staph aureus is superior to Mannitol salt for detection of *Staphylococcus aureus* in complex mixed infections. *Am. J. Clin. Pathol* 123, 806–808. 10.1309/FVHRF3GRLEQXGBAG [PubMed: 15899769]
- Flayhart D, Lema C, Borek A, Carroll KC, 2004 Comparison of Two Prototypes of BBL™ CHROMagar™ MRSA to Conventional Media for the Detection of Methicillin Resistant *Staphylococcus aureus* in Clinical Samples. *J. Clin. Microbiol* 42, 3566–3569. 10.1128/JCM.42.8.3566 [PubMed: 15297498]

- Gaillot O, Wetsch M, Fortineau N, Berche P, 2006 Evaluation of a new chromogenic medium for isolation and presumptive identification of methicillin-resistant *Staphylococcus aureus* from human clinical specimens. *Eur. J. Clin. Microbiol. Infect. Dis* 25, 407–409. 10.1007/s10096-006-0144-9 [PubMed: 16767492]
- Hajjalilo M, Ghorbanihaghjo A, Khabbazi A, Valizadeh H, Raeisi S, Hasani A, Varshochi M, Kolahi S, Nakhjavani MR, 2015 Nasal carriage rate of *Staphylococcus aureus* among patients with systemic lupus erythematosus and its correlation with disease relapse. *Egypt. Rheumatol* 37, 81–84. 10.1016/j.ejr.2014.06.006
- Han Z, Lautenbach E, Fishman N, Nachamkin I, 2007 Evaluation of mannitol salt agar, CHROMagar *Staph aureus* and CHROMagar MRSA for detection of methicillin-resistant *Staphylococcus aureus* from nasal swab specimens. *J. Med. Microbiol* 56, 43–46. 10.1099/jmm.0.46777-0 [PubMed: 17172515]
- Luteijn JM, Hubben GAA, Pechlivanoglou P, Bonten MJ, Postma MJ, 2011 Diagnostic accuracy of culture-based and PCR-based detection tests for methicillin-resistant *Staphylococcus aureus*: A meta-analysis. *Clin. Microbiol. Infect* 17, 146–154. 10.1111/j.1469-0691.2010.03202.x [PubMed: 20219085]
- Meylan P, Lang C, Mermoud S, Johannsen A, Norrenberg S, Hohl D, Vial Y, Prod'homme G, Greub G, Kypriotou M, Christen-Zaech S, 2017 Skin Colonization by *Staphylococcus Aureus* Precedes the Clinical Diagnosis of Atopic Dermatitis in Infancy. *J. Invest. Dermatol* 2497–2504. 10.1016/j.jid.2017.07.834 [PubMed: 28842320]
- Ng CY, Huang YH, Chu CF, Wu TC, Liu SH, 2017 Risks for *Staphylococcus aureus* colonization in patients with psoriasis: a systematic review and meta-analysis. *Br. J. Dermatol* 177, 967–977. 10.1111/bjd.15366 [PubMed: 28160277]
- de L. Ribeiro de Souza da Cunha M, 2017 Methods for the Identification, Characterization, and Tracking the Spread of *Staphylococcus aureus*, *Staphylococcus aureus*. Elsevier Inc. 10.1016/b978-0-12-809671-0.00006-1
- Shittu A, Lin J, Morrison D, Kolawole D, 2006 Identification and molecular characterization of mannitol salt positive, coagulase-negative staphylococci from nasal samples of medical personnel and students. *J. Med. Microbiol* 55, 317–324. 10.1099/jmm.0.46072-0 [PubMed: 16476796]
- Wertheim HFL, Melles DC, Vos MC, Van Leeuwen W, Van Belkum A, Verbrugh HA, Nouwen JL, 2005 The role of nasal carriage in *Staphylococcus aureus* infections. *Lancet* 5, 751–762. 10.1016/S1473-3099(05)70295-4 [PubMed: 16310147]
- Williams MR, Gallo RL, 2017 Evidence that Human Skin Microbiome Dysbiosis Promotes Atopic Dermatitis. *J. Invest. Dermatol* 137, 2460–2461. 10.1016/j.jid.2017.09.010 [PubMed: 29169458]
- Williams MR, Nakatsuji T, Gallo RL, 2017 *Staphylococcus aureus*: Master Manipulator of the Skin. *Cell Host Microbe* 22, 579–581. 10.1016/j.chom.2017.10.015 [PubMed: 29120738]
- Wollina U, 2017 Microbiome in atopic dermatitis. *Clin. Cosmet. Investig. Dermatol* 10, 51–56. 10.2147/CCID.S130013
- Zhang K, Sparling J, Chow BL, Elsayed S, Hussain Z, Church DL, Gregson DB, Louie T, Conly JM, 2004 New Quadriplex PCR Assay for Detection of Methicillin and Mupirocin Resistance and Simultaneous Discrimination of *Staphylococcus aureus* from Coagulase-Negative Staphylococci 42, 4947–4955. 10.1128/JCM.42.11.4947

Highlights:

- Chromagar is more sensitive than MSA for identifying *S. aureus* from clinical swab samples.
- Both Chromagar and MSA demonstrate high false negative results
- Secondary confirmation is critical to obtain accurate colonization rates.

Fig. 1.

A. Flowchart for *S. aureus* confirmation methodology. **B.** Single colonies of confirmed *S. aureus* positive isolates were re-streaked on Chromagar to clearly show color differences. **C.** Single colonies of *S. aureus* negative (via PCR) isolates were restreaked on Chromagar to clearly show color differences. Photographs were taken at 24 hours post inoculation.

A. Atypical *S. aureus* colonies at 24 hrs

B. Atypical *S. aureus* colonies at 48 hrs

C. *S. aureus* negative colonies at 24 hrs

D. *S. aureus* negative colonies at 48 hrs

Fig. 2.

A. Single colonies of confirmed *S. aureus* positive isolates that were atypical on Chromagar were restreaked to show color differences at 24 hrs. **B.** Single colonies of confirmed *S. aureus* positive isolates that were atypical on Chromagar were restreaked to show color differences at 48 hrs. **C.** *S. aureus* negative colonies that were atypical on Chromagar were restreaked to show color differences at 24 hrs. **D.** *S. aureus* negative colonies that were atypical on Chromagar were restreaked to show color differences at 48 hrs.

Table. 1.

Comparison of Chromagar, MSA and *S. aureus nuc* Colony PCR for detection.

Selection on MSA			False Positive	False Negative
MSA (+)	Chromagar (+) 48 hrs	PCR (+)		
26	23	23	11.5%	
MSA (–)	Chromagar (+) 48 hrs	PCR (+)		
28	8	11		39.3%
Selection on Chromagar				
Chromagar (+)	MSA (+)	PCR (+)		
24 hours 28	23	26	7.1%	
48 hours 31	23	27	12.9%	
Chromagar (–)	MSA (+)	PCR (+)		
24 hours 26	3	8		30.8%
48 hours 23	3	7		30.4%

Total number of samples tested 54.