Received:
24 October 2018

Revised:
26 February 2019

Accepted:
18 March 2019

<https://doi.org/10.1259/bjr.20180918>

Cite this article as:

King AJ, Dudderidge T, Darekar A, Schimitz K, Heard R, Everitt C, et al. Establishing MRI-guided prostate intervention at a UK Centre. *Br J Radiol* 2019; **92**: 20180918.

SHORT COMMUNICATION

Establishing MRI-guided prostate intervention at a UK Centre

ALEXANDER J KING, FRCR, TIM DUDDERIDGE, FRCS(urol), ANGELA DAREKAR, KATYA SCHIMITZ, RORY HEARD, CHRIS EVERITT, ROBERT CHAMBERS, FRCA and DAVID BREEN, FRCR

University Hospital Southampton NHS Foundation Trust, Hampshire

Address correspondence to: Dr Alexander J King
E-mail: alexking@doctors.org.uk

Objectives: To describe our preliminary experience in establishing an MRI suite capable to deliver targeted prostate biopsy and cryoablation.

Methods: This article includes a description of the necessary infrastructure alterations, scanning sequence suggestions, anaesthetic advice, and practical procedural considerations. We aim to examine the anticipated issues most UK centres would encounter and offer our experience in overcoming them. During this process we will also explore some of the technical aspects of MRI-guided prostate biopsy and cryoablation.

Results: The clinical indication, treatment rationale, intervention strategy, and initial clinical outcomes are described for our first series of patients.

Conclusion: MRI-guided prostate intervention provides many theoretical advantages over traditional TRUS guidance. This article demonstrates some of the complexities encountered in establishing this technique in a UK centre, and the proposed solutions.

Advances in knowledge: This article gives an account of establishing the first MRI intervention suite in the UK. It demonstrates some of the logistical considerations, and offers the unit's early experience.

INTRODUCTION

This report describes the baseline requirements in establishing an MRI-guided prostate intervention service. Specifically, it details some of the logistical and technical considerations to adapt an MRI suite to be capable of delivering MRI-guided transperineal prostate biopsy, and MRI-guided cryoablation. Herein we outline the enablement work required and practical considerations involved in instigating an MR-guided prostate intervention and cryoablation service.

RATIONALE FOR MRI-GUIDED PROSTATE INTERVENTION

MRI has become a central part of prostate cancer investigation and monitoring. Its high sensitivity for clinically significant prostate cancer has recently been evaluated in the PROMIS¹ trial and PRECISION STUDY.² Consequently, the ability of MRI to identify focal disease has been widely recognised by radiologists, urologists and industrial partners, with implications for less-invasive targeted therapy and its delivery. To date, prostate biopsy and focal therapies have largely been delivered under TRUS-guidance. Whilst ultrasound provides a portable, real time imaging modality, it is limited in its diagnostic sensitivity and often shows poor correlation with MRI detected lesions.³

Similarly, appreciation of collateral structures at risk and their visualisation is inevitably compromised during the delivery of ablative energies. This acknowledgment has led to the development of several MRI-TRUS fusion platforms that attempt to match the MRI detected lesion to a realtime TRUS image for targeted biopsy. Despite the sophistication offered by many of these packages, there are unavoidable approximations in fusion imaging with either deformable or rigid registration. Inherently it makes sense to preferentially use the imaging modality that best identifies the lesion, to ultimately target the lesion.

Parallel to the ascendance of prostate MRI, prostate cryoablation has undergone a renaissance in recent years with the advent of 17G third generation argon systems. MR guidance permits excellent appreciation of ice-ball distribution and its accurate, controlled delivery. As a TRUS-guided procedure prostate cryoablation is relatively fast, but has some limitations. A major concern is that once the ice begins to form, the only visible margin is the ice-edge moving towards the rectum. The anterior ice margin, and to a lesser extent the lateral margins, are inconspicuous. Consequently, the confirmation of adequate ice coverage of the prostate anteriorly is reliant on proxy measurements with thermal sensors. Inevitably this raises concerns regarding

treatment margins and inadvertent collateral damage. In theory, unrecognised incomplete treatment and intraprocedural complications encountered with TRUS-guided cryoablation, could potentially be avoided or anticipated earlier with MRI guidance.

In short, MRI guidance should confer several advantages over TRUS including:

- true intraprocedural visualisation of the target lesion
- appreciation of the entire circumferential ice-ball margin
- identification and avoidance of vulnerable adjacent structures
- precise probe placement without the angulation constraints of a transperineal grid
- protective prostate-rectal space hydrodissection or haemodissection⁴ without loss of image quality

ROOM ALTERATIONS AND CRYOABLATION EQUIPMENT REQUIREMENTS

There are multiple infrastructure and equipment considerations to take into account when establishing a cryoablation-compatible MRI suite.

Several pieces of the required equipment cannot be taken into the scan room and so must be operated externally, with RF-cage penetration panels and adequate wave-guide ports required to be built into the scanner room wall to permit the passage of connecting tubes. The gas cylinders and control unit that generate the ice ('Seednet' – BTG/Galil) must not enter the scan room, thus the gas pipes are required to be ducted through the scanner room wall. Once these panels have been installed, a junction box can be sited inside the room which allows gas-carrying hoses to connect to a pedestal nearer the patient, into which the cryoprobes are attached (Figure 1). In order to reduce the risk of ice injury to the urethra, a catheter is sited which circulates warm fluid through the urethra, within a closed system. The fluid warmer and associated pump for the 'urethral warming catheter' also need to be located remote from the MR field and require extension tubing to be passed through wave-guides into the room (Figure 2).

Figure 1. Plan view of scan room layout and pedestal.

Figure 2. Urethral warming catheter, tubing and pump.

Given the nature of the procedure the patient will require a general anaesthetic. Most major centres will have addressed the increasing need for anaesthetic support in the Radiology department as the number and complexity of interventions have grown. Similarly, the challenges of anaesthetising patients in remote sites and the risks posed from the MRI suite are well documented.^{5,6} For these particular procedures the patient is anaesthetised for ~3 h which is generally longer than for most GA MRI patients. This requires a secure airway so the patient is intubated. The procedure is relatively non-stimulating so maintaining adequate blood pressure requires vassopressor infusion delivered via a syringe pump. This is also sited outside the scan room and the drug is delivered via long extension tubing. The patient is supine with partial knee flexion and hip adduction for a number of hours so careful positioning and padding is essential to avoid pressure areas and nerve injury. The risks of thrombo-embolic disease were relatively unknown, but in these cases the length of procedure and patient factors will usually require TED stockings and LMWH as prophylaxis. Intermittent pneumatic compression devices would have been preferred however the pumps available to us are not MRI compatible. The ability to have an anaesthetist and operative department practitioner (ODP) familiar with delivering invasive procedures to anaesthetised patients within the MRI/radiology department cannot be underestimated.

Specific perineal grid and patient positioning adjuncts have been developed by specialist international MR-guided prostate

Figure 3. (A) Perineal access device, with central square slot for placement of a perineal grid, if required. (B) The grid can be placed beneath the inferior aspect of the coil to aid perineal access.

cryoablation centres. These can attach to the scanner bed and may help to position the legs. They can also hold an integrated coil. Due to legislative limitations this equipment was not available to us, so as an alternative we constructed a plastic template that permitted access to the perineum, with the ability to insert a perineal grid if required (Figure 3).

The use of an in-room audio-visual monitor inevitably reduces the procedural time, by removing the need to intermittently exit the room and review the recent scans.

PROCEDURE AND PATIENT CONSIDERATIONS

All patients and staff must undergo the usual MRI safety checks and questionnaires. In addition, we undertake a pre-procedure team briefing detailing specific patient concerns, assigning specific responsibilities and procedures in case of a clinical emergency, which may necessitate evacuation of the scanner (cardiac arrest, equipment failure etc.).

The patient is anaesthetised in a bay adjacent to the MRI scan room, and then transferred directly onto the detachable MRI scanner bed. At this point the patient is positioned with the knees slightly spread and the perineum exposed. Standard surgical skin preparation and draping can be undertaken and, if required, skin incisions are made to avoid taking a scalpel into the scan room.

Figure 4. Probe tip blooming artefact in axial plane.

In our experience the rigidity of the standard 17g perineal grid was counterproductive during probe or needle placement. Using a freehand technique permitted a greater range of freedom in terms of oblique angulations (to adequately treat all aspects of the prostate), and consequently a more refined probe arrangement. A freehand technique should be comfortable to radiologists skilled in probe placement, cognitive targeting and image-guided intervention and permits bespoke approaches to gland shape and ice ball sculpture.

IMAGING SEQUENCES AND LESION TARGETING

As has already been described, one of the major advantages in MRI prostate intervention is the ability to see a target lesion. In theory, a complete multiparametric MRI could be performed at the time of the procedure but we elected to initially acquire T2 sequences (standard and volumetric) for target delineation. The axially acquired volumetric sequence (T2-SPACE) was automatically reformatted into sagittal and coronal planes and used for subsequent planning.

Once the target has been identified or the ablation zone planned, the needle or probe is initially placed into the perineal fat using a cognitive targeting approach. A subsequent axial acquisition (T2-TRUFISP – balanced gradient echo) starting from the perineum, scanning superiorly through the prostate, allows identification of the needle and prediction of its path. The tip of the needle appears as a small spherical ‘blooming’ artefact (Figure 4), and enables localisation of the insertion point of the needle. The path of the needle can be tracked from this point through the more superior axial images (assuming these images are perpendicular to the needle’s path) to estimate the final target point. The acquisition time of these sequences is less than 1 min, which means they can be repeated as necessary (to aid initial repositioning) without adding too much delay to the procedure.

Figure 5. (A) Still images from 'fluoro' sequences permit almost real time depth and angulation perception. MRI also permits delineation of the target lesion and confirmation of accurate sampling (sagittal and coronal planes). (B) MPR of extended left anterior hemiablation. Note iceball void incorporating left gland (long arrow) with periurethral protection from urethral catheter (short arrow).

Subsequently, lower resolution, single-slice, dynamic TRUFISP sequences in the sagittal or coronal plane, parallel to the needle, allow almost real time 'MR fluoroscopy' imaging while the needle is advanced to its final position (Figure 5). Images are acquired every 0.8 sec to provide satisfactory dynamic time resolution with adequate spatial resolution. As the needle is inserted parallel to the B_0 magnetic field, the needle-related artefact is actually quite small. However the use of a balanced gradient echo sequence (TRUFISP), which is sensitive to magnetic susceptibility effects such as those produced by the needle, enhances this artefact and results in clearer visualisation of the needle.

Final needle placement is assessed using another T2 volume acquisition and once satisfied, delivery of the cryogen through the needle can commence. Iceball formation is monitored using fast T2-TRUFISP acquisitions (<1 min acquisition time) and T2-SPACE volume images are also acquired at specified points throughout the procedure, including assessing final iceball coverage at the end to ensure adequate ablation is achieved (Figure 5).

INITIAL CASE SELECTION

Our initial experience includes a short series of targeted biopsy or cryoablation in the prostate or deep pelvis. The clinical indications have been limited to patients with strongly suspected prostate cancer or post-operative recurrence in the deep pelvis but who are without rectal access due to previous surgery or prior treatment.

Immediate intra procedural imaging has demonstrated the ability to accurately place needles and probes via MRI guidance. So far, this targeted approach has yielded a high level of positive biopsies and representative histology results. Not only does this point towards the accuracy of MRI guidance to successfully target small, often sub centimetre lesions,⁷ but also acts as further testament for MRI as a lesion detecting tool.

Figure 6. 'Hydrodissection' of the rectum by injecting saline into the rectoprostatic space.

The resultant signal void from the ice ball appears to be a good surrogate for the genuine ice ball, with evidence from other centres suggesting good correlation between the signal void and ice edge, which consequently confers a standard 5 mm margin to the lethal ice.⁸ In line with the theoretically improved safety profile of MRI guidance, the signal voids are clearly seen to encompass the target in all planes without extending to the vulnerable adjacent structures. Should additional safety measures be required, it is possible to displace vulnerable structures by injecting saline at the appropriate position (Figure 6).

At the time of submission we have undertaken six prostate or pelvic cryoablations, and four biopsies. Representative tissue samples have been obtained in all cases and short term follow up of our cryoablation patients has not shown a significant complication or collateral damage to adjacent structures; suggesting a safe and well tailored ablation zone. As anticipated, the ablated

Figure 7. Ablated tissue involution over time. Note how the residual transition zone appears to shift leftwards as the ablated tissue contracts.

tissue steadily involutes over time (Figure 7) leaving a poorly enhancing scar.

DISCUSSION

Our early experience with MRI-guided intervention has so far demonstrated an apparently accurate and safe method for targeting and treating prostatic or deep pelvic disease. The inherent advantages of direct lesion targeting without the need for MRI/TRUS fusion, coupled with the ability to visualise the entirety of the ice ball and the surrounding structures, seem to have inherent logical advantages over a standard TRUS-guided procedure.

However, these procedures take a considerable time to plan and prepare for, as well as to perform. This process has involved MR

scan room infrastructure changes, new MRI sequence optimisation, and education of a novel technique to multiple staff specialties. The enthusiastic cross-collaboration between Radiographers, MRI Physicists, Anaesthetic staff, Urologists, Radiologists and their industrial partners has been essential to achieving this level of success. At University Hospital Southampton we have undoubtedly benefitted from having a mature image-guided ablation service within Radiology, and an established prostatic focal therapy practice in our Urology department. However, the importance of a cohesive partnership between Radiologist and Urologist, and a shared aim, cannot be understated when considering this type of development.

So far, our cases have each had a clinical need for an unconventional approach to prostate intervention (*i.e.* AP resection, anal stenosis, and previous radiotherapy), but this minimally invasive technique has offered a diagnostic and therapeutic option that was previously not available.

Aside from our current patient selection it will be important to evaluate whether the theoretical advantages of MRI guidance can transfer into clinical practice when compared directly to traditional TRUS guidance. Our local expectation is that ultimately some patients will remain suited to a TRUS-guided approach but there will be a significant number who benefit from the more precise targeting MRI can offer, and the real time simultaneous visualisation of the ice ball, adequate tumour coverage, and surrounding vulnerable anatomy.

REFERENCES

1. Ahmed HU, El-Shater Bosaily A, Brown LC, Gabe R, Kaplan R, Parmar MK, et al. Diagnostic accuracy of multi-parametric MRI and TRUS biopsy in prostate cancer (PROMIS): a paired validating confirmatory study. *Lancet* 2017; **389**: 815–22. doi: [https://doi.org/10.1016/S0140-6736\(16\)32401-1](https://doi.org/10.1016/S0140-6736(16)32401-1)
2. Kasivisvanathan V, Rannikko AS, Borghi M, Panebianco V, Mynderse LA, Vaarala MH, et al. MRI-Targeted or standard biopsy for prostate-cancer diagnosis. *N Engl J Med* 2018; **378**: 1767–77. doi: <https://doi.org/10.1056/NEJMoa1801993>
3. van de Ven WJM, Sedelaar JPM, van der Leest MMG, Hulsbergen-van de Kaa CA, Barentsz JO, Fütterer JJ, et al. Visibility of prostate cancer on transrectal ultrasound during fusion with multiparametric magnetic resonance imaging for biopsy. *Clin Imaging* 2016; **40**: 745–50. doi: <https://doi.org/10.1016/j.clinimag.2016.02.005>
4. Garnon J, Cazzato RL, Koch G, Uri IF, Tsoumakidou G, Caudrelier J, et al. Trans-rectal ultrasound-guided autologous blood injection in the Interprostato-rectal space prior to percutaneous MRI-guided cryoablation of the prostate. *Cardiovasc Intervent Radiol* 2018; **41**: 653–9. doi: <https://doi.org/10.1007/s00270-017-1853-4>
5. Amin A, Lane J, Cutter T. An anesthesiologist's view of tumor ablation in the radiology suite. *Anesthesiol Clin* 2017; **35**: 611–5. doi: <https://doi.org/10.1016/j.andclin.2017.07.006>
6. Tan TK, Goh J. The anaesthetist's role in the setting up of an intraoperative MR imaging facility. *Singapore Med J* 2009; **50**: 4–10.
7. Overduin CG, Fütterer JJ, Barentsz JO. MRI-guided biopsy for prostate cancer detection: a systematic review of current clinical results. *Curr Urol Rep* 2013; **14**: 209–13. doi: <https://doi.org/10.1007/s11934-013-0323-z>
8. Overduin CG, Jenniskens SFM, Sedelaar JPM, Bomers JGR, Fütterer JJ. Percutaneous MR-guided focal cryoablation for recurrent prostate cancer following radiation therapy: retrospective analysis of iceball margins and outcomes. *Eur Radiol* 2017; **27**: 4828–36. doi: <https://doi.org/10.1007/s00330-017-4833-9>