A Review of Commercially Available Point-of-Care Devices to Concentrate Bone Marrow for the Treatment of Osteoarthritis and Focal Cartilage Lesions

CARTILAGE
2019, Vol. 10(4) 387–394
© The Author(s) 2018
Article reuse guidelines:
sagepub.com/journals-permissions
DOI: 10.1177/1947603518768080
journals.sagepub.com/home/CAR

Florian Gaul^{1,2,3}, William D. Bugbee⁴, Heinz R. Hoenecke Jr.⁴,
and Darryl D. D'Lima^{1,2}

Abstract

Objective. Mesenchymal stem cells (MSCs) are a promising cell-based therapy treatment option for several orthopedic indications. Because culture expansion of MSC is time and cost intensive, a bedside concentration of bone marrow (BM) aspirate is used as an alternative. Many commercial systems are available but the available literature and knowledge regarding these systems is limited. We compared different point-of-care devices that concentrate BM (BMC) by focusing on technical features and quality parameters to help surgeons make informed decisions while selecting the appropriate device. **Methods.** We compared published data on the BMC devices of Arteriocyte, Arthrex, Celling Biosciences, EmCyte, Exactech, ISTO Tech, Harvest Tech/Terumo BCT, and Zimmer/Biomet regarding technical features (centrifugation speed/time, input/output volume, kit components, type of aspiration syringes, filter usage) and quality parameters of their final BMC product (hematocrit, concentration of platelets and total nucleated cells, concentration of MSC and connective tissue progenitor cells). **Results.** The systems differ significantly in their technical features and centrifugation parameters. Only the fully automated systems use universal kits, which allow processing different volumes of BM. Only the Arthrex system allows selection of final hematocrit. There was no standardized reporting method to describe biologic potency. **Conclusions.** Based on the data obtained in this review, recommending a single device is not possible because the reported data could not be compared between devices. A standardized reporting method is needed for valid comparisons. Furthermore, clinical outcomes are required to establish the true efficacy of these systems. We are conducting additional studies for more careful comparison among the devices.

Keywords

bone marrow aspirate concentrate, bone marrow concentrate, osteoarthritis, diagnosis, mesenchymal stem cells, cells, cartilage defect

Introduction

Autologous cell-based therapies like platelet-rich plasma (PRP) and bone marrow aspirate concentrate (BMAC)/bone marrow concentrate (BMC) are promising alternative treatment options for several orthopedic indications. Both are generated by density gradient centrifugation of either whole blood or bone marrow aspirate (BMA), which leads to a concentration of platelets and their important growth factors. The major difference and possible advantage of BMC in comparison to PRP is the content of mesenchymal stem cells (MSCs) that have the ability to differentiate into different connective tissue cell types (including chondrocytes) and therefore represent promising candidates to repair damaged tissue. Apart from that, MSC also possess immunomodulatory, anti-inflammatory, antiapoptotic, proliferative

and chemoattractive functions and can coordinate the differentiation process of functional tissue regeneration in host cells.^{1,2} Many studies show that the use of culture expanded MSC and BMC for the treatment of different orthopedic

¹Shiley Center for Orthopaedic Research and Education at Scripps Clinic, La Jolla, CA, USA

²Department of Molecular Medicine, The Scripps Research Institute, La Jolla, CA, USA

³Department of Orthopaedic, Trauma and Plastic Surgery, University Hospital Leipzig, German

⁴Division of Orthopaedic Surgery, Scripps Clinic, La Jolla, CA, USA

Corresponding Author:

Darryl D. D'Lima, Shiley Center for Orthopaedic Research and Education at Scripps Clinic, 10550 North Torrey Pine Road, MEM-1116, La Jolla, CA 92037, USA.
Email: ddlima@scripps.edu

conditions, for example, tendinopathy, osteonecrosis of the femoral head and especially osteoarthritis and focal cartilage lesions is feasible, safe, and can provide some clinical efficacy in terms of pain reduction, joint function and quality of life.³⁻¹⁰ Furthermore, in a recent study Cassano *et al.*¹¹ found a significant concentration of interleukin 1 receptor antagonist (IL-1ra) in BMC while absent in PRP. IL-1ra can inhibit the IL-1 catabolism that could even result in a higher anti-inflammatory potential.

Although MSC can be found and extracted from different tissues, such as skeletal muscles, umbilical cord blood and Wharton's jelly, the most common sources in orthopedic interventions are bone marrow and fat tissue. It has to be clarified that bone marrow-derived MSC represent only 0.001% to 0.01% of the mononuclear cells in the bone marrow whereas MSC occur at a 100- to 1000-fold higher concentration in adipose tissue.¹ Despite the higher concentration of stem cells in adipose tissue, the extraction of cells requires collagenase digestion, which takes 1 to 2 hours and is not approved by the Food and Drug Administration (FDA). While culture expansion of bone marrow generates more MSC than centrifugation,¹² it is a highly time- and cost-intensive procedure, requires GMP (Good Manufacturing Practice) facilities, and is not yet approved by the FDA. In contrast, a point-of-care concentration of BMA may be a reasonable alternative for the clinical practice and is permitted by the FDA due to minimal manipulation of the cells. Many commercial systems from several manufacturers are available; however, it is challenging to select the most appropriate system.

Because of the lack of clinical trials and outcome data after BMC treatments, in this review we compared published data of 8 different point-of-care devices that concentrate BM with the focus on technical features and quality of the harvested product. The primary objective was to provide surgeons with a single resource to enable selection of the appropriate device in clinical practice.

Methods

An initial search was conducted from February until July 2016 using existing brochures and a Google search to find companies that manufactured point-of-care devices to concentrate bone marrow. All companies were contacted by email or phone to get detailed information about their devices. Devices and companies were excluded if they were not FDA approved, not approved for intra-articular injection, not considered point-of-care devices, did not provide information, or did not provide information in English. The following devices were included in our comparison: the MAGELLAN MAR0Max device from Arteriocyte Medical Systems, the Arthrex Angel System Bone Marrow Concentration, the ART BMC device from Celling Biosciences, the PureBMC device from EmCyte Corporation,

the Accelerate: Autologous Platelet Concentrating System from Exactech, the CellPoint Concentrated Bone Marrow Aspirate System from ISTO Technologies, the BMAC 2 device from Harvest Technologies/Terumo BCT, and the BioCUE Blood and Bone Marrow Aspirate Concentration System from Zimmer/Biomet. We then performed a literature search in PubMed to find publications in which the devices were used or in which different devices were compared. The search strategy included (BMAC or bone marrow aspirate concentrate), (BMC or bone marrow concentrate) as well as only the name and the company of the different devices. Finally, based on the information obtained from our searches, all devices were compared for technical features as follows: centrifugation speed and time, input and output volumes, kit components, type of syringe used to aspirate, and usage of a filter before processing the bone marrow. We also assessed each device on the quality of their BMC product as follows: hematocrit (HCT), concentration of platelets and total nucleated cells (TNC), concentration of MSC, and colony forming unit fibroblasts/connective tissue progenitor cells (CFU-F/CTP). We excluded manufacturer specifications and publications that were not peer-reviewed data. Because of our focus on the treatment of orthopedic conditions (especially osteoarthritis and focal cartilage lesions), we did not compare the presence of hematopoietic stem cells (HSCs). "Enrichment" is the relative increase of cell concentration in BMC relative to BMA using commercially available cell counters. "Percentage of recovery" was calculated by dividing the total number of cells recovered in the BMC by the total number present in the volume of BMA processed. CFU assays are commonly used to quantify the number of progenitor or MSC, in which bone marrow cells are seeded and cultured for a certain period of time, and the resulting cell colonies are counted. Information regarding CFU could be found for almost all systems, although the results were reported in different units, which precluded direct comparisons between devices.

Finally, all information was summarized. **Tables 1** and **2** list the complete detailed comparison of all devices, including the sources from where this information was gathered. **Table 1** compares the technical features of the devices and **Table 2** lists the independent peer-reviewed data regarding the quality criteria of the BMC. The asterisk (*) symbol identifies the sources from which the information was gathered. The absence of any symbol indicates that the source of the information was from manufacturers' specifications. This review was independent and was not funded by industry.

Clinical Use of BMC

There have been several encouraging reports of the use of autologous BMC as intra-articular injections for the treatment of OA and as implants for the treatment of full-thickness

Table 1. Detailed Comparison of Different Point-of-Care Devices to Concentrate Bone Marrow (Technical Aspects).

Company	Product Name	Centrifugation Time (Minutes)	Centrifugation Speed (rpm)	Input Volume (mL BMA)	Output Volume (mL BMC)	Filter	BMA Needle Included	Syringe Included in the Kit to Aspirate Bone Marrow
Arteriocyte	MAGELLAN® MAR0Max	12-17 (depends on the input volume)	Dual spin protocol (~8 min at 2800 rpm and ~8 min at 3800 rpm)	30-60 (adjustable; different kits available)	adjustable, 3-10	Yes (200-µm filter)	Yes (different ones, kit depending)	30 mL VacLok™ Syringes (MAR0Max™ Kit)
Arthrex	Arthrex Angel™ System Bone Marrow Concentration	15-26 (depends on the input volume)	Depends on default volume to process (3000-4000 rpm)	40-180 (adjustable, universal kit)	Adjustable (automatically, depending on input volume)	Yes	Yes	30 mL VacLok™ Syringes (ABS-10062 kit)
Celling Biosciences	ART BMC	15	N/A	60***	3.5-4.0	Yes (150-µm filter)	No	10, 30, 60 mL back-lock syringes
EmCyme Corporation	PureBMC®	7.5***	Double spin protocol (2.5 min and 5 min at 3800 rpm)***	30/60 (different kits); 75 mL***	3-4/7 (kit depending); 7.5 mL***	Yes	Yes	VacLok™ Syringes
Exactech, Inc.	Accelerate®: Autologous Platelet Concentrating System	12 (Drucker centrifuge)/10 (Eppendorf centrifuge)	2400 rpm (Drucker centrifuge)/3600 rpm (Eppendorf centrifuge)	60 mL	6 mL (2 mL of plasma + 4 mL of buffy coat)	Yes	Yes (5 holes, 11 G)	1 x 60 mL VacLok™ Syringe + 1 x 60 mL standard syringe (Bone marrow harvesting kit)
Harvest Technologies Corp./Terumo BCT, Inc.	BMAC® 2	12	Double spin protocol (4 min at 1000 x G and 8 min at 900 x G)	30/60/120/180/240 (different kits)	3-4/7-10/14-20/21-30/28-40 (kit depending)	Yes (200-µm filter)	5 port needle (60+120 mL kits), 15G BMA needle with depth stop (30 mL kit)	30 mL and 60 mL back-lock syringes
ISTO Technologies, Inc.	CellPoint™ Concentrated Bone Marrow Aspirate System	<20	N/A	30-220 (adjustable, universal kit)	7-20 (adjustable)	N/A	N/A	N/A
Zimmer/Biomet	BioCUE™ Blood and Bone Marrow Aspirate Concentration System	15	3200 rpm	30 (mini kit)/60 (standard kit)	3/6 (kit depending)	No (dual buoy design eliminates need for filter)	Yes (6 holes)	30 mL and 60 mL standard syringes (kit depending)

N/A = not available, BMA = bone marrow aspirate, BMC = bone marrow concentrate.

No symbol indicates the manufacturer specifications.

*Data from: Hegde et al. A Prospective Comparison of 3 Approved Systems for Autologous Bone Marrow Concentration Demonstrated Nonequivalency in Progenitor Cell Number and Concentration, 2014.

**Data from: Karli et al. Autologous Regenerative Therapies: Rapid Concentration of Progenitor Cells, Platelets, and Proteins at the Point-of-Care (peer reviewed, poster presented on TERMIS annual meeting, September 2015, Boston, MA).

***Data from: Comparison Data Independent Evaluation; Prepared by R. J. Mandle, PhD; Biosciences Research Associates, Cambridge, MA; Study Date March 6, 2014.

****Data from: Analysis of Accelerate System: Human Bone Marrow Concentration; Prepared by R. J. Mandle, PhD; Research Associates, Cambridge, MA; Study Date August 2006.

*****Data from: Hermann et al. Concentration of Bone Marrow Total Nucleated Cells by a Point-of-Care Device Provides a High Yield and Preserves Their Functional Activity. Cell Transplantation, Vol. 16, pp. 1059-1069, 2008.

Table 2. Detailed Comparison of Different Point-of-Care Devices to Concentrate Bone Marrow (BMC Quality Criteria).

Company	Product Name	TNC (% Recovery or Fold of Enrichment)	Platelets (% Recovery or Fold of Enrichment)	CFU-F/CTP (% Recovery or Fold of Enrichment)	MSC (Fold of Enrichment)
Arteriocyte	MAGELLAN® MAROMax	32.6% ± 13.7%*, 3.3× enrichment****	4.8× enrichment****	39.4% ± 22.2% (1.4-2.3 CTP/cm ²)*	No significant changes in CD markers****
Arthrex	Arthrex Angel™ System Bone Marrow Concentration	N/A	N/A	N/A	N/A
Celling Biosciences	ART BMC	80%, 6× enrichment**	N/A	7.7× enrichment**	N/A
EmCyte Corporation	PureBMC®	N/A	N/A	N/A	N/A
Exactech, Inc.	Accelerate®: Autologous Platelet Concentrating System	N/A	N/A	N/A	N/A
Harvest Technologies Corp./ Terumo BCT, Inc.	BMAC® 2	66.5% ± 19.5% and 65.1% ± 19.3%*, 61.8% ± 10.4%***, 4.1× enrichment****	3.5× enrichment****	57.2% ± 25.6% and 56.1% ± 28.0% (1.4-1.8 CTP/cm ² and 1.7-2.2 CTP/cm ²)*	N/A
ISTO Technologies, Inc.	CellPoint™ Concentrated Bone Marrow Aspirate System	N/A	N/A	N/A	N/A
Zimmer/BIOOMET	BioCUE™ Blood and Bone Marrow Aspirate Concentration System	77.5% (62.3% ± 28.3%)*	N/A	44.5% ± 37.0% (0.4-0.6 CTP/cm ²)*	N/A

N/A = no peer reviewed data available, BMA = bone marrow aspirate, BMC = bone marrow concentrate, HCT = hematocrit, TNC = total nucleated cells, CFU-F = colony forming unit fibroblasts, CTP = connective tissue progenitors, MSC = mesenchymal stem cells, CD = cluster of differentiation. No symbol indicates the manufacturer specifications.

*Data from: Hegde *et al.* A Prospective Comparison of 3 Approved Systems for Autologous Bone Marrow Concentration Demonstrated Nonequivalency in Progenitor Cell Number and Concentration, 2014.

**Data from: Karli *et al.* Autologous Regenerative Therapies: Rapid Concentration of Progenitor Cells, Platelets, and Proteins at the Point-of-Care (peer reviewed, poster presented on TERMIS annual meeting, September 2015, Boston, MA).

***Data from: Hermann *et al.* Concentration of Bone Marrow Total Nucleated Cells by a Point-of-Care Device Provides a High Yield and Preserves Their Functional Activity, Cell Transplantation, Vol. 16, pp. 1059-1069, 2008.

****Data from: Cassano *et al.* Bone marrow concentrate and platelet-rich plasma differ in cell distribution and interleukin 1 receptor antagonist protein concentration, 2016.

*****Data from: Zhong *et al.* In vivo comparison of the bone regeneration capability of human bone marrow concentrates vs. platelet-rich plasma, 2012.

chondral defects.^{3,13-20} Even though long-term results are pending, a recent prospective randomized study showed no significant differences in short-term pain relief between injection of BMC and saline in knees of patients with OA.²¹ Nevertheless, BMC is extensively used for orthopedic applications by surgeons worldwide and there are diverse commercial systems from several manufacturers that incorporate a vast number of technical features. Given the lack of high-quality studies and definitive evidence regarding clinical efficacy, the important parameters that are likely to be relevant to the clinician are safety, convenience, time involved, and reproducibility. It is not yet known if differences in the technical features of the different devices have an impact on the quality of patient care and clinical outcomes. We conducted

this review to assist the surgeon in making informed comparisons when choosing among the different systems.

Comparison of Technical Features

The harvesting systems can be broadly classified into 2 groups. The first includes the computer-controlled, fully automated, and closed-loop systems with an integrated centrifuge, which are used by the devices from Arteriocyte, Arthrex, and ISTO Technologies. An important feature of these devices is the optical sensor to detect the different wavelengths corresponding to the various blood fractions to separate the buffy coat after centrifuging. In a fully automated system, after connecting the BMA syringe to the

device and after selecting appropriate parameters in the computer software, the system will deliver the BMC ready to use in a syringe without any syringe switching or further operations required by the user. In contrast, the second group of systems uses a manual extraction of the buffy coat. Therefore, the systems of Celling Biosciences, EmCyte, Exactech, Harvest Technologies/Terumo BMC, and Zimmer/BIOMET provide different closed sterile processing disposables in which the BMA is injected. This disposable has to be placed in an external centrifuge to separate the various blood fractions. Next, the device has to be taken out of the centrifuge, a syringe has to be connected, and the buffy coat has to be extracted manually according to respective manufacturer's specifications.

The advantages of a closed-loop system are greater convenience and possibly less risk of infection or contamination, although no publications could be found to prove scientific evidence of likelihood of these risks. The advantage of an integrated centrifuge is convenience and reduced equipment in the procedure room. Manual extraction gives the surgeon greater control over the harvested material, but with the disadvantage of variation due to differences in extracted material.

Input/Output Volumes

One important criterion for selection is the volume of bone marrow the surgeon is able to harvest and the quantity of BMC needed for the treatment. As depicted in **Table 1**, the input volumes of BMA for the different devices range between 30 and 240 mL and the output volumes of BMC between 3 and 40 mL. For clinical applications requiring variable volumes of BMC, the fully automated devices possess certain advantages. The systems of Arteriocyte and ISTO Technologies allow selection of the final BMC volume, whereas the Arthrex device automatically adjusts the output volume depending on the volume of BMA input. Furthermore, the fully automated systems provide universal disposable kits, which allow processing different volumes (30–60 mL with the Arteriocyte device, 40–180 mL with the Arthrex device, and 30–220 mL with the ISTO Technologies device). All other companies provide individual kits, which allow only processing and delivery of predefined volumes of BMA and BMC, respectively.

Centrifugation Parameters

In **Table 1**, the centrifugation parameters of the different devices are shown. The centrifugation times range between 7.5 and 20 minutes. The centrifugation speed ranges between 2400 and 4000 rpm (Harvest Technologies: 900 and $1000 \times G$). Three systems (Arthrex, EmCyte, and Harvest Technologies) use a double spin protocol in which

a first centrifugation cycle separates the red blood cell fraction followed by a second cycle to separate the buffy coat. Except Celling Biosciences and ISTO Technologies, which did not provide information regarding the centrifugation speed, all other devices employ a single centrifugation cycle. We only found one study that compared different centrifugation protocols to separate BMC.²² In that study, single and double centrifugation of BMA for sinus floor elevation were compared but no statistically significant differences were found. Therefore, no evidence-based superiority of a specific device using a specific protocol to concentrate bone marrow can be determined at this time.

Kit Content

The quality of the final BMC likely depends on the quality of the bone marrow and the aspiration technique. Fast and continuous aspiration of bone marrow with a high differential pressure, using smaller volume syringes, results in a better quality specimen and in more stem cells.^{23,24} Furthermore, multiple advancements of the bone marrow needle and multiple site aspiration increases initial nucleated cell counts, CFU-Fs, and generates more MSC at the first passage.²⁵ Therefore, so-called back-lock syringes, which can be locked at a certain volume, are able to hold a continuous high differential pressure while aspirating and could be beneficial to improve BMA quality. The devices of Celling Biosciences, Arteriocyte, Arthrex, Harvest Technologies, Exactech, and Emcyte have these syringes included in their kits, which could be an advantage compared with systems that provide standard syringes.

Six devices (Arteriocyte, Arthrex, Celling Biosciences, Emcyte, Exactech, and Harvest Technologies) require the use of special filters after harvesting and before processing the bone marrow. This filter serves to remove bone marrow fat cells, clots, and small bone fragments from the bone marrow before centrifugation.²⁶ Hegde *et al.*²⁷ compared 3 different BMC devices (Arteriocyte, BIOMET and Harvest Technologies), and reported that the BIOMET system, which uses a dual buoy feature instead of a filter, resulted in lower CTP counts.²⁷ ISTO Technologies did not provide information about the type of syringe included in their kit, or if a filter is used while processing.

Heparin Coating

Another aspect that could affect BMC quality is the heparin coating of the syringe, which is recommended in protocols of some devices (Arteriocyte, Arthrex, Exactech, and Harvest Technologies). Heparin could reduce the binding of CTPs to the surface of the syringes or transfer bags used with the different systems, which could lead in a smaller loss of CTPs during processing.^{27–29}

Comparison of Different Blood Components and Quality Criteria

For orthopedic treatments, especially for the treatment of OA and focal cartilage defects, the primary possible advantage of BMC, besides the presence of platelets and their released growth factors, is the content of tissue progenitor cells, which could have the ability to form new cartilage. We selected the following surrogate criteria to compare the quality of the BMC: (1) the HCT; (2) the platelet count, because they release the important growth factors; (3) the TNC, which include MSC but, as mentioned above, only in a very small number (most of these cells are leukocytes); (4) the number of CFU-F or CTP, which represent progenitor cells; and (5) the concentration of MSC.

Hematocrit

The amount of residual red blood cells in the final BMC is a measure of the quality of cell separation. A recent *in vivo* study analyzing PRP (extracted from whole blood and injected into human knee joints) showed that a high HCT resulted in more cell death of human synoviocytes and production of pro-inflammatory cytokines.³⁰ Thus, it is desirable to have low HCT in the BMC, especially if it is used for the treatment of osteoarthritis or focal cartilage defects. Unfortunately, no independent peer-reviewed data regarding the HCT in the final BMC of the different devices could be found.

Enrichment and Recovery of Platelets and TNC

Table 2 displays the information regarding the recovery and enrichment of platelets and TNC in the BMC delivered by these systems. Direct comparisons were not possible because analyses were done on different volumes of BMA and BMC, using different cell counting machines, and using different counting protocols.

Only one study could be found that compared different devices using the same cell counter.¹¹ In that study, Cassano *et al.*¹¹ compared the devices of Arteriocyte and Harvest Technologies. The authors showed that the Arteriocyte device generated a significantly greater platelet concentration and that there were no significant differences in the TNC concentration between both devices. However, more independent research comparing devices under controlled conditions are needed.

Amount of Colony Forming Unit Fibroblasts/Connective Tissue Progenitor Cells

As a measure of the concentration of MSC in the final BMC, CFU-F/CTPs assay results were analyzed. In a CFU-assay, bone marrow is seeded into culture and the plastic-adherent

colonies are counted after incubation for several days or weeks. **Table 2** shows the information that could be gathered regarding the concentration of MSC and number of CFU-F/CTP in the BMC created by the different devices. Unfortunately, significant differences in the protocols such as the definition of a colony, incubation time, and reported units (CFU/mL, CTP/cm², CTP/cm³, percentage of enrichment), precluded direct comparisons.

Only 2 independent research studies could be found that compared different devices using the same CFU-assay protocol.^{11,27} Hegde *et al.*²⁷ showed that the Harvest Technologies device resulted in a higher CTP number and concentration in comparison to the systems of Arteriocyte and BIOMET. Cassano *et al.* compared the systems of Harvest Technologies and Arteriocyte and found that the increase of CFU was not significantly different between the devices. Cassano *et al.*¹¹ concluded that the varying results in these 2 studies are likely explained due to different study methodologies. Hegde *et al.*²⁷ used a defined number of TNC for culture whereas Cassano *et al.*¹¹ used a uniform volume of BMC, which could be the more practical approach in clinical practice.

Amount of Mesenchymal Stem Cells

The CFU assay is a commonly used method, which is reproducible and relatively inexpensive. However, CFU assays may not be the most quantitative measure of the presence of bone marrow MSC. The International Society for Cellular Therapy proposed minimal criteria for identification of MSC: MSC must be plastic-adherent under culture conditions, must express CD105, CD73, and CD90 while lacking expression of CD45, CD34, CD14 or CD11b, CD79 α , or CD19 and HLA-DR surface molecules, and must differentiate into different cell types such as osteoblasts, adipocytes, and chondroblasts *in vitro*.³¹ Studies involving the BIOMET device³² and the Arteriocyte device³³ reported no significant increases in the MSC CD-marker expression after centrifugation. Although flow cytometry is a widely used method to identify MSC, surface markers may not be the most relevant measure of the clinical potency of progenitor cells or their osteogenic, chondrogenic, or antiarthritic capacity. No specific *in vivo* marker exists to identify MSC without involving isolation and cell culture. A possible useful intraoperative method to quantify BM quality with respect to MSC content was reported recently.³⁴ Cuthbert *et al.*³⁴ could show that this rapid procedure (less than 40 minutes) by quantifying CD45^{-low} CD271^{bright} cells can reliably predict CFU yields in BMA. However, CFU may not correlate with clinical outcome and this new technique still needs to be validated. Therefore, more studies are needed for validating these flow cytometry results with clinical efficacy or for identifying more clinically relevant markers.

Limitation of the Review

One weakness of this study was the limited availability of independent peer-reviewed publications. This is even more important because all devices reviewed are FDA approved and commercially available for clinical use. Nevertheless, this comparison summarizes the most relevant information in a single source to assist the surgeon in making more informed decisions when choosing among different devices. Cost is also important because intra-articular injections of BMC are not covered by medical insurance. However, each manufacturer negotiates a different price with the health care provider, which is why we could not provide any prices or cost of the compared devices or their kits. We are planning a more controlled comparison of the performance of different BMC devices followed by assessment of clinical efficacy.

Conclusion

Based on the data obtained in this review, recommending a single harvesting system is not possible. In our opinion, the fully automated systems have some technical advantages (among others, universal kits including small volume back-lock syringes and the use of a filter) and might be more convenient. In terms of “chondrogenic or anti-arthritis potency” of the BMC, we cannot recommend a single system because the reported data could not be compared between devices. A standardized reporting method is needed for valid comparison among devices. Furthermore, clinical outcomes are required to establish the true efficacy of these systems. We are conducting additional studies for more careful comparison among the devices.

Acknowledgments and Funding

The author(s) received no financial support for the research, authorship, and/or publication of this article.

Declaration of Conflicting Interests

The author(s) declared the following potential conflicts of interest with respect to the research, authorship, and/or publication of this article: WDB consulted for Arthrex for unrelated product development.

References

- Kristjansson B, Honsawek S. Current perspectives in mesenchymal stem cell therapies for osteoarthritis. *Stem Cells Int*. 2014;2014:194318.
- Steinert AF, Rackwitz L, Gilbert F, Noth U, Tuan RS. Concise review: the clinical application of mesenchymal stem cells for musculoskeletal regeneration: current status and perspectives. *Stem Cells Transl Med*. 2012;1:237-47.
- Chahla J, Dean CS, Moatshe G, Pascual-Garrido C, Serra Cruz R, LaPrade RF. Concentrated bone marrow aspirate for the treatment of chondral injuries and osteoarthritis of the knee: a systematic review of outcomes. *Orthop J Sports Med*. 2016;4:2325967115625481.
- Kamel NS, Arafa MM, Nadim A, Amer H, Amin IR, Samir N, *et al*. Effect of intra-articular injection of mesenchymal stem cells in cartilage repair in experimental animals. *Egypt Rheumatologist*. 2014;36:179-86.
- Kehoe O, Cartwright A, Askari A, El Haj AJ, Middleton J. Intra-articular injection of mesenchymal stem cells leads to reduced inflammation and cartilage damage in murine antigen-induced arthritis. *J Transl Med*. 2014;12:157.
- Song F, Tang J, Geng R, Hu H, Zhu C, Cui W, *et al*. Comparison of the efficacy of bone marrow mononuclear cells and bone mesenchymal stem cells in the treatment of osteoarthritis in a sheep model. *Int J Clin Exp Pathol*. 2014;7:1415-26.
- Emadedin M, Ghorbani Liastani M, Fazeli R, Mohseni F, Moghadasali R, Mardpour S, *et al*. Long-term follow-up of intra-articular injection of autologous mesenchymal stem cells in patients with knee, ankle, or hip osteoarthritis. *Arch Iran Med*. 2015;18:336-44.
- Vega A, Martin-Ferrero MA, Del Canto F, Alberca M, García V, Munar A, *et al*. Treatment of knee osteoarthritis with allogeneic bone marrow mesenchymal stem cells: a randomized controlled trial. *Transplantation*. 2015;99:1681-90.
- Singh A, Gangwar DS, Singh S. Bone marrow injection: a novel treatment for tennis elbow. *J Nat Sci Biol Med*. 2014;5:389-91.
- Hendrich C, Franz E, Waertel G, Krebs R, Jager M. Safety of autologous bone marrow aspiration concentrate transplantation: initial experiences in 101 patients. *Orthop Rev (Pavia)*. 2009;1:e32.
- Cassano JM, Kennedy JG, Ross KA, Fraser EJ, Goodale MB, Fortier LA. Bone marrow concentrate and platelet-rich plasma differ in cell distribution and interleukin 1 receptor antagonist protein concentration. *Knee Surg Sports Traumatol Arthrosc*. 2018;26(1):333-42.
- Veronesi F, Giavaresi G, Tschon M, Borsari V, Nicoli Aldini N, Fini M. Clinical use of bone marrow, bone marrow concentrate, and expanded bone marrow mesenchymal stem cells in cartilage disease. *Stem Cells Dev*. 2013;22:181-92.
- Centeno CJ, Al-Sayegh H, Bashir J, Goodyear S, Freeman MD. A dose response analysis of a specific bone marrow concentrate treatment protocol for knee osteoarthritis. *BMC Musculoskelet Disord*. 2015;16:258.
- Centeno C, Pitts J, Al-Sayegh H, Freeman M. Efficacy of autologous bone marrow concentrate for knee osteoarthritis with and without adipose graft. *Biomed Res Int*. 2014;2014:370621.
- Centeno CJ, Al-Sayegh H, Bashir J, Goodyear S, Freeman MD. A prospective multi-site registry study of a specific protocol of autologous bone marrow concentrate for the treatment of shoulder rotator cuff tears and osteoarthritis. *J Pain Res*. 2015;8:269-76.
- Kim JD, Lee GW, Jung GH, Kim CK, Kim T, Park JH, *et al*. Clinical outcome of autologous bone marrow aspirates concentrate (BMAC) injection in degenerative arthritis of the knee. *Eur J Orthop Surg Traumatol*. 2014;24:1505-11.
- Gobbi A, Karnatzikos G, Scotti C, Mahajan V, Mazzucco L, Grigolo B. One-step cartilage repair with bone marrow

- aspirate concentrated cells and collagen matrix in full-thickness knee cartilage lesions: results at 2-year follow-up. *Cartilage*. 2011;2:286-99.
18. Gobbi A, Scotti C, Karnatzikos G, Mudhigere A, Castro M, Peretti GM. One-step surgery with multipotent stem cells and Hyaluronan-based scaffold for the treatment of full-thickness chondral defects of the knee in patients older than 45 years. *Knee Surg Sports Traumatol Arthrosc*. 2017;25:2494-501.
 19. Gobbi A, Whyte GP. One-stage cartilage repair using a hyaluronic acid-based scaffold with activated bone marrow-derived mesenchymal stem cells compared with microfracture: five-year follow-up. *Am J Sports Med*. 2016;44:2846-54.
 20. Cotter EJ, Wang KC, Yanke AB, Chubinskaya S. Bone marrow aspirate concentrate for cartilage defects of the knee: from bench to bedside evidence. *Cartilage*. Epub 2017 Nov 1.
 21. Shapiro SA, Kazmerchak SE, Heckman MG, Zubair AC, O'Connor MI. A prospective, single-blind, placebo-controlled trial of bone marrow aspirate concentrate for knee osteoarthritis. *Am J Sports Med*. 2017;45:82-90.
 22. de Oliveira TA, Aloise AC, Orosz JE, de Mello E, Oliveria R, de Carvalho P, Pelegrine AA. Double centrifugation versus single centrifugation of bone marrow aspirate concentrate in sinus floor elevation: a pilot study. *Int J Oral Maxillofac Implants*. 2016;31:216-22.
 23. Hernigou P, Homma Y, Flouzat Lachaniette CH, Poignard A, Allain J, Chevallier N, *et al*. Benefits of small volume and small syringe for bone marrow aspirations of mesenchymal stem cells. *Int Orthop*. 2013;37:2279-87.
 24. Gronkjaer M, Hasselgren CF, Ostergaard AS, Johansen P, Korup J, Bøgsted M, *et al*. Bone marrow aspiration: a randomized controlled trial assessing the quality of bone marrow specimens using slow and rapid aspiration techniques and evaluating pain intensity. *Acta Haematol*. 2016;135:81-7.
 25. Peters AE, Watts AE. Biopsy needle advancement during bone marrow aspiration increases mesenchymal stem cell concentration. *Front Vet Sci*. 2016;3:23.
 26. Hernigou P, Poignard A, Beaujean F, Rouard H. Percutaneous autologous bone-marrow grafting for nonunions. Influence of the number and concentration of progenitor cells. *J Bone Joint Surg Am*. 2005;87:1430-7.
 27. Hegde V, Shonuga O, Ellis S, Fragomen A, Kennedy J, Kudryashov V, *et al*. A prospective comparison of 3 approved systems for autologous bone marrow concentration demonstrated nonequivalency in progenitor cell number and concentration. *J Orthop Trauma*. 2014;28:591-8.
 28. Fannon M, Forsten KE, Nugent MA. Potentiation and inhibition of bFGF binding by heparin: a model for regulation of cellular response. *Biochemistry*. 2000;39:1434-45.
 29. Newman DR, Li CM, Simmons R, Khosla J, Sannes PL. Heparin affects signaling pathways stimulated by fibroblast growth factor-1 and -2 in type II cells. *Am J Physiol Lung Cell Mol Physiol*. 2004;287:L191-200.
 30. Braun HJ, Kim HJ, Chu CR, Dragoo JL. The effect of platelet-rich plasma formulations and blood products on human synoviocytes: implications for intra-articular injury and therapy. *Am J Sports Med*. 2014;42:1204-10.
 31. Dominici M, Le Blanc K, Mueller I, Slaper-Cortenbach I, Marini F, Krause D, *et al*. Minimal criteria for defining multipotent mesenchymal stromal cells. The International Society for Cellular Therapy position statement. *Cytotherapy*. 2006;8:315-7.
 32. Woodell-May JE, Tan ML, King WJ, Swift MJ, Welch ZR, Murphy MP, *et al*. Characterization of the cellular output of a point-of-care device and the implications for addressing critical limb ischemia. *Biores Open Access*. 2015;4:417-24.
 33. Zhong W, Sumita Y, Ohba S, Kawasaki T, Nagai K, Ma G, *et al*. In vivo comparison of the bone regeneration capability of human bone marrow concentrates vs. platelet-rich plasma. *PLoS One*. 2012;7:e40833.
 34. Cuthbert R, Boxall SA, Tan HB, Giannoudis PV, McGonagle D, Jones E. Single-platform quality control assay to quantify multipotential stromal cells in bone marrow aspirates prior to bulk manufacture or direct therapeutic use. *Cytotherapy*. 2012;14:431-40.