Published in final edited form as:

Eur Radiol. 2019 April ; 29(4): 1665–1673. doi:10.1007/s00330-018-5703-9.

Initial Clinical Evaluation of Stationary Digital Chest Tomosynthesis in Adult Patients with Cystic Fibrosis

Elias Taylor Gunnell, MD¹, Dora K. Franceschi, MD¹, Christina R. Inscoe, PhD², Allison Hartman, PhD², Jennifer L. Goralski, MD^{4,5,6}, Agathe Ceppe, MS⁴, Brian Handly, MD⁷, Cassandra Sams, MD⁷, Lynn Ansley Fordham, MD⁷, Jianping Lu, PhD², Otto Zhou, PhD², Yueh Z. Lee, MD/PhD^{2,3,4,7,8}

¹School of Medicine, The University of North Carolina at Chapel Hill, Chapel Hill, NC 27599

²Department of Physics and Astronomy, The University of North Carolina at Chapel Hill, Chapel Hill, NC 27599

³Lineberger Comprehensive Cancer Center, The University of North Carolina at Chapel Hill, Chapel Hill, NC 27599

⁴Marsico Lung Institute, The University of North Carolina at Chapel Hill, Chapel Hill, NC 27599

⁵Division of Pulmonary and Critical Care Medicine, The University of North Carolina at Chapel Hill, Chapel Hill, NC 27599

⁶Division of Pediatric Pulmonary, The University of North Carolina at Chapel Hill, Chapel Hill, NC 27599

⁷Department of Radiology, The University of North Carolina at Chapel Hill, Chapel Hill, NC 27599

⁸Biomedical Research Imaging Center, The University of North Carolina at Chapel Hill, Chapel Hill, NC 27599

Corresponding Author Contact information: 920-901-6132, egunnell@med.unc.edu, 207 Brookgreen Dr. Chapel Hill, NC 27516. We are requesting dual first authorship for E. Taylor Gunnell and Dora Franceschi. These students contributed equally to the design, implementation, analysis, and write up of this study.

Guarantor:

The scientific guarantor of this publication is Dr. Yueh Lee MD, PhD.

Conflict of interest:

The authors of this manuscript declare relationships with the following companies: Drs. Lee, Zhou and Lu are co-inventors of the stationary chest tomosynthesis imaging system evaluated in this study. Dr. Zhou has equity ownership and serves on the board of directors of Xintek, Inc., to which the technologies used or evaluated in this article have been or will be licensed. Dr. Lu has equity ownership in Xintek, Inc. All of these relationships are under management by the University of North Carolina's COI committees.

Statistics and biometry:

Mrs. Agathe Ceppe, MS kindly provided statistical advice for this manuscript.

Marsico Lung Institute, The University of North Carolina at Chapel Hill, Chapel Hill, NC 27599

Informed consent:

Written informed consent was obtained from all subjects (patients) in this study.

Ethical approval:

Institutional Review Board approval was obtained.

Methodology:

- experimental
- performed at one institution

Abstract

Objective: The imaging evaluation of cystic fibrosis currently relies on chest radiography or computed tomography. Recently, digital chest tomosynthesis has been proposed as an alternative. We have developed a stationary digital chest tomosynthesis [s-DCT] system based on a carbon nanotube [CNT] linear x-ray source array. This system enables tomographic imaging without movement of the x-ray tube and allows for physiological gating. The goal of this study was to evaluate the feasibility of clinical CF imaging with the s-DCT system.

Materials and Methods: CF patients undergoing clinically indicated chest radiography were recruited for the study and imaged on the s-DCT system. Three board-certified radiologists reviewed both the CXR and s-DCT images for image quality relevant to CF. CF disease severity was assessed by Brasfield score on CXR and chest tomosynthesis score on s-DCT. Disease severity measures were also evaluated against subject pulmonary function tests.

Results: Fourteen patients underwent s-DCT imaging within 72 hours of their chest radiograph imaging. Readers scored the visualization of proximal bronchi, small airways and vascular pattern higher on s-DCT than CXR. Correlation between the averaged Brasfield score and averaged tomosynthesis disease severity score for CF was -0.73 , $p=.0033$. The CF disease severity score system for tomosynthesis had high correlation with FEV1 [$r = -0.685$] and, FEF 25–75% [$r = -0.719$], as well as good correlation with FVC [$r = -0.582$].

Conclusion: We demonstrate the potential of CNT x-ray based s-DCT for use in the evaluation of cystic fibrosis disease status in the first clinical study of s-DCT.

Keywords

Tomography; Nanotubes, Carbon; X-rays; Cystic Fibrosis; Scoring Methods

INTRODUCTION

Cystic fibrosis [CF] is one of the most common life-limiting genetic disorders within the Caucasian population. Progressive lung disease is the main cause of morbidity and mortality in patients with CF, with damage to the lung parenchyma caused by repeated infections and chronic inflammation. Therefore, early diagnosis of infection is important to enable early intervention and prevent irreversible lung damage.[1] Though computed tomography [CT] offers the highest resolution 3-D imaging of the lungs, relatively high radiation doses limit repeated measures. Chest radiographs are generally used to monitor disease progression but can be relatively insensitive to detection of early disease.[2] Radiation dose considerations are increasingly important, especially since patients born with CF today now have a life expectancy of 70 years.[3] New short echo-time techniques paired with hyperpolarized gas inhalation are improving the potential utility of magnetic resonance imaging [MRI], however, high cost and limited availability of these advanced techniques restricts their utility. [4]

Digital chest tomosynthesis [DCT] is an imaging technique in which a 3D image of the chest is reconstructed using x-ray projection images acquired over a limited scanning angle, typically 30–60 degrees.[5,6] The use of DCT in thoracic imaging has advantages over both

conventional chest radiography and computed tomography. Digital chest tomosynthesis generates coronal slices through the thorax with a spatial resolution superior to coronal CT images.[5,6] The utility of DCT has been explored as an alternative imaging technique for lung cancer screening and the evaluation of cystic fibrosis.[7,8] Studies by Steyern and colleagues have demonstrated the potential value of DCT as an alternative to CT.[9–11]

We have developed a stationary digital chest tomosynthesis [s-DCT] system based on a carbon nanotube [CNT] field emission x-ray source technology. In contrast to conventional x-ray tubes, our x-ray source relies on field emission to generate electrons at room temperature. This technology allows two important advantages over a conventional x-ray source. First, multiple x-ray sources can be packed in close proximity, enabling the creation of an array of x-ray focal spots. Second, the field emission source allows generation of precise x-ray pulses from the individual sources on demand. Thus, a linear x-ray imaging array may be substituted for the conventional moving x-ray source in DCT to provide x-ray projections from multiple directions. In turn, this can shorten imaging times and reduce blur caused by gantry motion. The ability of the system to generate x-ray pulses on demand enables the potential of respiratory gating.[12]

The goal of this study was to evaluate the potential of CNT linear x-ray source arrays to perform CF imaging. In the first-in-human prospective imaging study we compared the s-DCT system against conventional radiographs in the evaluation of CF disease status. Our secondary aim was to explore the correlation between s-DCT findings and pulmonary function tests [PFTs].

MATERIALS AND METHODS

This study was approved by the UNC Institutional Review Board. Informed consent was obtained from each patient before enrollment.

Subject Characteristics

CF subjects were enrolled through collaboration with our institutions Cystic Fibrosis and Pulmonary Diseases Research and Treatment Center. Adult patients with a history of CF who were undergoing clinically indicated conventional chest radiographs and PFTs were recruited for the study. Exclusion criteria included [1] pregnancy, [2] age less than 18 years old, [3] body mass index greater than 33, and [4] patients with unstable lung disease. Patients were imaged with the s-DCT system within 3 days of the clinical chest radiograph.

s-DCT and Chest Radiography Imaging Systems

The s-DCT is based on a repurposed linear CNT x-ray source array combined with a high speed digital x-ray detector. [Figure 1] The linear x-ray array consists of 29 cathodes spaced an average of 10mm apart. The anode consists of a tungsten target. There is 2.5mm of aluminum filtration. The imaging setup uses a source to detector distance of 130 cm to achieve a 12-degree angular coverage. All twenty-nine projections are acquired over a six second scan time with subjects in a supine [AP] position, during a full inspiratory breath hold. The tube voltage is 80 kVp and can produce an x-ray pulse width up to 130ms. The tube voltage limitation is a function of this particular x-ray tube which was repurposed from

other, non-medical applications. The measured focal spot size is 2.5 mm by 0.5 mm.[13,14] The flat panel silicon detector is 35 cm by 43 cm with a 139 mm pixel size and acquisition time of 5 fps. The system resolution was 1.7 cycles/mm in scan direction and 3.4 cycles/mm in the perpendicular direction. The entrance dose for adult patients was 0.6 mGy.[14] Conventional chest radiography was performed on a Fuji machine with a Siemens digital detector. Clinically acquired upright PA images used for the study were acquired with a typical source to detector distance of 182 cm and automatic exposure. Typical tube voltages used ranged from 117 – 125 kVp and 11–14 mAs.

s-DCT Image Reconstruction

The acquired projection data were reconstructed using Adapted Fan-beam Volume Reconstruction [AFVR] with custom code written in Matlab [Mathworks Inc]. The method was optimized for our linear source array by transforming the 3D cone-beam reconstruction into a series of 2D fan-beam volume reconstructions. The AFVR provides fast reconstruction, under 6 minutes, while utilizing a model-based iterative algorithm for improved image quality.[15] Slice width of 2.5mm was used. This generated 75 image slices per imaging stack with no gap.

s-DCT Reader Study and Qualitative Analysis

Three board-certified radiologists with no prior experience in interpreting chest tomosynthesis were trained in image interpretation via case study of images not included in this analysis. The radiologists had an average of 10 +/- 11.3 years' experience. Readers were trained with ten s-DCT cases from another concurrent study that included both CF related and non-CF related findings. For study images the readers independently rated the imaging technique, quality of the s-DCT images, and contemporaneous radiographs. Imaging technique was scored based on patient positioning, presence of motion blur, and quality of full inspiration effort. Image quality was subjectively quantified by rating the reproduction of the lung fields, trachea, proximal bronchi [including main stem bronchus and primary branches], small airways, diaphragm and lateral costophrenic angles [LCA] as well as vascular pattern. Each observer assigned a numerical score from 1 to 5, with 1 representing poor quality and 5 representing high quality.

The observers then scored radiographs using the Brasfield score[16] for disease severity and s-DCT images based on a CF disease severity score system for tomosynthesis.[17] This tomosynthesis scoring system evaluated the severity and extent of five common pathological findings in cystic fibrosis: overinflation, bronchial wall thickening, parenchymal lesions, bronchiectasis and mucus plugging. The overinflation score evaluated the lungs as a whole; all other components were scored by quadrant. The lungs were divided into quadrants at the level of the first division of the left main bronchus. The maximum scores were 4 for overinflation, 16 for bronchial wall thickening, 16 for parenchymal lesions, 32 for bronchiectasis and 32 for mucus plugging, with a maximum total score of 100, with higher score indicating more severe disease.[17]

The readers were blinded to patient demographics, clinical status and spirometry data. Image series were presented in random order; radiographs and s-DCT images of the same patient

were shown non-consecutively. Tomosynthesis sets were viewed by scrolling through the entire image stack.

Statistical Analysis

Statistical analysis was performed by a dedicated biostatistician using SAS/STAT [SAS] software. A reader factor was included in the model. Ordinal data [Image technique data] were analyzed with a repeated ordinal logistic model [PROC GENMOD]. The model takes into account the ordinal rank [1 to 5]. The Score method was applied to estimate p-values which extends the linear model to a multinomial model. The interaction of Reader and Technique was included in a model only when significant. The association between the Brasfield and Tomosynthesis scores, and correlation of scores with spirometry are reported as Pearson correlation coefficients. Spirometry data was not available for one patient. The intra-class correlation coefficients [ICCs], quantifying the reliability of the measurement for Brasfield and tomosynthesis scoring were calculated using a twoway random effects model, with absolute agreement and single rater. Comparison of ICCs was performed using the Konishi-Gupta modified Z-test.

RESULTS

A total of 14 patients with cystic fibrosis, aged 19 to 47, mean age 27.38 $\pm$ 10.02 years, were imaged using the s-DCT system [Figure 1] following clinically indicated chest radiography and pulmonary function tests [Table 1]. All subjects underwent s-DCT imaging within 24 hours of CXR, except one subject whose imaging was performed within 72 hours [see Video, Supplemental Material 1, which demonstrates complete s-DCT image sequence].

TECHNIQUE EVALUATION

Overall, readers gave significantly higher scores for CXR than s-DCT when evaluating technical parameters including full inspiration, motion blur and patient positioning. [Table 2] These scores were reader independent for all three parameters. In general readers gave a wider range of scores for s-DCT.

IMAGE QUALITY EVALUATION

Stationary tomosynthesis had significantly higher reader quality scores when evaluating proximal bronchi, small airways and vascular pattern. [Table 3] Readers tended to give higher image quality scores to CXR than s-DCT for lung fields, and diaphragm/LCA. The techniques were not significantly different when evaluating the trachea. Reader effects were not significant except when evaluating proximal bronchi, diaphragm and LCA. Regarding proximal bronchi score, the difference of scores between the methods is not similar across readers, which could explain the interaction of reader and technique. Reader 1 gave only scores of 5 [excellent] to all subjects while Reader 2 and 3 graded the images more severely. Both reader 2 and 3 gave significantly higher rates for s-DCT than CXR: Reader 2 gave a majority of '3' to CXR and a majority of '5' to s-DCT, while Reader 3 gave respectively '3' and '4'. When evaluating diaphragm and LCA, Reader 1 and 3 gave higher rates to CXR than s-DCT, while Reader 2 gave generally higher rates to s-DCT than CXR.

BRASFIELD SCORE AND TOMOSYNTHESIS SCORE

Correlation coefficient between the averaged Brasfield score and the averaged reader tomosynthesis disease severity score for CF was -0.73 , p-value of $.0033$. The negative correlation is due to the inverted direction of the tomosynthesis severity scores and Brasfield scores. Lower Brasfield scores correspond to more severe disease, while lower tomosynthesis scores are seen in less severe disease. [Figures 2 & 3]. Intraclass correlation coefficients for tomosynthesis and Brasfield scoring were 0.80 and 0.79 , respectively. These values were not significantly different, comparison yielded a p-value of 0.51 .

PULMONARY FUNCTION TESTS

The CF disease severity score system for s-DCT has significant correlation with almost all pulmonary function tests, including forced expiratory volume during the first second [FEV1] $[-0.685]$, forced expiratory flow at 25–75% of the pulmonary volume [FEF 25–75%] $[-0.719]$, with a more modest correlation with forced vital capacity [FVC] $[-0.582]$. [Table 4]

DISCUSSION

This study is the first evaluation of stationary tomosynthesis imaging in CF patients. Tomosynthesis generates image slices with high fidelity replication of in-plane structures, and has been shown to offer more accurate assessment of subtle lung lesions as well as changes in the airways and parenchyma.[8,18,19] Bronchial wall thickening, mucus plugging, and bronchiectasis are known to be better delineated with conventional tomosynthesis as compared to radiography.[17]

In our application, CNT s-DCT received significantly higher scores on image quality than conventional radiography in visualizing structures of interest in cystic fibrosis, such as the proximal bronchi, small airways, and vascular pattern. We found no difference in image quality between s-DCT and CXR when visualizing the trachea. These findings are similar to reports of conventional chest tomosynthesis and CF, including Choo et al., who reported the sensitivity of conventional DCT was higher than chest radiography in detecting airway lesions.[18] The diagnostic accuracy of DCT [90.91 %; 94.70 %] was also reported as significantly better than that of radiography [78.03 %; 82.58 %, all $p < 0.05$].[18] In the work by Galea and colleagues the sensitivity and specificity was 0.65 and 0.39 for CXR and 0.91 and 1 for DCT in detecting pulmonary and hilar lesions compared to chest x-ray.[20] These results suggest that a s-DCT system with the same geometry as a conventional tomosynthesis system would be capable of more accurately scoring disease severity than conventional radiography. There are no technical or engineering limitations to building a carbon nanotube x-ray source tube that enables this geometry.

CT provides improved sensitivity and specificity compared to tomosynthesis, but the high cost and high effective dose make regular CT examinations prohibitive. The effective dose of published DCT tomosynthesis systems, 0.13 mSv, is equivalent to that of a standard two view radiograph [0.06 to .25 mSv], but much less than even a low-dose thoracic CT examination [typically 1–4 mSv].[21–26] In the work by Johnsson et al. use of DCT instead

of CT resulted in a substantial dose reduction of almost 2 mSv.[27] Our system has an entrance dose of 0.6 mGy, this is comparable to the entrance dose of standard chest x-ray systems [0.31 to 0.88mGy] and conventional chest tomosynthesis systems [0.31 to 1.27mGy].[28,29] Our system can achieve further dose reductions by developing more advanced reconstruction techniques and utilizing a more optimized x-ray tube.

Our results show a high correlation between the Brasfield Score and the tomosynthesis score for cystic fibrosis, suggesting a similar, but not exactly equivalent evaluation of disease severity. With other studies showing higher sensitivity and specificity of tomosynthesis compared to conventional thoracic radiographs we hypothesize that the correlation is not closer to 1 due to our s-DCT system detecting lung features that are not evident on CXR. [18,20] This trend of correlation coefficients was found in other CF tomosynthesis evaluations as well.[17] Despite the new nature of the technique, intraclass correlation coefficients were found to be comparable for tomosynthesis scores. More detailed assessments and larger patient populations, however, are necessary to further evaluate differences in the two approaches.

The correlation of the tomosynthesis disease severity score with pulmonary function tests further indicates the reliability of the s-DCT approach. Both imaging methods were found to have statistically significant correlations with PFTs [FVC, FEV1, and FEF 25–75%] with no statistical difference between the correlation coefficients of the two modalities. This data confirms the potential of s-DCT to provide greater imaging detail while still correlating well with other diagnostic information. While the measurements from PFTs are typically considered the most clinically relevant in CF patients[30,31], our study suggests the potential utility of s-DCT to characterize disease in patients who are unable to perform PFTs, typically young children.[32–34] Recent studies have shown that detecting disease at an early age, prior to clinical presentation may be important in altering the disease course of CF.[35] In contrast to conventional tomosynthesis, our CNT based x-ray tubes can also perform prospective gated imaging, which would increase temporal resolution [and thus disease detection] in young populations that cannot perform breath holds, at a minimum radiation dose and relatively low cost.[36] Since our study only included a small range of the disease severity scale [the highest tomosynthesis score was 38 out of 100], further study must be done to quantify the correlation of tomosynthesis with PFTs in extreme disease states. This will allow us to better understand the relationship between the two measures and especially how it may relate to more subtle areas of disease. Ultimately, this technique may need to be evaluated compared with high-resolution CT.

Readers scored CXR as having higher image quality than s-DCT when evaluating full inspiration and the degree of motion blur. Our prototype s-DCT system acquires images in the AP configuration which likely contributes to this perception, despite equivalent breathing instructions. Motion blur in s-DCT images is also likely introduced by a combination of longer imaging times, cardiac motion, [which cannot be compensated without more time consuming cardiac gating], and imperfect breath holds, which may be reduced by improved subject training and operator experience. Differences in patient positioning were also likely due to the technologist learning curve for the new modality. The higher scores given to CXR when evaluating lung fields, diaphragm and LCA are presumably due to motion blur at the

lung bases as a result of imperfect breath hold and cardiac motion. Proximal bronchi, small airways and vascular pattern, all of which are important when assessing disease severity in CF patients, scored better on s-DCT than CXR. Conventional DCT relies on the translation of a standard x-ray tube across an arc in front of the patient. This translation time typically requires 10 to 15 seconds, acquiring approximately 60 images. As a result, conventional DCT systems are severely limited in their ability to image patients who are unable to hold their breath for the entire imaging time [6]. These limitations preclude its use in patients such as young children or patients with severe respiratory compromise. A method to perform low dose, gated tomosynthesis imaging could have the potential to offer clinicians insight into the disease status of these patients without excessive doses of radiation. We are currently in the process of evaluating the system for gated tomosynthesis imaging, which could potentially be used to assess disease presence in young patients before they become symptomatic.[37]

Study Limitations

The main limitation of the current s-DCT system is the small angular span [12° vs $30\text{--}60^\circ$ in conventional DCT systems], a function of the repurposed tube. The small angular span leads to less effective out-of-plane suppression, or increased artifact spread function. In a study by Gomi et al. increased angular span was shown to significantly increase the detectability index of artificial pulmonary nodules.[38] Interestingly, small angular spans have been shown to be better for evaluation of small structures, such as microcalcifications in breast tomosynthesis.[39,40] We anticipate that future use of a tube that offers wider angular span will offer clinical results comparable to other published chest tomosynthesis studies. Longer length tubes have been manufactured for non-medical imaging applications, and we do not anticipate any difficulty adapting these for medical use in the future. Furthermore, faster detector readout could also substantially reduce imaging time without causing the resolution penalty in conventional moving sources. It should be emphasized that our s-DCT system is a prototype that would require modifications to be ready to be evaluated against approved chest imaging system. Another study limitation involves the lack of effective dose estimations. We were unable to acquire these values which makes dose comparisons to other systems difficult. Entrance dose for this system however is similar to conventional DCT systems. Additionally, the small number of patients in this trial makes robust conclusions difficult to attain.

CONCLUSIONS

We demonstrate the potential for CNT based linear x-ray source arrays for use in clinical studies in this first-in-human study of stationary digital chest tomosynthesis for cystic fibrosis. Despite the limitations imposed by this repurposed tube, this study has shown that stationary chest tomosynthesis imaging has the potential to become a useful tool in the evaluation of CF patients, and opens the door to low-cost early detection of CF in the pediatric population. Prospective respiratory gated imaging trials with the s-DCT imaging system are ongoing.

Supplementary Material

Refer to Web version on PubMed Central for supplementary material.

Funding:

This project was supported by a *Translational Team Science Award* with funding provided by the UNC School of Medicine and the National Center for Advancing Translational Sciences (NCATS), National Institutes of Health, through Grant Award Number UL1TR001111. This work is partially supported by NCI-funded Carolina Center for Cancer Nanotechnology Excellence (5U54CA151652), a grant from the NCI (5R21CA185741), and The University of North Carolina School of Medicine Team Translation Science Award (TTSA001P2). Dr. Goralski was supported by the Cystic Fibrosis Foundation, Grant Award Number GORALS16C0 and GORALS12L0. This project was further supported by the Howard Holderness Distinguished Medical Scholar grant. Hardware support was also provided by Carestream Health Inc (Rochester, NY).

Abbreviations

AFVR	Adapted Fan-beam Volume Reconstruction
CF	Cystic Fibrosis
CNT	Carbon nanotube
DCT	Digital Chest Tomosynthesis
FEF 25–75%	Forced Expiratory Flow at 25–75% of the pulmonary volume
FEV1	Forced Expiratory Volume During the First Second
FVC	Forced Vital Capacity
ICCs	Intra-class Correlation Coefficients
LCA	Lateral Costophrenic Angles
PFTs	Pulmonary Function Tests
s-DCT	Stationary Digital Chest Tomosynthesis

REFERENCES

1. Sly PD, Gangell CL, Chen L, Ware RS, Ranganathan S, Mott LS, et al. (2013) Risk factors for bronchiectasis in children with cystic fibrosis. *N Engl J Med*. 2013 5 23;368[21]:1963–70. [PubMed: 23692169]
2. Cleveland RH, Sawicki GS, Stamoulis C. (2015) Similar performance of Brasfield and Wisconsin scoring systems in young children with cystic fibrosis. *Pediatr Radiol*. 2015 10;45[11]:1624–8. [PubMed: 26022611]
3. 2016 Patient Registry Annual Data Report. Cyst Fibros Found Patient Regist. Available via <https://www.cff.org/Research/Researcher-Resources/Patient-Registry/2016-Patient-Registry-Annual-Data-Report.pdf>
4. Dournes G, Menut F, Macey J, Fayon M, Chateil J-F, Salel M, et al. (2016) Lung morphology assessment of cystic fibrosis using MRI with ultra-short echo time at submillimeter spatial resolution. *Eur Radiol*. 2016 11 1;26[11]:3811–20. [PubMed: 26843010]
5. Chou S-HS, Kicska GA, Pipavath SN, Reddy GP. (2014) Digital tomosynthesis of the chest: current and emerging applications. *Radiogr Rev Publ Radiol Soc N Am Inc*. 2014 4;34[2]:359–72.

6. Dobbins JT, McAdams HP, Godfrey DJ, Li CM. (2008) Digital tomosynthesis of the chest. *J Thorac Imaging*. 2008 5;23[2]:86–92. [PubMed: 18520565]
7. Terzi A, Bertolaccini L, Viti A, Comello L, Ghirardo D, Priotto R, et al. (2013) Lung cancer detection with digital chest tomosynthesis: baseline results from the observational study SOS. *J Thorac Oncol Off Publ Int Assoc Study Lung Cancer*. 2013 6;8[6]:685–92.
8. Vult von Steyern K, Björkman-Burtscher I, Geijer M. (2011) Tomosynthesis in pulmonary cystic fibrosis with comparison to radiography and computed tomography: a pictorial review. *Insights Imaging*. 2011 11 11;3[1]:81–9. [PubMed: 22696001]
9. Steyern KV von, Björkman-Burtscher IM, Geijer M. (2013) Radiography, tomosynthesis, CT and MRI in the evaluation of pulmonary cystic fibrosis: an untangling review of the multitude of scoring systems. *Insights Imaging*. 2013 12 1;4[6]:787–98. [PubMed: 24065629]
10. Steyern KV von, Björkman-Burtscher I, Geijer M. (2012) Tomosynthesis in pulmonary cystic fibrosis with comparison to radiography and computed tomography: a pictorial review. *Insights Imaging*. 2012 2 1;3[1]:81–9. [PubMed: 22696001]
11. Steyern KV von, Björkman-Burtscher IM, Weber L, Höglund P, Geijer M. (2014) Effective dose from chest tomosynthesis in children. *Radiat Prot Dosimetry*. 2014 2 1;158[3]:290–8. [PubMed: 24026899]
12. Lee YZ, Burk LM, Wang K, Cao G, Volmer J, Lu J, et al. (2011) Prospective respiratory gated carbon nanotube micro computed tomography. *Acad Radiol*. 2011 5;18[5]:588–93. [PubMed: 21377908]
13. Shan J, Tucker AW, Lee YZ, Heath MD, Wang X, Foos DH, et al. (2015) Stationary chest tomosynthesis using a carbon nanotube x-ray source array: a feasibility study. *Phys Med Biol*. 2015;60[1]:81. [PubMed: 25478786]
14. Hartman Allison E., Shan Jing, Wu Gongting, Lee Yueh Z., Zhou Otto, Lu Jianping, et al. (2016) Initial clinical evaluation of stationary digital chest tomosynthesis. In 2016. p. 978366–9783–8. Available from: 10.1117/12.2216850
15. Wu Gongting, Inscoe Christine, Calliste Jabari, Lee Yueh Z., Zhou Otto, Lu Jianping. (2015) Adapted fan-beam volume reconstruction for stationary digital breast tomosynthesis. In 2015. p. 94123J-9412–10. Available from: 10.1117/12.2081931
16. Brasfield D, Hicks G, Soong S, Peters J, Tiller R. (1980) Evaluation of scoring system of the chest radiograph in cystic fibrosis: a collaborative study. *Am J Roentgenol*. 1980 6 1;134[6]:1195–8. [PubMed: 6770630]
17. Vult von Steyern K, Björkman-Burtscher IM, Höglund P, Bozovic G, Wiklund M, Geijer M. (2012) Description and validation of a scoring system for tomosynthesis in pulmonary cystic fibrosis. *Eur Radiol*. 2012 12;22[12]:2718–28. [PubMed: 22752406]
18. Choo JY, Lee KY, Yu A, Kim J-H, Lee SH, Choi JW, et al. (2016) A comparison of digital tomosynthesis and chest radiography in evaluating airway lesions using computed tomography as a reference. *Eur Radiol*. 2016 9 1;26[9]:3147–54. [PubMed: 26662264]
19. Yamada Y, Jinzaki M, Hashimoto M, Shiomi E, Abe T, Kuribayashi S, et al. (2013) Tomosynthesis for the early detection of pulmonary emphysema: diagnostic performance compared with chest radiography, using multidetector computed tomography as reference. *Eur Radiol*. 2013 8;23[8]:2118–26. [PubMed: 23515916]
20. Galea A, Dubbins P, Riordan R, Adlan T, Roobottom C, Gay D. (2015) The value of digital tomosynthesis of the chest as a problem-solving tool for suspected pulmonary nodules and hilar lesions detected on chest radiography. *Eur J Radiol*. 2015 5;84[5]:1012–8. [PubMed: 25757629]
21. Kalra MK, Maher MM, Rizzo S, Kanarek D, Shephard J-AO. (2004) Radiation exposure from Chest CT: Issues and Strategies. *J Korean Med Sci*. 2004 4;19[2]:159–66. [PubMed: 15082885]
22. Diederich S, Lenzen H. (2000) Radiation exposure associated with imaging of the chest: comparison of different radiographic and computed tomography techniques. *Cancer*. 2000 12 1;89[11 Suppl]:2457–60. [PubMed: 11147626]
23. 1990 Recommendations of the International Commission on Radiological Protection. *Ann ICRP*. 1991;21[1–3]:1–201. Available from: <http://www.icrp.org/publication.asp?id=icrp%20publication%2060>

24. Wiest PW, Locken JA, Heintz PH, Mettler FA. (2002) CT scanning: a major source of radiation exposure. *Semin Ultrasound CT MR*. 2002 10;23[5]:402–10. [PubMed: 12509110]
25. Sabol JM. (2009) A Monte Carlo estimation of effective dose in chest tomosynthesis. *Med Phys*. 2009 12;36[12]:5480–7. [PubMed: 20095260]
26. Båth M, Svåkvist A, von Wrangel A, Rismyhr-Olsson H, Cederblad Å. (2010) Effective dose to patients from chest examinations with tomosynthesis. *Radiat Prot Dosimetry*. 2010 4 1;139[1–3]: 153–8. [PubMed: 20233755]
27. Johnsson ÅA, Vikgren J, Båth M. (2014) A retrospective study of chest tomosynthesis as a tool for optimizing the use of computed tomography resources and reducing patient radiation exposure. *Acad Radiol*. 2014 11;21[11]:1427–33. [PubMed: 25097012]
28. Aliasgharzadeh A, Miandoost E, Masoumbeigi M, Salimian M, Mohseni M. (2015) Measurement of Entrance Skin Dose and Calculation of Effective Dose for Common Diagnostic X-Ray Examinations in Kashan, Iran. *Glob J Health Sci*. 2015 9;7[5]:202–7. [PubMed: 26156930]
29. Hwang HS, Chung MJ, Lee KS. (2013) Digital Tomosynthesis of the Chest: Comparison of Patient Exposure Dose and Image Quality between Standard Default Setting and Low Dose Setting. *Korean J Radiol*. 2013;14[3]:525–31. [PubMed: 23690724]
30. Corey M, Edwards L, Levison H, Knowles M. (1997) Longitudinal analysis of pulmonary function decline in patients with cystic fibrosis. *J Pediatr*. 1997 12;131[6]:809–14. [PubMed: 9427882]
31. Kerem E, Reisman J, Corey M, Canny GJ, Levison H. (1992) Prediction of mortality in patients with cystic fibrosis. *N Engl J Med*. 1992 4 30;326[18]:1187–91. [PubMed: 1285737]
32. Tepper RS, Hiatt PW, Eigen H, Smith J. Total Respiratory System Compliance in Asymptomatic Infants with Cystic Fibrosis. (1987) *Am Rev Respir Dis*. 1987 5 1;135[5]:1075–9. [PubMed: 3579006]
33. Beardsmore C (1996) Lung function from infancy to school age in cystic fibrosis. *Arch Dis Child*. 1996 1 1;73:519–23.
34. Beydon N, Amsallem F, Bellet M, Boulé M, Chaussain M, Denjean A, et al. (2002) Pulmonary Function Tests in Preschool Children with Cystic Fibrosis. *Am J Respir Crit Care Med*. 2002 10 15;166[8]:1099–104. [PubMed: 12379554]
35. Logan PM, O'Laoide RM, Mulherin D, O'Mahony S, FitzGerald MX, Masterson JB. (1996) High resolution computed tomography in cystic fibrosis: correlation with pulmonary function and assessment of prognostic value. *Ir J Med Sci*. 1996 3;165[1]:27–31. [PubMed: 8867494]
36. Vikgren J, Zachrisson S, Svåkvist A, Johnsson ÅA, Boijesen M, Flinck A, et al. (2008) Comparison of Chest Tomosynthesis and Chest Radiography for Detection of Pulmonary Nodules: Human Observer Study of Clinical Cases. *Radiology*. 2008 12 1;249[3]:1034–41. [PubMed: 18849504]
37. Shan J, Burk L, Wu G, Lee YZ, Heath MD, Wang X, et al. (2015) Prospective gated chest tomosynthesis using CNT X-ray source array. In 2015 [cited 2016 Nov 22]. p. 941242–941242–7. Available from: 10.1117/12.2081701
38. Gomi T, Nakajima M, Fujiwara H, Takeda T, Saito K, Umeda T, et al. (2012) Comparison between chest digital tomosynthesis and CT as a screening method to detect artificial pulmonary nodules: a phantom study. *Br J Radiol*. 2012 9;85[1017]:e622–9. [PubMed: 22422390]
39. Chawla AS, Lo JY, Baker JA, Samei E. (2009) Optimized image acquisition for breast tomosynthesis in projection and reconstruction space. *Med Phys*. 2009 11;36[11]:4859–69. [PubMed: 19994493]
40. Chan H-P, Goodsitt MM, Helvie MA, Zelakiewicz S, Schmitz A, Noroozian M, et al. (2014) Digital Breast Tomosynthesis: Observer Performance of Clustered Microcalcification Detection on Breast Phantom Images Acquired with an Experimental System Using Variable Scan Angles, Angular Increments, and Number of Projection Views. *Radiology*. 2014 12;273[3]:675–85. [PubMed: 25007048]

Key Points

1. Carbon nanotube based linear array x-ray tomosynthesis systems have the potential to provide diagnostically relevant information for patients with cystic fibrosis without the need for a moving gantry.
2. Despite the short angular span in this prototype system, lung features such as the proximal bronchi, small airways, and pulmonary vasculature have improved visualization on s-DCT compared to CXR. Further improvements are anticipated with longer linear x-ray array tubes.
3. Evaluation of disease severity in CF patients is possible with s-DCT, yielding improved visualization of important lung features and high correlation with pulmonary function tests at a relatively low dose.

FIGURE 1: s-DCT SYSTEM

The stationary digital chest tomosynthesis system linear x-ray array is aligned with the patient imaging bed [stretcher]. The detector is integrated into the stretcher bed.

Figure 2: Mild CF

Mild severity CF, corresponding to a 21 on Brasfield scoring system and 14 on tomosynthesis scoring system. A single image from the tomosynthesis exam [a] demonstrates improved visualization of bronchial wall thickening and mucous plugging compared to conventional radiograph [b].

Figure 3: Severe CF

Severe CF, corresponding to a 15 on Brasfield scoring system and 33 on tomosynthesis scoring system. A single image from the tomosynthesis exam [a] demonstrates improved visualization of bronchial wall thickening, bronchiectasis, and mucous plugging compared to conventional radiograph [b].

Table 1:
PATIENT PULMONARY FUNCTION TESTS

Pulmonary function test results for recruited patients. Values for FVC and FEV1 are in liters. Values for FEF 25–75% are in liters/second. Subject population n = 13.

	FVC	FEV1	FEV1/FVC Ratio	FEF 25–75%
CF001	5.18	3.34	64	2.09
CF002	4.8	3.45	72	2.40
CF003	6.5	4.37	67	2.69
CF004	4.15	3.24	78	2.93
CF005	4.41	3.07	70	1.82
CF006	3.62	2.66	74	2.06
CF007	4.98	3.43	69	2.05
CF008	4.26	3.59	84	4.15
CF009	1.96	1.32	67	0.78
CF010	4.82	3.81	81	3.58
CF011	2.71	1.77	65	1.11
CF012	5.71	4.2	74	3.23
CF013	3.81	3.18	84	3.54

Table 2:**IMAGING TECHNIQUE**

Readers gave significantly higher scores for CXR when evaluating imaging technique, likely due to the more simplified imaging process [single view versus 29]. Reader scores for these parameters were reader independent for all three parameters.

		Median Score			p-values	
		Reader 1	Reader 2	Reader 3	Technique	Reader
Full Inspiration	CXR	5	5	5	0.0015	0.0213
	s-DCT	5	4	3		
Motion Blur	CXR	5	5	5	0.002	0.0044
	s-DCT	4.5	5	3		
Patient Positioning	CXR	5	5	5	0.0246	0.0083
	s-DCT	5	5	4		

Table 3:**IMAGE QUALITY**

Readers gave s-DCT higher scores when evaluating image quality for proximal bronchi, small airways, and vascular pattern. P values with a * indicate statistically significant differences in technique, favoring tomosynthesis. P values marked with a † indicate statistically significant differences in technique, favoring CXR. Image quality of the trachea was not statistically different between CXR and s-DCT. Reader effects were not significant except when evaluating proximal bronchi and diaphragm and LCA.

	Lung Fields	Trachea	Proximal Bronchi	Small Airways	Diaphragm and LCA	Vascular Pattern
Technique	0.0249†	0.3608	0.0329*	0.0028*	0.0303†	0.0071*
Reader	0.0035	<.0001	<.0001	0.0012	0.019	0.001
Interaction of Reader and Technique	NS	NS	0.0033	NS	0.0095	NS

Table 4:
Pulmonary Function Correlation

Tomosynthesis has high correlation with FEV1, FEF 25–75%, and good correlation with FVC. P-value < 0.05 indicates that the correlation coefficient is significantly different than 0. Subject population n = 13.

Pulmonary Function Tests Pearson Correlation Coefficients					
		FVC	FEV1	FEV1 FVC Ratio	FEF 25–75
Average s-DCT	CC	−0.5818	−0.68506	−0.41015	−0.71886
	p-value	0.037	0.0098	0.1639	0.0056
Average Brasfield	CC	0.68117	0.79454	0.47496	0.78121
	p-value	0.0104	0.0012	0.101	0.0016