Monte Carlo simulation using PRIMO code as a tool for checking the credibility of commissioning and quality assurance of 6 MV TrueBeam STx varian LINAC

M. Arif Efendi^{a,*}, Amporn Funsian^a, Thawat Chittrakarn^b, Tripob Bhongsuwan^b

^a Department of Radiology, Faculty of Medicine, Prince of Songkla University, Hatyai, Thailand

^b Department of Physics, Faculty of Science, Prince of Songkla University, Hatyai, Thailand

ARTICLE INFO

Article history:

Received 8 December 2018

Received in revised form 3 October 2019

Accepted 18 December 2019

Available online 20 December 2019

Keywords:

Commissioning

PRIMO code

TrueBeam STx varian LINAC

Quality assurance

Varian phase-space files

ABSTRACT

Aim: To validate and implement Monte Carlo simulation using PRIMO code as a tool for checking the credibility of measurements in LINAC initial commissioning and routine Quality Assurance (QA). Relative and absolute doses of 6 MV photon beam from TrueBeam STx Varian Linear Accelerator (LINAC) were simulated and validated with experimental measurement, Analytical Anisotropic Algorithm (AAA) calculation, and golden beam.

Methods and Materials: Varian phase-space files were imported to the PRIMO code and four blocks of jaws were simulated to determine the field size of the photon beam. Water phantom was modeled in the PRIMO code with water equivalent density. Golden beam data, experimental measurement, and AAA calculation results were imported to PRIMO code for gamma comparison.

Results: PRIMO simulations of Percentage Depth Dose (PDD) and in-plane beam profiles had good agreement with experimental measurements, AAA calculations and golden beam. However, PRIMO simulations of cross-plane beam profiles have a better agreement with AAA calculation and golden beam than the experimental measurement. Furthermore, PRIMO simulations of absolute dose agreed well with experimental results with $\pm 0.8\%$ uncertainty.

Conclusion: The PRIMO code has good accuracy and is appropriate for use as a tool to check the credibility of beam scanning and output measurement in initial commissioning and routine QA.

© 2019 Greater Poland Cancer Centre. Published by Elsevier B.V. All rights reserved.

1. Introduction

Accuracy of dose delivered to the patient is directly related to the beam data commissioning used in a Treatment Planning System (TPS). These data are verified periodically in the Quality Assurance (QA) of LINAC beam performance by a medical physicist.¹ Lack of standards and need for precision are challenges to the medical physicist, in the credibility of measurements in initial commissioning and routine QA. The TG-106 report of the American Association of Physicists in Medicine (AAPM)² gives two recommendations to verify initial commissioning results. The first recommendation is to use golden beam data for checking the quality of the beam scanning. Varian provides golden beam data, which is a reference dosimetric dataset. It could be used to evaluate the beam reproducibility of

LINAC. The golden beam dataset is available at www.myvarian.com with the name “TrueBeam Representative Beam Data for Eclipse”. Another recommendation is to verify beam scanning with Monte Carlo (MC) simulation. However, experimental measurements or golden beam are still needed to validate such MC simulation.

MC simulation is the most powerful tool for dose calculations in radiotherapy. MC simulation is implemented in several available codes, e.g. MCNP,³ Geant,⁴ PENELOPE,⁵ FLUKA⁶ for general applications; and PEREGRINE,⁷ EGSnrc,⁸ DPM,⁹ VMC++,¹⁰ ORANGE,¹¹ PRIMO¹² for medical applications. The availability of MC simulation tools has led to clinics using it in clinical radiation oncology.^{13,14} Nevertheless, the selection of a suitable and practical MC simulation code, and accurate validation of the simulation parameters, are necessary.

PRIMO is a MC code specific for radiotherapy. The computation engine of PRIMO is based on the PENELOPE general MC code. It can use parallel processors and variance reduction techniques to make the simulation quicker in real time. PRIMO simulates medical LINAC and estimates absorbed dose distributions in a water phantom or

* Corresponding author at: Department of Radiology, Faculty of Medicine, Prince of Songkla University, 90110 Hatyai city, Songkla, Thailand.

E-mail address: marife@psu.ac.th (M.A. Efendi).

in Computed Tomography (CT) data. PRIMO can simulate almost all Varian and Elekta LINAC, with their electron applicators and MLC.¹²

There are two ways to characterize the source of radiation for an MC simulation, to inform about particle type, energy, position, direction of movement, etc.¹⁵ The first is by using phase-space files,¹⁶ and the second is by providing detailed geometry of the LINAC head that should be modelled.¹⁷ Phase space files of TrueBeam 6 MV photon beam have been made available by Constantin and co-workers at www.myvarian.com/montecarlo. The phase-space files are formatted according to practices of the International Atomic Energy Agency (IAEA). Simulations were divided into two parts, (1) accelerator part using the Parmela code, (2) LINAC head using the Geant4 V4.9.2.p01 MC code. The first phase space result was used as input for the second simulation.¹⁸

Percentage Depth Dose (PDD) and beam profiles (cross-plane and in-plane) are two important concerns affecting the accuracy of MC modeling in a relative dose measurement. PDD is used as a beam quality parameter, and beam profiles are used for radiation dose reliability assessment.^{19–21} Aside from the relative dose, also the absolute dose is an important dosimetry aspect pursued by initial commissioning and routine QA. Experiments on absolute dose are also called output measurements.²²

Belosi and co-workers simulated 6 and 10 MV FFF TrueBeam using Varian phase-space files. Seven experimental measurements and golden beam dataset were used to validate the MC results.²³ Our previous work has published 6 MV FFF photon beam simulation of TrueBeam using the PRIMO code. In this publication authors generate phase space files with 10^9 histories. The results were in a better than 84% agreement with experimental measurements, using 1% 1 mm gamma criteria.²⁴

In most clinical practice QA of beam performance is done by matching the flatness and symmetry with commissioning data, although this is still a rough approach. Verification of all points in PDD and beam profiles would be needed for a precise QA of beam performance. Toutaoui et al. verified photon dose calculation by multi-grid superposition in a commercial TPS (CMS Xio, Elekta) using MC EGSnrc code.²⁵ Another study concerned the electron mode.²⁶ MC EGS4/BEAM code was used to commission the electron beam of Varian Clinac 2100C. However, the MC software used was not well suited for regular use as a verification tool in initial commissioning and routine QA. Moreover, the precise, reliable, and standard model of MC simulation is important if it is used as a verification tool of experimental measurements.

The goal of this work was to validate and implement MC simulation using the PRIMO code for 6 MV photon beam of TrueBeam STx Varian LINAC as a tool to verify the initial commissioning and routine QA. PRIMO simulation results are compared with results from experimental measurements, AAA calculations, and golden beam. TrueBeam STx Varian LINAC Radiotherapy facility at Songklanagarind Hospital, Prince of Songkla University, will be utilized in experimental measurements with the AAA Eclipse treatment planning system.

2. Materials and methods

2.1. Varian phase-space files

The phase-space files of 6 MV photon beam of TrueBeam STx as provided by Varian are available at www.myvarian.com/montecarlo.¹⁹ These serve as the first phase-space files in Fig. 1. The phase-space files can be used to describe the radiation source for dose estimation by MC simulation. Photon beam 6 MV of TrueBeam phase-space files contains 4.95×10^{10} histories and in total 1.79×10^9 particles.

Fig. 1. Sketch of Geometry of TrueBeam STx Varian LINAC head and water phantom in relation to simulation stages S1, S2, and S3.

2.2. Monte Carlo simulation using the PRIMO code

Simulation processes in the PRIMO code were divided into 3 parts. The first segment is called S1, which models from the electron beam to the ion chamber. The second segment called S2 has four blocks of jaws that can be adjusted to determine the field size of the radiation beam. The third segment is S3, which can be a water phantom or a patient's Digital Imaging and Communications in Medicine (DICOM) data. TrueBeam Varian phase-space files for the 6 MV photon beam were imported in the S1 part and simulated in the S2 and S3 parts using the PRIMO code, as shown in Fig. 1. Head of LINAC parts and estimated materials in each component are summarized in Table 1. The voxel size $3 \times 3 \times 3 \text{ mm}^3$ was used to simulate the water phantom.

2.3. AAA calculation

The Analytical Anisotropic Algorithm (AAA) is used in the Eclipse treatment planning system. It was developed by Waldemar Ulmer and Wolfgang Kaissl, in the publication "A Triple Gaussian Pencil beam Model for Photon beam Treatment Planning", in 1995.²⁷ Photon dose in the water phantom was calculated using AAA in Eclipse

Fig. 2. Experimental set-up (a) in a photograph and (b) in a schematic drawing.²¹

Table 1

Head of TrueBeam STx Varian LINAC parts and estimated materials.

No	Component	Composition	Symbol
1	Target	Tungsten, Tungsten-Copper	W, W-Cu
2	Primary collimator	Tungsten	W
3	Flattening filter	Lead-Steel-Copper	Pb-Fe-Cu
4	Ion chamber	Copper	Cu
5	Secondary collimator (Jaws)	Tungsten	W

software. PDD, cross-plane beam, and in-plane beam profiles were calculated along each axis. The AAA calculation in Eclipse software generated $0.25 \times 0.25 \times 0.25 \text{ mm}^3$ voxels.

2.4. Experiment setup

2.4.1. Relative dose measurement

TrueBeam STx Varian LINAC facility at Songklanagarind Hospital, Prince of Songkla University, was used in this experiment. Fig. 2a shows the 6 MV photon beam irradiating a water phantom. IBA blue phantom was used in the relative dose measurement by using IBA cc13. The ionization chamber scanned along each axis, as shown in Fig. 2b. When scanning along the Z, X, and Y axes, the PDD, cross-plane and in-plane beam profiles were obtained. The ionization chamber was set to move in 1 mm steps along each axis.

2.4.2. Absolute dose measurement

Experimental measurements of the absolute dose followed the IAEA Technical Report Series No. 398.²⁸ In this work, the Source to Surface Distance (SSD) of water phantom was 100 cm. However, all LINACs in Songklanagarind Hospital calibrate to 1 Gy at Dmax by using the Source Axis Distance (SAD) technique. A comparison between SAD and SSD techniques is shown in Fig. 3. IBA FC65-G with 0.65 cm^3 cavity volume, an ionization chamber, was used to measure output radiation along the central axis at depths of 1.5, 5, 10, and 20 cm.

2.5. Relative dose comparisons

Experimental measurement, AAA calculation, and golden beam data in ASCII format were imported to the PRIMO software. Comparisons between PRIMO simulation results and experimental

Fig. 3. Comparison of SAD and SSD techniques.

measurement, AAA calculation, and golden beam were performed using gamma analysis as proposed by Low et al.²⁹ Experimental measurement, AAA calculation, and golden beam were used as reference data and PRIMO simulation results as evaluation data. Gamma index (Γ) for reference point (p) and the dose at that point $d_e(p)$ are calculated based on Eq. (1).

$$\Gamma = \min \left\{ \sqrt{\left(\frac{\Delta d_i}{\Delta D} \right)^2 + \left(\frac{\Delta S_i}{\Delta S} \right)^2} \right\} \quad (1)$$

where:

ΔD = acceptance criteria for the dose difference,

ΔS = acceptance criteria for the distance to agreement,

Δd_i = the difference between $d_e(p)$ and the PRIMO simulated dose at a certain point p_i ,

ΔS_i = the distance between p and p_i .

In this work, the lowest acceptance criteria of 1% for the dose difference (ΔD) and 1 mm for the distance to agreement (ΔS) were

Table 2
PDD compared between PRIMO simulation and experimental measurement.

Field Size (cm ²)	Average Gamma Index		Percentage Passing Criteria (%)
	Build up Region	Equilibrium Region	
4 × 4	0.80	0.20	98.60
10 × 10	0.82	0.19	98.33
20 × 20	0.87	0.30	98.13
30 × 30	0.72	0.57	97.19
40 × 40	1.32	0.78	82.11

Table 3
PDD compared between PRIMO simulation and AAA calculation.

Field Size (cm ²)	Average Gamma Index		Percentage Passing Criteria (%)
	Build up Region	Equilibrium Region	
4 × 4	1.59	0.18	98.89
10 × 10	1.37	0.21	98.44
20 × 20	1.00	0.15	99.28
30 × 30	0.87	0.16	98.89
40 × 40	1.14	0.55	97.19

used. Based on Belosi et al.,²³ gamma passing criteria of 1%, 1 mm will show higher sensitivity than 2%, 2 mm.

2.6. Absolute dose comparisons

Absolute dose from the PRIMO simulation was compared with experimental measurements in field sizes 4 × 4, 10 × 10, and 20 × 20 cm² with and without MLC, and the field size 40 × 40 cm² was without MLC. Eq. 2 proposed by Zavgorodni et al.³⁰ was used to convert the absolute dose of PRIMO simulation from eV/g to Gy units.

$$D_{xyz,abs} = D_{xyz} \cdot S_b \cdot \frac{D_{xyz,abs}^{cal}}{D_{xyz}^{cal}} \quad (2)$$

where:

- $D_{xyz,abs}$ is the absolute dose of MC simulation in Gy after conversion,
- D_{xyz} is the normalized dose of MC simulation in eV/g,
- S_b is the monitor backscatter factors,
- $D_{xyz,abs}^{cal}$ is the absolute dose of experimental measurement in Gy at a reference point,
- D_{xyz}^{cal} is the normalized dose of MC simulation at a reference point.

3. Results

3.1. Percentage depth dose

PRIMO simulation of PDD gave a good agreement with experimental measurements, AAA calculations, and golden beam for field sizes up to 40 × 40 cm² as shown in Fig. 4a–e. Details of gamma agreement between the PRIMO simulation with experimental results, AAA calculations, and golden beam are presented in Tables 2–4, respectively.

3.2. Cross-plane beam profile

Beam profiles were determined with field size 10 × 10 cm² inside the water phantom at depths of 1.5, 5, 10, and 20 cm. Fig. 5a–d shows the comparison between the PRIMO simulation, experiment, AAA calculation, and golden beam. The PRIMO simulation has the poorest agreement with the experimental measurements, which only pass 68–71% gamma agreement as shown in Table 5, while the PRIMO simulation results compared with AAA calculation and

Table 4
PDD compared between PRIMO simulation and Golden Beam.

Field Size (cm ²)	Average Gamma Index		Percentage Passing Criteria (%)
	Build up Region	Equilibrium Region	
4 × 4	1.47	0.16	98.85
10 × 10	1.16	0.17	98.99
20 × 20	1.01	0.24	98.56
30 × 30	1.28	0.24	98.16
40 × 40	1.33	0.22	96.55

Table 5
Cross-plane beam profile compared between PRIMO simulation and experimental measurement for field size 10 × 10 cm².

Depth (cm)	Average Gamma Index			Percentage Passing Criteria (%)
	Inside field	Penumbra	Outside field	
1.5	0.59	1.82	0.99	68.96
5	0.66	1.77	0.94	70.55
10	0.70	1.36	0.95	68.90
20	0.61	1.16	0.94	71.02

Table 6
Cross-plane beam profile compared between PRIMO simulation and AAA calculation for field size 10 × 10 cm².

Depth (cm)	Average Gamma Index			Percentage Passing Criteria (%)
	Inside field	Penumbra	Outside field	
1.5	0.43	1.10	0.32	98.40
5	0.49	0.86	0.43	98.10
10	0.57	0.87	0.53	96.53
20	0.51	0.66	0.61	87.21

Table 7
Cross-plane beam profile compared between PRIMO simulation and Golden Beam for field size 10 × 10 cm².

Depth (cm)	Average Gamma Index			Percentage Passing Criteria (%)
	Inside field	Penumbra	Outside field	
1.5	0.93	1.91	0.38	77.32
5	0.65	1.55	0.30	88.15
10	0.65	1.26	0.28	92.51
20	0.46	0.71	0.30	95.62

Table 8
In-plane beam profile compared between PRIMO simulation and experimental measurement for field size 10 × 10 cm².

Depth (cm)	Average Gamma Index			Percentage Passing Criteria (%)
	Inside field	Penumbra	Outside field	
1.5	0.59	1.99	0.46	92.33
5	0.53	1.31	0.40	92.81
10	0.56	0.71	0.28	92.37
20	0.43	0.50	0.18	95.22

golden beam give a better than 77% passing gamma agreement, as shown in Tables 6 and 7.

3.3. In-plane beam profile

PRIMO simulation had good agreement with experimental results and AAA calculations, as shown in Fig. 6a–d. It had a better than 92% agreement with the experimental measurements and over 71% agreement with the AAA calculations, as shown in Tables 8 and 9.

Fig. 4. PDD compared for field sizes: (a) $4 \times 4 \text{ cm}^2$; (b) $10 \times 10 \text{ cm}^2$; (c) $20 \times 20 \text{ cm}^2$; (d) $30 \times 30 \text{ cm}^2$; and (e) $40 \times 40 \text{ cm}^2$.

Table 9

In-plane beam profile compared between PRIMO simulation and AAA calculation for field size $10 \times 10 \text{ cm}^2$.

Depth (cm)	Average Gamma Index			Percentage Passing Criteria (%)
	Inside field	Penumbra	Outside field	
1.5	0.36	0.81	0.76	80.82
5	0.51	1.38	0.82	78.37
10	0.58	1.14	0.82	87.57
20	0.47	0.87	0.96	71.36

Table 10

Absolute dose compared for field size $4 \times 4 \text{ cm}^2$.

Depth (cm)	Non MLC		MLC	
	Simulation (Gy)		Simulation (Gy)	
	Dose	2σ	Dose	2σ
1.5	0.9005	0.0065	0.9014	0.0066
5	0.7466	0.0055	0.7549	0.0055
10	0.5523	0.0044	0.5565	0.0044
20	0.2980	0.0033	0.3008	0.0033

3.4. Absolute dose

PRIMO simulations of the absolute dose with about $\pm 0.8\%$ uncertainty agreed well with the experimental results. The comparisons between the PRIMO simulation and experimental results are shown in Tables 10–13 for the respective field sizes.

4. Discussion

TrueBeam STx Varian LINAC has been simulated using the PRIMO code. Varian phase-space files of 6 MV photon beam were

imported into the S1 part. PRIMO code begin simulated the 6 MV photon beam of TrueBeam STx Varian LINAC from four block of jaws and water phantom in the S2 and S3 parts. PRIMO code appears suitable for use in small field dosimetry.³¹

However, Geometry of the TrueBeam STx Varian LINAC is still not distributed freely, instead the Varian Medical Systems provide the phase space files. Accuracy of MC simulations will depend on these phase-space files.

Fig. 5. Cross-plane beam profile compared for the field size $10 \times 10 \text{ cm}^2$ at the depths: (a) 1.5 cm; (b) 5 cm; (c) 10 cm; and (d) 20 cm.

Table 11
Absolute dose compared for field size $10 \times 10 \text{ cm}^2$.

Depth (cm)	Non MLC		MLC		
	Simulation (Gy)		Experiment (Gy)		
	Dose	2σ	Dose	2σ	Dose
1.5	0.9692	0.0066	0.9692	0.9714	0.0066
5	0.8284	0.0055	0.8373	0.8334	0.0055
10	0.6403	0.0055	0.6464	0.6424	0.0055
20	0.3663	0.0033	0.3701	0.3687	0.0033

Table 12
Absolute dose compared for field size $20 \times 20 \text{ cm}^2$.

Depth (cm)	Non MLC		MLC		
	Simulation (Gy)		Experiment (Gy)		
	Dose	2σ	Dose	2σ	Dose
1.5	0.9938	0.0065	0.9928	1.0252	0.0067
5	0.8657	0.0054	0.8691	0.8919	0.0056
10	0.6880	0.0054	0.6943	0.7087	0.0056
20	0.4178	0.0043	0.4226	0.4299	0.0044

4.1. Relative dose

The PDD curve is divided into two parts. The first part is from the surface of water to the depth of dose maximum, called the build-up region. The other part is the equilibrium region, which are the depths after dose maximum. The PDD was determined up

Table 13
Absolute dose compared for field size $40 \times 40 \text{ cm}^2$.

Depth (cm)	Non MLC		
	Simulation (Gy)		Experiment (Gy)
	Dose	2 σ	
1.5	1.0196	0.0064	1.0166
5	0.8996	0.0054	0.8972
10	0.7312	0.0054	0.7263
20	0.4632	0.0043	0.4588

to 30 cm under the surface of water phantom. Fig. 4 shows that the differences are bigger in the build-up region than in the equilibrium region. This is because in the build-up region there are free electrons interacting with Compton scattering, photoelectric effect, and pair production. High energy electrons will be ejected when a high energy photon interacts with water. These electrons will deposit their energy to the water, as explained in our previous article.²⁴

PRIMO simulation of PDD agreed well with the experiment, AAA calculation, and golden beam with over 96.55% passing the gamma criteria, except for in comparison with the experiment at the field size $40 \times 40 \text{ cm}^2$. The poorer gamma agreement with the field size $40 \times 40 \text{ cm}^2$ was probably caused by the length limitation of IBA blue phantom, making it not appropriate for scanning beam profiles up to 40 cm.

There are two kinds of beam profiles, a cross-plane beam profile in the X axis direction and an in-plane beam profile in the Y axis direction. The beam profile curve is further divided into three

Fig. 6. In-plane beam profiles compared for field size $10 \times 10 \text{ cm}^2$ at the depths: (a) 1.5 cm; (b) 5 cm; (c) 10 cm; and (d) 20 cm.

regions, namely the inside field region, the penumbra region where the dose falls rapidly towards the beam edge, and the outside field region out of the field. Dose difference is higher in the penumbra region than in the inside field and the outside field regions, as shown in Figs. 5 and 6. This is probably so because a lower number of simulated particles will be found in the penumbra region, and this will cause large stochastic fluctuations in the MC simulation and influence the gamma index.^{32,33}

The PRIMO simulations of the cross-plane beam profiles had a better agreement with AAA calculations than with experimental measurements. Experimental results of the cross-plane beam profiles had the poorest agreement with PRIMO simulations, which in turn had a good agreement with AAA calculations and golden beam of the cross-plane beam profiles. PRIMO simulation had a better than 87% agreement (i.e., passing the gamma criteria) with AAA calculations, as shown in Table 6.

Experimental measurement of the cross-plane beam profiles deviate from the ideal profiles. These beam profiles should be examined further, as they may contribute to inaccurate measurements of the LINAC configuration itself. Inaccurate beam profiles cause discrepancies with commissioning beam reference data, which is ultimately used by TPS. It will cause dosimetric characteristics of TPS not compatible with the real dose distribution received by patients.

4.2. Absolute dose

PRIMO simulation produces the absolute dose in electron-volts per gram (eV/g), and conversion to Gray (Gy) units is possible using

the monitor backscatter factor.³⁰ Absolute dose at 1.5 cm depth in field size $10 \times 10 \text{ cm}^2$ was determined using the SSD technique as 0.9692 Gy for both simulation and experimental measurement, while it is 1 Gy using the SAD technique. This is because the different positions of ionization chamber, as shown in Fig. 3, as the dose in the SSD technique decreases by the inverse-square law. The absolute dose was higher with Multi-leaf Collimator (MLC) than without it. This is because the MLC scattered radiation. In bigger field sizes the differences were much larger.

The good agreement in this work between the PRIMO simulations and experimental measurements of absolute dose with MLC show that the MLC model in the PRIMO code is precise and appropriate for use as a patient-specific QA of IMRT and VMAT plan. Moreover, the PRIMO code could be more practical and more accurate than that in a recent publication.³⁴

5. Conclusions

PRIMO simulations were in a good agreement with experimental measurements, AAA calculations, and golden beam in PDD and in-plane beam profiles. However, the PRIMO simulations had the poorest agreement with experiments in the cross-plane beam profiles. This probably indicates that the cross-plane beam profiles in experimental measurements deviate from the ideal profiles.

PRIMO simulation of the absolute dose ($\pm 0.8\%$ uncertainty) was in a good agreement with the experimental measurements. The TrueBeam STx Varian LINAC with MLC had more scattered radiation than without the MLC and gave higher absorbed dose in the ionization chamber or scoring region. This study indicates that the

PRIMO code can be used as a tool for checking the credibility of measurements in initial commissioning and routine QA (beam scanning and output). Moreover, superior to golden beam data, PRIMO simulation results provide dose distribution in three dimensions.

Financial disclosure

None declared.

Conflict of interest

None declared.

Acknowledgments

The authors would like to thank all medical physicists at Songklanagarind Hospital, radiation physicist from Department of Medical Sciences, Ministry of Public Health, Thailand, and engineers from Varian Medical Systems for their support and assistance to this project. We thank the PRIMO authors Lorenzo Brualla, Miguel Rodriguez, and Josep Sempau for their assistance with the PRIMO code. The authors would also like to thank Daren Sawkey for his help with downloading and using TrueBeam phase space files from the Varian website. Sergei Zavgorodni is acknowledged for his helpful advice on converting the absolute dose to Gray units. We thank Seppo Karrila, Faculty of Science and Industrial Technology, and the Research and Development Office of the Prince of Songkla University, for English proof reading service.

This work was carried out at the Department of Physics, Faculty of Science, Computer Center, and the Department of Radiology, Faculty of Medicine, Prince of Songkla University (PSU). Thailand's Education Hub for Southern Region of ASEAN Countries (TEH-AC) is acknowledged for scholarship to the first author during M.Sc. study at PSU in 2015–2017.

References

- Kutcher GJ, Coia L, Gillin M, et al. Comprehensive QA for radiation oncology: report of AAPM radiation therapy committee task group 40. *Med Phys.* 1994;21(4):581–618.
- Das IJ, Cheng C-W, Watts RJ, et al. Accelerator beam data commissioning equipment and procedures: report of the TG-106 of the Therapy Physics Committee of the AAPM. *Med Phys.* 2008;35(9):4186–4215.
- Forster RA, Godfrey TNK. MCNP—a general Monte Carlo code for neutron and photon transport. In: *Monte-Carlo methods and applications in Neutronics, photonics and statistical physics*. Berlin Heidelberg: Springer; 1985:33–55.
- Goossens M, Giani S, Ravndal S. GEANT: detector description and simulation tool. *CERN Program Libr Long Writeup W.* 1994;5013:1993.
- Baro J, Sempau J, Fernández-Varea JM, Salvat F. PENELOPE: an algorithm for Monte Carlo simulation of the penetration and energy loss of electrons and positrons in matter. *Nucl Instrum Methods Phys Res Sect B Beam Interact Mater At.* 1995;100(1):31–46.
- Fassó A, Ferrari A, Sala PR, Ranft J. FLUKA: Status and prospects for hadronic applications. In: *Advanced monte carlo for radiation physics, particle transport simulation and applications*. Berlin Heidelberg: Springer; 2001:955–960.
- Hartmann Siantar CL, Chandler WP, Rathkopf JA, Svatos MM, White RM. *PEREGRINE: An all-particle Monte Carlo code for radiation therapy*. CA (United States): Lawrence Livermore National Lab.; 1994:2537.
- Kawrakow I. Accurate condensed history Monte Carlo simulation of electron transport. I. EGSnrc, the new EGS4 version. *Med Phys.* 2000;27(3):485–498.
- Sempau J, Wilderman SJ, Bielajew AF. DPM, a fast, accurate Monte Carlo code optimized for photon and electron radiotherapy treatment planning dose calculations. *Phys Med Biol.* 2000;45(8):2263.
- Kawrakow I. VMC++, electron and photon Monte Carlo calculations optimized for radiation treatment planning. In: *Advanced monte carlo for radiation physics, particle transport simulation and applications*. Berlin Heidelberg: Springer; 2001:229–236.
- Van der Zee W, Hogenbirk A, Van der Marck SC. ORANGE: a Monte Carlo dose engine for radiotherapy. *Phys Med Biol.* 2005;50(4):625.
- Rodriguez M, Sempau J, Brualla L. PRIMO: a graphical environment for the Monte Carlo simulation of varian and Elekta linacs. *Strahlenther Onkol.* 2013;189(10):881–886.
- Chetty IJ, Curran B, Cygler JE, et al. Report of the AAPM Task Group No. 105: issues associated with clinical implementation of Monte Carlo-based photon and electron external beam treatment planning. *Med Phys.* 2007;34(12):4818–4853.
- Rogers DWO. Fifty years of Monte Carlo simulations for medical physics. *Phys Med Biol.* 2006;51(13):R287–R301.
- González W, García-Ferreira I-B, Anguiano M, Lallena AM. A general photon source model for clinical linac heads in photon mode. *Radiat Phys Chem.* 2015;117:140–152.
- Rucci A, Carletti C, Cravero W, Strbac B. Use of IAEA's phase-space files for the implementation of a clinical accelerator virtual source model. *Phys Med.* 2014;30(2):242–248.
- Sardari D, Maleki R, Samavat H, Esmaeeli A. Measurement of depth-dose of linear accelerator and simulation by use of Geant4 computer code. *Rep Pract Oncol Radiother.* 2010;15(3):64–68.
- Constantin M, Perl J, LoSasso T, et al. Modeling the TrueBeam linac using a CAD to Geant4 geometry implementation: dose and IAEA-compliant phase space calculations. *Med Phys.* 2011;38(7):4018–4024.
- Abdul Haneefa K, Cyriac TS, Musthafa MM, et al. FLUKA Monte Carlo for basic dosimetric studies of dual energy medical linear accelerator. *J Radiother.* 2014;2014.
- Eldib AAR, ELGohary MI, Fan J, et al. Dosimetric characteristics of an electron multileaf collimator for modulated electron radiation therapy. *J Appl Clin Med Phys.* 2010;11(2):5–22.
- Tartar A. Monte Carlo simulation approaches to dose distributions for 6 MV photon beams in clinical linear accelerator. *Biocybern Biomed Eng.* 2014;34(2):90–100.
- Popescu IA, Shaw CP, Zavgorodni SF, Beckham WA. Absolute dose calculations for Monte Carlo simulations of radiotherapy beams. *Phys Med Biol.* 2005;50(14):3375.
- Belosi MF, Rodriguez M, Fogliata A, et al. Monte Carlo simulation of TrueBeam flattening-filter-free beams using varian phase-space files: comparison with experimental data. *Med Phys.* 2014;41(5):051707.
- Efendi MA, Funsian A, Chittarakarn T, Bhongsuwan T. Monte carlo simulation of 6 MV flattening filter free photon beam of TrueBeam STx LINAC at Songklanagarind Hospital. *Sains Malays.* 2017;46(9):1407–1411.
- Toutaoui A, Ait chikh S, Khelassi-Toutaoui N, Hattali B. Monte Carlo photon beam modeling and commissioning for radiotherapy dose calculation algorithm. *Phys Med.* 2014;30(7):833–837.
- Antolak JA, Bieda MR, Hogstrom KR. Using Monte Carlo methods to commission electron beams: a feasibility study. *Med Phys.* 2002;29(5):771–786.
- Ulmer W, Harder D. A triple gaussian pencil beam model for photon beam treatment planning. *Z Für Med Phys.* 1995;5(1):25–30.
- Andreo P, Burns DT, Hohlfield K, et al. Absorbed dose determination in external beam radiotherapy: An international code of practice for dosimetry based on standards of absorbed dose to water. *IAEA TRS.* 2000;398.
- Low DA, Harms WB, Mutic S, Purdy JA. A technique for the quantitative evaluation of dose distributions. *Med Phys.* 1998;25(5):656–661.
- Zavgorodni S, Alhakeem E, Townson R. Monitor backscatter factors for the varian 21EX and TrueBeam linear accelerators: measurements and Monte Carlo modelling. *Phys Med Biol.* 2014;59(4):911.
- Benmakhlof H, Brualla L, Andreo P, Sempau J. SU-F-T-367: using PRIMO, a PENELOPE-based software, to improve the small field dosimetry of linear accelerators. *Med Phys.* 2016;43(6Part18), 3547–3547.
- Graves YJ, Jia X, Jiang SB. Effect of statistical fluctuation in Monte Carlo based photon beam dose calculation on gamma index evaluation. *Phys Med Biol.* 2013;58(6):1839.
- Saidi P, Sadeghi M, Tenreiro C. Variance reduction of Monte carlo simulation in nuclear engineering field. In: *Theory and applications of monte carlo simulations*. IntechOpen; 2013.
- Kandlakunta P, Momin S, Sloop A, Zhang T, Khan R. Characterizing a Geant4 Monte Carlo model of a multileaf collimator for a TrueBeam™ linear accelerator. *Phys Med.* 2019;59:1–12.