Correspondence

Comment on “Large area CMOS active pixel sensor x-ray imager for digital breast tomosynthesis: Analysis, modeling, and characterization” [Med. Phys. 42, 6294–6308 (2015)]

(Received 5 December 2015; accepted for publication 18 January 2016; published 1 March 2016)

[<http://dx.doi.org/10.1118/1.4942487>]

To the Editor,

In the paper by Zhao *et al.*,¹ the authors cited our paper by Lu *et al.*,² on the image quality of microcalcification in digital breast tomosynthesis (DBT). They compared the contrast-to-noise ratio (CNR) of microcalcifications calculated by their simulation model of a Dexela 2923 MAM detector using metal-oxide-semiconductor active pixel sensor (CMOS APS) technology with our experimentally measured CNR values using a GE GEN2 prototype DBT system with an amorphous silicon cesium-iodide active matrix flat panel (denoted as a-Si:H TFT PPS in Zhao *et al.*) detector. Zhao *et al.* selected several CNR values from our paper and compared the CNRs in Fig. 11 to demonstrate that the CMOS APS detector could produce a higher CNR (about ten in CNR value) at a lower mean glandular dose (MGD) than our system over a range of microcalcification sizes. We would like to point out some errors and key differences that make the comparison unreliable and misleading.

(1) Zhao *et al.* described that the CNR values they cited from our paper² were acquired with an MGD of 2.5 mGy. However, they overlooked the fact that although we acquired the original DBT scan of the American College of Radiology (ACR) phantom at 60° tomosynthesis angle and 21 projection views (PVs) having an MGD of 2.5 mGy, we used only 11 PVs of the 21 PVs for reconstruction when we simulated the other six different PV distributions of DBT systems in our study. Therefore, six of the seven conditions shown in Fig. 4 of Lu *et al.* would have an MGD of about 1.31 mGy (estimated as $11/21 = 52.4\%$ of 2.5); only the CNRs labeled as full set (FS) had an MGD of 2.5 mGy. As seen in Fig. 11 of Zhao *et al.*, they selected our CNR values obtained from the DBT reconstructions with 11 PVs but described our MGD as 2.5 mGy. They emphasized in their paper that the CMOS APS detector could provide higher CNR values at a lower MGD of 2 mGy than the a-Si:H PPS detector at an MGD of 2.5 mGy, which is incorrect because the MGD they should compare to was 1.31 mGy. Another minor error is that the phantom images in the Lu *et al.* study were acquired at 29 kVp but they cited it as 29–33 kVp.

(2) Zhao *et al.* calculated the contrast of a microcalcification under an idealized condition, namely, a single projection image of a microcalcification embedded in a thin (1 mm) slice of breast tissue material [Sec. 2.D and Eq. (20) in Zhao *et al.*¹] without reconstruction. They did not take into account the scattered radiation from a thick breast phantom that would degrade the image contrast. They further ignored the fact that

the electronic noise of the detector would contribute to every projection of a DBT scan. By assuming that a DBT scan was acquired by a single x-ray exposure, they could underestimate the total electronic noise in a DBT scan from the detector significantly. Furthermore, due to the oblique incidence angle of DBT projections to the detector and to a reconstructed slice of finite thickness (1 mm in this case), there is additional blurring in the reconstructed microcalcifications caused by the detector blur and the voxel size used in the reconstruction. The imperfect accuracy in the geometric parameters of a real DBT system (uncertainties in the projection angle and geometric distances), the finite focal spot size, and the limited-angle reconstruction further contribute to blurring in the reconstructed image. Zhao *et al.* ignored all these factors that can degrade the CNR of a small object such as a microcalcification in DBT and directly compared their idealized, calculated CNR values with experimental data obtained from real reconstructed DBT images by Lu *et al.*² and Park *et al.*³ The study by Park *et al.* used a DBT system with a Dexela 2923 MAM CMOS APS detector (the same model as that simulated by Zhao *et al.*) to image a breast phantom containing simulated microcalcifications using various DBT acquisition parameters. Park *et al.* reported, among other results, that the measured CNR values from the reconstructed DBT images of a 400 μm speck ranged approximately from 28 to 43 at a fixed MGD of 2 mGy, depending on the angular distribution of the PVs and the dose distribution among the PVs. Despite the idealized model used in their CNR calculation (described above), Zhao *et al.* showed some experimental CNR values by Park *et al.* that were close to their calculated CNR values. They further compared these CNR values to those by Lu *et al.* and stated “in comparison to the GE GEN2 a-Si:H based PPS x-ray imager, the CNR values achieved by the Dexela 2923 MAM CMOS APS x-ray imager are higher (by around 10) due to its low electronic noise.” They attributed the difference in the CNR values simply to the detector performance, disregarding other factors such as differences in the imaging conditions, the material of the simulated microcalcifications [see (3)], the DBT acquisition parameters (tomosynthesis angle, angular distributions of PVs, and dose distribution among the PVs), and the reconstruction methods between the studies by Lu *et al.* and Park *et al.* The conclusion is misleading.

(3) Park *et al.*³ used a BR3D phantom (Model 020; CIRS, Inc.) that contained simulated microcalcifications made of calcium carbonate (CaCO_3) specks. Lu *et al.* used an ACR

phantom in which the simulated microcalcifications were made of aluminum oxide (Al_2O_3) specks. In the mammography energy range (10–35 keV), the linear attenuation coefficient of CaCO_3 is, on average, about 1.7 times that of the Al_2O_3 . Therefore, the CNR values of Lu *et al.* (cited in Fig. 11 of Zhao *et al.*) should be increased by a factor of about 1.7 to account for the difference in the speck material and by a factor of about 1.24 (=square root of 2 mGy/1.31 mGy) to account for the difference in MGD, i.e., increased by a factor of about 2.1 ($=1.7 \times 1.24$) before being compared to the CNR values of Park *et al.*, if the CNR values in the two studies could be compared at all. The apparent CNR advantage that Zhao *et al.* showed for the Dexela detector would practically disappear after these corrections, especially considering the error bars and the range of CNR measurements from which these average values were derived.

(4) In Fig. 11, Zhao *et al.* reproduced an image of a simulated microcalcification cluster (nominal speck size $240\ \mu\text{m}$) from Fig. 3 of the Lu *et al.* paper and commented “ $165\ \mu\text{m}$ microcalcifications are still detectable using the Dexela 2923 MAM CMOS APS x-ray imager at MGD of 2.0 mGy, while $240\ \mu\text{m}$ microcalcifications are almost invisible using the GE GEN2 PPS x-ray imager at an even higher MGD of 2.5 mGy. Therefore, smaller microcalcifications can be detected using studied CMOS APS x-ray imager at lower dose. From Fig. 11, we can speculate that it would be very difficult for microcalcifications less than $200\ \mu\text{m}$ to be detected using the conventional a-Si:H based PPS x-ray imager.” We would like to point out that it is not reliable to judge the visibility of small signals from a picture in a published journal paper, especially for signals such as subtle microcalcifications. We had published the results of an observer study⁴ that evaluated the detectability of microcalcifications in reconstructed DBT of breast phantoms. In that study, heterogeneous breast phantoms that were 5 cm in thickness containing simulated microcalcification clusters (CaCO_3 specks) were imaged at seven acquisition geometries using our GE GEN2 prototype DBT system with an a-Si:H TFT PPS detector. The exposure condition was kept constant at Rh anode/Rh filter 29 kVp and 50 mAs with an estimated MGD of 1.1 mGy for each DBT scan. Six experienced radiologists participated as observers to search for the microcalcification clusters in the reconstructed DBT volumes. We found that their average sensitivities of detecting subtle microcalcification clusters (nominal speck size $150\text{--}180\ \mu\text{m}$) ranged from $80.4\% \pm 6.7\%$ at 60° tomosynthesis angle to $95.7\% \pm 6.6\%$ at

16° tomosynthesis angle. The speculation by Zhao *et al.* is therefore incorrect.

In summary, the conclusion by Zhao *et al.* that “In comparison to a General Electric GEN2 prototype DBT system (at MGD of 2.5 mGy), an increased CNR (by ~ 10) for microcalcifications was observed using the Dexela 2923MAM CMOS APS x-ray imager at a lower MGD (2.0 mGy).” is unwarranted.

The authors’ work is supported by a National Institutes of Health Grant No. RO1 CA151443. The prototype digital breast tomosynthesis system was developed by the GE Global Research Group, with input and some revisions from the University of Michigan investigators, through the Biomedical Research Partnership (NIH Grant No. R01 CA91713, PI: Paul Carson, Ph.D.) collaboration. The content of this letter does not necessarily reflect the position of the funding agencies and no official endorsement of any equipment and product of any companies mentioned should be inferred. The author would like to acknowledge that Mitch Goodsitt, Ph.D., noticed the difference in the materials of the simulated microcalcifications between the BR3D phantom (CaCO_3) used by Park *et al.* and the ACR phantom (Al_2O_3) used by Lu *et al.* and estimated the 1.7 times difference in the linear attenuation coefficients between the two types of specks.

^{a)} Author to whom correspondence should be addressed. Electronic mail: chanhp@umich.edu; Telephone: 734-936-4357.

¹ C. M. Zhao, J. Kanicki, A. C. Konstantinidis, and T. Patel, “Large area CMOS active pixel sensor x-ray imager for digital breast tomosynthesis: Analysis, modeling, and characterization,” *Med. Phys.* **42**, 6294–6308 (2015).

² Y. Lu, H.-P. Chan, J. Wei, M. M. Goodsitt, P. L. Carson, L. Hadjiiski, A. Schmitz, J. W. Eberhard, and B. E. H. Claus, “Image quality of microcalcifications in digital breast tomosynthesis: Effects of projection-view distributions,” *Med. Phys.* **38**, 5703–5712 (2011).

³ H. S. Park, Y. S. Kim, H. J. Kim, Y. W. Choi, and J. G. Choi, “Optimization of configuration parameters in a newly developed digital breast tomosynthesis system,” *J. Radiat. Res.* **55**, 589–599 (2014).

⁴ H.-P. Chan, M. M. Goodsitt, M. A. Helvie, S. Zelakiewicz, A. Schmitz, M. Noroozian, C. Paramagul, M. A. Roubidoux, A. V. Nees, C. H. Neal, P. Carson, Y. Lu, L. Hadjiiski, and J. Wei, “Digital breast tomosynthesis: Observer performance of clustered microcalcification detection on breast phantom images acquired with an experimental system using variable scan angles, angular increments, and number of projection views,” *Radiology* **273**, 675–685 (2014).

Heang-Ping Chan^{a)}

Department of Radiology, University of Michigan,
1500 E. Medical Center Drive, Med Inn Building C477,
Ann Arbor, Michigan 48109-5842