Published in final edited form as:

Proc SPIE Int Soc Opt Eng. 2020 February ; 11222: . doi:10.1117/12.2545292.

Fluorescence Image-Guided Surgery – a Perspective on Contrast Agent Development

Connor W. Barth^a, Summer L. Gibbs^{a,b,c}

^aBiomedical Engineering Department, Oregon Health & Science University, Portland, OR 97201.

^bKnight Cancer Institute, Oregon Health & Science University, Portland, OR 97201.

^cOHSU Center for Spatial Systems Biomedicine, Oregon Health & Science University, Portland, OR 97201.

Abstract

In the past several decades, a number of novel fluorescence image-guided surgery (FGS) contrast agents have been under development, with many in clinical translation and undergoing clinical trials. In this review, we have identified and summarized the contrast agents currently undergoing clinical translation. In total, 39 novel FGS contrast agents are being studied in 85 clinical trials. Four FGS contrast agents are currently being studied in phase III clinical trials and are poised to reach FDA approval within the next two to three years. Among all novel FGS contrast agents, a wide variety of probe types, targeting mechanisms, and fluorescence properties exists. Clinically available FGS imaging systems have been developed for FDA approved FGS contrast agents, and thus further clinical development is required to yield FGS imaging systems tuned for the variety of contrast agents in the clinical pipeline. Additionally, study of current FGS contrast agents for additional disease types and development of anatomy specific contrast agents is required to provide surgeons FGS tools for all surgical specialties and associated comorbidities. The work reviewed here represents a significant effort from many groups and further development of this promising technology will have an enormous impact on surgical outcomes across all specialties.

Keywords

fluorescence; image-guided surgery; fluorescence imaging system; clinical development; clinical trial; contrast agent; near-infrared fluorescence

1. Introduction

Fluorescence image-guided surgery (FGS) technologies have been under development for the past three decades. A movement that began with the development of clinically approved near-infrared (NIR) fluorescent agents indocyanine green (ICG) and methylene blue as vascular tracers has grown to become a multidisciplinary field studied by dozens of groups worldwide.^{1–4} Through the creation of targeted contrast agents and sensitive imaging systems, FGS has the potential to revolutionize surgery, improving surgical outcomes by enhancing visualization of tissues for resection, such as tumors, or preservation, such as nerves or vasculature. Utilizing compact and relatively low-cost imaging systems, FGS can be readily implemented into many procedures to bridge the gap between preoperative

imaging, such as magnetic resonance imaging (MRI) and computed tomography (CT), and the current intraoperative reality.^{5–15} FGS systems have been successfully utilized in a wide variety of clinical applications to improve outcomes, including tumor resection, sentinel lymph node mapping, angiography, lymphography, and ureter and bile duct anatomic imaging.^{16–28} Additionally, a number of targeted contrast agents are under development to expand the applications of this promising technology. For example, 5-aminomino-levulinic acid (5-ALA) and its fluorescent metabolite protoporphyrin IX (PpIX) has garnered clinical approval for glioma resection and has significantly enhanced complete resection rates and revolutionized neurosurgical treatment of brain tumors over the past decade.²⁹ Many more targeted contrast agents are currently in the clinical pipeline, representing substantial effort by the FGS community to translate this promising technology to many surgical indications.

Considered new drugs by the FDA, novel FGS contrast agents face a lengthy and expensive approval process to reach clinical use. Completion of preclinical development to enable Investigational New Drug (IND) approval for first-in-human trials represents the first major hurdle for FGS contrast agent development. Subsequently, completion of Phase I, II, and III clinical trials requires exponentially more time, effort, and money to complete successfully prior to market approval. Finding the necessary investments and strategic partners to navigate this long and costly process is difficult, especially for FGS contrast agents as non-curative diagnostic agents for single use, which provide a low financial incentive for investment from commercial sources.^{30,31} Nonetheless, many startups, groups in academia, and even mid-scale biotechnology and medical device companies have worked towards clinical translation of a variety of novel FGS contrast agents. Herein, we review the current landscape of novel FGS contrast agents undergoing clinical translation and outline their integration with the current ecosystem of clinical, and pre-clinical, fluorescence imaging systems.

2. Clinically Approved Fluorescence Image-Guided Surgery Contrast Agents

A complete picture of the current clinical translation of FGS contrast agents cannot be achieved without including fluorescent agents with current clinical approval. Although few in number and most lacking specific targeting, these agents represent the majority of clinical work using FGS to date (Table 1). The majority of these contrast agents, Methylene Blue (MB), ICG, and Fluorescein, have been used as far back as 1891 for the treatment of diseases like malaria and as colorimetric reporters for tissue perfusion and angiography.^{32,33} Initial use of these agents for FGS dates back to 1947, when fluorescein was used to guide brain tumor resection.³⁴ However, the agent used perhaps most frequently for FGS, ICG, has dominated the majority of clinical applications in recent years.

ICG is a water soluble tricyanocyanine fluorophore that possesses bright NIR fluorescence properties with excitation at 780 nm and emission at 820 nm.³⁵ The production and use of ICG dates back to 1955, when ICG was manufactured for use by Kodak Research Laboratories, and clinical approval dates back to 1956, when ICG was approved by the FDA for retinal angiography.³³ ICG clears rapidly from the body and possesses an excellent

safety profile, with LD₅₀s between 50–80 mg/kg in animals. ICG is currently used for sentinel lymph node mapping, angiography, reconstructive surgery, cholangiography, and tumor imaging among other uses for FGS and other clinical practices.^{16,35–38}

One FGS contrast agent was only recently approved, 5-ALA, and is unique among the clinically approved agents in that it provides targeted fluorescence imaging of cancer instead of providing contrast via passive mechanisms. 5-ALA is metabolized to the fluorescent porphyrin molecule PpIX following uptake in cancer cells, possessing visible fluorescence with excitation at 380–440 nm and emission at 620–640 nm.^{39,40} 5-ALA was first studied for FGS in 1998 and reached FDA approval in 2017.⁴¹ Remarkably, 5-ALA enabled FGS has improved high grade glioma complete resection rates almost two-fold over white light imaging alone, resulting in increased overall progression free survival.²³ These FGS contrast agents, while few in number and possessing suboptimal fluorescent and/or tissue uptake properties, have demonstrated the incredible utility of FGS to improve surgical outcomes. Looking forward, many new FGS contrast agents are under development and possess advanced targeting, physiochemical, and tissue uptake characteristics.

2. Novel Fluorescence Image-Guided Surgery Contrast Agents Under Clinical Development

At present, 39 novel FGS contrast agents are undergoing clinical trials in the United States. These agents account for 85 clinical trials registered in clinicaltrials.gov across a broad range of indications. These agents and their corresponding clinical trials are outlined in Table 2. These clinical trials represent a diverse field of researchers, clinicians, and industry partners undertaking the clinical translation of a variety of unique targeted FGS contrast agents. A number of probes utilize targeting moieties such as antibodies or peptides to obtain highly specific fluorescent signal.⁴² Others generate specific contrast via activatable fluorescence mechanisms using fluorophore quenching groups that are cleaved via enzymatic processes.⁴³ Others still possess structure inherent targeting mechanisms where the fluorophore itself acts as the targeting moiety and fluorescent reporter.^{44–47} This diversity in probe design is evident in the molecular type of each contrast agent (Fig. 1), where the majority of contrast agents are small molecule based, followed closely by peptide and antibody based contrast agents.

From this diverse group of contrast agents for FGS, a handful have completed significant development through phase II clinical trials on the pathway to FDA approval. The clinical trial phase of each probe is outlined in Fig. 2. Currently, SGM-101, BLZ-100, LUM015, and OTL38 have reached phase III clinical trials ([NCT03659448](https://clinicaltrials.gov/ct2/show/study/NCT03659448), [NCT03579602](https://clinicaltrials.gov/ct2/show/study/NCT03579602), [NCT03686215](https://clinicaltrials.gov/ct2/show/study/NCT03686215), and [NCT03180307](https://clinicaltrials.gov/ct2/show/study/NCT03180307), respectively). SGM-101 is a CEA targeting antibody labeled with a NIR fluorophore under phase III testing for colorectal neoplasms. SGM-101 has demonstrated a high degree of specificity in primary tumor tissue and metastases, with mean tumor to background ratios of 1.6 and 1.7, respectively.⁵⁰ BLZ-100 is composed of a chlorotoxin peptide covalently bound to ICG that demonstrates high affinity for matrix metalloproteinase 2 (MMP-2) under phase III testing for pediatric central nervous system (CNS) tumors. In phase I testing, BLZ-100 demonstrated low toxicity, with no observed

adverse events, and positive fluorescence signal in tumor tissue that was positively correlated with the dose of the imaging agent and grade of the cancer.⁷⁴ OTL38 is a small molecule probe consisting of folic acid, a ligand for folate receptor alpha, conjugated to the NIR fluorophore S0456 that is in phase III clinical studies for ovarian cancer. In phase II studies OTL38 identified cancer lesions with a sensitivity of 97.97%, where 48.3% of patients had at least one additional lesion identified using FGS over white light alone.¹¹⁰ These probes represent every major type of targeted FGS contrast agent and are leading the way for clinical translation of many others.

For all other FGS contrast agents under clinical translation, probes in phase I clinical trials make up the majority of clinical studies, followed closely by phase II studies (Fig. 2). One option for easing the regulatory burden for first-in-human studies of new agents is to utilize the food and drug administration's (FDA's) exploratory investigational new drug (eIND) pathway. First in human (FIH) studies conducted under an eIND require significantly less preclinical toxicology testing by allowing researchers to administer "microdoses" denoted as less than 100 μ g or 30 nmol per administration for small molecule or protein products, respectively.¹¹³ Due to the lower administered dose, substantially fewer preclinical toxicology studies are required prior to FIH clinical trials as compared to traditional translation under an IND, allowing proof of concept phase 0 studies to be performed with relative ease and significantly decreased financial burden. This alternative route to clinical use has been utilized with success in the recent phase 0 clinical trials of ABY-029, a promising tumor targeted FGS affibody probe, BBN-IRDye800CW, a dual modality PET/FGS contrast agent for brain cancer resection, fluorescein conjugated wisteria floribunda (WFA), a topically applied lectin for colon cancer resection, and EC17, a small molecule folate-FITC conjugate targeting folate receptor for renal carcinoma resection.^{66,67} The eIND pathway offers a promising alternative for ease of clinical translation for new FGS imaging probes, decreasing the financial burden to obtaining FIH results.

3. Fluorescence Spectral Properties and Compatibility with Clinical Imaging Systems

Additional diversity among the novel FGS contrast agents undergoing clinical translation is found in analysis of the fluorescence reporters' spectral properties (Fig. 3). A number of fluorophores have been utilized to provide contrast for each agent and those with reported excitation and emission are graphed in Fig. 3A along with the number of probes using each fluorophore. With fluorescence centered around 800 nm, IRDye800CW is the most often utilized fluorophore, providing contrast for more than double the number of probes than any other fluorophore. Examining the distribution of all 39 probes among the respective fluorescence NIR imaging channels (700 and 800 nm) and visible channel, just over half (20 probes) utilize fluorophores with excitation and emission wavelengths centered around 800 nm, with 9 probes using fluorophores with excitation and emission wavelengths centered around 700 nm and 10 probes using fluorophores with excitation and/or emission wavelengths in the visible range (Fig. 3B).

Assessing the compatibility of these fluorophores with clinical imaging systems requires a review of the currently clinically approved fluorescence imaging systems. Table 3 outlines the FGS imaging systems that are FDA approved or under development/clinical translation for human use. Notably, all FDA approved imaging systems are developed with fluorescence imaging capabilities compatible with the current FDA approved FGS contrast agents. Thus, the majority of systems are tuned for ICG fluorescence imaging, with excitation and emission wavelengths centered around 800 nm. Few FDA approved imaging systems possess capabilities for imaging fluorescein and PpIX fluorescence, and only two approved systems, the Fluobeam and Quest Spectrum, possess capabilities for imaging MB in the 700 nm channel. Thus, there exists a gap between novel FGS contrast agents' fluorescence properties and the imaging capabilities of clinically approved FGS imaging systems, where those contrast agents outside the 800 nm channel are lacking adequate options for spectrally tuned imaging systems. This gap exists likely due to the regulatory process for FGS imaging systems, where the majority of systems have achieved 510(k) clearance via predicate devices by showing substantial equivalence with already approved systems, most often the Stryker SPY imaging system that first obtained approval in 2005. However, a number of imaging systems are under development preclinically that could provide imaging capabilities for a wide range of fluorophores and incentive for approval of these systems will grow as novel FGS contrast agents with fluorescence outside of the 800 nm channel reach FDA approval (Table 3). For example, of the four novel FGS contrast agents in phase III clinical trials, two possess fluorescence properties outside the 800 nm channel including SGM-101, which uses BM104 as its fluorescent reporter with excitation and emission in the 700 nm channel, and LUM015, which utilizes Cy5 as its fluorescent reporter with excitation and emission in visible/NIR 700 nm channel.

4. Clinical Indications of Novel Fluorescence-Guided Surgery Contrast Agents

Analysis of the clinical indication studied in all clinical trials using novel FGS contrast agent reveals an overwhelming majority of agents targeted to cancer for enhanced detection and/or resection. 75, or 88%, of all novel FGS contrast agent clinical trials are underway for cancer related indications, while 6, or 7%, are underway for other disease related indications and only 4, or 5%, are underway for enhanced anatomical preservation. Thus, while great progress has been made in the field of FGS contrast agent development for improved cancer detection and resection, there remains a need for surgical treatment of non-cancer diseases and preservation of normal tissue function that can be enhanced using FGS. Interestingly, some cancer targeting agents have been employed in the treatment of other diseases, such as hyperparathyroidism (EC17, [NCT01996072](#)), carotid plaque instability (Bevacizumab-IRDye800CW, [NCT03757507](#)), or rheumatoid arthritis (OTL38, [NCT03938701](#)). Continued expansion of the many novel cancer targeted FGS contrast agents for other diseases could provide an excellent foundation for increasing the impact of FGS outside surgical oncology. Developing probes targeted to important anatomical structures, such as ureters or nerves, however, requires separate development efforts to identify targeting moieties for these non-diseased and intact tissues. Although, such development efforts possess a strong value and are worth undertaking, as injury to non-diseased tissues is responsible for a plethora of

surgical comorbidities that plague patient outcomes and present an enormous cost to the healthcare system. For instance, intraoperative nerve damage affects up to 63 million patients worldwide annually, causing pain or loss of function and significantly affecting quality of life.^{114,115} These rates remain high despite efforts to improve nerve sparing through complex surgical techniques and nerve detection technologies in procedures that have a high incidence of injury.^{115–123} Due to this need, several classes of nerve specific fluorophores have been studied for FGS preclinical.^{10,124–134} Further clinical translation of these and other anatomy targeted FGS contrast agents will improve surgical outcomes overall and could be used in synergy with cancer or disease specific agents to comprehensively benefit surgical goals.

5. Conclusion

Remarkable progress has been made in the past several decades not only in the field of optical imaging and biophotonics, but in the fields of fluorophore chemistry and FGS contrast agent development. An array of novel FGS contrast agents are under clinical translation, and many more are in preclinical development, providing evidence for a rapidly expanding field that is poised to significantly affect the surgical practice. There is an incredible diversity of FGS contrast agent molecular type, targeting mechanism, and fluorescence properties, owing to the efforts of a diverse field of physicists, chemists, clinicians, biologists, and engineers. Continued development towards clinical approval of these novel FGS contrast agents and FIH clinical studies of non-cancerous disease specific and anatomy specific FGS contrast agents will enable significant advancement in the field of surgery as a whole and ultimately improve patient outcomes across many surgical specialties.

Acknowledgements

This work was funded by the National Institute of Biomedical Imaging and Bioengineering (R01EB021362).

References

1. Sheridan RL, et al. Burn depth estimation by use of indocyanine green fluorescence: initial human trial. *J Burn Care Rehabil* 16, 602–604 (1995). [PubMed: 8582937]
2. Hongo K, Kobayashi S, Okudera H, Hokama M & Nakagawa F Noninvasive cerebral optical spectroscopy: depth-resolved measurements of cerebral haemodynamics using indocyanine green. *Neurol Res* 17, 89–93 (1995). [PubMed: 7609855]
3. Scheider A [Indocyanine green angiography with an infrared scanning laser ophthalmoscope. Initial clinical experiences]. *Ophthalmologie* 89, 27–33 (1992). [PubMed: 1581689]
4. Arens C, Malzahn K, Dias O, Andrea M & Glanz H [Endoscopic imaging techniques in the diagnosis of laryngeal carcinoma and its precursor lesions]. *Laryngorhinootologie* 78, 685–691 (1999). [PubMed: 10666695]
5. Lee BT, et al. The FLARE intraoperative near-infrared fluorescence imaging system: a first-in-human clinical trial in perforator flap breast reconstruction. *Plastic and reconstructive surgery* 126, 1472–1481 (2010). [PubMed: 21042103]
6. Tummers QR, et al. Real-time intraoperative detection of breast cancer using near-infrared fluorescence imaging and Methylene Blue. *European journal of surgical oncology: the journal of the European Society of Surgical Oncology and the British Association of Surgical Oncology* 40, 850–858 (2014).

7. Troyan SL, et al. The FLARE intraoperative near-infrared fluorescence imaging system: a first-in-human clinical trial in breast cancer sentinel lymph node mapping. *Annals of surgical oncology* 16, 2943–2952 (2009). [PubMed: 19582506]
8. Ashitate Y, Stockdale A, Choi HS, Laurence RG & Frangioni JV Real-time simultaneous near-infrared fluorescence imaging of bile duct and arterial anatomy. *The Journal of surgical research* 176, 7–13 (2012). [PubMed: 21816414]
9. Verbeek FP, et al. Intraoperative near infrared fluorescence guided identification of the ureters using low dose methylene blue: a first in human experience. *The Journal of urology* 190, 574–579 (2013). [PubMed: 23466242]
10. Gibbs-Strauss SL, et al. Nerve-highlighting fluorescent contrast agents for image-guided surgery. *Molecular imaging* 10, 91–101 (2011). [PubMed: 21439254]
11. Hirche C, et al. An experimental study to evaluate the Fluobeam 800 imaging system for fluorescence-guided lymphatic imaging and sentinel node biopsy. *Surgical innovation* 20, 516–523 (2013). [PubMed: 23275469]
12. Gotoh K, et al. A novel image-guided surgery of hepatocellular carcinoma by indocyanine green fluorescence imaging navigation. *Journal of surgical oncology* 100, 75–79 (2009). [PubMed: 19301311]
13. Kitagawa H, et al. Visualization of the Stomach's Arterial Networks During Esophageal Surgery Using the HyperEye Medical System. *Anticancer research* 35, 6201–6205 (2015). [PubMed: 26504051]
14. Weissleder R & Pittet MJ Imaging in the era of molecular oncology. *Nature* 452, 580–589 (2008). [PubMed: 18385732]
15. Frangioni JV New technologies for human cancer imaging. *Journal of clinical oncology: official journal of the American Society of Clinical Oncology* 26, 4012–4021 (2008). [PubMed: 18711192]
16. Kitai T, Inomoto T, Miwa M & Shikayama T Fluorescence navigation with indocyanine green for detecting sentinel lymph nodes in breast cancer. *Breast Cancer* 12, 211–215 (2005). [PubMed: 16110291]
17. Peek MC, Charalampoudis P, Anninga B, Baker R & Douek M Blue dye for identification of sentinel nodes in breast cancer and malignant melanoma: a systematic review and meta-analysis. *Future Oncol* 13, 455–467 (2017). [PubMed: 27578614]
18. Jeschke S, et al. Visualisation of the lymph node pathway in real time by laparoscopic radioisotope- and fluorescence-guided sentinel lymph node dissection in prostate cancer staging. *Urology* 80, 1080–1086 (2012). [PubMed: 22990053]
19. Chang DW, Suami H & Skoracki R A prospective analysis of 100 consecutive lymphovenous bypass cases for treatment of extremity lymphedema. *Plastic and reconstructive surgery* 132, 1305–1314 (2013). [PubMed: 24165613]
20. Yamamoto M, et al. Indocyanine green angiography for intra-operative assessment in vascular surgery. *Eur J Vasc Endovasc Surg* 43, 426–432 (2012). [PubMed: 22264591]
21. Boni L, et al. Clinical applications of indocyanine green (ICG) enhanced fluorescence in laparoscopic surgery. *Surg Endosc* 29, 2046–2055 (2015). [PubMed: 25303914]
22. Stummer W, et al. Intraoperative detection of malignant gliomas by 5-aminolevulinic acid-induced porphyrin fluorescence. *Neurosurgery* 42, 518–525; discussion 525–516 (1998). [PubMed: 9526986]
23. Stummer W, et al. Fluorescence-guided surgery with 5-aminolevulinic acid for resection of malignant glioma: a randomised controlled multicentre phase III trial. *Lancet Oncol* 7, 392–401 (2006). [PubMed: 16648043]
24. van Dam GM, et al. Intraoperative tumor-specific fluorescence imaging in ovarian cancer by folate receptor- α targeting: first in-human results. *Nat Med* 17, 1315–1319 (2011). [PubMed: 21926976]
25. van der Vorst JR, et al. Near-infrared fluorescence imaging of a solitary fibrous tumor of the pancreas using methylene blue. *World J Gastrointest Surg* 4, 180–184 (2012). [PubMed: 22905287]

26. Al-Taher M, van den Bos J, Schols RM, Bouvy ND & Stassen LP Fluorescence Ureteral Visualization in Human Laparoscopic Colorectal Surgery Using Methylene Blue. *J Laparoendosc Adv Surg Tech A* 26, 870–875 (2016). [PubMed: 27575463]
27. Ankersmit M, et al. Fluorescent Imaging With Indocyanine Green During Laparoscopic Cholecystectomy in Patients at Increased Risk of Bile Duct Injury. *Surgical innovation* 24, 245–252 (2017). [PubMed: 28178882]
28. Samkoe KS, et al. Application of Fluorescence-Guided Surgery to Subsurface Cancers Requiring Wide Local Excision: Literature Review and Novel Developments Toward Indirect Visualization. *Cancer Control* 25, 1073274817752332 (2018). [PubMed: 29334791]
29. Hadjipanayis CG, Widhalm G & Stummer W What is the Surgical Benefit of Utilizing 5-Aminolevulinic Acid for Fluorescence-Guided Surgery of Malignant Gliomas? *Neurosurgery* 77, 663–673 (2015). [PubMed: 26308630]
30. Rosenthal EL, et al. Successful Translation of Fluorescence Navigation During Oncologic Surgery: A Consensus Report. *J Nucl Med* 57, 144–150 (2016). [PubMed: 26449839]
31. Pogue BW, et al. Advancing Molecular-Guided Surgery through probe development and testing in a moderate cost evaluation pipeline. *Proc SPIE Int Soc Opt Eng* 9311(2015).
32. Schirmer RH, Adler H, Pickhardt M & Mandelkow E “Lest we forget you — methylene blue ...”. *Neurobiology of aging* 32, 2325.e2327–2325.e2316 (2011).
33. Nagaya T, Nakamura YA, Choyke PL & Kobayashi H Fluorescence-Guided Surgery. *Frontiers in Oncology* 7(2017).
34. Moore GE Fluorescein as an Agent in the Differentiation of Normal and Malignant Tissues. *Science* 106, 130–131 (1947).
35. Alander JT, et al. A Review of Indocyanine Green Fluorescent Imaging in Surgery. *International Journal of Biomedical Imaging* 2012, 1–26 (2012).
36. Lamby P, et al. Evaluation of the vascular integrity of free flaps based on microcirculation imaging techniques. *Clinical Hemorheology and Microcirculation* 39, 253–263 (2008). [PubMed: 18503134]
37. Newman M & Samson M The Application of Laser-Assisted Indocyanine Green Fluorescent Dye Angiography in Microsurgical Breast Reconstruction. *Journal of Reconstructive Microsurgery* 25, 021–026 (2008).
38. Cho SS, Salinas R & Lee JYK Indocyanine-Green for Fluorescence-Guided Surgery of Brain Tumors: Evidence, Techniques, and Practical Experience. *Frontiers in Surgery* 6(2019).
39. Widhalm G, et al. The value of visible 5-ALA fluorescence and quantitative protoporphyrin IX analysis for improved surgery of suspected low-grade gliomas. *Journal of Neurosurgery*, 1–10 (2019).
40. Stepp H & Stummer W 5-ALA in the management of malignant glioma. *Lasers in Surgery and Medicine* 50, 399–419 (2018). [PubMed: 29737540]
41. Walter S, et al. Intraoperative Detection of Malignant Gliomas by 5-Aminolevulinic Acid-induced Porphyrin Fluorescence. *Neurosurgery* 42, 518–526 (1998). [PubMed: 9526986]
42. Staderini M, Megia-Fernandez A, Dhaliwal K & Bradley M Peptides for optical medical imaging and steps towards therapy. *Bioorganic & medicinal chemistry* 26, 2816–2826 (2018). [PubMed: 29042225]
43. Mochida A, Ogata F, Nagaya T, Choyke PL & Kobayashi H Activatable fluorescent probes in fluorescence-guided surgery: Practical considerations. *Bioorganic & medicinal chemistry* 26, 925–930 (2018). [PubMed: 29242021]
44. Farnam RW, Arms RG, Klaassen AH & Sorger JM Intraoperative ureter visualization using a near-infrared imaging agent. *Journal of biomedical optics* 24(2019).
45. Munhenzva IR, et al. Assessment of human pancreas cancer tissue and precursor lesions via a fluorophore with inherent PDAC selectivity. *Methods* 168, 35–39 (2019). [PubMed: 31185273]
46. Shieh J-J, Rochelle Riley I & Dorshow RB Clinical analysis and quantitation of MB-102, a novel fluorescence tracer agent, in human plasma. *Analytical Methods* 10, 2376–2383 (2018).
47. Achilefu S, et al. Modeling of transdermal fluorescence measurements from first-in-human clinical trials for renal function determination using fluorescent tracer agent MB-102. in *Reporters, Markers, Dyes, Nanoparticles, and Molecular Probes for Biomedical Applications IX* (2017).

48. Gutowski M, et al. SGM-101: An innovative near-infrared dye-antibody conjugate that targets CEA for fluorescence-guided surgery. *Surgical Oncology* 26, 153–162 (2017). [PubMed: 28577721]
49. Boogerd LSF, et al. Safety and effectiveness of SGM-101, a fluorescent antibody targeting carcinoembryonic antigen, for intraoperative detection of colorectal cancer: a dose-escalation pilot study. *The Lancet Gastroenterology & Hepatology* 3, 181–191 (2018). [PubMed: 29361435]
50. Hoogstins CES, et al. Image-Guided Surgery in Patients with Pancreatic Cancer: First Results of a Clinical Trial Using SGM-101, a Novel Carcinoembryonic Antigen-Targeting, Near-Infrared Fluorescent Agent. *Annals of surgical oncology* 25, 3350–3357 (2018). [PubMed: 30051369]
51. Heath CH, Deep NL, Sweeny L, Zinn KR & Rosenthal EL Use of Panitumumab-IRDye800 to Image Microscopic Head and Neck Cancer in an Orthotopic Surgical Model. *Annals of surgical oncology* 19, 3879–3887 (2012). [PubMed: 22669455]
52. Korb ML, et al. Use of monoclonal antibody-IRDye800CW bioconjugates in the resection of breast cancer. *Journal of Surgical Research* 188, 119–128 (2014). [PubMed: 24360117]
53. Gao RW, et al. Safety of panitumumab-IRDye800CW and cetuximab-IRDye800CW for fluorescence-guided surgical navigation in head and neck cancers. *Theranostics* 8, 2488–2495 (2018). [PubMed: 29721094]
54. Marston JC, et al. Panitumumab-IRDye800CW for Fluorescence-Guided Surgical Resection of Colorectal Cancer. *Journal of Surgical Research* 239, 44–51 (2019). [PubMed: 30798171]
55. Day KE, Sweeny L, Kulbersh B, Zinn KR & Rosenthal EL Preclinical Comparison of Near-Infrared-Labeled Cetuximab and Panitumumab for Optical Imaging of Head and Neck Squamous Cell Carcinoma. *Molecular Imaging and Biology* 15, 722–729 (2013). [PubMed: 23715932]
56. Warram JM, et al. Fluorescence-guided resection of experimental malignant glioma using cetuximab-IRDye 800CW. *British Journal of Neurosurgery* 29, 850–858 (2015). [PubMed: 26073144]
57. Rosenthal EL, et al. Safety and Tumor Specificity of Cetuximab-IRDye800 for Surgical Navigation in Head and Neck Cancer. *Clinical Cancer Research* 21, 3658–3666 (2015). [PubMed: 25904751]
58. Rosenthal EL, et al. Sensitivity and Specificity of Cetuximab-IRDye800CW to Identify Regional Metastatic Disease in Head and Neck Cancer. *Clinical Cancer Research* 23, 4744–4752 (2017). [PubMed: 28446503]
59. Tummers WS, et al. Intraoperative Pancreatic Cancer Detection using Tumor-Specific Multimodality Molecular Imaging. *Annals of surgical oncology* 25, 1880–1888 (2018). [PubMed: 29667116]
60. Terwisscha van Scheltinga AGT, et al. Intraoperative Near-Infrared Fluorescence Tumor Imaging with Vascular Endothelial Growth Factor and Human Epidermal Growth Factor Receptor 2 Targeting Antibodies. *Journal of Nuclear Medicine* 52, 1778–1785 (2011). [PubMed: 21990576]
61. Harlaar NJ, et al. Molecular fluorescence-guided surgery of peritoneal carcinomatosis of colorectal origin: a single-centre feasibility study. *The Lancet Gastroenterology & Hepatology* 1, 283–290 (2016). [PubMed: 28404198]
62. Lamberts LE, et al. Tumor-Specific Uptake of Fluorescent Bevacizumab-IRDye800CW Microdosing in Patients with Primary Breast Cancer: A Phase I Feasibility Study. *Clinical Cancer Research* 23, 2730–2741 (2017). [PubMed: 28119364]
63. Hekman MCH, et al. Detection of Micrometastases Using SPECT/Fluorescence Dual-Modality Imaging in a CEA-Expressing Tumor Model. *J Nucl Med* 58, 706–710 (2017). [PubMed: 28126888]
64. Hekman MC, et al. Tumor-targeted Dual-modality Imaging to Improve Intraoperative Visualization of Clear Cell Renal Cell Carcinoma: A First in Man Study. *Theranostics* 8, 2161–2170 (2018). [PubMed: 29721070]
65. Samkoe KS, et al. Measuring microdose ABY-029 fluorescence signal in a primary human soft-tissue sarcoma resection. *Proc SPIE Int Soc Opt Eng* 10862(2019).
66. Samkoe KS, et al. Toxicity and Pharmacokinetic Profile for Single-Dose Injection of ABY-029: a Fluorescent Anti-EGFR Synthetic Affibody Molecule for Human Use. *Mol Imaging Biol* 19, 512–521 (2017). [PubMed: 27909986]

67. Elliott JT, et al. Microdose fluorescence imaging of ABY-029 on an operating microscope adapted by custom illumination and imaging modules. *Biomed Opt Express* 7, 3280–3288 (2016). [PubMed: 27699098]
68. Veiseh M, et al. Tumor Paint: A Chlorotoxin: Cy5.5 Bioconjugate for Intraoperative Visualization of Cancer Foci. *Cancer Research* 67, 6882–6888 (2007). [PubMed: 17638899]
69. Butte PV, et al. Near-infrared imaging of brain tumors using the Tumor Paint BLZ-100 to achieve near-complete resection of brain tumors. *Neurosurgical Focus* 36(2014).
70. Kittle DS, et al. Fluorescence-guided Tumor Visualization Using the Tumor Paint BLZ-100. *Cureus* (2014).
71. Fidel J, et al. Preclinical Validation of the Utility of BLZ-100 in Providing Fluorescence Contrast for Imaging Spontaneous Solid Tumors. *Cancer Research* 75, 4283–4291 (2015). [PubMed: 26471914]
72. Baik FM, et al. Fluorescence Identification of Head and Neck Squamous Cell Carcinoma and High-Risk Oral Dysplasia With BLZ-100, a Chlorotoxin-Indocyanine Green Conjugate. *JAMA Otolaryngology–Head & Neck Surgery* 142(2016).
73. Dintzis SM, et al. Real-time Visualization of Breast Carcinoma in Pathology Specimens From Patients Receiving Fluorescent Tumor-Marking Agent Tozuleristide. *Archives of Pathology & Laboratory Medicine* 143, 1076–1083 (2019). [PubMed: 30550350]
74. Patil CG, et al. Phase 1 Safety, Pharmacokinetics, and Fluorescence Imaging Study of Tozuleristide (BLZ-100) in Adults With Newly Diagnosed or Recurrent Gliomas. *Neurosurgery* 85, E641–E649 (2019). [PubMed: 31069381]
75. Miampamba M, et al. Sensitive in vivo Visualization of Breast Cancer Using Ratiometric Protease-activatable Fluorescent Imaging Agent, AVB-620. *Theranostics* 7, 3369–3386 (2017). [PubMed: 28900516]
76. Unkart JT, et al. Intraoperative Tumor Detection Using a Ratiometric Activatable Fluorescent Peptide: A First-in-Human Phase 1 Study. *Annals of surgical oncology* 24, 3167–3173 (2017). [PubMed: 28699134]
77. Burggraaf J, et al. Detection of colorectal polyps in humans using an intravenously administered fluorescent peptide targeted against c-Met. *Nature Medicine* 21, 955–961 (2015).
78. Li Z, et al. In vivo fluorescence imaging of hepatocellular carcinoma xenograft using near-infrared labeled epidermal growth factor receptor (EGFR) peptide. *Biomedical Optics Express* 7(2016).
79. Chen J, Gao Z, Li G & Wang TD Dual-modal in vivo fluorescence and photoacoustic imaging using a heterodimeric peptide. *Chemical Communications* 54, 13196–13199 (2018). [PubMed: 30334022]
80. Joshi BP, et al. Design and Synthesis of Near-Infrared Peptide for in Vivo Molecular Imaging of HER2. *Bioconjugate Chemistry* 27, 481–494 (2015). [PubMed: 26709709]
81. Joshi BP, et al. Detection of Sessile Serrated Adenomas in the Proximal Colon Using Wide-Field Fluorescence Endoscopy. *Gastroenterology* 152, 1002–1013.e1009 (2017). [PubMed: 28012848]
82. Jones JE, Busi SB, Mitchem JB, Amos-Landgraf JM & Lewis MR Evaluation of a Tumor-Targeting, Near-Infrared Fluorescent Peptide for Early Detection and Endoscopic Resection of Polyps in a Rat Model of Colorectal Cancer. *Molecular imaging* 17(2018).
83. Achilefu S, et al. Synthesis, In Vitro Receptor Binding, and In Vivo Evaluation of Fluorescein and Carbocyanine Peptide-Based Optical Contrast Agents. *Journal of medicinal chemistry* 45, 2003–2015 (2002). [PubMed: 11985468]
84. Huang R, et al. Integrin α_3 -Targeted IRDye 800CW Near-Infrared Imaging of Glioblastoma. *Clinical Cancer Research* 18, 5731–5740 (2012). [PubMed: 22914772]
85. Verbeek FPR, et al. Near-Infrared Fluorescence Imaging of Both Colorectal Cancer and Ureters Using a Low-Dose Integrin Targeted Probe. *Annals of surgical oncology* 21, 528–537 (2014).
86. Cheng H, et al. Precise integrin-targeting near-infrared imaging-guided surgical method increases surgical qualification of peritoneal carcinomatosis from gastric cancer in mice. *Oncotarget* 8(2016).
87. Handgraaf HJM, et al. Real-time near-infrared fluorescence imaging using cRGD-ZW800–1 for intraoperative visualization of multiple cancer types. *Oncotarget* 8(2017).

88. Liu Y, et al. Near-Infrared Fluorescent Peptides with High Tumor Selectivity: Novel Probes for Image-Guided Surgical Resection of Orthotopic Glioma. *Molecular pharmaceutics* 16, 108–117 (2018). [PubMed: 30517013]
89. Sun Y, et al. Novel dual-function near-infrared II fluorescence and PET probe for tumor delineation and image-guided surgery. *Chem Sci* 9, 2092–2097 (2018). [PubMed: 29675250]
90. Choi HS, et al. Targeted zwitterionic near-infrared fluorophores for improved optical imaging. *Nature biotechnology* 31, 148–153 (2013).
91. Deng H, Wang H, Wang M, Li Z & Wu Z Synthesis and Evaluation of ⁶⁴Cu-DOTA-NT-Cy5.5 as a Dual-Modality PET/Fluorescence Probe to Image Neurotensin Receptor-Positive Tumor. *Molecular pharmaceutics* 12, 3054–3061 (2015). [PubMed: 26162008]
92. Witjes M, et al. Fluorescence Guided Surgery Using the pH-Activated Micellar Tracer ONM-100: First-In-Human Proof of Principle in Head and Neck Squamous Cell Carcinoma. *Journal of Oral and Maxillofacial Surgery* 77(2019).
93. Wirth D, Smith TW, Moser R & Yaroslavsky AN Demeclocycline as a contrast agent for detecting brain neoplasms using confocal microscopy. *Physics in Medicine and Biology* 60, 3003–3011 (2015). [PubMed: 25790138]
94. van den Bos J, Al-Taher M, Bouvy ND & Stassen LPS Near-infrared fluorescence laparoscopy of the ureter with three preclinical dyes in a pig model. *Surgical Endoscopy* 33, 986–991 (2018). [PubMed: 30478696]
95. Whitley MJ, et al. A mouse-human phase 1 co-clinical trial of a protease-activated fluorescent probe for imaging cancer. *Science Translational Medicine* 8, 320ra324–320ra324 (2016).
96. Smith BL, et al. Real-time, intraoperative detection of residual breast cancer in lumpectomy cavity walls using a novel cathepsin-activated fluorescent imaging system. *Breast Cancer Research and Treatment* 171, 413–420 (2018). [PubMed: 29948401]
97. Cuneo KC, et al. Imaging Primary Mouse Sarcomas After Radiation Therapy Using Cathepsin-Activatable Fluorescent Imaging Agents. *International Journal of Radiation Oncology*Biophysics* 86, 136–142 (2013).
98. Irwin CP, et al. PARPi-FL - a Fluorescent PARP1 Inhibitor for Glioblastoma Imaging. *Neoplasia* 16, 432–440 (2014). [PubMed: 24970386]
99. Kossatz S, et al. Detection and delineation of oral cancer with a PARP1 targeted optical imaging agent. *Scientific Reports* 6(2016).
100. Sobol RW, Kossatz S, Weber WA & Reiner T Optical Imaging of PARP1 in Response to Radiation in Oral Squamous Cell Carcinoma. *PloS one* 11(2016).
101. Barrott Jared J., et al. Optical and Radioiodinated Tethered Hsp90 Inhibitors Reveal Selective Internalization of Ectopic Hsp90 in Malignant Breast Tumor Cells. *Chemistry & Biology* 20, 1187–1197 (2013). [PubMed: 24035283]
102. Crouch BT, et al. Exploiting heat shock protein expression to develop a non-invasive diagnostic tool for breast cancer. *Scientific Reports* 9(2019).
103. van Dam GM, et al. Intraoperative tumor-specific fluorescence imaging in ovarian cancer by folate receptor- α targeting: first in-human results. *Nature Medicine* 17, 1315–1319 (2011).
104. Guzzo TJ, et al. Intraoperative Molecular Diagnostic Imaging Can Identify Renal Cell Carcinoma. *Journal of Urology* 195, 748–755 (2016). [PubMed: 26541080]
105. Tummers QRJG, et al. Intraoperative imaging of folate receptor alpha positive ovarian and breast cancer using the tumor specific agent EC17. *Oncotarget* 7(2016).
106. Hoogstins CES, et al. A Novel Tumor-Specific Agent for Intraoperative Near-Infrared Fluorescence Imaging: A Translational Study in Healthy Volunteers and Patients with Ovarian Cancer. *Clinical Cancer Research* 22, 2929–2938 (2016). [PubMed: 27306792]
107. Keating JJ, et al. Intraoperative near-infrared fluorescence imaging targeting folate receptors identifies lung cancer in a large-animal model. *Cancer* 123, 1051–1060 (2017). [PubMed: 28263385]
108. Mahalingam SM, et al. Evaluation of Novel Tumor-Targeted Near-Infrared Probe for Fluorescence-Guided Surgery of Cancer. *Journal of medicinal chemistry* 61, 9637–9646 (2018). [PubMed: 30296376]

109. Predina JD, et al. An open label trial of folate receptor-targeted intraoperative molecular imaging to localize pulmonary squamous cell carcinomas. *Oncotarget* 9(2018).
110. Randall LM, Wenham RM, Low PS, Dowdy SC & Tanyi JL A phase II, multicenter, open-label trial of OTL38 injection for the intra-operative imaging of folate receptor-alpha positive ovarian cancer. *Gynecologic Oncology* 155, 63–68 (2019). [PubMed: 31362825]
111. Thukkani AK & Jaffer FA Intravascular near-infrared fluorescence molecular imaging of atherosclerosis. *Am J Nucl Med Mol Imaging* 3, 217–231 (2013). [PubMed: 23638334]
112. Jaffer FA, et al. Two-Dimensional Intravascular Near-Infrared Fluorescence Molecular Imaging of Inflammation in Atherosclerosis and Stent-Induced Vascular Injury. *Journal of the American College of Cardiology* 57, 2516–2526 (2011). [PubMed: 21679853]
113. Yamashita S & Sugiyama Y New strategy for drug development with exploratory IND studies: scientific basis and future directions. *Advanced drug delivery reviews* 63, 493 (2011). [PubMed: 21601603]
114. Bouyer-Ferullo S Preventing perioperative peripheral nerve injuries. *AORN J* 97, 110–124 e119 (2013). [PubMed: 23265653]
115. Antoniadis G, et al. Iatrogenic nerve injuries: prevalence, diagnosis and treatment. *Dtsch Arztebl Int* 111, 273–279 (2014). [PubMed: 24791754]
116. Ficarra V, et al. Systematic review and meta-analysis of studies reporting urinary continence recovery after robot-assisted radical prostatectomy. *Eur Urol* 62, 405–417 (2012). [PubMed: 22749852]
117. Ghavamian R, Knoll A, Boczek J & Melman A Comparison of operative and functional outcomes of laparoscopic radical prostatectomy and radical retropubic prostatectomy: single surgeon experience. *Urology* 67, 1241–1246 (2006). [PubMed: 16678887]
118. Lopez A, et al. Intraoperative Optical Biopsy during Robotic Assisted Radical Prostatectomy Using Confocal Endomicroscopy. *The Journal of urology* 195, 1110–1117 (2016). [PubMed: 26626214]
119. Boyette LB, et al. Fiberoptic imaging of cavernous nerves in vivo. *The Journal of urology* 178, 2694–2700 (2007). [PubMed: 17945279]
120. Rais-Bahrami S, et al. Optical coherence tomography of cavernous nerves: a step toward real-time intraoperative imaging during nerve-sparing radical prostatectomy. *Urology* 72, 198–204 (2008). [PubMed: 18280549]
121. Aron M, et al. Second prize: preliminary experience with the Niris optical coherence tomography system during laparoscopic and robotic prostatectomy. *J Endourol* 21, 814–818 (2007). [PubMed: 17867934]
122. Ukimura O, Magi-Galluzzi C & Gill IS Real-time transrectal ultrasound guidance during laparoscopic radical prostatectomy: impact on surgical margins. *The Journal of urology* 175, 1304–1310 (2006). [PubMed: 16515987]
123. Ponnusamy K, Sorger JM & Mohr C Nerve mapping for prostatectomies: novel technologies under development. *J Endourol* 26, 769–777 (2012). [PubMed: 22142311]
124. Wu C, et al. Molecular probes for imaging myelinated white matter in CNS. *Journal of medicinal chemistry* 51, 6682–6688 (2008). [PubMed: 18844339]
125. Wang C, et al. In situ fluorescence imaging of myelination. *The journal of histochemistry and cytochemistry: official journal of the Histochemistry Society* 58, 611–621 (2010). [PubMed: 20354147]
126. Gibbs SL, et al. Structure-activity relationship of nerve-highlighting fluorophores. *PloS one* 8, e73493 (2013). [PubMed: 24039960]
127. Stankoff B, et al. Imaging of CNS myelin by positron-emission tomography. *Proceedings of the National Academy of Sciences of the United States of America* 103, 9304–9309 (2006). [PubMed: 16754874]
128. Cotero VE, et al. Intraoperative fluorescence imaging of peripheral and central nerves through a myelin-selective contrast agent. *Molecular imaging and biology: MIB: the official publication of the Academy of Molecular Imaging* 14, 708–717 (2012). [PubMed: 22488576]

129. Cotero VE, et al. Improved Intraoperative Visualization of Nerves through a Myelin-Binding Fluorophore and Dual-Mode Laparoscopic Imaging. *PloS one* 10, e0130276 (2015). [PubMed: 26076448]
130. Bajaj A, et al. Identification of the protein target of myelin-binding ligands by immunohistochemistry and biochemical analyses. *The journal of histochemistry and cytochemistry: official journal of the Histochemistry Society* 61, 19–30 (2013). [PubMed: 23092790]
131. Gibbs-Strauss SL, et al. Molecular imaging agents specific for the annulus fibrosus of the intervertebral disk. *Molecular imaging* 9, 128–140 (2010). [PubMed: 20487679]
132. Meyers JR, et al. Lighting up the senses: FM1–43 loading of sensory cells through nonselective ion channels. *The Journal of neuroscience: the official journal of the Society for Neuroscience* 23, 4054–4065 (2003). [PubMed: 12764092]
133. Wang C, et al. Longitudinal near-infrared imaging of myelination. *The Journal of neuroscience: the official journal of the Society for Neuroscience* 31, 2382–2390 (2011). [PubMed: 21325505]
134. Park MH, et al. Prototype nerve-specific near-infrared fluorophores. *Theranostics* 4, 823–833 (2014). [PubMed: 24955143]

Figure 1.

A pie chart of novel FGS contrast agent currently undergoing clinical translation split by probe type. The number of each agent type is listed next to the type.

Figure 2.
A pie chart highlighting the clinical trial phase of the novel FGS contrast agents currently undergoing clinical translation. The number of agents in each phase is listed.

Figure 3. A.

Fluorescence excitation and emission wavelengths and **B.** imaging channel distribution of all fluorophores utilized by all novel FGS contrast agents undergoing clinical translation. The number of probes using each fluorophore are listed next to their names and the number of fluorophores in each channel are listed as well as the percentage of the total that number represents.

Figure 4.

A pie chart highlighting the broad clinical indication of the novel FGS contrast agents currently undergoing clinical translation. The number of agents targeted for each indication is listed next to it.

Table 1.

FDA-approved FGS contrast agents.

Name	Description	Chemical Class	Excitation Wavelength	Emission Wavelength
Fluorescein	Visible fluorophore often used for angiography and glioma resection. No specific targeting mechanism	Xanthene	494 nm	521 nm
Methylene Blue (MB)	Near-infrared fluorophore used in many FGS applications. No specific targeting mechanism.	Thiazine	668 nm	688 nm
Indocyanine Green (ICG)	Near-infrared fluorophore with the broadest clinical adoption and use. No specific targeting mechanism	Cyanine	780 nm	820 nm
5-aminominolevulinic acid (5-ALA)	Pro-drug that is metabolized to the visible fluorophore protoporphyrin IX (PpIX) in cancer cells. Specifically targeted to cancer tissue.	Porphyrin	380–440 nm	620–640 nm

Table 2.

Novel FGS contrast agents under clinical development.

Probe Name	Description	Fluorophore	Company/ Group	Indications	Clinical Trial Phase	NCT#	Status
Antibody							
SGM-101 ^{48–50}	Monoclonal antibody to carcinoembryonic antigen (CEA) labelled with 700nm BM104fluorophore. IV administration 4 days before surgery	BM104	Surgimab	Colon Cancer	Phase 2	NCT02973672	completed
				Rectum Cancer	Phase 2		
				Pancreatic Cancer	Phase 2		
				Metastatic Colorectal Cancer	Phase 2		
				Recurrent Colorectal Carcinoma	Phase 2		
				Colorectal Neoplasms	Phase 3	NCT03659448	recruiting
				Peritoneal Carcinomatosis	Phase 1	NCT02784028	recruiting
Panitumumab-IRDye800CW ^{51–54}	EGFR targeting antibody conjugated with IRDye800. IV administered over 60min, 1–5 days prior to surgery	IRDye-800	Rosenthal-Stanford	Pediatric Neoplasms	Phase 2	NCT04085887	not yet recruiting
				Malignant Glioma	Phase 2	NCT03510208	recruiting
				Head and Neck Cancer	Phase 2	NCT03405142	recruiting
				Lung Cancer	Phase 2	NCT03582124	suspended (business decision)
				Pancreatic Cancer	Phase 2	NCT03384238	recruiting
Cetuximab-IRDye800CW ^{52,53,55–59}	EGFR targeting antibody conjugated with IRDye800. IV administered 4 days prior to surgery	IRDye-800	Rosenthal-Stanford	Head and Neck Cancer	Phase 2	NCT03134846	recruiting
				Pancreatic Cancer	Phase 2	NCT02736578	terminated (logistics)
				Esophageal Cancer	Phase 1	NCT04161560	recruiting
				Brain Cancer	Phase 2	NCT02855086	terminated (logistics)
Bevacizumab-IRDye800CW ^{52,60–62}	VEGF targeting antibody conjugated to IRDye800. IV administration 3 days prior to surgery at doses of 10, 25, or 50 mg	IRDye-800	van Dam-Groningen	Adenomatous Polyposis	Phase 1	NCT02113202	completed
				Rectal Cancer	Phase 1	NCT01972373	completed
				Breast Cancer	Phase 2	NCT02583568	completed
				Esophageal Cancer	Phase 2	NCT03877601	recruiting
				Hilar Cholangiocarcinoma	Phase 2	NCT03620292	recruiting
				Soft Tissue Sarcoma	Phase 2	NCT03913806	recruiting
				Pancreatic Cancer	Phase 2	NCT02743975	recruiting
				Endometriosis	Phase 1	NCT02975219	recruiting
				Inverted Papilloma	Phase 1	NCT03925285	recruiting
				Pituitary Adenoma	Phase 1	NCT04212793	not yet recruiting
				Carotid plaque instability	N/A	NCT03757507	not yet recruiting

Probe Name	Description	Fluorophore	Company/ Group	Indications	Clinical Trial Phase	NCT#	Status
Indium-111-DOTA-Labetuzumab-IRDye800CW ⁶³	CEA targeting antibody conjugated with dual modality (SPECT/CT and fluorescence) tracers. Administered 6–7 days before surgery.	IRDye-800	Boerman-Radboud University	Colorectal Cancer	Phase 2	NCT03699332	recruiting
Indium-111-DOTA-Girentuximab-IRDye800CW ⁶⁴	Carbonic anhydrase IX (CAIX) targeting antibody conjugated with dual modality (SPECT/CT and fluorescence) tracers. Administered 7 days before surgery.	IRDye-800	Boerman-Radboud University	Renal Carcinoma	Phase 1	NCT02497599	active, not recruiting
MDX1201-A488	PSMA targeting antibody conjugated with Alexa Fluor 488. IV administration 4 days prior to surgery	AF488	Zhumkhawala - City of Hope Medical	Prostate Cancer	Phase 1	NCT02048150	active, not recruiting
ProstaFluor	PSMA targeting antibody huJ-591 conjugated to IR-800CW	IRDye-800	Spectros	Prostate Cancer	Phase 1	NCT01173146	withdrawn - cost of antibody production increased, no supplemental funding obtained
Protein							
Fluorescein conjugated wisteria floribunda	Fluorescein conjugated lectin. Sprayed onto colonic surface during surgery	Fluorescein	Yeung - Oxford	Colorectal Cancer	Phase 0	NCT03070613	enrolling by invitation
Affibody							
ABY-029 ^{28,65–67}	EGFR targeting affibody labeled with IRDye800CW. Microdose injection 1–3 hours prior to surgery	IRDye-800	Pogue - Dartmouth	Glioma	Phase 0	NCT02901925	suspended - awaiting FDA approval of eIND amendment
				Primary soft tissue Sarcoma	Phase 0	NCT03154411	suspended - awaiting FDA approval of eIND amendment
				Head and Neck Cancer	Phase 0	NCT03282461	suspended - awaiting FDA approval of eIND amendment
Peptide							

Probe Name	Description	Fluorophore	Company/ Group	Indications	Clinical Trial Phase	NCT#	Status
BLZ-100 (tozuleristide) ⁶⁸⁻⁷⁴	Chlorotoxin (scorpion venom) with high affinity to matrix metalloproteinase 2 (MMP-2) conjugated with NIR fluorophore. IV administration at least 1 hour before surgery	ICG	Blaze Bioscience	Soft Tissue Sarcoma	Phase 1	NCT02464332	withdrawn - not enough subjects enrolled
				CNS Tumors	Phase 3	NCT03579602	recruiting
				Glioma	Phase 1	NCT02234297	completed
				Breast Cancer	Phase 1	NCT02496065	completed
				Skin Neoplasms	Phase 1	NCT02097875	completed
AVB-620 ^{75,76}	Ratiometric MMP activatable peptide labeled with cy5 and cy7. Upon activation, cy5 is cleaved, and cy5 fluorescence increases. IV administration up to 24 hr before	Cy5& Cy7 (cy5 detected)	Avelas Biosciences	Breast Cancer	Phase 2	NCT03113825	recruiting
BBN-IRDye800CW ⁷⁶	Peptide targeting gastrin-releasing peptide receptor (GRPR) labeled with IRDye 800 and radiotracer. IV administration 2 hr before surgery at 1 mg dose	IRDye-800	Chen-NIH	Brain Cancer	Phase 0	NCT02910804	recruiting
EMI-137 ⁷⁷	Human hepatocyte growth factor receptor (c-MET) targeting peptide conjugated to cyanine based fluorophore (Cy5?). IV administration 1-3 hours before surgery	Cy5	Edinburgh Molecular Imaging	Colon Cancer	Phase 2	NCT03360461	recruiting
				Thyroid Cancer	Phase 1	NCT03470259	completed
				Barret Esophagus	Phase 1	NCT03205501	recruiting
				Lung Cancer	Phase 1	NCT02676050	not yet recruiting
QRH-882260 Heptapeptide ⁷⁸	Seven amino acid long peptide that binds to EGFR labeled with cy5. Orally administered and binds to tumor cells in GI tract.	Cy5	Wang-University of Michigan	Colon Cancer	Phase 1	NCT03148119	terminated (QRH to be used for other indications)
				Cholangiocarcinoma	Phase 1	NCT03438435	recruiting
				Barret Esophagus	Phase 1	NCT03589443	completed
				Safety study	Phase 1	NCT02574858	completed
KSP/QRH peptide dimer ⁷⁹	EGFR and HER2 targeting peptide labelled with IRDye800. Orally administered or sprayed onto area of interest	IRDye-800	Wang-University of Michigan	Barret Esophagus	Phase 1	NCT03852576	recruiting
GI heptapeptide ⁸⁰	Heptapeptide labeled with FITC. Orally administered or sprayed onto area of interest	FITC	Wang-University of Michigan	Barret Esophagus	Phase 1	NCT01630798	completed

Probe Name	Description	Fluorophore	Company/ Group	Indications	Clinical Trial Phase	NCT#	Status
KCC heptapeptide ⁸¹	Heptapeptide labeled with FITC. Orally administered or sprayed onto area of interest	FITC	Wang- University of Michigan	Colorectal Cancer	Phase 1	NCT02156557	completed
LS301 ^{82,83}	integrin receptor targeting octapeptide conjugated to NIR fluorophore cypate. IV administration 1 day prior to surgery.	Cypate	Achilefu - Washington University	Breast Cancer	Phase 2	NCT02807597	not yet recruiting
				Pancreatic Cancer	Phase 2	NCT04105062	not yet recruiting
				Liver Cancer	Phase 2		
				Gastric Cancer	Phase 2		
				Gastrointestinal Stromal Cancer	Phase 2		
				Metastatic Cancer	Phase 2		
RGD peptide -cy7 (ORL-1) ^{84–88}	Alpha(v) beta(3) integrin targeting peptide labeled with cy7. Topically applied to skin surface.	Cy7	Orlucant	Melanoma		NCT03535077	recruiting
Nanoparticle							
cRGDY-PEG-Cy5.5- C ^{*89}	Integrin-targeting, dual modality (PET & fluorescent) silica nanoparticle labeled with cy5.5. Injected at site of tumor before or during surgery	Cy5.5	Patel - Sloan Kettering	Head and Neck Cancer	Phase 2	NCT02106598	recruiting
				Breast Cancer	Phase 2		
				Colorectal Cancer	Phase 2		
cRGD-ZW800–1 ^{87,90}	Integrin-targeting silica nanoparticle labelled with ZW800. Injected 4–24 hours prior to surgery.	ZW-800	Keereweer - Erasmus Medical Center	Head and Neck Cancer	Phase 2	NCT04191460	not yet recruiting
⁶⁴ Cu-NOTA-PSMAi-PEG-Cy5.5-C ^{*91}	PSMA targeting nanoparticle, multi-modality (PET, MRI, fluorescent) labelled with Cy5.5.	Cy5.5	Memorial Sloan Kettering	Prostate Cancer	Phase 1	NCT04167969	recruiting
ONM-100 ⁹²	pH-sensitive micelles conjugated with ICG. IV administration on day of surgery.	ICG	OncoNano Medicine	Breast Cancer	Phase 2	NCT03735680	recruiting
				Head and Neck Cancer	Phase 2		
				Colorectal Cancer	Phase 2		
				Bladder Cancer	Phase 2		
				Prostate Cancer	Phase 2		
				Ovarian Cancer	Phase 2		
Small Molecule							
Demeclocycline ⁹³	Antibiotic with UV abs/yellow fl. Administered orally for 2 day	Demeclocycline	Curry - Massachuset ts General	Brain Tumor	Phase 1	NCT02740933	not yet recruiting

Probe Name	Description	Fluorophore	Company/ Group	Indications	Clinical Trial Phase	NCT#	Status
IRDye-800BK ⁹⁴	injected into urethra and imaged immediately	IRDye-800BK	Barnes - Oxford	Ureter Injury	Phase 2	NCT03387410	completed
LUM015 ⁹⁵⁻⁹⁷	Cathepsin-activatable labeled with cy5 and fluorescent quencher linked by a pan-cathepsin protease cleavable peptide. IV injection 2-6 hrs prior to surgery at 1 mg/kg dose.	Cy5	Lumicell	Breast Cancer	Phase 3	NCT03686215	recruiting
				Colorectal Cancer	Phase 2	NCT02584244	recruiting
				Barret Esophagus	Phase 2		
				Pancreatic Cancer	Phase 2		
				Brain Cancer	Phase 1	NCT03717142	recruiting
				Prostate Cancer	Phase 1	NCT03441464	recruiting
IS-001 ⁴⁴	Agent for use with daVinci robot	IS-001	Intuitive Surgical	Hysterectomy - Ureter Injury	Phase 2	NCT03937505	recruiting
PARPi-FL ^{98,100}	PARP1 inhibitor (olaparib) labeled with BODIPY-FL. Applied topically for basal cell carcinoma	BODIPY-FL	Reiner - Sloan Kettering	Oral Squamous Cell Carcinoma	Phase 2	NCT03085147	recruiting
HS-196 ^{101,102}	Heat shock protein 90 (HSP90) inhibitor labelled with NIRdye. IV administration	N/A	Lyerly - Duke	Solid Tumor	Phase 1	NCT03333031	recruiting
TMVP1-ICG	Used for SLN mapping, injected into the cervix	ICG	Ding Ma - Huazhong University	Cervical Cancer	Phase 1	NCT03320772	recruiting
EC17 ¹⁰³⁻¹⁰⁵	Folate-FITC conjugate targeting folate receptor. IV administration 2-4 hours before surgery at 0.1 mg/kg dose	FITC	Singhal - University of Pennsylvania	Hyperparathyroidism	Phase 1	NCT01996072	completed
				Breast Cancer	Phase 1	NCT01994369	completed
				Ovarian Cancer	Phase 1	NCT02000778	completed
				Renal Carcinoma	Phase 0	NCT01778933	completed
				Lung Cancer	Phase 1	NCT01778920	completed
OTL38 ¹⁰⁶⁻¹¹⁰	Folate receptor alpha (Fra) targeting ligand (folic acid) conjugated to NIR fluorophore S0456. IV administration 2-3 hrs before surgery	S0456	On Target Laboratories	Ovarian Cancer	Phase 3	NCT03180307	recruiting
				Lung Cancer	Phase 2	NCT02872701	completed
				Lung Cancer	Phase 1	NCT02769156	recruiting
				Lung Cancer	Phase 1	NCT02602119	recruiting
				pituitary adenoma	Phase 1	NCT02629549	terminated (recruitment fulfilled)
				Malignancies in pituitary gland	Phase 1	NCT02769533	completed
				Bladder Cancer	Phase 1	NCT02852252	recruiting
				Rheumatoid Arthritis	Phase 1	NCT03938701	not yet recruiting
				Renal Carcinoma	Phase 1	NCT02645409	completed
Li-COR Ureter Agent	IV administration during surgery at 0.06 mg/kg	IRDye-800BK	Li-COR	Ureter Injury	Phase 2	NCT03106038	completed
MB-102 ^{46,47}	Human plasma fluorescence tracer.	MB-102	MediBeacon	Acute Kidney Injury	Phase 2	NCT02772276	recruiting

Probe Name	Description	Fluorophore	Company/ Group	Indications	Clinical Trial Phase	NCT#	Status
Aftobetin - HCl	Amyloid beta binding ligand. Administered via ophthalmic ointment, eye lens imaged.	Aftobetin	Cognoptix	Alzheimer's Disease	Phase 1	NCT02928211	recruiting
HS201	HSP90 inhibitor connected by linker to verteporfin (imaging agent and photosensitizing agent).	Verteporfin	Lyerly - Duke	Solid Tumor	Phase 1	NCT03906643	not yet recruiting
LuminoMark	Fluorescence localization in patients with nonpalpable breast lesions - not clear if this is a functionalized version of ICG or tagged molecule	ICG	Hanlim Pharm. Co. Ltd.	Breast Cancer	Phase 2	NCT03743259	completed
Prosense/VM110 ^{111,112}	Near infrared fluorophore self-quenched and activated when cleaved by proteases, cathepsin B,L, and S and plasmin in cancer cells, fluorescent cleavage product detected. Combination product.	Cy5.5	Weissleder - Harvard	Ovarian Cancer	Phase 1	NCT03286062	active, not recruiting
				Pancreatic Cancer	Phase 1		

Table 3.

Current FGS imaging systems.

System Name	Description	Company/Group	Excitation (nm)	Emission (nm)	Approved?
Fluobeam	Hand held NIR imaging system (800nm channel approved, 700nm channel (680nm ex - 700 nm em developed)	FluoOptics Imaging	750	800 longpass	Yes
PDE/Photodynamic EyeNeo(II)	Hand held NIR imaging, uses LEDs	Hamamatsu/Mitaka	760	820 longpass	Yes
Firefly	Endoscopic fluorescence imaging system for robotic surgery	Intuitive Surgical	805	805 blocking	Yes
Image 1S Camera System	Endoscopic fluorescence imaging system	Karl Storz	N/A	N/A	Yes
Vitom II	Karl Storz Image 1s camera system, but used in open surgery as a microscope	Karl Storz	N/A	N/A	Yes
GLOW800	Leica Surgical Microscope accessory (M530)	Leica	790	835	Yes
FL560	Leica Surgical Microscope accessory (M530)	Leica	460–500	510 longpass	Yes
FL800	Leica Surgical Microscope accessory (M530)	Leica	700–800	820–860	Yes
FL400	Leica Surgical Microscope accessory (M530)	Leica	380–430	444 longpass	Yes
Pinpoint	Endoscopic fluorescence imaging system (“deep red” [for MB] system used in clinical trials)	Novadaq/Stryker	805	825–850	Yes
SPY PHI	Hand held fluorescence imaging system	Novadaq/Stryker	805	825–850	Yes
SPY Elite	On cart for open surgery fluorescence imaging system	Novadaq/Stryker	805	825–850	Yes
AIM (Advanced Imaging Modality)	Stryker’s AIM camera combined with SPY, SPY Elite, Pinpoint	Stryker	805	825–850	Yes
Visera Elite II	NIR and narrow band imaging in endoscope (Approved in Europe but not US?)	Olympus	N/A	N/A	Yes
Artemis	Hand held NIR fluorescence imaging system	Quest Medical Imaging	N/A	N/A	Yes
Spectrum	Replaced Artemis, hand held with 700 and 800 nm imaging channel, uses LEDs	Quest Medical Imaging	N/A	700–830, 830–1000	Yes
VS3 IR/Iridium System	3D endoscope or open surgical system with NIR fluorescence channel	VisionSense/Medtronic	805	825–850	Yes
Yellow 560	Zeiss Surgical Microscope accessory (Kinevo 900)	Zeiss	460–500	550–700	Yes
Infrared 800	Zeiss Surgical Microscope accessory (Kinevo 900)	Zeiss	700–780	820–900	Yes
Blue 400	Zeiss Surgical Microscope accessory (Kinevo 900)	Zeiss	400–410	620–710	Yes
Fluorescence Goggle System	Augmented reality goggle system for FGS	Achilefu (Wash U)	780	N/A	No

System Name	Description	Company/Group	Excitation (nm)	Emission (nm)	Approved?
OPAL	Light projection system for visualization on tissue surface in open surgery	Akers (Wash U)	780	785 longpass	No
GXMI Navigator	Cart based system	Chinese Academy of Sciences	760	810–870	No
FLARE	Cart based system for open surgery, 700 & 800 nm channel (mini-FLARE latest iteration)	Curadel	656–678, 745–779	689–725, 803–853	No
IC-flow	Handheld imaging system	Diagnostic Green	780	N/A	No
HyperEye Medical System	Hand held fluorescence imaging system	Mizuho Medical Company	760–780	800–850	No
Solaris	Filtered LED light source, on cart	PerkinElmer	488, 667, 743, 757	516–523, 692–742, 770–809, 784 LP	No
Visual Navigator	Hand held system	SH System	740	820	No
Explorer Air	Multispectral imaging platform (currently in clinical trials in EU, formerly SurgOptix T-3)	Surgvision	520, 800	N/A	No