Published in final edited form as:

J Biophotonics. 2018 March ; 11(3): . doi:10.1002/jbio.201700303.

Polarization-multiplexed, dual-beam swept source optical coherence tomography angiography

Jianlong Yang, Rahul Chandwani, Rui Zhao, Zhuo Wang, Yali Jia, David Huang, Gangjun Liu*

Casey Eye Institute, Oregon Health and Science University, 3375 SW Terwilliger Blvd, Portland, Oregon, 97239, USA

Abstract

A polarization-multiplexed, dual-beam setup is proposed to expand the field of view for a swept source optical coherence tomography angiography (OCTA) system. This method used a Wollaston prism to split sample path light into two orthogonal-polarized beams. This allowed two beams to shine on the cornea at an angle separation of ~ 14 degrees, which led to a separation of ~ 4.2 mm on the retina. A 3-mm glass plate was inserted into one of the beam paths to set a constant path length difference between the two polarized beams so the interferogram from the two beams are coded at different frequency bands. The resulting OCTA images from the two beams were coded with a depth separation of ~ 2 mm. 5×5 mm² angiograms from the two beams were obtained simultaneously in 4 seconds. The two angiograms then were montaged to get a wider field of view (FOV) of $\sim 5 \times 9.2$ mm².

Graphical Abstract

*Corresponding Author: liga@ohsu.edu.

We demonstrated a dual-beam swept-source OCT system based on polarization-multiplexing to enlarge the FOV of OCTA. A Wollaston prism was used to create two orthogonally-polarized beams. By montaging two OCTA volumes that were simultaneously captured in 4 seconds, we obtained *en face* angiograms of the human retina with a FOV of $5 \times 9.2 \text{ mm}^2$.

Keywords

Optical coherence tomography; ophthalmic optics and devices; medical and biological imaging

1. Introduction

Optical coherence tomography angiography (OCTA) is an emerging imaging modality that can be used to image microvasculature noninvasively [1–2]. However, commercially available OCTA systems usually have an A-line speed of 70 ~ 100 kHz and the field of view (FOV) of high-resolution OCTA is limited. A faster system could be used to obtain wider FOV OCTA [3]. However, the degraded signal to noise ratio (SNR) could not allow clear visualization of capillaries. Currently, the most popular approach increasing the FOV of OCTA is to montage multiple scans from different areas acquired at different times [4–6]. However, this method can be time-consuming and impractical in some clinical practices.

Two or more scan beams have been employed in many OCT applications. In the 1990s, a dual-beam setup enabled high-precision biometry of the human eye *in vivo* [7]. Dual and multiple beam setups have been used to increase the effective A-line rate of the OCT system [8–10]. Multiple beam systems have also been utilized in Doppler OCT to measure the absolute retinal blood flow rate or total retinal blood flow [11–13]. Makita *et al.* demonstrated a time-delayed dual beam system for Doppler OCTA with high sensitivity and high speed [14]. Here, we demonstrate a polarization-multiplexed dual-beam system to

expand the FOV of OCTA. The system only uses one swept laser, one interferometer, and one balanced detector. Two probe beams with two orthogonal polarization states are used to image different regions of the retina simultaneously. The separation between the two beams is ~ 4.2 mm and polarization crosstalk between the two beams is prevented for such a large separation. The interferograms from the two probe beams are detected with a single detector and coded at different frequency bands. The OCTA images from the two probe beams are montaged to obtain a wider FOV OCTA image. Due to the concurrent acquisition of the two OCTA images, the need for registration procedures for montaging the OCTA images is eliminated in the post-processing as the OCTA images are already registered.

2. Materials and Methods

2.1. System Configuration and Characterization

Figure 1(a) shows the schematic of the dual-beam OCTA system. We employed a swept source laser (Axsun Technologies, Billerica, MA, USA) with an A-line speed of 100 kHz, a central wavelength of 1050 nm, and a tuning range of 110 nm. The theoretical axial resolution of the system is ~ 5.7 μm . The built-in k-clock of the swept source laser enabled an axial imaging range of ~ 3 mm, which was insufficient for our dual-beam OCT setup. Thus, we adopted a frequency doubling circuit [6, 15, 16] to increase the axial imaging range to ~ 6 mm.

A 75:25 fiber coupler (AC Photonics, Santa Clara, CA, USA) was used to split the laser light into the sample arm (25%) and reference arm (75%). In the reference arm, a fiber-coupled optical delay line was used to fine tune the optical path length difference between the two arms and an iris diaphragm (D5S, Thorlabs, Newton, NJ, USA) was mounted on the delay line to control the reference power. A 50:50 fiber coupler (TW1064R5F2A, Thorlabs, Newton, NJ, USA) was used for combining the reference light and the back-scattered sample light, which ultimately were sent into a balanced detector (PDB481C-AC, Thorlabs, Newton, NJ, USA). The interferometric signal was digitalized by a high-speed digitizer (ATS9350, AlazarTech, Pointe-Claire, QC, Canada) and then processed by a computer.

In the sample arm, the light output from the fiber was collimated by a collimator and then passed through an electronically tunable lens (EL-10-30-C, Optotune, Dietikon, Switzerland), which was used to tune the focal plane location. A fast automatic search algorithm was used for rapid focusing [16]. After passing the tunable lens, the laser light traveled through a dual beam manipulation component to achieve two spatially-separated orthogonally-polarized beams. The dual beam manipulation component included a Wollaston prism and two golden mirrors (Figure 1b). The Wollaston prism (WP10, Thorlabs, Newton, NJ, USA) separated the collimated light into two orthogonally polarized outputs with a separation angle of 20° . The two 2-inch gold mirrors (PF20-03-M01, Thorlabs, Newton, NJ, USA) converged the two beams with an 6° separation angle. A 3-mm glass plate (WG10530-C, Thorlabs, Newton, NJ, USA) was inserted into one of the beams to set an optical path length difference of ~ 2 mm between two probe beams. After passing a two-axes galvo scanner (6215H, Cambridge Technology, Bedford, MA, USA) and a telescope system, the beams were able to shine on the cornea at a separation angle of 14 degrees. The telescope included a scan lens (AC508-100-B, Thorlabs, Newton, NJ, USA) with a focus

length of ~100 mm and a compound ocular lens (two 49–380, Edmund Optics Inc., Barrington, New Jersey) with an effective focal length of ~40 mm. Since the two probe beams shared the same reference arm and had an optical path length difference of ~2 mm, the images from the two probe beams were coded in the same image at different depths. The right side of Fig. 1(a) shows an example of the acquired retinal B-scan OCT image overlaid with its angiogram.

To guarantee the performance similarity of the two probing beams, two polarization controllers (PCs) (FPC030, Thorlabs, Newton, NJ, USA) were inserted, respectively, in the sample and reference arms. In the sample arm, the PC was used to even the probe powers of the two orthogonally-polarized beams. The amplitudes of the two orthogonally polarized beams were monitored with a digital optical power meter and a photodiode power sensor (PM100D and S122C, Thorlabs, Newton, NJ, USA). In the reference arm, the PC optimized the image quality by matching the polarization states to the sample arm. The beam profiles for the two probe beams were measured by a scanning slit optical beam profiler (BP209-IR, Thorlabs, Newton, NJ, USA) (Figure 2a). The results show a full-width-at-half-maximum beam diameter of ~1 mm for both beams. The sensitivity roll-off performance for the two beams is shown in Figures 2b and 2c. The measured peak sensitivities for the two beams were 95.5 dB and 95.9 dB, respectively. The 6-dB roll-off depth for both beams was ~4.3 mm. These results show a comparable performance for the two beams.

2.2. Scan Schematic and Laser Safety

To apply this method for imaging the human retina with the goal of expanding the FOV of OCTA, the scanning scheme and protocol have to be carefully selected. Figure 3 shows the scan schematic and protocol of the dual-beam OCTA system. As shown in Fig. 3, the corresponding distance between the two focal spots on the retina was ~4.2 mm. A scanning protocol of 500×2×400 A-lines (500 A-lines per B-scan, two repetitions at the same B-scan location and 400 B-scan locations) with a FOV of 5×5 mm² was used. This separation angle (the distance between the two focal spots on the retina) and scanning protocol were specifically chosen based on the system speed and the total scanning time (i.e. 4 seconds) for a single 3D volumetric angiography scan. Because the separation between the two beams was 4.2 mm and the total scanning size for a single beam was 5×5 mm², there was an overlap area of 0.8×5 mm² between the two areas scanned by the two beams. Therefore, the total scanning area by the two beams was 5×9.2 mm². To calculate the OCTA signal, we used two repeated B-scans at the same FOV. The total scan time was 4 seconds and the transverse sampling step size was ~12.5 μm for both directions.

A power of 1.6 mW was used for each of probe beams. The two probe beams had a separation of ~14 degrees (244 mrad) on the retina, which is larger than the “limiting angle of acceptance” (100 mrad) set by the International Standard for Safety of Laser Products (IEC 60825–1) [17]. Thus, the two beams were treated independently when considering ocular laser safety. The 1.6 mW power of each beam is below the most conservative limit for the 1-μm laser radiation used in retinal imaging (1.9 mW) [8]. On the cornea, the two beams were overlapped and the total power was 3.2 mW. This power is below the laser safety limit for anterior segment that was set by the American National Standard for the Safe Use of

Lasers (ANSI Z136.1–2014) [18]. This concludes that the power was safe for both the anterior segment and the retina according to the IEC 60825–1 [17] and ANSI Z136.1–2014 [18]. For these reasons, the system has been approved by the Oregon Health & Science University Institutional Review Board for clinical usage. All the studies performed using this system follow the tenets of the Declaration of Helsinki for the treatment of human subjects.

3. Results

We demonstrate the results of this dual-beam OCTA system in Fig. 4 and Fig. 5. A healthy male volunteer of age 29 years was involved in this research. Retinal fovea and optic nerve head (ONH) regions were imaged. Figure 4 shows the 3D intensity volumes (Figures 4a and 4d) and cross-sectional images (Figures 4b, 4c, 4e, 4f). In the cross-sectional images, red-color angiograms were overlaid on gray-scale reflectance images. It can be seen that the two beams delivered very similar performance. Since the scan regions of the two beams have an overlapping of $\sim 0.8 \times 5$ mm, they can be effectively montaged. Figure 5 shows the montaged *en face* angiograms of fovea (a) and ONH (b). The red dash lines in the middle indicate the stitching location of the images.

The current setup had a separation of ~ 4.2 mm on human retina, which can be altered by properly tuning the distance between the Wollaston prism and the two gold mirrors as well as tilting the angles of the two mirrors. This could allow more versatile setups for other applications such as the time-delayed dual beam Doppler OCT and OCTA [14]. In addition, the current probe beams were separated vertically on the retina and they simultaneously image the retina superiorly and inferiorly. By properly rearranging the Wollaston prism and two mirrors, the two beams can be separated horizontally on the retina. This could allow for imaging the macular and optic disk regions simultaneously in a single scan.

The polarization states of the reference arm cannot be perfectly aligned with those of the sample arm by manually tuning the PC, which may degrade the detection sensitivity [9]. Because non-polarization maintained fiber was used in this setup, the polarization state was not always maintained. This requires re-tuning of the PCs for long-term usage of the system. We will further improve the system by replacing the manually tunable PC to automatically tunable PC. By developing an algorithm to monitor and evaluate the brightness of the two images, we will adjust the PC automatically.

4. Conclusion

In summary, we demonstrated a dual-beam swept-source OCT system based on polarization-multiplexing and employed this method to enlarge the FOV of OCTA. A Wollaston prism was used to create two orthogonally-polarized beams while a thin glass sheet allowed for separating the images from these two beams in a single B frame. Our system avoided the use of multiple interferometers and detectors, plate and cube polarizing beam splitters, and related extra optics, which substantially simplified the system's design. The two beams showed very similar performance and gave comparable OCT image quality. By montaging two OCTA volumes that were simultaneously captured in 4 seconds, we obtained *en face* angiograms of the human retina with a FOV of 5×9.2 mm². These results show the

feasibility of further applying this design in faster OCT systems to shorten the acquiring time of wide-field OCTA.

Acknowledgments

Funding This research was funded by Oregon Health & Science University Foundation, National Institutes of Health Grants R01EY027833, R01 EY024544, R01 EY023285 and R01 EY018184, unrestricted departmental funding from Research to Prevent Blindness (New York, NY), and P30 EY010572 from the National Institutes of Health (Bethesda, MD).

References

- [1]. Gao Simon S., Jia Yali, Zhang Miao, Su Johnny P.; Liu Gangjun, Hwang Thomas S., Bailey Steven T., Huang David, "Optical coherence tomography angiography," *Invest. Ophthalmol. Vis. Sci.* 57(6), OCT27–OCT36 (2016). [PubMed: 27409483]
- [2]. Yang Jianlong, Su Johnny, Wang Jie, Men Silu, Jia Yali, Huang David, and Liu Gangjun, "Hematocrit dependence of flow signal in optical coherence tomography angiography," *Biomed. Opt. Express* 8(2), 776–789 (2017). [PubMed: 28270984]
- [3]. Blatter Cedric, Grajciar Branislav, Schmoll Tilman, Leitgeb Rainer A., Klein Thomas, Wieser Wolfgang, André Raphael J., Huber Robert., "Ultrahigh-speed non-invasive widefield angiography," *J. Biomed. Opt.* 17(7), 070505 (2012). [PubMed: 22894461]
- [4]. Zhang Qinqin, Huang Yanping, Zhang Thomas, Kubach Sophie, An Lin, Laron Michal, Sharma Utkarsh, Wang Ruikang K., "Wide-field imaging of retinal vasculature using optical coherence tomography-based microangiography provided by motion tracking," *J. Biomed. Opt.* 20(6), 066008 (2015). [PubMed: 26102573]
- [5]. Zhang Qinqin, Huang Yanping, Zhang Thomas, Kubach Sophie, An Lin, Laron Michal, Sharma Utkarsh, Wang Ruikang K., "Wide-field optical coherence tomography based microangiography for retinal imaging," *Sci. Rep.* 6, 22017 (2016). [PubMed: 26912261]
- [6]. Liu Gangjun, Yang Jianlong, Wang Jie, Li Yan, Zang Pengxiao, Jia Yali, Huang David, "Extended axial imaging range, widefield swept source optical coherence tomography angiography," *J. Biophotonics* doi:10.1002/jbio.201600325 (2017).
- [7]. Baumgartner Angela, Hitzenberger Christoph K., Sattmann Harald, Drexler Wolfgang, Adolf Friedrich Fercher, "Signal and Resolution Enhancements in Dual Beam Optical Coherence Tomography of the Human Eye," *J. Biomed. Opt.* 3(45) 45–54 (1998). [PubMed: 23015005]
- [8]. Potsaid Benjamin, Baumann Bernhard, Huang David, Barry Scott, Cable Alex E., Schuman Joel S., Duker Jay S., and Fujimoto James G., "Ultrahigh Speed 1050nm Swept Source / Fourier Domain OCT Retinal and Anterior Segment Imaging at 100,000 to 400,000 Axial Scans Per Second," *Opt. Express* 18(19), 20029–20048 (2010). [PubMed: 20940894]
- [9]. Nankivil Derek, Dhalla Al-Hafeez, Gahm Niklas, Shia Kevin, Farsiu Sina, and Izatt Joseph A., "Coherence revival multiplexed, buffered swept source optical coherence tomography: 400 kHz imaging with a 100 kHz source," *Opt. Lett.* 39(13), 3740–3743 (2014). [PubMed: 24978725]
- [10]. Suehira Nobuhito, Ooto Sotaro, Hangai Masanori, Matsumoto Kazuhiro, Tomatsu Nobuhiro, Yuasa Takeshi, Yamada Kazuro, Yoshimura Nagahisa, "Three-beam spectral-domain optical coherence tomography for retinal imaging," *J. of Biomed. Opt.* 17(10), 106001 (2012). [PubMed: 23224000]
- [11]. Veronika Doblhoff-Dier Leopold Schmetterer, Vilser Walthard, Gerhard Garhöfer Martin Gröschl, Leitgeb Rainer A., and Werkmeister René M., "Measurement of the total retinal blood flow using dual beam Fourier-domain Doppler optical coherence tomography with orthogonal detection planes," *Biomed. Opt. Express* 5(2), 630–642 (2014). [PubMed: 24575355]
- [12]. Haindl R, Trasischker W, Baumann B, Pircher M, and Hitzenberger CK, "Three-beam Doppler optical coherence tomography using a facet prism telescope and MEMS mirror for improved transversal resolution," *J. Mod. Opt.* 62(21):1781–1788 (2015). [PubMed: 26689672]
- [13]. Haindl Richard, Trasischker Wolfgang, Wartak Andreas, Baumann Bernhard, Pircher Michael, and Hitzenberger Christoph K., "Total retinal blood flow measurement by three beam Doppler

- optical coherence tomography,” *Biomed. Opt. Express* 7(2), 287–301 (2016). [PubMed: 26977340]
- [14]. Makita Shuichi, Jaillon Franck, Yamanari Masahiro, Miura Masahiro, and Yasuno Yoshiaki, “Comprehensive in vivo micro-vascular imaging of the human eye by dual-beam-scan Doppler optical coherence angiography,” *Opt. Express* 19(2), 1271–1283 (2011). [PubMed: 21263668]
- [15]. Wang Zhao, Lee Hsiang-Chieh, Osman Oguz Ahsen ByungKun Lee, Choi WooJhon, Potsaid Benjamin, Liu Jonathan, Jayaraman Vijaysekhar, Cable Alex, Kraus Martin F., Liang Kaicheng, Hornegger Joachim, and Fujimoto James G., “Depth-encoded all-fiber swept source polarization sensitive OCT,” *Biomed. Opt. Express* 5(9), 2931–2949 (2014). [PubMed: 25401008]
- [16]. Yang Jianlong, Liu Liang, Campbell J. Peter, Huang David, and Liu Gangjun, “Handheld optical coherence tomography angiography,” *Biomed. Opt. Express* 8(4), 2287–2300 (2017). [PubMed: 28736672]
- [17]. International Electrotechnical Commission, “Safety of laser products—part1: equipment classification and requirements,” International Standard IEC 60825–1, 2nd ed., International Electrotechnical Commission, Geneva (2007).
- [18]. Laser Institute of America, “American National Standard for Safe Use of Lasers,” ANSI Z136.1–2014 (2014).

Figure 1.
 (a) Schematic of the dual-beam OCTA system. (b) Photograph of the dual beam manipulation components.

Figure 2. (a) 2-dimensional beam profiles at the position of cornea. (b) and (c) are the sensitivity roll-offs of the two orthogonally-polarized probe beams.

Figure 3.
Scan schematic and protocol of the dual-beam OCTA system.

Figure 4. Captured OCTA image data sets at around fovea and optic nerve head (ONH). (a) and (d) show the 3D intensity volumes. (b), (c), (e), and (f) show the cross-sectional intensity images (gray) overlaid with angiograms (red).

Figure 5. Montaged en face angiograms of retina regions at around (a) fovea and (b) optic nerve head. The red dash lines in the middle give the positions of stitching images.