Published in final edited form as:

IEEE Trans Biomed Eng. 2020 February ; 67(2): 441–452. doi:10.1109/TBME.2019.2915165.

Pose Estimation of Periacetabular Osteotomy Fragments With Intraoperative X-Ray Navigation

Robert B. Grupp,

Department of Computer Science, Johns Hopkins University, Baltimore, MD 21218 USA

Rachel A. Hegeman,

Research and Exploratory Development Department, Johns Hopkins University Applied Physics Laboratory

Ryan J. Murphy,

Research and Exploratory Development Department, Johns Hopkins University Applied Physics Laboratory. He is now with Auris Surgical Robotics

Clayton P. Alexander,

Department of Orthopaedic Surgery, Johns Hopkins Medicine

Yoshito Otake,

Graduate School of Information Science, Nara Institute of Science and Technology

Benjamin A. McArthur,

Washington Orthopedics and Sports Medicine. He is now with the Dell Medical School, University of Texas

Mehran Armand,

Research and Exploratory Development Department, Johns Hopkins University Applied Physics Laboratory and also with the Department of Mechanical Engineering, Johns Hopkins University

Russell H. Taylor [Life Fellow, IEEE]

Department of Computer Science, Johns Hopkins University

Abstract

Objective—State-of-the-art navigation systems for pelvic osteotomies use optical systems with external fiducials. In this paper, we propose the use of X-ray navigation for pose estimation of periacetabular fragments without fiducials.

Methods—A two-dimensional/three-dimensional (2-D/3-D) registration pipeline was developed to recover fragment pose. This pipeline was tested through an extensive simulation study and six cadaveric surgeries. Using osteotomy boundaries in the fluoroscopic images, the preoperative plan was refined to more accurately match the intraoperative shape.

Personal use is permitted, but republication/redistribution requires IEEE permission. See http://www.ieee.org/publications_standards/publications/rights/index.html for more information.

Corresponding author: Robert Bruce Grupp. grupp@jhu.edu.

This paper has supplementary downloadable material available at <http://ieeexplore.ieee.org>, provided by the authors. The file size is 111 MB).

Results—In simulation, average fragment pose errors were $1.3^\circ/1.7$ mm when the planned fragment matched the intraoperative fragment, $2.2^\circ/2.1$ mm when the plan was not updated to match the true shape, and $1.9^\circ/2.0$ mm when the fragment shape was intraoperatively estimated. In cadaver experiments, the average pose errors were $2.2^\circ/2.2$ mm, $3.8^\circ/2.5$ mm, and $3.5^\circ/2.2$ mm when registering with the actual fragment shape, a preoperative plan, and an intraoperatively refined plan, respectively. Average errors of the lateral center edge angle were less than 2° for all fragment shapes in simulation and cadaver experiments.

Conclusion—The proposed pipeline is capable of accurately reporting femoral head coverage within a range clinically identified for long-term joint survivability.

Significance—Human interpretation of fragment pose is challenging and usually restricted to rotation about a single anatomical axis. The proposed pipeline provides an intraoperative estimate of rigid pose with respect to all anatomical axes, is compatible with minimally invasive incisions, and has no dependence on external fiducials.

Keywords

Two-dimensional/three-dimensional (2-D/3-D) registration; orthopedics; periacetabular osteotomy; X-ray navigation

I. Introduction

DEVELOPMENTAL hip dysplasia (DDH) is a congenital condition which may cause greater contact pressures between the femoral head and acetabular cartilage due to a deformed acetabulum or femoral head. These conditions may cause pain and limit the amount of physical activity an individual may perform. When DDH is not addressed, severe arthritis by age 30 is typical, and requires surgical intervention [1].

In 1984, the periacetabular osteotomy (PAO) was developed by Ganz for the purpose of treating DDH [2]. Standing AP and false profile radiographs are used to preoperatively assess the condition and develop an initial surgical plan. Osteotomies along the pubis, ilium, ischium, and posterior column of the pelvis are performed, typically with fluoroscopic guidance [2]. The ischial and posterior cuts present significant challenges, including using difficult to interpret fluoroscopic views, potentially affecting the Sciatic nerve, and risking joint breakage due to the osteotome's proximity to the acetabulum. After completing the osteotomies, the acetabular fragment is freed from the pelvis and repositioned to increase femoral coverage and improve joint contact pressure [2]–[6] (Fig. 1).

Mentally resolving the 3D pose of the acetabular fragment using intraoperative fluoroscopy is challenging, especially for novice surgeons [7]. An acceptable repositioning is often determined using the lateral center edge (LCE) angle [8], which estimates the amount of lateral femoral head coverage provided by the acetabulum. In order to achieve long-term survivability of the joint, a LCE angle between 30° – 40° is desired [9]. Mentally resolving the LCE angle from 2D radiographs, without any other tools, was shown to have a large variance [10].

To address the challenges with performing PAO, several computer assisted systems using optical tracking technology have been developed; e.g. [11]–[16]. However, intraoperative optical tracking systems are not yet standard equipment in most operating rooms, and have several technical disadvantages, such as a sensitivity to occlusion and a limited workspace. Moreover, the need to digitize points on the ilium and iliac crest for registration of the patient's pelvis to a preoperative model, requires more invasive incisions than typically needed for a PAO [12], [17], [18]. These reasons, and the universal availability of intraoperative fluoroscopic imaging, provide motivation for the use of an X-Ray based navigation system. In place of specialized 3D tracking devices, computer-assisted X-Ray navigation systems use the fluoroscopic or radiographic imagers already present during many interventions [19]–[32]. Since the acetabular fragment is created intraoperatively, an accurate model of the fragment shape is not available, and invalidates the assumptions used in existing multiple-object registration methods [22], [24], [28].

We propose a fiducialless approach that performs multiple-object, multiple-view, 2D/3D X-Ray to CT registration to resolve the pose of an acetabular fragment with respect to the anterior pelvic plane (APP). After the clinician performs a repositioning of the fragment, the pelvis and fragment are registered using several fluoroscopic images, and the fragment's pose and LCE angle are reported. The fragment shape is estimated after registering the patient's non-acetabular portion of the pelvis. Cut lines present in the 2D images are used to approximate the ilium and pubis osteotomies, while the ischial and posterior osteotomies remain set according to a preoperative plan. The pose and LCE values may be interpreted by the clinician to determine whether the fragment is in an acceptable pose, or whether further adjustment is needed. When the fragment needs to be moved, the current pose estimate may help determine the direction in which the fragment should next be adjusted. Additionally, but beyond the scope of this work, biomechanical indicators may also be computed once the fragment's relative pose is known [3]–[6]. The high-level workflow of our proposal is shown in Fig. 2.

We believe this is the first computer assisted system for intraoperatively tracking an acetabular fragment with X-Ray navigation without the use of artificial fiducial objects. Compared to other multiple-view registration approaches in literature, our approach does not require a tracked, encoded, or motorized, C-Arm. Unlike existing multiple-object registration solutions, an inaccurate model of the fractured shape is allowed and a prior distribution over fragment shapes is not required. Moreover, non-standard equipment is not required to perform the surgery and no burdensome steps are added to the operating workflow.

We have evaluated the proposed method with a large simulation study and six cadaver surgeries. In both experiments we report fragment pose and LCE errors when knowing the true fragment shape, a preoperatively planned shape, and an intraoperatively estimated shape. To evaluate the performance of our method, we examine rotation and translation components of the pose differences from ground truth, as well as the error in LCE angle measurement. Since it is expected that fragment pose and LCE angle errors will increase as the shapes used for registration match intraoperative shapes less, analysis of the simulation

and cadaver experiments will determine if the proposed methods are sufficiently accurate to assist with intraoperative evaluation of dysplasia.

II. Related Work

Initial work on computer aided systems for pelvic osteotomy focused on performing the actual osteotomies, rather than tracking the mobilized acetabular fragment. Langlotz demonstrated the use of preoperative CT and optically tracked surgical instruments to assist with PAOs, but did not provide estimates on the relocated pose of the acetabular fragment [11], [33]. Mayman used optical tracking to intraoperatively place guiding screws on the patient's pelvis and matching osteotomies from a preoperative plan, but did not provide intraoperative tracking of the fragment [13]. Akiyama also tracked osteotomes during a curved periacetabular osteotomy (CPO) [34] using optical tracking [15]. In [35], Radermacher constructed patient specific cutting guides for use in a triple osteotomy [36], however the pose of the fragment was not reported during the procedure. Similarly, Otsuki developed patient specific cutting guides to perform CPO and did not track the pose of the acetabular fragment [37].

Also using preoperative CT and optical tracking devices, Murphy did not track the osteotomes, but instead tracked the acetabular movements and computed intraoperative biomechanics [12], [38]. Through repeated digitizations of points on the acetabular fragment body, the fragment was tracked and the appropriate biomechanical properties were presented to the clinician [12]. Fragment pose errors of $1.4\text{--}1.8^\circ$ and $1.0\text{--}2.2\text{ mm}$ were reported. The process of manually digitizing the bone bur points during each fragment reposition adds a small amount of time to the overall procedure and may be subject to some error [38]. Liu also developed a system for preoperative PAO planning and intraoperative tracking of the fragment using optical tracking and a separate rigid body attached to the fragment [14], [39]. However, fixing a separate rigid body to the fragment is not necessarily practical when using a state-of-the-art, minimally invasive, approach such as [17] and [18]. For rotational acetabular osteotomies (RAO) [40], Takao used an optically tracked system to monitor the osteotomes and fragment movement [16]. The fragment pose was intraoperatively estimated by digitizing the anterior edge of the acetabulum, which resulted in some difficulty distinguishing between rotation and translation.

X-Ray navigation has been used to assist with the reduction of traumatic bone fractures, by providing real-time 3D visualizations of relocated bone fragments [24] and feedback corresponding to a 3D preoperative plan [22]. Methods to automatically annotate intraoperative images have been used to avoid wrong-level spinal surgery [30] and mitigate the malpositioning of surgical implants [19]. Using intraoperative X-Ray imaging, a surgical robot may be guided into an optimal location for the milling [23], [31], or drilling [32], of bone. Automatic 3D visualization and kinematic analysis of the wrist [20] and knee [21], [25] have also been demonstrated with X-Ray navigation. To our knowledge, no existing method based on X-Ray navigation, without fiducials, is able to localize a bone fragment without accurate preoperative knowledge or a statistical prior of the fragment shape.

At the core of an X-Ray based navigation system is a 2D/3D registration algorithm [26]. The goal of 2D/3D registration is to determine the pose of 3D objects with respect to a 3D coordinate frame using a series of 2D X-Ray images. Typically, a preoperative 3D model, such as a CT scan is used to represent the patient, and the information contained in the X-Ray image is used to determine the pose of the patient with respect to the intraoperative X-Ray imager. The majority of 2D/3D X-Ray registration methods may be classified as either “intensity-based” or “feature-based,” however we limit discussion to intensity-based methods in this paper.

Intensity-based registration performs an optimization over the relevant pose parameters, using an objective function that compares simulated radiographs, commonly referred to as digitally reconstructed radiographs (DRRs), with the intraoperative image [29]. The comparison is performed using a mathematical construct known as a similarity measure [29]. Due to the differences in X-Ray energy between preoperative CT and intraoperative fluoroscopy, the most effective similarity metrics compare the 2D gradients of a measured radiograph and a DRR, such as normalized cross-correlation between the Sobel gradient images (Grad-NCC) [29]. Robustness to metallic objects and bone fractures may be improved by taking a weighted sum of similarity measures computed in local regions of interest [41], [42]. Registration with multiple 2D views, and known relative poses between each view, is accomplished by creating DRRs at each view and summing the similarity scores for each view [27]. In order to register multiple objects with known shape, each object may be treated as a separate volume, and DRRs for each object are summed together to create a single DRR [28]. The registration problem for N object poses: $\theta_1, \dots, \theta_N$, with M intraoperative views: $I_1 \dots I_M$, a pre-operative CT: V , a DRR operator: $\mathcal{P}$, similarity metric: $\mathcal{S}$, and regularizer over plausible poses: $\mathcal{R}$, is concisely stated in (1).

$$\min_{\theta_1, \dots, \theta_N \in SE(3)} \sum_{m=1}^M \mathcal{S} \left(I_m, \sum_{n=1}^N \mathcal{P}(V; \theta_n) \right) + \mathcal{R}(\theta_1, \dots, \theta_N) \quad (1)$$

With the advent of general purpose GPU programming resources, Otake was able to efficiently form many DRRs simultaneously and use a state of the art “Covariance Matrix Adaptation: Evolutionary Search” (CMA-ES) optimization strategy [43] to carry out registration of a single femur using three views in under 22 seconds [27]. Relative pose information was computed using an external fiducial for a non-motorized C-Arm and was preoperatively calibrated when using a motorized C-Arm.

Several groups have demonstrated registration of multiple objects with intensity-based objective functions and accurate shape models, or with a statistical prior of the shape distributions.

In [28], Otake’s framework was extended to multiple objects for the knee joint (distal femur, patella, proximal tibia) tracking with bi-plane fluoroscopy. Initial registration times at the start of each sequence for the femur, tibia, and patella bones were between 2 and 5 minutes. All femur and tibia poses were estimated within 2° and 2 mm, and 74% of patella poses were estimated within the same thresholds.

Gong proposed to use intensity-based registration to intraoperatively estimate the position of bone fragments resulting from a distal radius fracture [22]. The approach requires preoperative knowledge of the bone fragment shapes and uses a preoperative, but post-trauma, CT scan [22]. Two 3D printed phantoms with synthetic fractures were used to test the method with four fluoroscopic views from a tracked C-Arm. Target registration errors (TREs) smaller than 3 mm were achieved when using a manual, interactive, initialization of the registration. Execution times of 3–9 minutes were reported using modest hardware.

In order to localize and determine the shape of carpal bones in the hand, Chen *et al.* use a 2D/3D registration of a single fluoroscopic view to 3D statistical shape and pose models of the carpal bones, radius, and ulna [20]. TREs of 2.45 mm were reported in simulation, and TREs from 0.93–2.37 were reported in the fluoroscopic experiments. Registration times were approximately 3 minutes per frame.

Although related by the motivation to track multiple objects with intensity-based registration, the aforementioned works do not provide a complete solution for the localization of an acetabular fragment. Most importantly, the fragment shape is not completely known pre-operatively and, to our knowledge, no statistical priors of pelvic osteotomies exist. In contrast to the fluoroscopic data collected of the phantom in [22], the fluoroscopic images of a human have more heterogeneous hard and soft tissue distributions, usually resulting in a more challenging registration. Additionally, objects such as metallic screws, wires, or tools may confound a registration strategy, and were not present in the fluoroscopic views used by [20], [22], [28].

Without any external fiducial objects or optical tracking systems, the methods proposed in this paper intraoperatively estimate an acetabular fragment's pose using X-Ray navigation. The pose estimation problem is solved by extending previous 2D/3D X-Ray registration techniques to partially estimate the acetabular fragment shape during surgery, and then localize the pelvis and fragment without a tracked, motorized, or encoded, X-Ray imager. Furthermore, our method aims to provide intraoperative feedback in a reasonable timeframe, with computation times on the order of seconds.

III. Materials and Methods

A. Preoperative Processing and Planning

Lower torso preoperative CT images were acquired and resampled to have 1 mm isotropic voxel spacing. Segmentation of the pelvis, and left and right femurs was performed using an automatic method [44], followed by manual touch up. The manual touch up was occasionally necessary to distinguish between the acetabulum of a pelvis and the corresponding femoral head. Using manually annotated landmarks, the APP was computed as described by [45] and the origin was relocated to the ipsilateral femoral head. The APP coordinate axes are aligned with the left/right (LR), inferior/superior (IS), and anterior/posterior (AP) anatomical axes. An example of the APP axis orientations is shown in Fig. 1. In order to report LCE angles intraoperatively, the most lateral points of the acetabulum are digitized from coronal slices of the preoperative CT volume.

Other landmarks identified for registration purposes were the medial and inferior points of maximal curvature on the obturator foramen, the greater sciatic notch, and inferior symphysis. To allow for accurate initialization of full pelvis registration in the presence of a mobilized fragment, the registration landmarks should not be located on portions of the pelvis that may become fragment.

A preoperative plan of the osteotomies is created by manual selection of several landmarks on the pelvis surface and cutting planes are fit to pass through the landmarks. Next, a label map is computed that indicates whether a voxel represents the acetabular bone fragment, the non-fragment pelvis, a portion of bone removed due to the chiseling action, left femur, or right femur. The osteotomy action is simulated by moving a virtual chisel, of width 1 mm, along the virtual cutting planes defined by the preoperative plan. Starting from the original segmentation, all pelvis labels contained within the convex hull defined by the virtual cutting planes are marked as a candidate fragment voxel. Furthermore, if a candidate fragment voxel is within a chisel width of any virtual cut plane, it is also marked as a candidate “cut” voxel. The set of candidate fragment voxels is divided into connected components, with the largest component kept as candidate fragment voxels, and the others reverted to pelvis labels. Of the remaining candidate fragment voxels, any marked as candidate cut voxels are labeled as cuts, and are otherwise labeled as fragment. Tight bounding boxes are computed about the pelvis labels, fragment labels, and femur labels to create sub-volumes for each object.

B. Intraoperative Registration Strategy

Our registration strategy is based on the methods described in [27] and [28]. However, the methods proposed here do not rely on external information to recover the multiple-view geometry, provide a quick single-landmark initialization, use a similarity metric robust to mismatches between the preoperative models and intraoperative reality, perform a faster local search at higher resolution levels, and attempt to recover the shape of intraoperatively created bone fragments without an additional CT.

DRRs are formed by ray casting through each object’s attenuation volume with a ray step size of 1 mm and trilinear interpolation. Grad-NCC scores are computed throughout image patches, with the mean value used as the similarity measure for all intensity-based registrations in this work [42]. Ray casting and similarity metric computation are computed on a GPU through the use of OpenCL [46], while optimization logic is conducted on the CPU. Image parameters and the projection geometry are derived from a Siemens CIOS Fusion C-Arm with 30 cm flat panel detector. Using DICOM metadata populated by a CIOS Fusion, we created a naïve set of intrinsic parameters for the C-Arm: image dimensions of 1536×1536 , isotropic pixel spacing of 0.194 mm/pixel, a source-to-detector distance of 1020 mm, no distortion, and a principal point at the center of the image.

We use two 2D image resolution levels in this work, starting from coarse ($8\times$ downsampling in each 2D dimension) and moving to a finer resolution ($4\times$ downsampling). At the coarsest resolution level, all optimizations are conducted with the CMA-ES strategy, and the Bounded Optimization by Quadratic Approximation (BOBYQA) strategy [47] is used at the second level. The CMA-ES search strategy is a derivative free method which requires a large number of objective function evaluations per iteration, but approximates a traditional

second-order algorithm while maintaining robustness to local minima [43]. Since the BOBYQA strategy requires only a single objective function evaluation per iteration, it has significantly faster runtimes compared to CMA-ES for this work. Even though BOBYQA does not provide the same robustness to local minima as CMA-ES, we justify its use at the second resolution level due to its efficiency, and because we assume that pose estimates after the first resolution level lie within a smooth, convex, region about the true minima.

The C-Arm is not required to be motorized, encoded, or capable of reporting relative pose information. In order to obtain the relative pose information required by the multi-object fragment registration, we treat the patient's anatomy as a "fiducial" object. Single-view registrations of the pelvis are performed for each view and the pelvis coordinate frame is then used as the C-Arm world frame.

Assuming the first fluoroscopic view is approximately AP, a single landmark is manually annotated in 2D to initialize pelvis registration. The 3D landmark location, with respect to the C-Arm, is estimated by traversing the source-to-detector ray of this landmark 85% of the source-to-detector distance. Orientation of the pelvis is computed using the APP coordinate frame described in Section III-A. Translation is recovered by aligning the landmark points in the two coordinate frames. The point of maximal curvature on the medial portion of the ipsilateral obturator foramen was chosen as the initialization landmark used in this work, and an example annotation is shown in the top left of Fig. 3. If the first view is not approximately AP, an initial pose estimate of the pelvis may be obtained by annotating more landmarks in 2D and solving the PnP problem [48]. Once the initial pose is estimated, a single-view intensity-based registration of the pelvis is performed.

By restricting the possible movements of the C-Arm for subsequent views, initial pelvis pose estimates for the remaining views are automatically obtained through an exhaustive search starting from the registered pose of the pelvis in the first view. More precisely, we limit the geometries of the remaining views to only differ from the first view by a rotation along the C-Arm's orbit. DRRs are computed by adjusting the registered pelvis pose of the initial view by C-Arm orbital rotations of $\pm 90^\circ$ in 1° increments, for a total of 181 DRRs. Similarity computations are made between the second and third fluoroscopic views, and the poses corresponding to the best similarity scores are used to initialize intensity-based registrations. An offline calibration process was conducted to determine center of orbital rotation and rotation axis for the CIOS Fusion.

Prior to any further registrations, the preoperative plan of the fragment may be refined as described in Section III-C.

Once the single view pelvis registrations are complete, the multiple-view geometry is recovered using the pelvis coordinate frame and the multiple-object registration is conducted. Specifically, we attempt to solve problem (1) through successive optimizations of individual objects, followed by a simultaneous optimization. At each resolution level, the pelvis pose estimate is refined starting from its current estimate. The pose of the pelvis is kept fixed, while an optimization over the pose of the femur is performed. Keeping the poses of the pelvis and femur fixed, the pose of the acetabular fragment is estimated. At the coarse

level, the femur and fragment poses are initialized with the current estimate of the pelvis pose, while previous poses are used for initialization at the second level. At the second level, a “simultaneous” optimization is performed after the “sequential” optimizations; which optimizes over all object poses simultaneously. An example of three views used for a fragment localization is shown in Fig 3.

The registration process produces pose estimates of the extrinsic C-Arm frame (C) with respect to the pelvis (PV) and fragment (FV) volumes: T_C^{PV} and T_C^{FV} . The relative pose of the fragment with respect to the APP is computed as:

$$\Delta_{APP} = T_{PV}^{APP} T_C^{FV} T_{PV}^C T_{APP}^{PV}. \quad (2)$$

T_{APP}^{PV} is obtained when estimating the APP from preoperative landmarks and maps points in the APP to the preoperative pelvis and fragment volumes. Using Δ_{APP} the intraoperative fragment movement may be visualized in 3D (see Fig. 1) and shown to the clinician. Additional information useful for the assessment of dysplasia may also be displayed, such as the LCE angle and a decomposition of the pose with respect to each anatomical axis. Similar to the approach in [49], the LCE angle is computed by applying the fragment’s relative pose to transform the 3D preoperative lateral trace of the acetabulum.

Given initial estimates for poses of each object, the optimization for each object is performed over the $\mathfrak{se}(3)$ Lie Algebra [50], with the $SE(3)$ reference point set from the initialization point. For a single view registration, an intermediate coordinate frame that is axis aligned with the camera projection frame, with origin at the initial estimate of the femoral head, is used. For multiple view registrations, the APP with origin at ipsilateral femoral head is used as an intermediate frame during the optimization. To discourage implausible poses, regularization is applied when using the CMA-ES optimization strategy. For single view registrations, regularization is applied separately to each translation component and the Euler decompositions of the rotation matrix, whereas the total rotation and translation magnitudes are used for multiple view registrations. Box constraints are used for all optimizations using the BOBYQA optimization strategy. Since the pelvis is initially registered with single-views, the optimization parameters used during the multi-view/multi-object registration for the pelvis are set for quicker execution time and tighter constraints in comparison to those used to register the femur and fragment.

Exact parameter values are listed in supplementary section S-1-B.

C. Fragment Shape Estimation

Since the fracture line introduced during bone chiseling is dependent on bone quality, the cut does not always follow a planned path. Therefore, it is not realistic for a clinician to exactly reproduce the planned cuts, even when the osteotome is guided by a navigation system. To account for the potential uncertainties associated with the osteotomies, we have developed a method to estimate the fragment shape; starting from a preoperative plan and refined from intraoperative fluoroscopic images.

For each osteotomy to be estimated, a post-osteotomy fluoroscopic view clearly showing the cut lines is required. The full pelvis shape is registered to this view and a user manually annotates 2D pixels along the cut lines visible on the main pelvis object. For each 2D label, a ray is cast from the C-Arm detector towards the X-Ray source, and the 3D intersection points with the pelvis surface are computed. The osteotomy is estimated by fitting a plane to the recovered 3D points. When looking approximately down the cut line, the ray is nearly tangent to the entire cut and 3D intersection points entering and leaving the pelvis surface are used. Otherwise, a label must indicate whether the ray intersects the 3D osteotomy when entering or exiting the pelvis surface; in this case only the entry or exit intersection point is used. Fig. 4 shows an example of the 2D labeling of ilium and pubis cut pixels.

If an appropriate view and corresponding pelvis registration is not available, then the preoperative plan for that osteotomy is used for shape estimation. For the ischial and posterior osteotomies, it is challenging to obtain a view that clearly shows the cut lines, and for which a pelvis registration may be successfully performed. Therefore, only the ilium and pubis osteotomies will be estimated in the following experiments, and the ischium and posterior osteotomies will be set at the planned values.

D. Simulation Study

In order to determine the feasibility of this registration strategy, we conducted a large simulation study using the CT scans of 6 non-dysplastic cadaveric specimens. The group of specimens consisted of 4 male and 2 female subjects, with ages ranging from 57 to 94 years (81 ± 14). For the left and right sides of each specimen, we created simulated acetabular fragments, simulated fragment movements, and simulated intraoperative fluoroscopic images.

All PAOs were performed bilaterally by a single surgeon (B.A.M.). In order to create nominal fragment shapes, planes were fit to the osteotomy contours in postoperative CT volumes and then mapped into preoperative CT coordinates. Random surgeries were simulated by applying random rigid transformations to the cutting planes associated with each nominal osteotomy. The random adjustments were manually verified to create plausible PAO fragments and any invalid fragments were rejected. This resulted in 15 simulated fragments for each side of each patient and 180 for the entire study.

Random movements of each fragment and ipsilateral femur were sampled. Random rotations of the fragment and femur were sampled along with a random translation component for both the fragment and femur. The translation component is shared, since the fragment and femur tend to move together. To mimic clinically relevant movements of the fragment and femur, Euler angles and translations about each anatomical axis were sampled. All transformations are with respect to the APP with appropriate femoral head origin. Collision detection was performed to ensure that different bones did not overlap in 3D space. When a collision occurred, the set of transformations was rejected, and another set was sampled. A total of 20 movements for each fragment were sampled, with a mean rotation of $15.49^\circ \pm 6.77^\circ$ and mean translation of $4.35 \text{ mm} \pm 1.57 \text{ mm}$. This yields a total of 3,600 fragment movements for this study.

Three views are used for each repositioned fragment: a perturbed AP view, followed by two views offset at random rotations about the C-Arm's orbit. The simulated fluoroscopy incorporates soft-tissues by piece-wise rigidly warping the original CT volume according to the pelvis, femur, and fragment labels as described in [42]. A temporary fixation of the fragment is simulated by inserting two random K-wires into the volume, each at a random pose intersecting the ilium and fragment. Using a protocol similar to [51], fluoroscopic images are created from this new volume and view geometries. An example set of simulated fluoroscopic images is shown in the bottom row of Fig. 3.

The registration process proceeds according to the strategy described in III-B. Registrations are initialized by simulating a 2D and 3D point picking process for the single landmark strategy. Using the registration point manually identified in each CT, the corresponding 2D pixel location is computed for each set of fluoroscopic views. Random noise is added to the 2D and 3D points to simulate variation in human point picking behavior. The landmarks are offset by a random distance along a random direction. The offset directions are drawn uniformly from the appropriate (2D or 3D) sphere. The 3D offset magnitudes are sampled, in mm, from $\mathcal{N}(0, 3)$, and the 2D magnitudes are sampled, in pixels, from $\mathcal{N}(0, 7.5)$. Five different initializations are estimated for each set of fluoroscopic views.

These simulations result in a total of 18,000 multiple-object/multiple-view registrations that need to be evaluated. Each registration is run over four scenarios; each with a different fragment shape used during the registration. The first case considered is when the fragment shape in the fluoroscopic views exactly matches the preoperative plan. We also considered the case of tracking a fragment with an unknown, or uncertain shape. To achieve this we assumed that the nominal fragments represent preoperative plans for the randomly adjusted fragments present in the simulated fluoroscopic images. The remaining two cases examined the effect of using intraoperative information to update the planned cut model. For the first case of incorporating intraoperative knowledge, it was assumed that the ilium and pubis osteotomies would be perfectly recovered; the planned ilium and pubis cuts were replaced with the actual values. In the second case, synthetic 2D cut annotations were used to simulate the fragment shape estimation described in Section III-C.

Further details regarding the simulated datasets are described in supplementary section S-1-D.

E. Cadaver Study

In order to further evaluate the proposed methods, intraoperative fluoroscopy was collected during six of the PAOs previously used for simulation studies. Four specimens, three male and one female, were used. Each surgery was performed under fluoroscopic guidance with a Siemens CIOS Fusion C-Arm with 30 cm flat panel detector. Fig. 5 shows a photograph of the operating theater. Each fluoroscopy image of the repositioned fragments contains screws, K-wires, or a combination of the two.

Fragment registrations were evaluated using the true fragment shape, preoperative planned fragments, and intraoperatively estimated fragment shapes. Postoperative CT scans were taken, and the fragment shape was manually segmented, to obtain the true fragment shape.

Fifteen preoperative plans for each surgery were created by fitting cutting planes to the manually segmented fragment, followed by random adjustments. Each preoperative plan was manually verified to be a valid PAO, with any invalid plan rejected and resampled. The 2D cut lines of the ilium and pubis osteotomies were manually annotated in the fluoroscopic images, and were used to compute the ilium and pubis cutting planes of the estimated fragment shape. Each estimated shape is completed with the planned ischium and posterior osteotomies.

Included with this paper is a supplementary video, which illustrates the workflow used to perform the fragment shape and pose estimation for the left side of the specimen shown in Figures 1, 3, 4, and 5. This will be available at <http://ieeexplore.ieee.org>.

Ground truth poses of each fragment were obtained using metallic BBs implanted into the pelvis. The postoperative CT volumes contain the relocated fragment which had been intraoperatively fixed in place using screws and K-wire. Prior to osteotomy, BBs were implanted onto the interior ilium surface in a region expected to lie on the fragment, and additional BBs were implanted onto the interior ilium surface in a region expected to not lie on the fragment. The ground truth fragment poses are calculated with 3D/3D paired-point registrations [52] between the 3D locations of the fragment BBs before osteotomy, and the 3D locations after osteotomy. For two specimens, and three PAOs, the fragment BB locations were computed using 3D CT imaging only. Eight BBs were placed on the ilium region and another eight placed on the fragment, with each eight BB constellation made of four 1.5 mm diameter BBs and four 1 mm diameter BBs. A Halifax Biomedical Inc. injection device was used to insert the 1 mm diameter BBs. One of our clinical co-authors, C.P.A., inserted these BBs after performing soft-tissue dissections. An additional preoperative CT of the specimens was obtained to recover the pre-osteotomy BB locations. For the remaining two specimens, and remaining three PAOs, four BBs (1.5 mm diameter) were inserted on the ilium region and four BBs (1.5 mm diameter) inserted on the fragment region. No additional preoperative CT was obtained, however a series of fluoroscopic images were taken prior to osteotomy. Using the BB constellations obtained from the postoperative CT, the pre and post-osteotomy 3D locations of the BBs were obtained through a series of 2D/3D landmark registrations on the fluoroscopic images. A similar approach, using a plastic pelvis phantom with injected BBs and three views from CBCT projections, showed that the fragment pose may be recovered within 1°/1 mm of CBCT ground truth [53], [54]. Details of BB insertion and the methods used to obtain ground truth poses are described in supplementary section S-1-E. The BBs were not present in the preoperative CTs used to drive the intensity-based registrations and did not aid the proposed methods in any way.

IV. Results

A. Simulation Study

Fragment pose estimation was most accurate when the exact fragment shape was used during registration, and accuracy monotonically decreased as knowledge of the shape decreased. For each simulation in which the osteotomies were at least partially known or estimated, the mean rotation and translation errors were less than 2° and 2 mm, respectively.

Mean fragment pose errors of 2.2° and 2.1 mm were obtained from the simulation using preoperatively planned shapes. For all fragment shapes, mean LCE errors were less than 1° . Table I lists the mean and standard deviations of the fragment pose and LCE errors for each simulation study. Fig. 6 shows the joint histograms for the fragment pose estimation errors. The shape of the error distribution “widens” as the amount of uncertainty associated with the fragment shape increases, particularly with respect to rotation error. The percentages of registrations with rotation errors below 3° were 96%, 83%, 92%, and 91%, for the cases of all cuts known, no cuts known (preoperatively planned), known ilium and pubis cuts, and estimated ilium and pubis cuts, respectively. Similarly, the corresponding percentages of trials with LCE errors below 3° were 99%, 97%, 98%, and 98%.

Single-tailed Mann-Whitney U-Tests were performed to determine any statistical significance between the errors obtained when running registration with different fragment shapes. The p-values are shown in Table II, and indicate that pose and LCE estimation errors associated with the various categories of fragment shape are statistically different. Moreover, the single-tailed test indicates that the errors associated with fragment shapes using less intraoperative osteotomy information are statistically larger than errors determined by fragment shapes incorporating more knowledge of the intraoperative cuts.

B. Cadaver Study

A summary of the fragment pose and LCE error statistics is provided in Table III. Two results, one corresponding to a planned fragment shape and another to an estimated shape, produced fragment rotation errors above 20° and were discarded from analysis. The pose estimations with manual segmented fragments, planned fragments, and estimated fragments resulted in mean rotation/translation errors of $2.2^\circ/2.2$ mm, $3.8^\circ/2.5$ mm, and $3.5^\circ/2.5$ mm, respectively. Mean LCE errors were below 2° for all fragment shapes. Additionally, the standard deviations of all rotation and LCE errors strictly increased as knowledge of the fragment’s 3D shape decreased. Rotation errors were less than 3° for 83%, 39%, and 39% of the registration trials when using a manual segmentation of the fragment, a planned fragment, and an estimated fragment, respectively. The corresponding proportions of LCE errors less than 3° were 100%, 85%, and 85%.

Table IV shows p-values from statistical tests of significance conducted on the rotation, translation, and LCE errors. The protocol was identical to that used for evaluating the simulation study results. Using a threshold of $p = 0.005$, no statistical differences are indicated between the rotation, translation, or LCE errors when comparing the various fragment shapes.

The single-landmark initialization strategy was used for four of the six cases, and multiple-landmark initialization was used for the remaining two. In the two cases requiring multiple-landmark initialization, the initial views were taken with the detector parallel to the operating table and did not match a typical AP view due to the pelvic tilt of the specimens.

An expected runtime of 25.3 seconds was obtained by re-running the estimated-fragment pose estimation pipeline 20 times for each surgery. The expected time required to manually choose a single 2D landmark, required for registration initialization, was 2 seconds; the

expected time for manual annotation of 2D cut lines, required for fragment shape estimation, was 51.2 seconds.

The ground truth fragment poses from each surgery, the error statistics for each individual surgery, violin plots of the error distributions, and a breakdown of timings across the individual components of the registration pipeline are provided in supplementary section S-2-B.

V. DISCUSSION

The simulation study yielded promising results for the localization of the acetabular fragment. As one would expect, the pose errors were smallest on average when the exact shape of each bone was known during registration, and degraded as less information about the osteotomies was known. This increase in average error was consistent across total rotation and translation magnitudes, the decompositions about anatomical axes, and for the LCE angle estimates. The increases in rotation error magnitudes and LCE angle errors were statistically significant for all comparisons and highlight the impact of preoperative/intraoperative fragment shape mismatch on registration accuracy. Increases in the translation error magnitudes were statistically significant in all but one comparison: using true pubis and ilium osteotomies, mixed with planned ischial and posterior osteotomies, compared with estimating the pubis and ilium osteotomies. This also indicates that the fragment localization is partially robust to differences between the planned and intraoperative osteotomies for the ischial and posterior cuts.

To our knowledge, no bounds on the accuracy of fragment pose or LCE angle have been identified in previous literature, therefore we consider 3° to be an acceptable upper bound on LCE angle error. This threshold allows the clinician to target LCE angles between 33° – 37° , while having confidence that the true LCE angle is within the 30° – 40° criteria put forth by [9]. At this threshold, the lowest success rate of 97% corresponds to a preoperatively planned fragment shape, and indicates that an arbitrary preoperative plan is sufficient to obtain a successful registration. However, if the LCE error threshold is lowered to 2° , successful pose estimates were found in 98%, 92%, 97%, and 96% of trials, for the cases of the fragment shape with all cuts known, the preoperatively planned fragment shape, the shape with known ilium and pubis cuts, and the shape with estimated ilium and pubis cuts, respectively. Therefore, a clinician may attach more confidence to pose estimates obtained after estimating the fragment's ilium and pubis cuts.

In cadaver surgeries, registrations using each fragment shape yielded average LCE errors less than 2° . This indicates that the proposed pipeline is capable of producing clinically usable estimates on real fluoroscopic imagery. In terms of average rotational and LCE error, registrations using the manually segmented fragment shape outperformed registrations using preoperatively planned shapes and intraoperatively refined shapes. However, when examining translation dimensions, the manually segmented fragment did not offer a noticeable advantage over the remaining two shapes. The performance across all error metrics was roughly equivalent when comparing the registrations using the preoperative plan and estimated fragments. Moreover, when comparing the errors produced through

registrations using differing shape types, no statistically significant differences were found. This is in stark contrast to the simulation study results, which imply that an estimated shape should perform statistically better than a preoperatively planned fragment.

We believe this is primarily caused by low bone qualities from two specimens used for three surgeries. Prior to any osteotomies made to free the acetabular fragment, a small osteotomy of the anterior superior iliac spine (ASIS) is performed in order to provide sufficient access to the ilium. At the conclusion of the surgery, the ASIS is reattached in its original location. The ASIS osteotomy was excessively large for two specimens, resulting in a fragment shape with an additional osteotomy not representable by our four-cut model. When these three surgeries are removed from the results, 100% of the registrations using manually segmented fragment shapes, 95% using preoperatively planned fragment shapes, and 100% using estimated shapes produce LCE errors less than 3° . Furthermore, mean rotation errors of the fragment pose were 2.4° , 4.0° , and 3.1° when using a manually segmented fragment shape, a preoperative plan, and estimated shape, respectively. Although no statistical difference was found between these error distributions, the results are more consistent with the trend established by the simulation study of increasing errors, corresponding to registrations using more mismatched fragment shapes. Since most patients undergoing PAO are young adults, bone quality is not expected to be an issue in clinical use. Considering these factors, we can say that the fragment pose estimation using an estimated shape provided clinically relevant results to determine femoral head coverage, with LCE errors less than 3° .

The mean rotation error of 3.5° , when using the estimated fragment shape during the cadaver surgeries, is noticeably larger than the 1.8° error reported by a state-of-the-art optically tracked solution [12]. Translation errors found with the proposed system are roughly equivalent to the 2.2 mm identified by the optically tracked method. Although the X-Ray navigation approach has inferior rotation performance compared to the optically tracked method, the previous analysis of LCE angle errors indicate it is capable of providing clinically appropriate pose estimates.

The amount of femoral head coverage is primarily derived from the series of rotations applied to the acetabular fragment about the femoral head. For this reason, we report errors using rotation and translation components of the pose differences between registered estimates and ground truth poses. Furthermore, large TREs at regions distant from the cartilage surface, such as the ilium osteotomy, do not imply an incorrect estimate of the femoral head coverage.

Given the promising results in cadaveric experiments, and the diversity of data used in simulation, we believe this method should perform well on human subjects. Although we did not evaluate the proposed methods on dysplastic hips, our method should not suffer any performance degradation when run on dysplastic hips of younger patients. We have not made any assumptions in our processing specific to normal hip anatomy, and we only require an intensity volume of the pelvis that we may compute DRRs from. Some dysplastic hips may be subject to severe arthritis, which would likely cause difficulty during preoperative segmentation and also create a less-smooth registration similarity metric,

however these patients would most likely undergo a total hip replacement (THR) instead of pelvic osteotomy.

The expected computation time of approximately 25 seconds is reasonable for intraoperative operation, however this does not include the time required to perform any manual annotations. Since the fragment will not change shape after the osteotomies are completed, annotation of the cut lines only needs to be performed once. It is likely that several adjustments to the fragment's pose will be required, therefore we believe that the cut annotation time is amortized over each of these adjustments. A similar amortization is not applicable to registration re-initialization via landmark identification. However, other techniques using neural networks for recognizing fluoroscopic landmarks [55], or for approximate pose regression [56], could be applied to avoid this bottleneck.

Although implicitly penalized by the image similarity metric, the proposed registration method does not penalize 3D collision or overlap of bones. It is possible that the method's performance may improve by including collision and overlap information into the regularization term of (1). When the fragment shape used during registration does not match the true fragment shape, it is possible for collisions to occur, particularly about the ischial and posterior osteotomies. Therefore, it may be possible to recover more accurate shapes by updating the volumetric segmentation to remove collisions after registration. Shapes could be refined by iteratively switching between registration and shape updates.

The six cadaver surgeries have served as proof-of-concept for the methods proposed in this work, however we believe further efforts should include clinical PAO cases. This would include the acquisition of clinical PAO fluoroscopic data used for further validation, and move towards a comparison of patient outcomes and operative times of non-navigated, optically navigated, and X-Ray navigated cases. The system could also integrate a preoperative plan of the biomechanically optimal fragment reposition, and intraoperatively provide feedback consisting of adjustments that would maximize the likelihood of achieving the plan.

Although the experiments in this paper were restricted to PAO, we have not made any assumptions that would preclude application to other pelvic osteotomies, such as RAO, CPO, or triple osteotomy. Forty percent of THR surgeries result in a mal-positioned (with respect to the "safe zone") acetabular component [57], therefore we believe that our fragment tracking approach would also be useful for the intraoperative visualization, and guidance, of the THR acetabular implant.

Although the use of a preoperative CT scan is an increasingly common practice, it is not considered the current standard of care at all centers. When CT scans are used for diagnosis or preoperative planning, slice spacings are typically in the range of 2–3 mm [10]. It is not clear what effect lower resolution CT data will have on the accuracy of the proposed system and remains an important topic to be investigated. Furthermore, higher resolution preoperative CT scans may add additional cost and expose the patient to, potentially hazardous, radiation. In the future, a "CT free" approach could be implemented using a statistical deformation model (SDM) created from a database of existing pelvic CTs [58].

The SDM would be used in a deformable 2D/3D registration to recover the patient's 3D anatomy from standard preoperative standing radiographs and several pre-osteotomy fluoroscopic images [59]. However, the SDM would most likely be unable to recover a sufficiently accurate cartilage model required for biomechanical measures and analysis. This could be overcome through the use of a partial preoperative CT, scanning only the acetabulum and extrapolating the remaining anatomy with the SDM, similar to the methods proposed in [60] and [61].

VI. Conclusion

The proposed method for intraoperative localization of a mobilized acetabular fragment, with uncertain shape, has been demonstrated to report LCE angles within clinical tolerance in simulated cases and also in cadaveric cases for PAO. Less common equipment used in other approaches, such as an encoded or motorized CBCT C-Arm or an optical tracking device, is not required by this method. Additionally, the processing runs in an amount of time which is reasonable for incorporation into an intraoperative workflow. As entry-level flat panel imaging technology becomes more common in operating theaters, our method should be directly applicable to pelvic osteotomies or other similar fragment tracking tasks. The proposed X-ray navigation could make pelvic osteotomies more accessible to novice surgeons, reduce the number of times a fragment needs to be adjusted during a procedure, and possibly reduce the radiation exposure to the surgical team by reducing the number of fluoroscopic images used.

Supplementary Material

Refer to Web version on PubMed Central for supplementary material.

Acknowledgment

The authors would like to thank D. Boston for his assistance during cadaveric testing. Prof. Y. Sato provided productive and insightful discourse during the algorithmic development of this paper, for which we are very grateful. In addition, the authors are indebted to Dr. M. Takao and Dr. N. Sugano for allowing us to observe an RAO procedure and become aware of the challenges involved. The authors also thank Prof. M. Unberath for helpful discussions. Finally, we are appreciative of the constructive comments and critiques provided by the anonymous reviewers, which have resulted in an improved paper. This feedback led to the single-landmark registration initialization strategy, and has greatly improved the system's ease of use. Part of this research project was conducted using computational resources at the Maryland Advanced Research Computing Center (MARCC).

This work was supported in part by NIH/NIBIB under Grants R01EB006839 and R21EB020113, in part by the MEXT/JSPS KAKENHI under Grant 26108004, in part by the Johns Hopkins University Internal Funds, and in part by the Johns Hopkins University Applied Physics Laboratory Graduate Student Fellowship.

References

- [1]. Murphy SB, Ganz R, and Müller M, "The prognosis in untreated dysplasia of the hip. A study of radiographic factors that predict the outcome," *J. Bone Joint Surg. Amer.*, vol. 77, no. 7, pp. 985–989, 1995. [PubMed: 7608241]
- [2]. Ganz R et al., "A new periacetabular osteotomy for the treatment of hip dysplasias technique and preliminary results," *Clin. Orthopaedics Related Res*, vol. 232, pp. 26–36, 1988.
- [3]. Armand M et al., "Outcome of periacetabular osteotomy: joint contact pressure calculation using standing AP radiographs, 12 patients followed for average 2 years," *Acta Orthopaedics*, vol. 76, no. 3, pp. 303–313, 2005.

- [4]. Armiger RS et al., "Three-dimensional mechanical evaluation of joint contact pressure in 12 periacetabular osteotomy patients with 10-year follow-up," *Acta Orthopaedics*, vol. 80, no. 2, pp. 155–161, 2009.
- [5]. Niknafs N et al., "Biomechanical factors in planning of periacetabular osteotomy," *Front Bioeng. Biotechnol*, vol. 1, p. 20, 2013. [PubMed: 25152876]
- [6]. Hipp JA et al., "Planning acetabular redirection osteotomies based on joint contact pressures," *Clin. Orthopaedics Related Res*, vol. 364, pp. 134–143, 1999.
- [7]. Troelsen A, "Surgical advances in periacetabular osteotomy for treatment of hip dysplasia in adults," *Acta Orthopaedics*, vol. 80, no. suppl. 332, pp. 1–33, 2009.
- [8]. Wiberg G, "Studies on dysplastic acetabulum and congenital subluxation of the hip joint with special reference to the complications of osteoarthritis," *Acta Chir. Scand*, vol. 83, no. 58, pp. 53–68, 1939.
- [9]. Hartig-Andreasen C et al., "What factors predict failure 4 to 12 years after periacetabular osteotomy?" *Clin. Orthopaedics Related Res*, vol. 470, no. 11, pp. 2978–2987, 2012.
- [10]. Troelsen A et al., "Assessment of hip dysplasia and osteoarthritis: Variability of different methods," *Acta Radiol*, vol. 51, no. 2, pp. 187–193, 2010. [PubMed: 20144145]
- [11]. Langlotz F et al., "Computer assistance for pelvic osteotomies," *Clin. Orthopaedics Related Res*, vol. 354, pp. 92–102, 1998.
- [12]. Murphy RJ et al., "Development of a biomechanical guidance system for periacetabular osteotomy," *Int. J. Comput. Assisted Radiol. Surgery*, vol. 10, no. 4, pp. 497–508, 2015.
- [13]. Mayman DJ et al., "The kingston periacetabular osteotomy utilizing computer enhancement: A new technique," *Comput. Aided Surg*, vol. 7, no. 3, pp. 179–186, 2002. [PubMed: 12362378]
- [14]. Liu L et al., "Computer assisted planning and navigation of periacetabular osteotomy with range of motion optimization," in *Proc. Med. Image Comput. Comput. Assisted Intervention*, 2014, pp. 643–650.
- [15]. Akiyama H et al., "Computed tomography-based navigation for curved periacetabular osteotomy," *J. Orthopaedics Sci*, vol. 15, no. 6, pp. 829–833, 2010.
- [16]. Takao M et al., "Comparison of rotational acetabular osteotomy performed with navigation by surgeons with different levels of experience of osteotomies," *Int. J. Comput. Assisted Radiol. Surgery*, vol. 12, no. 5, pp. 841–853, 2017.
- [17]. Troelsen A, Elmengaard B, and Søballe K, "A new minimally invasive transsartorial approach for periacetabular osteotomy," *J. Bone Joint Surg. Amer*, vol. 90, no. 3, pp. 493–498, 2008. [PubMed: 18310698]
- [18]. Troelsen A, Elmengaard B, and Søballe K, "Comparison of the minimally invasive and ilioinguinal approaches for periacetabular osteotomy 263 single-surgeon procedures in well-defined study groups," *Acta Orthopaedics*, vol. 79, no. 6, pp. 777–784, 2008.
- [19]. Belei P et al., "Fluoroscopic navigation system for hip surface replacement," *Comput. Aided Surg*, vol. 12, no. 3, pp. 160–167, 2007. [PubMed: 17538788]
- [20]. Chen X et al., "Automatic inference and measurement of 3D carpal bone kinematics from single view fluoroscopic sequences," *IEEE Trans. Med. Imag*, vol. 32, no. 2, pp. 317–328, Feb. 2013.
- [21]. Dennis DA et al., "In vivo determination of normal and anterior cruciate ligament-deficient knee kinematics," *J. Biomechan*, vol. 38, no. 2, pp. 241–253, 2005.
- [22]. Gong RH, Stewart J, and Abolmaesumi P, "Multiple-object 2-D–3-D registration for noninvasive pose identification of fracture fragments," *IEEE Trans. Biomed. Eng*, vol. 58, no. 6, pp. 1592–1601, Jun. 2011. [PubMed: 21233043]
- [23]. Guéziec A et al., "Anatomy-based registration of CT-scan and intraoperative X-ray images for guiding a surgical robot," *IEEE Trans. Med. Imag*, vol. 17, no. 5, pp. 715–728, Oct. 1998.
- [24]. Joskowicz L et al., "FRACAS: A system for computer-aided image-guided long bone fracture surgery," *Comput. Aided Surg*, vol. 3, no. 6, pp. 271–288, 1998. [PubMed: 10379977]
- [25]. Mahfouz MR et al., "A robust method for registration of three-dimensional knee implant models to two-dimensional fluoroscopy images," *IEEE Trans. Med. Imag*, vol. 22, no. 12, pp. 1561–1574, Dec. 2003.

- [26]. Markelj P et al., "A review of 3D/2D registration methods for image-guided interventions," *Med. Image Anal.*, vol. 16, no. 3, pp. 642–661, 2012. [PubMed: 20452269]
- [27]. Otake Y et al., "Intraoperative image-based multi-view 2D/3D registration for image-guided orthopaedic surgery: Incorporation of fiducial-based C-arm tracking and GPU-acceleration," *IEEE Trans. Med. Imag.*, vol. 31, no. 4, pp. 948–962, Apr. 2012.
- [28]. Otake Y. Robust patella motion tracking using intensity-based 2D-3D registration on dynamic bi-plane fluoroscopy: Toward quantitative assessment in MPFL reconstruction surgery; *Proc. SPIE*; 2016. vol. 9786Art. no. 97 860B.
- [29]. Penney GP et al., "A comparison of similarity measures for use in 2-D-3-D medical image registration," *IEEE Trans. Med. Imag.*, vol. 17, no. 4, pp. 586–595, Aug. 1998.
- [30]. Sheng-fu LL et al., "Automatic localization of target vertebrae in spine surgery: Clinical evaluation of the LevelCheck registration algorithm," *Spine*, vol. 40, no. 8, 2015, Art. no. E476.
- [31]. Yao J et al., "A C-arm fluoroscopy-guided progressive cut refinement strategy using a surgical robot," *Comput. Aided Surg.*, vol. 5, no. 6, pp. 373–390, 2000. [PubMed: 11295851]
- [32]. Yi T et al., "Robotic drill guide positioning using known-component 3D–2D image registration," *J. Med. Imag.*, vol. 5, no. 2, 2018, Art. no. 021212.
- [33]. Langlotz F et al., "The first twelve cases of computer assisted periacetabular osteotomy," *Comput. Aided Surg.*, vol. 2, no. 6, pp. 317–326, 1997. [PubMed: 9587693]
- [34]. Naito M et al., "Curved periacetabular osteotomy for treatment of dysplastic hip," *Clin. Orthopaedics Related Res.*, vol. 433, pp. 129–135, 2005.
- [35]. Radermacher K et al., "Computer assisted orthopaedic surgery with image based individual templates," *Clin. Orthopaedics Relat. Res.*, vol. 354, pp. 28–38, 1998.
- [36]. Tönnis D et al., "Triple pelvic osteotomy," *J. Pediatric Orthopaedics B*, vol. 3, no. 1, pp. 54–67, 1994.
- [37]. Otsuki B et al., "Developing a novel custom cutting guide for curved periacetabular osteotomy," *Int. Orthopaedics*, vol. 37, no. 6, pp. 1033–1038, 2013.
- [38]. Murphy RJ et al., "Clinical evaluation of a biomechanical guidance system for periacetabular osteotomy," *J. Orthopaedics Surg. Res.*, vol. 11, no. 1, p. 36, 2016.
- [39]. Liu L et al., "Periacetabular osteotomy through the pararectus approach: technical feasibility and control of fragment mobility by a validated surgical navigation system in a cadaver experiment," *Int. Orthopaedics*, vol. 40, no. 7, pp. 1389–1396, 2016.
- [40]. Ninomiya S and Tagawa H, "Rotational acetabular osteotomy for the dysplastic hip," *J. Bone Joint Surg. Amer.*, vol. 66, no. 3, pp. 430–436, 1984. [PubMed: 6699061]
- [41]. Knaan D and Joskowicz L, "Effective intensity-based 2D/3D rigid registration between fluoroscopic X-ray and CT," in *Medical Image Computing and Computer-Assisted Intervention*. Berlin Germany: Springer, 2003, pp. 351–358.
- [42]. Grupp RB, Armand M, and Taylor RH, "Patch-based image similarity for intraoperative 2D/3D pelvis registration during periacetabular osteotomy," in *Proc. Int. Workshop Clin. Image Based Procedures*, 2018, pp. 153–163.
- [43]. Hansen N and Ostermeier A, "Completely derandomized self-adaptation in evolution strategies," *Evol. Comput.*, vol. 9, no. 2, pp. 159–195, 2001. [PubMed: 11382355]
- [44]. Kr ah M, Székely G, and Blanc R, "Fully automatic and fast segmentation of the femur bone from 3D-CT images with no shape prior," in *Proc. IEEE Int. Symp. Biomed. Imag.*, 2011, pp. 2087–2090.
- [45]. Nikou C et al., "Description of anatomic coordinate systems and rationale for use in an image-guided total hip replacement system," in *Proc. Med. Image Comput. Comput. Assisted Intervention*, 2000, pp. 1188–1194.
- [46]. Stone JE, Gohara D, and Shi G, "OpenCL: A parallel programming standard for heterogeneous computing systems," *Comput. Sci. Eng.*, vol. 12, no. 3, pp. 66–73, 2010. [PubMed: 21037981]
- [47]. Powell MJ, "The BOBYQA algorithm for bound constrained optimization without derivatives," *University of Cambridge, Cambridge, U.K.*, Rep. NA2009/06, 2009.
- [48]. Hartley R and Zisserman A, *Multiple View Geometry in Computer Vision*. Cambridge, U.K.: Cambridge Univ. Press, 2003.

- [49]. Armiger RS et al., "Evaluation of a computerized measurement technique for joint alignment before and during periacetabular osteotomy," *Comput. Aided Surg*, vol. 12, no. 4, pp. 215–224, 2007. [PubMed: 17786597]
- [50]. Murray RM et al., *A Mathematical Introduction to Robotic Manipulation*. Boca Raton, FL, USA: CRC Press, 1994.
- [51]. Markelj P, Likar B, and Pernuš F, "Standardized evaluation methodology for 3D/2D registration based on the visible human data set," *Med. Phys*, vol. 37, no. 9, pp. 4643–4647, 2010. [PubMed: 20964183]
- [52]. Horn BK, "Closed-form solution of absolute orientation using unit quaternions," *J. Opt. Soc. Amer. A*, vol. 4, no. 4, pp. 629–642, 1987.
- [53]. Armand M et al., "Biomechanical guidance system for periacetabular osteotomy," in *Intelligent Orthopaedics*. Berlin, Germany: Springer, 2018, pp. 169–179.
- [54]. Murphy R et al., "Computer-assisted X-ray image-based navigation of periacetabular osteotomy with fiducial based 3D acetabular fragment tracking," *Orthopaedics Proc*, vol. 95, no. SUPP_28, 2013, Art. no. 84.
- [55]. Bier B et al., "X-ray-transform invariant anatomical landmark detection for pelvic trauma surgery," in *Medical Image Computing and Computer Assisted Intervention*. Berlin, Germany: Springer, 2018, pp. 55–63.
- [56]. Kendall A, Grimes M, and Cipolla R, "PoseNet: A convolutional network for real-time 6-DOF camera relocation," in *Proc. IEEE Int. Conf. Comput. Vis*, 2015, pp. 2938–2946.
- [57]. Callanan MC et al., "The john charnley award: Risk factors for cup malpositioning: Quality improvement through a joint registry at a tertiary hospital," *Clin. Orthopaedics Relat. Res*, vol. 469, no. 2, pp. 319–329, 2011.
- [58]. Otake Y. Comparison of optimization strategy and similarity metric in Atlas-to-subject registration using statistical deformation model; *Proc. SPIE*; 2015. vol. 94 15Art. no. 94 150Q.
- [59]. Sadowsky O, Chintalapani G, and Taylor R, "Deformable 2D-3D registration of the pelvis with a limited field of view, using shape statistics," in *Proc. Med. Image Comput. Comput. Assisted Interv*, 2007, pp. 519–526.
- [60]. Grupp R et al., "Pelvis surface estimation from partial CT for computer-aided pelvic osteotomies," *Orthopaedics Proc*, vol. 98, no. SUPP_5, 2016, Art. no. 55–55.
- [61]. Chintalapani G. Statistical atlas based extrapolation of CT data; *Proc. SPIE*; 2010. vol. 7625Art. no. 762 539.

Fig. 1.

Example of periacetabular osteotomies and a fragment reposition. This is taken from the cadaver experiments and represents the ground truth fragment pose. The manually segmented fragment shape is shown. See Fig. 3 for the corresponding intraoperative fluoroscopic images. The orientations of each anatomical axis for the anterior pelvic plane are also depicted: left/right x axis (LR), inferior/superior y axis (IS), and anterior/posterior z axis (AP).

Fig. 2. High level workflow detailing the steps performed during the surgery and the data required for each. The registration and shape estimation steps are the primary focus of this paper.

Fig. 3.

Top Row: Intraoperative fluoroscopic images used for registering the fragment on the left side of a cadaveric specimen. Note the four BBs on the mobilized fragment and four BBs on the ilium used for ground truth computation. The left image shows an approximate AP view and the yellow circle indicates the single-landmark used for initialization of the intraoperative pipeline. Middle Row: DRRs created as part of the intensity-based registration using manual segmentation of the fragment; the final pose estimates are shown. Prior to performing any similarity comparisons, the images in the top-row will be log-corrected. Bottom Row: An example of simulated fluoroscopic data generated using the protocol of the simulation study. The image includes soft-tissue and inserted K-wires, with object poses determined by the registrations from the middle row. Despite the visual similarity with the top row, the challenge of visually determining a fragment's pose is confirmed by an approximate 3°/3 mm pose error difference from the top row's ground truth. See Fig. 1 for the ground truth 3D view.

Fig. 4.

An example of the 2D cut manual annotations from fluoroscopic images corresponding to the case shown in Fig. 1. Labels identifying ilium cuts are shown in (a), and labels identifying pubis cuts are shown in (b). Zoomed displays of the regions of interest in (a) and (b) are shown in (c) and (d), respectively. In (a) and (c) beige labels indicate rays that enter the pelvis surface when moving from the detector to source, while blue labels indicate rays that exit the pelvis surface. In (b) and (d), red labels indicate rays that both enter and exit the pelvis surface; the view looks approximately down the cut line.

Fig. 5. Intraoperative setup for the PAO procedure performed on the right side of a cadaveric specimen. The current fluoroscopic view is shown in the bottom left. The enlarged region in the top right shows the markings used to control the orbital rotation of the C-Arm.

Fig. 6.

Normalized 2D histograms of fragment pose error for the simulation studies. (a) The actual fragment shape is known and matches the planned fragment shape, (b) the actual fragment shape is not known and none of the osteotomies match the planned fragment, (c) the actual fragment is partially known, with the ilium and pubis osteotomies matching the planned fragment, (d) the actual fragment is not known, but the ilium and pubis osteotomies are estimated from 2D cut lines.

TABLE I

Simulation Study Fragment Average Pose and Lateral Center Edge Angle Errors

Study	Rotation (°)				Translation (mm)				LCE (°)
	AP	LR	IS	Total	AP	LR	IS	Total	
All Cuts Known	0.47 ± 1.48	0.73 ± 1.35	0.67 ± 1.14	1.25 ± 2.24	1.54 ± 1.29	0.32 ± 0.59	0.43 ± 0.59	1.73 ± 1.42	0.50 ± 1.53
No Cuts Known	0.60 ± 1.62	1.53 ± 1.77	1.07 ± 1.29	2.19 ± 2.58	1.84 ± 1.46	0.38 ± 0.63	0.50 ± 0.58	2.05 ± 1.56	0.88 ± 1.72
Ilium & Pubis Known	0.50 ± 1.59	1.07 ± 1.64	0.83 ± 1.28	1.63 ± 2.54	1.74 ± 1.38	0.33 ± 0.61	0.49 ± 0.60	1.94 ± 1.49	0.61 ± 1.66
Ilium & Pubis Est.	0.58 ± 1.67	1.16 ± 1.78	1.00 ± 1.24	1.86 ± 2.62	1.75 ± 1.48	0.34 ± 0.57	0.49 ± 0.57	1.96 ± 1.56	0.73 ± 1.82

Means and standard deviations of the fragment pose errors and lateral center edge (LCE) angle errors for each simulation study. The errors are organized by the type of fragment shape used for pose estimation, the total rotation and translation error magnitudes, the decompositions of the errors about each anatomical axis, and the LCE angle. Small LCE errors for all fragment shapes indicate that any shape is reliable for reporting lateral coverage of the femoral head.

TABLE II

Simulation Study Statistical Test Results

Fragment Shapes		p-Value		
1	2	Rot.	Trans.	LCE
All Cuts Known	No Cuts Known	$< 10^{-6}$	$< 10^{-6}$	$< 10^{-6}$
All Cuts Known	Ilium & Pubis Known	$< 10^{-6}$	$< 10^{-6}$	$< 10^{-6}$
All Cuts Known	Ilium & Pubis Est.	$< 10^{-6}$	$< 10^{-6}$	$< 10^{-6}$
Ilium & Pubis Known	Ilium & Pubis Est.	$< 10^{-6}$	0.74	$< 10^{-6}$
Ilium & Pubis Est.	No Cuts Known	$< 10^{-6}$	$< 10^{-6}$	$< 10^{-6}$

Results of the Mann-Whitney U-Tests on fragment pose errors for the simulation studies. The total rotation and translation magnitudes of the pose errors, along with the error of the lateral center edge (LCE) angle are examined. The null hypothesis indicates that both sets of errors are drawn from the same distribution with identical medians and the alternative hypothesis indicates that the errors are drawn from different distributions with the second distribution (fragment shape with label 2) having larger median. Applying a threshold of 0.005 to the above p-values results in a rejection of the null-hypothesis for all but one of the tests. This implies that errors associated with fragment shapes incorporating less information of the true cuts are statistically larger than the errors associated with cuts that more closely match intraoperative cuts, except when comparing translation errors when ilium and pubis cuts are known versus estimated.

TABLE III

Cadaver Experiment Fragment Pose and Lateral Center Edge Angle Errors

Study	Rotation Errors (°)				Translation Error (mm)				LCE (°)
	AP	LR	IS	Total	AP	LR	IS	Total	
Man. Seg.	0.51 ± 0.25	1.76 ± 0.93	1.02 ± 0.51	2.22 ± 0.73	1.22 ± 0.50	1.03 ± 0.79	1.29 ± 1.01	2.18 ± 1.10	1.08 ± 0.52
Planned	1.17 ± 1.10	2.59 ± 1.70	1.89 ± 1.48	3.77 ± 1.88	1.49 ± 1.40	1.06 ± 0.90	1.16 ± 0.99	2.48 ± 1.51	1.80 ± 1.39
Estimated	0.98 ± 0.80	2.62 ± 1.38	1.55 ± 1.21	3.46 ± 1.47	0.98 ± 0.83	1.07 ± 0.83	1.15 ± 0.90	2.17 ± 0.94	1.80 ± 1.18

Means and standard deviations of fragment pose and lateral center edge angle (LCE) errors for the cadaver surgeries. For each type of fragment shape used during registration, the total rotation and translation error magnitudes are reported, along with the errors about each anatomical axis and the LCE angle error. The manual segmented fragment shape, planned fragment shape, and estimated fragment shapes roughly correspond to the simulation cases of a fragment shape with all cuts known, a preoperatively planned fragment shape with no cuts known, and a shape with ilium and pubis cuts estimated, respectively, as shown in Table I.

TABLE IV

Cadaver Experiments Statistical Test Results

Fragment Shapes		p-Value		
1	2	Rot.	Trans.	LCE
Man. Seg.	Planned	0.0194	0.3798	0.1194
Man. Seg.	Ilium & Pubis Cuts Est.	0.0187	0.5208	0.0741
Ilium & Pubis Cuts Est.	Planned	0.1678	0.1023	0.7167

Results of the Mann-Whitney U-Tests on fragment pose and lateral center edge (LCE) angle errors for the cadaver experiments. See Table II for a discussion of the hypotheses. Using a threshold of 0.005, no significant difference is reported for any of the errors and for each comparison of shapes.