Published in final edited form as:

J Neurosci Methods. 2020 January 01; 329: 108461. doi:10.1016/j.jneumeth.2019.108461.

Implantable Wireless Device for Study of Entrapment Neuropathy

Ronit Malka, MD^{a,b,1}, Diego L Guarin, PhD^b, Suresh Mohan, MD^b, Iván Coto Hernández, PhD^b, Pavel Gorelik, MS^c, Ofer Mazor, PhD^c, Tessa Hadlock, MD^b, Nate Jowett, MD^b

^aHealth Science and Technology Division, Harvard Medical School/Massachusetts Institute of Technology, Boston, MA

^bSurgical Photonics and Engineering Laboratory, Department of Otolaryngology – Head and Neck Surgery, Massachusetts Eye and Ear and Harvard Medical School, Boston, MA

^cResearch Instrumentation Core, Department of Neurobiology, Harvard Medical School, Boston, MA

Abstract

Background—Disease processes causing increased neural compartment pressure may induce transient or permanent neural dysfunction. Surgical decompression can prevent and reverse such nerve damage. Owing to insufficient evidence from controlled studies, the efficacy and optimal timing of decompression surgery remains poorly characterized for several entrapment syndromes.

New Method—We describe the design, manufacture, and validation of a device for study of entrapment neuropathy in a small animal model. This device applies graded extrinsic pressure to a peripheral nerve and wirelessly transmits applied pressure level in real-time. We implanted the device in rats applying low (under 100 mmHg), intermediate (200–300 mmHg) and high (above 300 mmHg) pressures to induce entrapment neuropathy of the facial nerve to mimic Bell's palsy. Facial nerve function was quantitatively assessed by tracking whisker displacements before, during, and after compression.

Results—At low pressure, no functional loss was observed. At intermediate pressure, partial functional loss developed with return of normal function several days after decompression. High pressure demonstrated complete functional loss with incomplete recovery following decompression. Histology demonstrated uninjured, Sunderland grade III, and Sunderland grade V injury in nerves exposed to low, medium, and high pressure, respectively.

Comparison with Existing Methods—Existing animal models of entrapment neuropathy are limited by inability to measure and titrate applied pressure over time.

¹Corresponding author permanent address: Department of Otolaryngology, Brooke Army Medical Center, 3551 Roger Brooke Dr., San Antonio, TX, 78234, USA. ronit.e.malka.mil@mail.mil.

Publisher's Disclaimer: This is a PDF file of an unedited manuscript that has been accepted for publication. As a service to our customers we are providing this early version of the manuscript. The manuscript will undergo copyediting, typesetting, and review of the resulting proof before it is published in its final form. Please note that during the production process errors may be discovered which could affect the content, and all legal disclaimers that apply to the journal pertain.

Conclusions—Described is a miniaturized, wireless, fully implantable device for study of entrapment neuropathy in a murine model, and may be broadly employed to induce various degrees of neural dysfunction and functional recovery in live animal models.

Graphical Abstract

Keywords

entrapment neuropathy; compression neuropathy; Bell's Palsy; surgical decompression; bioimplantable device

1. Introduction

Entrapment neuropathies are vexing clinical conditions affecting over 40 million people worldwide yearly [1]. These conditions may yield chronic pain, numbness, muscle weakness, and spasticity [2, 3]. Many cranial and peripheral nerves and their roots are susceptible to entrapment, including those innervating regions of the head, neck, trunk, and extremities [4]. The pathophysiology of dysfunction in entrapment neuropathy arises from increased pressure within a neural compartment secondary to intrinsic edema or extrinsic compression [5]. Acute entrapment results in neural conduction block and focal demyelination; symptoms are fully reversible if neural compartment pressure normalizes before axonal degeneration occurs. With long-term compression and high compartment pressures, axonotmesis and Wallerian degeneration occur [6]. Though peripheral and cranial nerve axons are capable of regeneration following axonotmesis, misdirection of extending axons to distal target tissues may occur, resulting in permanent neural dysfunction [7–12]. Management of entrapment neuropathy may depend on several factors, including symptom severity and chronicity, results of electrophysiology studies, and patient co-morbidities. Surgical decompression of neural compartments has proven effective for management of several entrapment neuropathies, including carpal [13] and cubital tunnel syndromes [14]. However, the optimal timing of surgical decompression varies widely among different entrapment neuropathies, and is often not at the time of diagnosis but rather after more conservative medical management has failed or poor long-term functional prognosis has been established.

Bell's palsy (BP) is the most common acute cranial neuropathy. It is believed to result from entrapment of the facial nerve within its bony canal secondary to inflammatory edema following viral reactivation [15, 16]. In two prospective surgical studies, Fisch and Gantz demonstrated clinically-significant improvement in long-term facial functional recovery among patients presenting with severe BP who underwent decompression surgery within

two weeks of palsy onset [17, 18]. Though several subsequent studies have reported similar benefits of acute facial nerve decompression in the setting of severe BP [19–24], the procedure remains controversial [25]. Heretofore, clinical studies of facial nerve decompression have been limited by patient selection bias, inconsistent surgical technique, and suboptimal evaluation of long-term recovery. Due to its close proximity to the inner ear, surgical decompression of the facial nerve carries risk of permanent hearing loss. Owing to procedural risk and lack of evidence from prospective controlled studies to support its effectiveness, clinical practice guidelines remain neutral on facial nerve decompression for acute BP [26]. A controlled animal model of facial nerve decompression could inform clinical management of acute facial palsy.

Prior work has employed animal models for study of entrapment and related neuropathies, including Bell's palsy, carpal tunnel syndrome, spinal stenosis, sciatica, and diabetic neuropathy [27–30]. Entrapment of peripheral nerves in animal models has been induced using tourniquets, inflatable cuffs, spring-loaded clamps, implanted silicone tubing, and adhesive molds [27–39]. However, precise pressure titration is an important feature that permits modeling of various degrees of severity of nerve entrapment, while continual measurement enables monitoring of pressure changes arising from biological responses to extrinsic compression, including demyelination, fluid shifts, and inflammatory cell migration. Thus, the study of nerve entrapment in animal models to date has been limited by the inability to measure and titrate applied pressure over the long term *in vivo*. Herein, we present the development and validation of an implantable device for controlled study of entrapment neuropathy. The device allows for graded extrinsic compression of surgically accessible nerves in small animals, concurrent with wireless monitoring of applied pressure for up to several months. Devices were implanted in live rats for compression of facial nerve branches to simulate BP. Facial nerve function was longitudinally assessed by quantification of whisking activity and correlated to applied pressure levels. After sacrifice, facial nerve sections were examined to evaluate for histological evidence of Wallerian degeneration and axonal regeneration.

2. Materials and Methods

2.1 Design

We designed a biocompatible, sterilizable, and reusable implant that communicates wirelessly with an external receiver via radiofrequency. The devices wirelessly relays *in vivo* temperature and pressure data of an enclosed hydraulic system. Data may be stored in a datalogger for offline analysis or transmitted to a computer for online analysis (Figure 1). The specifications of the implant and receiver design are described in greater detail in the supplementary material.

2.1.1 Implant—The implant consists of a microcontroller, wireless transmitter, pressure sensor, and battery integrated on a printed circuit board (PCB), connected to a fluid injection port and pressurizable cuff as shown in Figure 2. In this example, the cuff is secured around a peripheral nerve. Pressure within the system is titrated by injection of liquid through a biocompatible injection port and is monitored by a pressure sensor. The entire assembly

measures 25mm × 20mm × 15mm and weighs under 15g after biocompatible silicone embedding.

2.1.2 Pressure Validation—As the device is required to hold and detect changes in pressure within the hydraulic system over time, validation of dynamic and static pressure measurements was performed. The precision of the device's sensor pressure measurement was quantified by comparing the difference in sensor measurements and analog pressure regulator readings when connected in parallel to compressed air. Dynamic pressure change was additionally evaluated by measuring the time the sensor measurements took to equilibrate to analog readings.

2.1.2 Receiver—A standalone external microcontroller acts as a datalogging receiver for data broadcasted by the implant. The receiver can be configured as a portable device that accompanies the animal's cage and stores data in an external drive for long-term experiments.

2.2 Fabrication

2.2.1 Printed Circuit Board—A customized PCB was created to minimize device footprint while integrating the microcontroller, sensor, and coin battery. The board was designed in free circuit board design software (EAGLE PCB Design Software, Autodesk Inc., San Rafael, CA) and printed using a commercial vendor (Sunstone Circuits, Mulino, OR). All hardware on the board were commercially available and reusable across multiple experiments. PCB design and component list are provided in the supplementary material.

2.2.2 Cuff and Injection Port—The occlusion cuff employed for nerve compression (Vascular Occlude Cuff, Access Technologies, Skokie, IL) was connected to the pressure sensor and a subcutaneous injection port (Rat-o-port, Access Technologies, Skokie, IL) using inelastic biocompatible hollow tubing to enable titration and monitoring of system pressure *in vivo*. A custom coupler for connecting hydraulic system components was designed in SolidWorks (SolidWorks 3D CAD, Dassault Systèmes, Waltham, MA) and 3D printed. Coupler design is available as supplementary material.

The occlusion cuff employed is air and water permeable, allowing for osmotic gradients following injection of non-permeable glycerol into the system to dictate pressure changes. The gradual shift of water into the system and resultant increase in hydrostatic pressure models that postulated to occur in entrapment neuropathies secondary to inflammation in a confined space [40].

2.2.3 Silicone Coating—The device body was coated in a thin layer of epoxy (Loctite QuickSet Epoxy, Loctite Industrial, Westlake, OH) to waterproof the board. It was then coated with 2 mm of medical-grade silicone (A-103 Medical Grade Elastomer, Factor II Inc., Lakeside, AZ) to protect the device from corrosion and render the implant biocompatible. A deconstructed diagram of the implant, as well as the final device, is shown in Figure 3.

2.2.4 Sterilization—The silicone-coated device was sterilized with ethylene oxide (7 hours at 46°C) and degassed under vacuum for 5 days prior to implantation.

2.3 Surgical Procedures

Experiments were performed in accordance with the National Institutes of Health Guide for the Care and Use of Laboratory Animals, with Institutional Animal Care and Use Committee approval.

2.3.1 Head fixation—Three female Lewis rats, aged 65–90 days, were anesthetized with isoflurane (3.5% during induction, 2% during operation) until unresponsive to toe-pinch. Meloxicam (1 mg/kg) and buprenorphine (0.1 mg/kg) were injected subcutaneously. A 2 cm incision was made midline over the calvarium, and bone cleared of overlying tissues. A percutaneous titanium head-fixation device was affixed to the calvarium using five titanium screws as previously described [41] and the incision closed primarily using 5–0 polyglactin 910 suture (Vicryl Rapide, Ethicon US, LLC., Bridgewater, NJ). Topical bacitracin zinc antibiotic ointment was applied to the wound, and animals recovered from general anesthesia and returned to their cages. Meloxicam (1 mg/kg/day) was administered subcutaneously for 72 hours post-operatively.

2.4 Behavioral Conditioning and Facial Nerve Functional Testing

Two weeks following head fixation device implantation, animals underwent daily conditioning to restraint in a cloth sack until able to tolerate twenty minutes of continuous restraint without distress. Animals were then conditioned to head and body restraint for high-frame rate video capture of whisker movements. Facial nerve function was assessed by comparison of mean amplitude of whisking movements between sides. The design, setup, and signal processing underlying this functional assessment system is explained in greater detail in the supplementary section.

2.3.2 Implantation—Device hydraulic systems were filled with normal saline to neutral pressure (0 mmHg above atmospheric). Once conditioned, animals were again placed under isoflurane general anesthesia as above. After administration of analgesia, a 1.5 cm transverse incision was made in the left cheek, and skin-muscle flaps elevated atop the buccal and marginal mandibular branches of the left facial nerve (which carry motor efferents innervating the whisker pad musculature). The nerve branches were then circumferentially dissected in meticulous fashion without transection. Next, a 3 cm incision was made over the scapular portion of the dorsum, and a subcutaneous pocket created to receive the body of the device and injection port. The occlusion cuff with attached tubing was then tunneled subcutaneously between the dorsal and left facial incisions. The cuff was then secured around the dissected buccal and marginal mandibular branches with a 4–0 polypropylene suture (Figure 4b). The device body and injection port were then secured to the fascia of the latissimus dorsi muscles using 5–0 absorbable polygalactin 910 sutures, and skin incisions closed in a single layer using identical suture (Figure 4c). Animals were recovered from general anesthesia and perioperative care delivered as described above. Awake whisking function was documented by high-frame rate video capture immediately prior to surgery and

again 3 days post-operatively to ensure no iatrogenic neural insult occurred secondary to device implant.

2.5 Modeling Nerve Entrapment

Three days following implantation, a 50% glycerol solution (0.01, 0.02 or 0.03 mL) was injected into device ports of the three animals. Injection volumes employed avoided pressure spikes; the colloid osmotic pressure of the injected glycerol induced gradual pressure rise by the movement of water across the semipermeable membranes of the occlusion cuffs. Additional small volumes (0.01 to 0.02 mL of 50%–90% glycerol solution) were again injected on post-operative day (POD) # 3 and 4 until gauge pressure ranges under 100 mmHg ('low-pressure'), between 200 and 300 mmHg ('medium-pressure'), and between 300 and 500 mmHg ('high-pressure') were achieved.

2.7 Device Explant

Nerve decompression via device explant was planned after five days of compression for low, medium, and high-pressure groups. This duration is clinically relevant, as it falls within the 3–14 day window wherein facial nerve decompression is currently considered in humans with acute Bell's palsy. Decompression was achieved by removal of fluid from the injection ports, followed by immediate removal of the devices under general anesthesia to simulate decompression surgery. The occlusion cuff was meticulously freed from left facial nerve branches, and the device body and injection port with attached tubing extirpated through the dorsal incision. Incisions were closed in a single layer, and post-operative care delivered as described above. Whisker function was assessed daily for 7 days following decompression, followed by weekly assessment until a recovery plateau was noted for up to 12 weeks.

2.8 Nerve Tissue Processing

Upon study termination, animals were humanely euthanized in accordance with institutional animal care protocols. Corpses were immediately transcardially perfused with 4% phosphate-buffered paraformaldehyde (PFA). Left facial nerves were harvested and post-fixed by immersion in 4% PFA for 48h, followed by cryoprotection in 30% sucrose for 24h. Nerves were sharply divided into proximal, lesional (in the area of cuff compression), and distal segments. Axial cryosections at 1 μ m were obtained and stained with FluoroMyelin® Green (1:300, Molecular Probes, Eugene, OR), similar to previously described methods [42, 43].

2.9 Tissue Imaging

Fluorescent images of nerve cross-sections were collected with a confocal microscope (Leica Microsystems SP8, Wetzlar, Germany) equipped with a resonant scanner and 64X (1.3 NA) oil-immersion objective lens. FluoroMyelin® Green was excited with a 488 nm Argon laser, and fluorescent signal collected at 550/50 nm using Leica HyD hybrid detectors controlled with the Leica Application Suite X software. Proprietary deconvolution software (Lightning, Leica Microsystems) was uniformly employed for image enhancement. Image brightness and contrast were uniformly adjusted for enhanced visualization using ImageJ software (Fiji Distribution, Version 1.52e) [44, 45].

2.10 Image Processing

High-powered fields (63X) were obtained for nerve histomorphometry. Mean myelinated axon count and diameter were quantified using a trainable random forest machine learning algorithm in commercial software (DRVISION Technologies, Bellevue, WA).

3. Results

3.1 Pressure Validation

Figure 5a demonstrates measurements obtained from analog device sensors during rapid pressurization and depressurization. The implantable sensor captured rapid pressure changes and demonstrated near perfect agreement with analog gauge pressure within seconds. Figure 5b and c demonstrate measurements obtained from analog and implantable sensors over a long-term measurement, with an initial injection pressure of 300 mmHg. The device maintained and wirelessly transmitted pressure measurements in close agreement with analog measurements (mean difference of 19 mmHg, SD, 6.9 mmHg) over six days.

3.2 Device Lifespan and Implantation

Herein, devices were well tolerated during the eight day implantation with lossless data transmission prior to surgical explant (Figure 4c). No signs of device extrusion occurred over the course of the study. As described in Supplementary Materials, in a second subset of animals, devices were well tolerated *in vivo* and provided continuous transmission of pressure and temperature readings once per hour for over 10 weeks using a 12 mm coin battery (Eunicell™ CR1220, 40mAh, East Technology Co., Shenzhen District, China). Use of a larger battery (Eunicell™ CR2032, 235mAh, East Technology Co., Shenzhen District, China) in the devices allowed for continuous operation of the devices for more than five months. These durations of neural compression are sufficient for modeling various nerve entrapment syndromes [31, 40]. Reuse of the devices is possible by careful removal of the silicone and resin coatings, replacement of the single-use battery, followed by recoating with resin and silicone and re-sterilization. Future designs will seek to incorporate rechargeable batteries paired with externalized waterproof pin connectors for *ex vivo* recharging of the devices, or the addition of inductive or resonant inductive wireless charging components for *in vivo* recharging.

3.3 Pressure Threshold Determination

As demonstrated in Figure 6, low-pressure facial nerve compression (less than 100 mmHg) resulted in no detectable impairment in whisking function following 5 days of compression. Rapid and complete loss of whisking activity was noted for the rat whose facial nerve was compressed in the range of 300 mmHg - 500 mmHg within 24 hours; only partial recovery of whisking function was noted 12 weeks following decompression. Gradual and incomplete loss of whisking function occurred for the rat whose facial nerve was compressed between 200–300 mmHg over the course of 72 hours; subsequent decompression resulted in complete recovery of normal whisking activity within 48 hours.

3.4 Histology

At time of device removal, healthy appearing nerve was grossly noted within the cuff inflated to a low gauge pressure (under 100 mm Hg) (Figure 7a), while some thinning and partial clarification was noted for facial nerve branches compressed at medium pressure (200 mmHg – 300 mmHg) (Figure 7b-c). Nerve branches compressed between 300 mmHg and 500 mmHg demonstrated extreme thinning and discontinuity, believed secondary to ischemic necrosis from prolonged high-pressure compression.

Histologic evaluation of harvested facial nerves 12 weeks after nerve decompression and device explant demonstrated varying degrees of neural injury and regeneration, correlating with the level of applied pressure (Figure 8). In nerve compressed at low-pressure, myelin, thickness and axon caliber appeared normal across proximal, lesional, and distal segments. Nerve compressed at medium pressure demonstrated sporadic axonotmesis and occasional endoneurial disruption, suggestive of a Sunderland grade III neural injury [46]. As demonstrated in Figure 8, partial axonotmesis was evidenced by occasional Wallerian degeneration in lesional segments and the putative presence of ‘regenerating units’ in the distal segment of nerve compressed at medium pressure. Regenerating units comprise multiple tightly-spaced small-diameter and thinly-myelinated ‘daughter’ axons, which sprout from a single large myelinated axon following axotomy [47]. Regenerating units are bounded by connective tissues forming the endoneurial tubules, having outer diameters similar to those of undamaged large myelinated fibers. Distal sections of nerve compressed at high pressure demonstrated extensive regeneration consistent with a prior Sunderland grade V injury, as evidenced by the complete replacement of normal caliber myelinated axons in the distal segment presence with regenerating units. Quantitative nerve histomorphometry demonstrated minimal change in myelinated axon caliber and counts between proximal and distal segments for nerve compressed at low pressure. At medium pressure, a moderate decrease in mean myelinated nerve caliber and increase in axon count was noted in distal sections as compared to proximal sections, consistent with Grade III Sunderland injury. At high pressure, a larger decrease in mean myelinated axon caliber and larger increase in axon counts was observed for lesional and distal as compared to proximal sections, consistent with neurotmesis and subsequent spontaneous nerve regeneration across the gap.

4. Discussion

4.1 Summary

Described is the design, manufacture, and *ex vivo* and *in vivo* validation of a miniaturized, wireless, implantable device for study of entrapment neuropathy in a murine model. This approach is the first to enable titration and continuous monitoring of applied pressure to a peripheral nerve over the long term. Use of this device may be paired with functional assessments to correlate the degree and duration of nerve compression with nerve function in a variety of animal models. Herein, the device was employed to induce entrapment neuropathy of facial nerve branches controlling whisking function in the rat. Functional loss induced in this model spanned mild neurapraxia to complete paralysis, demonstrating the utility of this approach to model a broad range of neural entrapment severity. This device

could be employed to study the long-term effects of extrinsic compression and subsequent decompression of other accessible nerves in small animals. This approach provides opportunity to investigate the mechanics of entrapment neuropathy in greater depth than previously possible, carrying potential to enhance understanding and management of this devastating disease spectrum.

4.2 Limitations

Though the device and rodent model employed herein offers a cost-effective means for controlled study of entrapment neuropathy, caution is warranted with clinical translation of future research findings. Nerve entrapment syndromes may involve additional pathophysiologic mechanisms beyond the simple mechanical compression modeled herein, including altered local microvascular environment or viral reactivation [4, 5, 31, 40, 48, 49]. Further, rodents are known to undergo Wallerian degeneration at rates at least two-fold higher than humans, while exhibiting exceptionally robust neural regeneration [6, 50]. Nevertheless, we believe this model is likely to yield important insights into the role and timing of surgical decompression of peripheral nerves.

This model requires placement of a silicone cuff around a peripheral nerve. Neural dysfunction might occur in the absence of fluid injection into the hydraulic system via foreign body response, iatrogenic nerve injury at time of device implantation, or delayed traction injury during long-term implantation. Though capsule formation around the devices occurred in all animals, minimal capsule growth into the lumen of the occlusion cuff was noted after several days of implantation. Neuropraxic injury was avoided with use of meticulous surgical technique at time of device implantation, and traction injury was avoided by securing of the occlusion cuffs to the underlying masseter muscle using surgical suture. Though device extrusion may occur, it was not observed in this series of animals.

4.3 Future Directions

Though modelling of entrapment neuropathy of small peripheral nerves (1–2mm diameter) in a rodent model has been demonstrated herein, this device could be fitted with a larger cuff for characterizing entrapment of larger nerves in higher order mammals. This device and its microcontroller backbone may be repurposed for cost-effective *in vivo* study of a wide range of disease processes. By placing the cuff around a vessel as opposed to a nerve, this device could even be employed for study of a variety of ischemic or vaso-occlusive disorders, including cerebrovascular accidents, hemodynamic shock, and free flap thrombosis. By virtue of its wireless transmission capabilities and open-source design, this device could be readily redesigned with alternative sensors for monitoring of other physiologic signals. For example, the pressure transducer and hydraulic system could be replaced with a pH sensor for continuous monitoring of acid-base shifts in tissues, or with an accelerometer for monitoring of gait and vestibular function. Device lifespan can be additionally extended via incorporation of components enabling wireless power transfer. Depending on the application, device power consumption could be optimized by adjusting the sampling rate of data acquisition and transmission.

Though a murine model of BP has been previously characterized [51], no controlled animal model has been developed to characterize the impact of facial nerve decompression surgery. Future work will employ the approach described herein to characterize long-term facial functional recovery following facial nerve compression at various pressure levels and durations in a small animal model.

Supplementary Material

Refer to Web version on PubMed Central for supplementary material.

Acknowledgements

This work was supported by the NIH NINDS (R01NS071067); a philanthropic gift from the Berthiaume Family Foundation; and the American Academy of Otolaryngology Head and Neck Surgery Foundation (CORE Research Grant). The Research Instrumentation Core Facility is funded in part by a National Eye Institute P30 Core Grant for Vision Research (EY012196) and by the Bertarelli Program in Translational Neuroscience and Neuroengineering at Harvard Medical School. Declarations of interest: none.

Abbreviations

BP	Bell's Palsy
PCB	Printed Circuit Board
PFA	phosphate-buffered paraformaldehyde
POD	postoperative day

References

- [1]. Latinovic R, Gulliford MC, and Hughes RAC, "Incidence of common compressive neuropathies in primary care," (in eng), *Journal of neurology, neurosurgery, and psychiatry*, vol. 77, no. 2, pp. 263–265, 2006, doi: 10.1136/jnnp.2005.066696.
- [2]. Gilden DH, "Clinical Practice: Bell's Palsy," *New England Journal of Medicine*, vol. 351, pp. 1323–1331, 2004. [PubMed: 15385659]
- [3]. Latinovic R, Gulliford MC, & Hughes RAC, "Incidence of common compressive neuropathies in primary care," *Journal of Neurology, Neurosurgery, and Psychiatry*, vol. 77, pp. 263–265, 2006.
- [4]. Smith BE, "Focal and entrapment neuropathies," in *Handbook of Clinical Neurology; Diabetes and the nervous system*, vol. 126, no. 3), 2014, ch. 3.
- [5]. Arnold WDEBH, "Entrapment Neuropathies," *Neurologic Clinics*, vol. 31, no. 2, pp. 405–424, 2013. [PubMed: 23642716]
- [6]. Gaudet AD, Popovich PG, and Ramer MS, "Wallerian degeneration: Gaining perspective on inflammatory events after peripheral injury," *Journal of Neuroinflammation*, vol. 8, no. 1, p. 1, 2011. [PubMed: 21208419]
- [7]. Valls-Sole J, Tolosa ES, and Pujol M, "Myokymic discharges and enhanced facial nerve reflex responses after recovery from idiopathic facial palsy," (in eng), *Muscle & nerve*, vol. 15, no. 1, pp. 37–42, 1 1992, doi: 10.1002/mus.880150107. [PubMed: 1732759]
- [8]. Montserrat L and Benito M, "Facial synkinesis and aberrant regeneration of facial nerve," (in eng), *Advances in neurology*, vol. 49, pp. 211–24, 1988. [PubMed: 3278542]
- [9]. Sumner AJ, "Aberrant reinnervation," *Muscle & nerve*, vol. 13, no. 9, pp. 801–3, 9 1990, doi: 10.1002/mus.880130905. [PubMed: 2233866]

- [10]. Weber ED and Newman SA, "Aberrant regeneration of the oculomotor nerve: implications for neurosurgeons," (in eng), *Neurosurg Focus*, vol. 23, no. 5, p. E14, 2007, doi: 10.3171/foc-07/11/e14.
- [11]. Teasdall RD and Salman SD, "Mass facial movements: electromyographic evidence for misdirection," (in eng), *Neurology*, vol. 21, no. 6, pp. 652–4, 6 1971, doi: 10.1212/wnl.21.6.652. [PubMed: 5105417]
- [12]. de Ruiters GCW et al., "Misdirection of regenerating motor axons after nerve injury and repair in the rat sciatic nerve model," (in eng), *Experimental neurology*, vol. 211, no. 2, pp. 339–350, 2008, doi: 10.1016/j.expneurol.2007.12.023. [PubMed: 18448099]
- [13]. Phalen GS and Kendrick J. L., "COMPRESSION NEUROPATHY OF THE MEDIAN NERVE IN THE CARPAL TUNNEL," *JAMA*, vol. 164, no. 5, pp. 524–530, 1957, doi: 10.1001/jama.1957.02980050014005.
- [14]. Adelaar RS, Foster WC, and McDowell C, "The treatment of the cubital tunnel syndrome," *The Journal of hand surgery*, vol. 9, no. 1, pp. 90–95, 1984/01/01/ 1984, doi: 10.1016/S0363-5023(84)80193-8.
- [15]. Fisch U and Esslen E, "Total intratemporal exposure of the facial nerve. Pathologic findings in Bell's palsy," *Arch Otolaryngol*, vol. 95, no. 4, pp. 335–41, 4 1972. [PubMed: 5018255]
- [16]. Murakami S, Mizobuchi M, Nakashiro Y, Doi T, Hato N, and Yanagihara N, "Bell palsy and herpes simplex virus: identification of viral DNA in endoneurial fluid and muscle," (in English), *Annals of Internal Medicine*, Research Support, Non-U.S. Gov't vol. 124, no. 1 Pt 1, pp. 27–30, 1 1 1996. [PubMed: 7503474]
- [17]. Fisch U, "Surgery for Bell's Palsy," *Arch Otolaryngol*, vol. 107, 1981.
- [18]. Gantz BJ, Rubinstein JT, Gidley P, and Woodworth GG, "Surgical Management of Bell's Palsy," *The Laryngoscope*, vol. 109, 1999.
- [19]. Liu S et al., "Study on relationship between operation timing and clinical prognosis of cases with Bell palsy," *Journal Of Clinical Otorhinolaryngology, Head, & Neck Surgery*, vol. 27, no. 13, pp. 698–700, 2013.
- [20]. Shapira Y, Migirov L, and Kronenberg J, "Facial nerve decompression," *Harefuah*, vol. 145, no. 8, pp. 557–60, 632, 2006. [PubMed: 16983835]
- [21]. Wu SH, Chen X, Wang J, Liu H, Qian XZ, and Pan XL, "Subtotal facial nerve decompression in preventing further recurrence and promoting facial nerve recovery of severe idiopathic recurrent facial palsy," (in English), *European Archives of Oto-Rhino-Laryngology*, vol. 272, no. 11, pp. 3295–8, 11 2015, doi:10.1007/s00405-014-2991-9. [PubMed: 24619203]
- [22]. Hato N et al., "Facial nerve decompression surgery using bFGF-impregnated biodegradable gelatin hydrogel in patients with Bell palsy," (in English), *Otolaryngology - Head & Neck Surgery*, vol. 146, no. 4, pp. 641–6, 4 2012, doi:10.1177/0194599811431661. [PubMed: 22166965]
- [23]. Cannon RB, Gurgel RK, Warren FM, and Shelton C, "Facial nerve outcomes after middle fossa decompression for Bell's palsy," (in English), *Otol Neurotol*, vol. 36, no. 3, pp. 513–8, 3 2015, doi:10.1097/MAO.0000000000000513. [PubMed: 25058839]
- [24]. Bodenez C, Bernat I, Willer JC, Barre P, Lamas G, and Tankere F, "Facial nerve decompression for idiopathic Bell's palsy: report of 13 cases and literature review," (in English), *J Laryngol Otol*, Review vol. 124, no. 3, pp. 272–8, 3 2010 doi:10.1017/S0022215109991265.
- [25]. Smouha E, Toh E, and Schaitkin BM, "Surgical treatment of Bell's palsy: current attitudes," (in English), *Laryngoscope*, vol. 121, no. 9, pp. 1965–70, 9 2011 doi:10.1002/lary.21906.
- [26]. Baugh RF et al., "Clinical practice guideline: Bell's palsy," (in eng), *Otolaryngol Head Neck Surg*, vol. 149, no. 3 Suppl, pp. S1–27, 11 2013, doi: 10.1177/0194599813505967.
- [27]. Barac S, Jiga LP, Barac B, Honioiu T, Dellon AL, and Ionac M, "Hindpaw withdrawal from a painful thermal stimulus after sciatic nerve compression and decompression in the diabetic rat," *J Reconstr Microsurg*, vol. 29, no. 1, pp. 63–66, 2013. [PubMed: 23161393]
- [28]. Cheung PWH, Hu Y, and Cheung JPY, "Novel compression rat model for developmental spinal stenosis," *J Orthop Res*, vol. 37, no. 5, pp. 1090–1100, 5 2019, doi: 10.1002/jor.24221. [PubMed: 30644588]

- [29]. Gupta R, Rowshan K, Chao T, Mozaffar T, and Steward O, "Chronic nerve compression induces local demyelination and remyelination in a rat model of carpal tunnel syndrome," *Experimental Neurology*, vol. 187, no. 2, pp. 500–508, 2004. [PubMed: 15144876]
- [30]. Hazama H, Tamaki H, Nomura A, Furukawa Y, and Uetsuka H, "Compression of the facial nerve," *Archives of Otolaryngology*, vol. 95, no. 4, pp. 346–349, 1972. [PubMed: 5018257]
- [31]. Rempel DM, Dahlin L, and Lundborg G, "Pathophysiology of Nerve Compression Syndromes: Response of peripheral nerves to loading," *The Journal of Bone and Joint Surgery*, vol. 81-A, no. 11, 1999.
- [32]. Dellon AL and Mackinnon S, "Chronic nerve compression model for the double crush hypothesis," *Ann Plast Surg*, vol. 26, no. 3, pp. 259–264, 1991. [PubMed: 2029136]
- [33]. Mackinnon S, Dellon AL, Hudson AR, and Hunter DA, "A primate model for chronic nerve compression," *J Reconstr Microsurg*, vol. 1, no. 3, pp. 185–195, 1985. [PubMed: 4057158]
- [34]. O'Brien JP, Mackinnon S, MacLean AR, Hudson AR, Dellon AL, and Hunter DA, "A model of chronic nerve compression in the rat," *Ann Plast Surg*, vol. 19, no. 5, pp. 430–435, 1987. [PubMed: 3688790]
- [35]. Liu Z, Chen Z, and Chen J, "A novel chronic nerve compression model in the rat," *Neural Regeneration Research*, vol. 13, no. 8, pp. 1477–1485, 2018. [PubMed: 30106062]
- [36]. Chao T, Pham K, Steward O, and Gupta R, "Chronic nerve compression injury induces a phenotypic switch of neurons within the dorsal root ganglia," *J Comp Neurol*, vol. 506, no. 2, pp. 180–193, 1 10 2008, doi: 10.1002/cne.21537. [PubMed: 18022951]
- [37]. Aguayo A, Cherunada PVN, and Midgley R, "Experimental Compression Neuropathy in the Rabbit; Histologic and electrophysiologic studies," *Arch Neurol*, vol. 24, 1971.
- [38]. Dahlin LB and McLean WG, "Effects of graded experimental compression on slow and fast axonal transport in rabbit vagus nerve," vol. 72, no. 1, pp. 19–30, 1986.
- [39]. Dyck PJ, Lais AC, Giannini C, and Engelstad JK, "Structural alterations of nerve during cuff compression," *Neurobiology*, vol. 87, pp. 9828–9832, 1990.
- [40]. Rempel DM and Diao E, "Entrapment neuropathies: pathophysiology and pathogenesis," *J Electromyogr Kinesiol*, vol. 14, no. 1, pp. 71–5, 2 2004, doi: 10.1016/j.jelekin.2003.09.009. [PubMed: 14759752]
- [41]. Heaton JT, Kowaleski JM, Bermejo R, Zeigler HP, Ahlgren DJ, and Hadlock TA, "A system for studying facial nerve function in rats through simultaneous bilateral monitoring of eyelid and whisker movements," *Journal of neuroscience methods*, vol. 171, no. 2, pp. 197–206, 6 30 2008, doi: 10.1016/j.jneumeth.2008.02.023. [PubMed: 18442856]
- [42]. Wang W, Kang S, Coto Hernandez I, and Jowett N, "A Rapid Protocol for Intraoperative Assessment of Perineal Nerve Myelinated Axon Count and Its Application to Cross-Facial Nerve Grafting," *Plast Reconstr Surg*, vol. 143, no. 3, pp. 771–778, 3 2019, doi: 10.1097/PRS.0000000000005338. [PubMed: 30601328]
- [43]. Mohan S et al., "Fluorescent Reporter Mice for Nerve Guidance Conduit Assessment: A High-Throughput in vivo Model," *Laryngoscope*, vol. 128, no. 11, pp. E386–E392, 11 2018, doi: 10.1002/lary.27439. [PubMed: 30098047]
- [44]. Schindelin J et al., "Fiji: an open-source platform for biological-image analysis," (in Eng), *Nature methods*, vol. 9, no. 7, pp. 676–82, 6 28 2012, doi: 10.1038/nmeth.2019. [PubMed: 22743772]
- [45]. Schneidei CA, Rasband WS, and Eliceiri KW, "NIH Image to ImageJ: 25 years of image analysis," *Nature methods*, vol. 9, no. 7, pp. 671–5, 7 2012. [PubMed: 22930834]
- [46]. Sunderland S, "A classification of peripheral nerve injuries producing loss of function," *Brain*, vol. 74, no. 4, pp. 491–516, 12 1951, doi: 10.1093/brain/74.4.491. [PubMed: 14895767]
- [47]. Morris JH, Hudson AR, and Weddell G, "A study of degeneration and regeneration in the divided rat sciatic nerve based on electron microscopy. II. The development of the "regenerating unit"," *Zeitschrift fur Zellforschung und mikroskopische Anatomie*, vol. 124, no. 1, pp. 103–30, 1972. [PubMed: 5011137]
- [48]. Pham K and Gupta R, "Understanding the mechanisms of entrapment neuropathies," *Neurosurgical Focus*, vol. 26, no. 2, 2009, doi: 10.3171/FOC.2009.26.2.E7).

- [49]. Wasiu Wahab K, Sanya EO, Adebayo PB, Babalola MO, & Ibraheem HG, "Carpal Tunnel Syndrome and Other Entrapment Neuropathies," *Oman Medical Journal*, vol. 32, no. 6, pp. 449–454, 2017. [PubMed: 29218119]
- [50]. Coleman MP and Freeman MR, "Wallerian degeneration, wld(s), and nmnat," *Annual review of neuroscience*, vol. 33, pp. 245–67, 2010, doi: 10.1146/annurev-neuro-060909-153248.
- [51]. Sugita T, Fujiwara Y, Murakami S, Hirata Y, Yanagihara N, & Kurata T, "Facial Nerve Paralysis Induced by Herpes Simplex Virus in Mice: An Animal Model of Acute and Transient Facial Paralysis," *Annals of Otology, Rhinology, & Laryngology*, vol. 104, no. 7, pp. 574–581, 1995.

Highlights

- Described is the design and validation of a device for study of nerve entrapment.
- The device is miniaturized, wireless, and fully implantable in small animal models.
- Neural compartment pressure is continually measured and wirelessly transmitted.
- Injection port allows for dynamic modification of pressure applied to nerve.
- Quantitative correlation of function with degree & duration of compression achieved.

Figure 1: Workflow diagram showing implanted device (A) wirelessly transmitting data collected in vivo to a datalogger (B), which then logs the data for further analysis (C).

Figure 2:

Diagram of connections between implantable injection port (1), 3D printed fitting (2) PCB including integrated sensor (diagrammed), microcontroller, and battery (not shown), and occlusion cuff (3) surrounding nerve. Fluid (blue) can flow freely between the port, sensor, and cuff in this closed hydraulic system.

Figure 3:

(a) Implant deconstructed without silicone encasement, showing (1) hydraulic cuff, (2) microcontroller (Simblee), (3) pressure sensor, (4) printed circuit board, (5) magnetic switch, and (6) subcutaneous injection port, (b) Implant with parts assembled and encased in silicone. The coin cell battery is integrated onto the opposite side of the PCB (not shown).

Figure 4:

Device implantation: (a) The occlusion cuff and device body with adjacent injection port are shown prior to securing to underlying tissues and wound closure. (b) The occlusion cuff is shown in position around the marginal and buccal branches of the facial nerve. (c) Animal 3 days post-operatively with implant body (anterior) and injection port (posterior) visible under the skin of the dorsum.

Figure 5:

Pressure measurement validation using an analog pressure gauge attached to the device (a) Simulation of rapid pressure change. Wireless device pressure (blue line, sampling rate of 500 Hz) and analog pressure (red line) measurements are shown for a single injection bolus. (b) Device pressure (sampling rate of 1/60 Hz) and analog pressure measurements (daily) over six days following a single injection bolus into the system; (c) histogram of differences between device and analog pressure readings over six days.

Figure 6:

Comparison of nerve occlusion cuff pressure and whisking function for three different pressurization levels. Shown in blue are device pressure readings (sampling rate of once per hour) for low (under 100 mmHg), medium (200–300 mmHg), and high (300–500 mmHg) pressurization targets. Shown in orange are nerve functional measurements, calculated as a percentage of awake behavioral whisking amplitude compared to the control (unoperated) side. At low compression pressure (under 100 mmHg), no functional whisking impairment is demonstrated. At high compression pressure, rapid and complete loss of whisking activity occurred, with incomplete recovery noted 12 weeks following nerve decompression. At medium compression pressure, gradual and incomplete loss of whisking activity is noted, with complete recovery of function noted within 2 days of decompression.

Figure 7:

Gross appearance of facial nerve branches at time of device explant. (a) In nerve compressed at low pressure (under 100 mmHg), normal appearing nerve is demonstrated. (b) Upon cuff release, nerve compressed at medium pressure (200 mmHg – 300 mmHg) demonstrates thinning and partial clarification (nerve borders identified by dashed lines, arrows delineate proximal to distal course).

Figure 8:

Fluorescence imaging (above) of myelinated axons in cross-sections of facial nerve branches harvested several weeks after compression at low, medium, and high-pressures, and corresponding histomorphometry (below). Following compression at low pressure (under 100 mmHg), myelinated axons in proximal, lesional, and distal segments appeared similar. Nerve branches compressed between 200 mmHg and 300 mmHg demonstrated occasional Wallerian degeneration (arrowhead) in the area of compression, and several regenerating units (yellow circles) in the distal segment. Nerve previously compressed at high pressure

(300 mmHg to 500 mmHg) demonstrated extensive regenerating units (yellow circles) in the lesional and distal segments (Scale bar: 30 μ m, Color bar: normalized intensity value (0–1). FluoroMyelin Green®, 1 μ m sections, 63X).

Author Manuscript

Author Manuscript

Author Manuscript

Author Manuscript