Published in final edited form as:

Nat Chem. 2021 May ; 13(5): 392–393. doi:10.1038/s41557-021-00692-4.

Signal transduction with a swing

Kevin J. Cash^{1,2}, Kevin W. Plaxco^{3,4}

¹Department of Chemical and Biological Engineering, Colorado School of Mines, Golden, CO, USA,

²Quantitative Biosciences and Engineering Program, Colorado School of Mines, Golden, CO, USA.

³Department of Chemistry and Biochemistry, University of California Santa Barbara, Santa Barbara, California, USA,

⁴Center for Bioengineering, University of California Santa Barbara, Santa Barbara, California, USA.

Standfirst:

The continuous monitoring of proteins is a current challenge in medical diagnostics. A new electrochemical approach aiming to address this has been described. The method uses antibodies as a recognition element to achieve the real-time measurement of proteins in saliva in the mouth.

The detection and measurement of biomarkers in patients is essential in the diagnosis and treatment of many diseases. For millennia¹ the standard of care has been removal of a sample from the patient (urine prior to several hundred years ago; blood more recently) followed by offline analysis. This *in vitro* diagnostic approach, which remains the cornerstone of clinical analysis, works well for many analytes. For some analytes, however, continuous *in vivo* monitoring would be of significant additional value for clinicians and their patients. In a recent paper in *Nature Chemistry*, a team led by Edward H. Sargent and Shana O. Kelley and co-workers have demonstrated a new approach towards this challenging goal².

The majority of successful *in vivo* biosensing approaches reported to date rely on one of two mechanisms for transducing binding events into a detectable output. In one, recognition is predicated on the specific chemical or enzymatic reactivity of target molecules. An example is the *in vivo* monitoring of glucose via its enzymatic degradation by glucose oxidase to produce electrochemically-detectable products. This method is highly successful; it is the basis for the continuous glucose monitors worn by many diabetics. But because it is reliant on the specific enzymatic reactivity of its target, this approach is not generalizable to arbitrary targets, including, importantly, protein biomarkers. The second approach, in contrast, which employs a binding induced conformational change in the recognition

kwp@chem.ucsb.edu.

Competing Interests

The author declares no competing interests.

element to produce an optical or electrochemical output, is independent of the enzymatic reactivity of its targets and thus is generalizable as long as one can engineer such a conformational change into the receptor of choice. And while this signal transduction mechanism has proven applicable to receptors comprising short polypeptides, single-domain proteins, or nucleic acid aptamers³, it has not been demonstrated to work with larger recognition elements, including antibodies, which are perhaps the most widespread and important class of bioreceptors used for diagnostics *in vitro*. Using a newly described signal transduction mechanism², however, the antibody barrier has now been breached, at least for their use in real-time detection in saliva.

The newly described signal transduction mechanism is based on a molecular pendulum that converts target binding into an electrochemical readout. The approach (Figure 1), which builds on earlier optical⁴ and electrochemical⁵ approaches, employs a short, electrode-bound, double-stranded DNA scaffold that is modified on its distal end with an antibody, (which recognizes the sensor's target), and a ferrocene redox reporter, (which generates an electrochemical signal). Changing the electrode surface to a positive electric potential (Figure 1, t_1) pulls the negatively charged DNA pendulum towards the surface, with the dynamics of this motion being controlled by the hydrodynamics of the pendulum. This arrangement means that the kinetics of the pendulum slow when the target binds to the antibody, increasing the hydrodynamic radius of the complex and causing drag (Figure 1, right). On its own, this change would be challenging to measure. The included ferrocene probe enables electrochemical readout of this kinetic event, with the time constant of the electron transfer relating monotonically to the occupancy of the antibody.

This inverted pendulum approach is general, selective, and sensitive. By swapping the target-recognizing antibody, for example, the team achieved the detection of 10 different proteins varying in size, charge, and function. The approach also achieves good limits of detection even when challenged in complex media. For example, using the protein Cardiac Troponin I (Trop I) as their target, Sargent, Kelley and co-workers demonstrated the ability to measure this clinically important marker of heart failure down to picograms per milliliter *in vitro* in a variety of biofluids, including saliva, sweat, tears, urine, and blood.

As the dynamics of the pendulum recover if the analyte is released from the antibody, the new signal transduction is reversible, suggesting that it should support continuous, real-time measurements. When taken with its selectivity and its lack of any requirement for endogenous reagents, this might imply the ability to monitor specific proteins *in vivo*. As the first exploration of this, the team placed sensors in the mouths of anesthetized mice, where they achieved the continuous monitoring of Trop I levels in saliva under the influence of intravenous injection of the target as well as injections of drugs that cause cardiac damage and resulting changes in levels of the Trop I biomarker.

The ability to use unmodified antibodies to monitor proteins continuously and in real-time in saliva represents a major step forward in our ability to monitor molecular markers of health and disease. This said, the use of saliva as a sample matrix for *quantitative* analysis (as opposed, for example, to the qualitative, yes/no detection of a pathogen or antibodies indicative of a pathogen) remains largely unproven due to concentration differences between

blood and saliva that can vary with changing saliva production⁶. Thus, a still greater advance will be the adaptation of such systems to continuous measurements in the body itself, such as intravenously or subcutaneously. Though many technologies have floundered when moving into these spaces, the ability of this new scaffold approach to measure in blood *in vitro* suggests this exciting possibility may yet come to pass.

References

1. Hippocrates, On the Prognostics, (~400 B.C.E.) translation Adams, F (1886).
2. Das J et al., Nat. Chem 10.1038/s41557-021-00644-y (2021).
3. Dunn M, Jimenez R and Chaput J Nat Rev Chem 1, 0076 (2017).
4. Langer A, et al. Nature Communications, 4, 2099 (2013).
5. Cash KJ, Ricci F and Plaxco KW J. Am. Chem. Soc, 131, 6955–6957 (2009) [PubMed: 19413316]
6. Malamud D Det. Clin. N. Am, 55, 159–178 (2011)

Figure 1.

The inverted pendulum approach converts antibody binding into an electrochemical signal in a reversible manner and without the use of reagents. It is composed of a short, double-stranded DNA scaffold that is attached to an electrode and modified on its distal end with an antibody and a redox reporter. When the electrode is held at a positive potential (time = t_1), the negatively-charged pendulum is attracted to the surface, but the kinetics with which it approaches are modulated by the hydrodynamic size of the target. Specifically, the target-free pendulum (left) migrates to the surface more rapidly than the target-bound pendulum (right), thus causing the observed electron transfer rate to depend on receptor occupancy. Created with [BioRender.com](https://www.biorender.com)